



PUBLIC NOTICE

Federal Communications Commission
445 12th St., S.W.
Washington, D.C. 20554

News Media Information 202 / 418-0500
Internet: <http://www.fcc.gov>
TTY: 1-888-835-5322

DA 03-2408

Released: July 23, 2003

**Biotronik, Inc. Waiver Request
(ET Docket No. 03-92)
Modification of *Ex Parte* Status**

On March 27, 2003, Biotronik, Inc. filed a Request for Waiver of the frequency monitoring requirements in the Medical Implant Communications Service (MICS) rules in order to receive certification for its cardiac implant devices that emit periodic scheduled transmissions. (See 47 C.F.R. § 95.628.) Biotronik further requested that such waiver also apply to other like devices that may be developed in the future with a periodic transmission function.

On May 27, 2003, the Commission issued a Public Notice, DA 03-1778, inviting comment on the waiver request. The comment period expired on June 18, 2003.

Pursuant to Section 1.1200(a) of the Commission's rules, 47 C.F.R. § 1.1200(a), this proceeding is designated as permit-but-disclose for purposes of the application of the Commission's *ex parte* rules.

For further information, please contact the Office of Engineering and Technology, Bruce Romano, (202) 418-2470, email Bruce.Romano@fcc.gov.

By the Chief, Office of Engineering and Technology