

**United States of America
FEDERAL COMMUNICATIONS COMMISSION
EXPERIMENTAL
RADIO STATION CONSTRUCTION PERMIT
AND LICENSE**

EXPERIMENTAL

(Nature of Service)

WK2XZX

(Call Sign)

XT MO

(Class of Station)

0385-EX-CN-2020

(File Number)

NAME Envoy Medical Corporation

Subject to the provisions of the Communications Act of 1934, subsequent acts, and treaties, and all regulations heretofore or hereafter made by this Commission, and further subject to the conditions and requirements set forth in this license, the licensee hereof is hereby authorized to use and operate the radio transmitting facilities hereinafter described for radio communications in accordance with the program of experimentation described by the licensee in its application for license.

Operation: In accordance with Sec. 5.3(i) of the Commission's Rules

Station Locations

- (1) MOBILE: Envoy Medical Corp Headquarters, within 160 km, centered around NL 45-05-26; WL 93-02-28
- (2) MOBILE: Envoy Medical Corp Headquarters, within 160 km, centered around NL 45-05-26; WL 93-02-28
- (3) MOBILE: Envoy Medical Corp Headquarters, within 160 km, centered around NL 45-05-26; WL 93-02-28

Frequency Information

MOBILE: Envoy Medical Corp Headquarters, within 160 km, centered around NL 45-05-26; WL 93-02-28

Frequency	Station Class	Emission Designator	Authorized Power	Frequency Tolerance (+/-)
40-80 kHz	MO	10H0N0N 20K0A1D	8 W (Output Power)	
2402-2480 MHz	MO	1M00F1D	6.3 mW (Output Power)	

This authorization effective June 12, 2020 and
will expire 3:00 A.M. EST June 01, 2022

**FEDERAL
COMMUNICATIONS
COMMISSION**



Frequency Information

MOBILE: Envoy Medical Corp Headquarters, within 160 km, centered around NL 45-05-26; WL 93-02-28

Frequency	Station Class	Emission Designator	Authorized Power	Frequency Tolerance (+/-)
40-80 kHz	MO	20K0A1D 10H0N0N	8 W (Output Power)	
2402-2480 MHz	MO	1M00F1D	6.3 mW (Output Power)	

MOBILE: Envoy Medical Corp Headquarters, within 160 km, centered around NL 45-05-26; WL 93-02-28

Frequency	Station Class	Emission Designator	Authorized Power	Frequency Tolerance (+/-)
40-80 kHz	MO	10H0N0N	8 W (Output Power)	
2402-2480 MHz	MO	1M00F1D	6.3 mW (Output Power)	

Special Conditions:

- (1) All customers participating in the experiment shall be informed that this operation is for experimental purposes only and can be cancelled at any time.
- (2) Licensee is required to ensure that trial devices are either rendered inoperable or retrieved by them from trial participants at the conclusion of the trial. Licensee is required to notify trial participants in advance that operation of the trial device is subject to this condition.
- (3) Licensee should be aware that other stations may be licensed on these frequencies and if any interference occurs, the licensee of this authorization will be subject to immediate shut down.
- (4) The occupied bandwidth of the emission shall not extend beyond the band limits set forth above.
- (5) All transmitting and/or receiving equipment used in the study shall be owned by the licensee.
- (6) Envoy Medical Corporation shall coordinate with the local Society of Broadcast Engineers frequency coordinator and the affected Fixed Microwave Service licensee prior to operation in 2450-2480 MHz band.
- (7) Non-approved equipment must be labeled.
- (8) Non-approved equipment must be retrieved.