

From: Joe Clark

To: Doug Young

Date: January 05, 2007

Subject:

Request for Info - STA File #0004-EX-ST-2007

---

Message:

Testing of Laboratory Equipment for IEC Standard IEC 61000-6-2 9TQBLE 1 PAGE 21.) has added requirements for testing in frequency range 2.0 to 2.7 GHZ at a level of 1 v/m when tested to IEC 61000-4-3.

PLEASE ADVISE IF ADDITIONAL INFORMATION IS NEEDED.

Thank you.

Joe Clark