

From: Deependra Pokharel

To: Anthony Serafini

Date: May 25, 2018

Subject: Request for Info - File # 0004-EX-CN-2018

---

Message:

As requested, a diagram titled "Orbital Facility diagram" has been attached, showing location and distances.

We would like to also add that the anechoic material of our chamber provides 30 dB of shielding effectiveness at the GPS band, and the building walls provide an additional 40-50 dB of attenuation, based on measurements taken 5/24/2018.

Thank you.