


OrthoPulse®

Biolux Technology GmbH
Neubaugasse 31
3462 Absdorf, Austria

Patented: orthopulse.com/patents

Charging Case: IP 34
OrthoPulse device: IP 37



SN (01) 09120118090067
(11) 220519
(21) 20FB0918F0E0
2022-05-19
REF OPI2-100 P/N:100-018

Charger case status

- Charging
- Error
- Charging completed

Caution/Attention:
Plug cable to standard USB receptacle
1A 5V

OrthoPulse® Status Light Guide

In hand

- Ready for treatment
- Low battery, not ready
- Treatment in progress
- Treatment completed
- Treatment paused
- Error, email

In charging case (plugged in)

- Charging, ready
- Charging, not ready
- Fully charged, ready
- Restarting
- Ready for Bluetooth*
- Syncing

support@orthopulse.com

