

## System Performance check

## Exposure Conditions

Phantom Section, TSL	Position, Test Distance [mm]	Band	Group, UID	Frequency [MHz]	Conversion Factor	TSL Conductivity [S/m]	TSL Permittivity
Flat, HSL	,		CW, 0--	13.0	16.85	0.787	53.4

## Hardware Setup

Phantom	TSL (Tissue Simulating Liquid)	Probe, Calibration Date	DAE, Calibration Date
ELI V6.0 (20deg probe tilt) - 2005	HBBL-4-250	EX3DV4 - SN7646, 2024-03-15	DAE4 Sn912, 2023-11-17

## Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 90.0	32.0 x 32.0 x 30.0
Grid Steps [mm]	10.0 x 15.0	6.0 x 6.0 x 1.5
Sensor Surface [mm]	3.0	1.4

## Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	0.059	0.057
psSAR10g [W/Kg]	0.048	0.035
Power Drift [dB]		-0.01
M2/M1 [%]		73.3
Dist 3dB Peak [mm]		15.6

