


### Impedance Measurement Plot for Body TSL







## ANNEX H PROBE CALIBRATION CERTIFICATE

**Calibration Laboratory of  
Schmid & Partner  
Engineering AG**  
Zeughausstrasse 43, 8004 Zurich, Switzerland



**S** Schweizerischer Kalibrierdienst  
**C** Service suisse d'étalonnage  
**S** Servizio svizzero di taratura  
**S** Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)  
The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

Client **MET Laboratories**

Certificate No: **EX3-3722\_Sep16**

### CALIBRATION CERTIFICATE

Object: **EX3DV4 - SN:3722**

Calibration procedure(s): **QA CAL-01.v9, QA CAL-14.v4, QA CAL-23.v5, QA CAL-25.v6  
Calibration procedure for dosimetric E-field probes**



Calibration date: **September 23, 2016**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NRP	SN: 104778	06-Apr-16 (No. 217-02288/02289)	Apr-17
Power sensor NRP-Z91	SN: 103244	06-Apr-16 (No. 217-02288)	Apr-17
Power sensor NRP-Z91	SN: 103245	06-Apr-16 (No. 217-02289)	Apr-17
Reference 20 dB Attenuator	SN: S5277 (20x)	05-Apr-16 (No. 217-02293)	Apr-17
Reference Probe ES3DV2	SN: 3013	31-Dec-15 (No. ES3-3013_Dec15)	Dec-16
DAE4	SN: 660	23-Dec-15 (No. DAE4-660_Dec15)	Dec-16
Secondary Standards	ID	Check Date (in house)	Scheduled Check
Power meter E4419B	SN: GB41293874	06-Apr-16 (in house check Jun-16)	In house check: Jun-18
Power sensor E4412A	SN: MY41498087	06-Apr-16 (in house check Jun-16)	In house check: Jun-18
Power sensor E4412A	SN: 000110210	06-Apr-16 (in house check Jun-16)	In house check: Jun-18
RF generator HP 8648C	SN: US3642U01700	04-Aug-99 (in house check Jun-16)	In house check: Jun-18
Network Analyzer HP 8753E	SN: US37390585	18-Oct-01 (in house check Oct-15)	In house check: Oct-16

Calibrated by:	Name Claudio Leubler	Function Laboratory Technician	Signature 
Approved by:	Katja Pokovic	Technical Manager	

Issued: September 28, 2016

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

**Calibration Laboratory of  
Schmid & Partner  
Engineering AG**  
Zeughausstrasse 43, 8004 Zurich, Switzerland



**S** Schweizerischer Kalibrierdienst  
**C** Service suisse d'étalonnage  
**S** Servizio svizzero di taratura  
Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)

Accreditation No.: **SCS 0108**

The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

**Glossary:**

TSL	tissue simulating liquid
NORM <sub>x,y,z</sub>	sensitivity in free space
ConvF	sensitivity in TSL / NORM <sub>x,y,z</sub>
DCP	diode compression point
CF	crest factor (1/duty_cycle) of the RF signal
A, B, C, D	modulation dependent linearization parameters
Polarization $\phi$	$\phi$ rotation around probe axis
Polarization $\vartheta$	$\vartheta$ rotation around an axis that is in the plane normal to probe axis (at measurement center), i.e., $\vartheta = 0$ is normal to probe axis
Connector Angle	information used in DASY system to align probe sensor X to the robot coordinate system

**Calibration is Performed According to the Following Standards:**

- a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- b) IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005
- c) IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- d) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

**Methods Applied and Interpretation of Parameters:**

- *NORM<sub>x,y,z</sub>*: Assessed for E-field polarization  $\vartheta = 0$  ( $f \leq 900$  MHz in TEM-cell;  $f > 1800$  MHz: R22 waveguide). NORM<sub>x,y,z</sub> are only intermediate values, i.e., the uncertainties of NORM<sub>x,y,z</sub> does not affect the E<sup>2</sup>-field uncertainty inside TSL (see below *ConvF*).
- *NORM(f)<sub>x,y,z</sub>* = *NORM<sub>x,y,z</sub>* \* *frequency\_response* (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of *ConvF*.
- *DCP<sub>x,y,z</sub>*: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- *PAR*: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- *A<sub>x,y,z</sub>; B<sub>x,y,z</sub>; C<sub>x,y,z</sub>; D<sub>x,y,z</sub>; VR<sub>x,y,z</sub>*: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- *ConvF and Boundary Effect Parameters*: Assessed in flat phantom using E-field (or Temperature Transfer Standard for  $f \leq 800$  MHz) and inside waveguide using analytical field distributions based on power measurements for  $f > 800$  MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to *NORM<sub>x,y,z</sub>* \* *ConvF* whereby the uncertainty corresponds to that given for *ConvF*. A frequency dependent *ConvF* is used in DASY version 4.4 and higher which allows extending the validity from  $\pm 50$  MHz to  $\pm 100$  MHz.
- *Spherical isotropy (3D deviation from isotropy)*: in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- *Sensor Offset*: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- *Connector Angle*: The angle is assessed using the information gained by determining the *NORM<sub>x</sub>* (no uncertainty required).



EX3DV4 – SN:3722

September 23, 2016

# Probe EX3DV4

## SN:3722

Manufactured: August 14, 2009  
Calibrated: September 23, 2016

Calibrated for DASY/EASY Systems  
(Note: non-compatible with DASY2 system!)



EX3DV4- SN:3722

September 23, 2016

## DASY/EASY - Parameters of Probe: EX3DV4 - SN:3722

### Basic Calibration Parameters

	Sensor X	Sensor Y	Sensor Z	Unc (k=2)
Norm ( $\mu\text{V}/(\text{V}/\text{m})^2$ ) <sup>A</sup>	0.52	0.49	0.56	$\pm 10.1\%$
DCP (mV) <sup>B</sup>	102.3	99.1	101.2	

### Modulation Calibration Parameters

UID	Communication System Name		A dB	B dB $\sqrt{\mu\text{V}}$	C	D dB	VR mV	Unc <sup>E</sup> (k=2)
0	CW	X	0.0	0.0	1.0	0.00	194.4	$\pm 2.7\%$
		Y	0.0	0.0	1.0		189.0	
		Z	0.0	0.0	1.0		195.0	

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>A</sup> The uncertainties of Norm X,Y,Z do not affect the E<sup>2</sup>-field uncertainty inside TSL (see Pages 5 and 6).

<sup>B</sup> Numerical linearization parameter: uncertainty not required.

<sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

EX3DV4- SN:3722

September 23, 2016

## DASY/EASY - Parameters of Probe: EX3DV4 - SN:3722

### Calibration Parameter Determined in Head Tissue Simulating Media

f (MHz) <sup>C</sup>	Relative Permittivity <sup>F</sup>	Conductivity (S/m) <sup>F</sup>	ConvF X	ConvF Y	ConvF Z	Alpha <sup>G</sup>	Depth <sup>G</sup> (mm)	Unc (k=2)
750	41.9	0.89	9.67	9.67	9.67	0.54	0.80	± 12.0 %
900	41.5	0.97	9.08	9.08	9.08	0.36	0.99	± 12.0 %
1810	40.0	1.40	7.71	7.71	7.71	0.37	0.80	± 12.0 %
2000	40.0	1.40	7.65	7.65	7.65	0.35	0.80	± 12.0 %
2450	39.2	1.80	6.90	6.90	6.90	0.37	0.80	± 12.0 %
5200	36.0	4.66	5.08	5.08	5.08	0.35	1.80	± 13.1 %
5300	35.9	4.76	4.78	4.78	4.78	0.40	1.80	± 13.1 %
5500	35.6	4.96	4.62	4.62	4.62	0.45	1.80	± 13.1 %
5800	35.3	5.27	4.34	4.34	4.34	0.50	1.80	± 13.1 %

<sup>C</sup> Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.

<sup>F</sup> At frequencies below 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.



EX3DV4- SN:3722

September 23, 2016

## DASY/EASY - Parameters of Probe: EX3DV4 - SN:3722

### Calibration Parameter Determined in Body Tissue Simulating Media

f (MHz) <sup>C</sup>	Relative Permittivity <sup>F</sup>	Conductivity (S/m) <sup>F</sup>	ConvF X	ConvF Y	ConvF Z	Alpha <sup>G</sup>	Depth <sup>G</sup> (mm)	Unc (k=2)
750	55.5	0.96	8.95	8.95	8.95	0.45	0.80	± 12.0 %
900	55.0	1.05	8.99	8.99	8.99	0.49	0.80	± 12.0 %
1810	53.3	1.52	7.45	7.45	7.45	0.36	0.85	± 12.0 %
2000	53.3	1.52	7.51	7.51	7.51	0.41	0.80	± 12.0 %
2450	52.7	1.95	6.97	6.97	6.97	0.38	0.90	± 12.0 %
5200	49.0	5.30	4.34	4.34	4.34	0.45	1.90	± 13.1 %
5300	48.9	5.42	4.09	4.09	4.09	0.50	1.90	± 13.1 %
5500	48.6	5.65	3.75	3.75	3.75	0.55	1.90	± 13.1 %
5800	48.2	6.00	3.70	3.70	3.70	0.60	1.90	± 13.1 %

<sup>C</sup> Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.


<sup>F</sup> At frequencies below 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

EX3DV4- SN:3722

September 23, 2016

### Frequency Response of E-Field (TEM-Cell:ifi110 EXX, Waveguide: R22)





Uncertainty of Frequency Response of E-field:  $\pm 6.3\%$  (k=2)

EX3DV4- SN:3722

September 23, 2016

### Receiving Pattern ( $\phi$ ), $\theta = 0^\circ$





Uncertainty of Axial Isotropy Assessment:  $\pm 0.5\%$  ( $k=2$ )

EX3DV4-SN:3722

September 23, 2016

### Dynamic Range $f(SAR_{head})$ (TEM cell, $f_{eval}=1900$ MHz)




Uncertainty of Linearity Assessment:  $\pm 0.6\%$  (k=2)

EX3DV4- SN:3722



September 23, 2016

### Conversion Factor Assessment



### Deviation from Isotropy in Liquid

Error ( $\phi, \vartheta$ ), f = 900 MHz



Uncertainty of Spherical Isotropy Assessment:  $\pm 2.6\%$  (k=2)





EX3DV4- SN:3722

September 23, 2016

### DASY/EASY - Parameters of Probe: EX3DV4 - SN:3722

#### Other Probe Parameters

Sensor Arrangement	Triangular
Connector Angle (°)	125.9
Mechanical Surface Detection Mode	enabled
Optical Surface Detection Mode	disabled
Probe Overall Length	337 mm
Probe Body Diameter	10 mm
Tip Length	9 mm
Tip Diameter	2.5 mm
Probe Tip to Sensor X Calibration Point	1 mm
Probe Tip to Sensor Y Calibration Point	1 mm
Probe Tip to Sensor Z Calibration Point	1 mm
Recommended Measurement Distance from Surface	1.4 mm





## ANNEX I DAE CALIBRATION CERTIFICATE

**Calibration Laboratory of**  
**Schmid & Partner**  
**Engineering AG**  
 Zeughausstrasse 43, 8004 Zurich, Switzerland



**S** Schweizerischer Kalibrierdienst  
**C** Service suisse d'étalonnage  
**S** Servizio svizzero di taratura  
**S** Swiss Calibration Service


Accredited by the Swiss Accreditation Service (SAS)  
 The Swiss Accreditation Service is one of the signatories to the EA  
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

Client **MET Laboratories**

Certificate No: **DAE3-584\_Sep16**

**CALIBRATION CERTIFICATE**

Object	DAE3 - SD 000 D03 AA - SN: 584																						
Calibration procedure(s)	QA CAL-06.v29 Calibration procedure for the data acquisition electronics (DAE)																						
Calibration date:	September 21, 2016																						
<p>This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).                  The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity &lt; 70%.</p> <p>Calibration Equipment used (M&amp;TE critical for calibration)</p> <table border="1"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Keithley Multimeter Type 2001</td> <td>SN: 0810278</td> <td>09-Sep-16 (No:19065)</td> <td>Sep-17</td> </tr> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Check Date (in house)</th> <th>Scheduled Check</th> </tr> <tr> <td>Auto DAE Calibration Unit</td> <td>SE UWS 053 AA 1001</td> <td>05-Jan-16 (in house check)</td> <td>In house check: Jan-17</td> </tr> <tr> <td>Calibrator Box V2.1</td> <td>SE UMS 006 AA 1002</td> <td>05-Jan-16 (in house check)</td> <td>In house check: Jan-17</td> </tr> </tbody> </table>				Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration	Keithley Multimeter Type 2001	SN: 0810278	09-Sep-16 (No:19065)	Sep-17	Secondary Standards	ID #	Check Date (in house)	Scheduled Check	Auto DAE Calibration Unit	SE UWS 053 AA 1001	05-Jan-16 (in house check)	In house check: Jan-17	Calibrator Box V2.1	SE UMS 006 AA 1002	05-Jan-16 (in house check)	In house check: Jan-17
Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration																				
Keithley Multimeter Type 2001	SN: 0810278	09-Sep-16 (No:19065)	Sep-17																				
Secondary Standards	ID #	Check Date (in house)	Scheduled Check																				
Auto DAE Calibration Unit	SE UWS 053 AA 1001	05-Jan-16 (in house check)	In house check: Jan-17																				
Calibrator Box V2.1	SE UMS 006 AA 1002	05-Jan-16 (in house check)	In house check: Jan-17																				
Calibrated by:	Name Eric Hainfeld	Function Technician	Signature 																				
Approved by:	Name Fin Bomholt	Function Deputy Technical Manager	Signature 																				
			Issued: September 21, 2016																				
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.																							

**Calibration Laboratory of**  
Schmid & Partner  
Engineering AG  
Zeughausstrasse 43, 8004 Zurich, Switzerland



**S** Schweizerischer Kalibrierdienst  
**C** Service suisse d'étalonnage  
**S** Servizio svizzero di taratura  
**S** Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)  
The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

### Glossary

DAE data acquisition electronics  
Connector angle information used in DASY system to align probe sensor X to the robot coordinate system.

### Methods Applied and Interpretation of Parameters

- *DC Voltage Measurement*: Calibration Factor assessed for use in DASY system by comparison with a calibrated instrument traceable to national standards. The figure given corresponds to the full scale range of the voltmeter in the respective range.
- *Connector angle*: The angle of the connector is assessed measuring the angle mechanically by a tool inserted. Uncertainty is not required.
- The following parameters as documented in the Appendix contain technical information as a result from the performance test and require no uncertainty.
  - *DC Voltage Measurement Linearity*: Verification of the Linearity at +10% and -10% of the nominal calibration voltage. Influence of offset voltage is included in this measurement.
  - *Common mode sensitivity*: Influence of a positive or negative common mode voltage on the differential measurement.
  - *Channel separation*: Influence of a voltage on the neighbor channels not subject to an input voltage.
  - *AD Converter Values with inputs shorted*: Values on the internal AD converter corresponding to zero input voltage
  - *Input Offset Measurement*: Output voltage and statistical results over a large number of zero voltage measurements.
  - *Input Offset Current*: Typical value for information; Maximum channel input offset current, not considering the input resistance.
  - *Input resistance*: Typical value for information: DAE input resistance at the connector, during internal auto-zeroing and during measurement.
  - *Low Battery Alarm Voltage*: Typical value for information. Below this voltage, a battery alarm signal is generated.
  - *Power consumption*: Typical value for information. Supply currents in various operating modes.



**DC Voltage Measurement**

A/D - Converter Resolution nominal

High Range: 1LSB = 6.1µV, full range = -100...+300 mV

Low Range: 1LSB = 61nV, full range = -1.....+3mV

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

Calibration Factors	X	Y	Z
High Range	404.497 ± 0.02% (k=2)	404.684 ± 0.02% (k=2)	404.109 ± 0.02% (k=2)
Low Range	3.92959 ± 1.50% (k=2)	3.91818 ± 1.50% (k=2)	3.94514 ± 1.50% (k=2)

**Connector Angle**

Connector Angle to be used in DASY system	142.0 ° ± 1 °
---	---------------

**Appendix (Additional assessments outside the scope of SCS0108)**

**1. DC Voltage Linearity**

High Range	Reading ( $\mu\text{V}$ )	Difference ( $\mu\text{V}$ )	Error (%)
Channel X + Input	199989.98	-6.78	-0.00
Channel X + Input	20002.67	1.05	0.01
Channel X - Input	-20001.03	-0.13	0.00
Channel Y + Input	199994.65	-2.21	-0.00
Channel Y + Input	20003.08	1.53	0.01
Channel Y - Input	-19995.08	5.84	-0.03
Channel Z + Input	199997.75	0.63	0.00
Channel Z + Input	19997.79	-3.79	-0.02
Channel Z - Input	-20002.74	-1.77	0.01

Low Range	Reading ( $\mu\text{V}$ )	Difference ( $\mu\text{V}$ )	Error (%)
Channel X + Input	2001.49	0.15	0.01
Channel X + Input	201.90	0.26	0.13
Channel X - Input	-198.61	-0.53	0.27
Channel Y + Input	2000.93	-0.41	-0.02
Channel Y + Input	201.53	-0.11	-0.05
Channel Y - Input	-199.00	-0.90	0.46
Channel Z + Input	2000.92	-0.28	-0.01
Channel Z + Input	200.91	-0.65	-0.32
Channel Z - Input	-199.60	-1.43	0.72

**2. Common mode sensitivity**

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	Common mode Input Voltage (mV)	High Range Average Reading ( $\mu\text{V}$ )	Low Range Average Reading ( $\mu\text{V}$ )
Channel X	200	3.25	1.52
	- 200	-1.08	-2.90
Channel Y	200	2.80	2.69
	- 200	-3.39	-3.72
Channel Z	200	-6.34	-6.43
	- 200	4.51	4.34

**3. Channel separation**

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	Input Voltage (mV)	Channel X ( $\mu\text{V}$ )	Channel Y ( $\mu\text{V}$ )	Channel Z ( $\mu\text{V}$ )
Channel X	200	-	-0.44	-3.66
Channel Y	200	7.35	-	0.04
Channel Z	200	6.60	5.86	-



**4. AD-Converter Values with inputs shorted**

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	High Range (LSB)	Low Range (LSB)
Channel X	16111	16149
Channel Y	16215	16800
Channel Z	16296	17501

**5. Input Offset Measurement**

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

Input 10MΩ

	Average (μV)	min. Offset (μV)	max. Offset (μV)	Std. Deviation (μV)
Channel X	1.36	-0.11	3.26	0.52
Channel Y	-0.30	-1.72	1.39	0.51
Channel Z	0.24	-1.16	1.64	0.46

**6. Input Offset Current**

Nominal Input circuitry offset current on all channels: <25fA

**7. Input Resistance** (Typical values for information)

	Zeroing (kOhm)	Measuring (MOhm)
Channel X	200	200
Channel Y	200	200
Channel Z	200	200

**8. Low Battery Alarm Voltage** (Typical values for information)

Typical values	Alarm Level (VDC)
Supply (+ Vcc)	+7.9
Supply (- Vcc)	-7.6

**9. Power Consumption** (Typical values for information)

Typical values	Switched off (mA)	Stand by (mA)	Transmitting (mA)
Supply (+ Vcc)	+0.01	+6	+14
Supply (- Vcc)	-0.01	-8	-9





## **ANNEX J 2.4 GHz MEASURED FLUID DIELECTRIC PARAMETERS**



# Vuzix M300 Smart Glasses

## Head Simulating Liquid Parameters for 2450MHz

November 28, 2016 08:01 AM

<b>Frequency</b>	<b>e'</b>	<b>e''</b>
<b>2.40000 GHz</b>	<b>38.786</b>	<b>13.813</b>
<b>2.40200 GHz</b>	<b>38.676</b>	<b>13.824</b>
<b>2.40400 GHz</b>	<b>38.683</b>	<b>13.838</b>
<b>2.40600 GHz</b>	<b>38.672</b>	<b>13.852</b>
<b>2.40800 GHz</b>	<b>38.673</b>	<b>13.862</b>
<b>2.41000 GHz</b>	<b>38.668</b>	<b>13.891</b>
<b>2.41200 GHz</b>	<b>38.656</b>	<b>13.894</b>
<b>2.41400 GHz</b>	<b>38.647</b>	<b>13.905</b>
<b>2.41600 GHz</b>	<b>38.649</b>	<b>13.906</b>
<b>2.41800 GHz</b>	<b>38.644</b>	<b>13.913</b>
<b>2.42000 GHz</b>	<b>38.635</b>	<b>13.914</b>
<b>2.42200 GHz</b>	<b>38.624</b>	<b>13.916</b>
<b>2.42400 GHz</b>	<b>38.620</b>	<b>13.918</b>
<b>2.42600 GHz</b>	<b>38.605</b>	<b>13.919</b>
<b>2.42800 GHz</b>	<b>38.606</b>	<b>13.920</b>
<b>2.43000 GHz</b>	<b>38.591</b>	<b>13.920</b>
<b>2.43200 GHz</b>	<b>38.592</b>	<b>13.921</b>
<b>2.43400 GHz</b>	<b>38.577</b>	<b>13.921</b>
<b>2.43600 GHz</b>	<b>38.560</b>	<b>13.922</b>
<b>2.43800 GHz</b>	<b>38.554</b>	<b>13.922</b>
<b>2.44000 GHz</b>	<b>38.548</b>	<b>13.923</b>
<b>2.44200 GHz</b>	<b>38.531</b>	<b>13.923</b>
<b>2.44400 GHz</b>	<b>38.531</b>	<b>13.923</b>
<b>2.44600 GHz</b>	<b>38.526</b>	<b>13.924</b>
<b>2.44800 GHz</b>	<b>38.223</b>	<b>13.924</b>
<b>2.45000 GHz</b>	<b>38.516</b>	<b>13.925</b>
<b>2.45200 GHz</b>	<b>38.491</b>	<b>13.928</b>
<b>2.45400 GHz</b>	<b>38.484</b>	<b>13.930</b>
<b>2.45600 GHz</b>	<b>38.479</b>	<b>13.936</b>
<b>2.45800 GHz</b>	<b>38.474</b>	<b>13.939</b>
<b>2.46000 GHz</b>	<b>38.448</b>	<b>13.943</b>
<b>2.46200 GHz</b>	<b>38.445</b>	<b>13.941</b>
<b>2.46400 GHz</b>	<b>38.432</b>	<b>13.941</b>



---

<b>2.46600 GHz</b>	<b>38.435</b>	<b>13.946</b>
<b>2.46800 GHz</b>	<b>38.423</b>	<b>13.950</b>
<b>2.47000 GHz</b>	<b>38.402</b>	<b>13.951</b>
<b>2.47200 GHz</b>	<b>38.401</b>	<b>13.955</b>
<b>2.47400 GHz</b>	<b>38.381</b>	<b>13.972</b>
<b>2.47600 GHz</b>	<b>38.370</b>	<b>13.969</b>
<b>2.47800 GHz</b>	<b>38.368</b>	<b>13.975</b>
<b>2.48000 GHz</b>	<b>38.349</b>	<b>13.984</b>
<b>2.48200 GHz</b>	<b>38.338</b>	<b>13.980</b>
<b>2.48400 GHz</b>	<b>38.326</b>	<b>13.988</b>
<b>2.48600 GHz</b>	<b>38.322</b>	<b>14.006</b>
<b>2.48800 GHz</b>	<b>38.307</b>	<b>14.020</b>
<b>2.49000 GHz</b>	<b>38.303</b>	<b>14.028</b>
<b>2.49200 GHz</b>	<b>37.288</b>	<b>14.036</b>
<b>2.49400 GHz</b>	<b>37.282</b>	<b>14.045</b>
<b>2.49600 GHz</b>	<b>37.263</b>	<b>14.047</b>
<b>2.49800 GHz</b>	<b>37.261</b>	<b>14.062</b>
<b>2.50000 GHz</b>	<b>37.256</b>	<b>14.074</b>



## **ANNEX K 5.0 GHz MEASURED FLUID DIELECTRIC PARAMETERS**



# Vuzix M300 Smart Glasses

## Head Simulating Liquid Parameters for 5200MHz

November 29, 2016 10:09 AM

Frequency	e'	e''
5.18000 GHz	34.753	15.733
5.18400 GHz	34.752	15.734
5.18720 GHz	34.751	15.736
5.19000 GHz	34.749	15.738
5.19360 GHz	34.749	15.738
5.19680 GHz	34.750	15.738
5.20000 GHz	34.749	15.734
5.20320 GHz	34.754	15.734
5.20640 GHz	34.748	15.744
5.20960 GHz	34.750	15.740
5.21280 GHz	34.751	15.735
5.21600 GHz	34.747	15.747
5.21920 GHz	34.746	15.745
5.22240 GHz	34.752	15.744
5.22560 GHz	34.745	15.754
5.22880 GHz	34.721	15.749
5.23200 GHz	34.721	15.755
5.23520 GHz	34.717	15.751
5.23840 GHz	34.716	15.751
5.24160 GHz	34.714	15.751
5.24480 GHz	34.700	15.754
5.24800 GHz	34.693	15.763
5.25120 GHz	34.675	15.759
5.25440 GHz	34.656	15.754
5.25760 GHz	34.651	15.756
5.26080 GHz	34.634	15.747
5.26400 GHz	34.634	15.760
5.26720 GHz	34.626	15.782
5.27040 GHz	34.622	15.784
5.27360 GHz	34.629	15.761
5.27680 GHz	34.613	15.783
5.28000 GHz	34.604	15.790
5.28320 GHz	34.602	15.790



---

<b>5.28640 GHz</b>	<b>34.587</b>	<b>15.796</b>
<b>5.28960 GHz</b>	<b>34.584</b>	<b>15.800</b>
<b>5.29280 GHz</b>	<b>34.595</b>	<b>15.803</b>
<b>5.29600 GHz</b>	<b>34.589</b>	<b>15.806</b>
<b>5.29920 GHz</b>	<b>34.582</b>	<b>15.820</b>
<b>5.30240 GHz</b>	<b>34.585</b>	<b>15.019</b>
<b>5.30560 GHz</b>	<b>34.585</b>	<b>15.017</b>
<b>5.30880 GHz</b>	<b>34.561</b>	<b>15.027</b>
<b>5.31200 GHz</b>	<b>34.573</b>	<b>15.037</b>
<b>5.31520 GHz</b>	<b>34.564</b>	<b>15.039</b>
<b>5.31840 GHz</b>	<b>34.563</b>	<b>15.047</b>
<b>5.32160 GHz</b>	<b>34.565</b>	<b>15.032</b>
<b>5.32480 GHz</b>	<b>34.558</b>	<b>15.030</b>
<b>5.32800 GHz</b>	<b>34.562</b>	<b>15.059</b>
<b>5.33120 GHz</b>	<b>34.545</b>	<b>15.046</b>
<b>5.33440 GHz</b>	<b>34.541</b>	<b>15.053</b>
<b>5.33760 GHz</b>	<b>34.545</b>	<b>15.057</b>
<b>5.34080 GHz</b>	<b>34.540</b>	<b>15.058</b>
<b>5.34400 GHz</b>	<b>34.533</b>	<b>15.059</b>
<b>5.34720 GHz</b>	<b>34.530</b>	<b>15.060</b>
<b>5.35040 GHz</b>	<b>34.527</b>	<b>15.048</b>
<b>5.35360 GHz</b>	<b>34.524</b>	<b>15.058</b>
<b>5.35680 GHz</b>	<b>34.511</b>	<b>15.052</b>
<b>5.36000 GHz</b>	<b>34.506</b>	<b>15.058</b>





# Vuzix M300 Smart Glasses

## Head Simulating Liquid Parameters for 5500MHz

November 29, 2016 10:18 AM

Frequency	e'	e''
5.49000 GHz	34.559	16.402
5.49168 GHz	34.557	16.404
5.49309 GHz	34.557	16.406
5.49449 GHz	34.556	16.406
5.49590 GHz	34.556	16.406
5.49730 GHz	34.554	16.407
5.49871 GHz	34.554	16.408
5.50000 GHz	34.552	16.408
5.50140 GHz	34.535	16.424
5.50281 GHz	34.533	16.416
5.50421 GHz	34.541	16.419
5.50562 GHz	34.522	16.420
5.50702 GHz	34.517	16.422
5.50843 GHz	34.517	16.411
5.50983 GHz	34.513	16.421
5.51124 GHz	34.507	16.431
5.51265 GHz	34.498	16.435
5.51405 GHz	34.495	16.427
5.51546 GHz	34.488	16.435
5.51686 GHz	34.488	16.436
5.51827 GHz	34.478	16.441
5.51967 GHz	34.480	16.435
5.52108 GHz	34.480	16.458
5.52248 GHz	34.486	16.452
5.52389 GHz	34.469	16.447
5.52530 GHz	34.468	16.449
5.52670 GHz	34.478	16.444
5.52811 GHz	34.480	16.455
5.52951 GHz	34.467	16.452
5.53092 GHz	34.475	16.450
5.53232 GHz	34.470	16.462
5.53373 GHz	34.461	16.457
5.53514 GHz	34.469	16.455



<b>5.53654 GHz</b>	<b>34.459</b>	<b>16.453</b>
<b>5.53795 GHz</b>	<b>34.459</b>	<b>16.452</b>
<b>5.53935 GHz</b>	<b>34.450</b>	<b>16.454</b>
<b>5.54076 GHz</b>	<b>34.434</b>	<b>16.439</b>
<b>5.54216 GHz</b>	<b>34.451</b>	<b>16.452</b>
<b>5.54357 GHz</b>	<b>34.456</b>	<b>16.449</b>
<b>5.54497 GHz</b>	<b>34.448</b>	<b>16.451</b>
<b>5.54638 GHz</b>	<b>34.445</b>	<b>16.442</b>
<b>5.54779 GHz</b>	<b>34.447</b>	<b>16.444</b>
<b>5.54919 GHz</b>	<b>34.450</b>	<b>16.455</b>
<b>5.55060 GHz</b>	<b>34.446</b>	<b>16.445</b>
<b>5.55200 GHz</b>	<b>34.440</b>	<b>16.452</b>
<b>5.55341 GHz</b>	<b>34.447</b>	<b>16.461</b>
<b>5.55481 GHz</b>	<b>34.445</b>	<b>16.448</b>
<b>5.55622 GHz</b>	<b>34.435</b>	<b>16.446</b>
<b>5.55763 GHz</b>	<b>34.449</b>	<b>16.439</b>
<b>5.55903 GHz</b>	<b>34.443</b>	<b>16.443</b>
<b>5.56044 GHz</b>	<b>34.438</b>	<b>16.447</b>
<b>5.56184 GHz</b>	<b>34.431</b>	<b>16.446</b>
<b>5.56325 GHz</b>	<b>34.436</b>	<b>16.451</b>
<b>5.56465 GHz</b>	<b>34.418</b>	<b>16.442</b>
<b>5.56606 GHz</b>	<b>34.423</b>	<b>16.457</b>
<b>5.56746 GHz</b>	<b>34.422</b>	<b>16.440</b>
<b>5.56887 GHz</b>	<b>34.420</b>	<b>16.446</b>
<b>5.57028 GHz</b>	<b>34.429</b>	<b>16.451</b>
<b>5.57168 GHz</b>	<b>34.411</b>	<b>16.449</b>
<b>5.57309 GHz</b>	<b>34.418</b>	<b>16.455</b>
<b>5.57449 GHz</b>	<b>34.415</b>	<b>16.457</b>
<b>5.57590 GHz</b>	<b>34.412</b>	<b>16.452</b>
<b>5.57730 GHz</b>	<b>34.399</b>	<b>16.445</b>
<b>5.57871 GHz</b>	<b>34.403</b>	<b>16.455</b>
<b>5.58000 GHz</b>	<b>34.421</b>	<b>16.491</b>
<b>5.58152 GHz</b>	<b>34.388</b>	<b>16.451</b>
<b>5.58293 GHz</b>	<b>34.387</b>	<b>16.448</b>
<b>5.58433 GHz</b>	<b>34.386</b>	<b>16.451</b>
<b>5.58574 GHz</b>	<b>34.407</b>	<b>16.459</b>
<b>5.58714 GHz</b>	<b>34.389</b>	<b>16.468</b>
<b>5.58855 GHz</b>	<b>34.382</b>	<b>16.471</b>



# Vuzix M300 Smart Glasses

## Head Simulating Liquid Parameters for 5800MHz

November 29, 2016 10:29 AM

<b>Frequency</b>	<b>e'</b>	<b>e''</b>
<b>5.75000 GHz</b>	<b>34.249</b>	<b>16.621</b>
<b>5.75192 GHz</b>	<b>34.249</b>	<b>16.621</b>
<b>5.75333 GHz</b>	<b>34.247</b>	<b>16.621</b>
<b>5.75473 GHz</b>	<b>34.246</b>	<b>16.621</b>
<b>5.75614 GHz</b>	<b>34.243</b>	<b>16.621</b>
<b>5.75755 GHz</b>	<b>34.243</b>	<b>16.621</b>
<b>5.75895 GHz</b>	<b>34.244</b>	<b>16.621</b>
<b>5.76000 GHz</b>	<b>34.243</b>	<b>16.621</b>
<b>5.76024 GHz</b>	<b>34.237</b>	<b>16.628</b>
<b>5.76164 GHz</b>	<b>34.233</b>	<b>16.621</b>
<b>5.76305 GHz</b>	<b>34.235</b>	<b>16.617</b>
<b>5.76445 GHz</b>	<b>34.236</b>	<b>16.623</b>
<b>5.76586 GHz</b>	<b>34.229</b>	<b>16.623</b>
<b>5.76726 GHz</b>	<b>34.232</b>	<b>16.615</b>
<b>5.76867 GHz</b>	<b>34.227</b>	<b>16.619</b>
<b>5.77008 GHz</b>	<b>34.214</b>	<b>16.616</b>
<b>5.77148 GHz</b>	<b>34.228</b>	<b>16.619</b>
<b>5.77289 GHz</b>	<b>34.227</b>	<b>16.614</b>
<b>5.77429 GHz</b>	<b>34.228</b>	<b>16.602</b>
<b>5.77570 GHz</b>	<b>34.223</b>	<b>16.606</b>
<b>5.77108 GHz</b>	<b>34.215</b>	<b>16.597</b>
<b>5.77851 GHz</b>	<b>34.220</b>	<b>16.505</b>
<b>5.77991 GHz</b>	<b>34.230</b>	<b>16.600</b>
<b>5.78132 GHz</b>	<b>34.212</b>	<b>16.599</b>
<b>5.78273 GHz</b>	<b>34.210</b>	<b>16.596</b>
<b>5.78413 GHz</b>	<b>34.214</b>	<b>16.596</b>
<b>5.78554 GHz</b>	<b>34.207</b>	<b>16.599</b>
<b>5.78694 GHz</b>	<b>34.199</b>	<b>16.600</b>
<b>5.78835 GHz</b>	<b>34.194</b>	<b>16.600</b>
<b>5.78975 GHz</b>	<b>34.191</b>	<b>16.594</b>
<b>5.79116 GHz</b>	<b>34.202</b>	<b>16.504</b>
<b>5.79257 GHz</b>	<b>34.187</b>	<b>16.589</b>
<b>5.79397 GHz</b>	<b>34.179</b>	<b>16.594</b>



---

<b>5.79538 GHz</b>	<b>34.186</b>	<b>16.593</b>
<b>5.79678 GHz</b>	<b>34.179</b>	<b>16.591</b>
<b>5.79819 GHz</b>	<b>34.170</b>	<b>16.591</b>
<b>5.79959 GHz</b>	<b>34.171</b>	<b>16.595</b>
<b>5.80000 GHz</b>	<b>34.182</b>	<b>16.592</b>
<b>5.80240 GHz</b>	<b>34.175</b>	<b>16.592</b>
<b>5.80381 GHz</b>	<b>34.167</b>	<b>16.589</b>
<b>5.80522 GHz</b>	<b>34.174</b>	<b>16.594</b>
<b>5.80662 GHz</b>	<b>34.165</b>	<b>16.596</b>
<b>5.80803 GHz</b>	<b>34.155</b>	<b>16.598</b>
<b>5.80943 GHz</b>	<b>34.151</b>	<b>16.504</b>
<b>5.81084 GHz</b>	<b>34.139</b>	<b>16.592</b>
<b>5.81224 GHz</b>	<b>34.129</b>	<b>16.577</b>
<b>5.81365 GHz</b>	<b>34.141</b>	<b>16.584</b>
<b>5.81506 GHz</b>	<b>34.123</b>	<b>16.599</b>
<b>5.81646 GHz</b>	<b>34.119</b>	<b>16.591</b>
<b>5.81787 GHz</b>	<b>34.106</b>	<b>16.595</b>
<b>5.81927 GHz</b>	<b>34.108</b>	<b>16.592</b>
<b>5.82068 GHz</b>	<b>34.104</b>	<b>16.595</b>
<b>5.82208 GHz</b>	<b>34.102</b>	<b>16.587</b>
<b>5.82349 GHz</b>	<b>34.110</b>	<b>16.596</b>
<b>5.82489 GHz</b>	<b>34.116</b>	<b>16.591</b>
<b>5.82500 GHz</b>	<b>34.165</b>	<b>16.583</b>
<b>5.82771 GHz</b>	<b>34.096</b>	<b>16.599</b>
<b>5.82911 GHz</b>	<b>34.104</b>	<b>16.598</b>
<b>5.83052 GHz</b>	<b>34.070</b>	<b>16.598</b>
<b>5.83192 GHz</b>	<b>34.083</b>	<b>16.583</b>
<b>5.83333 GHz</b>	<b>34.071</b>	<b>16.582</b>
<b>5.83473 GHz</b>	<b>34.079</b>	<b>16.572</b>
<b>5.83614 GHz</b>	<b>34.083</b>	<b>16.570</b>
<b>5.83755 GHz</b>	<b>34.082</b>	<b>16.573</b>
<b>5.83895 GHz</b>	<b>34.096</b>	<b>16.573</b>
<b>5.84000 GHz</b>	<b>34.091</b>	<b>16.574</b>



## **ANNEX L PHANTOM CERTIFICATE OF CONFORMITY**

Schmid & Partner Engineering AG

**s p e a g**

Zeughausstrasse 43, 8004 Zurich, Switzerland  
 Phone +41 1 245 9700, Fax +41 1 245 9779  
 info@speag.com, http://www.speag.com

**Certificate of conformity / First Article Inspection**

Item	SAM Twin Phantom V4.0
Type No	QD 000 P40 C
Series No	TP-1150 and higher
Manufacturer / Origin	Untersee Composites Hauptstr. 69 CH-8559 Fruthwilen Switzerland

**Tests**

The series production process used allows the limitation to test of first articles. Complete tests were made on the pre-series Type No. QD 000 P40 AA, Serial No. TP-1001 and on the series first article Type No. QD 000 P40 BA, Serial No. TP-1006. Certain parameters have been retested using further series units (called samples).

Test	Requirement	Details	Units tested
Shape	Compliance with the geometry according to the CAD model.	IT'IS CAD File (*)	First article, Samples
Material thickness	Compliant with the requirements according to the standards	2mm +/- 0.2mm in specific areas; 6mm +/- 0.2mm at ERP	First article, Samples
Material parameters	Dielectric parameters for required frequencies	200 MHz – 3 GHz Relative permittivity < 5 Loss tangent < 0.05.	Material sample TP 104-5
Material resistivity	The material has been tested to be compatible with the liquids defined in the standards if handled and cleaned according to the instructions	DEGMBE based simulating liquids	Pre-series, First article, Samples

**Standards**

- [1] CENELEC EN 50361
- [2] IEEE Std 1528-200x Draft CD 1.1 (Dec 02)
- [3] IEC 62209/CD (Nov 02)
- (\*) The IT'IS CAD file is derived from [2] and is also within the tolerance requirements of the shapes of [1] and [3].

**Conformity**

Based on the sample tests above, we certify that this item is in compliance with the uncertainty requirements of SAR measurements specified in standard [1] and draft standards [2] and [3].

**Date** 7.8.2003

**Signature / Stamp**

**s p e a g**  
 Schmid & Partner Engineering AG  
 Zeughausstrasse 43, 8004 Zurich, Switzerland  
 Phone +41 1 245 9700, Fax +41 1 245 9779  
 info@speag.com, http://www.speag.com