

### Declaration of the Modular Approval

<b>Applicant / Grantee</b>	EnzyTek Technology Inc.
<b>FCC ID:</b>	2AABGQ4020-1M
<b>Model:</b>	BTA-QA4020-1

**The single module transmitter has been evaluated then tested meeting the requirements under Part 15C Section 212 as below:**

<b>Modular approval requirement</b>	<b>EUT Condition</b>	<b>Comply</b>
(a) The radio elements of the modular transmitter must have their own shielding. The physical crystal and tuning capacitors may be located external to the shielded radio elements.	Modular transmitter has their own shielding.	Y
(b) The modular transmitter must have buffered modulation/data inputs (if such inputs are provided) to ensure that the module will comply with part 15 requirements under conditions of excessive data rates or over-modulation.	Chipset "QCA4020" have buffered modulation/data inputs ( can refer the schematics)and comply with part 15 requirements under conditions of excessive data rates or over-modulation.	Y
(c) The modular transmitter must have its own power supply regulation.	Chipset "QCA4020" have its own power supply regulation ( can refer the schematics)	Y
(d) The modular transmitter must comply with the antenna and transmission system requirements of Sections 15.203, 15.204(b) and 15.204(c). The antenna must either be permanently attached or employ a "unique" antenna coupler (at all connections between the module and the antenna, including the cable). The "professional installation" provision of Section 15.203 is not applicable to modules but can apply to limited modular approvals under paragraph (b) of this section.	Can refer the internal photo.	Y
(e) The modular transmitter must be tested in a stand-alone configuration, <i>i.e.</i> , the module must not be inside another device during testing for compliance with part 15	Can refer the test setup photo.	Y

<p>requirements. Unless the transmitter module will be battery powered, it must comply with the AC line conducted requirements found in Section 15.207. AC or DC power lines and data input/output lines connected to the module must not contain ferrites, unless they will be marketed with the module (see Section 15.27(a)). The length of these lines shall be the length typical of actual use or, if that length is unknown, at least 10 centimeters to insure that there is no coupling between the case of the module and supporting equipment. Any accessories, peripherals, or support equipment connected to the module during testing shall be unmodified and commercially available (see Section 15.31(i)) must not be inside another device during testing.</p>		
<p>(f) The modular transmitter must be equipped with either a permanently affixed label or must be capable of electronically displaying its FCC identification number.</p>	Can refer the label.	Y
<p>(g) The modular transmitter must comply with any specific rules or operating requirements that ordinarily apply to a complete transmitter and the manufacturer must provide adequate instructions along with the module to explain any such requirements. A copy of these instructions must be included in the application for equipment authorization requirements, which are based on the intended use/configurations.</p>	Can refer the user manual.	Y
<p>(h) The modular transmitter must comply with any applicable RF exposure requirements in its final configuration.</p>	Can refer the MPE report.	Y

**EnzyTek**

**EnzyTek Technology, Inc.**

**Dated** 2019/7/2

**By:** \_\_\_\_\_

Peter Hsieh

Peter Hsieh

\_\_\_\_\_  
**Signature**

\_\_\_\_\_  
**Printed**

**Title:** Manager

**On behalf of :** EnzyTek Technology Inc.

**Telephone:** 886-3-5736708