



BUREAU
VERITAS

Test Report No.: FM2102WSZ0052-2

RF EXPOSURE REPORT

Applicant	Medtrum Technologies Inc.
Address	7F, Building 8, No.200 Niudun Road, Shanghai 201203, China

Manufacturer or Supplier	Medtrum Technologies Inc.
Address	Building 3 and 6F~7F, Building 8, NO.200, Niudun Road, Shanghai 201203, China
Product	Pump Base
Brand Name	N/A
Model	MD0201
Additional Model & Model Difference	N/A
Date of tests	Feb. 05, 2021 ~ Mar. 09, 2021

FCC Part 2 (Section 2.1093)

KDB 447498 D01

IEEE C95.1

CONCLUSION: The submitted sample was found to COMPLY with the test requirement

Tested by Aaron Liang Project Engineer / EMC Department	Approved by David Huang Supervisor / EMC Department

Date: Mar. 10, 2021

This report is governed by, and incorporates by reference, CPS Conditions of Service as posted at the date of issuance of this report at <http://www.bureauveritas.com/home/about-us/our-business/cps/about-us/terms-conditions/> and is intended for your exclusive use. Any copying or replication of this report to or for any other person or entity, or use of our name or trademark, is permitted only with our prior written permission. This report sets forth our findings solely with respect to the test samples identified herein. The results set forth in this report are not indicative or representative of the quality or characteristics of the lot from which a test sample was taken or any similar or identical product unless specifically and expressly noted. Our report includes all of the tests requested by you and the results thereof based upon the information that you provided to us. Measurement uncertainty is only provided upon request for accredited tests. You have 60 days from date of issuance of this report to notify us of any material error or omission caused by our negligence or if you require measurement uncertainty; provided, however, that such notice shall be in writing and shall specifically address the issue you wish to raise. A failure to raise such issue within the prescribed time shall constitute you unqualified acceptance of the completeness of this report, the tests conducted and the correctness of the report contents.

Bureau Veritas (Shenzhen)
Consumer Products Services Co., Ltd.

Zone A, Floor 1, Building 2, Wan Ye Long Technology
Park, South Side of Zhoushi Road, Bao'an District,
Shenzhen Guangdong, 518108, China.

Tel: +86-755-26014629 Ext.800
Email: customerservice.dg@cn.bureauveritas.com



Table of Contents

RELEASE CONTROL RECORD	3
1. CERTIFICATION.....	4
2. RF EXPOSURE DEFINE	5
3. CLASSIFICATION	5
4. SAR TEST EXCLUSION THRESHOLDS	6



BUREAU
VERITAS

Test Report No.: FM2102WSZ0052-2

RELEASE CONTROL RECORD

ISSUE NO.	REASON FOR CHANGE	DATE ISSUED
FM2102WSZ0052-2	Original release	Mar. 10, 2021

**Bureau Veritas (Shenzhen)
Consumer Products Services Co., Ltd.**

Zone A, Floor 1, Building 2, Wan Ye Long Technology
Park, South Side of Zhoushi Road, Bao'an District,
Shenzhen Guangdong, 518108, China.

Tel: +86-755-26014629 Ext.800
Email: customerservice.dg@cn.bureauveritas.com



**BUREAU
VERITAS** Test Report No.: FM2102WSZ0052-2

1. CERTIFICATION

FCC ID:	TG3-HXEP575
PRODUCT:	Pump Base
BRAND NAME:	N/A
MODEL NO.:	MD0201
ADDITIONAL NO.:	N/A
TEST SAMPLE:	Engineering Sample
APPLICANT:	Medtrum Technologies Inc.
STANDARDS:	FCC Part 2 (Section 2.1093)
	KDB 447498 D01
	IEEE C95.1



BUREAU
VERITAS

Test Report No.: FM2102WSZ0052-2

2. RF EXPOSURE DEFINE

The corresponding SAR Exclusion Threshold condition, listed below:

- 1) The 1-g and 10-g SAR test exclusion thresholds for 100 MHz to 6 GHz at test separation distances ≤ 50 mm are determined by:
[(max. power of channel, including tune-up tolerance, mW)/(min. test separation distance, mm)] $\cdot [\sqrt{f(\text{GHz})}] \leq 3.0$ for 1-g SAR and ≤ 7.5 for 10-g extremity SAR, 16 where
 - $f(\text{GHz})$ is the RF channel transmit frequency in GHz
 - Power and distance are rounded to the nearest mW and mm before calculation
 - The result is rounded to one decimal place for comparison

The test exclusions are applicable only when the minimum test separation distance is ≤ 50 mm and for transmission frequencies between 100 MHz and 6 GHz. When the minimum test separation distance is < 5 mm, a distance of 5 mm is applied to determine SAR test exclusion.

- 2) At 100 MHz to 6 GHz and for test separation distances > 50 mm, the SAR test exclusion threshold is determined according to the following:
 - a) [Threshold at 50 mm in step 1) + (test separation distance - 50 mm) $\cdot (f(\text{MHz})/150)$] mW, at 100MHz to 1500 MHz
 - b) [Threshold at 50 mm in step 1) + (test separation distance - 50 mm) $\cdot 10$] mW at > 1500 MHz and ≤ 6 GHz
- 3) At frequencies below 100 MHz, the following may be considered for SAR test exclusion.
 - a) The threshold at the corresponding test separation distance at 100 MHz in step 2) is multiplied by $[1 + \log(100/f(\text{MHz}))]$ for test separation distances > 50 mm and < 200 mm.
 - b) The threshold determined by the equation in a) for 50 mm and 100 MHz is multiplied by $\frac{1}{2}$ for test separation distances ≤ 50 mm.
 - c) SAR measurement procedures are not established below 100 MHz. When SAR test exclusion cannot be applied, a KDB inquiry is required to determine SAR evaluation requirements for any test results to be acceptable.

3. CLASSIFICATION

The antenna of this product, under normal use condition, is at less than 20cm away from the body of the user. So, this device is classified as **Portable Device**.



4. SAR TEST EXCLUSION THRESHOLDS

The tuned conducted Average Power (declared by client)

Mode	Frequency (MHz)	Target Power (dBm)	Tolerance (dBm)	Lower Tolerance (dBm)	Upper Tolerance (dBm)
GFSK	2402-2480	-6	+2	-8	-4

The measured conducted Average Power

Mode	Frequency (MHz)	Averaged Power (dBm)
GFSK	2480	-6.08

SAR Test Exclusion Thresholds

Frequency (MHz)	Maximum source-based time averaged conducted output power (dBm)	Minimum separation distance (mm)	Result of Eq. 1	Limit for 1-g SAR	Limit for 10-g extremity SAR	Verdict
2402-2480	-4	5	0.125	3	7.5	Exempt from SAR

Conclusion

Therefore this device complies with FCC's RF radiation exposure limits for general population without SAR evaluation.