


VER-REV	ZONE	DESCRIPTION OF CHANGE (SEE ACCS FOR APPROVALS)	DRFT	DATE
01-A	ALL	INITIAL RELEASE	BJM	11/11/13
02-B	ALL	CHANGE ID NUMBERS FOR NEW RFID READER. CORRECTED RADII FROM .05" TO .10" PREVIOUS SPECIFIED RADII DID NOT MATCH RADII OF ARTWORK. 3M #9672 MP ADHESIVE IS DISCONTINUED. DRAWING CHANGED TO REFLECT EQUIVALENT 3M REPLACEMENT ADHESIVE ON NOTE 2. ADDED PERFORMANCE INSPECTION POINT AND NEW SUPPLIER'S UL RECOGNIZED MATERIAL TO NOTE 7. ADDED PERFORMANCE INSPECTION POINT TO NOTE 8.	LPR	05/01/20



FCC ID: 2AAZF-ACCRX02
 IC: 11508A-ACCRX02
 Model: ACCRX02

Artwork is for reference only.

NOTES: (UNLESS OTHERWISE SPECIFIED)

- MATERIAL: .010" FINE TEXTURED LEXAN.
- ADHESIVE: .005" 3M #9672 (PERM) WITH PAPER LINER
- BACKGROUND: WHITE
- TEXT: BLACK
- SUPPLIER TO PROVIDE CERTIFICATE OF CONFORMANCE
- USE .AI FILE FOR ARTWORK
- P01 P/F SUPPLIER TO PROVIDE UL RECOGNIZED COMPONENT SYMBOL WITH COMPANY NAME AND MATERIAL. "SUBSURFACE POLYCARBONATE 2" or "SS POLYCARB1974" DESIGNATION ON SMALLEST BAG OF LABELS.
- P02 P/F THE SUPPLIER IS TO PROVIDE A RoHS COMPLIANCE CERTIFICATE OF CONFORMANCE WITH EVERY SHIPMENT.
- BAG AND TAG WITH PART NUMBER, REVISION LEVEL AND SHELF LIFE DURATION.

UNLESS OTHERWISE SPECIFIED:		DRAWN BY:		Intuitive Surgical		
LINEAR: XXX = ± 0.02 XXXX = ± 0.01 ANGLES: ± 0.5		B. MAY				
INTUITIVE SURGICAL CONFIDENTIAL <small>THIS DOCUMENT CONTAINS CONFIDENTIAL AND PROPRIETARY INFORMATION. UNLESS GRANTED WRITTEN PERMISSION FROM INTUITIVE SURGICAL, SUCH INFORMATION SHALL NOT BE PUBLISHED OR DISCLOSED TO OTHERS OR USED FOR ANY PURPOSE OTHER THAN ON BEHALF OF INTUITIVE SURGICAL.</small>		DESIGN ENGINEER:		DRAWING TITLE: LABEL, IS4000,RFID, ACCRX02		
		J. LATHROP				
		MFG ENGINEER:		DRAWING NUMBER - VERSION:		
		B. BLAIR		510196-02		
		DATE:	DRW SIZE:	DRW SCALE:	REV:	SHEET:
		05/01/20	B	2:1	B	1 OF 1