



## RF exposure Estimation for 2ABAFMD20XYSERIES

### 1. Introduction

FCC ID: 2ABAFMD20XYSERIES, MD2070 is a Digital Automatic Blood Pressure Monitor, which contain Bluetooth function inside.

Note: models << MD2070 >> and << MD2020, MD2030, MD2060, MD2080, MD2021, MD2031, MD2061, MD2071 >> have the same technical construction including circuit diagram, PCB Layout, components and component layout, all electrical construction and mechanical construction. The difference lies only in Backlight, LCD Type, WHO LED, DC Jack of the different models. MD2070 was selected as representative models to perform RF exposure estimation, other models are deemed to fulfill RF exposure estimation.

### 2. Limit and Guidelines on Exposure to Electromagnetic Fields

According to §15.247(e)(i) and §1.1307(b)(1), systems operating under the provisions of this section shall be operated in a manner that ensure that the public is not exposed to radio frequency energy level in excess of the Commission's guideline.

According to KDB 447498 D01 Mobile Portable RF Exposure v05r01, no SAR required if power is lower than the flowing threshold:

The 1-g and 10-g SAR test exclusion thresholds for 100 MHz to 6 GHz at test separation distances  $\leq 50$  mm are determined by:

$[(\text{max. power of channel, including tune-up tolerance, mW}) / (\text{min. test separation distance, mm})] \cdot \sqrt{f(\text{GHz})} \leq 3.0$  for 1-g SAR and  $\leq 7.5$  for 10-g extremity SAR, where

- $f(\text{GHz})$  is the RF channel transmit frequency in GHz
- Power and distance are rounded to the nearest mW and mm before calculation
- The result is rounded to one decimal place for comparison
- 3.0 and 7.5 are referred to as the numeric thresholds in the step 2 below

The test exclusions are applicable only when the minimum test separation distance is  $\leq 50$  mm and for transmission frequencies between 100 MHz and 6 GHz. When the minimum test separation distance is  $< 5$  mm, a distance of 5 mm according to 5) in section 4.1 is applied to determine SAR test exclusion.

### 3. Calculation method

$[(\text{max. power of channel, including tune-up tolerance, mW}) / (\text{min. test separation distance, mm})] \cdot \sqrt{f(\text{GHz})} \leq 3.0$

Conducted Power + tune up tolerance = 0.69mW

Distance = 5 mm

$f = 2.402$  GHz

$[0.69/5] \cdot \text{SQRT}(2.402) = 0.214$

$0.214 \leq 3.0$

Therefore, excluded from SAR testing.



Product Service

TÜV SÜD Certification and Testing (China) Co., Ltd. Shenzhen Branch

Reviewed by:

A handwritten signature in black ink, appearing to be 'Ken Li'.

Ken Li/EMC Project Manager  
Date: 2014-01-21

Prepared By:

A handwritten signature in blue ink, appearing to be 'Felix Li'.

Felix Li/EMC Project Engineer  
Date: 2014-01-21