

# Declaration of Compliance

FCC ID: 2ABTAHNL

Product Name: Automatic Wrist Blood Pressure Monitor

Model No: HL158HM

1. According to FCC CFR 47 §1.1310, the criteria listed in the following table shall be used to evaluate the environmental impact of human exposure to radio frequency (RF) radiation as specified in 1.1307(b).

Table 1 Limits for Maximum Permissible Exposure

Frequency Range (MHz)	Electric Field Strength (V/m)	Magnetic Field Strength (A/m)	Power Density (mW/cm <sup>2</sup> )	Average Time (Minutes)
(A) Limits For Occupational / Control Exposures (f = frequency)				
30-300	61.4	0.163	1.0	6
300-1500	...	...	f/300	6
1500-100,000	...	...	5.0	6
(B) Limits For General Population / Uncontrolled Exposure (f = frequency)				
30-300	27.5	0.073	0.2	30
300-1500	...	...	f/1500	30
1500-100,000	...	...	1.0	30

2. MPE Calculation

INVENTEC BESTA CO., LTD declares that the product described above has been evaluated and found to comply with the RF exposure limits for humans, as specified based on ANSI/FCC recommendation.

Based on safety distance 20cm, the antenna gain is 2.66 dBi, and the power output is 1.469mW, the power density is  $7.48 \times 10^{-5} \text{mW/cm}^2$

RF Exposure Calculations:

$$S = (P * G) / (4 * \pi * r^2) \text{ or } r = \sqrt{(P * G) / (4 * \pi * S)}$$

Where S = Power Density in mW/cm<sup>2</sup>

P=1.671dBm=1.469mw

G =2.66dBi=1.845Numerical

r = 20cm

$$S = 1.469 * 1.845 / 4 * \pi * 20^2 = 2.756 / 5026.548 = 7.48 \times 10^{-5} \text{mW/cm}^2$$

Sincerely Yours,




---

Mr. Ben Cheng  
 Manager  
 AUDIX Technology Corporation