

APPLICATION FOR CERTIFICATION  
On Behalf of  
Health & Life Co., Ltd.  
Automatic Upper Arm Blood Pressure Monitor  
Model No.: HL858CB  
FCC ID: 2ABTAHNL85CB

Prepared for : Health & Life Co., Ltd.  
9F, No.186, Jian Yi Road, Zhonghe  
District, New Taipei City, Taiwan

Prepared by : AUDIX Technology Corporation  
EMC Department  
No. 53-11, Dingfu, Linkou Dist.,  
New Taipei City 244, Taiwan

Tel : (02) 2609-9301, 2609-2133  
Fax: (02) 2609-9303

File Number : C1M1409047  
Report Number : EM-F140556  
Date of Test : 2014. 09. 11 ~ 17  
Date of Report : 2014. 09. 18

## TABLE OF CONTENTS

Description	Page
TEST REPORT CERTIFICATION .....	4
<b>1. DESCRIPTION OF REVISION HISTORY .....</b>	<b>5</b>
<b>2. GENERAL INFORMATION .....</b>	<b>6</b>
2.1. Description of Device (EUT).....	6
2.2. Tested Supporting System Details.....	7
2.3. Description of Test Facility .....	8
2.4. Measurement Uncertainty.....	8
<b>3. CONDUCTED EMISSION MEASUREMENT .....</b>	<b>9</b>
3.1. Test Equipment.....	9
3.2. Block Diagram of Test Setup.....	9
3.3. Powerline Conducted Emission Limit (§15.207).....	9
3.4. Operating Condition of EUT .....	10
3.5. Test Procedure .....	10
3.6. Powerline Conducted Emission Measurement Results.....	10
<b>4. RADIATED EMISSION MEASUREMENT .....</b>	<b>13</b>
4.1. Test Equipment.....	13
4.2. Test Setup .....	13
4.3. Radiated Emission Limits (§15.209) .....	15
4.4. Operating Condition of EUT .....	15
4.5. Test Procedure .....	16
4.6. Test Results.....	17
<b>5. 6dB BANDWIDTH MEASUREMENT .....</b>	<b>25</b>
5.1. Test Equipment.....	25
5.2. Block Diagram of Test Setup.....	25
5.3. Specification Limits [§15.247(a)(2)] .....	25
5.4. Operating Condition of EUT .....	25
5.5. Test Procedure .....	25
5.6. Test Results.....	26
<b>6. MAXIMUM PEAK OUTPUT POWER MEASUREMENT .....</b>	<b>29</b>
6.1. Test Equipment.....	29
6.2. Block Diagram of Test Setup.....	29
6.3. Specification Limits [§15.247(b)-(3)].....	29
6.4. Operating Condition of EUT .....	29
6.5. Test Procedure .....	29
6.6. Test Results.....	30
<b>7. EMISSION LIMITATIONS MEASUREMENT .....</b>	<b>31</b>
7.1. Test Equipment.....	31
7.2. Block Diagram of Test Setup.....	31
7.3. Specification Limits (§15.247(c)).....	31
7.4. Operating Condition of EUT .....	31
7.5. Test Procedure .....	31
7.6. Test Results.....	32
<b>8. BAND EDGES MEASUREMENT .....</b>	<b>41</b>
8.1. Test Equipment.....	41
8.2. Block Diagram of Test Setup.....	41
8.3. Specification Limits [§15.247(c)].....	41
8.4. Operating Condition of EUT .....	41
8.5. Test Procedure .....	41

8.6. Test Results..... 42

**9. POWER SPECTRAL DENSITY MEASUREMENT ..... 43**

9.1. Test Equipment..... 43

9.2. Block Diagram of Test Setup..... 43

9.3. Specification Limits [§15.247(d)]..... 43

9.4. Operating Condition of EUT ..... 43

9.5. Test Procedure ..... 43

9.6. Test Results..... 43

**10. DEVIATION TO TEST SPECIFICATIONS..... 46**

**11. PHOTOGRAPHS..... 47**

11.1. Photos of Conducted Disturbance Measurement..... 47

11.2. Photos of Radiated Measurement at Semi-Anechoic Chamber ..... 48

11.3. Photo of Section RF Conducted Measurement ..... 50

## TEST REPORT CERTIFICATION

Applicant : Health & Life Co., Ltd.  
 Manufacturer : Health & Life Co., Ltd.  
 EUT Description : Automatic Upper Arm Blood Pressure Monitor  
 FCC ID : 2ABTAHNL85CB  
           (A) Model No. : HL858CB  
           (B) Serial No. : N/A  
           (C) Power Supply : DC 6V  
   (Via Batteries or AC Adapter)  
           (D) Test Voltage : (1)DC 6V (Via Batteries)  
   (2)AC 120V, 60Hz (Via AC Adapter)

Measurement Procedure Used:

FCC Rules and Regulations Part 15 Subpart C, Oct. 2013  
 (FCC CFR 47 Part 15C, §15.205, §15.207, §15.209 and §15.247)  
 AND ANSI C63.4:2003

The device described above was tested by AUDIX Technology Corporation to determine the maximum emission levels emanating from the device. The maximum emission levels were compared to the FCC Part 15 subpart C limits.

The measurement results are contained in this test report and AUDIX Technology Corporation is assumed full responsibility for the accuracy and completeness of these measurements. Also, this report shows that the EUT to be technically compliant with the requirements of FCC standards.

This report applies to above tested sample only. This report shall not be reproduced in part without written approval of AUDIX Technology Corporation.

Date of Test: 2014. 09. 11 ~ 17

Date of Report: 2014. 09. 18

Producer:   
 (Annie Yu/Administrator)

Signatory:   
 (Ben Cheng/Manager)

## 1. DESCRIPTION OF REVISION HISTORY

Edition No.	Date of Rev.	Revision Summary	Report No.
0	2014. 09. 18	Original Report	EM-F140556

## 2. GENERAL INFORMATION

### 2.1. Description of Device (EUT)

Product	Automatic Upper Arm Blood Pressure Monitor
Model Number	HL858CB
Serial Number	N/A
Applicant	Health & Life Co., Ltd. 9F, No.186, Jian Yi Road, Zhonghe District, New Taipei City, Taiwan
Manufacturer	Health & Life Co., Ltd. 9F, No.186, Jian Yi Road, Zhonghe District, New Taipei City, Taiwan
FCC ID	2ABTAHNL85CB
Fundamental Range	Bluetooth Low Energy: 2402MHz ~ 2480MHz
Frequency Channel	40 channels
Radio Technology	GFSK
Data Transfer Rate	1Mbps
Antenna Type	PCB Antenna, 4.33dBi(Peak)
AC Adapter	Fuhua, M/N UE08WCP-060100SPA Input: 100-240V~, 50-60Hz, 400mA Output: 6.0V, 1.0A Cord: Non-Shielded, Undetachable, 1.8m
Date of Receipt of Sample	2014. 09. 05
Date of Test	2014. 09. 11 ~ 17

## 2.2. Tested Supporting System Details

### 2.2.1. Support Peripheral Unit

No.	Product	Brand	Model No.	Serial No.	FCC ID
1.	Notebook PC	DELL	P20G	P20G001	N/A
2.	Power Socket	N/A	N/A	N/A	N/A
3.	Test Jig	N/A	N/A	N/A	N/A

### 2.2.2. Cable Lists

No.	Cable Description Of The Above Support Units
1.	USB Cable: Shielded, Detachable, 1.0m Adapter: DELL, M/N AA90PM111 AC Power Code: Non-Shielded, Detachable, 1.8m DC Power Cable: Non-Shielded, Undetachable, 1.8m, Bonded a ferrite core
2.	AC Power Code: Non-Shielded, Detachable, 1.8m
3.	Bus Cable: Non-Shielded, Undetachable, 0.1m

### 2.3. Description of Test Facility

Name of Firm : **AUDIX Technology Corporation**  
 EMC Department  
 No. 53-11, Dingfu, Linkou Dist.,  
 New Taipei City 244, Taiwan

Test Site : **No. 8 Shielded Room &**  
 (C8/Semi-AC) No. 53-11, Dingfu, Linkou Dist.,  
 New Taipei City 244, Taiwan

**Semi-Anechoic Chamber**  
 No. 53-11, Dingfu, Linkou Dist.,  
 New Taipei City 244, Taiwan  
 May 11, 2012 Renewal on  
 Federal Communication Commission  
 Registration Number: 90993

NVLAP Lab. Code : 200077-0

TAF Accreditation No : 1724

### 2.4. Measurement Uncertainty

Test Item	Frequency Range	Uncertainty (dB)
Conduction Test	150kHz~30MHz	±3.43dB
Radiation Test (Distance: 3m)	30MHz~300MHz	± 2.91dB
	300MHz~1000MHz	± 2.74dB
	Above 1GHz	± 5.02dB

Remark : Uncertainty =  $ku_c(y)$

Test Item	Uncertainty
6dB Bandwidth	± 0.05kHz
Maximum peak output power	± 0.33dBm
Emission Limitations	± 0.13dB
Band edges	± 0.13dB
Power spectral density	± 0.13dB




### 3. CONDUCTED EMISSION MEASUREMENT

#### 3.1. Test Equipment

The following test equipment was used during the powerline conducted emission measurement: (No. 8 Shielded Room)

Item	Type	Manufacturer	Model No.	Serial No.	Cal. Due Date
1.	Test Receiver	R&S	ESR3	101774	2015. 02. 18
2.	A.M.N.	R&S	ESH2-Z5	100366	2015. 06. 20
3.	L.I.S.N.	Kyoritsu	KNW-407	8-855-9	2014. 12. 25

#### 3.2. Block Diagram of Test Setup



**EUT: Automatic Upper Arm Blood Pressure Monitor**

#### 3.3. Powerline Conducted Emission Limit (§15.207)

Frequency	Maximum RF Line Voltage	
	Quasi-Peak Level	Average Level
150kHz ~ 500kHz	66 ~ 56 dB $\mu$ V	56 ~ 46 dB $\mu$ V
500kHz ~ 5MHz	56 dB $\mu$ V	46 dB $\mu$ V
5MHz ~ 30MHz	60 dB $\mu$ V	50 dB $\mu$ V

- Remark: 1. If the average limit is met when using a quasi-peak detector, the EUT shall be deemed to meet both limits and measurement with the average detector is unnecessary.  
 2. The lower limit applies at the band edges.

### 3.4. Operating Condition of EUT

- 3.4.1. Set up the EUT and simulator as shown on 3.2.
- 3.4.2. Set to EUT (Automatic Upper Arm Blood Pressure Monitor) on transmitting and receiving during all testing.

### 3.5. Test Procedure

The EUT (link Power Socket) was placed on the table which was above the ground by 80cm and Power Socket's power cord connected to the AC mains through an Artificial Mains Network (A.M.N.). This provided a 50 ohm coupling impedance for the measuring equipment. (Please refer to the block diagram of the test setup and photographs.)

Both sides of A.C. line were checked for maximum conducted interference. In order to find the maximum emission, the relative positions simulators of the interface cables should be manipulated according to ANSI C63.4-2003, regulation during conducted measurement.

The bandwidth of the R&S Test Receiver ESR3 was set at 9kHz.

The frequency range from 150kHz to 30MHz was checked.

All the final readings from Test Receiver were measured with the Quasi-Peak detector and Average detector. Remark: If the Average limit is met when using a Quasi-Peak detector, the Average detector is unnecessary)

### 3.6. Powerline Conducted Emission Measurement Results

**PASSED.** All emissions not reported below are too low against the prescribed limits.

The EUT was measured during this section testing and all the test results are listed in next pages.

EUT : Automatic Upper Arm Blood Pressure Monitor

Model No. : HL858CB

Test Date : 2014. 09. 17      Temperature : 26      Humidity : 64%


The details are as follows :

Mode	Reference Test Data	
	Neutral	Line
1.	# 2	# 1



AUDIX TECHNOLOGY Corp. EMC Department  
 No.53-11, Dingfu, Linkou Dist., New Taipei City  
 24442, Taiwan R.O.C.  
 Tel:+886-2-26092133 Fax:+886-2-26099303  
 Email:emc@audixtech.com

Data: 2 File: D:\test data\REPORT\2014\1M1409XXX\1M1409047-C-D.EM6 (2) Date: 2014-09-17



Site no. : No.8 Shielded Room Data no. : 2  
 Condition : ESH2-Z5 366 Phase : NEUTRAL  
 Limit : FCC 15B-C  
 Env. / Ins. : 26°C / 64% ESR3 (1774) Engineer : John  
 EUT : HL858CB  
 Power Rating : 120Vac/60Hz  
 Test Mode : OPERATING


	AMN	Cable	Pulse	Emission		Limits	Margin	Remark	
Freq. (MHz)	Factor (dB)	Loss (dB)	Att. (dB)	Reading (dBμV)	Level (dBμV)	(dBμV)	(dB)		
1	0.164	0.21	0.02	9.85	19.76	29.84	65.25	35.41	QP
2	0.377	0.23	0.03	9.84	18.35	28.45	58.34	29.89	QP
3	0.564	0.23	0.03	9.86	22.64	32.76	56.00	23.24	QP
4	3.364	0.32	0.07	9.86	8.17	18.42	56.00	37.58	QP
5	11.198	0.52	0.15	9.90	6.32	16.89	60.00	43.11	QP
6	19.532	0.80	0.20	9.93	12.64	23.57	60.00	36.43	QP

Remarks: 1. Emission Level= AMN Factor + Cable Loss + Pulse Att. + Reading.  
 2. If the average limit is met when using a quasi-peak detector, the EUT shall be deemed to meet both limits and measurement with average detector is unnecessary.



AUDIX TECHNOLOGY Corp. EMC Department  
 No.53-11, Dingfu, Linkou Dist., New Taipei City  
 24442, Taiwan R.O.C.  
 Tel:+886-2-26092133 Fax:+886-2-26099303  
 Email:emc@audixtech.com

Data: 1 File: D:\test data\REPORT\2014\1M1409XXX\1M1409047-C-D.EM6 (2) Date: 2014-09-17



Site no. : No.8 Shielded Room Data no. : 1  
 Condition : ESH2-Z5 366 Phase : LINE  
 Limit : FCC 15B-C  
 Env. / Ins. : 26°C / 64% ESR3 (1774) Engineer : John  
 EUT : HL858CB  
 Power Rating : 120Vac/60Hz  
 Test Mode : OPERATING

	AMN	Cable	Pulse	Emission			Margin	Remark
Freq. (MHz)	Factor (dB)	Loss (dB)	Att. (dB)	Reading (dBμV)	Level (dBμV)	Limits (dBμV)	(dB)	
1	0.166	0.18	0.02	9.85	19.49	29.54	35.62	QP
2	0.567	0.20	0.03	9.86	23.64	33.73	22.27	QP
3	0.862	0.21	0.04	9.85	13.34	23.44	32.56	QP
4	1.433	0.23	0.05	9.85	10.94	21.07	34.93	QP
5	2.736	0.26	0.07	9.86	7.35	17.54	38.46	QP
6	18.820	0.67	0.19	9.93	12.38	23.17	36.83	QP

Remarks: 1. Emission Level= AMN Factor + Cable Loss + Pulse Att. + Reading.  
 2. If the average limit is met when using a quasi-peak detector, the EUT shall be deemed to meet both limits and measurement with average detector is unnecessary.

## 4. RADIATED EMISSION MEASUREMENT

### 4.1. Test Equipment

The following test equipment was used during the radiated emission measurement:

#### 4.1.1. For Frequency Range 30MHz~1000MHz (at Semi-Anechoic Chamber)


Item	Type	Manufacturer	Model No.	Serial No.	Cal. Due Date
1	Spectrum Analyzer	Agilent	N9030A-544	US51350140	2015. 07. 24
2	Test Receiver	R & S	ESCS30	100338	2015. 06. 23
3	Amplifier	HP	8447D	2944A06305	2015. 02. 17
4	Bilog Antenna	CHASE	CBL6112D	33821	2015. 08. 01

#### 4.1.2. For Frequency Above 1GHz (at Semi-Anechoic Chamber)


Item	Type	Manufacturer	Model No.	Serial No.	Cal. Due Date
1	Spectrum Analyzer	Agilent	N9030A-544	US51350140	2015. 07. 24
2	Test Receiver	R & S	ESCS30	100338	2015. 06. 23
3	Amplifier	Agilent	8449B	3008A02676	2015. 02. 20
4	2.4GHz Notch Filter	K&L	7NSL10-2441. 5E130.5-00	1	2015. 06. 12
5	3G High Pass Filter	Microwave Circuits	H3G018G1	484796	2015. 06. 12
6	Horn Antenna	EMCO	3115	9609-4927	2015. 06. 16
7	Horn Antenna	EMCO	3116	2653	2014. 10. 10

### 4.2. Test Setup


#### 4.2.1. Block Diagram of connection between EUT and simulators



4.2.2. Semi-Anechoic Chamber (3m) Setup Diagram for 30-1000MHz



4.2.3. Semi-Anechoic Chamber (3m) Setup Diagram for above 1GHz



### 4.3. Radiated Emission Limits (§15.209)

FREQUENCY MHz	DISTANCE Meters	FIELD STRENGTHS LIMITS	
		$\mu\text{V/m}$	$\text{dB}\mu\text{V/m}$
30 ~ 88	3	100	40.0
88 ~ 216	3	150	43.5
216 ~ 960	3	200	46.0
Above 960	3	500	54.0
Above 1000	3	74.0 $\text{dB}\mu\text{V/m}$ (Peak) 54.0 $\text{dB}\mu\text{V/m}$ (Average)	

- Remark :
- (1) Emission level ( $\text{dB}\mu\text{V/m}$ ) =  $20 \log$  Emission level ( $\mu\text{V/m}$ )
  - (2) The tighter limit applies at the edge between two frequency bands.
  - (3) Distance refers to the distance in meters between the measuring instrument antenna and the closed point of any part of the device or system.
  - (4) The limits in this table are based on CFR 47 Part 15.205(a)(b) and Part 15.209 (a).
  - (5) The over 1GHz limit, FCC limit is used based on CFR 47 Part 15.35(b) and Part 15.205(b) & Part 15.209(e) and Part 15.207(c).

### 4.4. Operating Condition of EUT

- 4.4.1. The Automatic Upper Arm Blood Pressure Monitor (EUT) can be operated with battery (DC 6V) or AC adapter (120Vac transfer to DC 6V). We do TX test and associated with battery (DC 6V).
- 4.4.2. Set up the EUT and simulator as shown on 4.2.
- 4.4.3. Set to EUT (Automatic Upper Arm Blood Pressure Monitor) on transmitting and receiving during all testing.

#### 4.5. Test Procedure

The EUT and its simulators were placed on a turn table which was 0.8 meter above the ground. The turn table rotated 360 degrees to determine the position of the maximum emission level. EUT was set 3 meters away from the receiving antenna which was mounted on an antenna tower. The antenna moved up and down between 1 to 4 meters to find out the maximum emission level. Broadband antenna such as bilog antenna or horn antenna were used as a receiving antenna. Both horizontal and vertical polarization of the antenna were set on measurement. In order to find the maximum emission, all of the interface cables were manipulated according to ANSI C63.4-2003 regulation.

The bandwidth of the R&S Test Receiver ESCS30 was set at 120kHz. (For 30MHz to 1000MHz)

The resolution bandwidth and video bandwidth of test spectrum analyzer is 1MHz for peak detection (PK) at frequency above 1GHz.

The resolution bandwidth of test spectrum analyzer is 1MHz and the video bandwidth is 10Hz for average detection (AV) at frequency above 1GHz.

The frequency range from 30MHz to 25GHz (Up to 10<sup>th</sup> harmonics from fundamental frequency) was checked. 30MHz to 1000MHz was measured with Quasi-Peak detector.

Pursuant to ANSI C63.4 8.3.1.2, when peak value complies with the average limit, we didn't perform measurement in average detector.



#### 4.6. Test Results

##### **PASSED.**

(All emissions not reported for there is no emission be found.)

EUT: Automatic Upper Arm Blood Pressure Monitor M/N: HL858CB

Test Date: 2014. 09. 17 Temperature: 23 Humidity: 42%

##### **For Frequency Range 30MHz~1000MHz:**

The EUT with following test modes was performed during this section testing and all the test results are listed in section 4.6.1.

Mode	Test Voltage	Channel	Frequency	Test Mode	Reference Test Data	
					Horizontal	Vertical
1.	DC 6V (Via Batteries)	CH 0	2402MHz	Transmit	# 1	# 2
2.		CH 19	2440MHz		# 2	# 1
3.		CH 39	2480MHz		# 1	# 2

\* Above all final readings were measured with Quasi-Peak detector.

##### **For Frequency above 1GHz:**

The emissions (up to 25GHz) not reported are too low to be measured.

##### **For Restricted Bands:**

The EUT was tested in restricted bands and all the test results are listed in section 4.6.2. (The restricted bands defined in part 15.205(a))

Mode	Test Voltage	Channel	Frequency	Test Mode	Reference Test Data No.	
					Horizontal	Vertical
1	DC 6V (Via Batteries)	CH 0	2402MHz	Transmit	# 3, # 4	# 1, # 2
2		CH 39	2480MHz		# 5, # 6	# 7, # 8

4.6.1. For 30-1000MHz Frequency Range Measurement Results

**Bluetooth Low Energy, Transmit, Frequency: 2402MHz**

Site no. : Audix NO.1 Chamber Data no. : 1  
 Dis. / Ant. : 3m CBL6112D 33821 Ant. pol. : HORIZONTAL  
 Limit : 30M-1G  
 Env. / Ins. : 23°C/42% N9030A(140) Engineer : Ken\_chen  
 EUT : HL858CB  
 Power Rating : DC 6V  
 Test Mode : Tx2402

	Freq. (MHz)	Ant. Factor (dB/m)	Cable Loss (dB)	Reading (dBµV)	Emission Level (dBµV/m)	Limits (dBµV/m)	Margin (dB)	Remark
1	93.05	9.67	3.17	8.53	21.37	43.50	22.13	Peak
2	101.78	11.03	3.23	10.44	24.70	43.50	18.80	Peak

Remarks: 1. Emission Level= Antenna Factor + Cable Loss + Reading.  
 2. The emission levels that are 20dB below the official limit are not reported.

Site no. : Audix NO.1 Chamber Data no. : 2  
 Dis. / Ant. : 3m CBL6112D 33821 Ant. pol. : VERTICAL  
 Limit : 30M-1G  
 Env. / Ins. : 23°C/42% N9030A(140) Engineer : Ken\_chen  
 EUT : HL858CB  
 Power Rating : DC 6V  
 Test Mode : Tx2402

	Freq. (MHz)	Ant. Factor (dB/m)	Cable Loss (dB)	Reading (dBµV)	Emission Level (dBµV/m)	Limits (dBµV/m)	Margin (dB)	Remark
1	92.08	9.49	3.17	6.15	18.81	43.50	24.69	
2	97.90	10.53	3.21	5.42	19.16	43.50	24.34	
3	101.78	11.03	3.23	7.95	22.21	43.50	21.29	

Remarks: 1. Emission Level= Antenna Factor + Cable Loss + Reading.  
 2. The emission levels that are 20dB below the official limit are not reported.

**Bluetooth Low Energy, Transmit, Frequency: 2440MHz**

Site no. : Audix NO.1 Chamber Data no. : 2  
 Dis. / Ant. : 3m CBL6112D 33821 Ant. pol. : HORIZONTAL  
 Limit : 30M-1G  
 Env. / Ins. : 23°C/42% N9030A(140) Engineer : Ken\_chen  
 EUT : HL858CB  
 Power Rating : DC 6V  
 Test Mode : Tx2440

	Freq. (MHz)	Ant. Factor (dB/m)	Cable Loss (dB)	Reading (dB $\mu$ V)	Emission Level (dB $\mu$ V/m)	Limits (dB $\mu$ V/m)	Margin (dB)	Remark
1	93.05	9.67	3.17	8.24	21.08	43.50	22.42	Peak
2	101.78	11.03	3.23	10.38	24.64	43.50	18.86	Peak
3	116.33	12.06	3.34	9.22	24.62	43.50	18.88	Peak

Remarks: 1. Emission Level= Antenna Factor + Cable Loss + Reading.  
 2. The emission levels that are 20dB below the official limit are not reported.

Site no. : Audix NO.1 Chamber Data no. : 1  
 Dis. / Ant. : 3m CBL6112D 33821 Ant. pol. : VERTICAL  
 Limit : 30M-1G  
 Env. / Ins. : 23°C/42% N9030A(140) Engineer : Ken\_chen  
 EUT : HL858CB  
 Power Rating : DC 6V  
 Test Mode : Tx2440

	Freq. (MHz)	Ant. Factor (dB/m)	Cable Loss (dB)	Reading (dB $\mu$ V)	Emission Level (dB $\mu$ V/m)	Limits (dB $\mu$ V/m)	Margin (dB)	Remark
1	92.08	9.49	3.17	5.69	18.35	43.50	25.15	Peak
2	100.81	10.98	3.23	4.20	18.41	43.50	25.09	Peak
3	101.78	11.03	3.23	6.73	20.99	43.50	22.51	Peak
4	111.48	11.73	3.31	3.32	18.36	43.50	25.14	Peak

Remarks: 1. Emission Level= Antenna Factor + Cable Loss + Reading.  
 2. The emission levels that are 20dB below the official limit are not reported.

**Bluetooth Low Energy, Transmit, Frequency: 2480MHz**

Site no. : Audix NO.1 Chamber Data no. : 1  
 Dis. / Ant. : 3m CBL6112D 33821 Ant. pol. : HORIZONTAL  
 Limit : 30M-1G  
 Env. / Ins. : 23\*C/42% N9030A(140) Engineer : Ken\_chen  
 EUT : HL858CB  
 Power Rating : DC 6V  
 Test Mode : Tx2480

	Freq. (MHz)	Ant. Factor (dB/m)	Cable Loss (dB)	Reading (dBμV)	Emission Level (dBμV/m)	Limits (dBμV/m)	Margin (dB)	Remark
1	93.05	9.67	3.17	7.98	20.82	43.50	22.68	Peak
2	101.78	11.03	3.23	10.30	24.56	43.50	18.94	Peak

Remarks: 1. Emission Level= Antenna Factor + Cable Loss + Reading.  
 2. The emission levels that are 20dB below the official limit are not reported.

Site no. : Audix NO.1 Chamber Data no. : 2  
 Dis. / Ant. : 3m CBL6112D 33821 Ant. pol. : VERTICAL  
 Limit : 30M-1G  
 Env. / Ins. : 23\*C/42% N9030A(140) Engineer : Ken\_chen  
 EUT : HL858CB  
 Power Rating : DC 6V  
 Test Mode : Tx2480

	Freq. (MHz)	Ant. Factor (dB/m)	Cable Loss (dB)	Reading (dBμV)	Emission Level (dBμV/m)	Limits (dBμV/m)	Margin (dB)	Remark
1	92.08	9.49	3.17	6.22	18.88	43.50	24.62	Peak
2	97.90	10.53	3.21	5.49	19.23	43.50	24.27	Peak
3	101.78	11.03	3.23	7.16	21.42	43.50	22.08	Peak
4	147.37	10.72	3.58	8.74	23.04	43.50	20.46	Peak

Remarks: 1. Emission Level= Antenna Factor + Cable Loss + Reading.  
 2. The emission levels that are 20dB below the official limit are not reported.


4.6.2. Restricted Bands Measurement Results

Date of Test : 2014. 09. 17 Temperature : 23

EUT : Automatic Upper Arm Blood Pressure Monitor Humidity : 42%

Test Mode : Bluetooth Low Energy, Transmit, Channel 0, Frequency: 2402MHz

Data: 3 File: C:\Users\laudix\Desktop\未完成\IC1M1409047\IC1M1409047(FCC 捷典\ofb)\Out of bar




Site no. : Audix NO.1 Chamber Data no. : 3  
 Dis. / Ant. : 3m 3115(4927) Ant. pol. : HORIZONTAL  
 Limit : ABOVE 1GHZ(PK)  
 Env. / Ins. : 23°C/42% N9030A(140) Engineer : ken\_chen  
 EUT : HL853CB  
 Power Rating : DC 6V  
 Test Mode : Tx2402

Peak	Freq. (MHz)	Ant. Factor (dB/m)	Cable Loss (dB)	Reading (dBµV)	Emission Level (dBµV/m)	Limits (dBµV/m)	Margin (dB)	Remark
1	2388.84	28.20	5.24	10.35	43.79	74.00	30.21	Peak
2	2390.04	28.20	5.24	10.85	44.29	74.00	29.71	Peak
3	2402.28	28.21	5.28	45.78	79.25	74.00	-5.25	Peak

Remarks: 1. Emission Level= Antenna Factor + Cable Loss + Reading.  
 2. The emission levels that are 20dB below the official limit are not reported.

Data: 4 File: C:\Users\laudix\Desktop\未完成\IC1M1409047\IC1M1409047(FCC 捷典\ofb)\Out of bar



Site no. : Audix NO.1 Chamber Data no. : 4  
 Dis. / Ant. : 3m 3115(4927) Ant. pol. : HORIZONTAL  
 Limit : ABOVE 1GHZ(AV)  
 Env. / Ins. : 23°C/42% N9030A(140) Engineer : ken\_chen  
 EUT : HL853CB  
 Power Rating : DC 6V  
 Test Mode : Tx2402

Peak	Freq. (MHz)	Ant. Factor (dB/m)	Cable Loss (dB)	Reading (dBµV)	Emission Level (dBµV/m)	Limits (dBµV/m)	Margin (dB)	Remark
1	2388.84	28.20	5.24	-0.36	33.08	54.00	20.92	Average
2	2390.04	28.20	5.24	0.12	33.56	54.00	20.44	Average
3	2402.16	28.21	5.28	44.14	77.81	54.00	-23.81	Average


Remarks: 1. Emission Level= Antenna Factor + Cable Loss + Reading.  
 2. The emission levels that are 20dB below the official limit are not reported.

Date of Test : 2014. 09. 17 Temperature : 23

EUT : Automatic Upper Arm Blood Pressure Monitor Humidity : 42%

Test Mode : Bluetooth Low Energy, Transmit, Channel 0, Frequency: 2402MHz

Data: 1 File: C:\Users\audix\Desktop\未完成\1M1409047\1M1409047(FCC 捷典\ofb)\Out of bar




Site no. : Audix NO.1 Chamber Data no. : 1  
 Dis. / Ant. : 3m 3115(4927) Ant. pol. : VERTICAL  
 Limit : ABOVE 1GHZ(PK)  
 Env. / Ins. : 23°C/42% N9030A(140) Engineer : ken\_chen  
 EUT : HL858CB  
 Power Rating : DC 6W  
 Test Mode : Tx2402

	Freq. (MHz)	Ant. Factor (dB/m)	Cable Loss (dB)	Reading (dBμV)	Emission Level (dBμV/m)	Limits (dBμV/m)	Margin (dB)	Remark
1	2388.84	28.20	5.24	12.42	45.86	74.00	28.14	Peak
2	2390.04	28.20	5.24	12.87	46.31	74.00	27.69	Peak
3	2402.28	28.21	5.26	50.55	84.02	74.00	-10.02	Peak

Remarks: 1. Emission Level= Antenna Factor + Cable Loss + Reading.  
 2. The emission levels that are 20dB below the official limit are not reported.

Data: 2 File: C:\Users\audix\Desktop\未完成\1M1409047\1M1409047(FCC 捷典\ofb)\Out of bar



Site no. : Audix NO.1 Chamber Data no. : 2  
 Dis. / Ant. : 3m 3115(4927) Ant. pol. : VERTICAL  
 Limit : ABOVE 1GHZ(AV)  
 Env. / Ins. : 23°C/42% N9030A(140) Engineer : ken\_chen  
 EUT : HL858CB  
 Power Rating : DC 6W  
 Test Mode : Tx2402

	Freq. (MHz)	Ant. Factor (dB/m)	Cable Loss (dB)	Reading (dBμV)	Emission Level (dBμV/m)	Limits (dBμV/m)	Margin (dB)	Remark
1	2388.84	28.20	5.24	0.80	34.24	54.00	19.76	Average
2	2390.04	28.20	5.24	1.59	35.03	54.00	18.97	Average
3	2402.04	28.21	5.26	48.57	82.04	54.00	-28.04	Average


Remarks: 1. Emission Level= Antenna Factor + Cable Loss + Reading.  
 2. The emission levels that are 20dB below the official limit are not reported.

Date of Test : 2014. 09. 17 Temperature : 23

EUT : Automatic Upper Arm Blood Pressure Monitor Humidity : 42%

Test Mode : Bluetooth Low Energy, Transmit, Channel 39, Frequency: 2480MHz

Data: 5 File: C:\Users\audix\Desktop\未完成\C1M1409047\C1M1409047(FCC 捷典\ofb)Out of bar




Site no. : Audix NO.1 Chamber Data no. : 5  
 Dis. / Ant. : 3m 3115(4927) Ant. pol. : HORIZONTAL  
 Limit : ABOVE 1GHZ(PK)  
 Env. / Ins. : 23°C/42% N9030A(140) Engineer : ken\_chen  
 EUT : HL858CB  
 Power Rating : DC 8V  
 Test Mode : Tx2480

	Freq. (MHz)	Ant. Factor (dB/m)	Cable Loss (dB)	Reading (dBμV)	Emission Level (dBμV/m)	Limits (dBμV/m)	Margin (dB)	Remark
1	2479.80	28.28	5.36	43.62	77.28	74.00	-3.28	Peak
2	2483.50	28.29	5.37	14.54	48.20	74.00	25.80	Peak
3	2484.50	28.29	5.37	12.38	48.04	74.00	27.96	Peak

Remarks: 1. Emission Level= Antenna Factor + Cable Loss + Reading.  
 2. The emission levels that are 20dB below the official limit are not reported.

Data: 6 File: C:\Users\audix\Desktop\未完成\C1M1409047\C1M1409047(FCC 捷典\ofb)Out of bar



Site no. : Audix NO.1 Chamber Data no. : 6  
 Dis. / Ant. : 3m 3115(4927) Ant. pol. : HORIZONTAL  
 Limit : ABOVE 1GHZ(AV)  
 Env. / Ins. : 23°C/42% N9030A(140) Engineer : ken\_chen  
 EUT : HL858CB  
 Power Rating : DC 8V  
 Test Mode : Tx2480

	Freq. (MHz)	Ant. Factor (dB/m)	Cable Loss (dB)	Reading (dBμV)	Emission Level (dBμV/m)	Limits (dBμV/m)	Margin (dB)	Remark
1	2480.10	28.28	5.36	41.50	75.14	54.00	-21.14	Average
2	2483.50	28.29	5.37	3.78	37.44	54.00	16.56	Average
3	2484.50	28.29	5.37	1.31	34.97	54.00	19.03	Average


Remarks: 1. Emission Level= Antenna Factor + Cable Loss + Reading.  
 2. The emission levels that are 20dB below the official limit are not reported.

Date of Test : 2014. 09. 17 Temperature : 23

EUT : Automatic Upper Arm Blood Pressure Monitor Humidity : 42%

Test Mode : Bluetooth Low Energy, Transmit, Channel 39, Frequency: 2480MHz

Data: 7 File: C:\Users\audix\Desktop\未完成\IC1M1409047\IC1M1409047(FCC 捷興\ofb)\Out of bar




Site no. : Audix NO.1 Chamber Data no. : 7  
 Dis. / Ant. : 3m 3115(4927) Ant. pol. : VERTICAL  
 Limit : ABOVE 1GHZ(PK)  
 Env. / Ins. : 23°C/42% N9030A(140) Engineer : ken\_chen  
 EUT : HL858CB  
 Power Rating : DC 8W  
 Test Mode : Tx2480

	Freq. (MHz)	Ant. Factor (dB/m)	Cable Loss (dB)	Reading (dBμV)	Emission Level (dBμV/m)	Limit (dBμV/m)	Margin (dB)	Remark
1	2479.80	28.28	5.36	45.04	78.68	74.00	-4.68	Peak
2	2483.50	28.29	5.37	14.47	48.13	74.00	25.87	Peak
3	2484.50	28.29	5.37	14.00	47.68	74.00	26.34	Peak

Remarks: 1. Emission Level= Antenna Factor + Cable Loss + Reading.  
 2. The emission levels that are 20dB below the official limit are not reported.

Data: 8 File: C:\Users\audix\Desktop\未完成\IC1M1409047\IC1M1409047(FCC 捷興\ofb)\Out of bar



Site no. : Audix NO.1 Chamber Data no. : 8  
 Dis. / Ant. : 3m 3115(4927) Ant. pol. : VERTICAL  
 Limit : ABOVE 1GHZ(AV)  
 Env. / Ins. : 23°C/42% N9030A(140) Engineer : ken\_chen  
 EUT : HL858CB  
 Power Rating : DC 8W  
 Test Mode : Tx2480

	Freq. (MHz)	Ant. Factor (dB/m)	Cable Loss (dB)	Reading (dBμV)	Emission Level (dBμV/m)	Limit (dBμV/m)	Margin (dB)	Remark
1	2480.10	28.28	5.36	43.31	78.95	54.00	-22.95	Average
2	2483.50	28.29	5.37	4.89	38.55	54.00	15.45	Average
3	2484.50	28.29	5.37	1.75	35.41	54.00	18.59	Average

Remarks: 1. Emission Level= Antenna Factor + Cable Loss + Reading.  
 2. The emission levels that are 20dB below the official limit are not reported.




## 5. 6dB BANDWIDTH MEASUREMENT

### 5.1. Test Equipment

The following test equipment was used during the Emission Bandwidth measurement:

Item	Type	Manufacturer	Model No.	Serial No.	Cal. Due Date
1	Spectrum Analyzer	Agilent	N9030A-526	MY53310269	2014. 09. 22

### 5.2. Block Diagram of Test Setup



### 5.3. Specification Limits [§15.247(a)(2)]

The minimum 6dB bandwidth shall be at least 500kHz.

### 5.4. Operating Condition of EUT

The Notebook PC was running test program “SMARTRF Studio 7” to enable the EUT to transmit data at different channel frequency individually.

### 5.5. Test Procedure

The transmitter output was connected to the spectrum analyzer. The bandwidth of the fundamental frequency was measure by spectrum analyzer with 1.5% EBW,  $VBW \geq 3 \times RBW$ . The 6dB bandwidth is defined as the total spectrum the power of which is higher than peak power minus 6dB.

The measurement guideline was according to KDB 558074 D01 DTS Meas Guidance is v03r01.

## 5.6. Test Results


**PASSED.** All the test results are attached in next pages.

Test Date: 2014. 09. 11    Temperature: 25    Humidity: 43%


Mode	Type of Network	Channel	Frequency	<b>6dB Bandwidth</b>
1	Bluetooth Low Energy	CH0	2402MHz	<b>0.688 MHz</b>
2		CH19	2440MHz	<b>0.673 MHz</b>
3		CH39	2480MHz	<b>0.721 MHz</b>

**[Limit: least 500kHz]**


### Bluetooth Low Energy, Frequency: 2402MHz



### Bluetooth Low Energy, Frequency: 2440MHz



### Bluetooth Low Energy, Frequency: 2480MHz




## 6. MAXIMUM PEAK OUTPUT POWER MEASUREMENT

### 6.1. Test Equipment

The following test equipment was used during the maximum peak output power measurement:

Item	Type	Manufacturer	Model No.	Serial No.	Cal. Due Date
1.	Power Meter	Anritsu	ML2495A	1145008	2014. 10. 22
2.	Power Sensor	Anritsu	MA2411B	1126096	2014. 10. 22

### 6.2. Block Diagram of Test Setup



### 6.3. Specification Limits [§15.247(b)-(3)]

The Limits of maximum Peak Output Power for digital modulation in 2400-2483.5MHz is: 1Watt. (30dBm)

### 6.4. Operating Condition of EUT

The Notebook PC was running test program “SMARTRF Studio 7” to enable the EUT to transmit data at different channel frequency individually.

### 6.5. Test Procedure

The transmitter output was connected to the power sensor and record the reading of power meter.

The measurement guideline was according to KDB 558074 D01 DTS Meas Guidance is v03r01

## 6.6. Test Results

**PASSED.** All the test results are listed below.

Test Date: 2014. 09. 11    Temperature: 25    Humidity: 43%

Mode	Type of Network	Channel	Test Frequency	Output Power(dBm)	
				Peak	Average
1	Bluetooth Low Energy	CH0	2402MHz	<b>-1.33</b>	<b>-2.64</b>
2		CH19	2440MHz	<b>-0.35</b>	<b>-1.46</b>
3		CH39	2480MHz	<b>-0.52</b>	<b>-0.35</b>

**[Limit: 1Watt. (30dBm)]**

## 7. EMISSION LIMITATIONS MEASUREMENT

### 7.1. Test Equipment

The following test equipment was used during the emission limitations test :

Item	Type	Manufacturer	Model No.	Serial No.	Cal. Due Date
1.	Spectrum Analyzer	Agilent	N9030A-526	MY53310269	2014. 09. 22

### 7.2. Block Diagram of Test Setup

The same as section.5.2

### 7.3. Specification Limits (§15.247(c))

7.3.1. In any 100kHz bandwidth outside the frequency band in which the spread spectrum or digitally modulated intentional radiator is operating, the radio frequency power that is produced by the intentional radiator shall be at least 20dB below that in the 100kHz bandwidth within the band that contains the highest level of the desired power, based on either an RF conducted or a radiated measurement. Attenuation below the general limits specified in Section 15.209(a) is not required. In addition, radiated emissions which fall in restricted bands, as defined in Section 15.205(a), must also comply with the radiated emission limits specified in Section 15.209(a) (See Section 15.205(c)).( This test result attaching to §4.6.3)

7.3.2. The reference level for determining limit of emission limitations is according to the value measured indicated in plots at section 9.6.

### 7.4. Operating Condition of EUT

The Notebook PC was running test program “SMARTRF Studio 7” to enable the EUT to transmit data at different channel frequency individually.

### 7.5. Test Procedure

The RF output of EUT was connected to the spectrum analyzer. The bandwidth of the fundamental frequency was measure by spectrum analyzer with 100kHz RBW and 300kHz VBW.



The measurement guideline was according to KDB 558074 D01 DTS Meas Guidance is v03r01.

7.6. Test Results



**PASSED.** The testing data was attached in the next pages.



Test Date: 2014. 09. 11    Temperature: 25    Humidity: 43%

**Test Frequency: CH 0, 2402MHz**

















Test Frequency: CH 19, 2440MHz











Test Frequency: CH 39, 2480MHz











## 8. BAND EDGES MEASUREMENT

### 8.1. Test Equipment

The following test equipment was used during the band edges measurement:

Item	Type	Manufacturer	Model No.	Serial No.	Cal. Due Date
1	Spectrum Analyzer	Agilent	N9030A-526	MY53310269	2014. 09. 22

### 8.2. Block Diagram of Test Setup

The same as section.5.2.

### 8.3. Specification Limits [§15.247(c)]

8.3.1. In any 100kHz bandwidth outside the frequency band in which the spread spectrum or digitally modulated intentional radiator is operating, the radio frequency power that is produced by the intentional radiator shall be at least 20dB below that in the 100kHz bandwidth within the band that contains the highest level of the desired power, based on either an RF conducted or a radiated measurement. Attenuation below the general limits specified in Section 15.209(a) is not required. In addition, radiated emissions which fall in restricted bands, as defined in Section 15.205(a), must also comply with the radiated emission limits specified in Section 15.209(a) (See Section 15.205(c)).( This test result attaching to §4.6.3)

8.3.2. The reference level for determining limit of emission limitations is according to the value measured indicated in plots at section 9.6.

### 8.4. Operating Condition of EUT

The Notebook PC was running test program “SMARTRF Studio 7” to enable the EUT to transmit data at different channel frequency individually.

### 8.5. Test Procedure

The transmitter output was connected to the spectrum analyzer. Set both RBW=100 kHz and VBW to 300kHz with suitable frequency span including 100kHz bandwidth from band edge.


The measurement guideline was according to KDB 558074 D01 DTS Meas Guidance is v03r01.

### 8.6. Test Results


**PASSED.** All the test results are attached in next pages.

Test Date: 2014. 09. 17    Temperature: 25    Humidity: 45%

#### Bluetooth Low Energy, Upper Band edge



#### Below Band edge



## 9. POWER SPECTRAL DENSITY MEASUREMENT

### 9.1. Test Equipment

The following test equipment was used during the power spectral density measurement:

Item	Type	Manufacturer	Model No.	Serial No.	Cal. Due Date
1	Spectrum Analyzer	Agilent	N9030A-526	MY53310269	2014. 09. 22

### 9.2. Block Diagram of Test Setup

The same as section.5.2.

### 9.3. Specification Limits [§15.247(d)]

The peak power spectral density conducted from the intentional radiator to the antenna shall not be greater than 8dBm in any 3kHz band.

### 9.4. Operating Condition of EUT

The Notebook PC was running test program “SMARTRF Studio 7” to enable the EUT to transmit data at different channel frequency individually.

### 9.5. Test Procedure

The transmitter output was connected to the spectrum analyzer. The bandwidth of the fundamental frequency was measured with the spectrum analyzer using 100kHz RBW and  $\geq 300$ kHz VBW, set sweep time = Auto.

The measurement guideline was according to KDB 558074 D01 DTS Meas Guidance is v03r01.

### 9.6. Test Results


**PASSED.** All the test results are attached in next pages.

Test Date: 2014. 04. 11    Temperature: 25    Humidity: 43%


Mode	Type of Network	Channel	Frequency	Power Spectral Density
1	Bluetooth Low Energy	CH0	2402MHz	<b>-0.942 dBm</b>
2		CH19	2440MHz	<b>-1.808 dBm</b>
3		CH39	2480MHz	<b>-3.100 dBm</b>

[Limit: 8dBm]


### Bluetooth Low Energy, Frequency: 2402MHz



### Bluetooth Low Energy, Frequency: 2440MHz



### Bluetooth Low Energy, Frequency: 2480MHz



## **10.DEVIATION TO TEST SPECIFICATIONS**

**【NONE】**

## 11. PHOTOGRAPHS

### 11.1. Photos of Conducted Disturbance Measurement



FRONT VIEW OF CONDUCTED MEASUREMENT



BACK VIEW OF CONDUCTED MEASUREMENT

## 11.2. Photos of Radiated Measurement at Semi-Anechoic Chamber

### 11.2.1. Frequency Range 30MHz~1GHz

Operated with batteries



Operated with AC adapter





11.2.2.Frequency Range Above 1GHz

Operated with batteries



Operated with AC adapter



### 11.3. Photo of Section RF Conducted Measurement

