



FCC PART 15B

TEST REPORT

For

BMC Medical Co., Ltd.

5/F Main Building, No.19 Gucheng Street West, Shijingshan, Beijing 100043, P.R.China

FCC ID: 2ABWVYH600AYH600B

Report Type: Original Report	Product Type: Portable Diagnostic System
Test Engineer: <u>Ares Liu</u>	
Report Number: <u>R2BJ140225050-00B</u>	
Report Date: <u>2014-04-29</u> <u>Sula Huang</u>	
Reviewed By: <u>RF Engineer</u>	
Test Laboratory: Bay Area Compliance Laboratories Corp. (Dongguan) No.69 Pulongcun, Puxinhu Industrial Zone, Tangxia, Dongguan, Guangdong, China Tel: +86-769-86858888 Fax: +86-769-86858891 www.baclcorp.com.cn	

Note: This test report is prepared for the customer shown above and for the device described herein. It may not be duplicated or used in part without prior written consent from Bay Area Compliance Laboratories Corp. (Dongguan). This report is valid only with a valid digital signature. The digital signature may be available only under the Adobe software above version 7.0.

TABLE OF CONTENTS

GENERAL INFORMATION.....	3
PRODUCT DESCRIPTION FOR EQUIPMENT UNDER TEST (EUT)	3
OBJECTIVE	3
RELATED SUBMITTAL(S)/GRANT(S).....	3
TEST FACILITY	3
SYSTEM TEST CONFIGURATION.....	4
JUSTIFICATION	4
EUT EXERCISE SOFTWARE	4
EQUIPMENT MODIFICATIONS	4
SUPPORT EQUIPMENT LIST AND DETAILS	4
EXTERNAL I/O CABLE.....	4
BLOCK DIAGRAM OF TEST SETUP	5
SUMMARY OF TEST RESULTS	6
FCC §15.107 – AC LINE CONDUCTED EMISSIONS.....	7
MEASUREMENT UNCERTAINTY	7
EUT SETUP	7
EMI TEST RECEIVER SETUP.....	8
TEST PROCEDURE	8
CORRECTED AMPLITUDE & MARGIN CALCULATION	8
TEST EQUIPMENT LIST AND DETAILS.....	9
TEST RESULTS SUMMARY	9
TEST DATA	9
FCC §15.109 - RADIATED EMISSIONS	12
MEASUREMENT UNCERTAINTY	12
EUT SETUP	12
EMI TEST RECEIVER SETUP.....	13
TEST PROCEDURE	13
CORRECTED AMPLITUDE & MARGIN CALCULATION	14
TEST EQUIPMENT LIST AND DETAILS.....	14
TEST RESULTS SUMMARY	14
TEST DATA	14

GENERAL INFORMATION

Product Description for Equipment under Test (EUT)

The *BMC Medical Co., Ltd.*'s product, model number: *YH-600B Pro (FCC ID: 2ABWVYH600AYH600B)* (the "EUT") in this report was a *Portable Diagnostic System*, which was measured approximately: 8.0 cm (L) x 6.0cm (W) x 2.5 cm (H), rated input voltage: 1 × 1.5V AA alkaline battery. The highest operating frequency is 1.0 GHz.

* All measurement and test data in this report was gathered from production sample serial number: 140225050 (Assigned by BACL.Dongguan). The EUT was received on 2014-02-26.

Objective

This report is prepared on behalf of *BMC Medical Co., Ltd.* in accordance with Part 2, Subpart J, Part 15, Subparts A and B of the Federal Communication Commissions rules.

The objective of the manufacturer is to determine compliance with FCC Part 15B, Class B.

Related Submittal(s)/Grant(s)

FCC Part 15C DTS submissions with FCC ID: 2ABWVYH600AYH600B

Test Facility

The Test site used by Bay Area Compliance Laboratories Corp. (Dongguan) to collect test data is located on the No.69 Pulongcun, Puxinhu Industrial Zone, Tangxia, Dongguan, Guangdong, China

Test site at Bay Area Compliance Laboratories Corp. (Dongguan) has been fully described in reports submitted to the Federal Communications Commission (FCC). The details of these reports have been found to be in compliance with the requirements of Section 2.948 of the FCC Rules on February 02, 2012. The facility also complies with the radiated and AC line conducted test site criteria set forth in ANSI C63.4-2003.

The Federal Communications Commission has the reports on file and is listed under FCC Registration No.: 273710. The test site has been approved by the FCC for public use and is listed in the FCC Public Access Link (PAL) database.

SYSTEM TEST CONFIGURATION

Justification

The system was configured for testing in a typical fashion (as normally used by a typical user).

EUT operation mode: Downloading (data transfer with computer)

EUT Exercise Software

“BurnIn test v5.3” exercise software was used.

Equipment Modifications

No modification was made to the EUT.

Support Equipment List and Details


Manufacturer	Description	Model	Serial Number
DELL	Laptop	PP11L	QDS-BRCM1017
HP	Printer	C3941A	JPTVOB2337
DELL	Keyboard	L100	CNORH656658907BL05DC
AST	Modem	AEM-2100	0293
Kingston	Micro SD Card	4GB	/

External I/O Cable

Cable Description	Shielding Type	Ferrite Core	Length (m)	From Port	To
Parallel Cable	yes	No	1.2	Parallel Port of Laptop	Printer
USB Cable	yes	No	1.5	USB Port of Laptop	Keyboard
Serial Cable	yes	No	1.5	RS232 Port of Laptop	Modem
USB Cable	yes	No	1.2	USB Port of Laptop	EUT

Block Diagram of Test Setup

Downloading:



SUMMARY OF TEST RESULTS

FCC Rules	Description of Test	Results
§15.107	AC Line Conducted Emissions	Compliance
§15.109	Radiated Emissions	Compliance

FINAL

FCC §15.107 – AC LINE CONDUCTED EMISSIONS

Measurement Uncertainty

Compliance or non- compliance with a disturbance limit shall be determined in the following manner:

If U_{lab} is less than or equal to $U_{\text{cisp}}_{\text{pr}}$ of Table 1, then:

- compliance is deemed to occur if no measured disturbance level exceeds the disturbance limit;
- non - compliance is deemed to occur if any measured disturbance level exceeds the disturbance limit.

If U_{lab} is greater than $U_{\text{cisp}}_{\text{pr}}$ of Table 1, then:


- compliance is deemed to occur if no measured disturbance level, increased by $(U_{\text{lab}} - U_{\text{cisp}}_{\text{pr}})$, exceeds the disturbance limit;
- non - compliance is deemed to occur if any measured disturbance level, increased by $(U_{\text{lab}} - U_{\text{cisp}}_{\text{pr}})$, exceeds the disturbance limit.

Based on CISPR 16-4-2: 2011, measurement uncertainty of conducted disturbance at mains port using AMN at Bay Area Compliance Laboratories Corp. (Dongguan) is 3.46 dB (150 kHz to 30 MHz).

Table 1 – Values of $U_{\text{cisp}}_{\text{pr}}$

Measurement	$U_{\text{cisp}}_{\text{pr}}$
Conducted disturbance at mains port using AMN (150 kHz to 30 MHz)	3.4 dB

EUT Setup



Note:

1. Support units were connected to second LISN.
2. Both of LISNs (AMN) 80 cm from EUT and at the least 80 cm from other units and other metal planes support units.

The setup of EUT is according with per ANSI C63.4-2003 measurement procedure. The specification used was with the FCC Part 15.107 Class B limits.

The spacing between the peripherals was 10 cm.

The adapter of laptop was connected to a 120 VAC/60 Hz power source

EMI Test Receiver Setup

The EMI test receiver was set to investigate the spectrum from 150 kHz to 30 MHz.

During the conducted emission test, the EMI test receiver was set with the following configurations:

Frequency Range	IF B/W
150 kHz – 30 MHz	9 kHz

Test Procedure

During the conducted emission test, the adapter was connected to the outlet of the first LISN and the other support equipments were connected to the outlet of the second LISN.

Maximizing procedure was performed on the six (6) highest emissions of the EUT.

All data was recorded in the Quasi-peak and average detection mode.

Corrected Amplitude & Margin Calculation

The basic equation is as follows:

$$V_C = V_R + A_C + VDF$$
$$C_f = A_C + VDF$$

Herein,

V_C (cord. Reading): corrected voltage amplitude

V_R : reading voltage amplitude

A_C : attenuation caused by cable loss

VDF: voltage division factor of AMN

C_f : Correction Factor

The “Margin” column of the following data tables indicates the degree of compliance within the applicable limit. For example, a margin of 7dB means the emission is 7dB below the maximum limit. The equation for margin calculation is as follows:

$$\text{Margin} = \text{Limit} - \text{Corrected Amplitude}$$

Test Equipment List and Details

Manufacturer	Description	Model	Serial Number	Calibration Date	Calibration Due Date
R&S	EMI Test Receiver	ESCS 30	830245/006	2013-11-20	2014-11-19
R&S	Two-line V-network	ENV216	3560.6550.12	2014-01-22	2015-01-21
R&S	L.I.S.N	ESH3-Z5	100113	N/A	N/A
BACL	Test Software	BACL-EMC	V1.0-2010	N/A	N/A

* Statement of Traceability: Bay Area Compliance Laboratories Corp. (Dongguan) attests that all calibrations have been performed, traceable to National Primary Standards and International System of Units (SI).

Test Results Summary

According to the recorded data in following table, the EUT complied with the [FCC Part 15.107](#), with the worst margin reading of:

3.43 dB at 0.270 MHz in the **Line** conducted mode

Test Data


Environmental Conditions

Temperature:	25 °C
Relative Humidity:	48 %
ATM Pressure:	100.9 kPa

The testing was performed by Ares Liu on 2014-03-21.

Test mode: Downloading


AC120 V, 60 Hz, Line:



Frequency (MHz)	Reading (dB μ V)	Corr. Factor (dB)	Limit (dB μ V)	Margin (dB)	Detector (PK/AV/QP)	Comment
0.270	52.08	10.14	61.12	9.04	QP	Compliance
0.270	47.69	10.14	51.12	3.43*	AV	Compliance
0.300	53.55	10.12	60.24	6.69	QP	Compliance
0.300	43.48	10.12	50.24	6.76	AV	Compliance
0.540	43.56	9.92	56.00	12.44	QP	Compliance
0.540	34.14	9.92	46.00	11.86	AV	Compliance
0.920	43.42	9.75	56.00	12.58	QP	Compliance
0.920	30.78	9.75	46.00	15.22	AV	Compliance
1.210	39.04	9.72	56.00	16.96	QP	Compliance
1.210	30.99	9.72	46.00	15.01	AV	Compliance
1.480	39.40	9.72	56.00	16.60	QP	Compliance
1.480	32.45	9.72	46.00	13.55	AV	Compliance

*Within measurement uncertainty!

AC120 V, 60 Hz, Neutral:



Frequency (MHz)	Reading (dB μ V)	Corr. Factor (dB)	Limit (dB μ V)	Margin (dB)	Detector (PK/AV/QP)	Comment
0.270	52.06	10.60	61.12	9.06	QP	Compliance
0.270	47.57	10.60	51.12	3.55*	AV	Compliance
0.300	52.78	10.52	60.24	7.46	QP	Compliance
0.300	40.54	10.52	50.24	9.70	AV	Compliance
0.350	43.93	10.38	58.96	15.03	QP	Compliance
0.350	31.30	10.38	48.96	17.66	AV	Compliance
0.560	43.97	9.91	56.00	12.03	QP	Compliance
0.560	29.40	9.91	46.00	16.60	AV	Compliance
0.830	37.85	9.82	56.00	18.15	QP	Compliance
0.830	29.64	9.82	46.00	16.36	AV	Compliance
0.930	33.61	9.82	56.00	22.39	QP	Compliance
0.930	29.58	9.82	46.00	16.42	AV	Compliance

*Within measurement uncertainty!

FCC §15.109 - RADIATED EMISSIONS

Measurement Uncertainty

Compliance or non- compliance with a disturbance limit shall be determined in the following manner:

If U_{lab} is less than or equal to U_{cispr} of Table 2, then:

- compliance is deemed to occur if no measured disturbance level exceeds the disturbance limit;
- non - compliance is deemed to occur if any measured disturbance level exceeds the disturbance limit.

If U_{lab} is greater than U_{cispr} of Table 1, then:

- compliance is deemed to occur if no measured disturbance level, increased by $(U_{\text{lab}} - U_{\text{cispr}})$, exceeds the disturbance limit;
- non - compliance is deemed to occur if any measured disturbance level, increased by $(U_{\text{lab}} - U_{\text{cispr}})$, exceeds the disturbance limit.

Based on CISPR 16-4-2: 2011, measurement uncertainty of radiated emission at a distance of 3m at Bay Area Compliance Laboratories Corp. (Dongguan) is:

30M~200MHz: 5.0 dB

200M~1GHz: 6.2 dB

1G~6GHz: 4.45 dB


6G~18GHz: 5.23 dB


Table 2 – Values of U_{cispr}

Measurement	U_{cispr}
Radiated disturbance (electric field strength at an OATS or in a SAC) (30 MHz to 1000 MHz)	6.3 dB
Radiated disturbance (electric field strength in a FAR) (1 GHz to 6 GHz)	5.2 dB
Radiated disturbance (electric field strength in a FAR) (6 GHz to 18 GHz)	5.5 dB

EUT Setup

Below 1 GHz:



Above 1GHz:

The radiated emission tests were performed in the 3 meters chamber test site, using the setup accordance with the ANSI C63.4-2003. The specification used was the FCC Part 15.109, Class B limits.

The external I/O cables were draped along the test table and formed a bundle 30 to 40 cm long in the middle.

The spacing between the peripherals was 10 cm.

The adapter of laptop was connected to a 120 VAC/60 Hz power source

EMI Test Receiver Setup

According to FCC 15.33 requirements, the system was measured from 30 MHz to 5 GHz.

During the radiated emission test, the EMI test receiver was set with the following configurations:

Frequency Range	RBW	Video B/W	IF B/W	Detector
30 MHz – 1000 MHz	120 kHz	300 kHz	120 kHz	QP
Above 1 GHz	1 MHz	3 MHz	/	PK
	1 MHz	10 Hz	/	Ave.

Test Procedure

For the radiated emissions test, the adapter of laptop was connected to the first AC floor outlet and the other support equipments were connected to the second AC floor outlet.

Maximizing procedure was performed on the highest emissions to ensure that the EUT complied with all installation combinations.

The data was recorded in Quasi-peak detection mode for 30 MHz to 1 GHz, Peak and average detection mode above 1 GHz.

Corrected Amplitude & Margin Calculation

The Corrected Amplitude is calculated by adding the Antenna Loss and Cable Loss, and subtracting the Amplifier Gain from the Meter Reading. The basic equation is as follows:

$$\text{Corrected Amplitude} = \text{Meter Reading} + \text{Antenna Loss} + \text{Cable Loss} - \text{Amplifier Gain}$$

The “Margin” column of the following data tables indicates the degree of compliance with the applicable limit. For example, a margin of 7dB means the emission is 7dB below the limit. The equation for margin calculation is as follows:

$$\text{Margin} = \text{Limit} - \text{Corrected Amplitude}$$

Test Equipment List and Details

Manufacturer	Description	Model	Serial Number	Calibration Date	Calibration Due Date
R&S	EMI Test Receiver	ESCI	100224	2013-05-06	2014-05-05
Sunol Sciences	Antenna	JB3	A060611-1	2011-09-06	2014-09-05
HP	Amplifier	8447E	2434A02181	2013-09-06	2014-09-05
R&S	Spectrum Analyzer	FSEM	DE31388	2013-05-07	2014-05-06
ETS-Lindgren	Horn Antenna	3115	000 527 35	2012-09-06	2015-09-05
Mini-Circuit	Amplifier	ZVA-213-S+	054201245	2014-02-19	2015-02-18
Farad	Test Software	EZ-EMC	V1.1.4.2	N/A	N/A

* Statement of Traceability: Bay Area Compliance Laboratories Corp. (Dongguan) attests that all calibrations have been performed, traceable to National Primary Standards and International System of Units (SI).

Test Results Summary

According to the data in the following table, the EUT complied with the FCC §15.109, Class B, with the worst margin reading of:


8.80 dB at 165.8000 MHz in the Horizontal polarization

Test Data


Environmental Conditions

Temperature:	26 °C
Relative Humidity:	60%
ATM Pressure:	100.9 kPa


The testing was performed by Ares Liu on 2014-04-22.

1) Below 1GHz:*Test mode: Downloading***Horizontal:**


Frequency (MHz)	Receiver Reading (dB μ V)	Detector (PK/QP/Ave)	Correction Factor (dB/m)	Cord. Amp. (dB μ V/m)	Limit (dB μ V/m)	Margin (dB)
101.7800	39.59	QP	-9.19	30.40	43.50	13.10
146.4000	37.84	QP	-7.24	30.60	43.50	12.90
165.8000	42.31	QP	-7.61	34.70	43.50	8.80
499.4800	33.51	QP	-1.41	32.10	46.00	13.90
797.2700	33.75	QP	2.55	36.30	46.00	9.70
998.0600	31.62	QP	5.68	37.30	54.00	16.70

Vertical:

Frequency (MHz)	Receiver Reading (dB μ V)	Detector (PK/QP/Ave)	Correction Factor (dB/m)	Cord. Amp. (dB μ V/m)	Limit (dB μ V/m)	Margin (dB)
144.4600	41.66	QP	-7.06	34.60	43.50	8.90
165.8000	41.31	QP	-7.61	33.70	43.50	9.80
250.1900	39.97	QP	-7.57	32.40	46.00	13.60
385.0200	36.78	QP	-3.68	33.10	46.00	12.90
533.4300	32.93	QP	-1.23	31.70	46.00	14.30
796.3000	35.03	QP	2.57	37.60	46.00	8.40

2) Above 1GHz:*Test mode: Downloading***Horizontal:**

Frequency (MHz)	Receiver Reading (dB μ V)	Detector (PK/QP/Ave)	Correction Factor (dB/m)	Cord. Amp. (dB μ V/m)	Limit (dB μ V/m)	Margin (dB)
1200.401	49.00	peak	-1.20	47.80	74.00	26.20
1200.401	31.53	peak	-1.20	30.33	74.00	43.67
1851.703	44.99	peak	0.54	45.53	74.00	28.47
1851.703	27.31	AVG	0.54	27.85	54.00	26.15
2122.244	44.94	peak	1.57	46.51	74.00	27.49
2122.244	27.12	AVG	1.57	28.69	54.00	25.31

Vertical:

Frequency (MHz)	Receiver Reading (dB μ V)	Detector (PK/QP/Ave)	Correction Factor (dB/m)	Cord. Amp. (dB μ V/m)	Limit (dB μ V/m)	Margin (dB)
1320.641	48.16	peak	-0.83	47.33	74.00	26.67
1320.641	31.08	AVG	-0.83	30.25	54.00	23.75
1591.182	46.25	peak	-0.76	45.49	74.00	28.51
1591.182	28.61	AVG	-0.76	27.85	54.00	26.15
2492.986	44.92	peak	3.02	47.94	74.00	26.06
2492.986	27.32	AVG	3.02	30.34	54.00	23.66

***** END OF REPORT *****