



-		In Collaboration with		-	10
1	TT	y speag		CA	IC.
		CALIBRATION LABORATORY		1	
		No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China +86-10-62304633-2117			
	E-ma	ail: emf@caict.ac.cn http://www.caict.ac.cn			
10696	AAA	IEEE 802.11ax (40MHz, MCS1, 90pc dc)	WLAN	8.91	± 9.6 %
10697	AAA	IEEE 802.11ax (40MHz, MCS2, 90pc dc)	WLAN	8.61	± 9.6 %
10698	AAA	IEEE 802.11ax (40MHz, MCS3, 90pc dc)	WLAN	8.89	± 9.6 %
0699	AAA	IEEE 802.11ax (40MHz, MCS4, 90pc dc)	WLAN	8.82	± 9.6 %
10700	AAA	IEEE 802.11ax (40MHz, MCS5, 90pc dc)	WLAN WLAN	8.73	± 9.6 %
10701	AAA	IEEE 802.11ax (40MHz, MCS6, 90pc dc) IEEE 802.11ax (40MHz, MCS7, 90pc dc)	WLAN	8.70	± 9.6 %
10702	AAA	IEEE 802.11ax (40MHz, MCS1, 80pc dc)	WLAN	8.82	± 9.6 %
10704	AAA	IEEE 802.11ax (40MHz, MCS9, 90pc dc)	WLAN	8.56	± 9.6 %
10705	AAA	IEEE 802.11ax (40MHz, MCS10, 90pc dc)	WLAN	8.69	± 9.6 %
10706	AAC	IEEE 802.11ax (40MHz, MCS11, 90pc dc)	WLAN	8.66	± 9.6 %
10707	AAC	IEEE 802.11ax (40MHz, MCS0, 99pc dc)	WLAN	8.32	± 9.6 %
10708	AAC	IEEE 802.11ax (40MHz, MCS1, 99pc dc)	WLAN WLAN	8.55	± 9.6 %
10709	AAC	IEEE 802.11ax (40MHz, MCS2, 99pc dc) IEEE 802.11ax (40MHz, MCS3, 99pc dc)	WLAN	8.29	± 9.6 %
10710	AAC	IEEE 802.11ax (40MHz, MCS3, 99pc dc)	WLAN	8.39	± 9.6 %
10712	AAC	IEEE 802.11ax (40MHz, MCS5, 99pc dc)	WLAN	8.67	± 9.6 %
10713	AAC	IEEE 802.11ax (40MHz, MCS6, 99pc dc)	WLAN	8.33	± 9.6 %
10714	AAC	IEEE 802.11ax (40MHz, MCS7, 99pc dc)	WLAN	8.26	± 9.6 %
10715	AAC	IEEE 802.11ax (40MHz, MCS8, 99pc dc)	WLAN	8.45	± 9.6 %
10716	AAC	IEEE 802.11ax (40MHz, MCS9, 99pc dc)	WLAN WLAN	8.30	± 9.6 %
10717	AAC	IEEE 802.11ax (40MHz, MCS10, 99pc dc) IEEE 802.11ax (40MHz, MCS11, 99pc dc)	WLAN	8.24	± 9.6 %
10718	AAC	IEEE 802.11ax (40MHz, MCS11, 99pc dc)	WLAN	8.81	± 9.6 %
10720	AAC	IEEE 802.11ax (80MHz, MCS1, 90pc dc)	WLAN	8.87	± 9.6 %
10721	AAC	IEEE 802.11ax (80MHz, MCS2, 90pc dc)	WLAN	8.76	± 9.6 %
10722	AAC	IEEE 802.11ax (80MHz, MCS3, 90pc dc)	WLAN	8.55	± 9.6 %
10723	AAC	IEEE 802.11ax (80MHz, MCS4, 90pc dc)	WLAN	8.70	± 9.6 %
10724	AAC	IEEE 802.11ax (80MHz, MCS5, 90pc dc)	WLAN WLAN	8.90	± 9.6 %
10725	AAC	IEEE 802.11ax (80MHz, MCS6, 90pc dc) IEEE 802.11ax (80MHz, MCS7, 90pc dc)	WLAN	8.72	± 9.6 %
10726	AAC	IEEE 802.11ax (80MHz, MCS7, 90pc dc)	WLAN	8.66	± 9.6 %
10728	AAC	IEEE 802.11ax (80MHz, MCS9, 90pc dc)	WLAN	8.65	± 9.6 %
10729	AAC	IEEE 802.11ax (80MHz, MCS10, 90pc dc)	WLAN	8.64	± 9.6 %
10730	AAC	IEEE 802.11ax (80MHz, MCS11, 90pc dc)	WLAN	8.67	± 9.6 %
10731	AAC	IEEE 802.11ax (80MHz, MCS0, 99pc dc)	WLAN	8.42	± 9.6 %
10732	AAC	IEEE 802.11ax (80MHz, MCS1, 99pc dc)	WLAN WLAN	8.46	± 9.6 %
10733 10734	AAC	IEEE 802.11ax (80MHz, MCS2, 99pc dc) IEEE 802.11ax (80MHz, MCS3, 99pc dc)	WLAN	8.40	± 9.6 %
10735	AAC	IEEE 802.11ax (80MHz, MCS3, 99pc dc)	WLAN	8.33	± 9.6 %
10736	AAC	IEEE 802.11ax (80MHz, MCS5, 99pc dc)	WLAN	8.27	± 9.6 %
10737	AAC	IEEE 802.11ax (80MHz, MCS6, 99pc dc)	WLAN	8.36	± 9.6 %
10738	AAC	IEEE 802.11ax (80MHz, MCS7, 99pc dc)	WLAN	8.42	± 9.6 %
10739	AAC	IEEE 802.11ax (80MHz, MCS8, 99pc dc)	WLAN	8.29	± 9.6 %
10740	AAC	IEEE 802.11ax (80MHz, MCS9, 99pc dc)	WLAN WLAN	8.48	± 9.6 %
10741 10742	AAC	IEEE 802.11ax (80MHz, MCS10, 99pc dc) IEEE 802.11ax (80MHz, MCS11, 99pc dc)	WLAN	8.40	± 9.6 %
10742	AAC	IEEE 802.11ax (60MHz, MCS1, 99pc dc)	WLAN	8.94	± 9.6 %
10744	AAC	IEEE 802.11ax (160MHz, MCS1, 90pc dc)	WLAN	9.16	± 9.6 %
10745	AAC	IEEE 802.11ax (160MHz, MCS2, 90pc dc)	WLAN	8.93	± 9.6 %
10746	AAC	IEEE 802.11ax (160MHz, MCS3, 90pc dc)	WLAN	9.11	± 9.6 %
10747	AAC	IEEE 802.11ax (160MHz, MCS4, 90pc dc)	WLAN	9.04	± 9.6 %
10748	AAC	IEEE 802.11ax (160MHz, MCS5, 90pc dc)	WLAN	8.93	± 9.6 %
10749	AAC	IEEE 802.11ax (160MHz, MCS6, 90pc dc)	WLAN	8.90	± 9.6 %
10750 10751	AAC	IEEE 802.11ax (160MHz, MCS7, 90pc dc) IEEE 802.11ax (160MHz, MCS8, 90pc dc)	WLAN WLAN	8.79	± 9.6 %
10751	AAC	IEEE 802.11ax (160MHz, MCS6, 90pc dc)	WLAN	8.81	± 9.6 %
10753	AAC	IEEE 802.11ax (160MHz, MCS10, 90pc dc)	WLAN	9.00	± 9.6 %
10754	AAC	IEEE 802.11ax (160MHz, MCS11, 90pc dc)	WLAN	8.94	± 9.6 %

Page 18 of 22





-	and the second	In Collaboration with			-
1	TT.	<u>lspeag</u>		CA	IC
		CALIBRATION LABORATORY			
	A.4.4.	No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China			
		+86-10-62304633-2117			
	E-ma	il: emf@caict.ac.cn http://www.caict.ac.cn			
1					
10755	AAC	IEEE 802.11ax (160MHz, MCS0, 99pc dc)	WLAN WLAN	8.64	± 9.6 %
10756 10757	AAC	IEEE 802.11ax (160MHz, MCS1, 99pc dc) IEEE 802.11ax (160MHz, MCS2, 99pc dc)	WLAN	8.77 8.77	± 9.6 %
0758	AAC	IEEE 802.11ax (160MHz, MCS3, 99pc dc)	WLAN	8.69	± 9.6 %
10759	AAC	IEEE 802.11ax (160MHz, MCS4, 99pc dc)	WLAN	8.58	± 9.6 %
10760	AAC	IEEE 802.11ax (160MHz, MCS5, 99pc dc)	WLAN	8.49	± 9.6 %
0761	AAC	IEEE 802.11ax (160MHz, MCS6, 99pc dc)	WLAN	8.58	± 9.6 %
10762 10763	AAC	IEEE 802.11ax (160MHz, MCS7, 99pc dc) IEEE 802.11ax (160MHz, MCS8, 99pc dc)	WLAN WLAN	8.49 8.53	± 9.6 %
0763	AAC	IEEE 802.11ax (160MHz, MCS8, 99pc dc)	WLAN	8.54	± 9.6 %
0765	AAC	IEEE 802.11ax (160MHz, MCS10, 99pc dc)	WLAN	8.54	± 9.6 %
0766	AAC	IEEE 802.11ax (160MHz, MCS11, 99pc dc)	WLAN	8.51	± 9.6 %
0767	AAC	5G NR (CP-OFDM, 1 RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	7.99	± 9.6 %
0768	AAC	5G NR (CP-OFDM, 1 RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.01	± 9.6 %
10769 10770	AAC	5G NR (CP-OFDM, 1 RB, 15 MHz, QPSK, 15 kHz)	5G NR FR1 TDD 5G NR FR1 TDD	8.01 8.02	± 9.6 %
0770	AAC	5G NR (CP-OFDM, 1 RB, 20 MHz, QPSK, 15 kHz) 5G NR (CP-OFDM, 1 RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.02	± 9.6 %
0772	AAC	5G NR (CP-OFDM, 1 RB, 30 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.23	± 9.6 %
0773	AAC	5G NR (CP-OFDM, 1 RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.03	± 9.6 %
0774	AAC	5G NR (CP-OFDM, 1 RB, 50 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.02	± 9.6 %
0775	AAC	5G NR (CP-OFDM, 50% RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.31	± 9.6 %
0776	AAC	5G NR (CP-OFDM, 50% RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.30	± 9.6 %
0777	AAC	5G NR (CP-OFDM, 50% RB, 15 MHz, QPSK, 15 kHz)	5G NR FR1 TDD 5G NR FR1 TDD	8.30 8.34	± 9.6 %
0778	AAC	5G NR (CP-OFDM, 50% RB, 20 MHz, QPSK, 15 kHz) 5G NR (CP-OFDM, 50% RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.42	± 9.6 %
0780	AAC	5G NR (CP-OFDM, 50% RB, 30 MHz, QPSK, 15 KHz)	5G NR FR1 TDD	8.38	± 9.6 %
0781	AAC	5G NR (CP-OFDM, 50% RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.38	± 9.6 %
0782	AAC	5G NR (CP-OFDM, 50% RB, 50 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.43	± 9.6 %
10783	AAC	5G NR (CP-OFDM, 100% RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.31	± 9.6 %
10784	AAC	5G NR (CP-OFDM, 100% RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 TDD 5G NR FR1 TDD	8.29 8.40	± 9.6 %
10785 10786	AAC	5G NR (CP-OFDM, 100% RB, 15 MHz, QPSK, 15 kHz) 5G NR (CP-OFDM, 100% RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.35	± 9.6 %
10787	AAC	5G NR (CP-OFDM, 100% RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.44	± 9.6 %
10788	AAC	5G NR (CP-OFDM, 100% RB, 30 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.39	± 9.6 %
10789	AAC	5G NR (CP-OFDM, 100% RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.37	± 9.6 %
10790	AAC	5G NR (CP-OFDM, 100% RB, 50 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.39	± 9.6 %
10791	AAC	5G NR (CP-OFDM, 1 RB, 5 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.83	± 9.6 %
10792 10793	AAC	5G NR (CP-OFDM, 1 RB, 10 MHz, QPSK, 30 kHz) 5G NR (CP-OFDM, 1 RB, 15 MHz, QPSK, 30 kHz)	5G NR FR1 TDD 5G NR FR1 TDD	7.92	± 9.6 %
10793	AAC	5G NR (CP-OFDM, 1 RB, 20 MHz, QPSK, 30 KHz)	5G NR FR1 TDD	7.82	± 9.6 %
0795	AAC	5G NR (CP-OFDM, 1 RB, 25 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.84	± 9.6 %
0796	AAC	5G NR (CP-OFDM, 1 RB, 30 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.82	± 9.6 %
0797	AAC	5G NR (CP-OFDM, 1 RB, 40 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.01	± 9.6 %
0798	AAC	5G NR (CP-OFDM, 1 RB, 50 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.89	± 9.6 %
0799	AAC	5G NR (CP-OFDM, 1 RB, 60 MHz, QPSK, 30 kHz)	5G NR FR1 TDD 5G NR FR1 TDD	7.93	± 9.6 %
0801	AAC	5G NR (CP-OFDM, 1 RB, 80 MHz, QPSK, 30 kHz) 5G NR (CP-OFDM, 1 RB, 90 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.87	± 9.6 %
0802	AAE	5G NR (CP-OFDM, 1 RB, 100 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.93	± 9.6 %
0805	AAD	5G NR (CP-OFDM, 50% RB, 10 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.34	± 9.6 %
0806	AAD	5G NR (CP-OFDM, 50% RB, 15 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.37	± 9.6 %
0809	AAD	5G NR (CP-OFDM, 50% RB, 30 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.34	± 9.6 %
10810	AAD	5G NR (CP-OFDM, 50% RB, 40 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.34	± 9.6 %
10812 10817	AAD	5G NR (CP-OFDM, 50% RB, 60 MHz, QPSK, 30 kHz) 5G NR (CP-OFDM, 100% RB, 5 MHz, QPSK, 30 kHz)	5G NR FR1 TDD 5G NR FR1 TDD	8.35 8.35	± 9.6 %
10818	AAD	5G NR (CP-OFDM, 100% RB, 10 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.34	± 9.6 %
0819	AAD	5G NR (CP-OFDM, 100% RB, 15 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.33	± 9.6 %
10820	AAD	5G NR (CP-OFDM, 100% RB, 20 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.30	± 9.6 %
10020			5G NR FR1 TDD	8.41	± 9.6 %

Page 19 of 22





-	TT	In Collaboration with		CA	IC
	11	CALIBRATION LABORATORY			IC
		No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China +86-10-62304633-2117			
	E-ma	il: emf@caict.ac.cn http://www.caict.ac.cn			
0823	AAC	5G NR (CP-OFDM, 100% RB, 40 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.36	± 9.6 %
0824	AAD	5G NR (CP-OFDM, 100% RB, 50 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.39	± 9.6 %
0825	AAD	5G NR (CP-OFDM, 100% RB, 60 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.41	± 9.6 %
0827	AAD	5G NR (CP-OFDM, 100% RB, 80 MHz, QPSK, 30 kHz) 5G NR (CP-OFDM, 100% RB, 90 MHz, QPSK, 30 kHz)	5G NR FR1 TDD 5G NR FR1 TDD	8.42 8.43	± 9.6 %
0829	AAD	5G NR (CP-OFDM, 100% RB, 100 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.40	± 9.6 %
0830	AAD	5G NR (CP-OFDM, 1 RB, 10 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.63	± 9.6 %
0831	AAD	5G NR (CP-OFDM, 1 RB, 15 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.73	± 9.6 %
0832	AAD	5G NR (CP-OFDM, 1 RB, 20 MHz, QPSK, 60 kHz) 5G NR (CP-OFDM, 1 RB, 25 MHz, QPSK, 60 kHz)	5G NR FR1 TDD 5G NR FR1 TDD	7.74	± 9.6 %
0834	AAD	5G NR (CP-OFDM, 1 RB, 30 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.75	± 9.6 %
0835	AAD	5G NR (CP-OFDM, 1 RB, 40 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.70	± 9.6 %
0836	AAE	5G NR (CP-OFDM, 1 RB, 50 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.66	± 9.6 %
0837	AAD	5G NR (CP-OFDM, 1 RB, 60 MHz, QPSK, 60 kHz) 5G NR (CP-OFDM, 1 RB, 80 MHz, QPSK, 60 kHz)	5G NR FR1 TDD 5G NR FR1 TDD	7.68	± 9.6 %
0840	AAD	5G NR (CP-OFDM, 1 RB, 90 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.67	± 9.6 %
0841	AAD	5G NR (CP-OFDM, 1 RB, 100 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.71	± 9.6 %
0843	AAD	5G NR (CP-OFDM, 50% RB, 15 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.49	± 9.6 %
0844	AAD	5G NR (CP-OFDM, 50% RB, 20 MHz, QPSK, 60 kHz) 5G NR (CP-OFDM, 50% RB, 30 MHz, QPSK, 60 kHz)	5G NR FR1 TDD 5G NR FR1 TDD	8.34 8.41	± 9.6 %
0846	AAD	5G NR (CP-OFDM, 50% RB, 30 MHz, QPSK, 60 KHz)	5G NR FR1 TDD	8.34	± 9.6 %
0855	AAD	5G NR (CP-OFDM, 100% RB, 15 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.36	± 9.6 %
0856	AAD	5G NR (CP-OFDM, 100% RB, 20 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.37	± 9.6 %
0857	AAD	5G NR (CP-OFDM, 100% RB, 25 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.35	± 9.6 %
0858 0859	AAD	5G NR (CP-OFDM, 100% RB, 30 MHz, QPSK, 60 kHz) 5G NR (CP-OFDM, 100% RB, 40 MHz, QPSK, 60 kHz)	5G NR FR1 TDD 5G NR FR1 TDD	8.36 8.34	± 9.6 %
0860	AAD	5G NR (CP-OFDM, 100% RB, 50 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.41	± 9.6 %
0861	AAD	5G NR (CP-OFDM, 100% RB, 60 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.40	± 9.6 %
0863	AAD	5G NR (CP-OFDM, 100% RB, 80 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.41	± 9.6 %
0864	AAE	5G NR (CP-OFDM, 100% RB, 90 MHz, QPSK, 60 kHz) 5G NR (CP-OFDM, 100% RB, 100 MHz, QPSK, 60 kHz)	5G NR FR1 TDD 5G NR FR1 TDD	8.37	± 9.6 %
0866	AAD	5G NR (DFT-s-OFDM, 1 RB, 100 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	± 9.6 %
0868	AAD	5G NR (DFT-s-OFDM, 100% RB, 100 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.89	± 9.6 %
0869	AAD	5G NR (DFT-s-OFDM, 1 RB, 100 MHz, QPSK, 120 kHz)	5G NR FR2 TDD	5.75	± 9.6 %
0870	AAD	5G NR (DFT-s-OFDM, 100% RB, 100 MHz, QPSK, 120 kHz) 5G NR (DFT-s-OFDM, 1 RB, 100 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD 5G NR FR2 TDD	5.86 5.75	± 9.6 %
0872	AAD	5G NR (DFT-s-OFDM, 100% RB, 100 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD	6.52	± 9.6 %
0873	AAD	5G NR (DFT-s-OFDM, 1 RB, 100 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	6.61	± 9.6 %
0874	AAD	5G NR (DFT-s-OFDM, 100% RB, 100 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	6.65	± 9.6 %
0875		5G NR (CP-OFDM, 1 RB, 100 MHz, QPSK, 120 kHz) 5G NR (CP-OFDM, 100% RB, 100 MHz, QPSK, 120 kHz)	5G NR FR2 TDD 5G NR FR2 TDD	7.78	± 9.6 %
0876	AAD	5G NR (CP-OFDM, 100% RB, 100 MHz, QPSK, 120 KHz)	5G NR FR2 TDD	7.95	± 9.6 %
0878	AAD	5G NR (CP-OFDM, 100% RB, 100 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD	8.41	± 9.6 %
0879	AAD	5G NR (CP-OFDM, 1 RB, 100 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	8.12	± 9.6 %
0880	AAD	5G NR (CP-OFDM, 100% RB, 100 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	8.38	± 9.6 %
0881	AAD	5G NR (DFT-s-OFDM, 1 RB, 50 MHz, QPSK, 120 kHz) 5G NR (DFT-s-OFDM, 100% RB, 50 MHz, QPSK, 120 kHz)	5G NR FR2 TDD 5G NR FR2 TDD	5.75 5.96	± 9.6 %
0883	AAD	5G NR (DFT-s-OFDM, 1 RB, 50 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD	6.57	± 9.6 %
0884	AAD	5G NR (DFT-s-OFDM, 100% RB, 50 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD	6.53	± 9.6 %
0885	AAD	5G NR (DFT-s-OFDM, 1 RB, 50 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	6.61	± 9.6 %
0886 0887	AAD	5G NR (DFT-s-OFDM, 100% RB, 50 MHz, 64QAM, 120 kHz) 5G NR (CP-OFDM, 1 RB, 50 MHz, QPSK, 120 kHz)	5G NR FR2 TDD 5G NR FR2 TDD	6.65 7.78	± 9.6 %
0888	AAD	5G NR (CP-OFDM, 100% RB, 50 MHz, QPSK, 120 kHz)	5G NR FR2 TDD	8.35	± 9.6 %
0889	AAD	5G NR (CP-OFDM, 1 RB, 50 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD	8.02	± 9.6 %
0890	AAD	5G NR (CP-OFDM, 100% RB, 50 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD	8.40	± 9.6 %
0891 0892	AAD	5G NR (CP-OFDM, 1 RB, 50 MHz, 64QAM, 120 kHz) 5G NR (CP-OFDM, 100% RB, 50 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD 5G NR FR2 TDD	8.13 8.41	± 9.6 %
0897	AAD	5G NR (DFT-s-OFDM, 100% RB, 50 MHz, 040AM, 120 KHz)	5G NR FR1 TDD	5.66	± 9.6 %
0898	AAD	5G NR (DFT-s-OFDM, 1 RB, 10 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.67	± 9.6 %

Page 20 of 22





-	TT	In Collaboration with		CA	IC
-	TT	CALIBRATION LABORATORY			inc
	Add	No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China			
		+86-10-62304633-2117 ail: emf@caict.ac.cn http://www.caict.ac.cn			
	E-m	an ennigeater.ac.en http://www.cater.ac.en			
0899	AAD	5G NR (DFT-s-OFDM, 1 RB, 15 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.67	± 9.6 %
0900	AAD	5G NR (DFT-s-OFDM, 1 RB, 20 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	± 9.6 %
0901	AAD	5G NR (DFT-s-OFDM, 1 RB, 25 MHz, QPSK, 30 kHz) 5G NR (DFT-s-OFDM, 1 RB, 30 MHz, QPSK, 30 kHz)	5G NR FR1 TDD 5G NR FR1 TDD	5.68 5.68	± 9.6 %
0903	AAD	5G NR (DFT-s-OFDM, 1 RB, 40 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	± 9.6 %
0904	AAD	5G NR (DFT-s-OFDM, 1 RB, 50 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	± 9.6 %
0905	AAD	5G NR (DFT-s-OFDM, 1 RB, 60 MHz, QPSK, 30 kHz) 5G NR (DFT-s-OFDM, 1 RB, 80 MHz, QPSK, 30 kHz)	5G NR FR1 TDD 5G NR FR1 TDD	5.68	± 9.6 %
0907	AAD	5G NR (DFT-s-OFDM, 50% RB, 5 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.78	± 9.6 %
0908	AAD	5G NR (DFT-s-OFDM, 50% RB, 10 MHz, QPSK, 30 kHz)	5G NR FR1 TDD 5G NR FR1 TDD	5.93 5.96	± 9.6 %
0909	AAD	5G NR (DFT-s-OFDM, 50% RB, 15 MHz, QPSK, 30 kHz) 5G NR (DFT-s-OFDM, 50% RB, 20 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.83	± 9.6 %
0911	AAD	5G NR (DFT-s-OFDM, 50% RB, 25 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.93	± 9.6 %
0912	AAD	5G NR (DFT-s-OFDM, 50% RB, 30 MHz, QPSK, 30 kHz)	5G NR FR1 TDD 5G NR FR1 TDD	5.84 5.84	± 9.6 %
0913	AAD	5G NR (DFT-s-OFDM, 50% RB, 40 MHz, QPSK, 30 kHz) 5G NR (DFT-s-OFDM, 50% RB, 50 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.85	± 9.6 %
0915	AAD	5G NR (DFT-s-OFDM, 50% RB, 60 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.83	± 9.6 %
0916	AAD	5G NR (DFT-s-OFDM, 50% RB, 80 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.87	± 9.6 %
0917	AAD	5G NR (DFT-s-OFDM, 50% RB, 100 MHz, QPSK, 30 kHz) 5G NR (DFT-s-OFDM, 100% RB, 5 MHz, QPSK, 30 kHz)	5G NR FR1 TDD 5G NR FR1 TDD	5.86	± 9.6 %
0919	AAD	5G NR (DFT-s-OFDM, 100% RB, 10 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.86	± 9.6 %
0920	AAD	5G NR (DFT-s-OFDM, 100% RB, 15 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.87 5.84	± 9.6 %
0921	AAD	5G NR (DFT-s-OFDM, 100% RB, 20 MHz, QPSK, 30 kHz) 5G NR (DFT-s-OFDM, 100% RB, 25 MHz, QPSK, 30 kHz)	5G NR FR1 TDD 5G NR FR1 TDD	5.82	± 9.6 %
0923	AAD	5G NR (DFT-s-OFDM, 100% RB, 30 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.84	± 9.6 %
0924	AAD	5G NR (DFT-s-OFDM, 100% RB, 40 MHz, QPSK, 30 kHz)	5G NR FR1 TDD 5G NR FR1 TDD	5.84 5.95	± 9.6 %
0925	AAD	5G NR (DFT-s-OFDM, 100% RB, 50 MHz, QPSK, 30 kHz) 5G NR (DFT-s-OFDM, 100% RB, 60 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.84	± 9.6 %
0927	AAD	5G NR (DFT-s-OFDM, 100% RB, 80 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.94	± 9.6 %
0928	AAD	5G NR (DFT-s-OFDM, 1 RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.52	± 9.6 %
0929	AAD	5G NR (DFT-s-OFDM, 1 RB, 10 MHz, QPSK, 15 kHz) 5G NR (DFT-s-OFDM, 1 RB, 15 MHz, QPSK, 15 kHz)	5G NR FR1 FDD 5G NR FR1 FDD	5.52	± 9.6 %
0931	AAD	5G NR (DFT-s-OFDM, 1 RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.51	± 9.6 %
0932	AAB	5G NR (DFT-s-OFDM, 1 RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 FDD 5G NR FR1 FDD	5.51	± 9.6 %
0933	AAA	5G NR (DFT-s-OFDM, 1 RB, 30 MHz, QPSK, 15 kHz) 5G NR (DFT-s-OFDM, 1 RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.51	± 9.6 %
0935	AAA	5G NR (DFT-s-OFDM, 1 RB, 50 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.51	± 9.6 %
0936	AAC	5G NR (DFT-s-OFDM, 50% RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.90	± 9.6 %
0937	AAB	5G NR (DFT-s-OFDM, 50% RB, 10 MHz, QPSK, 15 kHz) 5G NR (DFT-s-OFDM, 50% RB, 15 MHz, QPSK, 15 kHz)	5G NR FR1 FDD 5G NR FR1 FDD	5.90	± 9.6 %
0939	AAB	5G NR (DFT-s-OFDM, 50% RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.82	± 9.6 %
0940	AAB	5G NR (DFT-s-OFDM, 50% RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.89	± 9.6 %
0941	AAB	5G NR (DFT-s-OFDM, 50% RB, 30 MHz, QPSK, 15 kHz) 5G NR (DFT-s-OFDM, 50% RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 FDD 5G NR FR1 FDD	5.83	± 9.6 %
0943	AAB	5G NR (DFT-s-OFDM, 50% RB, 50 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.95	± 9.6 %
0944	AAB	5G NR (DFT-s-OFDM, 100% RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.81	± 9.6 %
0945	AAB	5G NR (DFT-s-OFDM, 100% RB, 10 MHz, QPSK, 15 kHz) 5G NR (DFT-s-OFDM, 100% RB, 15 MHz, QPSK, 15 kHz)	5G NR FR1 FDD 5G NR FR1 FDD	5.85 5.83	± 9.6 %
0947	AAB	5G NR (DFT-s-OFDM, 100% RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.87	± 9.6 %
0948	AAB	5G NR (DFT-s-OFDM, 100% RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.94	± 9.6 %
0949	AAB	5G NR (DFT-s-OFDM, 100% RB, 30 MHz, QPSK, 15 kHz) 5G NR (DFT-s-OFDM, 100% RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 FDD 5G NR FR1 FDD	5.87 5.94	± 9.6 %
0951	AAB	5G NR (DFT-s-OFDM, 100% RB, 50 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.92	± 9.6 %
0952	AAB	5G NR DL (CP-OFDM, TM 3.1, 5 MHz, 64-QAM, 15 kHz)	5G NR FR1 FDD	8.25	± 9.6 %
0953	AAB	5G NR DL (CP-OFDM, TM 3.1, 10 MHz, 64-QAM, 15 kHz) 5G NR DL (CP-OFDM, TM 3.1, 15 MHz, 64-QAM, 15 kHz)	5G NR FR1 FDD 5G NR FR1 FDD	8.15 8.23	± 9.6 %
0955	AAB	5G NR DL (CP-OFDM, TM 3.1, 10 MHz, 64-QAM, 15 KHz)	5G NR FR1 FDD	8.42	± 9.6 %
0956	AAB	5G NR DL (CP-OFDM, TM 3.1, 5 MHz, 64-QAM, 30 kHz) 5G NR DL (CP-OFDM, TM 3.1, 10 MHz, 64-QAM, 30 kHz)	5G NR FR1 FDD 5G NR FR1 FDD	8.14 8.31	± 9.6 %

Page 21 of 22









Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2117 E-mail: emf@caict.ac.cn http://www.caict.ac.cn

10958	AAB	5G NR DL (CP-OFDM, TM 3.1, 15 MHz, 64-QAM, 30 kHz)	5G NR FR1 FDD	8.61	± 9.6 %
10959	AAB	5G NR DL (CP-OFDM, TM 3.1, 20 MHz, 64-QAM, 30 kHz)	5G NR FR1 FDD	8.33	±9.6 %
10960	AAB	5G NR DL (CP-OFDM, TM 3.1, 5 MHz, 64-QAM, 15 kHz)	5G NR FR1 TDD	9.32	± 9.6 %
10961	AAB	5G NR DL (CP-OFDM, TM 3.1, 10 MHz, 64-QAM, 15 kHz)	5G NR FR1 TDD	9.36	± 9.6 %
10962	AAB	5G NR DL (CP-OFDM, TM 3.1, 15 MHz, 64-QAM, 15 kHz)	5G NR FR1 TDD	9.40	± 9.6 %
10963	AAB	5G NR DL (CP-OFDM, TM 3.1, 20 MHz, 64-QAM, 15 kHz)	5G NR FR1 TDD	9.55	±9.6 %
10964	AAB	5G NR DL (CP-OFDM, TM 3.1, 5 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.29	± 9.6 %
10965	AAB	5G NR DL (CP-OFDM, TM 3.1, 10 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.37	± 9.6 %
10966	AAB	5G NR DL (CP-OFDM, TM 3.1, 15 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.55	± 9.6 %
10967	AAB	5G NR DL (CP-OFDM, TM 3.1, 20 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.42	±9.6 %
10968	AAB	5G NR DL (CP-OFDM, TM 3.1, 100 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.49	±9.6%
10972	AAB	5G NR (CP-OFDM, 1 RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	11.59	±9.6 %
10973	AAB	5G NR (DFT-s-OFDM, 1 RB, 100 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	9.06	± 9.6 %
10974	AAB	5G NR (CP-OFDM, 100% RB, 100 MHz, 256-QAM, 30 kHz)	5G NR FR1 TDD	10.28	± 9.6 %
10978	AAA	ULLA BDR	ULLA	1.16	± 9.6 %
10979	AAA	ULLA HDR4	ULLA	8.58	± 9.6 %
10980	AAA	ULLA HDR8	ULLA	10.32	± 9.6 %
10981	AAA	ULLA HDRp4	ULLA	3.19	± 9.6 %
10982	AAA	ULLA HDRp8	ULLA	3.43	± 9.6 %
10983	AAA	5G NR DL (CP-OFDM, TM 3.1, 40 MHz, 64-QAM, 15 kHz)	5G NR FR1 TDD	9.31	± 9.6 %
10984	AAA	5G NR DL (CP-OFDM, TM 3.1, 50 MHz, 64-QAM, 15 kHz)	5G NR FR1 TDD	9.42	± 9.6 %
10985	AAA	5G NR DL (CP-OFDM, TM 3.1, 40 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.54	± 9.6 %
10986	AAA	5G NR DL (CP-OFDM, TM 3.1, 50 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.50	± 9.6 %
10987	AAA	5G NR DL (CP-OFDM, TM 3.1, 60 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.53	± 9.6 %
10988	AAA	5G NR DL (CP-OFDM, TM 3.1, 70 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.38	± 9.6 %
10989	AAA	5G NR DL (CP-OFDM, TM 3.1, 80 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.33	± 9.6 %
10990	AAA	5G NR DL (CP-OFDM, TM 3.1, 90 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.52	± 9.6 %

<sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

Certificate No:Z22-60492

Page 22 of 22





# 835 MHz Dipole Calibration Certificate

**Calibration Laboratory of** 

Schweizerischer Kalibrierdienst S Schmid & Partner Service suisse d'étalonnage С **ac-MR Engineering AG** Servizio svizzero di taratura S Zeughausstrasse 43, 8004 Zurich, Switzerland Swiss Calibration Service Accreditation No.: SCS 0108 Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates Client CTTL Certificate No. D835V2-4d069 Jul23 Beijing CALIBRATION CERTIFICATE D835V2 - SN:4d069 Object QA CAL-05.v12 Calibration procedure(s) Calibration Procedure for SAR Validation Sources between 0.7-3 GHz Calibration date: July 14, 2023 This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%. Calibration Equipment used (M&TE critical for calibration) Primary Standards ID # Cal Date (Certificate No.) Scheduled Calibration SN: 104778 Power meter NRP2 30-Mar-23 (No. 217-03804/03805) Mar-24 Power sensor NRP-Z91 SN: 103244 30-Mar-23 (No. 217-03804) Mar-24 Power sensor NRP-Z91 SN: 103245 30-Mar-23 (No. 217-03805) Mar-24 SN: BH9394 (20k) Reference 20 dB Attenuator 30-Mar-23 (No. 217-03809) Mar-24 SN: 310982 / 06327 Type-N mismatch combination 30-Mar-23 (No. 217-03810) Mar-24 Reference Probe EX3DV4 SN: 7349 10-Jan-23 (No. EX3-7349\_Jan23) Jan-24 DAE4 SN: 601 19-Dec-22 (No. DAE4-601\_Dec22) Dec-23 Secondary Standards ID # Check Date (in house) Scheduled Check Power meter E4419B SN: GB39512475 30-Oct-14 (in house check Oct-22) In house check: Oct-24 Power sensor HP 8481A SN: US37292783 07-Oct-15 (in house check Oct-22) In house check: Oct-24 Power sensor HP 8481A SN: MY41093315 07-Oct-15 (in house check Oct-22) In house check: Oct-24 RF generator R&S SMT-06 SN: 100972 15-Jun-15 (in house check Oct-22) In house check: Oct-24 Network Analyzer Agilent E8358A SN: US41080477 31-Mar-14 (in house check Oct-22) In house check: Oct-24 Name Function Calibrated by: Michael Weber Laboratory Technician Approved by: Sven Kühn Technical Manager Issued: July 18, 2023 This calibration certificate shall not be reproduced except in full without written approval of the laboratory. Certificate No: D835V2-4d069 Jul23 Page 1 of 6





**Calibration Laboratory of** Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Schweizerischer Kalibrierdienst S Service suisse d'étalonnage С

Servizio svizzero di taratura S Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

## Glossary:

TSL	tissue simulating liquid
ConvF	sensitivity in TSL / NORM x,y,z
N/A	not applicable or not measured

#### Calibration is Performed According to the Following Standards:

- a) IEC/IEEE 62209-1528, "Measurement Procedure For The Assessment Of Specific Absorption Rate Of Human Exposure To Radio Frequency Fields From Hand-Held And Body-Worn Wireless Communication Devices - Part 1528: Human Models, Instrumentation And Procedures (Frequency Range of 4 MHz to 10 GHz)", October 2020.
- b) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

#### Additional Documentation:

c) DASY System Handbook

#### Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The source is mounted in a touch configuration below the . center marking of the flat phantom.
- Return Loss: This parameter is measured with the source positioned under the liquid filled phantom (as described in the measurement condition clause). The Return Loss ensures low reflected power. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: D835V2-4d069\_Jul23

Page 2 of 6





#### **Measurement Conditions**

DASY system configuration, as far as not given on page 1.

DASY Version	DASY52	V52.10.4
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom	
Distance Dipole Center - TSL	15 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	835 MHz ± 1 MHz	

Head TSL parameters The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	41.5	0.90 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	41.9 ± 6 %	0.92 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		

#### SAR result with Head TSL

SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition		
SAR measured	250 mW input power	2.44 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	9.62 W/kg ± 17.0 % (k=2)	
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	condition		
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL SAR measured	condition 250 mW input power	1.58 W/kg	

Certificate No: D835V2-4d069\_Jul23

Page 3 of 6





# Appendix (Additional assessments outside the scope of SCS 0108)

#### Antenna Parameters with Head TSL

Impedance, transformed to feed point	51.3 Ω - 1.2 jΩ	
Return Loss	- 35.2 dB	

#### **General Antenna Parameters and Design**

Electrical Delay (one direction)	1.393 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

#### **Additional EUT Data**

Manufactured by	SPEAG

Certificate No: D835V2-4d069\_Jul23

Page 4 of 6





## **DASY5 Validation Report for Head TSL**

Date: 14.07.2023

Test Laboratory: SPEAG, Zurich, Switzerland

#### DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN: 4d069


Communication System: UID 0 - CW; Frequency: 835 MHz Medium parameters used: f = 835 MHz;  $\sigma$  = 0.92 S/m;  $\epsilon_r$  = 41.9;  $\rho$  = 1000 kg/m<sup>3</sup> Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 SN7349; ConvF(9.69, 9.69, 9.69) @ 835 MHz; Calibrated: 10.01.2023
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 19.12.2022
- Phantom: Flat Phantom 4.9 (front); Type: QD 00L P49 AA; Serial: 1001
- DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501)

# Dipole Calibration for Head Tissue/Pin=250 mW, d=15mm/Zoom Scan (7x7x7)/Cube 0:

Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 63.54 V/m; Power Drift = -0.05 dB Peak SAR (extrapolated) = 3.68 W/kg **SAR(1 g) = 2.44 W/kg; SAR(10 g) = 1.58 W/kg** Smallest distance from peaks to all points 3 dB below = 16.8 mm Ratio of SAR at M2 to SAR at M1 = 65.9% Maximum value of SAR (measured) = 3.27 W/kg



0 dB = 3.27 W/kg = 5.15 dBW/kg


Certificate No: D835V2-4d069\_Jul23

Page 5 of 6





### Impedance Measurement Plot for Head TSL



Certificate No: D835V2-4d069\_Jul23

Page 6 of 6





# **1750 MHz Dipole Calibration Certificate**

Engineering AG Leughausstrasse 43, 8004 Zurich,		ACCENTA SC S	Servizio svizzero di taratura
Accredited by the Swiss Accreditation The Swiss Accreditation Service in Aultilateral Agreement for the rec	s one of the signatories		Accreditation No.: 3C3 0100
Client CTTL Beijing		Certificate No.	D1750V2-1003_Jul23
CALIBRATION C			
Object	D1750V2 - SN:10	103	
Calibration procedure(s)	QA CAL-05.v12 Calibration Proce	dure for SAR Validation Sources	between 0.7-3 GHz
Calibration date:	July 12, 2023		
The measurements and the uncerta	ainties with confidence pr	onal standards, which realize the physical uni obability are given on the following pages an y facility: environment temperature (22 ± 3)°C	d are part of the certificate.
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE	ainties with confidence pr	obability are given on the following pages an	d are part of the certificate.
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power meter NRP2	ainties with confidence pre- ed in the closed laborator critical for calibration) ID # SN: 104778	Cal Date (Certificate No.) 30-Mar-23 (No. 217-03804/03805)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Mar-24
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power meter NRP2 Power sensor NRP-Z91	ainties with confidence pre- ed in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244	Cal Date (Certificate No.) 30-Mar-23 (No. 217-03804)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Mar-24 Mar-24
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power meter NRP2	ainties with confidence pre- ed in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245	Cal Date (Certificate No.) 30-Mar-23 (No. 217-03804) 30-Mar-23 (No. 217-03804) 30-Mar-23 (No. 217-03805)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Mar-24
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91	ainties with confidence pre- ed in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244	Cal Date (Certificate No.) 30-Mar-23 (No. 217-03804)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Mar-24 Mar-24 Mar-24 Mar-24
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4	ainties with confidence pre- ed in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 103245 SN: BH9394 (20k) SN: 310982 / 06327 SN: 7349	Cal Date (Certificate No.)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03805)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03805)           30-Mar-23 (No. 217-03810)           10-Jan-23 (No. EX3-7349_Jan23)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Mar-24 Mar-24 Mar-24 Mar-24 Mar-24 Mar-24 Jan-24
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination	ainties with confidence pre- ed in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 103245 SN: BH9394 (20k) SN: 310982 / 06327	Cal Date (Certificate No.)         30-Mar-23 (No. 217-03804/03805)         30-Mar-23 (No. 217-03804)         30-Mar-23 (No. 217-03804)         30-Mar-23 (No. 217-03805)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Mar-24 Mar-24 Mar-24 Mar-24 Mar-24 Mar-24 Mar-24
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4	ainties with confidence pre- ed in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 103245 SN: BH9394 (20k) SN: 310982 / 06327 SN: 7349	Cal Date (Certificate No.)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03805)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03805)           30-Mar-23 (No. 217-03810)           10-Jan-23 (No. EX3-7349_Jan23)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Mar-24 Mar-24 Mar-24 Mar-24 Mar-24 Mar-24 Jan-24
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B	ainties with confidence pre- ed in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244 SN: 103244 SN: 103245 SN: BH9394 (20k) SN: 310982 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475	Cal Date (Certificate No.)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03805)           30-Mar-23 (No. 217-03805)           30-Mar-23 (No. 217-03805)           30-Mar-23 (No. 217-03809)           30-Mar-23 (No. 217-03809)           30-Mar-23 (No. 217-03810)           10-Jan-23 (No. EX3-7349_Jan23)           19-Dec-22 (No. DAE4-601_Dec22)           Check Date (in house)           30-Oct-14 (in house check Oct-22)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Mar-24 Mar-24 Mar-24 Mar-24 Mar-24 Mar-24 Jan-24 Jan-24 Jan-24 Dec-23 Scheduled Check In house check: Oct-24
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A	ainties with confidence presentation           ain the closed laborator           critical for calibration)           ID #           SN: 104778           SN: 103244           SN: 103245           SN: 8H9394 (20k)           SN: 310982 / 06327           SN: 7349           SN: 601           ID #           SN: GB39512475           SN: US37292783	Cal Date (Certificate No.)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03805)           30-Mar-23 (No. 217-03809)           30-Mar-23 (No. 217-03810)           10-Jan-23 (No. EX3-7349_Jan23)           19-Dec-22 (No. DAE4-601_Dec22)           Check Date (in house)           30-Oct-14 (in house check Oct-22)           07-Oct-15 (in house check Oct-22)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Mar-24 Mar-24 Mar-24 Mar-24 Mar-24 Mar-24 Jan-24 Jan-24 Dec-23 Scheduled Check In house check: Oct-24 In house check: Oct-24
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A	ainties with confidence pr ed in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244 SN: 103244 SN: 103245 SN: 103245 SN: 310982 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: MY41093315	Cal Date (Certificate No.)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03805)           30-Mar-23 (No. 217-03809)           30-Mar-23 (No. 217-03809)           30-Mar-23 (No. 217-03809)           30-Mar-23 (No. 217-03810)           10-Jan-23 (No. 217-03810)           10-Jan-23 (No. EX3-7349_Jan23)           19-Dec-22 (No. DAE4-601_Dec22)           Check Date (in house)           30-Oct-14 (in house check Oct-22)           07-Oct-15 (in house check Oct-22)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Mar-24 Mar-24 Mar-24 Mar-24 Mar-24 Jan-24 Jan-24 Dec-23 Scheduled Check In house check: Oct-24 In house check: Oct-24 In house check: Oct-24
The measurements and the uncerta All calibrations have been conducts Calibration Equipment used (M&TE Primary Standards Power meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06	ainties with confidence presentation           ain the closed laborator           critical for calibration)           ID #           SN: 104778           SN: 103244           SN: 103245           SN: 8H9394 (20k)           SN: 310982 / 06327           SN: 7349           SN: 601           ID #           SN: GB39512475           SN: US37292783	Cal Date (Certificate No.)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03805)           30-Mar-23 (No. 217-03809)           30-Mar-23 (No. 217-03810)           10-Jan-23 (No. EX3-7349_Jan23)           19-Dec-22 (No. DAE4-601_Dec22)           Check Date (in house)           30-Oct-14 (in house check Oct-22)           07-Oct-15 (in house check Oct-22)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Mar-24 Mar-24 Mar-24 Mar-24 Mar-24 Mar-24 Jan-24 Jan-24 Dec-23 Scheduled Check In house check: Oct-24 In house check: Oct-24
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4	ainties with confidence pr ed in the closed laborator critical for calibration) ID # SN: 104778 SN: 103245 SN: 81H9394 (20k) SN: 310982 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: MY41093315 SN: 100972 SN: US41080477	Cal Date (Certificate No.)         30-Mar-23 (No. 217-03804/03805)         30-Mar-23 (No. 217-03804/03805)         30-Mar-23 (No. 217-03804)         30-Mar-23 (No. 217-03804)         30-Mar-23 (No. 217-03809)         30-Mar-23 (No. 217-03809)         30-Mar-23 (No. 217-03809)         30-Mar-23 (No. 217-03810)         10-Jan-23 (No. EX3-7349_Jan23)         19-Dec-22 (No. DAE4-601_Dec22)         Check Date (in house)         30-Oct-14 (in house check Oct-22)         07-Oct-15 (in house check Oct-22)         15-Jun-15 (in house check Oct-22)         31-Mar-14 (in house check Oct-22)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Mar-24 Mar-24 Mar-24 Mar-24 Mar-24 Jan-
The measurements and the uncerta All calibrations have been conducts Calibration Equipment used (M&TE Primary Standards Power meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06	ainties with confidence pr ed in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244 SN: 103244 SN: 103245 SN: 103245 SN: 103245 SN: 310962 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: WY41093315 SN: 100972	Cal Date (Certificate No.)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03805)           30-Mar-23 (No. 217-03809)           30-Mar-23 (No. 217-03809)           30-Mar-23 (No. 217-03810)           10-Jan-23 (No. EX3-7349_Jan23)           19-Dec-22 (No. DAE4-601_Dec22)           Check Date (in house)           30-Oct-14 (in house check Oct-22)           07-Oct-15 (in house check Oct-22)           07-Oct-15 (in house check Oct-22)           15-Jun-15 (in house check Oct-22)	d are part of the certificate. 2 and humidity < 70%. Scheduled Calibration Mar-24 Mar-24 Mar-24 Mar-24 Mar-24 Mar-24 Jan-24 Jan-24 Dec-23 Scheduled Check In house check: Oct-24 In house check: Oct-24
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power meter NRP2 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent E8358A	ainties with confidence pr ed in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: BH9394 (20k) SN: 310982 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: MY41093315 SN: US37292783 SN: US41080477 Name	Cal Date (Certificate No.)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03805)           30-Mar-23 (No. 217-03809)           30-Mar-23 (No. 217-03809)           30-Mar-23 (No. 217-03809)           30-Mar-23 (No. 217-03809)           30-Mar-23 (No. EX3-7349_Jan23)           19-Dec-22 (No. DAE4-601_Dec22)           Check Date (in house)           30-Oct-14 (in house check Oct-22)           07-Oct-15 (in house check Oct-22)           15-Jun-15 (in house check Oct-22)           15-Jun-15 (in house check Oct-22)           31-Mar-14 (in house check Oct-22)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Mar-24 Mar-24 Mar-24 Mar-24 Mar-24 Jan-





#### Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



 S
 Schweizerischer Kalibrierdienst

 C
 Service suisse d'étalonnage

 S
 Servizio svizzero di taratura

 S
 Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

#### Glossary:

TSL	tissue simulating liquid
ConvF	sensitivity in TSL / NORM x,y,z
N/A	not applicable or not measured

#### Calibration is Performed According to the Following Standards:

- a) IEC/IEEE 62209-1528, "Measurement Procedure For The Assessment Of Specific Absorption Rate Of Human Exposure To Radio Frequency Fields From Hand-Held And Body-Worn Wireless Communication Devices - Part 1528: Human Models, Instrumentation And Procedures (Frequency Range of 4 MHz to 10 GHz)", October 2020.
- b) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

#### Additional Documentation:

c) DASY System Handbook


#### Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end
  of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The source is mounted in a touch configuration below the center marking of the flat phantom.
- Return Loss: This parameter is measured with the source positioned under the liquid filled phantom (as described in the measurement condition clause). The Return Loss ensures low reflected power. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: D1750V2-1003\_Jul23

Page 2 of 6







#### **Measurement Conditions**

DASY system configuration, as far as not given on page 1.

DASY Version	DASY52	V52.10.4	
Extrapolation	Advanced Extrapolation		
Phantom	Modular Flat Phantom		
Distance Dipole Center - TSL	10 mm	with Spacer	
Zoom Scan Resolution	dx, dy, dz = 5 mm		
Frequency	1750 MHz ± 1 MHz		

Head TSL parameters The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity	
Nominal Head TSL parameters	22.0 °C	40.1	1.37 mho/m	
Measured Head TSL parameters	(22.0 ± 0.2) °C	40.0 ± 6 %	1.34 mho/m ± 6 %	
Head TSL temperature change during test	< 0.5 °C			

# SAR result with Head TSL

SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	8.84 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	35.8 W/kg ± 17.0 % (k=2)
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	condition	
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL SAR measured	condition 250 mW input power	4.69 W/kg

Certificate No: D1750V2-1003\_Jul23

Page 3 of 6





#### Appendix (Additional assessments outside the scope of SCS 0108)

#### Antenna Parameters with Head TSL

Impedance, transformed to feed point	51.4 Ω - 0.2 jΩ	
Return Loss	- 37.1 dB	

#### **General Antenna Parameters and Design**

Electrical Delay (one direction)	1.214 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

#### **Additional EUT Data**

Manufactured by	SPEAG

Certificate No: D1750V2-1003\_Jul23

Page 4 of 6





## **DASY5 Validation Report for Head TSL**

Date: 12.07.2023

Test Laboratory: SPEAG, Zurich, Switzerland

### DUT: Dipole 1750 MHz; Type: D1750V2; Serial: D1750V2 - SN:1003


Communication System: UID 0 - CW; Frequency: 1750 MHz Medium parameters used: f = 1750 MHz;  $\sigma = 1.34$  S/m;  $\epsilon_r = 40$ ;  $\rho = 1000$  kg/m<sup>3</sup> Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 SN7349; ConvF(8.67, 8.67, 8.67) @ 1750 MHz; Calibrated: 10.01.2023
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 19.12.2022
- Phantom: Flat Phantom 5.0 (front); Type: QD 000 P50 AA; Serial: 1001
- DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501)

#### Dipole Calibration for Head Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0:

Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 104.6 V/m; Power Drift = 0.05 dB Peak SAR (extrapolated) = 16.1 W/kg SAR(1 g) = 8.84 W/kg; SAR(10 g) = 4.69 W/kg Smallest distance from peaks to all points 3 dB below = 10 mm Ratio of SAR at M2 to SAR at M1 = 55.4% Maximum value of SAR (measured) = 13.3 W/kg



0 dB = 13.3 W/kg = 11.24 dBW/kg

Certificate No: D1750V2-1003\_Jul23

Page 5 of 6





# Impedance Measurement Plot for Head TSL

Eile	⊻iew	<u>C</u> hannel	Sweep	Calibration	Irace	Scale	Marker	System	Window	Help	
					Ę		XXXX			1.750000 GHz 533.69 pF 1.750000 GHz	51.408 Ω -170.41 mΩ 13.989 mU -6.8032 °
		Ch 1 Avg = art 1.55000		_				E			Stop 1.95000 GHz
10.0 5.0 0.0	10	dB 811						>	1:	1.750000 GHz	-37.085 dB
-10 -15 -20	.00				/						
-25 -30 -35	.00						1	/			
135		Ch 1 Avg =	1.2				ſ	1			

Certificate No: D1750V2-1003\_Jul23

Page 6 of 6





# **1900 MHz Dipole Calibration Certificate**

Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich	y Of		Service suisse d'étalonnage Servizio svizzero di taratura
Accredited by the Swiss Accreditati The Swiss Accreditation Service Multilateral Agreement for the red	is one of the signatorie		Accreditation No.: SCS 010
Client CTTL Beijing		Certificate No.	D1900V2-5d101_Jul23
CALIBRATION C	ERTIFICATI		
Object	D1900V2 - SN:50	1101	
Calibration procedure(s)	QA CAL-05.v12 Calibration Proce	edure for SAR Validation Sources	s between 0.7-3 GHz
Calibration date:	July 17, 2023		
All calibrations have been conducted	ainties with confidence p ed in the closed laborator	robability are given on the following pages an y facility: environment temperature $(22 \pm 3)^{\circ}($	nd are part of the certificate.
The measurements and the uncert	ainties with confidence p ed in the closed laborator	robability are given on the following pages an	nd are part of the certificate.
The measurements and the uncert All calibrations have been conduct Calibration Equipment used (M&TE Primary Standards Power meter NRP2	ainties with confidence p ed in the closed laborator E critical for calibration) ID # SN: 104778	robability are given on the following pages an y facility: environment temperature $(22 \pm 3)^{\circ}($	nd are part of the certificate. C and humidity < 70%.
The measurements and the uncert All calibrations have been conduct Calibration Equipment used (M&TE Primary Standards Power meter NRP2 Power sensor NRP-Z91	ainties with confidence p ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244	Cal Date (Certificate No.) 30-Mar-23 (No. 217-03804/03805) 30-Mar-23 (No. 217-03804)	nd are part of the certificate. C and humidity < 70%. <u>Scheduled Calibration</u> Mar-24 Mar-24
The measurements and the uncert All calibrations have been conduct Calibration Equipment used (M&TE Primary Standards Power meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91	ainties with confidence p ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245	Cal Date (Certificate No.)         30-Mar-23 (No. 217-03804/03805)         30-Mar-23 (No. 217-03804)         30-Mar-23 (No. 217-03805)	nd are part of the certificate. C and humidity < 70%. <u>Scheduled Calibration</u> Mar-24 Mar-24 Mar-24 Mar-24
The measurements and the uncert All calibrations have been conduct Calibration Equipment used (M&TE Primary Standards Power meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator	ainties with confidence p ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: BH9394 (20k)	Cal Date (Certificate No.)         30-Mar-23 (No. 217-03804/03805)         30-Mar-23 (No. 217-03804)         30-Mar-23 (No. 217-03804)         30-Mar-23 (No. 217-03805)         30-Mar-23 (No. 217-03805)         30-Mar-23 (No. 217-03805)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration Mar-24 Mar-24 Mar-24 Mar-24 Mar-24
The measurements and the uncert All calibrations have been conduct Calibration Equipment used (M&TE Primary Standards Power meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91	ainties with confidence p ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245	Cal Date (Certificate No.)         30-Mar-23 (No. 217-03804/03805)         30-Mar-23 (No. 217-03804)         30-Mar-23 (No. 217-03804)         30-Mar-23 (No. 217-03805)         30-Mar-23 (No. 217-03805)         30-Mar-23 (No. 217-03805)         30-Mar-23 (No. 217-03805)         30-Mar-23 (No. 217-03809)         30-Mar-23 (No. 217-03810)	nd are part of the certificate. C and humidity < 70%. <u>Scheduled Calibration</u> Mar-24 Mar-24 Mar-24 Mar-24 Mar-24 Mar-24 Mar-24
The measurements and the uncert All calibrations have been conduct Calibration Equipment used (M&TE Primary Standards Power meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination	ainties with confidence p ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 103245 SN: BH9394 (20k) SN: 310982 / 06327	Cal Date (Certificate No.)         30-Mar-23 (No. 217-03804/03805)         30-Mar-23 (No. 217-03804)         30-Mar-23 (No. 217-03804)         30-Mar-23 (No. 217-03805)         30-Mar-23 (No. 217-03805)         30-Mar-23 (No. 217-03805)         30-Mar-23 (No. 217-03809)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration Mar-24 Mar-24 Mar-24 Mar-24 Mar-24
The measurements and the uncert All calibrations have been conducts Calibration Equipment used (M&TE Primary Standards Power meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards	ainties with confidence p ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 103245 SN: 310982 / 06327 SN: 7349 SN: 601 ID #	Cal Date (Certificate No.)         30-Mar-23 (No. 217-03804/03805)         30-Mar-23 (No. 217-03804)         30-Mar-23 (No. 217-03804)         30-Mar-23 (No. 217-03805)         30-Mar-23 (No. 217-03805)         30-Mar-23 (No. 217-03809)         30-Mar-23 (No. 217-03810)         10-Jan-23 (No. EX3-7349_Jan23)         19-Dec-22 (No. DAE4-601_Dec22)         Check Date (in house)	nd are part of the certificate. C and humidity < 70%. <u>Scheduled Calibration</u> Mar-24 Mar-24 Mar-24 Mar-24 Mar-24 Jan-24
The measurements and the uncert All calibrations have been conductor Calibration Equipment used (M&TE Primary Standards Power meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B	ainties with confidence p ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103245 SN: 103245 SN: BH3934 (20k) SN: 310982 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475	Cal Date (Certificate No.)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03805)           30-Mar-23 (No. 217-03809)           30-Mar-23 (No. 217-03810)           10-Jan-23 (No. EX3-7349_Jan23)           19-Dec-22 (No. DAE4-601_Dec22)           Check Date (in house)           30-Oct-14 (in house check Oct-22)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration Mar-24 Mar-24 Mar-24 Mar-24 Jan-24 Jan-24 Jan-24 Jec-23 Scheduled Check In house check: Oct-24
The measurements and the uncert All calibrations have been conducts Calibration Equipment used (M&TE Primary Standards Power meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A	ainties with confidence p ed in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 103245 SN: 310982 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475 SN: US37292783	Cal Date (Certificate No.)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03805)           30-Mar-23 (No. 217-03805)           30-Mar-23 (No. 217-03809)           30-Mar-23 (No. 217-03809)           30-Mar-23 (No. 217-03809)           30-Mar-23 (No. 217-03809)           30-Mar-23 (No. 217-03810)           10-Jan-23 (No. EX3-7349_Jan23)           19-Dec-22 (No. DAE4-601_Dec22)           Check Date (in house)           30-Oct-14 (in house check Oct-22)           07-Oct-15 (in house check Oct-22)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration Mar-24 Mar-24 Mar-24 Mar-24 Mar-24 Jan-24 Jan-24 Dec-23 Scheduled Check In house check: Oct-24 In house check: Oct-24
The measurements and the uncert All calibrations have been conduct Calibration Equipment used (M&TE Primary Standards Power meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A	ainties with confidence p ed in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: BH9394 (20k) SN: 310982 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: MY41093315	Cal Date (Certificate No.)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03805)           30-Mar-23 (No. 217-03809)           30-Mar-23 (No. 217-03809)           30-Mar-23 (No. 217-03810)           10-Jan-23 (No. 217-03810)           10-Jan-23 (No. 217-03810)           10-Jan-23 (No. EX3-7349_Jan23)           19-Dec-22 (No. DAE4-601_Dec22)           Check Date (in house)           30-Oct-14 (in house check Oct-22)           07-Oct-15 (in house check Oct-22)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration Mar-24 Mar-24 Mar-24 Mar-24 Jan-24 Jan-24 Jec-23 Scheduled Check In house check: Oct-24 In house check: Oct-24 In house check: Oct-24
The measurements and the uncert All calibrations have been conduct Calibration Equipment used (M&TE Primary Standards Power meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06	ainties with confidence p ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: BH9394 (20k) SN: 310982 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: WY41093315 SN: 100972	Cal Date (Certificate No.)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03805)           30-Mar-23 (No. 217-03805)           30-Mar-23 (No. 217-03809)           30-Mar-23 (No. 217-03810)           10-Jan-23 (No. 217-03810)           10-Jan-23 (No. 217-03810)           10-Jan-23 (No. 217-03810)           10-Jan-23 (No. EX3-7349_Jan23)           19-Dec-22 (No. DAE4-601_Dec22)           Check Date (in house)           30-Oct-14 (in house check Oct-22)           07-Oct-15 (in house check Oct-22)           07-Oct-15 (in house check Oct-22)           15-Jun-15 (in house check Oct-22)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration Mar-24 Mar-24 Mar-24 Mar-24 Jan-24 Jan-24 Jec-23 Scheduled Check In house check: Oct-24 In house check: Oct-24
The measurements and the uncert All calibrations have been conduct Calibration Equipment used (M&TE Primary Standards Power meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A	ainties with confidence p ed in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: BH9394 (20k) SN: 310982 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: MY41093315	Cal Date (Certificate No.)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03805)           30-Mar-23 (No. 217-03809)           30-Mar-23 (No. 217-03809)           30-Mar-23 (No. 217-03810)           10-Jan-23 (No. 217-03810)           10-Jan-23 (No. 217-03810)           10-Jan-23 (No. EX3-7349_Jan23)           19-Dec-22 (No. DAE4-601_Dec22)           Check Date (in house)           30-Oct-14 (in house check Oct-22)           07-Oct-15 (in house check Oct-22)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration Mar-24 Mar-24 Mar-24 Mar-24 Jan-24 Jan-24 Jec-23 Scheduled Check In house check: Oct-24 In house check: Oct-24 In house check: Oct-24
The measurements and the uncert All calibrations have been conduct Calibration Equipment used (M&TE Primary Standards Power meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent E8358A	ainties with confidence p ed in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244 SN: 103244 SN: 103245 SN: 310982 / 06327 SN: 310982 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: WY41093315 SN: 100972 SN: US41080477 Name	Cal Date (Certificate No.)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03805)           30-Mar-23 (No. 217-03809)           30-Mar-23 (No. 217-03810)           10-Jan-23 (No. EX3-7349_Jan23)           19-Dec-22 (No. DAE4-601_Dec22)           Check Date (in house)           30-Oct-14 (in house check Oct-22)           07-Oct-15 (in house check Oct-22)           07-Oct-15 (in house check Oct-22)           15-Jun-15 (in house check Oct-22)           31-Mar-14 (in house check Oct-22)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration Mar-24 Mar-24 Mar-24 Mar-24 Jan-24 Jan-24 Jec-23 Scheduled Check In house check: Oct-24 In house check: Oct-24
The measurements and the uncert All calibrations have been conduct Calibration Equipment used (M&TE Primary Standards Power meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06	ainties with confidence p ed in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 8H9394 (20k) SN: 310982 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: WY41093315 SN: 100972 SN: US41080477	Cal Date (Certificate No.)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03805)           30-Mar-23 (No. 217-03809)           30-Mar-23 (No. 217-03810)           10-Jan-23 (No. EX3-7349_Jan23)           19-Dec-22 (No. DAE4-601_Dec22)           Check Date (in house)           30-Oct-14 (in house check Oct-22)           07-Oct-15 (in house check Oct-22)           07-Oct-15 (in house check Oct-22)           15-Jun-15 (in house check Oct-22)           31-Mar-14 (in house check Oct-22)	and are part of the certificate. C and humidity < 70%. Scheduled Calibration Mar-24 Mar-24 Mar-24 Mar-24 Jan-24 Jan-24 Dec-23 Scheduled Check In house check: Oct-24 In house check: Oct-24
The measurements and the uncert All calibrations have been conduct Calibration Equipment used (M&TE Primary Standards Power meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent E8358A	ainties with confidence p ed in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244 SN: 103244 SN: 103245 SN: 310982 / 06327 SN: 310982 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: WY41093315 SN: 100972 SN: US41080477 Name	Cal Date (Certificate No.)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804/03805)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03804)           30-Mar-23 (No. 217-03805)           30-Mar-23 (No. 217-03809)           30-Mar-23 (No. 217-03810)           10-Jan-23 (No. EX3-7349_Jan23)           19-Dec-22 (No. DAE4-601_Dec22)           Check Date (in house)           30-Oct-14 (in house check Oct-22)           07-Oct-15 (in house check Oct-22)           07-Oct-15 (in house check Oct-22)           15-Jun-15 (in house check Oct-22)           31-Mar-14 (in house check Oct-22)	and are part of the certificate. C and humidity < 70%. Scheduled Calibration Mar-24 Mar-24 Mar-24 Mar-24 Jan-24 Jan-24 Dec-23 Scheduled Check In house check: Oct-24 In house check: Oct-24





### Calibration Laboratory of Schmid & Partner Engineering AG

Zeughausstrasse 43, 8004 Zurich, Switzerland

 S
 Schweizerischer Kalibrierdienst

 C
 Service suisse d'étalonnage

 S
 Servizio svizzero di taratura

 Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

## Glossary:

TSL	tissue simulating liquid
ConvF	sensitivity in TSL / NORM x,y,z
N/A	not applicable or not measured

#### Calibration is Performed According to the Following Standards:

- a) IEC/IEEE 62209-1528, "Measurement Procedure For The Assessment Of Specific Absorption Rate Of Human Exposure To Radio Frequency Fields From Hand-Held And Body-Worn Wireless Communication Devices - Part 1528: Human Models, Instrumentation And Procedures (Frequency Range of 4 MHz to 10 GHz)", October 2020.
- b) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

#### Additional Documentation:

c) DASY System Handbook

# Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The source is mounted in a touch configuration below the center marking of the flat phantom.
- *Return Loss:* This parameter is measured with the source positioned under the liquid filled phantom (as described in the measurement condition clause). The Return Loss ensures low reflected power. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: D1900V2-5d101\_Jul23

Page 2 of 6