

HAC_E_Dipole_1880**DUT: HAC Dipole 1880 MHz**

Communication System: CW; Frequency: 1880 MHz; Duty Cycle: 1:1

Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.5 °C

DASY5 Configuration:

- Probe: EF3DV3 - SN4047; ConvF(1, 1, 1) @ 1880 MHz; Calibrated: 2019/1/30

- Sensor-Surface: (Fix Surface)

- Electronics: DAE4 Sn854; Calibrated: 2018/6/14

- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;

- Measurement SW: DASY52, Version 52.10 (1); SEMCAD X Version 14.6.11 (7439)

E Scan - measurement distance from the probe sensor center to CD1880 = 10mm & 15mm/Hearing Aid Compatibility Test at 15mm distance (41x181x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 158.3 V/m; Power Drift = 0.00 dB

PMR not calibrated. PMF = 1.000 is applied.

E-field emissions = 87.45 V/m

Average value of Total=(86.98+87.45) / 2 = 87.22 V/m

PMF scaled E-field


Grid 1 M3 86.35 V/m	Grid 2 M3 86.98 V/m	Grid 3 M3 83.85 V/m
Grid 4 M3 64.67 V/m	Grid 5 M3 64.73 V/m	Grid 6 M3 63.17 V/m
Grid 7 M3 86.75 V/m	Grid 8 M3 87.45 V/m	Grid 9 M3 84.22 V/m

Cursor:

Total = 87.45 V/m

E Category: M3

Location: 1, 34, 9.7 mm



0 dB = 87.45 V/m = 38.84 dBV/m