

Appendix C for KSCR230600109703

Calibration Certificate

Object	Apply	No	Model	SN	Calibration Date	Due date of calibration
Dipole	<input type="checkbox"/>	1	CLA150	4025	2021/04/26	2024/04/25
	<input type="checkbox"/>	2	D450V3	1103	2021/04/21	2024/04/20
	<input type="checkbox"/>	3	D750V3	1188	2022/03/29	2025/03/28
	<input type="checkbox"/>	4	D835V2	4d114	2022/03/31	2025/03/30
	<input type="checkbox"/>	5	D900V2	1d079	2022/06/07	2025/06/06
	<input type="checkbox"/>	6	D1800V2	2d170	2022/03/31	2025/03/30
	<input type="checkbox"/>	7	D1900V2	5d136	2022/06/07	2025/06/06
	<input type="checkbox"/>	8	D2000V2	1041	2022/06/06	2025/06/05
	<input type="checkbox"/>	9	D2300V2	1096	2022/03/31	2025/03/30
	<input checked="" type="checkbox"/>	10	D2450V2	817	2022/04/01	2025/03/31
	<input type="checkbox"/>	11	D2600V2	1158	2022/03/31	2025/03/30
	<input type="checkbox"/>	12	D5GHzV2	1095	2022/06/01	2025/05/31
DAE	<input checked="" type="checkbox"/>	13	DAE4	1245	2023/04/25	2024/04/24
Probe	<input checked="" type="checkbox"/>	14	EX3DV4	7767	2023/10/26	2024/10/25

1 Dipole

1.1 CLA150 - SN 4025

<p>Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland</p> <p>Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p>Client: SGS-CN (Auden) Certificate No: CLA150-4025_Apr21</p> <p style="text-align: right;">Accreditation No.: SCS 0108</p> <hr/> <p style="text-align: center;">CALIBRATION CERTIFICATE</p> <p>Object: CLA150 - SN: 4025</p> <p>Calibration procedure(s): QA CAL-15.v9 Calibration Procedure for SAR Validation Sources below 700 MHz</p> <p>Calibration date: April 26, 2021</p> <p>The calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%.</p> <p>Calibration Equipment Used (M&E critical for calibration)</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Certificate No.)</th> <th>Schedule / Calibration</th> </tr> </thead> <tbody> <tr> <td>Power meter NRP</td> <td>SN: 104736</td> <td>09-Apr-21 (No. 217-03201.00292)</td> <td>Apr-22</td> </tr> <tr> <td>Power sensor NRP Z01</td> <td>SN: 103344</td> <td>09-Apr-21 (No. 217-03201)</td> <td>Apr-22</td> </tr> <tr> <td>Power sensor NRP Z01</td> <td>SN: 103245</td> <td>09-Apr-21 (No. 217-03202)</td> <td>Apr-22</td> </tr> <tr> <td>Reference 20 dB Attenuator</td> <td>SN: C22962 (20)</td> <td>09-Apr-21 (No. 217-03343)</td> <td>Apr-22</td> </tr> <tr> <td>Type-N mismatch combination</td> <td>SN: 310952 / 00327</td> <td>09-Apr-21 (No. 217-03344)</td> <td>Apr-22</td> </tr> <tr> <td>Reference Probe EX3004 (DIE4)</td> <td>SN: 3877</td> <td>30-Dec-20 (No. EX3-3877_Dec20)</td> <td>Dec-21</td> </tr> <tr> <td></td> <td>SN: 664</td> <td>26-Jun-20 (No. DMS4-656_Jun20)</td> <td>Jun-21</td> </tr> </tbody> </table> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Check Date (in house)</th> <th>Scheduled Check</th> </tr> </thead> <tbody> <tr> <td>Power meter E4118B</td> <td>SN: G814282874</td> <td>06-Apr-18 (in house check Jun-20)</td> <td>in house check Jun-22</td> </tr> <tr> <td>Power sensor E4112A</td> <td>SN: MY41498087</td> <td>06-Apr-18 (in house check Jun-20)</td> <td>in house check Jun-22</td> </tr> <tr> <td>Power sensor E4112A</td> <td>SN: 000100210</td> <td>06-Apr-18 (in house check Jun-20)</td> <td>in house check Jun-22</td> </tr> <tr> <td>RF generator HP 85940D</td> <td>SN: U53840310700</td> <td>04-Aug-09 (in house check Jun-20)</td> <td>in house check Jun-22</td> </tr> <tr> <td>Network Analyzer Agilent E8363A</td> <td>SN: U541000477</td> <td>31-Mar-14 (in house check Oct-20)</td> <td>in house check Oct-21</td> </tr> </tbody> </table> <p>Calibrated by: Jeffrey Katzman Function: Laboratory Technician Signature: <i>[Signature]</i></p> <p>Approved by: Katja Polovic Technical Manager <i>[Signature]</i></p> <p style="text-align: right;">Issued: April 26, 2021</p> <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p> <p>Certificate No: CLA150-4025_Apr21 Page 1 of 6</p>	Primary Standards	ID #	Cal Date (Certificate No.)	Schedule / Calibration	Power meter NRP	SN: 104736	09-Apr-21 (No. 217-03201.00292)	Apr-22	Power sensor NRP Z01	SN: 103344	09-Apr-21 (No. 217-03201)	Apr-22	Power sensor NRP Z01	SN: 103245	09-Apr-21 (No. 217-03202)	Apr-22	Reference 20 dB Attenuator	SN: C22962 (20)	09-Apr-21 (No. 217-03343)	Apr-22	Type-N mismatch combination	SN: 310952 / 00327	09-Apr-21 (No. 217-03344)	Apr-22	Reference Probe EX3004 (DIE4)	SN: 3877	30-Dec-20 (No. EX3-3877_Dec20)	Dec-21		SN: 664	26-Jun-20 (No. DMS4-656_Jun20)	Jun-21	Secondary Standards	ID #	Check Date (in house)	Scheduled Check	Power meter E4118B	SN: G814282874	06-Apr-18 (in house check Jun-20)	in house check Jun-22	Power sensor E4112A	SN: MY41498087	06-Apr-18 (in house check Jun-20)	in house check Jun-22	Power sensor E4112A	SN: 000100210	06-Apr-18 (in house check Jun-20)	in house check Jun-22	RF generator HP 85940D	SN: U53840310700	04-Aug-09 (in house check Jun-20)	in house check Jun-22	Network Analyzer Agilent E8363A	SN: U541000477	31-Mar-14 (in house check Oct-20)	in house check Oct-21	<p>Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland</p> <p>Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p>Accreditation No.: SCS 0108</p> <hr/> <p>Glossary:</p> <p>TSL: Issue simulating liquid sensitivity in TSL / NORM x,y,z</p> <p>ConvF: not applicable or not measured</p> <p>N/A: not applicable or not measured</p> <p>Calibration is Performed According to the Following Standards:</p> <ol style="list-style-type: none"> IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Technique", June 2013 IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016 IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010 KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz" <p>Additional Documentation:</p> <ol style="list-style-type: none"> DASY4/5 System Handbook <p>Methods Applied and Interpretation of Parameters:</p> <ul style="list-style-type: none"> Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in this certificate are valid at the frequency indicated. Antenna Parameters with TSL: The source is mounted in a touch configuration below the center marking of the flat phantom. Return Loss: This parameter is measured with the source positioned under the liquid filled phantom (as described in the measurement condition clause). The Return Loss ensures low reflected power. No uncertainty required. SAR measured: SAR measured at the stated antenna input power. SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector. SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result. <div style="border: 1px solid black; padding: 5px; margin-top: 10px;"> <p>The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.</p> </div> <p>Certificate No: CLA150-4025_Apr21 Page 2 of 6</p>						
Primary Standards	ID #	Cal Date (Certificate No.)	Schedule / Calibration																																																												
Power meter NRP	SN: 104736	09-Apr-21 (No. 217-03201.00292)	Apr-22																																																												
Power sensor NRP Z01	SN: 103344	09-Apr-21 (No. 217-03201)	Apr-22																																																												
Power sensor NRP Z01	SN: 103245	09-Apr-21 (No. 217-03202)	Apr-22																																																												
Reference 20 dB Attenuator	SN: C22962 (20)	09-Apr-21 (No. 217-03343)	Apr-22																																																												
Type-N mismatch combination	SN: 310952 / 00327	09-Apr-21 (No. 217-03344)	Apr-22																																																												
Reference Probe EX3004 (DIE4)	SN: 3877	30-Dec-20 (No. EX3-3877_Dec20)	Dec-21																																																												
	SN: 664	26-Jun-20 (No. DMS4-656_Jun20)	Jun-21																																																												
Secondary Standards	ID #	Check Date (in house)	Scheduled Check																																																												
Power meter E4118B	SN: G814282874	06-Apr-18 (in house check Jun-20)	in house check Jun-22																																																												
Power sensor E4112A	SN: MY41498087	06-Apr-18 (in house check Jun-20)	in house check Jun-22																																																												
Power sensor E4112A	SN: 000100210	06-Apr-18 (in house check Jun-20)	in house check Jun-22																																																												
RF generator HP 85940D	SN: U53840310700	04-Aug-09 (in house check Jun-20)	in house check Jun-22																																																												
Network Analyzer Agilent E8363A	SN: U541000477	31-Mar-14 (in house check Oct-20)	in house check Oct-21																																																												
<p>Measurement Conditions DASY system configuration, as far as not given on page 1.</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>DASY Version</th> <th>DASY5</th> <th>V52.10.4</th> </tr> </thead> <tbody> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> <td></td> </tr> <tr> <td>Phantom</td> <td>ELN Flat Phantom</td> <td>Shell thickness: 2 ± 0.2 mm</td> </tr> <tr> <td>EUT Positioning</td> <td>Touch Position</td> <td></td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy = 4.0 mm, dz = 1.4 mm</td> <td>Graded Ratio = 1.4 (Z direction)</td> </tr> <tr> <td>Frequency</td> <td>150 MHz ± 1 MHz</td> <td></td> </tr> </tbody> </table> <p>Head TSL parameters The following parameters and calculations were applied.</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Parameter</th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>52.3</td> <td>0.75 mho/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>51.1 ± 6 %</td> <td>0.75 mho/m ± 6 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td>< 0.5 °C</td> <td>---</td> <td>---</td> </tr> </tbody> </table> <p>SAR result with Head TSL</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>SAR averaged over 1 cm³ (1 g) of Head TSL</th> <th>Condition</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>1 W input power</td> <td>3.90 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>3.88 W/kg ± 18.4 % (k=2)</td> </tr> </tbody> </table> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>SAR averaged over 10 cm³ (10 g) of Head TSL</th> <th>condition</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>1 W input power</td> <td>2.60 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>2.59 W/kg ± 18.0 % (k=2)</td> </tr> </tbody> </table> <p>Certificate No: CLA150-4025_Apr21 Page 3 of 6</p>	DASY Version	DASY5	V52.10.4	Extrapolation	Advanced Extrapolation		Phantom	ELN Flat Phantom	Shell thickness: 2 ± 0.2 mm	EUT Positioning	Touch Position		Zoom Scan Resolution	dx, dy = 4.0 mm, dz = 1.4 mm	Graded Ratio = 1.4 (Z direction)	Frequency	150 MHz ± 1 MHz		Parameter	Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	52.3	0.75 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	51.1 ± 6 %	0.75 mho/m ± 6 %	Head TSL temperature change during test	< 0.5 °C	---	---	SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	Value	SAR measured	1 W input power	3.90 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	3.88 W/kg ± 18.4 % (k=2)	SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	Value	SAR measured	1 W input power	2.60 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	2.59 W/kg ± 18.0 % (k=2)	<p>Appendix (Additional assessments outside the scope of SCS 0108)</p> <p>Antenna Parameters with Head TSL</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Parameter</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>Impedance, transformed to feed point</td> <td>47.8 Ω ± 1.5 Ω</td> </tr> <tr> <td>Return Loss</td> <td>-31.4 dB</td> </tr> </tbody> </table> <p>Additional EUT Data</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Parameter</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>Manufactured by</td> <td>SPEAG</td> </tr> </tbody> </table> <p>Certificate No: CLA150-4025_Apr21 Page 4 of 6</p>	Parameter	Value	Impedance, transformed to feed point	47.8 Ω ± 1.5 Ω	Return Loss	-31.4 dB	Parameter	Value	Manufactured by	SPEAG
DASY Version	DASY5	V52.10.4																																																													
Extrapolation	Advanced Extrapolation																																																														
Phantom	ELN Flat Phantom	Shell thickness: 2 ± 0.2 mm																																																													
EUT Positioning	Touch Position																																																														
Zoom Scan Resolution	dx, dy = 4.0 mm, dz = 1.4 mm	Graded Ratio = 1.4 (Z direction)																																																													
Frequency	150 MHz ± 1 MHz																																																														
Parameter	Temperature	Permittivity	Conductivity																																																												
Nominal Head TSL parameters	22.0 °C	52.3	0.75 mho/m																																																												
Measured Head TSL parameters	(22.0 ± 0.2) °C	51.1 ± 6 %	0.75 mho/m ± 6 %																																																												
Head TSL temperature change during test	< 0.5 °C	---	---																																																												
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	Value																																																													
SAR measured	1 W input power	3.90 W/kg																																																													
SAR for nominal Head TSL parameters	normalized to 1W	3.88 W/kg ± 18.4 % (k=2)																																																													
SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	Value																																																													
SAR measured	1 W input power	2.60 W/kg																																																													
SAR for nominal Head TSL parameters	normalized to 1W	2.59 W/kg ± 18.0 % (k=2)																																																													
Parameter	Value																																																														
Impedance, transformed to feed point	47.8 Ω ± 1.5 Ω																																																														
Return Loss	-31.4 dB																																																														
Parameter	Value																																																														
Manufactured by	SPEAG																																																														

DASY5 Validation Report for Head TSL

Date: 26.04.2021

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: CLA150; Type: CLA150; Serial: CLA150 - SN: 4025

Communication System: UID 0 - CW; Frequency: 150 MHz
 Medium parameters used: $f = 150 \text{ MHz}$; $\sigma = 0.76 \text{ S/m}$; $\epsilon = 51.1$; $\rho = 1000 \text{ kg/m}^3$
 Phantom section: Flat Section
 Measurement Standard: DASY5 (IEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 - SN3877; ConvF(12.51, 12.51, 12.51) @ 150 MHz; Calibrated: 30.12.2020
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DA54 Snt54; Calibrated: 26.06.2020
- Phantom: ELI v4.0; Type: QDOVA001BB; Serial: TP.1003
- DASY52 52.10.4(1527); SEMCAD X 14.6.14(7483)

CLA Calibration for HSL-LF Tissue/CLA150, touch configuration, Pin=1W/Zoom Scan, dist=1.4mm (8x10x8)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
 Reference Value = 85.93 W/m; Power Drift = -0.02 dB
 Peak SAR (extrapolated) = 7.36 W/kg
SAR(1 g) = 3.90 W/kg; SAR(10 g) = 2.60 W/kg
 Smallest distance from peaks to all points 3 dB below: Larger than measurement grid (> 30mm)
 Ratio of SAR at M2 to SAR at M1 = 80.4%
 Maximum value of SAR (measured) = 5.48 W/kg

Certificate No: CLA150-4025_Apr21 Page 5 of 6

Certificate No: CLA150-4025_Apr21 Page 6 of 6

1.2 D450V3 - SN 1103

Calibration Laboratory of Schmid & Partner Engineering AG
 Zeughausstrasse 43, 8004 Zurich, Switzerland

Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client: **SGS-CN (Aude)** Certificate No: **D450V3-1103_Apr21**

CALIBRATION CERTIFICATE

Object: **D450V3 - SN:1103**

Calibration procedure(s): **QA CAL-15_v9**
 Calibration Procedure for SAR Validation Sources below 700 MHz

Calibration date: **April 21, 2021**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility; environment temperature (22 ± 0.1) °C and humidity < 70%.

Calibration Equipment used (MPE critical for calibration)

Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NRP	SN: 104778	09-Apr-21 (No. 217-03021/03030)	Apr-22
Power sensor NRP-291	SN: 103244	09-Apr-21 (No. 217-03021)	Apr-22
Power sensor NRP-291	SN: 103245	09-Apr-21 (No. 217-03020)	Apr-22
Reference 20 dB Attenuator	SN: CC2852 (200)	09-Apr-21 (No. 217-03345)	Apr-22
Type-N mission combination	SN: 310982 / 06327	09-Apr-21 (No. 217-03344)	Apr-22
Reference Probe E3030A	SN: 3877	30-Dec-20 (No. E303-2077_Decl2)	Dec-21
DAEA	SN: 654	26-Jan-20 (No. DAE4-654_Jan20)	Jan-21

Secondary Standards	ID #	Check Date (in house)	Scheduled Check
Power meter E4418B	SN: GB41200274	06-Apr-16 (in house check Jun-20)	In house check Jun-22
Power sensor E4412A	SN: MY41496027	06-Apr-16 (in house check Jun-20)	In house check Jun-22
Power sensor E4412A	SN: 000100210	06-Apr-16 (in house check Jun-20)	In house check Jun-22
RF generator HP 8648C	SN: U533400101700	04-Aug-09 (in house check Jun-20)	In house check Jun-22
Network Analyzer Agilent E8358A	SN: U541980427	31-Mar-14 (in house check Oct-20)	In house check Oct-21

Calibrated by: **Christof Leuber** (Function: Laboratory Technician)

Approved by: **Katja Polovic** (Function: Technical Manager)

This calibration certificate shall not be reproduced except in full without written approval of the laboratory. Issued: April 23, 2021

Certificate No: D450V3-1103_Apr21 Page 1 of 6

Calibration Laboratory of Schmid & Partner Engineering AG
 Zeughausstrasse 43, 8004 Zurich, Switzerland

Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0106**

Glossary:

TSL: Issue simulating liquid
 ConvF: sensitivity in TSL / NORM x,y,z
 N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices; Measurement Techniques", June 2013
- IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016
- IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- KDB 665664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

- DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k=2$, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: D450V3-1103_Apr21 Page 2 of 6

Measurement Conditions
 DASY5 system configuration, as far as not given on page 1.

DASY Version	DASY5	V82.10.4
Extrapolation	Advanced Extrapolation	
Phantom	ELJ4 Flat Phantom	Shell thickness: 2 ± 0.2 mm
Distance Dipole Center - TSL	15 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	450 MHz ± 1 MHz	

Head TSL parameters
 The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	43.5	0.57 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	43.1 ± 6 %	0.67 mho/m ± 8 %
Head TSL temperature change during test	< 0.5 °C	---	---

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	1.14 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	4.56 W/kg ± 18.1 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	250 mW input power	0.757 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	3.06 W/kg ± 17.6 % (k=2)

Certificate No: D450V3-1103_Apr21 Page 3 of 6

Appendix (Additional assessments outside the scope of SCS 0106)

Antenna Parameters with Head TSL

Impedance, transformed to feed point	57.1 Ω - 2.6 jΩ
Return Loss	-23.0 dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.346 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.
 The dipole is made of standard straight coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is set according to the Standard.
 No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
-----------------	-------

Certificate No: D450V3-1103_Apr21 Page 4 of 6

DASY5 Validation Report for Head TSL

Test Laboratory: SPEAG, Zurich, Switzerland Date: 21.04.2021

DUT: Dipole 450 MHz; Type: D450V3; Serial: D450V3 - SN:1103

Communication System: UID 0 - CW; Frequency: 450 MHz
 Medium parameters used: f = 450 MHz, α = 0.87 S/m; ε = 43.1; ρ = 1000 kg/m³
 Phantom section: Flat Section
 Measurement Standard: DASY5 (IEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 - SN3877; ConvF(10.64, 10.64, 10.64) @ 450 MHz; Calibrated: 30.12.2020
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn654; Calibrated: 26.06.2020
- Phantom: ELJ v4.0; Type: QDOVA001BB; Serial: TP:1003
- DASY52 52.10.4(1527); SEMCAD X 14.6.14(7483)

Dipole Calibration for Head Tissue/d=15mm, Pin=250mW/Zoom Scan (7x7x7)/Cube 0:
 Measurement grid: dx=5mm, dy=5mm, dz=5mm
 Reference Value = 39.18 V/m; Power Drift = -0.08 dB
 Peak SAR (extrapolated) = 1.76 W/kg
 SAR(1 g) = 1.14 W/kg; SAR(10 g) = 0.767 W/kg
 Smallest distance from peaks to all points 3 dB below: Larger than measurement grid
 Ratio of SAR at M2 to SAR at M1 = 64.9%
 Maximum value of SAR (measured) = 1.53 W/kg

0 dB = 1.53 W/kg = 1.85 dBW/kg

Certificate No: D450V3-1103_Apr21 Page 5 of 6

Certificate No: D450V3-1103_Apr21 Page 6 of 6

1.3 D750V3 - SN 1188

<div style="text-align: center;"> </div> <p style="font-size: small;"> Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-4236633-2112 Fax: +86-10-4236633-2504 E-mail: cti@chinaast.com http://www.chinaast.cn </p> <p> Client: SGS-CN Certificate No: Z22-60103 </p> <h3 style="text-align: center;">CALIBRATION CERTIFICATE</h3> <p> Object: D750V3 - SN: 1188 </p> <p> Calibration Procedure(s): FF-Z11-003-01 Calibration Procedures for dipole validation kits </p> <p> Calibration date: March 28, 2022 </p> <p> This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. </p> <p> All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity<70%. </p> <p> Calibration Equipment used (M&TE critical for calibration) </p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power Meter NRP2</td> <td>102277</td> <td>24-Sep-21 (CTTL No.J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Power sensor NRP88</td> <td>104291</td> <td>24-Sep-21 (CTTL No.J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Reference Probe EX3DV4</td> <td>SN 7307</td> <td>26-May-21(SPEAG.No.EX3-7307_May21)</td> <td>May-22</td> </tr> <tr> <td>DAE4</td> <td>SN 1556</td> <td>12-Jan-22(CTTL-SPEAG.No.Z22-60007)</td> <td>Jan-23</td> </tr> </tbody> </table> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Signal Generator E4439C</td> <td>MY49071430</td> <td>13-Jan-22 (CTTL No.J22X00409)</td> <td>Jan-23</td> </tr> <tr> <td>Network Analyzer E5071C</td> <td>MY46110973</td> <td>14-Jan-22 (CTTL No.J22X00409)</td> <td>Jan-23</td> </tr> </tbody> </table> <p> Calibrated by: Zhao Jing SAR Test Engineer </p> <p> Reviewed by: Lin Hao SAR Test Engineer </p> <p> Approved by: Qi Dianyuan SAR Project Leader </p> <p style="text-align: right;"> Issued: April 3, 2022 </p> <p style="font-size: x-small;"> This calibration certificate shall not be reproduced except in full without written approval of the laboratory. </p> <p style="font-size: x-small;"> Certificate No: Z22-60103 Page 1 of 6 </p>	Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Power Meter NRP2	102277	24-Sep-21 (CTTL No.J21X08326)	Sep-22	Power sensor NRP88	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22	Reference Probe EX3DV4	SN 7307	26-May-21(SPEAG.No.EX3-7307_May21)	May-22	DAE4	SN 1556	12-Jan-22(CTTL-SPEAG.No.Z22-60007)	Jan-23	Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Signal Generator E4439C	MY49071430	13-Jan-22 (CTTL No.J22X00409)	Jan-23	Network Analyzer E5071C	MY46110973	14-Jan-22 (CTTL No.J22X00409)	Jan-23	<div style="text-align: center;"> </div> <p style="font-size: small;"> Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-4236633-2079 Fax: +86-10-4236633-2504 E-mail: cti@chinaast.com http://www.chinaast.cn </p> <p> Glossary: TSL: tissue simulating liquid ConvF: sensitivity in TSL / NORMx.yz N/A: not applicable or not measured </p> <p> Calibration is Performed According to the Following Standards: a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices-Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020 b) KDB 865684, "SAR Measurement Requirements for 100 MHz to 6 GHz" </p> <p> Additional Documentation: c) DASy4/5 System Handbook </p> <p> Methods Applied and Interpretation of Parameters: </p> <ul style="list-style-type: none"> Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated. Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis. Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required. Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required. SAR measured: SAR measured at the stated antenna input power. SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector. SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result. <div style="border: 1px solid black; padding: 5px; font-size: x-small;"> The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%. </div> <p style="font-size: x-small;"> Certificate No: Z22-60103 Page 2 of 6 </p>																												
Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																										
Power Meter NRP2	102277	24-Sep-21 (CTTL No.J21X08326)	Sep-22																																																										
Power sensor NRP88	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22																																																										
Reference Probe EX3DV4	SN 7307	26-May-21(SPEAG.No.EX3-7307_May21)	May-22																																																										
DAE4	SN 1556	12-Jan-22(CTTL-SPEAG.No.Z22-60007)	Jan-23																																																										
Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																										
Signal Generator E4439C	MY49071430	13-Jan-22 (CTTL No.J22X00409)	Jan-23																																																										
Network Analyzer E5071C	MY46110973	14-Jan-22 (CTTL No.J22X00409)	Jan-23																																																										
<div style="text-align: center;"> </div> <p style="font-size: small;"> Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-4236633-2079 Fax: +86-10-4236633-2504 E-mail: cti@chinaast.com http://www.chinaast.cn </p> <p> Measurement Conditions DASy system configuration, as far as not given on page 1. </p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tbody> <tr> <td>DASY Version</td> <td>DASY52</td> <td>V52.10.4</td> </tr> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> <td></td> </tr> <tr> <td>Phantom</td> <td>Triple Flat Phantom 5.1C</td> <td></td> </tr> <tr> <td>Distance Dipole Center - TSL</td> <td>15 mm</td> <td>with Spacer</td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy, dz = 5 mm</td> <td></td> </tr> <tr> <td>Frequency</td> <td>750 MHz ± 1 MHz</td> <td></td> </tr> </tbody> </table> <p> Head TSL parameters The following parameters and calculations were applied. </p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th></th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>42.0</td> <td>0.90 mho/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>41.4 ± 0.6 %</td> <td>0.89 mho/m ± 6 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td><1.0 °C</td> <td>---</td> <td>---</td> </tr> </tbody> </table> <p> SAR result with Head TSL </p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>SAR averaged over 1 cm² (1 g) of Head TSL</th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>2.07 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>8.27 W/kg ± 18.8 % (k=2)</td> </tr> <tr> <th>SAR averaged over 10 cm² (10 g) of Head TSL</th> <th>Condition</th> <th></th> </tr> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>1.37 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>5.48 W/kg ± 18.7 % (k=2)</td> </tr> </tbody> </table> <p style="font-size: x-small;"> Certificate No: Z22-60103 Page 3 of 6 </p>	DASY Version	DASY52	V52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	15 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm		Frequency	750 MHz ± 1 MHz			Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	42.0	0.90 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	41.4 ± 0.6 %	0.89 mho/m ± 6 %	Head TSL temperature change during test	<1.0 °C	---	---	SAR averaged over 1 cm ² (1 g) of Head TSL	Condition		SAR measured	250 mW input power	2.07 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	8.27 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm ² (10 g) of Head TSL	Condition		SAR measured	250 mW input power	1.37 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	5.48 W/kg ± 18.7 % (k=2)	<div style="text-align: center;"> </div> <p style="font-size: small;"> Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-4236633-2079 Fax: +86-10-4236633-2504 E-mail: cti@chinaast.com http://www.chinaast.cn </p> <p> Appendix (Additional assessments outside the scope of CNAS L0570) </p> <p> Antenna Parameters with Head TSL </p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tbody> <tr> <td>Impedance, transformed to feed point</td> <td>53.60- 1.13jΩ</td> </tr> <tr> <td>Return Loss</td> <td>-28.7dB</td> </tr> </tbody> </table> <p> General Antenna Parameters and Design </p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tbody> <tr> <td>Electrical Delay (one direction)</td> <td>0.947 ns</td> </tr> </tbody> </table> <p> After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured. </p> <p> The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged. </p> <p> Additional EUT Data </p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tbody> <tr> <td>Manufactured by</td> <td>SPEAG</td> </tr> </tbody> </table> <p style="font-size: x-small;"> Certificate No: Z22-60103 Page 4 of 6 </p>	Impedance, transformed to feed point	53.60- 1.13jΩ	Return Loss	-28.7dB	Electrical Delay (one direction)	0.947 ns	Manufactured by	SPEAG
DASY Version	DASY52	V52.10.4																																																											
Extrapolation	Advanced Extrapolation																																																												
Phantom	Triple Flat Phantom 5.1C																																																												
Distance Dipole Center - TSL	15 mm	with Spacer																																																											
Zoom Scan Resolution	dx, dy, dz = 5 mm																																																												
Frequency	750 MHz ± 1 MHz																																																												
	Temperature	Permittivity	Conductivity																																																										
Nominal Head TSL parameters	22.0 °C	42.0	0.90 mho/m																																																										
Measured Head TSL parameters	(22.0 ± 0.2) °C	41.4 ± 0.6 %	0.89 mho/m ± 6 %																																																										
Head TSL temperature change during test	<1.0 °C	---	---																																																										
SAR averaged over 1 cm ² (1 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	2.07 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	8.27 W/kg ± 18.8 % (k=2)																																																											
SAR averaged over 10 cm ² (10 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	1.37 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	5.48 W/kg ± 18.7 % (k=2)																																																											
Impedance, transformed to feed point	53.60- 1.13jΩ																																																												
Return Loss	-28.7dB																																																												
Electrical Delay (one direction)	0.947 ns																																																												
Manufactured by	SPEAG																																																												

TTL Speaq CALIBRATION LABORATORY
 In Collaboration with **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62506633-2079 Fax: +86-10-62506633-2564
 E-mail: cti@china.ttl.com http://www.china.ttl.com

DASY5 Validation Report for Head TSL Date: 2022-03-29
 Test Laboratory: CTTL, Beijing, China
DUT: Dipole 750 MHz; Type: D750V3; Serial: D750V3 - SN: 1188
 Communication System: UID 0, CW; Frequency: 750 MHz; Duty Cycle: 1:1
 Medium parameters used: $f = 750 \text{ MHz}$; $\sigma = 0.888 \text{ S/m}$; $\epsilon_r = 41.36$; $\rho = 1000 \text{ kg/m}^3$
 Phantom section: Right Section
 Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)
 DASY5 Configuration:

- Probe: EX3DV4 - SN7307; ConvF(10.31, 10.31, 10.31) @ 750 MHz; Calibrated: 2021-05-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1556; Calibrated: 2022-01-12
- Phantom: MFP-V5.1C (20kg probe tilt); Type: QD 000 P51 Cx; Serial: 1062
- DASY52.52.10.4(1535); SEMCAD X 14.6.14(7501)

Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm
 Reference Value = 55.06 V/m; Power Drift = 0.00 dB
 Peak SAR (extrapolated) = 3.07 W/kg
 SAR(1 g) = 2.07 W/kg; SAR(10 g) = 1.37 W/kg
 Smallest distance from peaks to all points 3 dB below = 18.9 mm
 Ratio of SAR at M2 to SAR at M1 = 67.1%
 Maximum value of SAR (measured) = 2.74 W/kg

Certificate No: Z22-60103 Page 5 of 6

TTL Speaq CALIBRATION LABORATORY
 In Collaboration with **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62506633-2079 Fax: +86-10-62506633-2564
 E-mail: cti@china.ttl.com http://www.china.ttl.com

Impedance Measurement Plot for Head TSL

Certificate No: Z22-60103 Page 6 of 6

1.4 D835V2 - SN 4d114

TTL Speaq CALIBRATION LABORATORY
 In Collaboration with **CAICT** and **CNAS**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191
 Tel: +86-10-62506633-2079 Fax: +86-10-62506633-2564
 E-mail: cti@china.ttl.com http://www.china.ttl.com

Client: **SGS-CN** Certificate No: **Z22-60104**

CALIBRATION CERTIFICATE

Object: **D835V2 - SN: 4d114**

Calibration Procedure(s): **FF-Z11-003-01**
 Calibration Procedures for dipole validation kits

Calibration date: **March 31, 2022**

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity < 70%.

Calibration Equipment used (M&E critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration
Power Meter NRP2	106277	24-Sep-21 (CTTL, No.J21X08326)	Sep-22
Power sensor NRPBS	104291	24-Sep-21 (CTTL, No.J21X08326)	Sep-22
Reference Probe EX3DV4	SN 7307	26-May-21(SPEAG.No.EX3-7307_May21)	May-22
DAE4	SN 1556	12-Jan-22(CTTL-SPEAG.No.Z22-60007)	Jan-23

Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL, No.J22X00409)	Jan-23
Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL, No.J22X00406)	Jan-23

Calibrated by: **Zhao Jing** SAR Test Engineer
 Reviewed by: **Lin Hao** SAR Test Engineer
 Approved by: **Qi Dianyuan** SAR Project Leader

Issued: April 6, 2022
 This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60104 Page 1 of 6

TTL Speaq CALIBRATION LABORATORY
 In Collaboration with **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62506633-2079 Fax: +86-10-62506633-2564
 E-mail: cti@china.ttl.com http://www.china.ttl.com

Glossary:
 TSL: tissue simulating liquid
 ConvF: sensitivity in TSL / NORMx.yz
 N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:
 a) IEC/IEEE 62208-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020
 b) KDB 685864, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:
 c) DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60104 Page 2 of 6

In Collaboration with **TTL Speag** CALIBRATION LABORATORY **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504
 E-mail: cti@china.ttl.com http://www.china.ttl.com

Measurement Conditions
 DASY system configuration, as far as not given on page 1.

DASY Version	DASY52	VS2 10.4
Extrapolation	Advanced Extrapolation	
Phantom	Triple Flat Phantom 5.1C	
Distance Dipole Center - TSL	15 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	835 MHz ± 1 MHz	

Head TSL parameters
 The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	41.5	0.90 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	41.0 ± 5 %	0.91 mho/m ± 8 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	2.37 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	9.40 W/kg ± 18.6 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	250 mW input power	1.54 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	6.12 W/kg ± 18.7 % (k=2)

Certificate No: Z22-60104 Page 3 of 6

In Collaboration with **TTL Speag** CALIBRATION LABORATORY **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504
 E-mail: cti@china.ttl.com http://www.china.ttl.com

Appendix (Additional assessments outside the scope of CNAS L0570)

Antenna Parameters with Head TSL

Impedance, transformed to feed point	48.70 - 5.22jΩ
Return Loss	-25.3dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.307 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.

Additional EUT Data

Manufactured by	SPEAG
-----------------	-------

Certificate No: Z22-60104 Page 4 of 6

In Collaboration with **TTL Speag** CALIBRATION LABORATORY **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504
 E-mail: cti@china.ttl.com http://www.china.ttl.com

DASY5 Validation Report for Head TSL Date: 2022-03-21
 Test Laboratory: CTTI, Beijing, China
 DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN: 4d114
 Communication System: UID 0, CW; Frequency: 835 MHz; Duty Cycle: 1:1
 Medium parameters used: f = 835 MHz; σ = 0.907 S/m; ε_r = 40.98; ρ = 1000 kg/m³
 Phantom section: Right Section
 Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)
 DASY5 Configuration:

- Probe: EX3DV4 - SN7307; ConvF(10.13, 10.13, 10.13) @ 835 MHz; Calibrated: 2021-05-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DA E4 Sn 1556; Calibrated: 2022-01-12
- Phantom: MFP_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062
- DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501)

Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm
 Reference Value = 57.88 V/m; Power Drift = 0.04 dB
 Peak SAR (extrapolated) = 3.56 W/kg
SAR(1 g) = 2.37 W/kg; SAR(10 g) = 1.54 W/kg
 Smallest distance from peaks to all points 3 dB below = 15.8 mm
 Ratio of SAR at M2 to SAR at M1 = 66.2%
 Maximum value of SAR (measured) = 3.17 W/kg

Certificate No: Z22-60104 Page 5 of 6

In Collaboration with **TTL Speag** CALIBRATION LABORATORY **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504
 E-mail: cti@china.ttl.com http://www.china.ttl.com

Impedance Measurement Plot for Head TSL

Certificate No: Z22-60104 Page 6 of 6

1.5 D900V2 - SN 1d079

<div style="display: flex; justify-content: space-between; align-items: center;"> <div style="text-align: center;"> <small>In Collaboration with</small> TTL Calibration Laboratory <small>ADD: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China TEL: +86-10-42306633-2117 E-mail: cti@ttest.com</small> </div> <div style="text-align: center;"> <small>中国合格评定国家认可委员会</small> CNAS <small>CALIBRATION</small> <small>CMS 1070</small> </div> <div style="text-align: center;"> <small>中国信息通信研究院</small> CAICT </div> </div> <p style="text-align: center;">Client: SGS-CN Certificate No: Z22-60184</p> <div style="border: 1px solid black; padding: 5px;"> <p>CALIBRATION CERTIFICATE</p> <p>Object: D900V2 - SN: 1d079</p> <p>Calibration Procedure(s): FF-Z11-003-01 Calibration Procedures for dipole validation kits</p> <p>Calibration date: June 7, 2022</p> <p>The calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity <70%.</p> <p>Calibration Equipment used (M&TE critical for calibration)</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power Meter NRP2</td> <td>106277</td> <td>24-Sep-21 (CTTL No. J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Power sensor NRP8S</td> <td>104291</td> <td>24-Sep-21 (CTTL No. J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Reference Probe EX3DV4</td> <td>SN 7464</td> <td>26-Jan-22 (SPEAG No. EX3-7464_Jan22)</td> <td>Jan-23</td> </tr> <tr> <td>DAE4</td> <td>SN 1566</td> <td>12-Jan-22 (CTTL-SPEAG No. Z22-60007)</td> <td>Jan-23</td> </tr> </tbody> </table> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Signal Generator E4438C</td> <td>M146071430</td> <td>13-Jan-22 (CTTL No. J22X00409)</td> <td>Jan-23</td> </tr> <tr> <td>Network Analyzer E5071C</td> <td>M146110673</td> <td>14-Jan-22 (CTTL No. J22X00409)</td> <td>Jan-23</td> </tr> </tbody> </table> <p>Calibrated by: Zhao Jing, SAR Test Engineer, Signature: [Signature]</p> <p>Reviewed by: Lin Hao, SAR Test Engineer, Signature: [Signature]</p> <p>Approved by: Qi Diqiyuan, SAR Project Leader, Signature: [Signature]</p> <p style="text-align: right;">Issued: June 13, 2022</p> <p>The calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p> </div> <p style="text-align: center;">Certificate No: Z22-60184 Page 1 of 6</p>	Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Power Meter NRP2	106277	24-Sep-21 (CTTL No. J21X08326)	Sep-22	Power sensor NRP8S	104291	24-Sep-21 (CTTL No. J21X08326)	Sep-22	Reference Probe EX3DV4	SN 7464	26-Jan-22 (SPEAG No. EX3-7464_Jan22)	Jan-23	DAE4	SN 1566	12-Jan-22 (CTTL-SPEAG No. Z22-60007)	Jan-23	Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Signal Generator E4438C	M146071430	13-Jan-22 (CTTL No. J22X00409)	Jan-23	Network Analyzer E5071C	M146110673	14-Jan-22 (CTTL No. J22X00409)	Jan-23	<div style="display: flex; justify-content: space-between; align-items: center;"> <div style="text-align: center;"> <small>In Collaboration with</small> TTL Calibration Laboratory <small>ADD: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China TEL: +86-10-42306633-2117 E-mail: cti@ttest.com</small> </div> <div style="text-align: center;"> <small>中国信息通信研究院</small> CAICT </div> </div> <p>Glossary: TSL: tissue simulating liquid ConvF: sensitivity in TSL / NORM_{x,y,z} N/A: not applicable or not measured</p> <p>Calibration is Performed According to the Following Standards: a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020 b) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz" c) DASY4/S System Handbook</p> <p>Additional Documentation: c) DASY4/S System Handbook</p> <p>Methods Applied and Interpretation of Parameters:</p> <ul style="list-style-type: none"> Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in this certificate are valid at the frequency indicated. Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis. Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required. Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required. SAR measured: SAR measured at the stated antenna input power. SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector. SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result. <div style="border: 1px solid black; padding: 5px; margin-top: 10px;"> <p>The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.</p> </div> <p style="text-align: center;">Certificate No: Z22-60184 Page 2 of 6</p>																														
Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																												
Power Meter NRP2	106277	24-Sep-21 (CTTL No. J21X08326)	Sep-22																																																												
Power sensor NRP8S	104291	24-Sep-21 (CTTL No. J21X08326)	Sep-22																																																												
Reference Probe EX3DV4	SN 7464	26-Jan-22 (SPEAG No. EX3-7464_Jan22)	Jan-23																																																												
DAE4	SN 1566	12-Jan-22 (CTTL-SPEAG No. Z22-60007)	Jan-23																																																												
Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																												
Signal Generator E4438C	M146071430	13-Jan-22 (CTTL No. J22X00409)	Jan-23																																																												
Network Analyzer E5071C	M146110673	14-Jan-22 (CTTL No. J22X00409)	Jan-23																																																												
<div style="display: flex; justify-content: space-between; align-items: center;"> <div style="text-align: center;"> <small>In Collaboration with</small> TTL Calibration Laboratory <small>ADD: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China TEL: +86-10-42306633-2117 E-mail: cti@ttest.com</small> </div> <div style="text-align: center;"> <small>中国信息通信研究院</small> CAICT </div> </div> <p>Measurement Conditions DASY system configuration, as far as not given on page 1</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>DASY Version</th> <th>Configuration</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>Extrapolation</td> <td>DASY52</td> <td>52.10.4</td> </tr> <tr> <td>Phantom</td> <td>Advanced Extrapolation</td> <td></td> </tr> <tr> <td>Distance Dipole Center - TSL</td> <td>Triple Flat Phantom 5.1C</td> <td></td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dk, dz = 5 mm</td> <td>with Spacer</td> </tr> <tr> <td>Frequency</td> <td>900 MHz ± 1 MHz</td> <td></td> </tr> </tbody> </table> <p>Head TSL parameters The following parameters and calculations were applied.</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Parameter</th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>41.5</td> <td>0.07 mho/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>42.1 ± 6 %</td> <td>0.08 mho/m ± 6 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td><1.0 °C</td> <td>---</td> <td>---</td> </tr> </tbody> </table> <p>SAR result with Head TSL</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>SAR averaged over 1 cm³ (1 g) of Head TSL</th> <th>Condition</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>2.70 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>11.0 W/kg ± 18.8 % (k=2)</td> </tr> <tr> <td>SAR averaged over 10 cm³ (10 g) of Head TSL</td> <td>Condition</td> <td>Value</td> </tr> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>1.73 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>7.09 W/kg ± 18.7 % (k=2)</td> </tr> </tbody> </table> <p style="text-align: center;">Certificate No: Z22-60184 Page 3 of 6</p>	DASY Version	Configuration	Value	Extrapolation	DASY52	52.10.4	Phantom	Advanced Extrapolation		Distance Dipole Center - TSL	Triple Flat Phantom 5.1C		Zoom Scan Resolution	dk, dz = 5 mm	with Spacer	Frequency	900 MHz ± 1 MHz		Parameter	Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	41.5	0.07 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	42.1 ± 6 %	0.08 mho/m ± 6 %	Head TSL temperature change during test	<1.0 °C	---	---	SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	Value	SAR measured	250 mW input power	2.70 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	11.0 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	Value	SAR measured	250 mW input power	1.73 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	7.09 W/kg ± 18.7 % (k=2)	<div style="display: flex; justify-content: space-between; align-items: center;"> <div style="text-align: center;"> <small>In Collaboration with</small> TTL Calibration Laboratory <small>ADD: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China TEL: +86-10-42306633-2117 E-mail: cti@ttest.com</small> </div> <div style="text-align: center;"> <small>中国信息通信研究院</small> CAICT </div> </div> <p>Appendix (Additional assessments outside the scope of CNAS L0570)</p> <p>Antenna Parameters with Head TSL</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Parameter</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>Impedance, transformed to feed point</td> <td>48.10 - 8.48jΩ</td> </tr> <tr> <td>Return Loss</td> <td>-23.3 dB</td> </tr> </tbody> </table> <p>General Antenna Parameters and Design</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tbody> <tr> <td>Electrical Delay (one direction)</td> <td>1.312 ns</td> </tr> </tbody> </table> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p>Additional EUT Data</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tbody> <tr> <td>Manufactured by</td> <td>SPEAG</td> </tr> </tbody> </table> <p style="text-align: center;">Certificate No: Z22-60184 Page 4 of 6</p>	Parameter	Value	Impedance, transformed to feed point	48.10 - 8.48jΩ	Return Loss	-23.3 dB	Electrical Delay (one direction)	1.312 ns	Manufactured by	SPEAG
DASY Version	Configuration	Value																																																													
Extrapolation	DASY52	52.10.4																																																													
Phantom	Advanced Extrapolation																																																														
Distance Dipole Center - TSL	Triple Flat Phantom 5.1C																																																														
Zoom Scan Resolution	dk, dz = 5 mm	with Spacer																																																													
Frequency	900 MHz ± 1 MHz																																																														
Parameter	Temperature	Permittivity	Conductivity																																																												
Nominal Head TSL parameters	22.0 °C	41.5	0.07 mho/m																																																												
Measured Head TSL parameters	(22.0 ± 0.2) °C	42.1 ± 6 %	0.08 mho/m ± 6 %																																																												
Head TSL temperature change during test	<1.0 °C	---	---																																																												
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	Value																																																													
SAR measured	250 mW input power	2.70 W/kg																																																													
SAR for nominal Head TSL parameters	normalized to 1W	11.0 W/kg ± 18.8 % (k=2)																																																													
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	Value																																																													
SAR measured	250 mW input power	1.73 W/kg																																																													
SAR for nominal Head TSL parameters	normalized to 1W	7.09 W/kg ± 18.7 % (k=2)																																																													
Parameter	Value																																																														
Impedance, transformed to feed point	48.10 - 8.48jΩ																																																														
Return Loss	-23.3 dB																																																														
Electrical Delay (one direction)	1.312 ns																																																														
Manufactured by	SPEAG																																																														

In Collaboration with
CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-4230603-2117
E-mail: cti@chinaetl.com
http://www.chinaetl.com

DASY5 Validation Report for Head TSL Date: 2022-06-07

Test Laboratory: CCTL, Beijing, China
DUT: Dipole 900 MHz; Type: D900V2; Serial: D900V2 - SN: 14079
Communication System: UTD 0, CW; Frequency: 900 MHz; Duty Cycle: 1:1
Medium parameters used: f = 900 MHz; $\sigma = 0.98$ S/m; $\epsilon_r = 42.05$; $\rho = 1000$ kg/m³
Phantom section: Right Section
Measurement Standard: DASY5 (IEE/IEC/ANSI C63.19-2007)
DASY5 Configuration:

- Probe: EX3DV4 - SN7464; ConvF(9.72, 9.72) @ 900 MHz; Calibrated: 2022-01-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronic: DA44 SN1556; Calibrated: 2022-01-12
- Phantom: MFP_V5.1C (2ddeg probe tilt); Type: QD 000 P51 Cx; Serial: 1062
- DASY52.52.10.4(1535); SEMCAD X 14.6.14(7501)

Dipole Calibration/Zoom Scan (7x7) (7x7) Cube 0; Measurement grid: dx=5mm, dy=5mm, dz=5mm
Reference Value = 59.81 V/m; Power Drift = -0.01 dB
Peak SAR (extrapolated) = 4.20 W/kg
SAR(1g) = 2.78 W/kg; SAR(10g) = 1.78 W/kg
Smallest distance from peaks to all points 3 dB below = 16 mm
Ratio of SAR at M2 to SAR at M1 = 65.8%
Maximum value of SAR (measured) = 3.71 W/kg

Certificate No: Z22-60184 Page 6 of 6

In Collaboration with
CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-4230603-2117
E-mail: cti@chinaetl.com
http://www.chinaetl.com

Impedance Measurement Plot for Head TSL

Certificate No: Z22-60184 Page 6 of 6

1.6 D1800V2 - SN 2d170

In Collaboration with
CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-4230603-2117
E-mail: cti@chinaetl.com
http://www.chinaetl.com

中国认可
国家互认
CALIBRATION
CNAS L8570

Client: **SGS-CN** Certificate No: **Z22-60105**

CALIBRATION CERTIFICATE

Object: D1800V2 - SN: 2d170

Calibration Procedure(s): FF-Z11-003-01
Calibration Procedures for dipole validation kits

Calibration date: March 31, 2022

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility, environment temperature (22±3)°C and humidity <70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Power Meter NRP2	106277	24-Sep-21 (CCTL No.J21X08326)	Sep-22
Power sensor NRP5	104291	24-Sep-21 (CCTL No.J21X08326)	Sep-22
Reference Probe EX3DV4	SN 7307	26-May-21(SPEAG.No.EK3-7307_May21)	May-22
DAE4	SN 1556	12-Jan-22(CCTL-SPEAG.No.Z22-60007)	Jan-23

Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	13-Jan-22 (CCTL No.J22X00406)	Jan-23
Network Analyzer E5071C	MY46110973	14-Jan-22 (CCTL No.J22X00406)	Jan-23

Calibrated by: Zhao Jing, SAR Test Engineer, Signature: [Signature]

Reviewed by: Lin Hao, SAR Test Engineer, Signature: [Signature]

Approved by: Qi Diqiyuan, SAR Project Leader, Signature: [Signature]

Issued: April 6, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60105 Page 1 of 6

In Collaboration with
CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-4230603-2079
E-mail: cti@chinaetl.com
http://www.chinaetl.com

Glossary:

TSL: tissue simulating liquid
ConvF: sensitivity in TSL / NORMx.y.z
N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020
- KDB 865864, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

- DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60105 Page 2 of 6

In Collaboration with **TTL Speag** CALIBRATION LABORATORY **CAICT**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504
 E-mail: cti@chinatest.com http://www.chinatest.cn

Measurement Conditions
 DASYS system configuration, as far as not given on page 1.

DASY Version	DASY52	52.10.4
Extrapolation	Advanced Extrapolation	
Phantom	Triple Flat Phantom 5.1C	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	1800 MHz ± 1 MHz	

Head TSL parameters
 The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	40.0	1.40 mholm
Measured Head TSL parameters	(22.0 ± 0.2) °C	40.8 ± 8 %	1.41 mholm ± 8 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	9.73 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	38.9 W/kg ± 18.8 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	250 mW input power	5.11 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	20.4 W/kg ± 18.7 % (k=2)

Certificate No: Z22-60105 Page 3 of 6

In Collaboration with **TTL Speag** CALIBRATION LABORATORY **CAICT**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504
 E-mail: cti@chinatest.com http://www.chinatest.cn

Appendix (Additional assessments outside the scope of CNAS L0570)

Antenna Parameters with Head TSL

Impedance, transformed to feed point	47.90-2.54jΩ
Return Loss	-29.4dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.116 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.

Additional EUT Data

Manufactured by	SPEAG
-----------------	-------

Certificate No: Z22-60105 Page 4 of 6

In Collaboration with **TTL Speag** CALIBRATION LABORATORY **CAICT**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504
 E-mail: cti@chinatest.com http://www.chinatest.cn

DASY5 Validation Report for Head TSL Date: 2022-03-31
 Test Laboratory: CTTL, Beijing, China
 DUT: Dipole 1800 MHz; Type: D1800V2; Serial: D1800V2 - SN: 2d170
 Communication System: UID 0, CW; Frequency: 1800 MHz; Duty Cycle: 1:1
 Medium parameters used: f = 1800 MHz; σ = 1.411 S/m; ε = 40.62; ρ = 1000 kg/m³
 Phantom section: Right Section
 Measurement Standard: DASY5 (IEEE/ANSI C63.19-2007)
 DASY5 Configuration:

- Probe: EX3DV4 - SN7307; ConvF(8.34, 8.34) @ 1800 MHz; Calibrated: 2021-05-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1556; Calibrated: 2022-01-12
- Phantom: MFP_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062
- DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501)

Dipole Calibration/Zoom Scan (7x7x7) (Cube 0): Measurement grid: dx=5mm, dy=5mm, dz=5mm
 Reference Value = 98.14 V/m; Power Drift = 0.03 dB
 Peak SAR (extrapolated) = 18.2 W/kg
 SAR(1 g) = 9.73 W/kg; SAR(10 g) = 5.11 W/kg
 Smallest distance from peaks to all points 3 dB below = 10 mm
 Ratio of SAR at M2 to SAR at M1 = 54%
 Maximum value of SAR (measured) = 15.2 W/kg

Certificate No: Z22-60105 Page 5 of 6

In Collaboration with **TTL Speag** CALIBRATION LABORATORY **CAICT**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504
 E-mail: cti@chinatest.com http://www.chinatest.cn

Impedance Measurement Plot for Head TSL

Certificate No: Z22-60105 Page 6 of 6

1.7 D1900V2 - SN 5d136

<div style="display: flex; justify-content: space-between; align-items: center;"> </div> <p style="font-size: 8px;"> Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42204633-2117 E-mail: vt@china.ttl.com.cn http://www.caict.ac.cn </p> <p style="text-align: center;"> Client: SGS-CN Certificate No: Z22-60185 </p> <h3 style="text-align: center;">CALIBRATION CERTIFICATE</h3> <p> Object: D1900V2 - SN: 5d136 Calibration Procedure(s): FF-Z11-003-01 Calibration Procedures for dipole validation kits Calibration date: June 7, 2022 </p> <p style="font-size: 8px;"> This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. </p> <p style="font-size: 8px;"> All calibrations have been conducted in the closed laboratory facility; environment temperature (23±3)°C and humidity<70%. </p> <p style="font-size: 8px;"> Calibration Equipment used (M&TE critical for calibration) </p> <table border="1" style="width: 100%; font-size: 8px;"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power Meter NRP2</td> <td>106277</td> <td>24-Sep-21 (CTTL No. J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Power sensor NRP6S</td> <td>104291</td> <td>24-Sep-21 (CTTL No. J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Reference Probe EXSDV4</td> <td>SN 7464</td> <td>28-Jan-22 (SPEAG No. EX3-7464_Jan22)</td> <td>Jan-23</td> </tr> <tr> <td>DAE4</td> <td>SN 1656</td> <td>12-Jan-22 (CTTL-SPEAG No. Z22-90007)</td> <td>Jan-23</td> </tr> </tbody> </table> <table border="1" style="width: 100%; font-size: 8px;"> <thead> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Signal Generator E4438C</td> <td>MY48071430</td> <td>13-Jan-22 (CTTL No. J22X00409)</td> <td>Jan-23</td> </tr> <tr> <td>Network Analyser E5071C</td> <td>MY48110673</td> <td>14-Jan-22 (CTTL No. J22X00406)</td> <td>Jan-23</td> </tr> </tbody> </table> <table border="1" style="width: 100%; font-size: 8px;"> <thead> <tr> <th>Calibrated by:</th> <th>Name</th> <th>Function</th> <th>Signature</th> </tr> </thead> <tbody> <tr> <td></td> <td>Zhao Jing</td> <td>SAR Test Engineer</td> <td></td> </tr> <tr> <td>Reviewed by:</td> <td>Lin Hao</td> <td>SAR Test Engineer</td> <td></td> </tr> <tr> <td>Approved by:</td> <td>Qi Dianyuan</td> <td>SAR Project Leader</td> <td></td> </tr> </tbody> </table> <p style="font-size: 8px; text-align: center;"> Issued: June 13, 2022 This calibration certificate shall not be reproduced except in full without written approval of the laboratory. </p> <p style="font-size: 8px;"> Certificate No: Z22-60185 Page 1 of 6 </p>	Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Power Meter NRP2	106277	24-Sep-21 (CTTL No. J21X08326)	Sep-22	Power sensor NRP6S	104291	24-Sep-21 (CTTL No. J21X08326)	Sep-22	Reference Probe EXSDV4	SN 7464	28-Jan-22 (SPEAG No. EX3-7464_Jan22)	Jan-23	DAE4	SN 1656	12-Jan-22 (CTTL-SPEAG No. Z22-90007)	Jan-23	Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Signal Generator E4438C	MY48071430	13-Jan-22 (CTTL No. J22X00409)	Jan-23	Network Analyser E5071C	MY48110673	14-Jan-22 (CTTL No. J22X00406)	Jan-23	Calibrated by:	Name	Function	Signature		Zhao Jing	SAR Test Engineer		Reviewed by:	Lin Hao	SAR Test Engineer		Approved by:	Qi Dianyuan	SAR Project Leader		<div style="display: flex; justify-content: space-between; align-items: center;"> </div> <p style="font-size: 8px;"> Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42204633-2117 E-mail: vt@china.ttl.com.cn http://www.caict.ac.cn </p> <p>Glossary:</p> <p> TSL: tissue simulating liquid ConvF: sensitivity in TSL / NORMx.y.z N/A: not applicable or not measured </p> <p> Calibration is Performed According to the Following Standards: a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1526: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020 b) KDB 865984, "SAR Measurement Requirements for 100 MHz to 6 GHz" </p> <p>Additional Documentation: c) DASY4/S System Handbook</p> <p>Methods Applied and Interpretation of Parameters:</p> <ul style="list-style-type: none"> • Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated. • Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis. • Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required. • Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required. • SAR measured: SAR measured at the stated antenna input power. • SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector. • SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result. <div style="border: 1px solid black; padding: 5px; font-size: 8px; margin-top: 10px;"> The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%. </div> <p style="font-size: 8px;"> Certificate No: Z22-60185 Page 2 of 6 </p>												
Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																										
Power Meter NRP2	106277	24-Sep-21 (CTTL No. J21X08326)	Sep-22																																																										
Power sensor NRP6S	104291	24-Sep-21 (CTTL No. J21X08326)	Sep-22																																																										
Reference Probe EXSDV4	SN 7464	28-Jan-22 (SPEAG No. EX3-7464_Jan22)	Jan-23																																																										
DAE4	SN 1656	12-Jan-22 (CTTL-SPEAG No. Z22-90007)	Jan-23																																																										
Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																										
Signal Generator E4438C	MY48071430	13-Jan-22 (CTTL No. J22X00409)	Jan-23																																																										
Network Analyser E5071C	MY48110673	14-Jan-22 (CTTL No. J22X00406)	Jan-23																																																										
Calibrated by:	Name	Function	Signature																																																										
	Zhao Jing	SAR Test Engineer																																																											
Reviewed by:	Lin Hao	SAR Test Engineer																																																											
Approved by:	Qi Dianyuan	SAR Project Leader																																																											
<div style="display: flex; justify-content: space-between; align-items: center;"> </div> <p style="font-size: 8px;"> Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42204633-2117 E-mail: vt@china.ttl.com.cn http://www.caict.ac.cn </p> <p>Measurement Conditions DASY system configuration, as far as not given on page 1.</p> <table border="1" style="width: 100%; font-size: 8px;"> <thead> <tr> <th>DASY Version</th> <th>DASY52</th> <th>52.10.4</th> </tr> </thead> <tbody> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> <td></td> </tr> <tr> <td>Phantom</td> <td>Triple Flat Phantom 5.1C</td> <td></td> </tr> <tr> <td>Distance Dipole Center - TSL</td> <td>10 mm</td> <td>with Spacer</td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy, dz = 5 mm</td> <td></td> </tr> <tr> <td>Frequency</td> <td>1900 MHz ± 1 MHz</td> <td></td> </tr> </tbody> </table> <p>Head TSL parameters The following parameters and calculations were applied:</p> <table border="1" style="width: 100%; font-size: 8px;"> <thead> <tr> <th></th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>40.0</td> <td>1.40 mS/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>39.9 ± 6 %</td> <td>1.39 mS/m ± 6 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td><1.0 °C</td> <td>---</td> <td>---</td> </tr> </tbody> </table> <p>SAR result with Head TSL</p> <table border="1" style="width: 100%; font-size: 8px;"> <thead> <tr> <th>SAR averaged over 1 cm² (1 g) of Head TSL</th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>9.65 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>40.0 W/kg ± 18.8 % (k=2)</td> </tr> <tr> <td>SAR averaged over 10 cm² (10 g) of Head TSL</td> <th>Condition</th> <th></th> </tr> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>5.18 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>20.8 W/kg ± 18.7 % (k=2)</td> </tr> </tbody> </table> <p style="font-size: 8px;"> Certificate No: Z22-60185 Page 3 of 6 </p>	DASY Version	DASY52	52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	10 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm		Frequency	1900 MHz ± 1 MHz			Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	40.0	1.40 mS/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	39.9 ± 6 %	1.39 mS/m ± 6 %	Head TSL temperature change during test	<1.0 °C	---	---	SAR averaged over 1 cm² (1 g) of Head TSL	Condition		SAR measured	250 mW input power	9.65 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	40.0 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm² (10 g) of Head TSL	Condition		SAR measured	250 mW input power	5.18 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	20.8 W/kg ± 18.7 % (k=2)	<div style="display: flex; justify-content: space-between; align-items: center;"> </div> <p style="font-size: 8px;"> Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42204633-2117 E-mail: vt@china.ttl.com.cn http://www.caict.ac.cn </p> <p>Appendix (Additional assessments outside the scope of CNAS L0570)</p> <p>Antenna Parameters with Head TSL</p> <table border="1" style="width: 100%; font-size: 8px;"> <tbody> <tr> <td>Impedance, transformed to feed point</td> <td>51.2Ω ± 7.5Ω(j)</td> </tr> <tr> <td>Return Loss</td> <td>-22.4dB</td> </tr> </tbody> </table> <p>General Antenna Parameters and Design</p> <table border="1" style="width: 100%; font-size: 8px;"> <tbody> <tr> <td>Electrical Delay (one direction)</td> <td>1.109 ns</td> </tr> </tbody> </table> <p style="font-size: 8px;"> After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured. </p> <p style="font-size: 8px;"> The dipole is made of standard semi-rigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged. </p> <p>Additional EUT Data</p> <table border="1" style="width: 100%; font-size: 8px;"> <tbody> <tr> <td>Manufactured by</td> <td>SPEAG</td> </tr> </tbody> </table> <p style="font-size: 8px;"> Certificate No: Z22-60185 Page 4 of 6 </p>	Impedance, transformed to feed point	51.2Ω ± 7.5Ω(j)	Return Loss	-22.4dB	Electrical Delay (one direction)	1.109 ns	Manufactured by	SPEAG
DASY Version	DASY52	52.10.4																																																											
Extrapolation	Advanced Extrapolation																																																												
Phantom	Triple Flat Phantom 5.1C																																																												
Distance Dipole Center - TSL	10 mm	with Spacer																																																											
Zoom Scan Resolution	dx, dy, dz = 5 mm																																																												
Frequency	1900 MHz ± 1 MHz																																																												
	Temperature	Permittivity	Conductivity																																																										
Nominal Head TSL parameters	22.0 °C	40.0	1.40 mS/m																																																										
Measured Head TSL parameters	(22.0 ± 0.2) °C	39.9 ± 6 %	1.39 mS/m ± 6 %																																																										
Head TSL temperature change during test	<1.0 °C	---	---																																																										
SAR averaged over 1 cm² (1 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	9.65 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	40.0 W/kg ± 18.8 % (k=2)																																																											
SAR averaged over 10 cm² (10 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	5.18 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	20.8 W/kg ± 18.7 % (k=2)																																																											
Impedance, transformed to feed point	51.2Ω ± 7.5Ω(j)																																																												
Return Loss	-22.4dB																																																												
Electrical Delay (one direction)	1.109 ns																																																												
Manufactured by	SPEAG																																																												

In Collaboration with **TTL Speaq** Calibration Laboratory and **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62506633-3117 E-mail: cti@china.ttl.com.cn http://www.caict.ac.cn

DASY5 Validation Report for Head TSL Date: 2022-06-07
Test Laboratory: CTTL, Beijing, China
DUT: Dipole 1900 MHz; Type: D1900V2; Serial: D1900V2 - SN: 54136
Communication System: UTD 0, CW; Frequency: 1900 MHz; Duty Cycle: 1:1
Medium parameters used: $f = 1900 \text{ MHz}$; $\epsilon = 1.385 \text{ S/m}$; $\sigma = 39.85$; $\rho = 1000 \text{ kg/m}^3$
Phantom section: Right Section
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)
DASY5 Configuration:

- Probe: EX3DV4 - SN7464; ConvF(R, 1.8, 8.18, 8.18) @ 1900 MHz; Calibrated: 2022-01-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1556; Calibrated: 2022-01-12
- Phantom: MFP_V5_IC (20kg probe 0th); Type: QD 000 P51 Cx; Serial: 1062
- DASY5 52.10.4(1555); SEMCAD X.14.6.14(7501)

Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube D: Measurement grid: dx=5mm, dy=5mm, dz=5mm
Reference Value = 99.99 V/m; Power Drift = 0.04 dB
Peak SAR (extrapolated) = 18.6 W/kg
SAR(1 g) = 9.95 W/kg; SAR(10 g) = 5.18 W/kg
Smallest distance from peaks to all points 3 dB below = 9.2 mm
Ratio of SAR at M2 to SAR at M1 = 54.1%

Maximum value of SAR (measured) = 15.6 W/kg

0 dB = 15.6 W/kg = 11.93 dBW/kg

Certificate No: Z22-60185 Page 5 of 6

In Collaboration with **TTL Speaq** Calibration Laboratory and **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62506633-3117 E-mail: cti@china.ttl.com.cn http://www.caict.ac.cn

Impedance Measurement Plot for Head TSL

Certificate No: Z22-60185 Page 6 of 6

1.8 D2000V2 - SN 1041

In Collaboration with **TTL Speaq** Calibration Laboratory, **CEC-IRA**, **CNAS** (China National Accreditation Center for Calibration and Testing), and **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62506633-3117 E-mail: cti@china.ttl.com.cn http://www.caict.ac.cn

Client: **SGS-CN** Certificate No: **Z22-60188**

CALIBRATION CERTIFICATE

Object: D2000V2 - SN: 1041
Calibration Procedure(s): FF-Z11-003-01
Calibration Procedures for dipole validation kits
Calibration date: June 8, 2022

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (8). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility; environment temperature (22±)°C and humidity < 70%.

Calibration Equipment used (M&E critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Power Meter NRPZ	106277	24-Sep-21 (CTTL No. J21X06326)	Sep-22
Power sensor KRPS	104291	24-Sep-21 (CTTL No. J21X06326)	Sep-22
Reference Probe EX3DV4	SN 7464	26-Jan-22 (SPEAG No. EX3-7464_Jan22)	Jan-23
DAE4	SN 1556	12-Jan-22 (CTTL-SPEAG No. Z22-60007)	Jan-23

Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No. J22X00409)	Jan-23
Network Analyzer E5071C	MY48110673	14-Jan-22 (CTTL No. J22X00406)	Jan-23

Calibrated by: Zhao Jing SAR Test Engineer
Reviewed by: Lin Hao SAR Test Engineer
Approved by: Qi Dianyuan SAR Project Leader

Issued: June 13, 2022
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60188 Page 1 of 6

In Collaboration with **TTL Speaq** Calibration Laboratory and **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62506633-3117 E-mail: cti@china.ttl.com.cn http://www.caict.ac.cn

Glossary:
TSL: Issue simulating liquid
ConvF: sensitivity in TSL / NORMx.y.z
N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:
a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020
b) KDB 865964, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:
c) DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60188 Page 2 of 6

In Collaboration with **TTL S p e a g** CALIBRATION LABORATORY and **CAICT**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-4239683-2117
 E-mail: ott@china.ttl.com http://www.caict.ac.cn

Measurement Conditions
 DASYS system configuration, as far as not given on page 1.

DASY Version	DASY52	52.10.4
Extrapolation	Advanced Extrapolation	
Phantom	Triple Flat Phantom 5.1C	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	2000 MHz ± 1 MHz	

Head TSL parameters
 The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	40.0	1.40 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	40.2 ± 8 %	1.39 mho/m ± 8 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	10.4 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	41.8 W/kg ± 18.8 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	250 mW input power	5.30 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	21.3 W/kg ± 18.7 % (k=2)

Certificate No: Z22-60186 Page 3 of 6

In Collaboration with **TTL S p e a g** CALIBRATION LABORATORY and **CAICT**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-4239683-2117
 E-mail: ott@china.ttl.com http://www.caict.ac.cn

Appendix (Additional assessments outside the scope of CNAS L0570)

Antenna Parameters with Head TSL

Impedance, transformed to feed point	48.4Ω ± 0.74jΩ
Return Loss	-34.9dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.088 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.

The dipole is made of standard semi-rigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small and caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.

Additional EUT Data

Manufactured by	SPEAG
-----------------	-------

Certificate No: Z22-60186 Page 4 of 6

In Collaboration with **TTL S p e a g** CALIBRATION LABORATORY and **CAICT**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-4239683-2117
 E-mail: ott@china.ttl.com http://www.caict.ac.cn

DASY5 Validation Report for Head TSL Date: 2022-06-06

Test Laboratory: CTTL, Beijing, China
 DUT: Dipole 2000 MHz; Type: D2000V2; Serial: D2000V2 - SN: 1041
 Communication System: LIID 0, CW; Frequency: 2000 MHz; Duty Cycle: 1:1
 Medium parameters used: $f = 2000$ MHz; $\sigma = 1.392$ S/m; $\epsilon_r = 40.21$; $\rho = 1000$ kg/m³
 Phantom section: Right Section
 Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)
 DASY5 Configuration:

- Probe: EX3DV4 - SN7464; ConvF(R,2, 8.2, 8.2) @ 2000 MHz; Calibrated: 2022-01-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DA64 Sn1556; Calibrated: 2022-01-12
- Phantom: MFP_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062
- DASY5: S2.10.4(1535); SEMCAD X 14.6.14(7501)

Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube 0; Measurement grid: dx=5mm, dy=5mm, dz=5mm
 Reference Value = 103.4 V/m; Power Drift = 0.03 dB
 Peak SAR (extrapolated) = 19.6 W/kg
 SAR(1 g) = 10.4 W/kg; SAR(10 g) = 5.3 W/kg
 Smallest distance from peaks to all points 3 dB below = 9.1 mm
 Ratio of SAR at M2 to SAR at M1 = 53.6%
 Maximum value of SAR (measured) = 16.3 W/kg

Certificate No: Z22-60186 Page 5 of 6

In Collaboration with **TTL S p e a g** CALIBRATION LABORATORY and **CAICT**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-4239683-2117
 E-mail: ott@china.ttl.com http://www.caict.ac.cn

Impedance Measurement Plot for Head TSL

Certificate No: Z22-60186 Page 6 of 6

1.9 D2300V2 - SN 1096

<p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2512 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.chinatitl.cn</p>		<p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.chinatitl.cn</p>																			
<p>Client: SGS-CN Certificate No: Z22-60106</p>																					
<p>CALIBRATION CERTIFICATE</p>																					
Object	D2300V2 - SN 1096																				
Calibration Procedure(s)	FF-Z11-003-01 Calibration Procedures for dipole validation kits																				
Calibration date:	March 31, 2022																				
<p>This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity <70%.</p> <p>Calibration Equipment used (M&TE critical for calibration)</p>																					
Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration																		
Power Meter NRP2	108277	24-Sep-21 (CTTL No.J21X08328)	Sep-22																		
Power sensor NRP8S	104291	24-Sep-21 (CTTL No.J21X08328)	Sep-22																		
Reference Probe EX3DV4 DAE4	SN 7307 SN 1556	26-May-21(SPEAG.No.EK3-7307_May21) 12-Jan-22(CTTL-SPEAG.No.Z22-60007)	May-22 Jan-23																		
Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration																		
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No.J22X00406)	Jan-23																		
Network Analyzer E5071C	MY48110673	14-Jan-22 (CTTL No.J22X00406)	Jan-23																		
Calibrated by:	Name	Function	Signature																		
	Zhao Jing	SAR Test Engineer																			
Reviewed by:	Lin Hao	SAR Test Engineer																			
Approved by:	Qi Diaryuan	SAR Project Leader																			
<p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p>																					
Certificate No: Z22-60106		Page 1 of 6																			
<p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.chinatitl.cn</p>		<p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.chinatitl.cn</p>																			
<p>Measurement Conditions DASY system configuration, as far as not given on page 1</p> <table border="1"> <tr> <td>DASY Version</td> <td>DASY52</td> <td>52.10.4</td> </tr> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> <td></td> </tr> <tr> <td>Phantom</td> <td>Triple Flat Phantom 5.1C</td> <td></td> </tr> <tr> <td>Distance Dipole Center - TSL</td> <td>10 mm</td> <td>with Spacer</td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy, dz = 5 mm</td> <td></td> </tr> <tr> <td>Frequency</td> <td>2300 MHz ± 1 MHz</td> <td></td> </tr> </table>				DASY Version	DASY52	52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	10 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm		Frequency	2300 MHz ± 1 MHz	
DASY Version	DASY52	52.10.4																			
Extrapolation	Advanced Extrapolation																				
Phantom	Triple Flat Phantom 5.1C																				
Distance Dipole Center - TSL	10 mm	with Spacer																			
Zoom Scan Resolution	dx, dy, dz = 5 mm																				
Frequency	2300 MHz ± 1 MHz																				
<p>Head TSL parameters The following parameters and calculations were applied:</p> <table border="1"> <thead> <tr> <th></th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>39.5</td> <td>1.67 mho/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>39.8 ± 6 %</td> <td>1.70 mho/m ± 6 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td><1.0 °C</td> <td>—</td> <td>—</td> </tr> </tbody> </table>					Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	39.5	1.67 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	39.8 ± 6 %	1.70 mho/m ± 6 %	Head TSL temperature change during test	<1.0 °C	—	—		
	Temperature	Permittivity	Conductivity																		
Nominal Head TSL parameters	22.0 °C	39.5	1.67 mho/m																		
Measured Head TSL parameters	(22.0 ± 0.2) °C	39.8 ± 6 %	1.70 mho/m ± 6 %																		
Head TSL temperature change during test	<1.0 °C	—	—																		
<p>SAR result with Head TSL</p> <table border="1"> <thead> <tr> <th>SAR averaged over 1 cm³ (1g) of Head TSL</th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>12.4 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>49.2 W/kg ± 18.8 % (k=2)</td> </tr> <tr> <th>SAR averaged over 10 cm³ (10g) of Head TSL</th> <th>Condition</th> <th></th> </tr> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>5.88 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>23.4 W/kg ± 18.7 % (k=2)</td> </tr> </tbody> </table>				SAR averaged over 1 cm ³ (1g) of Head TSL	Condition		SAR measured	250 mW input power	12.4 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	49.2 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm ³ (10g) of Head TSL	Condition		SAR measured	250 mW input power	5.88 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	23.4 W/kg ± 18.7 % (k=2)
SAR averaged over 1 cm ³ (1g) of Head TSL	Condition																				
SAR measured	250 mW input power	12.4 W/kg																			
SAR for nominal Head TSL parameters	normalized to 1W	49.2 W/kg ± 18.8 % (k=2)																			
SAR averaged over 10 cm ³ (10g) of Head TSL	Condition																				
SAR measured	250 mW input power	5.88 W/kg																			
SAR for nominal Head TSL parameters	normalized to 1W	23.4 W/kg ± 18.7 % (k=2)																			
<p>Glossary:</p> <p>TSL: Issue simulating liquid ConvF: sensitivity in TSL / NORMx,y,z N/A: not applicable or not measured</p> <p>Calibration is Performed According to the Following Standards:</p> <p>a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020 b) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"</p> <p>Additional Documentation: c) DASY4/5 System Handbook</p> <p>Methods Applied and Interpretation of Parameters:</p> <ul style="list-style-type: none"> Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated. Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis. Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required. Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required. SAR measured: SAR measured at the stated antenna input power. SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector. SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result. <p>The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.</p>																					
Certificate No: Z22-60106		Page 2 of 6																			
<p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.chinatitl.cn</p>		<p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.chinatitl.cn</p>																			
<p>Appendix (Additional assessments outside the scope of CNAS L0570)</p> <p>Antenna Parameters with Head TSL</p> <table border="1"> <tr> <td>Impedance, transformed to feed point</td> <td>49.2Ω - 4.56jΩ</td> </tr> <tr> <td>Return Loss</td> <td>-26.6dB</td> </tr> </table> <p>General Antenna Parameters and Design</p> <table border="1"> <tr> <td>Electrical Delay (one direction)</td> <td>1.083 ns</td> </tr> </table> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p>Additional EUT Data</p> <table border="1"> <tr> <td>Manufactured by</td> <td>SPEAG</td> </tr> </table>				Impedance, transformed to feed point	49.2Ω - 4.56jΩ	Return Loss	-26.6dB	Electrical Delay (one direction)	1.083 ns	Manufactured by	SPEAG										
Impedance, transformed to feed point	49.2Ω - 4.56jΩ																				
Return Loss	-26.6dB																				
Electrical Delay (one direction)	1.083 ns																				
Manufactured by	SPEAG																				
Certificate No: Z22-60106		Page 3 of 6																			
Certificate No: Z22-60106		Page 4 of 6																			

In Collaboration with **TTL Speaq** CALIBRATION LABORATORY **CAICT**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
 E-mail: cti@china.ttl.com http://www.chinatit.com

DASY5 Validation Report for Head TSL Date: 2022-03-31
 Test Laboratory: CTTL, Beijing, China
 DUT: Dipole 2300 MHz; Type: D2300V2; Serial: D2300V2 - SN: 1096
 Communication System: UTD 0, CW; Frequency: 2300 MHz; Duty Cycle: 1:1
 Medium parameters used: $f = 2300 \text{ MHz}$; $\sigma = 1.702 \text{ S/m}$; $\epsilon = 39.77$; $\rho = 1000 \text{ kg/m}^3$
 Phantom section: Right Section
 Measurement Standard: DASY5 (IEEE/EC/ANSI C63.19-2007)
 DASY5 Configuration:

- Probe: EX3DV4 - SN7307; ConvF(8.01, 8.01) @ 2300 MHz; Calibrated: 2021-05-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sst1556; Calibrated: 2022-01-12
- Phantom: MFP V5.1C (2ldag probe tilt); Type: QD 000 P51 Cx; Serial: 1062
- DASY52 S2.10.4(1535); SEMCAD X 14.6;14(7501)

Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm
 Reference Value = 102.7 V/m; Power Drift = 0.00 dB
 Peak SAR (extrapolated) = 21.8 W/kg
 SAR(1 g) = 12.4 W/kg; SAR(10 g) = 5.88 W/kg
 Smallest distance from peaks to all points 3 dB below = 9 mm
 Ratio of SAR at M2 to SAR at M1 = 50.4%
 Maximum value of SAR (measured) = 20.3 W/kg

Certificate No: Z22-60106 Page 1 of 6

In Collaboration with **TTL Speaq** CALIBRATION LABORATORY **CAICT**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
 E-mail: cti@china.ttl.com http://www.chinatit.com

Impedance Measurement Plot for Head TSL

Certificate No: Z22-60106 Page 4 of 6

1.10 D2450V2 - SN 817

In Collaboration with **TTL Speaq** CALIBRATION LABORATORY **CAICT** **CNAS** **中国认可 国家互认 校准** **CAICT**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
 E-mail: cti@china.ttl.com http://www.chinatit.com

Client: **SGS-CN** Certificate No: **Z22-60107**

CALIBRATION CERTIFICATE

Object: D2450V2 - SN 817
 Calibration Procedure(s): FF-Z11-003-01
 Calibration date: April 1, 2022

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by: Certificate No.)	Scheduled Calibration
Power Meter	NRP2	24-Sep-21 (CTTL No.J21X08320)	Sep-22
Power sensor	NRP8S	104291 24-Sep-21 (CTTL No.J21X08320)	Sep-22
Reference Probe	EX3DV4	SN 7307 26-May-21(SPEAG.No.EX3-7307_May21)	May-22
DAE4	SN 1556	12-Jan-22(CTTL-SPEAG.No.Z22-60007)	Jan-23

Secondary Standards	ID #	Cal Date (Calibrated by: Certificate No.)	Scheduled Calibration
Signal Generator	E4438C	MY49071430 13-Jan-22 (CTTL No. J22X00406)	Jan-23
Network Analyzer	E5071C	MY46110873 14-Jan-22 (CTTL No. J22X00406)	Jan-23

Calibrated by: Zhao Jing SAR Test Engineer
 Reviewed by: Lin Hao SAR Test Engineer
 Approved by: Qi Dianyuan SAR Project Leader

Signature: [Signatures]
 Issued: April 6, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60107 Page 1 of 6

In Collaboration with **TTL Speaq** CALIBRATION LABORATORY **CAICT**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
 E-mail: cti@china.ttl.com http://www.chinatit.com

Glossary:

TSL: tissue simulating liquid
 ConvF: sensitivity in TSL / NORMx.yz
 N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020
- KDB 855664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:











- DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60107 Page 2 of 6

<div style="display: flex; justify-content: space-between; align-items: center;">   </div> <p style="font-size: small;">In Collaboration with Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.chinatit.com</p> <p>Measurement Conditions DASY system configuration, as far as not given on page 1</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td>DASY Version</td> <td>DASY52</td> <td>52.10.4</td> </tr> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> <td></td> </tr> <tr> <td>Phantom</td> <td>Triple Flat Phantom 5.1C</td> <td></td> </tr> <tr> <td>Distance Dipole Center - TSL</td> <td>10 mm</td> <td>with Spacer</td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy, dz = 5 mm</td> <td></td> </tr> <tr> <td>Frequency</td> <td>2450 MHz ± 1 MHz</td> <td></td> </tr> </table> <p>Head TSL parameters The following parameters and calculations were applied.</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th></th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>39.2</td> <td>1.80 mho/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>39.5 ± 6 %</td> <td>1.79 mho/m ± 6 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td><+1.0 °C</td> <td>---</td> <td>---</td> </tr> </tbody> </table> <p>SAR result with Head TSL</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>SAR averaged over 1 cm³ (1 g) of Head TSL</th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>13.2 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>53.0 W/kg ± 18.8 % (k=2)</td> </tr> <tr> <th>SAR averaged over 10 cm³ (10 g) of Head TSL</th> <th>Condition</th> <th></th> </tr> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>6.15 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>24.7 W/kg ± 18.7 % (k=2)</td> </tr> </tbody> </table> <p style="font-size: x-small;">Certificate No: Z22-60107 Page 3 of 6</p>	DASY Version	DASY52	52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	10 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm		Frequency	2450 MHz ± 1 MHz			Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	39.2	1.80 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	39.5 ± 6 %	1.79 mho/m ± 6 %	Head TSL temperature change during test	<+1.0 °C	---	---	SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition		SAR measured	250 mW input power	13.2 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	53.0 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition		SAR measured	250 mW input power	6.15 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	24.7 W/kg ± 18.7 % (k=2)	<div style="display: flex; justify-content: space-between; align-items: center;">   </div> <p style="font-size: small;">In Collaboration with Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.chinatit.com</p> <p>Appendix (Additional assessments outside the scope of CNAS L0570)</p> <p>Antenna Parameters with Head TSL</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td>Impedance, transformed to feed point</td> <td>52.10 ± 3.20jΩ</td> </tr> <tr> <td>Return Loss</td> <td>-28.5dB</td> </tr> </table> <p>General Antenna Parameters and Design</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td>Electrical Delay (one direction)</td> <td>1.066 ns</td> </tr> </table> <p style="font-size: x-small;">After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p style="font-size: x-small;">The dipole is made of standard serringrid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p>Additional EUT Data</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td>Manufactured by</td> <td>SPEAG</td> </tr> </table> <p style="font-size: x-small;">Certificate No: Z22-60107 Page 4 of 6</p>	Impedance, transformed to feed point	52.10 ± 3.20jΩ	Return Loss	-28.5dB	Electrical Delay (one direction)	1.066 ns	Manufactured by	SPEAG
DASY Version	DASY52	52.10.4																																																											
Extrapolation	Advanced Extrapolation																																																												
Phantom	Triple Flat Phantom 5.1C																																																												
Distance Dipole Center - TSL	10 mm	with Spacer																																																											
Zoom Scan Resolution	dx, dy, dz = 5 mm																																																												
Frequency	2450 MHz ± 1 MHz																																																												
	Temperature	Permittivity	Conductivity																																																										
Nominal Head TSL parameters	22.0 °C	39.2	1.80 mho/m																																																										
Measured Head TSL parameters	(22.0 ± 0.2) °C	39.5 ± 6 %	1.79 mho/m ± 6 %																																																										
Head TSL temperature change during test	<+1.0 °C	---	---																																																										
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	13.2 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	53.0 W/kg ± 18.8 % (k=2)																																																											
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	6.15 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	24.7 W/kg ± 18.7 % (k=2)																																																											
Impedance, transformed to feed point	52.10 ± 3.20jΩ																																																												
Return Loss	-28.5dB																																																												
Electrical Delay (one direction)	1.066 ns																																																												
Manufactured by	SPEAG																																																												
<div style="display: flex; justify-content: space-between; align-items: center;">   </div> <p style="font-size: small;">In Collaboration with Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.chinatit.com</p> <p>DASY5 Validation Report for Head TSL Date: 2022-04-01</p> <p>Test Laboratory: CTTL, Beijing, China DUT: Dipole 2450 MHz; Type: D2450V2 - SN: 817 Communication System: UID 0, CW; Frequency: 2450 MHz; Duty Cycle: 1:1 Medium parameters used: f = 2450 MHz; σ = 1.79 S/m; ε_r = 39.52; ρ = 1000 kg/m³ Phantom section: Right Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007) DASY5 Configuration:</p> <ul style="list-style-type: none"> Probe: EX3DV4 - SN7307; ConvF(7.75, 7.75, 7.75) @ 2450 MHz; Calibrated: 2021-05-26 Sensor-Surface: 1.4mm (Mechanical Surface Detection) Electronics: DA14 Sn1556; Calibrated: 2022-01-12 Phantom: MFP_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062 DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501) <p>Dipole Calibration Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 104.6 V/m; Power Drift = -0.03 dB Peak SAR (extrapolated) = 27.0 W/kg SAR(1 g) = 13.2 W/kg; SAR(10 g) = 6.15 W/kg Smallest distance from peaks to all points 3 dB below = 8.9 mm Ratio of SAR at M2 to SAR at M1 = -49.2% Maximum value of SAR (measured) = 22.1 W/kg</p>  <p style="font-size: x-small;">Certificate No: Z22-60107 Page 5 of 6</p>	<div style="display: flex; justify-content: space-between; align-items: center;">   </div> <p style="font-size: small;">In Collaboration with Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.chinatit.com</p> <p>Impedance Measurement Plot for Head TSL</p>  <p style="font-size: x-small;">Certificate No: Z22-60107 Page 6 of 6</p>																																																												

1.11 D2600V2 - SN 1158

<div style="text-align: center;"> </div> <p style="font-size: small;"> Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2512 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com.cn http://www.chinatitl.cn </p> <p> Client: SGS-CN Certificate No: Z22-60108 </p> <h3 style="text-align: center;">CALIBRATION CERTIFICATE</h3> <p> Object: D2600V2 - SN: 1158 </p> <p> Calibration Procedure(s): FF-Z11-003-01 Calibration Procedures for dipole validation kits </p> <p> Calibration date: March 31, 2022 </p> <p> This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. </p> <p> All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity<70%. </p> <p> Calibration Equipment used (M&TE critical for calibration) </p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power Meter NRP2</td> <td>102577</td> <td>24-Sep-21 (CTTL No.J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Power sensor NRP8S</td> <td>104291</td> <td>24-Sep-21 (CTTL No.J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Reference Probe EX3DV4</td> <td>SN 7307</td> <td>26-May-21(SPEAG.No.EX3-7307_May21)</td> <td>May-22</td> </tr> <tr> <td>DAE4</td> <td>SN 1556</td> <td>12-Jan-22(CTTL-SPEAG.No.Z22-60007)</td> <td>Jan-23</td> </tr> </tbody> </table> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Signal Generator E4438C</td> <td>MY49071430</td> <td>13-Jan-22 (CTTL No.Z22X00406)</td> <td>Jan-23</td> </tr> <tr> <td>Network Analyzer E5071C</td> <td>MY49110673</td> <td>14-Jan-22 (CTTL No.Z22X00406)</td> <td>Jan-23</td> </tr> </tbody> </table> <p> Calibrated by: Zhao Jing SAR Test Engineer </p> <p> Reviewed by: Lin Hao SAR Test Engineer </p> <p> Approved by: Qi Dianyuan SAR Project Leader </p> <p style="text-align: right;"> Issued: April 6, 2022 </p> <p style="font-size: x-small;"> This calibration certificate shall not be reproduced except in full without written approval of the laboratory. </p> <p style="font-size: x-small;"> Certificate No: Z22-60108 Page 1 of 6 </p>	Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration	Power Meter NRP2	102577	24-Sep-21 (CTTL No.J21X08326)	Sep-22	Power sensor NRP8S	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22	Reference Probe EX3DV4	SN 7307	26-May-21(SPEAG.No.EX3-7307_May21)	May-22	DAE4	SN 1556	12-Jan-22(CTTL-SPEAG.No.Z22-60007)	Jan-23	Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration	Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No.Z22X00406)	Jan-23	Network Analyzer E5071C	MY49110673	14-Jan-22 (CTTL No.Z22X00406)	Jan-23	<div style="text-align: center;"> </div> <p style="font-size: small;"> Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com.cn http://www.chinatitl.cn </p> <p> Client: SGS-CN Certificate No: Z22-60108 </p> <h3 style="text-align: center;">GLOSSARY</h3> <p> TSL: tissue simulating liquid ConvF: sensitivity in TSL / NORMx.y.z N/A: not applicable or not measured </p> <p> Calibration is Performed According to the Following Standards: </p> <ol style="list-style-type: none"> IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020 KDB 865864, "SAR Measurement Requirements for 100 MHz to 6 GHz" <p> Additional Documentation: c) DASY4/S System Handbook </p> <p> Methods Applied and Interpretation of Parameters: </p> <ul style="list-style-type: none"> Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated. Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis. Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required. Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required. SAR measured: SAR measured at the stated antenna input power. SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector. SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result. <div style="border: 1px solid black; padding: 5px; font-size: x-small;"> The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%. </div> <p style="font-size: x-small;"> Certificate No: Z22-60108 Page 2 of 6 </p>																														
Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration																																																												
Power Meter NRP2	102577	24-Sep-21 (CTTL No.J21X08326)	Sep-22																																																												
Power sensor NRP8S	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22																																																												
Reference Probe EX3DV4	SN 7307	26-May-21(SPEAG.No.EX3-7307_May21)	May-22																																																												
DAE4	SN 1556	12-Jan-22(CTTL-SPEAG.No.Z22-60007)	Jan-23																																																												
Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration																																																												
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No.Z22X00406)	Jan-23																																																												
Network Analyzer E5071C	MY49110673	14-Jan-22 (CTTL No.Z22X00406)	Jan-23																																																												
<div style="text-align: center;"> </div> <p style="font-size: small;"> Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com.cn http://www.chinatitl.cn </p> <p> Client: SGS-CN Certificate No: Z22-60108 </p> <h3 style="text-align: center;">MEASUREMENT CONDITIONS</h3> <p> DASY system configuration, as far as not given on page 1. </p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Parameter</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>DASY Version</td> <td>DASY52 52.10.4</td> </tr> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> </tr> <tr> <td>Phantom</td> <td>Triple Flat Phantom 5.1C</td> </tr> <tr> <td>Distance Dipole Center - TSL</td> <td>10 mm with Spacer</td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy, dz = 5 mm</td> </tr> <tr> <td>Frequency</td> <td>2600 MHz ± 1 MHz</td> </tr> </tbody> </table> <p> Head TSL parameters The following parameters and calculations were applied. </p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Parameter</th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>39.0</td> <td>1.96 mho/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>38.7 ± 6 %</td> <td>1.96 mho/m ± 6 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td><1.0 °C</td> <td>---</td> <td>---</td> </tr> </tbody> </table> <p> SAR result with Head TSL </p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>SAR averaged over 1 cm² (1 g) of Head TSL</th> <th>Condition</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>13.7 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>54.8 W/kg ± 18.8 % (k=2)</td> </tr> </tbody> </table> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>SAR averaged over 10 cm² (10 g) of Head TSL</th> <th>Condition</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>6.12 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>24.6 W/kg ± 18.7 % (k=2)</td> </tr> </tbody> </table> <p style="font-size: x-small;"> Certificate No: Z22-60108 Page 3 of 6 </p>	Parameter	Value	DASY Version	DASY52 52.10.4	Extrapolation	Advanced Extrapolation	Phantom	Triple Flat Phantom 5.1C	Distance Dipole Center - TSL	10 mm with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm	Frequency	2600 MHz ± 1 MHz	Parameter	Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	39.0	1.96 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	38.7 ± 6 %	1.96 mho/m ± 6 %	Head TSL temperature change during test	<1.0 °C	---	---	SAR averaged over 1 cm² (1 g) of Head TSL	Condition	Value	SAR measured	250 mW input power	13.7 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	54.8 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm² (10 g) of Head TSL	Condition	Value	SAR measured	250 mW input power	6.12 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	24.6 W/kg ± 18.7 % (k=2)	<div style="text-align: center;"> </div> <p style="font-size: small;"> Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com.cn http://www.chinatitl.cn </p> <p> Client: SGS-CN Certificate No: Z22-60108 </p> <h3 style="text-align: center;">APPENDIX (Additional assessments outside the scope of CNAS L0570)</h3> <p> Antenna Parameters with Head TSL </p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Parameter</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>Impedance, transformed to feed point</td> <td>49.90- 6.49jΩ</td> </tr> <tr> <td>Return Loss</td> <td>- 23.8dB</td> </tr> </tbody> </table> <p> General Antenna Parameters and Design </p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Parameter</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>Electrical Delay (one direction)</td> <td>1.053 ns</td> </tr> </tbody> </table> <p> After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured. </p> <p> The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged. </p> <p> Additional EUT Data </p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Manufacturer by</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>Manufacturer by</td> <td>SPEAG</td> </tr> </tbody> </table> <p style="font-size: x-small;"> Certificate No: Z22-60108 Page 4 of 6 </p>	Parameter	Value	Impedance, transformed to feed point	49.90- 6.49jΩ	Return Loss	- 23.8dB	Parameter	Value	Electrical Delay (one direction)	1.053 ns	Manufacturer by	Value	Manufacturer by	SPEAG
Parameter	Value																																																														
DASY Version	DASY52 52.10.4																																																														
Extrapolation	Advanced Extrapolation																																																														
Phantom	Triple Flat Phantom 5.1C																																																														
Distance Dipole Center - TSL	10 mm with Spacer																																																														
Zoom Scan Resolution	dx, dy, dz = 5 mm																																																														
Frequency	2600 MHz ± 1 MHz																																																														
Parameter	Temperature	Permittivity	Conductivity																																																												
Nominal Head TSL parameters	22.0 °C	39.0	1.96 mho/m																																																												
Measured Head TSL parameters	(22.0 ± 0.2) °C	38.7 ± 6 %	1.96 mho/m ± 6 %																																																												
Head TSL temperature change during test	<1.0 °C	---	---																																																												
SAR averaged over 1 cm² (1 g) of Head TSL	Condition	Value																																																													
SAR measured	250 mW input power	13.7 W/kg																																																													
SAR for nominal Head TSL parameters	normalized to 1W	54.8 W/kg ± 18.8 % (k=2)																																																													
SAR averaged over 10 cm² (10 g) of Head TSL	Condition	Value																																																													
SAR measured	250 mW input power	6.12 W/kg																																																													
SAR for nominal Head TSL parameters	normalized to 1W	24.6 W/kg ± 18.7 % (k=2)																																																													
Parameter	Value																																																														
Impedance, transformed to feed point	49.90- 6.49jΩ																																																														
Return Loss	- 23.8dB																																																														
Parameter	Value																																																														
Electrical Delay (one direction)	1.053 ns																																																														
Manufacturer by	Value																																																														
Manufacturer by	SPEAG																																																														

In Collaboration with
TTL Speaq
 CALIBRATION LABORATORY

CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62104633-2079 Fax: +86-10-62104633-2504
 E-mail: cti@china.ttl.com http://www.china.ttl.com

Date: 2022-03-31

DASY5 Validation Report for Head TSL
 Test Laboratory: CTTL, Beijing, China
 DUT: Dipole 2600 MHz; Type: D2600V2; Serial: D2600V2 - SN: 1158
 Communication System: UID 0, CW; Frequency: 2600 MHz; Duty Cycle: 1:1
 Medium parameters used: $f = 2600 \text{ MHz}$; $\sigma = 1.955 \text{ S/m}$; $\epsilon_r = 38.68$; $\rho = 1000 \text{ kg/m}^3$
 Phantom section: Right Section
 Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)
 DASY5 Configuration:

- Probe: EX3DV4 - SN7307; ConvF(7.5, 7.5, 7.5) @ 2600 MHz; Calibrated: 2021-05-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1556; Calibrated: 2022-01-12
- Phantom: MFP-V5.1C (2dkg probe fill); Type: QD 000 P51 Cx; Serial: 1062
- DASY5 52.10.4(1535); SEMCAD X 14.6.14(7501)

Dipole Calibration/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm
 Reference Value = 103.3 V/m; Power Drift = 0.04 dB
 Peak SAR (extrapolated) = 29.0 W/kg
 SAR(1 g) = 13.7 W/kg; SAR(10 g) = 6.12 W/kg
 Smallest distance from peaks to all points 3 dB below = 8.9 mm
 Ratio of SAR at M2 to SAR at M1 = 47.5%
 Maximum value of SAR (measured) = 23.4 W/kg

0 dB = 23.4 W/kg = 13.69 dBW/kg

Certificate No: Z22-60108 Page 5 of 6

In Collaboration with
TTL Speaq
 CALIBRATION LABORATORY

CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62104633-2079 Fax: +86-10-62104633-2504
 E-mail: cti@china.ttl.com http://www.china.ttl.com

Impedance Measurement Plot for Head TSL

Certificate No: Z22-60108 Page 6 of 6

1.12 D5GHZV2 - SN 1095

In Collaboration with
TTL Speaq
 CALIBRATION LABORATORY

CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62104633-2079 Fax: +86-10-62104633-2504
 E-mail: cti@china.ttl.com http://www.china.ttl.com

Client: **SGS-CN** Certificate No: **Z22-60187**

CALIBRATION CERTIFICATE

Object: D5GHZV2 - SN: 1095

Calibration Procedure(s): FF-Z11-003-01
 Calibration Procedures for dipole validation kits

Calibration date: June 1, 2022

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (23±3)°C and humidity<70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by: Certificate No.)	Scheduled Calibration
Power Meter NRP2	106277	24-Sep-21 (CTTL No.J21008326)	Sep-22
Power sensor NRP8S	104291	24-Sep-21 (CTTL No.J21008326)	Sep-22
Reference Probe EX3DV4	SN 7464	26-Jan-22(SPEAG.No.EX3-7464_Jan22)	Jan-23
DAE4	SN 1556	12-Jan-22(CTTL-SPEAG.No.Z22-60007)	Jan-23

Secondary Standards	ID #	Cal Date (Calibrated by: Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY46071430	13-Jan-22 (CTTL No. J22X00406)	Jan-23
Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL No. J22X00406)	Jan-23

Calibrated by: Zhao Jing SAR Test Engineer

Reviewed by: Lin Hao SAR Test Engineer

Approved by: Qi Dianyan SAR Project Leader

Issued: June 6, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60187 Page 1 of 10

In Collaboration with
TTL Speaq
 CALIBRATION LABORATORY

CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62104633-2079 Fax: +86-10-62104633-2504
 E-mail: cti@china.ttl.com http://www.china.ttl.com

Glossary:

TSL Issue simulating liquid
 ConF sensitivity in TSL / NORMx,y,z
 N/A not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices-Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020
- KDB 665664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

- DASY4/G System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60187 Page 2 of 10

In Collaboration with
TTL
CALIBRATION LABORATORY

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62382117
E-mail: cti@ttsintl.com http://www.ttsintl.com

Measurement Conditions
DASY system configuration, as far as not given on page 1.

DASY Version	DASY52	52.10.4
Extrapolation	Advanced Extrapolation	
Phantom	Triple Flat Phantom 5.1C	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy = 4 mm, dz = 1.4 mm	Graded Ratio = 1.4 (Z direction)
Frequency	5200 MHz ± 1 MHz 5200 MHz ± 1 MHz 5500 MHz ± 1 MHz 5800 MHz ± 1 MHz	

Head TSL parameters at 5200MHz
The following parameters and calculations were applied.

Nominal Head TSL parameters	Temperature	Permittivity	Conductivity
	22.0 °C	35.0	4.66 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.4 ± 6 %	4.82 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL at 5200MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	7.79 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	17.8 W/kg ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	250 mW input power	2.22 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	22.1 W/kg ± 24.2 % (k=2)

Certificate No: Z22-60187 Page 3 of 10

In Collaboration with
TTL
CALIBRATION LABORATORY

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62382117
E-mail: cti@ttsintl.com http://www.ttsintl.com

Head TSL parameters at 5300MHz
The following parameters and calculations were applied.

Nominal Head TSL parameters	Temperature	Permittivity	Conductivity
	22.0 °C	35.9	4.76 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.2 ± 6 %	4.73 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL at 5300MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	7.94 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	79.1 W/kg ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	100 mW input power	2.27 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	22.6 W/kg ± 24.2 % (k=2)

Head TSL parameters at 5500MHz
The following parameters and calculations were applied.

Nominal Head TSL parameters	Temperature	Permittivity	Conductivity
	22.0 °C	35.6	4.96 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	34.8 ± 6 %	4.94 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL at 5500MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	8.29 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	82.6 W/kg ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	100 mW input power	2.34 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	23.3 W/kg ± 24.2 % (k=2)

Certificate No: Z22-60187 Page 4 of 10

In Collaboration with
TTL
CALIBRATION LABORATORY

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62382117
E-mail: cti@ttsintl.com http://www.ttsintl.com

Head TSL parameters at 5600MHz
The following parameters and calculations were applied.

Nominal Head TSL parameters	Temperature	Permittivity	Conductivity
	22.0 °C	35.5	5.07 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	34.7 ± 6 %	5.05 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL at 5600MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	8.12 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	80.8 W/kg ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	100 mW input power	2.30 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	22.9 W/kg ± 24.2 % (k=2)

Head TSL parameters at 5800MHz
The following parameters and calculations were applied.

Nominal Head TSL parameters	Temperature	Permittivity	Conductivity
	22.0 °C	35.3	5.27 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	34.4 ± 6 %	5.25 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL at 5800MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	7.71 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	76.7 W/kg ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	100 mW input power	2.16 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	21.8 W/kg ± 24.2 % (k=2)

Certificate No: Z22-60187 Page 5 of 10

In Collaboration with
TTL
CALIBRATION LABORATORY

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62382117
E-mail: cti@ttsintl.com http://www.ttsintl.com

Appendix (Additional assessments outside the scope of CNAS L0570)

Antenna Parameters with Head TSL at 5200MHz

Impedance, transformed to feed point	46.10-5.03jΩ
Return Loss	-23.6dB

Antenna Parameters with Head TSL at 5300MHz

Impedance, transformed to feed point	47.80-2.42jΩ
Return Loss	-28.5dB

Antenna Parameters with Head TSL at 5500MHz

Impedance, transformed to feed point	50.30-4.26jΩ
Return Loss	-27.4dB

Antenna Parameters with Head TSL at 5600MHz

Impedance, transformed to feed point	54.50-4.80jΩ
Return Loss	-24.0dB

Antenna Parameters with Head TSL at 5800MHz

Impedance, transformed to feed point	51.50-5.61jΩ
Return Loss	-24.9dB

Certificate No: Z22-60187 Page 6 of 10

In Collaboration with
speaq
CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62302117
 E-mail: ott@speaq.com <http://www.caict.ac.cn>

General Antenna Parameters and Design

Electrical Delay (one direction)	1.101 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.

Additional EUT Data

Manufactured by	SPEAG
-----------------	-------

Certificate No: Z22-60187 Page 7 of 10

In Collaboration with
speaq
CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62302117
 E-mail: ott@speaq.com <http://www.caict.ac.cn>

Date: 2022-06-01

DASY5 Validation Report for Head TSL

Test Laboratory: CTTL, Beijing, China
DUT: Dipole 5GHz; Type: D5GHzV2; Serial: D5GHzV2 - SN: 1095

Communication System: CW; Frequency: 5200 MHz; Frequency: 5300 MHz; Frequency: 5500 MHz; Frequency: 5600 MHz; Frequency: 5800 MHz; Duty Cycle: 1:1
 Medium parameters used: f = 5200 MHz; $\sigma = 4.62$ S/m; $\epsilon_r = 35.38$; $\rho = 1000$ kg/m³
 Medium parameters used: f = 5300 MHz; $\sigma = 4.73$ S/m; $\epsilon_r = 35.19$; $\rho = 1000$ kg/m³
 Medium parameters used: f = 5500 MHz; $\sigma = 4.939$ S/m; $\epsilon_r = 34.83$; $\rho = 1000$ kg/m³
 Medium parameters used: f = 5600 MHz; $\sigma = 5.051$ S/m; $\epsilon_r = 34.68$; $\rho = 1000$ kg/m³
 Medium parameters used: f = 5800 MHz; $\sigma = 5.247$ S/m; $\epsilon_r = 34.42$; $\rho = 1000$ kg/m³
 Phantom section: Right Section
 Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)
 DASY5 Configuration:

- Probe: EX3DV4 - SN7484; ConvF(5.6, 5.6, 5.6) @ 5200 MHz; ConvF(5.32, 5.32, 5.32) @ 5300 MHz; ConvF(5.11, 5.11, 5.11) @ 5500 MHz; ConvF(4.91, 4.91, 4.91) @ 5600 MHz; ConvF(5, 5, 5) @ 5800 MHz; Calibrated: 2022-01-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DA4 Sn1556; Calibrated: 2022-01-12
- Phantom: MFP_V5.1C (20deg probe tilt); Type: GD 000 P51 Cx; Serial: 1062
- DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501)

Dipole Calibration /Pin=100mW, d=10mm, f=5200 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
 Reference Value = 60.80 V/m; Power Drift = -0.06 dB
 Peak SAR (extrapolated) = 29.8 W/kg
SAR(1 g) = 7.79 W/kg; SAR(10 g) = 2.22 W/kg
 Smallest distance from peaks to all points 3 dB below = 7.2 mm
 Ratio of SAR at M2 to SAR at M1 = 66.8%
 Maximum value of SAR (measured) = 18.3 W/kg

Dipole Calibration /Pin=100mW, d=10mm, f=5300 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
 Reference Value = 61.08 V/m; Power Drift = -0.07 dB
 Peak SAR (extrapolated) = 31.5 W/kg
SAR(1 g) = 7.94 W/kg; SAR(10 g) = 2.27 W/kg
 Smallest distance from peaks to all points 3 dB below = 7.2 mm
 Ratio of SAR at M2 to SAR at M1 = 65.5%
 Maximum value of SAR (measured) = 19.0 W/kg

Certificate No: Z22-60187 Page 8 of 10

In Collaboration with
speaq
CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62302117
 E-mail: ott@speaq.com <http://www.caict.ac.cn>

Dipole Calibration /Pin=100mW, d=10mm, f=5500 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
 Reference Value = 61.92 V/m; Power Drift = -0.08 dB
 Peak SAR (extrapolated) = 34.7 W/kg
SAR(1 g) = 8.28 W/kg; SAR(10 g) = 2.34 W/kg
 Smallest distance from peaks to all points 3 dB below = 7.2 mm
 Ratio of SAR at M2 to SAR at M1 = 63.9%
 Maximum value of SAR (measured) = 20.2 W/kg

Dipole Calibration /Pin=100mW, d=10mm, f=5600 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
 Reference Value = 65.08 V/m; Power Drift = -0.07 dB
 Peak SAR (extrapolated) = 35.2 W/kg
SAR(1 g) = 8.12 W/kg; SAR(10 g) = 2.3 W/kg
 Smallest distance from peaks to all points 3 dB below = 7.2 mm
 Ratio of SAR at M2 to SAR at M1 = 62.5%
 Maximum value of SAR (measured) = 19.1 W/kg

Dipole Calibration /Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
 Reference Value = 62.13 V/m; Power Drift = -0.06 dB
 Peak SAR (extrapolated) = 34.8 W/kg
SAR(1 g) = 7.71 W/kg; SAR(10 g) = 2.16 W/kg
 Smallest distance from peaks to all points 3 dB below = 7.2 mm
 Ratio of SAR at M2 to SAR at M1 = 61.6%
 Maximum value of SAR (measured) = 18.7 W/kg

0 dB = 18.7 W/kg = 12.72 dBW/kg

Certificate No: Z22-60187 Page 9 of 10

In Collaboration with
speaq
CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62302117
 E-mail: ott@speaq.com <http://www.caict.ac.cn>

Impedance Measurement Plot for Head TSL

Certificate No: Z22-60187 Page 10 of 10

2 DAE4 - SN 1245

<p>Schmid & Partner Engineering AG 2025-01-15-15:43:00 (UTC+08:00) Phone: +41 41 245 3700, Fax: +41 41 245 3779 www.speag.com, info@speag.com</p> <p style="text-align: center;">s p e a g</p> <p style="text-align: center;">IMPORTANT NOTICE</p> <p>USAGE OF THE DAE4</p> <p>The DAE unit is a delicate, high precision instrument and requires careful treatment by the user. There are no serviceable parts inside the DAE. Special attention shall be given to the following points:</p> <p>Battery Exchange: The battery cover of the DAE4 unit is fixed using a screw, over tightening the screw may cause the threads inside the DAE to wear out.</p> <p>Shipping of the DAE: Before shipping the DAE to SPEAG for calibration, remove the batteries and pack the DAE in an anti-static bag. This anti-static bag shall then be packed into a larger box or container which protects the DAE from impacts during transportation. The package shall be marked to indicate that a fragile instrument is inside.</p> <p>E-Stop Failures: Touch detection may be malfunctioning due to broken magnets in the E-stop. Rough handling of the E-stop may lead to damage of these magnets. Touch and collision errors are often caused by dust and get accumulated in the E-stop. To prevent E-stop failure, the customer shall always mount the probe to the DAE carefully and keep the DAE unit in a non-dusty environment if not used for measurements.</p> <p>Repair: Minor repairs are performed at no extra cost during the annual calibration. However, SPEAG reserves the right to charge for any repair especially if rough unprofessional handling caused the defect.</p> <p>DASY Configuration File: Since the exact values of the DAE input resistances, as measured during the calibration procedure of a DAE unit, are not used by the DASY software, a nominal value of 200 MΩ is given in the corresponding configuration file.</p> <p>Important Note: Warranty and calibration is void if the DAE unit is disassembled partly or fully by the Customer.</p> <p>Important Note: Never attempt to grease or oil the E-stop assembly. Cleaning and readjusting of the E-stop assembly is allowed by certified SPEAG personnel only and is part of the annual calibration procedure.</p> <p>Important Note: To prevent damage of the DAE probe connector pins, use great care when installing the probe to the DAE. Carefully connect the probe with the connector notch oriented in the while turning the locking nut of the connector. The same care shall be used when disconnecting the probe from the DAE.</p> <p>TL_EH190306AE DAE4.docx 07.03.2019</p>	<p>Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p style="text-align: right;">Schweizerischer Kalibrierdienst Service suisse d'étalonnage Servizio svizzero di taratura Swiss Calibration Service</p> <p style="text-align: right;">Accreditation No.: SCS 0108</p> <p>Client: SGS Certificate No: DAE4-1245_Apr23 Kunshan City, China</p> <p style="text-align: center;">CALIBRATION CERTIFICATE</p> <p>Object: DAE4 - SD 000 D04 BM - SN: 1245</p> <p>Calibration procedure(s): QA CAL-06.v30 Calibration procedure for the data acquisition electronics (DAE)</p> <p>Calibration date: April 25, 2023</p> <p>This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurement and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility, environment temperature (22 ± 3°C and humidity < 70%). Calibration Equipment used (MATE critical for calibration)</p> <table border="1"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Kathley Multimeter Type 2001</td> <td>SN: 0810278</td> <td>29-Aug-22 (No. 34389)</td> <td>Aug-23</td> </tr> </tbody> </table> <table border="1"> <thead> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Check Date (in house)</th> <th>Scheduled Check</th> </tr> </thead> <tbody> <tr> <td>Auto DAE Calibration Unit</td> <td>SE UMS 002 AA 1001</td> <td>27-Jan-23 (in house check)</td> <td>In house check: Jan-24</td> </tr> <tr> <td>Calibrator Box V2.1</td> <td>SE UMS 008 AA 1002</td> <td>27-Jan-23 (in house check)</td> <td>In house check: Jan-24</td> </tr> </tbody> </table> <p>Calibrated by: Dominique Sellen (Name) / Laboratory Technician (Function) / <i>[Signature]</i> (Signature)</p> <p>Approved by: Steen Kuhn (Name) / Technical Manager (Function) / <i>[Signature]</i> (Signature)</p> <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory. Issued: April 25, 2023</p> <p>Certificate No: DAE4-1245_Apr23 Page 1 of 5</p>	Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration	Kathley Multimeter Type 2001	SN: 0810278	29-Aug-22 (No. 34389)	Aug-23	Secondary Standards	ID #	Check Date (in house)	Scheduled Check	Auto DAE Calibration Unit	SE UMS 002 AA 1001	27-Jan-23 (in house check)	In house check: Jan-24	Calibrator Box V2.1	SE UMS 008 AA 1002	27-Jan-23 (in house check)	In house check: Jan-24
Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration																		
Kathley Multimeter Type 2001	SN: 0810278	29-Aug-22 (No. 34389)	Aug-23																		
Secondary Standards	ID #	Check Date (in house)	Scheduled Check																		
Auto DAE Calibration Unit	SE UMS 002 AA 1001	27-Jan-23 (in house check)	In house check: Jan-24																		
Calibrator Box V2.1	SE UMS 008 AA 1002	27-Jan-23 (in house check)	In house check: Jan-24																		
<p>Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p style="text-align: right;">Schweizerischer Kalibrierdienst Service suisse d'étalonnage Servizio svizzero di taratura Swiss Calibration Service</p> <p style="text-align: right;">Accreditation No.: SCS 0108</p> <p>Glossary</p> <p>DAE: data acquisition electronics</p> <p>Connector angle: information used in DASY system to align probe sensor X to the robot coordinate system.</p> <p>Methods Applied and Interpretation of Parameters</p> <ul style="list-style-type: none"> DC Voltage Measurement: Calibration Factor assessed for use in DASY system by comparison with a calibrated instrument traceable to national standards. The figure given corresponds to the full scale range of the voltmeter in the respective range. Connector angle: The angle of the connector is assessed measuring the angle mechanically by a tool inserted. Uncertainty is not required. The following parameters as documented in the Appendix contain technical information as a result from the performance test and require no uncertainty. <ul style="list-style-type: none"> DC Voltage Measurement Linearity: Verification of the Linearity at +10% and -10% of the nominal calibration voltage. Influence of offset voltage is included in this measurement. Common mode sensitivity: Influence of a positive or negative common mode voltage on the differential measurement. Channel separation: Influence of a voltage on the neighbor channels not subject to an input voltage. AD Converter Values with inputs shorted: Values on the internal AD converter corresponding to zero input voltage Input Offset Measurement: Output voltage and statistical results over a large number of zero voltage measurements. Input Offset Current: Typical value for information; Maximum channel input offset current, not considering the input resistance. Input resistance: Typical value for information; DAE input resistance at the connector, during internal auto-zeroing and during measurement. Low Battery Alarm Voltage: Typical value for information. Below this voltage, a battery alarm signal is generated. Power consumption: Typical value for information. Supply currents in various operating modes. <p>Certificate No: DAE4-1245_Apr23 Page 2 of 5</p>	<p>DC Voltage Measurement</p> <p>AD - Converter Resolution nominal</p> <p>High Range: 1LSB = 6.1μV, full range = -100...+300 mV</p> <p>Low Range: 1LSB = 61nV, full range = -1...+30mV</p> <p>DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec</p> <table border="1"> <thead> <tr> <th>Calibration Factors</th> <th>X</th> <th>Y</th> <th>Z</th> </tr> </thead> <tbody> <tr> <td>High Range</td> <td>405.243 ± 0.02% (k=2)</td> <td>403.938 ± 0.02% (k=2)</td> <td>405.064 ± 0.02% (k=2)</td> </tr> <tr> <td>Low Range</td> <td>3.99474 ± 1.50% (k=2)</td> <td>3.99478 ± 1.50% (k=2)</td> <td>4.00994 ± 1.50% (k=2)</td> </tr> </tbody> </table> <p>Connector Angle</p> <table border="1"> <tr> <td>Connector Angle to be used in DASY system</td> <td>32.0° ± 1°</td> </tr> </table> <p>Certificate No: DAE4-1245_Apr23 Page 3 of 5</p>	Calibration Factors	X	Y	Z	High Range	405.243 ± 0.02% (k=2)	403.938 ± 0.02% (k=2)	405.064 ± 0.02% (k=2)	Low Range	3.99474 ± 1.50% (k=2)	3.99478 ± 1.50% (k=2)	4.00994 ± 1.50% (k=2)	Connector Angle to be used in DASY system	32.0° ± 1°						
Calibration Factors	X	Y	Z																		
High Range	405.243 ± 0.02% (k=2)	403.938 ± 0.02% (k=2)	405.064 ± 0.02% (k=2)																		
Low Range	3.99474 ± 1.50% (k=2)	3.99478 ± 1.50% (k=2)	4.00994 ± 1.50% (k=2)																		
Connector Angle to be used in DASY system	32.0° ± 1°																				

Appendix (Additional assessments outside the scope of SCS0108)

1. DC Voltage Linearity

High Range

Channel	Reading (µV)	Difference (µV)	Error (%)
Channel X + Input	19998.60	2.90	0.00
Channel X - Input	20005.77	2.75	0.01
Channel X + Input	19998.60	2.19	-0.01
Channel Y + Input	19998.60	1.08	0.00
Channel Y - Input	20003.12	0.26	0.00
Channel Z + Input	-20000.51	0.53	-0.00
Channel Z - Input	199994.62	-1.05	-0.00
Channel Z + Input	20002.17	-0.70	-0.00
Channel Z - Input	-20001.94	-0.91	0.00

Low Range

Channel	Reading (µV)	Difference (µV)	Error (%)
Channel X + Input	2002.91	0.81	0.04
Channel X - Input	203.06	0.73	0.36
Channel X + Input	-196.56	0.88	-0.45
Channel Y + Input	2002.33	0.29	0.01
Channel Y - Input	201.91	-0.39	-0.19
Channel Z + Input	-196.22	-0.79	0.40
Channel Z - Input	2002.20	0.24	0.01
Channel Z + Input	201.28	-0.88	-0.44
Channel Z - Input	-196.93	-1.36	0.69

2. Common mode sensitivity

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

Channel	Common mode Input Voltage (mV)	High Range Average Reading (µV)	Low Range Average Reading (µV)
Channel X	200	-6.42	-8.27
Channel X	-200	8.81	6.00
Channel Y	200	7.04	7.13
Channel Y	-200	-14.70	-15.29
Channel Z	200	-4.52	-5.35
Channel Z	-200	3.50	3.52

3. Channel separation

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

Channel	Input Voltage (mV)	Channel X (µV)	Channel Y (µV)	Channel Z (µV)
Channel X	200	3.29	-3.29	-
Channel Y	200	9.00	-	4.00
Channel Z	200	10.03	7.20	-

Certificate No: DAE4-1245_Apr23 Page 4 of 5

4. AD-Converter Values with inputs shorted

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

Channel	High Range (LSB)	Low Range (LSB)
Channel X	16001	16100
Channel Y	16079	16051
Channel Z	16040	15991

5. Input Offset Measurement

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

Channel	Average (µV)	min. Offset (µV)	max. Offset (µV)	Std. Deviation (µV)
Channel X	0.77	-0.83	1.89	0.49
Channel Y	-0.24	-1.72	1.19	0.52
Channel Z	-0.85	-2.62	0.99	0.61

6. Input Offset Current

Nominal input circuitry offset current on all channels: <25µA

7. Input Resistance (Typical values for information)

Channel	Zeroing (MΩ)	Measuring (MΩ)
Channel X	200	200
Channel Y	200	200
Channel Z	200	200

8. Low Battery Alarm Voltage (Typical values for information)

Typical values	Alarm Level (VDC)
Supply (+ Vcc)	+7.9
Supply (- Vcc)	-7.6

9. Power Consumption (Typical values for information)

Typical values	Switched off (mA)	Stand by (mA)	Transmitting (mA)
Supply (+ Vcc)	+0.01	+6	+14
Supply (- Vcc)	-0.01	-8	-9

Certificate No: DAE4-1245_Apr23 Page 5 of 5

3 EX3DV4 - SN 7767

<p>Calibration Laboratory of Schmid & Partner Engineering AG Zugstrasse 43, 8004 Zurich, Switzerland</p> <p>Accredited by the Swiss Accreditation Service (SAS). The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p>Client: SOS Kunshan</p> <p>Certificate No.: EX-7767_Oct23</p> <p>CALIBRATION CERTIFICATE</p> <p>Object: EX3DV4 - SN:7767</p> <p>Calibration procedure(s): QA CAL-01 v10, QA CAL-12 v10, QA CAL-14 v7, QA CAL-23 v6, QA CAL-25 v6 Calibration procedure for dosimetric E-field probes</p> <p>Calibration date: October 26, 2023</p> <p>This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. All calibrations have been conducted in the closest laboratory facility; environment temperature (22 ± 3°C and humidity < 70%). Calibration Equipment used (MATE unless for calibration):</p> <table border="1"> <thead> <tr> <th>Primary Standards</th> <th>ID</th> <th>Cal Date (Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr><td>Power meter NRP2</td><td>SN: 104778</td><td>30-Mar-23 (No. 217-08854-05855)</td><td>Mar-24</td></tr> <tr><td>Power sensor NRP 301</td><td>SN: 103344</td><td>31-Mar-23 (No. 217-08854)</td><td>Mar-24</td></tr> <tr><td>DOF DAK-3.3 (single lead)</td><td>SN: 13489</td><td>05-Oct-23 (DOF-DAK3-1349, Oct23)</td><td>Mar-24</td></tr> <tr><td>DOF DAK-12</td><td>SN: 1819</td><td>05-Oct-23 (DOF-DAK12-1819, Oct23)</td><td>Oct-24</td></tr> <tr><td>Reference 30 dB Attenuator</td><td>SN: C23593 (20x)</td><td>30-Mar-23 (No. 217-08854)</td><td>Mar-24</td></tr> <tr><td>EMC</td><td>SN: 680</td><td>16-Mar-23 (No. DAE4-489_Mar23)</td><td>Mar-24</td></tr> <tr><td>Reference Probe ESD2V2</td><td>SN: 3013</td><td>06-Mar-23 (No. E33-3013_Apr23)</td><td>Apr-24</td></tr> </tbody> </table> <table border="1"> <thead> <tr> <th>Secondary Standards</th> <th>ID</th> <th>Check Date (in house)</th> <th>Scheduled Check</th> </tr> </thead> <tbody> <tr><td>Power meter E41198</td><td>SN: 0841983874</td><td>06-Apr-18 (in house check Jun-23)</td><td>in house check Jun-24</td></tr> <tr><td>Power sensor E4119A</td><td>SN: 0741983897</td><td>06-Apr-18 (in house check Jun-23)</td><td>in house check Jun-24</td></tr> <tr><td>Power sensor E4119A</td><td>SN: 000118219</td><td>06-Apr-18 (in house check Jun-23)</td><td>in house check Jun-24</td></tr> <tr><td>RF generator HP 8446C</td><td>SN: US8446C07100</td><td>04-Apr-18 (in house check Jun-23)</td><td>in house check Jun-24</td></tr> <tr><td>Network Analyzer E8358A</td><td>SN: US841886477</td><td>31-Mar-14 (in house check Oct-23)</td><td>in house check Oct-24</td></tr> </tbody> </table> <p>Calibrated by: Joanna Lischaj Laboratory Technician</p> <p>Approved by: Sven Kuhn Technical Manager</p> <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p> <p>Certificate No: EX-7767_Oct23 Page 1 of 9</p>	Primary Standards	ID	Cal Date (Certificate No.)	Scheduled Calibration	Power meter NRP2	SN: 104778	30-Mar-23 (No. 217-08854-05855)	Mar-24	Power sensor NRP 301	SN: 103344	31-Mar-23 (No. 217-08854)	Mar-24	DOF DAK-3.3 (single lead)	SN: 13489	05-Oct-23 (DOF-DAK3-1349, Oct23)	Mar-24	DOF DAK-12	SN: 1819	05-Oct-23 (DOF-DAK12-1819, Oct23)	Oct-24	Reference 30 dB Attenuator	SN: C23593 (20x)	30-Mar-23 (No. 217-08854)	Mar-24	EMC	SN: 680	16-Mar-23 (No. DAE4-489_Mar23)	Mar-24	Reference Probe ESD2V2	SN: 3013	06-Mar-23 (No. E33-3013_Apr23)	Apr-24	Secondary Standards	ID	Check Date (in house)	Scheduled Check	Power meter E41198	SN: 0841983874	06-Apr-18 (in house check Jun-23)	in house check Jun-24	Power sensor E4119A	SN: 0741983897	06-Apr-18 (in house check Jun-23)	in house check Jun-24	Power sensor E4119A	SN: 000118219	06-Apr-18 (in house check Jun-23)	in house check Jun-24	RF generator HP 8446C	SN: US8446C07100	04-Apr-18 (in house check Jun-23)	in house check Jun-24	Network Analyzer E8358A	SN: US841886477	31-Mar-14 (in house check Oct-23)	in house check Oct-24	<p>Calibration Laboratory of Schmid & Partner Engineering AG Zugstrasse 43, 8004 Zurich, Switzerland</p> <p>Accredited by the Swiss Accreditation Service (SAS). The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p>Client: SOS Kunshan</p> <p>Certificate No.: EX-7767_Oct23</p> <p>Glossary</p> <p>TSL: Issue simulating liquid NORM_{M,z,z'}: sensitivity in free space CorNF: sensitivity in TSL / NORM_{M,z,z'} DCP: diode compression point CF: crest factor (1.414_{rms} / cycle) of the RF signal A, B, C, D: modulation dependent (linearization) parameters Polarization ψ: ψ rotation around probe axis Polarization θ: θ rotation around an axis that is in the plane normal to probe axis (at measurement center), i.e., $\theta = 0$ is normal to probe axis Connector Angle: information used in DASY system to align probe/sensor X to the robot coordinate system</p> <p>Calibration is Performed According to the Following Standards:</p> <ul style="list-style-type: none"> IEC 61010-1:2010-1528: Measurement Procedure For The Assessment Of Specific Absorption Rate Of Human Exposure To Radio Frequency Fields From Hand Held And Body-Worn Wireless Communication Devices - Part 1528: Human Models, Instrumentation And Procedures (Frequency Range of 4 MHz to 115 GHz) - October 2020. KDB 865664: "SAR Measurement Requirements for 100 MHz to 6 GHz" <p>Methods Applied and Interpretation of Parameters:</p> <ul style="list-style-type: none"> NORM_{M,z,z'}: Assessed for E-field polarization $\theta = 0$ (≤ 900 MHz) in TEM cell; $\theta > 900$ MHz: R22 waveguide; NORM_{M,z,z'} are only intermediate values, i.e., the uncertainty of NORM_{M,z,z'} does not affect the E-field uncertainty inside TSL (see below CorNF). NORM_{M,z,z'} = NORM_{M,z,z'} * frequency_response (see Frequency Response Chart). This linearization is implemented in DASY software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of CorNF. DOF_{M,z,z'}: DCP are numerical linearization parameters assessed based on the data of power swept with CW signal. DCP does not depend on frequency nor media. RA_{M,z,z'}: RA is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics. A_{M,z,z'}, B_{M,z,z'}, C_{M,z,z'}, D_{M,z,z'}, VR_{M,z,z'}: A, B, C, D are numerical linearization parameters assessed based on the data of power swept for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode. CorNF and Boundary Effect Parameters: Assessed in flat phantom using E-field (or Temperature Transfer Standard for ≤ 60 MHz) and inside waveguide using analytical field distributions based on power measurements for > 900 MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha_{bc}, delta_{bc}) of which typical boundary. The sensitivity in TSL corresponds to NORM_{M,z,z'} * CorNF whereby the uncertainty corresponds to the given CorNF. A frequency dependent CorNF is used in DASY version 4.4 and higher which allows extending the validity from ≤ 50 MHz to ≤ 100 MHz. Spherical Isotropy (2D deviation from isotropy): in a field of low gradients realized using a flat phantom exposed by a patch antenna. Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No isolation required. Connector Angle: The angle is assessed using the information gained by determining the NORM_M (no uncertainty required). <p>Certificate No: EX-7767_Oct23 Page 2 of 9</p>
Primary Standards	ID	Cal Date (Certificate No.)	Scheduled Calibration																																																						
Power meter NRP2	SN: 104778	30-Mar-23 (No. 217-08854-05855)	Mar-24																																																						
Power sensor NRP 301	SN: 103344	31-Mar-23 (No. 217-08854)	Mar-24																																																						
DOF DAK-3.3 (single lead)	SN: 13489	05-Oct-23 (DOF-DAK3-1349, Oct23)	Mar-24																																																						
DOF DAK-12	SN: 1819	05-Oct-23 (DOF-DAK12-1819, Oct23)	Oct-24																																																						
Reference 30 dB Attenuator	SN: C23593 (20x)	30-Mar-23 (No. 217-08854)	Mar-24																																																						
EMC	SN: 680	16-Mar-23 (No. DAE4-489_Mar23)	Mar-24																																																						
Reference Probe ESD2V2	SN: 3013	06-Mar-23 (No. E33-3013_Apr23)	Apr-24																																																						
Secondary Standards	ID	Check Date (in house)	Scheduled Check																																																						
Power meter E41198	SN: 0841983874	06-Apr-18 (in house check Jun-23)	in house check Jun-24																																																						
Power sensor E4119A	SN: 0741983897	06-Apr-18 (in house check Jun-23)	in house check Jun-24																																																						
Power sensor E4119A	SN: 000118219	06-Apr-18 (in house check Jun-23)	in house check Jun-24																																																						
RF generator HP 8446C	SN: US8446C07100	04-Apr-18 (in house check Jun-23)	in house check Jun-24																																																						
Network Analyzer E8358A	SN: US841886477	31-Mar-14 (in house check Oct-23)	in house check Oct-24																																																						

EX3DV4 - SN:7767 October 26, 2023

Parameters of Probe: EX3DV4 - SN:7767

Basic Calibration Parameters

	Sensor X	Sensor Y	Sensor Z	Unc. (k=2)
Norm. $\mu V/(V/m)^2$ ^A	0.68	0.72	0.58	$\pm 10.1\%$
DCP $\mu V/m^2$ ^B	102.4	105.4	106.6	$\pm 4.7\%$

Calibration Results for Modulation Response

UID	Communication System Name	A	B	C	D	VR	Max dev.	Max Unc. (k=2)
		dB	dB \sqrt{V}		dB	mV		
0	CW	X: 0.00	0.00	1.00	0.00	100.0	$\pm 1.9\%$	$\pm 4.7\%$
		Y: 0.00	0.00	1.00		138.5		
		Z: 0.00	0.00	1.00		138.8		

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

^A The uncertainties of Norm. X, Y, Z do not meet the C^2 field uncertainty limits (3), (see Page 6).
^B Uncalibration parameter uncertainty for maximum specified field strength.
^C Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

EX3DV4 - SN:7767 October 26, 2023

Parameters of Probe: EX3DV4 - SN:7767

Other Probe Parameters

Sensor Arrangement	Triangular
Connector Angle	$\pm 40.9^\circ$
Mechanical Surface Detection Mode	enabled
Optical Surface Detection Mode	disabled
Probe Overall Length	337 mm
Probe Body Diameter	10 mm
Tip Length	9 mm
Tip Diameter	2.5 mm
Probe Tip to Sensor X Calibration Point	1 mm
Probe Tip to Sensor Y Calibration Point	1 mm
Probe Tip to Sensor Z Calibration Point	1 mm
Recommended Measurement Distance from Surface	1.4 mm

Note: Measurement distance from surface can be increased to 3.4 mm for an Area Scan job.

EX3DV4 - SN:7767 October 26, 2023

Parameters of Probe: EX3DV4 - SN:7767


Calibration Parameter Determined in Head Tissue Simulating Media

f (MHz) ^C	Relative Permittivity ^D	Conductivity ^E (S/m)	ConvF X	ConvF Y	ConvF Z	Alpha ^G	Depth ^H (mm)	Unc. (k=2)
150	52.3	0.76	13.37	13.37	13.37	0.00	1.25	$\pm 13.3\%$
450	43.5	0.87	11.62	11.62	11.62	0.16	1.36	$\pm 13.3\%$
750	41.9	0.89	10.54	10.54	10.57	0.36	1.27	$\pm 12.0\%$
835	41.5	0.90	10.38	9.66	10.00	0.38	1.27	$\pm 12.0\%$
900	41.5	0.97	10.23	9.78	9.99	0.38	1.27	$\pm 12.0\%$
1750	40.1	1.37	9.45	8.73	9.34	0.24	1.27	$\pm 12.0\%$
1900	40.0	1.40	8.54	7.96	8.42	0.28	1.27	$\pm 12.0\%$
3000	39.8	1.49	8.41	7.82	8.34	0.28	1.27	$\pm 12.0\%$
3200	39.6	1.67	8.70	8.12	8.61	0.39	1.27	$\pm 12.0\%$
2450	39.2	1.80	8.36	7.75	8.29	0.29	1.27	$\pm 12.0\%$
2600	39.0	1.98	8.01	7.49	7.95	0.27	1.27	$\pm 12.0\%$
3300	38.2	2.71	7.41	6.83	7.31	0.33	1.27	$\pm 14.0\%$
3500	37.9	2.91	7.64	7.02	7.52	0.34	1.27	$\pm 14.0\%$
3700	37.7	3.12	7.21	6.82	7.10	0.34	1.27	$\pm 14.0\%$
3900	37.5	3.82	7.17	6.99	7.08	0.34	1.27	$\pm 14.0\%$
4100	37.2	3.53	6.95	6.59	6.86	0.35	1.27	$\pm 14.0\%$
4200	37.1	3.63	6.64	6.09	6.55	0.35	1.27	$\pm 14.0\%$
4400	36.9	3.94	6.59	6.06	6.50	0.36	1.27	$\pm 14.0\%$
4600	36.7	4.04	6.05	6.11	6.07	0.36	1.27	$\pm 14.0\%$
4800	36.4	4.25	6.78	6.23	6.87	0.36	1.27	$\pm 14.0\%$
4900	36.3	4.40	6.43	5.87	6.30	0.39	1.36	$\pm 14.0\%$
5200	36.0	4.66	6.01	5.50	5.93	0.30	1.60	$\pm 14.0\%$
5300	35.9	4.76	5.90	5.36	5.75	0.29	1.67	$\pm 14.0\%$
5500	35.6	4.96	5.53	5.07	5.46	0.32	1.70	$\pm 14.0\%$
5600	35.5	5.07	5.25	4.78	5.16	0.34	1.75	$\pm 14.0\%$
5800	35.3	5.27	5.26	4.79	5.17	0.34	1.86	$\pm 14.0\%$

^C Frequency validity above 300 MHz of $\pm 10\%$ only applies for DASY v4.4 and higher (see Page 3), else it is restricted to $\pm 30\%$. The uncertainty in the RMS of the ConvF parameters at calibration frequencies and the uncertainty for the measured frequency band. Frequency validity below 300 MHz is $\pm 4\%$, 30, 40, 50 and 70 MHz for ConvF assessments at 30, 40, 50, 100 and 200 MHz respectively. Validity of ConvF assessment at 8 MHz to 4-MHz, and ConvF assessment at 1500 Hz to 10 Hz. Above 500 Hz frequency validity can be extended to 110 MHz.
^D The values are calculated using tissue simulating media (TSM) that deviate for x and y by less than 0.5% from the target values (typically better than 0.5%) and for z by 1%, with deviation from the target of less than 0.5%. For all cases, the calibration uncertainty was 1.1% for \sqrt{E} and 0.5% for \sqrt{H} .
^E Alpha-Delta was determined during calibration. SPREAD indicates that the remaining deviation due to the boundary effect after compensation is always less than 1.5% for frequencies below 3 GHz and below $\pm 2\%$ for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

EX3DV4 - SN:7767 October 26, 2023


Frequency Response of E-Field (TEM-Cell:0110 EXX, Waveguide:922)



Uncertainty of Frequency Response of E-Field: $\pm 0.3\%$ (k=2)

EX3DV4 - SN:7767

October 26, 2023




Certificate No: EX-7767_Oct23

Page 7 of 9

EX3DV4 - SN:7767

October 26, 2023




Certificate No: EX-7767_Oct23

Page 8 of 9

EX3DV4 - SN:7767


October 26, 2023

Conversion Factor Assessment



Deviation from Isotropy in Liquid

Error (ϕ, θ), $f = 900\text{MHz}$



Certificate No: EX-7767_Oct23

Page 9 of 9

4 Impedance and return loss

Dipole CLA150 SN 4025				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2021/4/26	-31.4	/	47.8	/
2022/4/26	-32.5	-3.5%	47.1	0.7
2023/4/26	-32.3	-2.87%	46.5	1.3
Dipole D450V3 SN 1103				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2021/4/21	-23	/	57.1	/
2022/4/26	-23.4	-1.74%	56.6	0.5
2023/4/26	-23.9	-3.91%	56.2	0.9
Dipole D750V3 SN 1188				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/3/29	-28.7	/	53.6	/
2023/3/29	-28.3	1.39%	53.2	0.4
Dipole D835V2 SN 4d114				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/3/31	-25.3	/	48.7	/
2023/3/31	-24.6	2.77%	49.1	0.4
Dipole D900V2 SN 1d079				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/6/7	-23.3	/	48.1	/
2023/6/7	-23.6	-1.29%	48.3	0.2
Dipole D1800V2 SN 2d170				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/3/31	-29.4	/	47.9	/
2023/3/31	-28.9	1.70%	47.2	0.7
Dipole D1900V2 SN 5d136				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/6/7	-22.4	/	51.2	/
2023/6/7	-22.9	-2.23%	51.6	-0.4
Dipole D2000V2 SN 1041				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/6/6	-34.9	/	48.4	/
2023/6/6	-33.5	4.01%	49.1	-0.7
Dipole D2300V2 SN 1096				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$

2022/3/31	-26.6	/	49.2	/
2023/3/31	-27.1	-1.88%	49.4	0.2
Dipole D2450V2 SN 817				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/4/1	-28.5	/	52.1	/
2023/4/1	-28.0	1.75%	51.6	0.5
Dipole D2600V2 SN 1158				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/3/31	-23.8	/	49.9	/
2023/3/31	-23.3	2.10%	50.3	0.4
Dipole D5GHzV2 SN 1095 for 5200				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/6/1	-23.6	/	46.1	/
2023/6/1	-23.1	2.12%	45.6	0.5
Dipole D5GHzV2 SN 1095 for 5300				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/6/1	-29.5	/	47.8	/
2023/6/1	-28.8	2.37%	46.9	0.9
Dipole D5GHzV2 SN 1095 for 5500				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/6/1	-27.4	/	50.3	/
2023/6/1	-27.6	-0.73%	50.8	-0.5
Dipole D5GHzV2 SN 1095 for 5600				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/6/1	-24.0	/	54.5	/
2023/6/1	-23.6	1.67%	54.9	-0.4
Dipole D5GHzV2 SN 1095 for 5800				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/6/1	-24.9	/	51.5	/
2023/6/1	-24.3	2.41%	51.0	0.5