

# SBO Hearing A/S

**Federal Communication Commission**  
Equipment Authorization Division, Application Processing Branch  
7435 Oakland Mills Road  
Columbia, MD 21048

**Certification and Engineering Bureau**  
Industry Canada  
Spectrum Engineering Branch  
3701 Carling Avenue, Building 94  
Ottawa, Ontario K2H 8S2, Canada

November 21th, 2018.

## **Declaration for IC and FCC Label Exemption for Philips ITE 312 hearing aid module**

With reference to Industry Canada, RSP-100 issue 11 chapter 3 Certification of Radio Apparatus and KDB 784748 D01v08 GUIDELINES FOR LABELLING AND USER INFORMATION FOR DEVICES SUBJECT TO PART 15 AND PART 18, Philips hereby declare an exemption regarding the labelling of the IC and FCC certification number on the Aurora ITE 312 wireless hearing aid families.

The granted exemption for the identical **Oticon Aurora ITE 312 radio module** under IC: **1350B-AUITE312** and FCC ID: **U28AUITE312** respectively is used for the Philips exemption declaration.

The Philips IC and FCC certification number: **11936A-AUITE312** and **2ACAH-AUITE312** is located next to the device model number in the user manual and/or on the product packaging material. The exemption declaration is based on the fact that the hearing aids will be too small (Final product size shells are individual to the patient ear) for the IC and FCC certification number to be legible without the aid of magnification while printed on a surface of the hearing aid area not exposed to wear and user functions.

For your information, our TCB is CTC Advanced GmbH in Germany.

Signed for  
SBO Hearing A/S

Smoerum, 21<sup>th</sup> of November 2018



---

Per Klaus Nielsen  
Regulatory Affairs Officer  
*RF approval specialist*