

SBO Hearing A/S

Federal Communication Commission
Equipment Authorization Division, Application Processing Branch
7435 Oakland Mills Road
Columbia, MD 21048

Certification and Engineering Bureau
Industry Canada
Spectrum Engineering Branch
3701 Carling Avenue, Building 94
Ottawa, Ontario K2H 8S2, Canada

November 19th, 2018.

Declaration for IC and FCC Label Exemption for Philips miniRITE T hearing aid module

With reference to Industry Canada, RSP-100 issue 11 chapter 3 Certification of Radio Apparatus and KDB 784748 D01v08 GUIDELINES FOR LABELLING AND USER INFORMATION FOR DEVICES SUBJECT TO PART 15 AND PART 18, Philips hereby declare an exemption regarding the labelling of the IC and FCC certification number on the Aurora BTE power wireless hearing aid families.

The granted exemption for the identical **Oticon Aurora miniRITE T radio module** under **IC: 1350B-AUMRTE** and **FCC ID: U28AUMRTE** respectively is used for the Philips exemption declaration.

The Philips IC and FCC certification number: **11936A-AUMRTE** and **2ACAH-AUMRTE** is located next to the device model number in the user manual and/or on the product packaging material. The exemption declaration is based on the fact that the hearing aids will be too small (Final product size 28.5 x 11.2 x 7.3 mm) for the IC and FCC certification number to be legible without the aid of magnification while printed on a surface of the hearing aid area not exposed to wear and user functions.

For your information, our TCB is CTC Advanced GmbH in Germany.

Signed for
SBO Hearing A/S

Smoerum, 19th of November 2018



Per Klaus Nielsen
Regulatory Affairs Officer
RF approval specialist