

# SBO Hearing A/S

**Federal Communication Commission**  
Equipment Authorization Division, Application Processing Branch  
7435 Oakland Mills Road  
Columbia, MD 21048

**Certification and Engineering Bureau**  
Industry Canada  
Spectrum Engineering Branch  
3701 Carling Avenue, Building 94  
Ottawa, Ontario K2H 8S2, Canada

May 2<sup>nd</sup> , 2022.

## **Declaration for IC and FCC Label Exemption for AM\_AU5\_KIT312 hearing aid module**

With reference to Industry Canada, RSP-100 issue 11, chapter 3 Certification of Radio Apparatus and KDB 784748 D01v08 GUIDELINES FOR LABELLING AND USER INFORMATION FOR DEVICES SUBJECT TO PART 15 AND PART 18, we, SBO Hearing A/S hereby declare an exemption regarding the labelling of the IC and FCC certification number on the AM\_AU5\_KIT10 wireless hearing aid families.

The SBO Hearing A/S IC and FCC certification number: **11936A-AU5ITE** and **2ACAHAU5ITE** is located next to the device model number in the user manual and/or on the product packaging material. The exemption declaration is based on the fact that the hearing aids will be too small (Final ear and canal shaped product size max. 25.3 x 26.0 x 19.0 mm) for the IC and FCC certification number to be legible without the aid of magnification while printed on a surface of the hearing aid area not exposed to wear and user functions.

For your information, our TCB is CTC Advanced GmbH in Germany.

Signed for  
SBO Hearing A/S



---

Per Klaus Nielsen  
Certification Manager, Regulatory Affairs  
Email: [pkni@sbohearing.com](mailto:pkni@sbohearing.com)