

# SBO Hearing A/S

## Class II Permissive Change (C2PC)

2021-04-15

### Federal Communication Commission

Equipment Authorization Division, Application Processing Branch  
7435 Oakland Mills Road  
Columbia, MD 21048

### Certification and Engineering Bureau

Industry Canada  
Spectrum Engineering Branch  
3701 Carling Avenue, Building 94  
Ottawa, Ontario K2H 8S2

### TO WHOM IT MAY CONCERN

Pursuant to the applicable paragraphs in ISED's RSP-100 §7 and CFR 47 FCC part §2.1043(b)(2) we hereby declare the following changes under Class II Permissive Change to the product:

Hearing Instrument amplifier module Aurora miniRITE

**FCC ID: 2ACAHSBMRTRC**

**IC: 11936A-SBMRTRC**

Original grant: February 20<sup>th</sup> -2019

We, SBO Hearing A/S hereby declare the following Class 2 Permissive Change to the amplifier module Aurora miniRITE related to the FCC part 15.247 / RSS-247 DTS radio:

- The change of Bluetooth BAW filter and ESD protection diode due to "End-of-Life" components changes has been performed with a minor PCB layout change handled in correlation with KDB 178919 D01, section III, subclause D + Page 3, Note 5, bullet 2.
- Only passive components have been changed.
- The Limited Modular Approval (LMA) still apply after the C2PC product update.
- No mechanical changes has been applied to the product for this C2PC.
- The radio module performance is equal to the previously certified module performance.

We, SBO Hearing A/S declare that the hearing aid amplifier module changes as documented in the test report will have a similar module performance as established in the FCC/ISED original certification.

# SBO Hearing A/S

If you have any questions, please feel free to contact me at the address shown below.

Sincerely,



Per Klaus Nielsen, Certification Manager  
SBO Hearing A/S  
Kongebakken 9  
DK-2765 Smørum, Denmark

Phone : +45 3917 7100

Email : [pkni@sbohearing.com](mailto:pkni@sbohearing.com)