

Appendix (Additional assessments outside the scope of SCS 0108)

Antenna Parameters with Head TSL

Impedance, transformed to feed point	51.7 Ω + 5.8 jΩ
Return Loss	- 24.5 dB

Antenna Parameters with Body TSL

Impedance, transformed to feed point	46.2 Ω + 6.6 jΩ
Return Loss	- 22.0 dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1 203 ps
Liectical Delay (one direction)	1.203 115

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
Manufactured on	March 28, 2008



DASY5 Validation Report for Head TSL

Date: 26.07.2017

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 1900 MHz; Type: D1900V2; Serial: D1900V2 - SN:5d101

Communication System: UID 0 - CW; Frequency: 1900 MHz Medium parameters used: f = 1900 MHz; σ = 1.39 S/m; ϵ_r = 40.7; ρ = 1000 kg/m³ Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 SN7349; ConvF(8.43, 8.43, 8.43); Calibrated: 31.05.2017;
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 28.03.2017
- Phantom: Flat Phantom 5.0 (front); Type: QD 000 P50 AA; Serial: 1001
- DASY52 52.10.0(1446); SEMCAD X 14.6.10(7417)

Dipole Calibration for Head Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 106.3 V/m; Power Drift = -0.04 dB Peak SAR (extrapolated) = 18.4 W/kg SAR(1 g) = 9.93 W/kg; SAR(10 g) = 5.23 W/kg Maximum value of SAR (measured) = 14.9 W/kg



0 dB = 14.9 W/kg = 11.73 dBW/kg

Certificate No: D1900V2-5d101_Jul17

Page 5 of 8



Impedance Measurement Plot for Head TSL



Certificate No: D1900V2-5d101_Jul17

Page 6 of 8



No. I17Z61226-SEM02 Page 94 of 86

DASY5 Validation Report for Body TSL

Date: 26.07.2017

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 1900 MHz; Type: D1900V2; Serial: D1900V2 - SN:5d101

Communication System: UID 0 - CW; Frequency: 1900 MHz Medium parameters used: f = 1900 MHz; $\sigma = 1.5$ S/m; $\epsilon_r = 54.1$; $\rho = 1000$ kg/m³ Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 SN7349; ConvF(8.2, 8.2, 8.2); Calibrated: 31.05.2017;
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 28.03.2017
- Phantom: Flat Phantom 5.0 (back); Type: QD 000 P50 AA; Serial: 1002
- DASY52 52.10.0(1446); SEMCAD X 14.6.10(7417)

Dipole Calibration for Body Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 101.8 V/m; Power Drift = -0.09 dB Peak SAR (extrapolated) = 17.6 W/kg SAR(1 g) = 10 W/kg; SAR(10 g) = 5.33 W/kg Maximum value of SAR (measured) = 14.4 W/kg



0 dB = 14.4 W/kg = 11.58 dBW/kg

Certificate No: D1900V2-5d101_Jul17

Page 7 of 8



Impedance Measurement Plot for Body TSL



Certificate No: D1900V2-5d101_Jul17

Page 8 of 8



2450 MHz Dipole Calibration Certificate

eughausstrasse 43, 8004 Zuric	ch, Switzerland	CONTRACTOR S	Schweizerischer Kalibrierdienst Service suisse d'étalonnage Servizio svizzero di taratura Swiss Calibration Service
Accredited by the Swiss Accredita The Swiss Accreditation Servic	ation Service (SAS) e is one of the signatori	es to the EA	ccreditation No.: SCS 0108
Multilateral Agreement for the r	ecognition of calibration	n certificates	
Client CTTL-BJ (Aud	en)	Certificate N	o: D2450V2-853_Jul17
CALIBRATION O	CERTIFICATI	Editoria	
Object	D2450V2 - SN:8	53	
Calibration procedure(s)	QA CAL-05.v9 Calibration proce	edure for dipole validation kits ab	ove 700 MHz
Calibration date:	July 21, 2017		
All calibrations have been conduct	cted in the closed laboratc	probability are given on the following pages are ny facility: environment temperature (22 \pm 3)°	nd are part of the certificate. C and humidity < 70%.
All calibrations have been conduc Calibration Equipment used (M& Primary Standards	ted in the closed laborate TE critical for calibration)	robability are given on the following pages ar ny facility: environment temperature (22 ± 3)° Cal Date (Certificate No.)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration
All calibrations have been conduc Calibration Equipment used (M& Primary Standards Power meter NRP	TE critical for calibration)	robability are given on the following pages ar ny facility: environment temperature (22 ± 3)° Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521/02522)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-18
All calibrations have been conduct Calibration Equipment used (M& Primary Standards Power meter NRP Power sensor NRP-Z91	TE critical for calibration)	Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-18 Apr-18
All calibrations have been conduc Calibration Equipment used (M& Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91	TE critical for calibration)	Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02522)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18
All calibrations have been conduc Calibration Equipment used (M& Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator	TE critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k)	Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02522)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 Apr-18
All calibrations have been conduc Calibration Equipment used (M& Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Power Sensor NRP-Z01/4	TE critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327	Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02522) 07-Apr-17 (No. 217-02528) 07-Apr-17 (No. 217-02529)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-18
All calibrations have been conduct Calibration Equipment used (M& Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Fype-N mismatch combination Reference Probe EX3DV4 DAE4	ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 5058 (20k) SN: 7349 SN: 601	Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02522) 04-Apr-17 (No. 217-02528) 07-Apr-17 (No. 217-02529) 31-May-17 (No. EX3-7349_May17) 28-Mar-17 (No. DAE4-601_Mar17)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 May-18 May-18 Mar-18
All calibrations have been conduct Calibration Equipment used (M& Calibration Equipment used (M& Primary Standards Power sensor NRP-Z91 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4	ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 7349 SN: 601	Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02522) 07-Apr-17 (No. 217-02522) 07-Apr-17 (No. 217-02529) 31-May-17 (No. EX3-7349_May17) 28-Mar-17 (No. DAE4-601_Mar17) Check Date (in house)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 May-18 May-18 Mar-18 Scheduled Check
All calibrations have been conduct Calibration Equipment used (M& Calibration Equipment used	ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 601 ID #	Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02522) 07-Apr-17 (No. 217-02522) 07-Apr-17 (No. 217-02529) 07-Apr-17 (No. 217-02529) 031-May-17 (No. EX3-7349_May17) 28-Mar-17 (No. DAE4-601_Mar17) Check Date (in house) 07-Oct-15 (in house check Oct-16)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 May-18 May-18 Mar-18 Scheduled Check In house check: Oct-18
All calibrations have been conduct Calibration Equipment used (M& Calibration Equipment used (M& Calibration Equipment used (M& Calibration Equipment used (M& Calibration Reference 20 dB Attenuator Fype-N mismatch combination Reference Probe EX3DV4 DAE4 Caecondary Standards Cower meter EPM-442A Cower sensor HP 8481A	ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 601 ID # SN: GB37480704 SN: US37292783	Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02522) 07-Apr-17 (No. 217-02522) 07-Apr-17 (No. 217-02529) 31-May-17 (No. DAE4-601_Mar17) 28-Mar-17 (No. DAE4-601_Mar17) Check Date (in house) 07-Oct-15 (in house check Oct-16) 07-Oct-15 (in house check Oct-16)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 May-18 May-18 Mar-18 Scheduled Check In house check: Oct-18 In house check: Oct-18
All calibrations have been conduct Calibration Equipment used (M& Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter EPM-442A Power sensor HP 8481A Power sensor HP 8481A Power sensor HP 8481A	ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 601 ID # SN: GB37480704 SN: U32247 SN: 7349 SN: 601 ID # SN: GB37480704 SN: WY41092317 SN: MY41092317	Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02522) 07-Apr-17 (No. 217-02522) 07-Apr-17 (No. 217-02528) 07-Apr-17 (No. 217-02529) 31-May-17 (No. EX3-7349_May17) 28-Mar-17 (No. DAE4-601_Mar17) Check Date (in house) 07-Oct-15 (in house check Oct-16) 07-Oct-15 (in house check Oct-16) 07-Oct-15 (in house check Oct-16)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 May-18 Mar-18 Scheduled Check In house check: Oct-18 In house check: Oct-18 In house check: Oct-18
All calibrations have been conduct Calibration Equipment used (M& Calibration Equipment (M& Calibration Equipment (M& Calibration Equipment (M& Calibration Equipment (M& Calibration Equipment (Calibration Equipment) Cover sensor NRP-291 Calibration Equipment (Calibration Equipment) Calibration Equipment (Calibration Equipment) Cover sensor HP 8481A Cover sensor HP 8481A Calibration Equipment (Calibration Equipment) Calibration Equipment (Calibrat	ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 7349 SN: 601 ID # SN: 0132442 SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 7349 SN: 601 ID # SN: GB37480704 SN: US37292783 SN: 10972 SN: 10972	Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02522) 07-Apr-17 (No. 217-02522) 07-Apr-17 (No. 217-02529) 31-May-17 (No. 217-02529) 31-May-17 (No. 217-02529) 31-May-17 (No. DAE4-601_Mar17) 28-Mar-17 (No. DAE4-601_Mar17) Check Date (in house) 07-Oct-15 (in house check Oct-16) 07-Oct-15 (in house check Oct-16) 07-Oct-15 (in house check Oct-16) 15-Jun-15 (in house check Oct-16) 16 (in house check Oct-16)	Are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 May-18 May-18 May-18 Scheduled Check In house check: Oct-18 In house
All calibrations have been conduc Calibration Equipment used (M& Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter EPM-442A Power sensor HP 8481A Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer HP 8753E	ID # ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 7349 SN: 601 ID # SN: 0137292783 SN: US37292783 SN: 10972 SN: US37390585	Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02522) 07-Apr-17 (No. 217-02528) 07-Apr-17 (No. 217-02529) 31-May-17 (No. 217-02529) 31-May-17 (No. 217-02529) 31-May-17 (No. EX3-7349_May17) 28-Mar-17 (No. DAE4-601_Mar17) Check Date (in house) 07-Oct-15 (in house check Oct-16) 07-Oct-15 (in house check Oct-16) 07-Oct-15 (in house check Oct-16) 15-Jun-15 (in house check Oct-16) 18-Oct-01 (in house check Oct-16)	Ad are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 May-18 May-18 Mar-18 Scheduled Check In house check: Oct-18 In house check: Oct-18 In house check: Oct-18 In house check: Oct-18 In house check: Oct-17
All calibrations have been conduc Calibration Equipment used (M& Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Fype-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power sensor HP 8481A Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer HP 8753E	ID # ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 5058 (20k) SN: 601 ID # SN: 601 ID # SN: 0537292783 SN: US37292783 SN: 100972 SN: US37390585 Name	Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02522) 07-Apr-17 (No. 217-02528) 07-Apr-17 (No. 217-02529) 31-May-17 (No. 217-02529) 31-May-17 (No. 217-02529) 31-May-17 (No. EX3-7349_May17) 28-Mar-17 (No. DAE4-601_Mar17) Check Date (in house) 07-Oct-15 (in house check Oct-16) 15-Jun-15 (in house check Oct-16) 18-Oct-01 (in house check Oct-16) Function	Are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 May-18 May-18 May-18 Scheduled Check In house check: Oct-18 In house check: Oct-18 In house check: Oct-18 In house check: Oct-18 In house check: Oct-17 Signature
All calibrations have been conduct Calibration Equipment used (M& Calibration Equipment used (M& Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter EPM-442A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer HP 8753E Calibrated by:	ID # ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 7349 SN: 601 ID # SN: 03245 SN: 5058 (20k) SN: 7349 SN: 601 ID # SN: 0B37480704 SN: US37292783 SN: 100972 SN: US37390585 Name Michael Weber	Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02522) 04-Apr-17 (No. 217-02522) 04-Apr-17 (No. 217-02522) 07-Apr-17 (No. 217-02529) 31-May-17 (No. 217-02529) 31-May-17 (No. 217-02529) 31-May-17 (No. EX3-7349_May17) 28-Mar-17 (No. DAE4-601_Mar17) Check Date (in house) 07-Oct-15 (in house check Oct-16) 07-Oct-15 (in house check Oct-16) 07-Oct-15 (in house check Oct-16) 15-Jun-15 (in house check Oct-16) 18-Oct-01 (in house check Oct-16) Function Laboratory Technician	Are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 May-18 May-18 Mar-18 Scheduled Check In house check: Oct-18 In house check: Oct-18 In house check: Oct-18 In house check: Oct-18 In house check: Oct-17 Signature
All calibrations have been conduct Calibration Equipment used (M& Calibration Equipment used (M& Primary Standards Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power sensor HP 8481A Power sensor HP 8481A Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer HP 8753E Calibrated by:	ID # SN: 104778 SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 5058 (20k) SN: 601 ID # SN: 601 ID # SN: GB37480704 SN: US37292783 SN: 10972 SN: 1037390585 Name Michael Weber Katja Pokovic	Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02522) 07-Apr-17 (No. 217-02528) 07-Apr-17 (No. 217-02529) 31-May-17 (No. 217-02529) 31-May-17 (No. EX3-7349_May17) 28-Mar-17 (No. DAE4-601_Mar17) Check Date (in house) 07-Oct-15 (in house check Oct-16) 07-Oct-15 (in house check Oct-16) 15-Jun-15 (in house check Oct-16) 18-Oct-01 (in house check Oct-16) Function Laboratory Technician	Ad are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 May-18 May-18 May-18 Mar-18 Scheduled Check In house check: Oct-18 In house check: Oct-17 Signature

Certificate No: D2450V2-853_Jul17

Page 1 of 8



No. I17Z61226-SEM02 Page 97 of 86

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst

C Service suisse d'étalonnage

Accreditation No.: SCS 0108

Servizio svizzero di taratura Suiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Glossary:

TSL	tissue simulating liquid
ConvF	sensitivity in TSL / NORM x,y,z
N/A	not applicable or not measured

Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- b) IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016
- c) IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- d) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

e) DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed
 point exactly below the center marking of the flat phantom section, with the arms oriented
 parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole
 positioned under the liquid filled phantom. The impedance stated is transformed from the
 measurement at the SMA connector to the feed point. The Return Loss ensures low
 reflected power. No uncertainty required.
- *Electrical Delay:* One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: D2450V2-853_Jul17

Page 2 of 8



Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V52.10.0
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	2450 MHz ± 1 MHz	

Head TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	39.2	1.80 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	37.8 ± 6 %	1.87 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	13.4 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	52.2 W/kg ± 17.0 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR averaged over 10 cm ³ (10 g) of Head TSL SAR measured	condition 250 mW input power	6.26 W/kg

Body TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	52.7	1.95 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	52.1 ± 6 %	2.04 mho/m ± 6 %
Body TSL temperature change during test	< 0.5 °C		

SAR result with Body TSL

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	250 mW input power	12.9 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	50.4 W/kg ± 17.0 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Body TSL	condition	
SAR averaged over 10 cm ³ (10 g) of Body TSL SAR measured	condition 250 mW input power	6.03 W/kg

Certificate No: D2450V2-853_Jul17



Appendix (Additional assessments outside the scope of SCS 0108)

Antenna Parameters with Head TSL

Impedance, transformed to feed point	52.0 Ω + 5.0 jΩ	
Return Loss	- 25.6 dB	

Antenna Parameters with Body TSL

Impedance, transformed to feed point	49.6 Ω + 6.3 jΩ		
Return Loss	- 24.0 dB		

General Antenna Parameters and Design

Electrical Delay (one direction)	1.161 ns

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG		
Manufactured on	November 10, 2009		

Certificate No: D2450V2-853_Jul17



DASY5 Validation Report for Head TSL

Date: 20.07.2017

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN: 853

Communication System: UID 0 - CW; Frequency: 2450 MHz Medium parameters used: f = 2450 MHz; σ = 1.87 S/m; ϵ_r = 37.8; ρ = 1000 kg/m³ Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 SN7349; ConvF(8.12, 8.12, 8.12); Calibrated: 31.05.2017;
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 28.03.2017
- Phantom: Flat Phantom 5.0 (front); Type: QD 000 P50 AA; Serial: 1001
- DASY52 52.10.0(1446); SEMCAD X 14.6.10(7417)

Dipole Calibration for Head Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 112.7 V/m; Power Drift = -0.00 dB Peak SAR (extrapolated) = 27.0 W/kg SAR(1 g) = 13.4 W/kg; SAR(10 g) = 6.26 W/kg Maximum value of SAR (measured) = 21.5 W/kg



0 dB = 21.5 W/kg = 13.32 dBW/kg

Certificate No: D2450V2-853_Jul17

Page 5 of 8



Impedance Measurement Plot for Head TSL



Certificate No: D2450V2-853_Jul17

Page 6 of 8



DASY5 Validation Report for Body TSL

Date: 21.07.2017

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN: 853

Communication System: UID 0 - CW; Frequency: 2450 MHz Medium parameters used: f = 2450 MHz; σ = 2.04 S/m; ϵ_r = 52.1; ρ = 1000 kg/m³ Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 SN7349; ConvF(8.1, 8.1, 8.1); Calibrated: 31.05.2017;
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 28.03.2017
- Phantom: Flat Phantom 5.0 (back); Type: QD 000 P50 AA; Serial: 1002
- DASY52 52.10.0(1446); SEMCAD X 14.6.10(7417)

Dipole Calibration for Body Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 104.1 V/m; Power Drift = -0.09 dB Peak SAR (extrapolated) = 25.5 W/kg SAR(1 g) = 12.9 W/kg; SAR(10 g) = 6.03 W/kg Maximum value of SAR (measured) = 20.0 W/kg



0 dB = 20.0 W/kg = 13.01 dBW/kg



Impedance Measurement Plot for Body TSL



Certificate No: D2450V2-853_Jul17

Page 8 of 8



ANNEX I SPOT CHECK

I.1 Conducted power of selected case

Table I.1-1: The conducted power results for GSM850/1900

0014	Conducted Power (dBm)					
	Channel 251(848.8MHz)	Channel 190(836.6MHz)	Channel 128(824.2MHz)			
	32.38	32.36	32.37			
Tune up	33	33	33			
0014	Conducted Power (dBm)					
	Channel 810(1909.8MHz)	Channel 661(1880MHz)	Channel 512(1850.2MHz)			
	29.43	29.38	29.29			
Tune up	30	30	30			

Table I.1-2: The conducted power results for GPRS

GSM 850	Measured Power (dBm)				
GPRS (GMSK)	251	190	128		
2 Txslots	29.94	29.92	29.84		
Tune up	30.5 30.5 30.5				
PCS1900	Measured Power (dBm)				
GPRS (GMSK)	810	661	512		
2 Txslots	26.83	26.80	26.75		
Tune up	27	27	27		

Table I.1-3: The conducted Power for WiFi

802.11b(dBm)							
Channel\data rate	1Mbps	2Mbps	5.5Mbps	11Mbps			
1(2412MHz)	15.26	/	/	/			
6(2437MHz)	15.18	/	/	/			
11(2462MHz)	15.09	/	/	/			

WiFi Tune Up: 15.5



I.2 Measurement results

Test Position	Phantom position L/R/F	Frequency Band	Channel Number	Frequency (MHz)	Test setup	EUT Measured Power (dBm)	Tune up (dBm)	Measured SAR 1g (W/kg)	Calculated SAR 1g (W/kg)	Measured SAR 10g (W/kg)	Calculated SAR 10g (W/kg)	Power Drift
Cheek	L	GSM850	251	848.8		32.38	33	0.649	0.75	0.486	0.56	-0.13
Body	F	GSM850	251	848.8	Rear GPRS 10mm	29.94	30.5	0.856	0.97	0.651	0.74	-0.07
Body	F	GSM850	190	836.6	Rear GPRS 10mm	29.92	30.5	0.634	0.72	0.484	0.55	-0.15
Body	F	GSM850	128	824.2	Rear GPRS 10mm	29.84	30.5	0.560	0.65	0.427	0.50	-0.09
Cheek	L	GSM1900	661	1880		29.38	30	0.345	0.40	0.219	0.25	0.05
Body	F	GSM1900	512	1850.2	Rear GPRS 10mm	26.75	27	0.703	0.74	0.415	0.44	-0.01
Cheek	L	WLAN	11	2462		15.09	16.5	0.211	0.29	0.098	0.14	-0.02
Body	F	WLAN	11	2462	Rear 10mm	15.09	16.5	0.076	0.11	0.039	0.05	0.06

Table I.2-1: SAR Values (WLAN - Head) – 802.11b (Scaled Reported SAR)

Ambient Temperature: 22.5 °C			Liquid Te	mperature: 23.3	°C		
Freque	ency	Side	Test	Actual duty	maximum	Reported SAR	Scaled reported SAR
MHz	Ch.		Position	factor	duty factor	(1g) (W/kg)	(1g) (W/kg)
2462	11	Left	Touch	99.52%	100%	0.29	0.29

Table I.2-2: SAR Values (WLAN - Body) – 802.11b (Scaled Reported SAR)

Ambient Temperature: 22.5 °C Liquid Temperature: 23.3°C						
Frequ	ency	Test	Actual duty	maximum duty	Reported SAR	Scaled reported SAR
MHz	Ch.	Position	factor	factor	(1g) (W/kg)	(1g) (W/kg)
2462	11	Rear	99.52%	100%	0.11	0.11





Picture I.1 The plot of duty factor for 802.11b

I.3 Reported SAR Comparison

Exposure Configuration	Technology Band	Reported SAR 1g (W/Kg): spot check	Reported SAR 1g (W/Kg): original
Head	GSM 850	0.75	0.68
(Separation Distance	PCS 1900	0.40	0.50
0mm)	WLAN 2.4 GHz	0.29	0.52
Hotspot	GSM 850	0.97	0.83
(Separation Distance	PCS 1900	0.74	0.70
10mm)	WLAN 2.4 GHz	0.11	0.24

Note: Spot Check results of GSM850/PCS1900 body, and GSM850 head, are larger than the original result. So they replace the original value and others are shared.



850 Left Cheek High

Date: 2017-8-4 Electronics: DAE4 Sn1331 Medium: Head 850 MHz Medium parameters used (interpolated): f = 848.8 MHz; $\sigma = 0.918$ mho/m; $\epsilon r = 40.21; \rho = 1000$ kg/m³ Ambient Temperature: 22.5°C Liquid Temperature: 22.4°C Communication System: GSM 850 Frequency: 848.8 MHz Duty Cycle: 1:8.3 Probe: EX3DV4 - SN3846 ConvF(9.33, 9.33, 9.33)

Area Scan (71x101x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm Maximum value of SAR (interpolated) = 0.707 W/kg

Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 7.786 V/m; Power Drift = -0.13 dB Peak SAR (extrapolated) = 0.820 W/kg SAR(1 g) = 0.649 W/kg; SAR(10 g) = 0.486 W/kg

Maximum value of SAR (measured) = 0.713 W/kg



Fig.I.1 850MHz



850 Body Rear High

Date: 2017-8-5 Electronics: DAE4 Sn1331 Medium: Body 850 MHz Medium parameters used (interpolated): f = 848.8 MHz; $\sigma = 0.987$ mho/m; $\epsilon r = 55.15$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.5°C Liquid Temperature: 22.4°C Communication System: GSM 850 GPRS Frequency: 848.8 MHz Duty Cycle: 1:4 Probe: EX3DV4 - SN3846 ConvF(9.52, 9.52, 9.52)

Area Scan (101x71x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm Maximum value of SAR (interpolated) = 0.943 W/kg

Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 28.83 V/m; Power Drift = -0.07 dB Peak SAR (extrapolated) = 1.09 W/kg SAR(1 g) = 0.856 W/kg; SAR(10 g) = 0.651 W/kg Maximum value of SAR (measured) = 0.894 W/kg



Fig.I.2 850 MHz



1900 Left Cheek Middle

Date: 2017-8-5 Electronics: DAE4 Sn1331 Medium: Head 1900 MHz Medium parameters use (interpolated): f = 1880 MHz; $\sigma = 1.371$ mho/m; $\epsilon r = 40.20$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.5°C Liquid Temperature: 22.4°C Communication System: GSM 1900MHz Frequency: 1880 MHz Duty Cycle: 1:8.3 Probe: EX3DV4 - SN3846 ConvF(7.89, 7.89, 7.89)

Area Scan (71x101x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm Maximum value of SAR (interpolated) = 0.398 W/kg

Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mmReference Value = 4.352 V/m; Power Drift = 0.05 dB Peak SAR (extrapolated) = 0.498 W/kg SAR(1 g) = 0.345 W/kg; SAR(10 g) = 0.219 W/kg

Maximum value of SAR (measured) = 0.397 W/kg



Fig.I.3 1900 MHz



1900 Body Rear Low

Date: 2017-8-5 Electronics: DAE4 Sn1331 Medium: Body 1900 MHz Medium parameters used (interpolated): f = 1850.2 MHz; $\sigma = 1.487$ mho/m; $\epsilon r = 54.76$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.5°C Liquid Temperature: 22.4°C Communication System: GSM 1900MHz GPRS Frequency: 1850.2 MHz Duty Cycle: 1:4 Probe: EX3DV4 - SN3846 ConvF(7.57, 7.57, 7.57)

Area Scan (111x71x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm Maximum value of SAR (interpolated) = 0.887 W/kg

Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 8.933 V/m; Power Drift = -0.01 dB Peak SAR (extrapolated) = 1.16 W/kg SAR(1 g) = 0.703 W/kg; SAR(10 g) = 0.415 W/kg Maximum value of SAR (measured) = 0.752 W/kg

W/kg 0.752 0.606 0.459 0.313 0.167 0.020

Fig.I.4 1900 MHz



Wifi 802.11n Left Cheek Channel 11

Date: 2017-8-7 Electronics: DAE4 Sn1331 Medium: Head 2450 MHz Medium parameters used (interpolated): f = 2462 MHz; $\sigma = 1.839$ S/m; $\epsilon_r = 39.45$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.5°C Liquid Temperature: 22.4°C Communication System: WLan 2450 Frequency: 2462 MHz Duty Cycle: 1:1 Probe: EX3DV4 – SN3846 ConvF(7.22, 7.22, 7.22)

Area Scan (81x131x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm Maximum value of SAR (interpolated) = 0.297 W/kg

Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mmReference Value = 7.161 V/m; Power Drift = -0.02 dB Peak SAR (extrapolated) = 0.456 W/kg SAR(1 g) = 0.211 W/kg; SAR(10 g) = 0.098 W/kg Maximum value of SAB (measured) = 0.270 W/kg

Maximum value of SAR (measured) = 0.270 W/kg



Fig.I.5 2450 MHz



Wifi 802.11b Body Rear Channel 11

Date: 2017-8-7 Electronics: DAE4 Sn1331 Medium: Body 2450 MHz Medium parameters used (interpolated): f = 2462 MHz; $\sigma = 1.986$ S/m; $\epsilon_r = 53.02$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.5°C Liquid Temperature: 22.4°C Communication System: WLan 2450 Frequency: 2462 MHz Duty Cycle: 1:1 Probe: EX3DV4 – SN3846 ConvF(7.31, 7.31, 7.31)

Area Scan (101x61x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm Maximum value of SAR (interpolated) = 0.179 W/kg

Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 3.689 V/m; Power Drift = 0.06 dB Peak SAR (extrapolated) = 0.461 W/kg SAR(1 g) = 0.076 W/kg; SAR(10 g) = 0.039 W/kg Maximum value of SAR (measured) = 0.173 W/kg



Fig.I.6 2450 MHz



ANNEX J Accreditation Certificate



END OF REPORT