



Date: 6/12/2020Electronics: DAE4 Sn777 Medium: Head 2600 MHz Medium parameters used: f = 2600 MHz;  $\sigma = 1.956$  mho/m;  $\epsilon_r = 39.01$ ;  $\rho = 1000$  kg/m<sup>3</sup> Ambient Temperature: 22.5°C Liquid Temperature: 22.3°C Communication System: CW Frequency: 2600 MHz Duty Cycle: 1:1 Probe: EX3DV4 – SN3617 ConvF(7.52,7.52,7.52)

System Validation /Area Scan (81x191x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm


Reference Value = 120.71 V/m; Power Drift = 0.02 **Fast SAR: SAR(1 g) = 13.77 W/kg; SAR(10 g) = 6.15 W/kg** Maximum value of SAR (interpolated) = 24.36 W/kg


System Validation /Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm


Reference Value =120.71 V/m; Power Drift = 0.02 dB Peak SAR (extrapolated) = 28.93 W/kg

SAR(1 g) = 13.91 W/kg; SAR(10 g) = 6.29 W/kg

Maximum value of SAR (measured) = 23.56 W/kg











Date: 6/13/2020Electronics: DAE4 Sn777 Medium: Head 2600 MHz Medium parameters used: f = 2600 MHz;  $\sigma = 1.96$  mho/m;  $\epsilon_r = 38.4$ ;  $\rho = 1000$  kg/m<sup>3</sup> Ambient Temperature: 22.5°C Liquid Temperature: 22.3°C Communication System: CW Frequency: 2600 MHz Duty Cycle: 1:1 Probe: EX3DV4 – SN3617 ConvF(7.52,7.52,7.52)

System Validation /Area Scan (81x191x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm

Reference Value = 118.41 V/m; Power Drift = 0.02 **Fast SAR: SAR(1 g) = 13.71 W/kg; SAR(10 g) = 6.23 W/kg** Maximum value of SAR (interpolated) = 24.49 W/kg


System Validation /Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm


Reference Value =118.41 V/m; Power Drift = 0.02 dB


Peak SAR (extrapolated) = 28.25 W/kg

SAR(1 g) = 13.67 W/kg; SAR(10 g) = 6.31 W/kg

Maximum value of SAR (measured) = 24.31 W/kg













Date: 6/14/2020Electronics: DAE4 Sn777 Medium: Head 5250 MHz Medium parameters used: f = 5250 MHz;  $\sigma = 4.729$  mho/m;  $\epsilon_r = 36.07$ ;  $\rho = 1000$  kg/m<sup>3</sup> Ambient Temperature: 22.5°C Liquid Temperature: 22.3°C Communication System: CW Frequency: 5250 MHz Duty Cycle: 1:1 Probe: EX3DV4 – SN3617 ConvF(5.39,539,5.39)

System Validation /Area Scan (81x191x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm

Maximum value of SAR (interpolated) = 17.8 W/kg

System Validation /Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value =75.93 V/m; Power Drift = 0.02 dB Peak SAR (extrapolated) = 27.55 W/kg SAR(1 g) = 19.98 W/kg; SAR(10 g) = 5.87 W/kg Maximum value of SAR (measured) = 18.01 W/kg










Date: 6/15/2020Electronics: DAE4 Sn777 Medium: Head 5600 MHz Medium parameters used: f = 5600 MHz;  $\sigma = 5.153$  mho/m;  $\epsilon_r = 35.75$ ;  $\rho = 1000$  kg/m<sup>3</sup> Ambient Temperature: 22.5°C Liquid Temperature: 22.3°C Communication System: CW Frequency: 5600 MHz Duty Cycle: 1:1 Probe: EX3DV4 – SN3617 ConvF(4.99,4.99,4.99)

System Validation /Area Scan (81x191x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm

Maximum value of SAR (interpolated) = 19.45 W/kg

System Validation /Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value =75.52 V/m; Power Drift = -0.1 dB Peak SAR (extrapolated) = 30.47 W/kg SAR(1 g) = 21.18 W/kg; SAR(10 g) = 6 W/kg Maximum value of SAR (measured) = 19.58 W/kg










Date: 6/16/2020Electronics: DAE4 Sn777 Medium: Head 5750 MHz Medium parameters used: f = 5750 MHz;  $\sigma = 5.201$  mho/m;  $\epsilon_r = 35.73$ ;  $\rho = 1000$  kg/m<sup>3</sup> Ambient Temperature: 22.5°C Liquid Temperature: 22.3°C Communication System: CW Frequency: 5750 MHz Duty Cycle: 1:1 Probe: EX3DV4 – SN3617 ConvF(5.10,5.10,5.10)

System Validation /Area Scan (81x191x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm

Maximum value of SAR (interpolated) = 19.28 W/kg

System Validation /Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value =72.33 V/m; Power Drift = -0.02 dB Peak SAR (extrapolated) = 31.37 W/kg SAR(1 g) = 19.9 W/kg; SAR(10 g) = 5.82 W/kg Maximum value of SAR (measured) = 18.56 W/kg







# **CAICT** No.I20Z60720-SEM01

The SAR system verification must be required that the area scan estimated 1-g SAR is within 3% of the zoom scan 1-g SAR.

Date	Band	Position	Area scan (1g)	Zoom scan (1g)	Drift (%)		
2020-6-5	750	Head	2.12	2.18	-2.75		
2020-6-6	835	Head	2.39	2.45	-2.45		
2020-6-7	835	Head	2.44	2.43	0.41		
2020-6-8	1750	Head	8.99	9.02	-0.33		
2020-6-9	1900	Head	9.76	9.73	0.31		
2020-6-10	1900	Head	9.75	9.79	-0.41		
2020-6-11	2450	Head	13.15	12.78	2.90		
2020-6-12	2600	Head	13.77	13.91	-1.01		
2020-6-13	2600	Head	13.71	13.67	0.29		

## Table B.1 Comparison between area scan and zoom scan for system verification






# ANNEX C SAR Measurement Setup

## C.1 Measurement Set-up

The Dasy4 or DASY5 system for performing compliance tests is illustrated above graphically. This system consists of the following items:



Picture C.1SAR Lab Test Measurement Set-up

- A standard high precision 6-axis robot (StäubliTX=RX family) with controller, teach pendant and software. An arm extension for accommodating the data acquisition electronics (DAE).
- An isotropic field probe optimized and calibrated for the targeted measurement.
- A data acquisition electronics (DAE) which performs the signal amplification, signal multiplexing, AD-conversion, offset measurements, mechanical surface detection, collision detection, etc. The unit is battery powered with standard or rechargeable batteries. The signal is optically transmitted to the EOC.
- The Electro-optical converter (EOC) performs the conversion from optical to electrical signals for the digital communication to the DAE. To use optical surface detection, a special version of the EOC is required. The EOC signal is transmitted to the measurement server.
- The function of the measurement server is to perform the time critical tasks such as signal filtering, control of the robot operation and fast movement interrupts.
- The Light Beam used is for probe alignment. This improves the (absolute) accuracy of the probe positioning.
- A computer running WinXP and the DASY4 or DASY5 software.
- Remote control and teach pendant as well as additional circuitry for robot safety such as
- warning lamps, etc.
- The phantom, the device holder and other accessories according to the targeted measurement.





# C.2 Dasy4 or DASY5 E-field Probe System

The SAR measurements were conducted with the dosimetric probe designed in the classical triangular configuration and optimized for dosimetric evaluation. The probe is constructed using the thick film technique; with printed resistive lines on ceramic substrates. The probe is equipped with an optical multifiber line ending at the front of the probe tip. It is connected to the EOC box on the robot arm and provides an automatic detection of the phantom surface. Half of the fibers are connected to a pulsed infrared transmitter, the other half to a synchronized receiver. As the probe approaches the surface, the reflection from the surface produces a coupling from the transmitting to the receiving fibers. This reflection increases first during the approach, reaches maximum and then decreases. If the probe is flatly touching the surface, the coupling is zero. The distance of the coupling maximum to the surface is independent of the surface reflectivity and largely independent of the surface to probe angle. The DASY4 or DASY5 software reads the reflection durning a software approach and looks for the maximum using 2<sup>nd</sup> ord curve fitting. The approach is stopped at reaching the maximum.

## **Probe Specifications:**

· · • • • • • • • • • • • • • • • • • •	
Model:	ES3DV3, EX3DV4
Frequency	10MHz — 6.0GHz(EX3DV4)
Range:	10MHz — 4GHz(ES3DV3)
Calibration:	In head and body simulating tissue at
	Frequencies from 835 up to 5800MHz
Linearity:	± 0.2 dB(30 MHz to 6 GHz) for EX3DV4
± 0.2 dB(30 MHz	to 4 GHz) for ES3DV3
DynamicRange:	10 mW/kg — 100W/kg
Probe Length:	330 mm
Probe Tip	
Length:	20 mm
<b>Body Diameter:</b>	12 mm
Tip Diameter:	2.5 mm (3.9 mm for ES3DV3)
Tip-Center:	1 mm (2.0mm for ES3DV3)
Application:SAF	R Dosimetry Testing
	Compliance tests of mobile phones
	Dosimetry in strong gradient fields

# /

Picture C.2Near-field Probe



## Picture C.3E-field Probe

## C.3 E-field Probe Calibration

Each E-Probe/Probe Amplifier combination has unique calibration parameters. A TEM cell calibration procedure is conducted to determine the proper amplifier settings to enter in the probe parameters. The amplifier settings are determined for a given frequency by subjecting the probe to a known E-field density (1 mW/cm<sup>2</sup>) using an RF Signal generator, TEM cell, and RF Power Meter.

The free space E-field from amplified probe outputs is determined in a test chamber. This calibration can be performed in a TEM cell if the frequency is below 1 GHz and inn a waveguide or



CAICT No.I20Z60720-SEM01

other methodologies above 1 GHz for free space. For the free space calibration, the probe is placed in the volumetric center of the cavity and at the proper orientation with the field. The probe is then rotated 360 degrees until the three channels show the maximum reading. The power density readings equates to 1 mW/cm<sup>2</sup>.

E-field temperature correlation calibration is performed in a flat phantom filled with the appropriate simulated brain tissue. The E-field in the medium correlates with the temperature rise in the dielectric medium. For temperature correlation calibration a RF transparent thermistor-based temperature probe is used in conjunction with the E-field probe.

$$SAR = C \frac{\Delta T}{\Delta t}$$

Where:

 $\Delta t$  = Exposure time (30 seconds), C = Heat capacity of tissue (brain or muscle),  $\Delta T$  = Temperature increase due to RF exposure.

$$SAR = \frac{\left|E\right|^2 \cdot \sigma}{\rho}$$

Where:

 $\sigma$  = Simulated tissue conductivity,

 $\rho$  = Tissue density (kg/m<sup>3</sup>).

# C.4 Other Test Equipment

# C.4.1 Data Acquisition Electronics(DAE)

The data acquisition electronics consist of a highly sensitive electrometer-grade preamplifier with auto-zeroing, a channel and gain-switching multiplexer, a fast 16 bit AD-converter and a command decoder with a control logic unit. Transmission to the measurement server is accomplished through an optical downlink for data and status information, as well as an optical uplink for commands and the clock.

The mechanical probe mounting device includes two different sensor systems for frontal and sideways probe contacts. They are used for mechanical surface detection and probe collision detection.

The input impedance of the DAE is 200 MOhm; the inputs are symmetrical and floating. Common mode rejection is above 80 dB.



PictureC.4: DAE





# C.4.2 Robot

The SPEAG DASY system uses the high precision robots (DASY4: RX90XL; DASY5: RX160L) type from Stäubli SA (France). For the 6-axis controller system, the robot controller version from Stäubli is used. The Stäubli robot series have many features that are important for our application:

- High precision (repeatability 0.02mm)
- High reliability (industrial design)
- > Low maintenance costs (virtually maintenance free due to direct drive gears; no belt drives)
- > Jerk-free straight movements (brushless synchron motors; no stepper motors)
- > Low ELF interference (motor control fields shielded via the closed metallic construction shields)



Picture C.5DASY 4

Picture C.6DASY 5

## C.4.3 Measurement Server

The Measurement server is based on a PC/104 CPU broad with CPU (dasy4: 166 MHz, Intel Pentium; DASY5: 400 MHz, Intel Celeron), chipdisk (DASY4: 32 MB; DASY5: 128MB), RAM (DASY4: 64 MB, DASY5: 128MB). The necessary circuits for communication with the DAE electronic box, as well as the 16 bit AD converter system for optical detection and digital I/O interface are contained on the DASY I/O broad, which is directly connected to the PC/104 bus of the CPU broad.

The measurement server performs all real-time data evaluation of field measurements and surface detection, controls robot movements and handles safety operation. The PC operating system cannot interfere with these time critical processes. All connections are supervised by a watchdog, and disconnection of any of the cables to the measurement server will automatically disarm the robot and disable all program-controlled robot movements. Furthermore, the measurement server is equipped with an expansion port which is reserved for future applications. Please note that this expansion port does not have a standardized pinout, and therefore only devices provided by SPEAG can be connected. Devices from any other supplier could seriously damage the measurement server.







Picture C.7 Server for DASY 4

Picture C.8 Server for DASY 5

# C.4.4 Device Holder for Phantom

The SAR in the phantom is approximately inversely proportional to the square of the distance between the source and the liquid surface. For a source at 5mm distance, a positioning uncertainty of  $\pm 0.5$ mm would produce a SAR uncertainty of  $\pm 20\%$ . Accurate device positioning is therefore crucial for accurate and repeatable measurements. The positions in which the devices must be measured are defined by the standards.


The DASY device holder is designed to cope with the different positions given in the standard. It has two scales for device rotation (with respect to the body axis) and device inclination (with respect to the line between the ear reference points). The rotation centers for both scales are the ear reference point (ERP). Thus the device needs no repositioning when changing the angles.

The DASY device holder is constructed of low-loss POM material having the following dielectric parameters: relative permittivity  $\ell = 3$  and loss tangent  $\delta = 0.02$ . The amount of dielectric material has been reduced in the closest vicinity of the device, since measurements have suggested that the influence of the clamp on the test results could thus be lowered.

<Laptop Extension Kit>

The extension is lightweight and made of POM, acrylic glass and foam. It fits easily on the upper part of the Mounting Device in place of the phone positioner. The extension is fully compatible with the Twin-SAM and ELI phantoms.





**Picture C.9-1: Device Holder** 

Picture C.9-2: Laptop Extension Kit

# C.4.5 Phantom

The SAM Twin Phantom V4.0 is constructed of a fiberglass shell integrated in a table. The shape of the shell is based on data from an anatomical study designed to

Represent the 90<sup>th</sup> percentile of the population. The phantom enables the dissymmetric evaluation





of SAR for both left and right handed handset usage, as well as body-worn usage using the flat phantom region. Reference markings on the Phantom allow the complete setup of all predefined phantom positions and measurement grids by manually teaching three points in the robot. The shell phantom has a 2mm shell thickness (except the ear region where shell thickness increases to 6 mm).

Shell Thickness:2±0. 2 mmFilling Volume:Approx. 25 litersDimensions:810 x 1000 x 500 mm (H x L x W)Available:Special



Picture C.10: SAM Twin Phantom







# ANNEX D Position of the wireless device in relation to the phantom

## **D.1 General considerations**

This standard specifies two handset test positions against the head phantom – the "cheek" position and the "tilt" position.






Picture D.2 Cheek position of the wireless device on the left side of SAM








Picture D.3 Tilt position of the wireless device on the left side of SAM

## D.2 Body-worn device

A typical example of a body-worn device is a mobile phone, wireless enabled PDA or other battery operated wireless device with the ability to transmit while mounted on a person's body using a carry accessory approved by the wireless device manufacturer.



Picture D.4Test positions for body-worn devices


# D.3 Desktop device

A typical example of a desktop device is a wireless enabled desktop computer placed on a table or desk when used.

The DUT shall be positioned at the distance and in the orientation to the phantom that corresponds to the intended use as specified by the manufacturer in the user instructions. For devices that employ an external antenna with variable positions, tests shall be performed for all antenna positions specified. Picture8.5 show positions for desktop device SAR tests. If the intended use is not specified, the device shall be tested directly against the flat phantom.







Picture D.5 Test positions for desktop devices



# **D.4 DUT Setup Photos**

Picture D.6

©Copyright. All rights reserved by CTTL.





# ANNEX E Equivalent Media Recipes

The liquid used for the frequency range of 800-3000 MHz consisted of water, sugar, salt, preventol, glycol monobutyl and Cellulose. The liquid has been previously proven to be suited for worst-case. The Table E.1 shows the detail solution. It's satisfying the latest tissue dielectric parameters requirements proposed by the IEEE 1528 and IEC 62209.

Frequency	835Head	835Body	1900	1900	2450	2450	5800	5800	
(MHz)	osoneau	03560Uy	Head	Body	Head	Body	Head	Body	
Ingredients (% by weight)									
Water	41.45	52.5	55.242	69.91	58.79	72.60	65.53	65.53	
Sugar	56.0	45.0	١	١	١	١	١	١	
Salt	1.45	1.4	0.306	0.13	0.06	0.18	١	١	
Preventol	0.1	0.1	١	١	١	١	١	١	
Cellulose	1.0	1.0	١	١	١	١	١	\	
Glycol	1	N	44.452	29.96	41.15	27.22	1	١	
Monobutyl	١	١	44.452	29.90	41.15	21.22	١	١	
Diethylenglycol	1	N	N	N	1	N	17.24	17.24	
monohexylether	١	١	١	١	١	١	17.24	17.24	
Triton X-100	١	١	١	١	١	١	17.24	17.24	
Dielectric	ε=41.5	ε=55.2	ε=40.0	ε=53.3	ε=39.2	ε=52.7	ε=35.3	ε=48.2	
Parameters									
Target Value	σ=0.90	σ=0.97	σ=1.40	σ=1.52	σ=1.80	σ=1.95	σ=5.27	σ=6.00	

## TableE.1: Composition of the Tissue Equivalent Matter

Note: There are a little adjustment respectively for 750, 1750, 2600, 5200, 5300 and 5600 based on the recipe of closest frequency in table E.1.





# ANNEX F System Validation

The SAR system must be validated against its performance specifications before it is deployed. When SAR probes, system components or software are changed, upgraded or recalibrated, these must be validated with the SAR system(s) that operates with such components.

	100	ie F.T. System vanu		
Probe SN.	Liquid name	Validation date	Frequency point	Status (OK or Not)
3617	Head 750MHz	Feb.17,2020	750 MHz	OK
3617	Head 850MHz	Feb.17,2020	835 MHz	OK
3617	Head 900MHz	Feb.17,2020	900 MHz	OK
3617	Head 1750MHz	Feb.17,2020	1750 MHz	OK
3617	Head 1810MHz	Feb.17,2020	1810 MHz	OK
3617	Head 1900MHz	Feb.18,2020	1900 MHz	OK
3617	Head 2000MHz	Feb.18,2020	2000 MHz	OK
3617	Head 2100MHz	Feb.18,2020	2100 MHz	OK
3617	Head 2300MHz	Feb.18,2020	2300 MHz	OK
3617	Head 2450MHz	Feb.18,2020	2450 MHz	OK
3617	Head 2600MHz	Feb.19,2020	2600 MHz	OK
3617	Head 3500MHz	Feb.19,2020	3500 MHz	OK
3617	Head 3700MHz	Feb.19,2020	3700 MHz	OK
3617	Head 5200MHz	Feb.19,2020	5250 MHz	OK
3617	Head 5500MHz	Feb.19,2020	5600 MHz	OK
3617	Head 5800MHz	Feb.19,2020	5800 MHz	OK
3617	Body 750MHz	Feb.19,2020	750 MHz	OK
3617	Body 850MHz	Feb.20,2020	835 MHz	OK
3617	Body 900MHz	Feb.20,2020	900 MHz	OK
3617	Body 1750MHz	Feb.20,2020	1750 MHz	OK
3617	Body 1810MHz	Feb.20,2020	1810 MHz	OK
3617	Body 1900MHz	Feb.20,2020	1900 MHz	OK
3617	Body 2000MHz	Feb.21,2020	2000 MHz	OK
3617	Body 2100MHz	Feb.21,2020	2100 MHz	OK
3617	Body 2300MHz	Feb.21,2020	2300 MHz	OK
3617	Body 2450MHz	Feb.21,2020	2450 MHz	OK
3617	Body 2600MHz	Feb.21,2020	2600 MHz	OK
3617	Body 3500MHz	Feb.22,2020	3500 MHz	OK
3617	Body 3700MHz	Feb.22,2020	3700 MHz	OK
3617	Body 5200MHz	Feb.22,2020	5250 MHz	OK
3617	Body 5500MHz	Feb.22,2020	5600 MHz	OK
3617	Body 5800MHz	Feb.22,2020	5800 MHz	OK

Tahla	E1. S.	istom V	alidation	for 3617
rapie	F. I: 3\	stem v	anuation	101 3017





# ANNEX G Probe Calibration Certificate

## **Probe 3617 Calibration Certificate**

Engineering AG Leughausstrasse 43, 8004 Zuri	ory of	S C S	Schweizerischer Kalibrierdiens Service suisse d'étalonnage Servizio svizzero di taratura Swiss Calibration Service
Accredited by the Swiss Accredit The Swiss Accreditation Servio Multilateral Agreement for the	ce is one of the signatories to	o the EA	reditation No.: SCS 0108
Client CTTL (Auden)	)	Certificate No:	EX3-3617_Jan20/2
CALIBRATION	CERTIFICATE	(Replacement of No: E)	(3-3617_Jan20)
Object	EX3DV4 - SN:3617	7	
Calibration procedure(s)	QA CAL-25.v7	CAL-12.v9, QA CAL-14.v5, QA ure for dosimetric E-field probes	CAL-23.v5,
This calibration certificate docur The measurements and the unc All calibrations have been cond	certainties with confidence prot	al standards, which realize the physical units ability are given on the following pages and acility: environment temperature $(22 \pm 3)^\circ C$	are part of the certificate.
The measurements and the uno	ments the traceability to nation certainties with confidence prot ucted in the closed laboratory f	bability are given on the following pages and	are part of the certificate.
This calibration certificate docur The measurements and the unc All calibrations have been cond	ments the traceability to nation certainties with confidence prot ucted in the closed laboratory f	bability are given on the following pages and	are part of the certificate.
This calibration certificate docur The measurements and the unc All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP	ments the traceability to nation certainties with confidence prot lucted in the closed laboratory f &TE critical for calibration)	Cal Date (Certificate No.) 03-Apr-19 (No. 217-02892/02893)	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-20
This calibration certificate docur The measurements and the unc All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91	ments the traceability to nation certainties with confidence prot ucted in the closed laboratory f &TE critical for calibration) ID SN: 104778 SN: 103244	Cal Date (Certificate No.) 03-Apr-19 (No. 217-02892)	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-20 Apr-20
This calibration certificate docur The measurements and the unc All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91	ID SN: 104778 SN: 103245	Cal Date (Certificate No.)       03-Apr-19 (No. 217-02892/02893)       03-Apr-19 (No. 217-02892)	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-20 Apr-20 Apr-20
This calibration certificate docur The measurements and the unc All calibrations have been cond Calibration Equipment used (M- Primary Standards Power sensor NRP- Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator	ID SN: 104778 SN: 103244 SN: SS277 (20x)	Cal Date (Certificate No.)       03-Apr-19 (No. 217-02892/02893)       03-Apr-19 (No. 217-02892)       03-Apr-19 (No. 217-02892)       03-Apr-19 (No. 217-02892)       03-Apr-19 (No. 217-02892)       03-Apr-19 (No. 217-02893)       04-Apr-19 (No. 217-02894)	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-20 Apr-20 Apr-20 Apr-20
This calibration certificate docur The measurements and the unc All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91	ID SN: 104778 SN: 103245	Cal Date (Certificate No.)       03-Apr-19 (No. 217-02892/02893)       03-Apr-19 (No. 217-02892)	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-20 Apr-20 Apr-20
This calibration certificate docur The measurements and the und All calibrations have been cond Calibration Equipment used (M- Primary Standards Power sensor NRP-Z91 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator DAE4	ments the traceability to nation certainties with confidence prot ucted in the closed laboratory f &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: S5277 (20x) SN: 660	Cal Date (Certificate No.)       03-Apr-19 (No. 217-02892/02893)       03-Apr-19 (No. 217-02892/02893)       03-Apr-19 (No. 217-02892)       03-Apr-19 (No. 217-02892)       03-Apr-19 (No. 217-02893)       04-Apr-19 (No. 217-02894)       27-Dec-19 (No. DAE4-660_Dec19)	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-20 Apr-20 Apr-20 Apr-20 Dec-20
This calibration certificate docur The measurements and the und All calibrations have been cond Calibration Equipment used (M- Primary Standards Power sensor NRP-Z91 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator DAE4	ments the traceability to nation certainties with confidence prot ucted in the closed laboratory f &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: S5277 (20x) SN: 660	Cal Date (Certificate No.)       03-Apr-19 (No. 217-02892/02893)       03-Apr-19 (No. 217-02892/02893)       03-Apr-19 (No. 217-02892)       03-Apr-19 (No. 217-02893)       04-Apr-19 (No. 217-02894)       27-Dec-19 (No. DAE4-660_Dec19)       31-Dec-19 (No. ES3-3013_Dec19)       Check Date (in house)	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-20 Apr-20 Apr-20 Apr-20 Dec-20
This calibration certificate docur The measurements and the uno All calibrations have been cond Calibration Equipment used (M- Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator DAE4 Reference Probe ES3DV2 Secondary Standards Power meter E4419B	ID SN: 104778 SN: 103244 SN: 103244 SN: 303245 SN: 55277 (20x) SN: 3013 ID SN: 3013 ID SN: GB41293874	Cal Date (Certificate No.)       03-Apr-19 (No. 217-02892/02893)       03-Apr-19 (No. 217-02892/02893)       03-Apr-19 (No. 217-02892)       03-Apr-19 (No. 217-02893)       04-Apr-19 (No. 217-02894)       27-Dec-19 (No. DAE4-660_Dec19)       31-Dec-19 (No. ES3-3013_Dec19)       06-Apr-16 (in house)       06-Apr-16 (in house check Jun-18)	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-20 Apr-20 Apr-20 Apr-20 Dec-20 Dec-20 Scheduled Check In house check: Jun-20
This calibration certificate docur The measurements and the unc All calibrations have been cond Calibration Equipment used (M- Primary Standards Power sensor NRP-Z91 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator DAE4 Reference Probe ES3DV2 Secondary Standards Power meter E4419B Power sensor E4412A	ments the traceability to nation certainties with confidence prot ucted in the closed laboratory f &TE critical for calibration) ID SN: 104778 SN: 104778 SN: 103244 SN: 103245 SN: 3013245 SN: 660 SN: 3013 ID SN: 6B41293874 SN: MY41498087	Cal Date (Certificate No.)       03-Apr-19 (No. 217-02892/02893)       03-Apr-19 (No. 217-02892/02893)       03-Apr-19 (No. 217-02892)       03-Apr-19 (No. 217-02892)       03-Apr-19 (No. 217-02893)       04-Apr-19 (No. 217-02894)       27-Dec-19 (No. DAE4-660_Dec19)       31-Dec-19 (No. ES3-3013_Dec19)       Check Date (in house)       06-Apr-16 (in house check Jun-18)       06-Apr-16 (in house check Jun-18)	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-20 Apr-20 Apr-20 Apr-20 Dec-20 Dec-20 Dec-20 Scheduled Check In house check: Jun-20 In house check: Jun-20
This calibration certificate docur The measurements and the und All calibrations have been cond Calibration Equipment used (M- Primary Standards Power sensor NRP-Z91 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator DAE4 Reference Probe ES3DV2 Secondary Standards Power meter E4419B Power sensor E4412A	ments the traceability to nation certainties with confidence prot ucted in the closed laboratory f &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103244 SN: 30245 SN: 35277 (20x) SN: 660 SN: 3013 ID SN: GB41293874 SN: MY41498087 SN: 000110210	Cal Date (Certificate No.)       03-Apr-19 (No. 217-02892/02893)       03-Apr-19 (No. 217-02892/02893)       03-Apr-19 (No. 217-02892)       03-Apr-19 (No. 217-02893)       04-Apr-19 (No. 217-02894)       27-Dec-19 (No. DAE4-660_Dec19)       31-Dec-19 (No. ES3-3013_Dec19)       06-Apr-16 (in house)       06-Apr-16 (in house check Jun-18)       06-Apr-16 (in house check Jun-18)	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-20 Apr-20 Apr-20 Dec-20 Dec-20 Dec-20 Scheduled Check In house check: Jun-20 In house check: Jun-20 In house check: Jun-20
This calibration certificate docur The measurements and the und All calibrations have been cond Calibration Equipment used (M- Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator DAE4 Reference Probe ES3DV2 Secondary Standards Power meter E4419B Power sensor E4412A Power sensor E4412A R generator HP 8648C	ments the traceability to nation certainties with confidence prot ucted in the closed laboratory f &TE critical for calibration)	Cal Date (Certificate No.)       03-Apr-19 (No. 217-02892/02893)       03-Apr-19 (No. 217-02892/02893)       03-Apr-19 (No. 217-02892)       03-Apr-19 (No. 217-02893)       04-Apr-19 (No. 217-02894)       27-Dec-19 (No. DAE4-660_Dec19)       31-Dec-19 (No. ES3-3013_Dec19)       Check Date (in house)       06-Apr-16 (in house check Jun-18)       06-Apr-16 (in house check Jun-18)       06-Apr-16 (in house check Jun-18)       04-Aug-99 (in house check Jun-18)	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-20 Apr-20 Apr-20 Dec-20 Dec-20 Dec-20 Scheduled Check In house check: Jun-20 In house check: Jun-20 In house check: Jun-20 In house check: Jun-20
This calibration certificate docur The measurements and the und All calibrations have been cond Calibration Equipment used (M- Primary Standards Power sensor NRP-Z91 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator DAE4 Reference Probe ES3DV2 Secondary Standards Power meter E4419B Power sensor E4412A	ments the traceability to nation certainties with confidence prot ucted in the closed laboratory f &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103244 SN: 30245 SN: 35277 (20x) SN: 660 SN: 3013 ID SN: GB41293874 SN: MY41498087 SN: 000110210	Cal Date (Certificate No.)       03-Apr-19 (No. 217-02892/02893)       03-Apr-19 (No. 217-02892/02893)       03-Apr-19 (No. 217-02892)       03-Apr-19 (No. 217-02893)       04-Apr-19 (No. 217-02894)       27-Dec-19 (No. DAE4-660_Dec19)       31-Dec-19 (No. ES3-3013_Dec19)       06-Apr-16 (in house)       06-Apr-16 (in house check Jun-18)       06-Apr-16 (in house check Jun-18)	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-20 Apr-20 Apr-20 Apr-20 Dec-20 Dec-20 Dec-20 Scheduled Check In house check: Jun-20 In house check: Jun-20
This calibration certificate docur The measurements and the unc All calibrations have been cond Calibration Equipment used (M- Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator DAE4 Reference Probe ES3DV2 Secondary Standards Power meter E4419B Power sensor E4412A Power sensor E4412A RF generator HP 8648C	ments the traceability to nation certainties with confidence prot ucted in the closed laboratory f &TE critical for calibration)	Cal Date (Certificate No.)       03-Apr-19 (No. 217-02892/02893)       03-Apr-19 (No. 217-02892/02893)       03-Apr-19 (No. 217-02892)       03-Apr-19 (No. 217-02893)       04-Apr-19 (No. 217-02894)       27-Dec-19 (No. DAE4-660_Dec19)       31-Dec-19 (No. ES3-3013_Dec19)       Check Date (in house)       06-Apr-16 (in house check Jun-18)       06-Apr-16 (in house check Jun-18)       06-Apr-16 (in house check Jun-18)       04-Aug-99 (in house check Jun-18)	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-20 Apr-20 Apr-20 Dec-20 Dec-20 Dec-20 Scheduled Check In house check: Jun-20 In house check: Jun-20 In house check: Jun-20 In house check: Jun-20
This calibration certificate docur The measurements and the und All calibrations have been cond Calibration Equipment used (M- Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator DAE4 Reference Probe ES3DV2 Secondary Standards Power meter E4419B Power sensor E4412A Power sensor E4412A R generator HP 8648C	ID SN: 104778 SN: 104778 SN: 103244 SN: 103244 SN: 303245 SN: 55277 (20x) SN: 660 SN: 3013 ID SN: GB41293874 SN: MY41498087 SN: 000110210 SN: US3642U01700 SN: US3642U01700 SN: US41080477	Cal Date (Certificate No.)       03-Apr-19 (No. 217-02892/02893)       03-Apr-19 (No. 217-02892/02893)       03-Apr-19 (No. 217-02892)       03-Apr-19 (No. 217-02893)       04-Apr-19 (No. 217-02894)       27-Dec-19 (No. 217-02894)<	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-20 Apr-20 Apr-20 Apr-20 Dec-20 Dec-20 Dec-20 Scheduled Check In house check: Jun-20 In

Certificate No: EX3-3617\_Jan20/2

Page 1 of 23





#### **Calibration Laboratory of** Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Schweizerischer Kalibrierdienst Service suisse d'étalonnage

Servizio svizzero di taratura

S

C

S

Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Glossary:	
TSL	tissue simulating liquid
NORMx,y,z	sensitivity in free space
ConvF	sensitivity in TSL / NORMx,y,z
DCP	diode compression point
CF	crest factor (1/duty_cycle) of the RF signal
A, B, C, D	modulation dependent linearization parameters
Polarization φ	φ rotation around probe axis
Polarization 9	9 rotation around an axis that is in the plane normal to probe axis (at measurement center),
	i.e., $\vartheta = 0$ is normal to probe axis
Connector Angle	information used in DASY system to align probe sensor X to the robot coordinate system

information used in DASY system to align probe sensor X to the robot coordinate system

#### Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, ", "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from handb)
- held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016 IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices C)
- used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010 d) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

#### Methods Applied and Interpretation of Parameters:

- NORMx,y,z: Assessed for E-field polarization 9 = 0 (f ≤ 900 MHz in TEM-cell; f > 1800 MHz: R22 waveguide). NORMx,y,z are only intermediate values, i.e., the uncertainties of NORMx,y,z does not affect the E<sup>2</sup>-field uncertainty inside TSL (see below ConvF).
- NORM(f)x,y,z = NORMx,y,z \* frequency\_response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- DCPx,y,z: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- PAR: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- Ax,y,z; Bx,y,z; Cx,y,z; Dx,y,z; VRx,y,z: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters: Assessed in flat phantom using E-field (or Temperature Transfer Standard for f ≤ 800 MHz) and inside waveguide using analytical field distributions based on power measurements for f > 800 MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORMx,y,z \* ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from ± 50 MHz to ± 100 MHz.
- Spherical isotropy (3D deviation from isotropy): in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle: The angle is assessed using the information gained by determining the NORMx (no uncertainty required).

Certificate No: EX3-3617\_Jan20/2





January 30, 2020

# DASY/EASY - Parameters of Probe: EX3DV4 - SN:3617

#### **Basic Calibration Parameters**

	Sensor X	Sensor Y	Sensor Z	Unc (k=2)
Norm $(\mu V/(V/m)^2)^A$	0.35	0.21	0.32	± 10.1 %
DCP (mV) <sup>B</sup>	104.3	93.8	97.1	

#### **Calibration Results for Modulation Response**

UID	Communication System Name		A dB	B dBõV	С	D dB	VR mV	Max dev.	Max Unc <sup>E</sup> (k=2)
0	CW	X	0.00	0.00	1.00	0.00	130.5	± 3.5 %	± 4.7 %
		Y	0.00	0.00	1.00		137.4		
		Z	0.00	0.00	1.00		129.2		
10352-	Pulse Waveform (200Hz, 10%)	X	5.74	74.31	15.16	10.00	60.0	± 2.6 %	± 9.6 %
AAA		Y	20.00	84.63	18.23		60.0		208
		Z	20.00	90.64	20.98		60.0		
10353-	Pulse Waveform (200Hz, 20%)	X	11.18	82.57	16.62	6.99	80.0	± 1.6 %	± 9.6 %
AAA		Y	11.60	81.13	15.97		80.0	1	
		Z	20.00	91.54	20.06		80.0		
10354-	Pulse Waveform (200Hz, 40%)	X	20.00	88.75	16.93	3.98	95.0	± 1.0 %	± 9.6 %
AAA		Y	1.22	64.13	8.17		95.0	1	
		Z	20.00	94.77	20.04	1	95.0	1	
10355-	Pulse Waveform (200Hz, 60%)	X	20.00	90.94	16.71	2.22	120.0	± 1.3 %	± 9.6 %
AAA		Y	0.41	60.00	4.32		120.0	1	
		Z	20.00	99.77	20.92	1	120.0	1	
10387-	QPSK Waveform, 1 MHz	X	0.73	63.23	9.65	0.00	150.0	± 4.1 %	±9.6 %
AAA		Y	0.47	60.00	5.82		150.0		
		Z	0.73	63.00	9.63	1	150.0	1	
10388-	QPSK Waveform, 10 MHz	X	2.46	70.66	17.17	0.00	150.0	± 1.7 %	±9.6 %
AAA		Y	2.10	68.37	15.67	1	150.0	1	
		Z	2.45	70.34	17.05		150.0	]	
10396-	64-QAM Waveform, 100 kHz	X	3.34	72.82	19.20	3.01	150.0	± 1.6 %	± 9.6 %
AAA		Y	3.57	72.45	19.52		150.0	]	
		Z	3.45	73.00	19.94		150.0		
10399-	64-QAM Waveform, 40 MHz	X	3.61	68.21	16.41	0.00	150.0	± 3.8 %	± 9.6 %
AAA		Y	3.40	67.13	15.82		150.0		
		Z	3.62	68.06	16.39		150.0		
10414-	WLAN CCDF, 64-QAM, 40MHz	Х	4.88	66.26	15.89	0.00	150.0	± 6.6 %	± 9.6 %
AAA		Y	4.57	64.95	15.35		150.0		
		Z	4.92	66.18	15.92		150.0		

Note: For details on UID parameters see Appendix

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

 <sup>A</sup> The uncertainties of Norm X,Y,Z do not affect the E<sup>2</sup>-field uncertainty inside TSL (see Pages 5 and 6).
<sup>B</sup> Numerical linearization parameter: uncertainty not required.
<sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value. field value.

Certificate No: EX3-3617\_Jan20/2

Page 3 of 23





January 30, 2020

# DASY/EASY - Parameters of Probe: EX3DV4 - SN:3617

#### Sensor Model Parameters

	C1 fF	C2 fF	α V <sup>-1</sup>	T1 ms.V⁻²	T2 ms.V <sup>-1</sup>	T3 ms	T4 V <sup>-2</sup>	T5 V <sup>-1</sup>	Т6
Х	41.2	299.64	34.06	12.13	0.82	5.00	1.88	0.20	1.00
Y	42.0	334.64	39.96	9.91	1.46	5.06	0.00	0.82	1.01
Ζ	42.8	318.14	35.45	11.95	0.73	5.04	1.02	0.40	1.01

#### **Other Probe Parameters**

Sensor Arrangement	1	Triangular
Connector Angle (°)		13
Mechanical Surface Detection Mode		enabled
Optical Surface Detection Mode		disabled
Probe Overall Length		337 mm
Probe Body Diameter		10 mm
Tip Length		9 mm
Tip Diameter		2.5 mm
Probe Tip to Sensor X Calibration Point		1 mm
Probe Tip to Sensor Y Calibration Point		1 mm
Probe Tip to Sensor Z Calibration Point		1 mm
Recommended Measurement Distance from Surface		1.4 mm

Certificate No: EX3-3617\_Jan20/2

Page 4 of 23





January 30, 2020

# DASY/EASY - Parameters of Probe: EX3DV4 - SN:3617

Calibration Parameter Determined in Head Tissue Simulating Media

f (MHz) <sup>c</sup>	Relative Permittivity <sup>F</sup>	Conductivity (S/m) <sup>F</sup>	ConvF X	ConvF Y	ConvF Z	Alpha <sup>G</sup>	Depth <sup>G</sup> (mm)	Unc (k=2)
64	54.2	0.75	12.37	12.37	12.37	0.00	1.00	± 13.3 %
150	52.3	0.76	11.63	11.63	11.63	0.00	1.00	± 13.3 %
300	45.3	0.87	11.41	11.41	11.41	0.08	1.20	± 13.3 %
450	43.5	0.87	10.84	10.84	10.84	0.12	1.40	± 13.3 %
750	41.9	0.89	10.07	10.07	10.07	0.61	0.80	± 12.0 %
835	41.5	0.90	9.66	9.66	9.66	0.54	0.84	± 12.0 %
900	41.5	0.97	9.56	9.56	9.56	0.54	0.80	± 12.0 %
1450	40.5	1.20	8.72	8.72	8.72	0.45	0.80	± 12.0 %
1640	40.2	1.31	8.50	8.50	8.50	0.25	0.80	± 12.0 %
1750	40.1	1.37	8.41	8.41	8.41	0.30	0.80	± 12.0 %
1810	40.0	1.40	8.20	8.20	8.20	0.15	1.26	± 12.0 %
1900	40.0	1.40	8.14	8.14	8.14	0.31	0.80	± 12.0 %
2000	40.0	1.40	8.25	8.25	8.25	0.40	0.81	± 12.0 %
2100	39.8	1.49	8.16	8.16	8.16	0.28	0.80	± 12.0 %
2300	39.5	1.67	7.95	7.95	7.95	0.35	0.86	± 12.0 %
2450	39.2	1.80	7.65	7.65	7.65	0.33	0.90	± 12.0 %
2600	39.0	1.96	7.52	7.52	7.52	0.38	0.90	± 12.0 %
3300	38.2	2.71	7.07	7.07	7.07	0.30	1.20	± 13.1 %
3500	37.9	2.91	7.02	7.02	7.02	0.35	1.30	± 13.1 %
3700	37.7	3.12	6.77	6.77	6.77	0.35	1.30	± 13.1 9
3900	37.5	3.32	6.62	6.62	6.62	0.40	1.60	± 13.1 9
4100	37.2	3.53	6.60	6.60	6.60	0.40	1.60	± 13.1 %
4200	37.1	3.63	6.50	6.50	6.50	0.40	1.60	± 13.1 9
4400	36.9	3.84	6.35	6.35	6.35	0.40	1.60	± 13.1 %
4600	36.7	4.04	6.30	6.30	6.30	0.40	1.60	± 13.1 9
4800	36.4	4.25	6.25	6.25	6.25	0.40	1.80	± 13.1 9
4950	36.3	4.40	6.10	6.10	6.10	0.40	1.80	± 13.1 9
5200	36.0	4.66	5.49	5.49	5.49	0.40	1.80	± 13.1 9
5250	35.9	4.71	5.39	5.39	5.39	0.40	1.80	± 13.1 9
5300	35.9	4.76	5.29	5.29	5.29	0.40	1.80	± 13.1 9
5500	35.6	4.96	5.14	5.14	5.14	0.40	1.80	± 13.1 9
5600	35.5	5.07	4.99	4.99	4.99	0.40	1.80	± 13.1 9
5750	35.4	5.22	5.10	5.10	5.10	0.40	1.80	± 13.1 9
5800	35.3	5.27	5.00	5.00	5.00	0.40	1.80	± 13.1 °

<sup>C</sup> Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Validity of ConvF assessed at 6 MHz is +9 MHz, and ConvF assessed at 13 MHz is 9-19 MHz. Above 5 GHz frequency validity can be extended to ± 110 MHz. <sup>F</sup> At frequencies below 3 GHz, the validity of tissue parameters (ε and σ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters (ε and σ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF assessed to the convF uncertainty for indicated tracet tissue parameters.

The conversion of the conversion of the values of the val

Certificate No: EX3-3617\_Jan20/2

Page 5 of 23





January 30, 2020

# DASY/EASY - Parameters of Probe: EX3DV4 - SN:3617

Calibration Parameter Determined in Body Tissue Simulating Media

f (MHz) <sup>C</sup>	Relative Permittivity <sup>F</sup>	Conductivity (S/m) <sup>F</sup>	ConvF X	ConvF Y	ConvF Z	Alpha <sup>G</sup>	Depth <sup>G</sup> (mm)	Unc (k=2)
750	55.5	0.96	9.80	9.80	9.80	0.50	0.80	± 12.0 %
835	55.2	0.97	9.53	9.53	9.53	0.43	0.80	± 12.0 %
900	55.0	1.05	9.49	9.49	9.49	0.42	0.80	± 12.0 %
1450	54.0	1.30	8.56	8.56	8.56	0.25	0.80	± 12.0 %
1640	53.7	1.42	8.44	8.44	8.44	0.32	0.80	± 12.0 %
1750	53.4	1.49	8.09	8.09	8.09	0.48	0.80	± 12.0 %
1810	53.3	1.52	8.05	8.05	8.05	0.44	0.80	± 12.0 %
1900	53.3	1.52	7.94	7.94	7.94	0.39	0.80	± 12.0 %
2000	53.3	1.52	7.92	7.92	7.92	0.37	0.86	± 12.0 %
2100	53.2	1.62	7.89	7.89	7.89	0.35	0.89	± 12.0 %
2300	52.9	1.81	7.78	7.78	7.78	0.39	0.85	± 12.0 %
2450	52.7	1.95	7.76	7.76	7.76	0.41	0.80	± 12.0 %
2600	52.5	2.16	7.45	7.45	7.45	0.32	0.80	± 12.0 %
3300	51.6	3.08	6.44	6.44	6.44	0.40	1.70	± 13.1 %
3500	51.3	3.31	6.30	6.30	6.30	0.40	1.70	± 13.1 %
3700	51.0	3.55	6.27	6.27	6.27	0.40	1.70	± 13.1 %
3900	51.2	3.78	6.24	6.24	6.24	0.40	1.70	± 13.1 %
4100	50.5	4.01	6.21	6.21	6.21	0.40	1.70	± 13.1 %
4200	50.4	4.13	6.20	6.20	6.20	0.40	1.70	± 13.1 %
4400	50.1	4.37	5.97	5.97	5.97	0.40	1.70	± 13.1 %
4600	49.8	4.60	5.83	5.83	5.83	0.40	1.70	± 13.1 %
4800	49.6	4.83	5.72	5.72	5.72	0.50	1.80	± 13.1 %
4950	49.4	5.01	5.41	5.41	5.41	0.50	1.90	± 13.1 %
5200	49.0	5.30	4.80	4.80	4.80	0.50	1.90	± 13.1 %
5250	48.9	5.36	4.70	4.70	4.70	0.50	1.90	± 13.1 %
5300	48.9	5.42	4.61	4.61	4.61	0.50	1.90	± 13.1 %
5500	48.6	5.65	4.32	4.32	4.32	0.50	1.90	± 13.1 %
5600	48.5	5.77	4.23	4.23	4.23	0.50	1.90	± 13.1 %
5750	48.3	5.94	4.36	4.36	4.36	0.50	1.90	± 13.1 %
5800	48.2	6.00	4.22	4.22	4.22	0.50	1.90	± 13.1 %

<sup>C</sup> Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Validity of ConvF assessed at 6 MHz is ± 9 MHz, and ConvF assessed at 13 MHz is 9-19 MHz. Above 5 GHz frequency validity can be extended to ± 110 MHz. <sup>F</sup> At frequencies below 3 GHz, the validity of tissue parameters (ε and σ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters (ε and σ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters. <sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

Certificate No: EX3-3617\_Jan20/2


Page 6 of 23





January 30, 2020





Uncertainty of Frequency Response of E-field: ± 6.3% (k=2)


Certificate No: EX3-3617\_Jan20/2

Page 7 of 23






January 30, 2020



# Receiving Pattern ( $\phi$ ), $\vartheta = 0^{\circ}$



Uncertainty of Axial Isotropy Assessment: ± 0.5% (k=2)


Certificate No: EX3-3617\_Jan20/2

Page 8 of 23





January 30, 2020




Certificate No: EX3-3617\_Jan20/2

Page 9 of 23





January 30, 2020



# **Conversion Factor Assessment**





#### January 30, 2020

### Appendix: Modulation Calibration Parameters

UID	Rev	Communication System Name	Group	PAR (dB)	Unc <sup>E</sup> (k=2)
0		CW	CW	0.00	± 4.7 %
10010	CAA	SAR Validation (Square, 100ms, 10ms)	Test	10.00	± 9.6 %
10011	CAB	UMTS-FDD (WCDMA)	WCDMA	2.91	± 9.6 %
10012	CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps)	WLAN	1.87	± 9.6 %
10013	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 6 Mbps)	WLAN	9.46	± 9.6 %
10021	DAC	GSM-FDD (TDMA, GMSK)	GSM	9.39	± 9.6 %
10023	DAC	GPRS-FDD (TDMA, GMSK, TN 0)	GSM	9.57	± 9.6 %
10024	DAC	GPRS-FDD (TDMA, GMSK, TN 0-1)	GSM	6.56	± 9.6 %
10025	DAC	EDGE-FDD (TDMA, 8PSK, TN 0)	GSM	12.62	± 9.6 %
10026	DAC	EDGE-FDD (TDMA, 8PSK, TN 0-1)	GSM	9.55	± 9.6 %
10027	DAC	GPRS-FDD (TDMA, GMSK, TN 0-1-2)	GSM	4.80	± 9.6 %
10028	DAC	GPRS-FDD (TDMA, GMSK, TN 0-1-2-3)	GSM	3.55	± 9.6 %
10029	DAC	EDGE-FDD (TDMA, 8PSK, TN 0-1-2)	GSM	7.78	± 9.6 %
10030	CAA	IEEE 802.15.1 Bluetooth (GFSK, DH1)	Bluetooth	5.30	± 9.6 %
10031	CAA	IEEE 802.15.1 Bluetooth (GFSK, DH3)	Bluetooth	1.87	± 9.6 9
10032	CAA	IEEE 802.15.1 Bluetooth (GFSK, DH5)	Bluetooth	1.16	± 9.6 9
10033	CAA	IEEE 802.15.1 Bluetooth (PI/4-DQPSK, DH1)	Bluetooth	7.74	± 9.6 9
10034	CAA	IEEE 802.15.1 Bluetooth (PI/4-DQPSK, DH3)	Bluetooth	4.53	± 9.6 9
10034	CAA	IEEE 802.15.1 Bluetooth (PI/4-DQPSK, DH5)	Bluetooth	3.83	± 9.6 9
10035	CAA	IEEE 802.15.1 Bluetooth (8-DPSK, DH3)	Bluetooth	8.01	± 9.6 9
10030	CAA	IEEE 802.15.1 Bluetooth (8-DPSK, DH3)	Bluetooth	4.77	± 9.6 9
10037	CAA	IEEE 802.15.1 Bluetooth (8-DPSK, DH5)	Bluetooth	4.10	± 9.6 9
10038	CAA	CDMA2000 (1xRTT, RC1)	CDMA2000	4.10	± 9.6 9
			AMPS	7.78	± 9.6 °
10042 10044	CAB CAA	IS-54 / IS-136 FDD (TDMA/FDM, PI/4-DQPSK, Halfrate)	AMPS	0.00	± 9.6 °
		IS-91/EIA/TIA-553 FDD (FDMA, FM)	DECT		
10048	CAA	DECT (TDD, TDMA/FDM, GFSK, Full Slot, 24)		13.80	± 9.6 °
10049	CAA	DECT (TDD, TDMA/FDM, GFSK, Double Slot, 12)	DECT TD-SCDMA	10.79	± 9.6 °
10056	CAA	UMTS-TDD (TD-SCDMA, 1.28 Mcps)			
10058	DAC	EDGE-FDD (TDMA, 8PSK, TN 0-1-2-3)	GSM	6.52	± 9.6
10059	CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 2 Mbps)	WLAN WLAN	2.12	± 9.6 °
10060	CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 5.5 Mbps)			± 9.6
10061	CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 11 Mbps)	WLAN	3.60	
10062	CAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 6 Mbps)	WLAN	8.68	± 9.6 °
10063	CAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 9 Mbps)	WLAN	8.63	± 9.6
10064	CAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 12 Mbps)	WLAN	9.09	± 9.6
10065	CAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 18 Mbps)	WLAN	9.00	± 9.6
10066	CAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 24 Mbps)	WLAN	9.38	± 9.6
10067	CAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 36 Mbps)	WLAN	10.12	± 9.6
10068	CAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 48 Mbps)	WLAN	10.24	± 9.6
10069	CAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 54 Mbps)	WLAN	10.56	± 9.6
10071	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 9 Mbps)	WLAN	9.83	± 9.6
10072	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 12 Mbps)	WLAN	9.62	± 9.6
10073	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 18 Mbps)	WLAN	9.94	± 9.6
10074	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 24 Mbps)	WLAN	10.30	± 9.6
10075	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 36 Mbps)	WLAN	10.77	± 9.6
10076	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 48 Mbps)	WLAN	10.94	± 9.6
10077	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 54 Mbps)	WLAN	11.00	± 9.6
10081	CAB	CDMA2000 (1xRTT, RC3)	CDMA2000	3.97	± 9.6
10082	CAB	IS-54 / IS-136 FDD (TDMA/FDM, PI/4-DQPSK, Fullrate)	AMPS	4.77	± 9.6
10090	DAC	GPRS-FDD (TDMA, GMSK, TN 0-4)	GSM	6.56	± 9.6
10097	CAB	UMTS-FDD (HSDPA)	WCDMA	3.98	± 9.6
10098	CAB	UMTS-FDD (HSUPA, Subtest 2)	WCDMA	3.98	± 9.6
10099	DAC	EDGE-FDD (TDMA, 8PSK, TN 0-4)	GSM	9.55	± 9.6
10100	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	5.67	± 9.6
10101	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM)	LTE-FDD	6.42	± 9.6
10102	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM)	LTE-FDD	6.60	± 9.6
10103	CAG	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-TDD	9.29	± 9.6
10104	CAG	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM)	LTE-TDD	9.97	± 9.6
10105	CAG	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM)	LTE-TDD	10.01	± 9.6
10108	CAG	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, QPSK)	LTE-FDD	5.80	± 9.6

Certificate No: EX3-3617\_Jan20/2

Page 11 of 23