



# SAR TEST REPORT

Applicant	iRay Technology Co., Ltd.	
FCC ID	2ACHK-01070189	
Product	Wireless Digital Flat Panel Detector	
Model	Mars1013X	
Report No.	R2209A0844-S1	
Issue Date	May 26, 2023	

TA Technology (Shanghai) Co., Ltd. tested the above equipment in accordance with the requirements in **IEEE 1528-2013**, **ANSI C95.1**: **1992**, **IEEE C95.1**: **1991**. The test results show that the equipment tested is capable of demonstrating compliance with the requirements as documented in this report.

Wei Fangying

Fan Guangchang

Prepared by: Wei Fangying

Approved by: Fan Guangchang

# TA Technology (Shanghai) Co., Ltd. Building 3, No.145, Jintang Rd, Pudong Shanghai, P.R.China

Eullaing 3, No.145, Jintang Ra, Pudong Shanghai, P.R.Chi TEL: +86-021-50791141/2/3 FAX: +86-021-50791141/2/3-8000



# **Table of Contents**

1	Tes	st Laboratory	. 3
	1.1	Notes of the Test Report	. 3
	1.2	Test Facility	. 3
	1.3	Testing Location	. 3
	1.4	Laboratory Environment	. 3
2	Sta	atement of Compliance	. 4
3	De	scription of Equipment Under Test	. 5
4	Tes	st Specification, Methods and Procedures	. 7
5	Ор	erational Conditions during Test	. 8
	5.1	Test Positions	. 8
	5.2	Measurement Variability	10
	5.3	Test Configuration	11
		5.2.1 Wi-Fi Test Configuration	1
6	SA	R Measurements System Configuration	12
	6.1	SAR Measurement Set-up	12
	6.2	DASY5 E-field Probe System	13
	6.3	SAR Measurement Procedure	14
7	Ма	in Test Equipment	16
8	Tis	sue Dielectric Parameter Measurements & System Check	17
	8.1	Tissue Verification	17
	8.2	System Check	18
	8.3	SAR System Validation	20
9	No	rmal and Maximum Output Power	21
	9.1	WLAN Mode	21
1	) Me	easured and Reported (Scaled) SAR Results	25
	10.1	EUT Antenna Locations	25
	10.2	Measured SAR Results	26
	10.3	Simultaneous Transmission Analysis	27
1	l Me	asurement Uncertainty	28
A	NNEX	(A: Test Layout	29
A	NNEX	KB: System Check Results	31
A	NNEX	C: Highest Graph Results	34
A	NNEX	(D: Probe Calibration Certificate (SN: 3677)	36
		E: D2450V2 Dipole Calibration Certificate	
A	NNEX	(F: D5GHzV2 Dipole Calibration Certificate	66
A	NNEX	G: DAE4 Calibration Certificate (SN: 1692)	75
		(H: The EUT Appearance	
A	NNEX	(I: Test Setup Photos	79

# 1 Test Laboratory

## 1.1 Notes of the Test Report

This report shall not be reproduced in full or partial, without the written approval of TA Technology

(Shanghai) Co., Ltd. The results documented in this report apply only to the tested sample, under the conditions and modes of operation as described herein .Measurement Uncertainties were not taken into account and are published for informational purposes only. This report is written to support regulatory compliance of the applicable standards stated above.

# 1.2 Test Facility

#### FCC (Designation number: CN1179, Test Firm Registration Number: 446626)

TA Technology (Shanghai) Co., Ltd. has been listed on the US Federal Communications Commission list of test facilities recognized to perform measurements.

#### A2LA (Certificate Number: 3857.01)

TA Technology (Shanghai) Co., Ltd. has been listed by American Association for Laboratory Accreditation to perform measurement.

## 1.3 Testing Location

Company:	TA Technology (Shanghai) Co., Ltd.
Address:	Building 3, No.145, Jintang Rd, Pudong Shanghai, P.R.China
City:	Shanghai
Post code:	201201
Country:	P. R. China
Contact:	Fan Guangchang
Telephone:	+86-021-50791141/2/3
Fax:	+86-021-50791141/2/3-8000
Website:	http://www.ta-shanghai.com
E-mail:	fanguangchang@ta-shanghai.com

## 1.4 Laboratory Environment

Temperature	Min. = 18°C, Max. = 25 °C		
Relative humidity	Min. = 30%, Max. = 70%		
Ground system resistance	< 0.5 Ω		
Ambient noise is checked and found very low and in compliance with requirement of standards.			
Reflection of surrounding objects is minimized and in compliance with requirement of standards.			



# 2 Statement of Compliance

The maximum results of Specific Absorption Rate (SAR) found during testing for the EUT are as follows: Table 1: Highest Reported SAR

		Highest Reported SAR (W/kg)		
	Mode	1g Body SAR		
		(Separation 0mm)		
	Wi-Fi (2.4G)	0.12		
Wi-Fi (5G)		0.23		
Da	Date of Testing: April 26, 2023			
Da	Date of Sample Received: September 28, 2022			
No	Note:			
1.	1. The device is in compliance with SAR for Uncontrolled Environment /General Population			
	exposure limits (1.6 W/kg ) specified in ANSI C95.1: 1992/IEEE C95.1: 1991, and had been			
	tested in accordance with the measurement methods and procedures specified in IEEE			
	1528-2013.			
2.	2. All indications of Pass/Fail in this report are opinions expressed by TA Technology (Shanghai)			
	Co., Ltd. based on interpretations and/or observations of test results. Measurement			
	Uncertainties were not taken into account and are published for informational purposes only.			

## Table 2: Highest Simultaneous Transmission SAR

Exposure Configuration	1g Body SAR (Separation 0mm)		
Highest Simultaneous Transmission SAR (W/kg)	0.37		
Note: The detail for simultaneous transmission consideration is described in chapter 10.3.			

# **3** Description of Equipment Under Test

#### **Client Information**

Applicant	iRay Technology Co., Ltd.
Applicant address	RM 202, Building 7, No. 590, Ruiqing RD.,Pudong, Shanghai, China
Manufacturer	iRay Technology Co., Ltd.
Manufacturer address	RM 202, Building 7, No. 590, Ruiqing RD.,Pudong, Shanghai, China

## **General Technologies**

Application Purpose	Original Grant	
EUT Stage	Identical Prototype	
Model	Mars1013X	
Lab Internal SN	R2209A0844/S01	
Hardware Version	A06	
Software Version	ARM: Kernel: 1.0.39.0 Core: 2.3.1.75 FPGA: 2.10.1.16 MCU: 2.10.0.17	
Antenna Type	Internal Antenna	
Device Class	В	
Wi-Fi Hotspot Wi-Fi 2.4G Wi-Fi 5G U-NII-1&U-NII-3		
EUT Accessory		
Adapter	Manufacturer: Shenzhen Longxc Power Supply Co., LTD Model: LXCP61-024300	
Rechargeable Li-ion	Manufacturer: iRay Technology Taicang Ltd.	
Battery Pack	Model: BATTERY-KX	
DC cable Manufacturer: iRay Technology Co. Ltd. Model: /		
Charger	Manufacturer: iRay Technology Taicang Ltd. Model: CHARGER-COMBO	
Note: The EUT is sent fro applicant.	om the applicant to TA and the information of the EUT is declared by the	



## Wireless Technology and Frequency Range

Wireless Technology		Modulation	Operating mode	Tx (MHz)	
	2.4G	DSSS, OFDM	802.11b/g/n HT20	2412 ~ 2462	
		OFDM	802.11n HT40	2422 ~ 2452	
Wi-Fi	5G	OFDM	802.11a/n HT20/ HT40/	5150 ~ 5250	
		OFDIVI	ac VHT20/ VHT40	5725 ~ 5850	
	Does this device support MIMO $\boxtimes$ Yes(2TX, 2RX) $\Box$ No				



# 4 Test Specification, Methods and Procedures

The tests documented in this report were performed in accordance with FCC 47 CFR § 2.1093, IEEE 1528- 2013, ANSI C95.1: 1992, IEEE C95.1: 1991, the following FCC Published RF exposure KDB procedures:

## **Reference Standards**


KDB 248227 D01 802.11Wi-Fi SAR v02r02 KDB 447498 D01 General RF Exposure Guidance v06 KDB 648474 D04 Handset SAR v01r03 KDB 690783 D01 SAR Listings on Grants v01r03 KDB 865664 D01 SAR measurement 100 MHz to 6 GHz v01r04 KDB 865664 D02 RF Exposure Reporting v01r02 KDB 941225 D06 Hotspot Mode v02r01 KDB 616217 D04 SAR for laptop and tablets v01r02



# 5 Operational Conditions during Test

# 5.1 Test Positions

According to KDB 616217 D04, SAR evaluation is required for back surface and edges of the devices. The back surface and edges of the tablet are tested with the tablet touching the phantom. Exposures from antennas through the front surface of the display section of a tablet are generally limited to the user's hands. Exposures to hands for typical consumer transmitters used in tablets are not expected to exceed the extremity SAR limit; therefore, SAR evaluation for the front surface of tablet display screens are generally not necessary. When voice mode is supported on a tablet and it is limited to speaker mode or headset operations only, additional SAR testing for this type of voice use is not required.



#### Fig-4.1 Illustration for Tablet Setup

According to KDB 447498 D01, the SAR test exclusion condition is based on source-based time-averaged maximum conducted output power, adjusted for tune-up tolerance, and the minimum test separation distance required for the exposure conditions. The SAR exclusion threshold is determined by the following formula.

(1) The SAR exclusion threshold for distances  $\leq$  50mm is defined by the following equation:

# (max. power of channel, including tune-up tolerance, mW) (min. test separation distance, mm) \*√ Frequency (GHz) ≤3.0

(2) The SAR exclusion threshold for distances >50mm is defined by the following equation, as illustrated in KDB 447498 D01 Appendix B:

a) at 100 MHz to 1500 MHz

# [Power allowed at numeric Threshold at 50 mm in step 1) + (test separation distance - 50 mm) · (f <sub>(MHz)</sub>/150)] mW


b) at > 1500 MHz and  $\leq$  6 GHz

[Power allowed at numeric Threshold at 50 mm in step 1) + (test separation distance - 50 mm) ·10] mW


SAR Test Report

TA

eurofins



			Front Side		
Band	Frequency (MHz)	Max. Tune-up Power (dBm)	Ant. To Surgace (mm)	Evaluation	Conclusion
Wi-Fi 2.4G Antenna 1	2462	13.00	5	6.26	Yes
Wi-Fi 2.4G Antenna 2	2462	15.00	5	9.92	Yes
	5240	14.00	5	11.50	Yes
Wi-Fi 5G Antenna 1	5825	14.50	5	13.60	Yes
Wi-Fi 5G Antenna 2	5240	13.50	5	10.25	Yes
wi-Fi 56 Antenna Z	5825	14.00	5	12.12	Yes





## 5.2 Measurement Variability

Per FCC KDB Publication 865664 D01, SAR measurement variability was assessed for each frequency band, which was determined by the SAR probe calibration point and tissue-equivalent medium used for the device measurements. When both head and body tissue-equivalent media were required for SAR measurements in a frequency band, the variability measurement procedures were applied to the tissue medium with the highest measured SAR, using the highest measured SAR configuration for that tissue-equivalent medium. These additional measurements were repeated after the completion of all measurements requiring the same head or body tissue-equivalent medium in a frequency band. The test device was returned to ambient conditions (normal room temperature) with the battery fully charged before it was re-mounted on the device holder for the repeated measurement(s) to minimize any unexpected variations in the repeated results.

SAR Measurement Variability was assessed using the following procedures for each frequency band: 1) When the original highest measured SAR is  $\geq$  0.80 W/kg, the measurement was repeated once.

2) A second repeated measurement was preformed only if the ratio of largest to smallest SAR for the original and first repeated measurements was > 1.20 or when the original or repeated measurement was  $\geq$  1.45 W/kg (~ 10% from the 1-g SAR limit).

3) A third repeated measurement was performed only if the original, first or second repeated measurement was  $\geq$  1.5 W/kg and the ratio of largest to smallest SAR for the original, first and second repeated measurements is > 1.20.

4) Repeated measurements are not required when the original highest measured SAR is < 0.80 W/kg

The same procedures should be adapted for measurements according to extremity and occupational exposure limits by applying a factor of 2.5 for extremity exposure and a factor of 5 for occupational exposure to the corresponding SAR thresholds.



SAR Test Report

TΔ

#### 5.3 Test Configuration

#### 5.2.1 Wi-Fi Test Configuration

SAR test reduction for 802.11 Wi-Fi transmission mode configurations are considered separately for DSSS and OFDM. An initial test position is determined to reduce the number of tests required for certain exposure configurations with multiple test positions. An initial test configuration is determined for each frequency band and aggregated band according to maximum output power, channel bandwidth, wireless mode configurations and other operating parameters to streamline the measurement requirements. For 2.4 GHz DSSS, either the initial test position or DSSS procedure is applied to reduce the number of SAR tests; These are mutually exclusive. For OFDM, an initial test position is only applicable to next to the ear, UMPC mini-tablet and hotspot mode configurations, which is tested using the initial test configuration to facilitate test reduction. For other exposure conditions with a fixed test position, SAR test reduction is determined using only the initial test configuration.

The multiple test positions require SAR measurements in head, hotspot mode or UMPC mini-tablet configurations may be reduced according to the highest reported SAR determined using the *initial test position(s)* by applying the DSSS or OFDM SAR measurement procedures in the required wireless mode test configuration(s). The *initial test position(s)* is measured using the highest measured maximum output power channel in the required wireless mode test configuration(s). When the *reported* SAR for the *initial test position* is:

- ≤ 0.4 W/kg, further SAR measurement is not required for the other test positions in that exposure configuration and wireless mode combination within the frequency band or aggregated band. DSSS and OFDM configurations are considered separately according to the required SAR procedures.
- 0.4 W/kg, SAR is repeated using the same wireless mode test configuration tested in the *initial test position* to measure the subsequent next closet/smallest test separation distance and maximum coupling test position, on the highest maximum output power channel, until the *reported* SAR is ≤ 0.8 W/kg or all required test positions are tested.
  - ✤ For subsequent test positions with equivalent test separation distance or when exposure is dominated by coupling conditions, the position for maximum coupling condition should be tested.
  - ♦ When it is unclear, all equivalent conditions must be tested.
- For all positions/configurations tested using the *initial test position* and subsequent test positions, when the *reported* SAR is > 0.8 W/kg, measure the SAR for these positions/configurations on the subsequent next highest measured output power channel(s) until the *reported* SAR is ≤ 1.2 W/kg or all required test channels are considered.
  - ♦ The additional power measurements required for this step should be limited to those necessary for identifying subsequent highest output power channels to apply the test reduction.

To determine the initial test position, Area Scans were performed to determine the position with the Maximum Value of SAR (measured). The position that produced the highest Maximum Value of SAR is considered the worst case position; thus used as the initial test position.


A Wi-Fi device must be configured to transmit continuously at the required data rate, channel bandwidth and signal modulation, using the highest transmission duty factor supported by the test mode tools for SAR measurement.



# 6 SAR Measurements System Configuration

# 6.1 SAR Measurement Set-up

The DASY system for performing compliance tests consists of the following items:



- A standard high precision 6-axis robot with controller, teach pendant and software. An arm extension for accommodating the data acquisition electronics (DAE).
- > An isotropic Field probe optimized and calibrated for the targeted measurement.
- A data acquisition electronics (DAE) which performs the signal amplification, signal multiplexing, AD-conversion, offset measurements, mechanical surface detection, collision detection, etc. The unit is battery powered with standard or rechargeable batteries. The signal is optically transmitted to the EOC.
- The Electro-optical converter (EOC) performs the conversion from optical to electrical signals for the digital communication to the DAE. To use optical surface detection, a special version of the EOC is required. The EOC signal is transmitted to the measurement server.
- > The function of the measurement server is to perform the time critical tasks such as signal filtering, control of the robot operation and fast movement interrupts.
- The Light Beam used is for probe alignment. This improves the (absolute) accuracy of the probe positioning.
- A computer running WinXP or Win7 and the DASY software.
- Remote control and teach pendant as well as additional circuitry for robot safety such as warning lamps, etc.
- > The phantom, the device holder and other accessories according to the targeted measurement.

## 6.2 DASY5 E-field Probe System

TΔ

The SAR measurements were conducted with the dosimetric probe EX3DV4 (manufactured by SPEAG), designed in the classical triangular configuration and optimized for dosimetric evaluation.

#### **EX3DV4 Probe Specification**

SAR Test Report

🔅 eurofins

Construction	Symmetrical design with triangular core Built-in shielding against static charges PEEK enclosure material (resistant to organic solvents, e.g., DGBE)	
Calibration	ISO/IEC 17025 calibration	_
	service available	111
Frequency	10 MHz to > 6 GHz	1111
	Linearity: ± 0.2 dB	
	(30 MHz to 6 GHz)	
Directivity	± 0.3 dB in HSL (rotation around probe	
	axis) $\pm$ 0.5 dB in tissue material (rotation	
	normal to probe axis)	
Dynamic	10 μW/g to > 100 mW/g Linearity:	100 1015
Range	± 0.2dB (noise: typically < 1 μW/g)	(P.2)
Dimensions	Overall length: 330 mm (Tip: 20 mm)	The second se
	Tip diameter: 2.5 mm (Body: 12 mm)	
	Typical distance from probe tip to dipole	a 2 8
	centers: 1 mm	
Application	High precision dosimetric	
	measurements in any exposure	
	Scenario (e.g., very strong gradient	
	fields). Only probe which enables	
	compliance testing for frequencies up to	
	6 GHz with precision of better 30%.	

#### **E-field Probe Calibration**

Each probe is calibrated according to a dosimetric assessment procedure with accuracy better than  $\pm$  10%. The spherical isotropy was evaluated and found to be better than  $\pm$  0.25dB. The sensitivity parameters (NormX, NormY, NormZ), the diode compression parameter (DCP) and the conversion factor (ConvF) of the probe are tested.

The free space E-field from amplified probe outputs is determined in a test chamber. This is performed in a TEM cell for frequencies bellow 1 GHz, and in a wave guide above 1 GHz for free space. For the free space calibration, the probe is placed in the volumetric center of the cavity and at the proper orientation with the field. The probe is then rotated 360 degrees.

E-field temperature correlation calibration is performed in a flat phantom filled with the appropriate simulated brain tissue. The measured free space E-field in the medium correlates to temperature rise in a dielectric medium. For temperature correlation calibration a RF transparent thermistor-based temperature probe is used in conjunction with the E-field probe.



SAR Test Report

**SAR=C** $\Delta$ **T**/ $\Delta$ **t** Where:  $\Delta$ t = Exposure time (30 seconds),

C = Heat capacity of tissue (brain or muscle),

 $\Delta T$  = Temperature increase due to RF exposure.

Or

## SAR=IEl<sup>2</sup>σ/ρ

Where:  $\sigma$  = Simulated tissue conductivity,

 $\rho$  = Tissue density (kg/m<sup>3</sup>).

## 6.3 SAR Measurement Procedure

#### **Power Reference Measurement**

The Power Reference Measurement and Power Drift Measurements are for monitoring the power drift of the device under test in the batch process. The minimum distance of probe sensors to surface determines the closest measurement point to phantom surface. This distance cannot be smaller than the distance of sensor calibration points to probe tip as defined in the probe properties.

### Area Scan

The area scan is used as a fast scan in two dimensions to find the area of high field values, before doing a fine measurement around the hot spot. The sophisticated interpolation routines implemented in DASY software can find the maximum found in the scanned area, within a range of the global maximum. The range (in dB) is specified in the standards for compliance testing. For example, a 2 dB range is required in IEEE standard 1528 and IEC 62209 standards, whereby 3 dB is a requirement when compliance is assessed in accordance with the ARIB standard (Japan), if only one zoom scan follows the area scan, then only the absolute maximum will be taken as reference. For cases where multiple maximums are detected, the number of zoom scans has to be increased accordingly. Area scan parameters extracted from FCC KDB 865664 D01 SAR measurement 100 MHz to 6 GHz.

	≤3 GHz	> 3 GHz
Maximum distance from closest		
measurement point (geometric center of	5 ± 1 mm	½·δ·ln(2) ± 0.5 mm
probe sensors) to phantom surface		
Maximum probe angle from probe axis to		
phantom surface normal at the	30° ± 1°	20° ± 1°
measurement location		
	≤ 2 GHz: ≤ 15 mm	3 – 4 GHz: ≤ 12 mm
	2 – 3 GHz: ≤ 12 mm	4 – 6 GHz: ≤ 10 mm
	When the x or y dimen	sion of the test device, in
Maximum area scan spatial resolution:	the measurement plar	ne orientation, is smaller
ΔxArea, ΔyArea	than the above, the n	neasurement resolution
	must be ≤ the correspo	nding x or y dimension of
	the test device with at	least one measurement
point on the test device.		



## Zoom Scan

Zoom scans are used assess the peak spatial SAR values within a cubic averaging volume containing 1 gram and 10 gram of simulated tissue. The zoom scan measures points (refer to table below) within a cube shoes base faces are centered on the maxima found in a preceding area scan job within the same procedure. When the measurement is done, the zoom scan evaluates the averaged SAR for 1 gram and 10 gram and displays these values next to the job's label.

Zoom scan parameters extracted from FCC KDB 865664 D01 SAR measurement 100 MHz to 6 GHz.

			≤3GHz	> 3 GHz				
Maximum zoom scan spatial resolution: $\triangle x_{zoom} \triangle y_{zoom}$			≤2GHz: ≤8mm	3 – 4GHz: ≤5mm*				
Maximum 200m	scan spa		2 – 3GHz: ≤5mm*	4 – 6GHz: ≤4mm*				
Maximum				3 – 4GHz: ≤4mm				
Maximum	U	niform grid: ∆z <sub>zoom</sub> (n)	≤5mm	4 – 5GHz: ≤3mm				
zoom scan				5 – 6GHz: ≤2mm				
spatial		$\triangle z_{zoom}(1)$ : between 1 <sup>st</sup> two		3 – 4GHz: ≤3mm				
resolution, normal to	Graded	points closest to phantom	≤4mm	4 – 5GHz: ≤2.5mm				
phantom		surface		5 – 6GHz: ≤2mm				
surface	grid	$ riangle z_{zoom}$ (n>1): between	<b>~1 F</b> a ^ -	z <sub>zoom</sub> (n-1)				
Sunace		subsequent points	≤1.3•△2	∠ <sub>zoom</sub> (11-1)				
Minimum				3 – 4GHz: ≥28mm				
zoom scan		X, y, z	≥30mm	4 – 5GHz: ≥25mm				
volume				5 – 6GHz: ≥22mm				
Note: $\delta$ is the pe	enetration	depth of a plane-wave at nor	mal incidence to the	tissue medium; see				
draft standard IE	EEE P152	8-2011 for details.						
* When zoom	* When zoom scan is required and the <i>reported</i> SAR from the area scan based 1-g SAR							
estimation proc	estimation procedures of KDB 447498 is ≤ 1.4W/kg, ≤8mm, ≤7mm and ≤5mm zoom scan							
resolution may b	be applied	, respectively, for 2GHz to 3G	Hz, 3GHz to 4GHz a	and 4GHz to 6GHz.				

#### **Volume Scan Procedures**

The volume scan is used for assess overlapping SAR distributions for antennas transmitting in different frequency bands. It is equivalent to an oversized zoom scan used in standalone measurements. The measurement volume will be used to enclose all the simultaneous transmitting antennas. For antennas transmitting simultaneously in different frequency bands, the volume scan is measured separately in each frequency band. In order to sum correctly to compute the 1g aggregate SAR, the EUT remain in the same test position for all measurements and all volume scan use the same spatial resolution and grid spacing. When all volume scan were completed, the software, SEMCAD postprocessor can combine and subsequently superpose these measurement data to calculating the multiband SAR.

## **Power Drift Monitoring**

All SAR testing is under the EUT install full charged battery and transmit maximum output power. In DASY measurement software, the power reference measurement and power drift measurement procedures are used for monitoring the power drift of EUT during SAR test. Both these procedures measure the field at a specified reference position before and after the SAR testing. The software will calculate the field difference in dB. If the power drifts more than 5%, the SAR will be retested.

# 7 Main Test Equipment

Name of Equipment	Manufacturer	Type/Model	Serial Number	Last Cal.	Cal. Due Date
Network Analyzer	Agilent	E5071B	MY42404014	2022-05-14	2023-05-13
Dielectric Probe Kit	SPEAG	DAK-12	1171	2022-10-29	2023-10-28
Power Meter	Agilent	E4417A	GB41291714	2022-05-14	2023-05-13
Power Sensor	Agilent	N8481H	MY50350004	2022-05-14	2023-05-13
Power Sensor	Agilent	E9327A	US40441622	2022-05-14	2023-05-13
Power Sensor	Agilent	NRP18S	101955	2022-05-14	2023-05-13
Signal Generator	Agilent	N5181A	MY50140143	2022-05-14	2023-05-13
Dual Directional Coupler	UCL	UCL-DDC0 56G-S	20010600118	1	/
Amplifier	INDEXSAR	TPA-005060 G01	13030502	2022-05-14	2023-05-13
E-field Probe	SPEAG	EX3DV4	3677	2022-07-08	2023-07-07
DAE	SPEAG	DAE4	1692	2022-11-18	2023-11-17
Validation Kit 2450MHz	SPEAG	D2450V2	786	2020-08-27	2023-08-26
Validation Kit 5GHz	SPEAG	D5GHzV2	1203	2022-12-09	2025-12-08
Software for Tissue	Agilent	85070	/	/	/
Temperature Probe	Tianjin jinming	JM222	381	2022-05-14	2023-05-13
Twin ELI Phantom	SPEAG	ELI v4.0	1058	/	/
Hygrothermograph	Anymetr	HTC - 1	TY2020A003	2022-05-14	2023-05-13
TX90 XL	SPEAG	Staubli TX90 XL	/	1	/
Software for Test	SPEAG	DASY52	52.10.4.1527	/	/



# 8 Tissue Dielectric Parameter Measurements & System Check

# 8.1 **Tissue Verification**

The temperature of the tissue-equivalent medium used during measurement must also be within 18°C to 25°C and within ± 2°C of the temperature when the tissue parameters are characterized. The dielectric parameters must be measured before the tissue-equivalent medium is used in a series of SAR measurements. The parameters should be re-measured after each 24 hours of use; or earlier if the dielectric parameters can become out of tolerance.

#### Target values

Frequency (MHz)	٤r	σ(s/m)
2450	39.2	1.80

#### **Measurements results**


Frequency	Test Date	Temp		Dielectric neters	•	vielectric neters		nit n ±5%)	
(MHz)	Test Date	°C	٤r	σ(s/m)	٤r	σ(s/m)	Dev ε <sub>r</sub> (%)	Dev σ(%)	
2450	2023/4/26	21.5	38.7	1.82	39.2	1.80	-1.28	1.11	
5250	2023/4/26	21.5	35.7	4.74	35.9	4.71	-0.56	0.64	
5750	2023/4/26	21.5	35.2	5.32	35.4	5.22	-0.56	1.92	
	Note: The depth of tissue-equivalent liquid in a phantom must be $\geq 15.0$ cm for SARmeasurements $\leq 3$ GHz and $\geq 10.0$ cm for measurements $> 3$ GHz.								



# 8.2 System Check

The manufacturer calibrates the probes annually. Dielectric parameters of the tissue simulates were measured using the dielectric probe kit and the network analyzer. A system check measurement for every day was made following the determination of the dielectric parameters of the Tissue simulates, using the dipole validation kit. The dipole antenna was placed under the flat section of the twin SAM phantom.

System check is performed regularly on all frequency bands where tests are performed with the DASY system.



Picture 1 System Check setup



#### **Picture 2 Setup Photo**

### Justification for Extended SAR Dipole Calibrations

Usage of SAR dipoles calibrated less than 3 years ago but more than 1 year ago were confirmed in maintaining return loss (< - 20 dB, within 20% of prior calibration) and impedance (within 5 ohm from prior calibration) requirements per extended calibrations in KDB 865664 D01:

Dipole		Date of Return Loss (dB)		Date of		A 0/		Impeda	nce (Ω)	
				Δ%	Real	ΔΩ	Imaginary	ΔΩ		
		8/27/2020	26.9	/	54.5	/	1.44	/		
	Dipole D2450V2 Head SN: 786 Liquid	8/26/2021	27.1	0.7	53.8	-0.7	1.43	-0.01		
SN. 700		8/25/2022	27.4	1.1	53.4	-0.4	1.43	0		

### System Check Results

Frequency (MHz)	Test Date	Temp ℃	250mW Measured SAR <sub>1g</sub> (W/kg)	1W Normalized SAR <sub>1g</sub> (W/kg)	1W Target SAR <sub>1g</sub> (W/kg)	Δ % (Limit ±10%)	Plot No.		
2450	2023/4/26	21.5	13.52	54.08	52.30	3.40	1		
Frequency (MHz)	Test Date	Temp ℃	100mW Measured SAR <sub>1g</sub> (W/kg)	1W Normalized SAR <sub>1g</sub> (W/kg)	1W Target SAR <sub>1g</sub> (W/kg)	Δ % (Limit ±10%)	Plot No.		
5250	2023/4/26	21.5	7.54	75.40	77.70	-2.96	2		
5750	2023/4/26	21.5	7.75	77.50	76.80	0.91	3		
Note: Target	Note: Target Values used derive from the calibration certificate data storage and evaluation.								

## 8.3 SAR System Validation

Per FCC KDB 865664 D02v01, SAR system verification is required to confirm measurement accuracy. The SAR systems (including SAR probes, system components and software versions) used for this device were validated against its performance specifications prior to the SAR measurements. Reference dipoles are used with the required tissue-equivalent media for system validation, according to the procedures outlined in FCC KDB 865664 D01 and IEEE 1528-2013. Since SAR probe calibrations are frequency dependent, each probe calibration point must be validated at a frequency within the valid frequency range of the probe calibration point, using the system that normally operates with the probe for routine SAR measurements and according to the required tissue-equivalent media.

A tabulated summary of the system validation status, measurement frequencies, SAR probes, calibrated signal type(s) and tissue dielectric parameters has been included.

Eroqueney		Probe	Droho			DI				PERM	COND	CW	Validatio	า
Frequency [MHz]	Date	SN	Probe Type	Probe C	Cal Point		(Σ)	Sensitivity	Probe	Probe				
		31	туре					Sensitivity	Linearity	Isotropy				
2450	2022/7/8	3677	EX3DV4	2450	Head	39.2	1.80	PASS	PASS	PASS				
5250	2022/7/8	3677	EX3DV4	5250	Head	35.9	4.71	PASS	PASS	PASS				
5750	2022/7/8	3677	EX3DV4	5750	Head	35.4	5.22	PASS	PASS	PASS				

NOTE: While the probes have been calibrated for both CW and modulated signals, all measurements were performed using communication systems calibrated for CW signals only. Modulations in the table above represent test configurations for which the measurement system has been validated per FCC KDB Publication 865664D01v01 for scenarios when CW probe calibrations are used with other signal types. SAR systems were validated for modulated signals with a periodic duty cycle, such as GMSK, or with a high peak to average ratio (>5dB), such as OFDM according to KDB 865664.



# 9 Normal and Maximum Output Power

KDB 447498 D01 at the maximum rated output power and within the tune-up tolerance range specified for the product, but not more than 2 dB lower than the maximum tune-up tolerance limit.

# 9.1 WLAN Mode

Wi-Fi 2.4G	Channel	Мах	kimum Output Power (dBm)
Antenna 1	/Frequency(MHz)	Tune-up	Meas.
Mode		Tune-up	ineas.
000 446	1/2412	13.00	12.37
802.11b (1M)	6/2437	13.00	11.54
(111)	11/2462	13.00	11.53
000.44.5	1/2412	12.50	11.95
802.11g (6M)	6/2437	12.50	10.83
(0101)	11/2462	12.50	10.78
	1/2412	12.50	11.97
802.11n-HT20 (MCS0)	6/2437	12.50	10.87
(10030)	11/2462	12.50	10.66
	3/2422	12.00	11.23
802.11n-HT40 (MCS0)	6/2437	12.00	11.11
(1000)	9/2452	12.00	10.61
Note: Initial test configu	uration is 802.11b mod	le.	

Wi-Fi 2.4G	Channel	Мах	kimum Output Power (dBm)				
Antenna 2	/Frequency(MHz)		Meas.				
Mode		Tune-up	INICAS.				
900 11h	1/2412	15.00	14.13				
802.11b (1M)	6/2437	15.00	13.94				
(111)	11/2462	15.00	13.50				
000.44-	1/2412	14.00	13.30				
802.11g (6M)	6/2437	14.00	12.17				
	11/2462	14.00	12.32				
	1/2412	14.00	13.26				
802.11n-HT20 (MCS0)	6/2437	14.00	12.29				
(1000)	11/2462	14.00	12.43				
	3/2422	13.50	12.65				
802.11n-HT40 (MCS0)	6/2437	13.50	11.91				
(1000)	9/2452	13.50	11.14				
Note: Initial test config	Note: Initial test configuration is 802.11b mode.						

TA Technology (Shanghai) Co., Ltd.

TA-MB-06-003S

This report shall not be reproduced except in full, without the written approval of TA Technology (Shanghai) Co., Ltd.



Wi-Fi 2.4G	Channel	Max	ximum Output	Power (dBm)	)			
MIMO	Channel /Frequency(MHz)	Tune-up	Meas.	Ant1	Ant2			
Mode		Tune-up	IVICAS.	Anti	Antz			
	1/2412	16.50	15.67	11.97	13.26			
802.11n-HT20 (MCS0)	6/2437	16.50	14.65	10.87	12.29			
(1000)	11/2462	16.50	14.64	10.66	12.43			
	3/2422	16.00	15.01	11.23	12.65			
802.11n-HT40 (MCS0)	6/2437	16.00	14.54	11.11	11.91			
(1000)	9/2452	16.00	13.89	10.61	11.14			
Note: Initial test config	Note: Initial test configuration is 802.11n-HT20 mode.							

Wi-Fi 5G (U-NII-1)	Channel	Мах	ximum Output Power (dBm)
Antenna 1	/Frequency(MHz)	Tune-up	Meas.
Mode		Tune-up	ivieas.
000 44-	36/5180	14.00	13.12
802.11a (6M)	44/5220	14.00	12.14
(0101)	48/5240	14.00	12.20
	36/5180	12.00	11.50
802.11n-HT20 (MCS0)	44/5220	12.00	10.18
(1000)	48/5240	12.00	10.63
802.11n-HT40	38/5190	11.50	10.58
(MCS0)	46/5230	11.50	10.30
	36/5180	12.00	11.62
802.11ac-VHT20 (MCS0)	44/5220	12.00	10.49
(1000)	48/5240	12.00	10.54
802.11ac-VHT40	38/5190	11.50	10.56
(MCS0)	46/5230	11.50	10.03
Note. Initial test configu	uration is 802.11a mod	le, since the high	est maximum output power.

Wi-Fi 5G (U-NII-1)	Charmel	Мах	kimum Output Power (dBm)
Antenna 2	Channel /Frequency(MHz)	Tune-up	Meas.
Mode		Tune-up	ivieas.
000 11-	36/5180	13.50	12.93
802.11a (6M)	44/5220	13.50	12.51
(0117)	48/5240	13.50	12.15
	36/5180	13.50	13.05
802.11n-HT20 (MCS0)	44/5220	12.00	10.77
(MC30)	48/5240	12.00	10.06
802.11n-HT40	38/5190	11.50	10.38

TA Technology (Shanghai) Co., Ltd.TA-MB-06-003SPage 22 of 79This report shall not be reproduced except in full, without the written approval of TA Technology (Shanghai) Co., Ltd.




SAR Test Report			Report No.: R2209A0844-S1
(MCS0)	46/5230	11.50	10.17
	36/5180	12.00	11.62
802.11ac-VHT20 (MCS0)	44/5220	12.00	10.84
(1000)	48/5240	12.00	10.45
802.11ac-VHT40	38/5190	11.50	10.29
(MCS0)	46/5230	11.50	10.46
Note. Initial test configu	iration is 802.11a mo	de, since the highes	t maximum output power.

Wi-Fi 5G (U-NII-1)	Channel	Мах	kimum Output	Power (dBm)	
MIMO	Channel · /Frequency(MHz)	Tune-up	Meas.	Ant1	Ant2
Mode		Tune-up	IVICAS.	Anti	Antz
900 11 <del>0</del> UT00	36/5180	16.00	15.35	11.50	13.05
802.11n-HT20 (MCS0)	44/5220	15.00	13.50	10.18	10.77
(1000)	48/5240	15.00	13.36	10.63	10.06
802.11n-HT40	38/5190	15.00	13.49	10.58	10.38
(MCS0)	46/5230	15.00	13.25	10.30	10.17
	36/5180	15.00	14.63	11.62	11.62
802.11ac-VHT20 (MCS0)	44/5220	15.00	13.68	10.49	10.84
(1000)	48/5240	15.00	13.51	10.54	10.45
802.11ac-VHT40	38/5190	15.00	13.44	10.56	10.29
(MCS0)	46/5230	15.00	13.26	10.03	10.46
Note. Initial test configu	uration is 802.11n-HT2	20 mode, since the	e highest max	imum output j	power.

Wi-Fi 5G (U-NII-3)	Channel	Мах	ximum Output Power (dBm)
Antenna 1	/Frequency(MHz)	Tune-up	Meas.
Mode		Tune-up	ivicas.
000 11-	149/5745	14.50	13.30
802.11a (6M)	157/5785	14.50	13.80
(0101)	165/5825	14.50	13.71
	149/5745	14.00	12.10
802.11n-HT20 (MCS0)	157/5785	14.00	12.27
(1000)	165/5825	14.00	12.04
802.11n-HT40	151/5755	13.50	11.98
(MCS0)	159/5795	13.50	12.60
	149/5745	14.00	12.24
802.11ac-VHT20 (MCS0)	157/5785	14.00	12.95
(1000)	165/5825	14.00	12.91
802.11ac-VHT40	151/5755	13.50	12.14
(MCS0)	159/5795	13.50	12.97
Note. Initial test configu	uration is 802.11a mod	le, since the high	est maximum output power.

TA Technology (Shanghai) Co., Ltd.TA-MB-06-003SPage 2This report shall not be reproduced except in full, without the written approval of TA Technology (Shanghai) Co., Ltd.



11.35

SAR Test Report			Report No.: R2209A0844
Wi-Fi 5G (U-NII-3)	Channel	Ма	ximum Output Power (dBm)
Antenna 2	/Frequency(MHz)	Tune-up	Meas.
Mode	, <u>-</u> )		inicuo.
802.11a	149/5745	14.00	12.38
802.11a (6M)	157/5785	14.00	12.84
(0111)	165/5825	14.00	13.57
	149/5745	13.00	11.05
802.11n-HT20 (MCS0)	157/5785	13.00	11.24
(1000)	165/5825	13.00	11.91
802.11n-HT40	151/5755	12.50	10.87
(MCS0)	159/5795	12.50	11.29
	149/5745	13.00	11.10
802.11ac-VHT20 (MCS0)	157/5785	13.00	11.37
(10000)	165/5825	13.00	11.99
802.11ac-VHT40	151/5755	12.50	10.83

Note. Initial test configuration is 802.11a mode, since the highest maximum output power.

159/5795

(MCS0)

Wi-Fi 5G (U-NII-3)	Channel	Мах	kimum Output	Power (dBm)	1
MIMO Mode	Channel /Frequency(MHz)	Tune-up	Meas.	Ant1	Ant2
000 44- 11700	149/5745	16.00	14.62	12.10	11.05
802.11n-HT20 (MCS0)	157/5785	16.00	14.80	12.27	11.24
(1000)	165/5825	16.00	14.99	12.04	11.91
802.11n-HT40	151/5755	16.00	14.47	11.98	10.87
(MCS0)	159/5795	16.00	15.00	12.60	11.29
	149/5745	16.50	14.72	12.24	11.10
802.11ac-VHT20 (MCS0)	157/5785	16.50	15.24	12.95	11.37
(10030)	165/5825	16.50	15.48	12.91	11.99
802.11ac-VHT40	151/5755	16.00	14.54	12.14	10.83
(MCS0)	159/5795	16.00	15.25	12.97	11.35
Note. Initial test configu	uration is 802.11ac-V⊢	IT20 mode, since	the highest m	aximum outp	ut power.

12.50



# 10 Measured and Reported (Scaled) SAR Results

## **10.1 EUT Antenna Locations**

#### The Detailed Antenna Locations Refer to Antenna Locations.

0	verall (Length x Wio	dth): 362.1 mm x 26	9.2 mm	
Area Scan Antenna	Scan 1	Scan 2	Scan 3	Scan 4
Antenna 1	Yes	Yes	Yes	Yes
Antenna 2	Yes	Yes	Yes	Yes
Note:				
1. Per FCC KDB 447498 D01,				
for each exposure position, testi	ng of other requise	d channels within th	e operating mode o	of a frequency
band is not required when the re	ported 1-g or 10-g	SAR for the mid-ba	nd or highest outpu	t power channel
is:				
$\rightarrow$		- I I		

a)  $\leq$  0.8 W/kg or 2.0 W/kg, for 1-g or 10-g respectively, when the transmission band is  $\leq$  100MHz

b)  $\leq$  0.6 W/kg or 1.5 W/kg, for1-g or 10-g respectively, when the transmission band is between 100 MHz and 200 MHz.

c)  $\leq$  0.4 W/kg or 1.0 Wkg, for 1-g or 10-g respectively, when the transmission band is  $\geq$  200 MHz.

2. When the original highest measured SAR is  $\geq$  0.80 W/kg, the measurement was repeated once.



## **10.2 Measured SAR Results**

Note:

1. The value with blue color is the maximum SAR Value of each test band.

#### **Body SAR**

Band	Antenna	Test Position	Dist. (mm)	Mode	Duty Cycle	Ch./Freq. (MHz)		Measured power (dBm)	Area Scan	2 Zone Area Scan SAR 1g	Area Scan	4 Zone Area Scan SAR 1g	Worst Zone Zoom Scan SAR 1g	Power Drift (dB)	Scaling Factor	Report SAR1g (W/kg)	Plot No.
	ANT 1	Front Side	0	802.11b	100.0%	1/2412	13.00	12.37	0.004	0.001	0.046	0.003	0.102	0.099	1.16	0.118	4
2.4G		Front Side	0	802.11n HT20	100.0%	1/2412	12.50	11.97	0.004	0.001	0.006	0.003	0.049	0.044	1.13	0.056	/
2.40	ANT 2	Front Side	0	802.11b	100.0%	1/2412	15.00	14.13	0.007	0.001	0.001	0.001	0.008	0.132	1.22	0.010	/
	ANT 2	Front Side	0	802.11n HT20	100.0%	1/2412	14.00	13.26	0.007	0.001	0.001	0.001	0.005	0.038	1.19	0.006	/
	ANT 1	Front Side	0	802.11a	100.0%	36/5180	14.00	13.12	0.045	0.042	0.064	0.054	0.080	0.105	1.22	0.098	/
U-NII-1	ANTI	Front Side	0	802.11n HT20	100.0%	36/5180	12.00	11.50	0.041	0.045	0.068	0.056	0.079	0.138	1.12	0.088	/
U-NII-T	ANT 2	Front Side	0	802.11a	100.0%	36/5180	13.50	12.93	0.067	0.046	0.045	0.046	0.095	0.032	1.14	0.109	/
	ANT 2	Front Side	0	802.11n HT20	100.0%	36/5180	13.50	13.05	0.077	0.045	0.049	0.049	0.095	0.030	1.11	0.105	/
	ANT 1	Front Side	0	802.11a	100.0%	157/5785	14.50	13.80	0.071	0.097	0.223	0.069	0.154	0.050	1.17	0.181	/
		Front Side	0	802.11n HT20	100.0%	157/5785	14.00	12.27	0.048	0.064	0.158	0.050	0.157	0.066	1.49	0.234	5
U-NII-3	ANT 2	Front Side	0	802.11a	100.0%	165/5825	14.00	13.57	0.124	0.069	0.053	0.051	0.100	0.030	1.10	0.110	/
	ANT 2	Front Side	0	802.11n HT20	100.0%	165/5825	13.00	11.91	0.120	0.072	0.063	0.060	0.107	0.171	1.29	0.138	/

## **10.3 Simultaneous Transmission Analysis**

Simultaneous Transmission Configurations	Body SAR
Wi-Fi 2.4G Antenna 1 + Wi-Fi 2.4G Antenna 2	Yes
Wi-Fi 5G Antenna 1 + Wi-Fi 5G Antenna 2	Yes
Wi-Fi 2.4G + Wi-Fi 5G	No

#### **General Note:**

1. The Scaled SAR summation is calculated based on the same configuration and test position.

2. Per KDB 447498 D01, simultaneous transmission SAR is compliant if,

i) Scalar SAR summation < 1.6W/kg, simultaneously transmission SAR measurement is not necessary.


ii) SPLSR =  $(SAR1 + SAR2)^{A1.5}$  / (min. separation distance, mm), and the peak separation distance is determined from the square root of  $[(x1-x2)^2 + (y1-y2)^2 + (z1-z2)^2]$ , where (x1, y1, z1) and (x2,

y2, z2) are the coordinates of the extrapolated peak SAR locations in the zoom scan.

iii) If SPLSR  $\leq$  0.04, simultaneously transmission SAR measurement is not necessary.

#### About Wi-Fi Antenna 1 and Antenna 2

	ront Side	SAR <sub>1g</sub>	(W/kg)	MAY SCAD.
F	Tont Side	Antenna 1	Antenna 2	MAX. ΣSAR <sub>1g</sub>
	Wi-Fi 2.4G	0.118	0.010	0.128
Body SAR	Wi-Fi 5G U-NII-1	0.098	0.109	0.207
	Wi-Fi 5G U-NII-3	0.234	0.138	0.372
Note:				
1. The value wi	th blue color is the maxin	num ΣSAR <sub>1g</sub> Value.		
2. MAX. ΣSAR	g =Unlicensed SAR <sub>MAX</sub> +	Licensed SAR <sub>MAX</sub>		
3. MAX. ΣSAR	<sub>lg</sub> = 0.372W/kg<1.6W/kg,	so the Simultaneous	transmission SAR v	vith volume scan
are not required	d for Wi-Fi Antenna 1 and	l Antenna 2.		





# **11 Measurement Uncertainty**

Per KDB 865664 D01 SAR Measurement 100 MHz to 6 GHz, when the highest measured 1-g SAR within a frequency band is < 1.5 W/kg, the extensive SAR measurement uncertainty analysis described in IEEE Std 1528- 2013 is not required in SAR reports submitted for equipment approval.

\*\*\*\*\*\*END OF REPORT \*\*\*\*\*\*



# **ANNEX A: Test Layout**





Rep

#### SAR Test Report Tissue Simulating Liquids

For the measurement of the field distribution inside the flat phantom with DASY, the phantom must be filled with around 25 liters of homogeneous tissue simulating liquid. For SAR testing, the liquid height from the center of the flat phantom to the liquid top surface is >15 cm, which is shown as below.



Picture 3: Liquid depth in the flat Phantom



# **ANNEX B: System Check Results**

Plot 1 System Performance Check at 2450 MHz TSL DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 Date: 2023/4/26 Communication System: CW; Frequency: 2450 MHz;Duty Cycle: 1:1 Medium parameters used: f = 2450 MHz;  $\sigma$  = 1.82 S/m;  $\epsilon_r$  = 38.7;  $\rho$  = 1000 kg/m<sup>3</sup> Ambient Temperature:22.3 °C Liquid Temperature: 21.5 °C Phantom section: Flat Section DASY5 Configuration: Sensor-Surface: 1.4mm (Mechanical Surface Detection) Probe: EX3DV4 - SN3677; ConvF(7.46, 7.46, 7.46); Calibrated: 2022/7/8 Electronics: DAE4 SN1692; Calibrated: 2022/11/18 Phantom: ELI v4.0; Type: QDOVA001BB; Serial:1058 Measurement SW: DASY52, Version 52.10 (4); SEMCAD X Version 14.6.14 (7483)

**d=10mm, Pin=250mW/Area Scan (4x7x1):** Measurement grid: dx=12mm, dy=12mm Maximum value of SAR (interpolated) = 14.01 W/kg

**d=10mm, Pin=250mW/Zoom Scan (7x7x7)/Cube 0:** Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 82.461V/m; Power Drift = 0.06 dB


Peak SAR (extrapolated) = 28.46 W/kg

SAR(1 g) = 13.52 W/kg; SAR(10 g) = 6.17 W/kg

Smallest distance from peaks to all points 3 dB below = 9.2 mm

Ratio of SAR at M2 to SAR at M1 = 50.1%

Maximum value of SAR (measured) = 15.35 W/kg





SAR Test Report

Plot 2 System Performance Check at 5250 MHz TSL

#### DUT: Dipole 5250 MHz; Type: D5GHzV2; Serial: D5GHzV2

Date: 2023/4/26 Communication System: CW; Frequency: 5250 MHz;Duty Cycle: 1:1 Medium parameters used: f = 5200 MHz;  $\sigma$  = 4.74 S/m;  $\epsilon_r$  = 35.7;  $\rho$  = 1000 kg/m<sup>3</sup> Ambient Temperature:22.3 °C Liquid Temperature: 21.5 °C Phantom section: Flat Section DASY5 Configuration: Sensor-Surface: 1.4mm (Mechanical Surface Detection) Probe: EX3DV4 - SN3677; ConvF(5.48, 5.48, 5.48); Calibrated: 2022/7/8 Electronics: DAE4 SN1692; Calibrated: 2022/11/18 Phantom: ELI v4.0; Type: QDOVA001BB; Serial:1058 Measurement SW: DASY52, Version 52.10 (4); SEMCAD X Version 14.6.14 (7483)

d=10mm, Pin=100mW/Area Scan (6x10x1): Measurement grid: dx=10mm, dy=10mm

Maximum value of SAR (interpolated) = 9.14 W/kg

**d=10mm, Pin=100mW/Zoom Scan (8x8x7)/Cube 0:** Measurement grid: dx=4mm, dy=4mm, dz=1.4mm

Reference Value = 36.428 V/m; Power Drift = -0.15 dB


Peak SAR (extrapolated) = 50.15 W/kg

### SAR(1 g) = 7.54 W/kg; SAR(10 g) = 2.27 W/kg

Smallest distance from peaks to all points 3 dB below = 7 mm

Ratio of SAR at M2 to SAR at M1 = 65.3%

Maximum value of SAR (measured) = 9.64 W/kg





SAR Test Report

Plot 3 System Performance Check at 5750 MHz TSL

#### DUT: Dipole 5750 MHz; Type: D5GHzV2; Serial: D5GHzV2

Date: 2023/4/26 Communication System: CW; Frequency: 5750 MHz;Duty Cycle: 1:1 Medium parameters used: f = 5750 MHz;  $\sigma$  = 5.32 S/m;  $\epsilon_r$  = 35.2;  $\rho$  = 1000 kg/m<sup>3</sup> Ambient Temperature:22.3 °C Liquid Temperature: 21.5 °C Phantom section: Flat Section DASY5 Configuration: Sensor-Surface: 1.4mm (Mechanical Surface Detection) Probe: EX3DV4 - SN3677; ConvF(5.00, 5.00, 5.00); Calibrated: 2022/7/8 Electronics: DAE4 SN1692; Calibrated: 2022/11/18 Phantom: ELI v4.0; Type: QDOVA001BB; Serial:1058 Measurement SW: DASY52, Version 52.10 (4); SEMCAD X Version 14.6.14 (7483)

d=10mm, Pin=100mW/Area Scan (6x10x1): Measurement grid: dx=10mm, dy=10mm

Maximum value of SAR (interpolated) = 7.84 W/kg

**d=10mm, Pin=100mW/Zoom Scan (8x8x7)/Cube 0:** Measurement grid: dx=4mm, dy=4mm, dz=1.4mm

Reference Value = 30.562 V/m; Power Drift = -0.018 dB


Peak SAR (extrapolated) = 22.26 W/kg


### SAR(1 g) = 7.75 W/kg; SAR(10 g) = 2.19 W/kg

Smallest distance from peaks to all points 3 dB below = 8 mm

Ratio of SAR at M2 to SAR at M1 = 55.6%

Maximum value of SAR (measured) = 8.58 W/kg








# **ANNEX C: Highest Graph Results**

# Plot 4 802.11b Front Side Low (Distance 0mm)

Date: 2023/4/26 Communication System: UID 0, 802.11b (0); Frequency: 2412 MHz;Duty Cycle: 1:1 Medium parameters used: f = 2412 MHz;  $\sigma$  = 1.801 S/m;  $\epsilon_r$  = 37.737;  $\rho$  = 1000 kg/m<sup>3</sup> Ambient Temperature:22.3 °C Liquid Temperature: 21.5 °C Phantom section: Flat Section DASY5 Configuration: Sensor-Surface: 1.4mm (Mechanical Surface Detection) Probe: EX3DV4 - SN3677; ConvF(7.46, 7.46, 7.46); Calibrated: 2022/7/8 Electronics: DAE4 SN1692; Calibrated: 2022/11/18 Phantom: ELI v4.0; Type: QDOVA001BB; Serial:1058 Measurement SW: DASY52, Version 52.10 (4); SEMCAD X Version 14.6.14 (7483)

**Front Side Low/Area Scan (14x14x1):** Measurement grid: dx=12mm, dy=12mm Maximum value of SAR (measured) = 0.105 W/kg

Front Side Low/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=85mm, dz=5mm Reference Value = 0 V/m; Power Drift = 0.099 dB Peak SAR (extrapolated) = 0.236 W/kg SAR(1 g) = 0.102 W/kg; SAR(10 g) = 0.047 W/kg Smallest distance from peaks to all points 3 dB below = 11.2 mm Ratio of SAR at M2 to SAR at M1 = 42.4% Maximum value of SAR (measured) = 0.173 W/kg






SAR Test Report

Plot 5 802.11a U-NII-3 Front Side Low (Distance 0mm) Date: 2023/4/26 Communication System: UID 0, 802.11nHT20(0); Frequency: 5785 MHz;Duty Cycle: 1:1 Medium parameters used: f = 5785 MHz;  $\sigma$  = 5.22 S/m;  $\epsilon_r$  = 35.343;  $\rho$  = 1000 kg/m<sup>3</sup> Ambient Temperature:22.3 °C Liquid Temperature: 21.5 °C Phantom section: Flat Section DASY5 Configuration: Sensor-Surface: 1.4mm (Mechanical Surface Detection) Probe: EX3DV4 - SN3677; ConvF(5.00, 5.00, 5.00); Calibrated: 2022/7/8 Electronics: DAE4 SN1692; Calibrated: 2022/11/18 Phantom: ELI v4.0; Type: QDOVA001BB; Serial:1058 Measurement SW: DASY52, Version 52.10 (4); SEMCAD X Version 14.6.14 (7483)

**Front Side Low/Area Scan (19x19x1):** Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (measured) = 0.259 W/kg

Front Side Low/Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 1.141 V/m; Power Drift = 0.066 dB Peak SAR (extrapolated) = 0.776 W/kg SAR(1 g) = 0.157 W/kg; SAR(10 g) = 0.088 W/kg Smallest distance from peaks to all points 3 dB below = 10.7 mm Ratio of SAR at M2 to SAR at M1 = 49.2% Maximum value of SAR (measured) = 0.304 W/kg



SAR Test Report

# ANNEX D: Probe Calibration Certificate (SN: 3677)

Add: No.52 HuaYuanBei Road. 1 Tel: +86-10-62304633-2117 E-mail: cttl@chinattl.com	http://www.caict.ac.en		261674
Client TA(Shan		PERSONAL PROPERTY ALTER.	60223
CALIBRATION CER	TIFICATE		
Object	EX3DV4 - S	N : 3677	
Calibration Procedure(s)			
	FF-Z11-004		
	Calibration F	Procedures for Dosimetric E-field Probes	
Calibration date:	July 08, 202	2	
pages and are part of the certifi		losed laboratory facility anyiranment tomas	rational and
All calibrations have been co humidity<70%.	onducted in the o	closed laboratory facility: environment tempe libration)	rature(22±3)℃ and
All calibrations have been co humidity<70%. Calibration Equipment used (M	onducted in the o	libration) Cal Date(Calibrated by, Certificate No.) Sch	rature(22±3)°C and eduled Calibration
All calibrations have been co humidity<70%. Calibration Equipment used (M. Primary Standards Power Meter NRP2	TE critical for ca	libration) Cal Date(Calibrated by, Certificate No.) Sch 14-Jun-22(CTTL, No.J22X04181)	
All calibrations have been co humidity<70%. Calibration Equipment used (M Primary Standards Power Meter NRP2 Power sensor NRP-Z91	TE critical for ca LD # 101919 101547	libration) Cal Date(Calibrated by, Certificate No.) Sch 14-Jun-22(CTTL, No.J22X04181) 14-Jun-22(CTTL, No.J22X04181)	eduled Calibration Jun-23 Jun-23
All calibrations have been co humidity<70%. Calibration Equipment used (M Primary Standards Power Meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91	&TE critical for ca ID # 101919 101547 101548	libration) Cal Date(Calibrated by, Certificate No.) Sch 14-Jun-22(CTTL, No.J22X04181) 14-Jun-22(CTTL, No.J22X04181) 14-Jun-22(CTTL, No.J22X04181)	eduled Calibration Jun-23 Jun-23 Jun-23
All calibrations have been co humidity<70%. Calibration Equipment used (M Primary Standards Power Meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 10dBAttenuator	TE critical for ca LD # 101919 101547 101548 18N50W-10dB	libration) Cal Date(Calibrated by, Certificate No.) Sch 14-Jun-22(CTTL, No.J22X04181) 14-Jun-22(CTTL, No.J22X04181) 14-Jun-22(CTTL, No.J22X04181) 20-Jan-21(CTTL, No.J21X00486)	eduled Calibration Jun-23 Jun-23 Jun-23 Jan-23
All calibrations have been co humidity<70%. Calibration Equipment used (M Primary Standards Power Meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 10dBAttenuator Reference 20dBAttenuator	TE critical for ca ID # 101919 101547 101548 18N50W-10dB 18N50W-20dB	libration) Cal Date(Calibrated by, Certificate No.) Sch 14-Jun-22(CTTL, No.J22X04181) 14-Jun-22(CTTL, No.J22X04181) 14-Jun-22(CTTL, No.J22X04181) 20-Jan-21(CTTL, No.J21X00486) 20-Jan-21(CTTL, No.J21X00485)	eduled Calibration Jun-23 Jun-23 Jun-23 Jan-23 Jan-23 Jan-23
All calibrations have been co humidity<70%. Calibration Equipment used (M Primary Standards Power Meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 10dBAttenuator	TE critical for ca LD # 101919 101547 101548 18N50W-10dB	libration) Cal Date(Calibrated by, Certificate No.) Sch 14-Jun-22(CTTL, No.J22X04181) 14-Jun-22(CTTL, No.J22X04181) 14-Jun-22(CTTL, No.J22X04181) 20-Jan-21(CTTL, No.J21X00486)	eduled Calibration Jun-23 Jun-23 Jun-23 Jan-23 Jan-23 Jan-23 Jan-23
All calibrations have been con humidity<70%. Calibration Equipment used (M. Primary Standards Power Meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 10dBAttenuator Reference 20dBAttenuator Reference Probe EX3DV4 DAE4	&TE critical for ca BD # 101919 101547 101548 18N50W-10dB 18N50W-20dB SN 7464	libration) Cal Date(Calibrated by, Certificate No.) Sch 14-Jun-22(CTTL, No.J22X04181) 14-Jun-22(CTTL, No.J22X04181) 14-Jun-22(CTTL, No.J22X04181) 20-Jan-21(CTTL, No.J21X00486) 20-Jan-21(CTTL, No.J21X00485) 26-Jan-22(SPEAG, No.EX3-7464_Jan22) 20-Aug-21(SPEAG, No.DAE4-1555_Aug21/2)	eduled Calibration Jun-23 Jun-23 Jun-23 Jan-23 Jan-23 Jan-23 Aug-22
All calibrations have been co humidity<70%. Calibration Equipment used (M. Primary Standards Power Meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 10dBAttenuator Reference 20dBAttenuator Reference Probe EX3DV4	8TE critical for ca ID # 101919 101547 101548 18N50W-10dB 18N50W-20dB SN 7464 SN 1555	libration) Cal Date(Calibrated by, Certificate No.) Sch 14-Jun-22(CTTL, No.J22X04181) 14-Jun-22(CTTL, No.J22X04181) 14-Jun-22(CTTL, No.J22X04181) 20-Jan-21(CTTL, No.J21X00486) 20-Jan-21(CTTL, No.J21X00485) 26-Jan-22(SPEAG, No.EX3-7464_Jan22) 20-Aug-21(SPEAG, No.DAE4-1555_Aug21/2) Cal Date(Calibrated by, Certificate No.) Sche	eduled Calibration Jun-23 Jun-23 Jan-23 Jan-23 Jan-23 Jan-23 Aug-22 duled Calibration
All calibrations have been conhumidity<70%. Calibration Equipment used (M. Primary Standards Power Meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 10dBAttenuator Reference 20dBAttenuator Reference Probe EX3DV4 DAE4 Secondary Standards	EXE critical for ca ID # 101919 101547 101548 18N50W-10dB 18N50W-20dB SN 7464 SN 1555 ID #	libration) Cal Date(Calibrated by, Certificate No.) Sch 14-Jun-22(CTTL, No.J22X04181) 14-Jun-22(CTTL, No.J22X04181) 14-Jun-22(CTTL, No.J22X04181) 20-Jan-21(CTTL, No.J21X00486) 20-Jan-21(CTTL, No.J21X00485) 26-Jan-22(SPEAG, No.EX3-7464_Jan22) 20-Aug-21(SPEAG, No.DAE4-1555_Aug21/2)	eduled Calibration Jun-23 Jun-23 Jun-23 Jan-23 Jan-23 Jan-23 Aug-22
All calibrations have been con- humidity<70%. Calibration Equipment used (M. Primary Standards Power Meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 10dBAttenuator Reference 20dBAttenuator Reference Probe EX3DV4 DAE4 Secondary Standards SignalGenerator MG3700A Network Analyzer E5071C	Example conducted in the observation           &TE critical for call           ID #           101919           101547           101548           18N50W-10dB           18N50W-20dB           SN 7464           SN 1555           ID #           6201052605	Ibibration)           Cal Date(Calibrated by, Certificate No.)         Sch           14-Jun-22(CTTL, No.J22X04181)         14-Jun-22(CTTL, No.J22X04181)           14-Jun-22(CTTL, No.J22X04181)         20-Jan-21(CTTL, No.J21X00486)           20-Jan-21(CTTL, No.J21X00485)         26-Jan-22(SPEAG, No.EX3-7464_Jan22)           20-Aug-21(SPEAG, No.DAE4-1555_Aug21/2)         20-Aug-21(SPEAG, No.DAE4-1555_Aug21/2)           Cal Date(Calibrated by, Certificate No.)         Sche           14-Jun-22(CTTL, No.J22X04182)         14-Jan-22(CTTL, No.J22X0406)	eduled Calibration Jun-23 Jun-23 Jan-23 Jan-23 Jan-23 Jan-23 Aug-22 duled Calibration Jun-23
All calibrations have been co humidity<70%. Calibration Equipment used (M. Primary Standards Power Meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 10dBAttenuator Reference 20dBAttenuator Reference Probe EX3DV4 DAE4 Secondary Standards SignalGenerator MG3700A Network Analyzer E5071C Na	Earlier         Second and the second se	Ibibration)           Cal Date(Calibrated by, Certificate No.)         Sch           14-Jun-22(CTTL, No.J22X04181)         14-Jun-22(CTTL, No.J22X04181)           14-Jun-22(CTTL, No.J22X04181)         20-Jan-21(CTTL, No.J21X00486)           20-Jan-21(CTTL, No.J21X00485)         26-Jan-22(SPEAG, No.EX3-7464_Jan22)           20-Aug-21(SPEAG, No.DAE4-1555_Aug21/2)         20-Aug-21(SPEAG, No.DAE4-1555_Aug21/2)           Cal Date(Calibrated by, Certificate No.)         Sche           14-Jun-22(CTTL, No.J22X04182)         14-Jan-22(CTTL, No.J22X0406)	eduled Calibration Jun-23 Jun-23 Jan-23 Jan-23 Jan-23 Jan-23 Aug-22 duled Calibration Jun-23 Jan-23
All calibrations have been co humidity<70%. Calibration Equipment used (M. Primary Standards Power Meter NRP2 Power sensor NRP-Z91 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 10dBAttenuator Reference 20dBAttenuator Reference Probe EX3DV4 DAE4 Secondary Standards SignalGenerator MG3700A Network Analyzer E5071C Na Calibrated by: Yu	Example conducted in the organization           &TE critical for ca           ID #           101919           101547           101548           18N50W-10dB           18N50W-20dB           SN 7464           SN 1555           ID #           6201052605           MY46110673	Ibibration)           Cal Date(Calibrated by, Certificate No.)         Sch           14-Jun-22(CTTL, No.J22X04181)         14-Jun-22(CTTL, No.J22X04181)           14-Jun-22(CTTL, No.J22X04181)         14-Jun-22(CTTL, No.J22X04181)           20-Jan-21(CTTL, No.J21X00486)         20-Jan-21(CTTL, No.J21X00485)           26-Jan-22(SPEAG, No.EX3-7464_Jan22)         20-Aug-21(SPEAG, No.DAE4-1555_Aug21/2)           20-Aug-21(SPEAG, No.DAE4-1555_Aug21/2)         Sche           14-Jun-22(CTTL, No.J22X00406)         Sche           14-Jun-22(CTTL, No.J22X00406)         Sche	eduled Calibration Jun-23 Jun-23 Jan-23 Jan-23 Jan-23 Jan-23 Aug-22 duled Calibration Jun-23 Jan-23

Certificate No: Z22-60223

Page 1 of 22





Add: No.52 Hua YuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2117 I-mail cftl a chinattl.com http://www.enict.ac.en

#### Glossary:

TSL	tissue simulating liquid
NORMx,y,z	sensitivity in free space
ConvF	sensitivity in TSL / NORMx,y,z
DCP	diode compression point
CF	crest factor (1/duty_cycle) of the RF signal
A.B.C.D	modulation dependent linearization parameters
Polarization Φ	Φ rotation around probe axis
Polarization 8	0 rotation around an axis that is in the plane nor

In Collaboration with

e CALIBRATION LABORATORY

plane normal to probe axis (at measurement center). I 0=0 is normal to probe axis

Connector Angle information used in DASY system to align probe sensor X to the robot coordinate system Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices Measurement Techniques", June 2013
- b) IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016
- c) IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010

d) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"


- Methods Applied and Interpretation of Parameters:
- NORMx, y.z: Assessed for E-field polarization 0=0 (IS900MHz in TEM-cell; I>1800MHz; waveguide) NORMx, y,z are only intermediate values, i.e., the uncertainties of NORMx, y,z does not effect the E<sup>2</sup>-field uncertainty inside TSL (see below ConvF).
- NORM(f)x,y,z = NORMx,y,z\* frequency\_response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- DCPx, y, z: DCP are numerical linearization parameters assessed based on the data of power sweep (no uncertainty required). DCP does not depend on frequency nor media.
- PAR: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics.
- Ax, y, z; Bx, y, z; Cx, y, z; VRx, y, z: A, B, C are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode
- ConvF and Boundary Effect Parameters: Assessed In flat phantom using E-field (or Temperature Transfer Standard for fs800MHz) and inside waveguide using analytical field distributions based on power measurements for f >800MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty valued are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORMx, y,z\* ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from±50MHz to±100MHz.
- Spherical isotropy (3D deviation from isotropy): in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required
- Connector Angle: The angle is assessed using the information gained by determining the NORMx (no uncertainty required).

Certificate No:Z22-60223

Page 2 of 22









Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2117 E-mail: cttl@chinattl.com http://www.caict.ac.cn

# DASY/EASY – Parameters of Probe: EX3DV4 – SN: 3677

#### **Basic Calibration Parameters**

	Sensor X	Sensor Y	Sensor Z	Unc (k=2)
Norm(µV/(V/m)²) A	0.42	0.46	0.41	±10,0%
DCP(mV) <sup>8</sup>	100.5	102.7	102.8	1.0

### Calibration Results for Modulation Response

UID	Communication System Name		dB	B dBõV	C	dB	WR mV	Max Dev.	Max Unc <sup>E</sup> (k=2)			
0	CW	X	0.0	0.0	1.0	0.00	150.8	±2.2%	±4.7%			
	1 · · · · · · · · · · · · · · · · · · ·	Y	0.0	0.0	1.0	22	161.2		100			
		Z	0.0	0.0	1.0		150.4					
10352-AAA	Pulse Waveform (200Hz, 10%)	X	1.64	60.07	6.04	The second	60	±4.8%	±9.6%			
	and the second second second	Y	1.81	60.93	6.48	10.00	60		100			
	and the second sec	Z	1.71	60.22	6.24	f frank	60					
10353-AAA	Pulse Waveform (200Hz, 20%)	X	1.21	60.00	5.26	1.00	80	±2.9%	±9.6%			
	and the state of the state of	Y	1.14	60.00	5.34	6.99	80		1000			
		Z	1.24	60.00	5.39		80					
10354-AAA	Pulse Waveform (200Hz, 40%)	X	0.78	60.00	4.62	1	95	±1.6%	±1.6%	±1.6%	±9.6%	
		Y	0.74	60.00	4.64	3.98	95					
	and the second s	Z	0.80	60.00	4.79		95		-			
10355-AAA	Pulse Waveform (200Hz, 60%)	X	0.51	60.00	3.94	0	120	±1.4%	±9.6%			
	Lander we show a true	Y	0.47	60.00	4.02	2.22	120					
		Z	0.51	60.00	4.20	in the second	120			100		
10387-AAA	QPSK Waveform, 1 MHz	X	1.24	63.61	12.00	2.00	200	150 ±3.1%	±3.1%	±3.1%	±3.1%	±9.6%
		Y	1.42	66.07	13.87	1.00	150	1				
	D. I. S. Martin Martin Street	Z	1.27	65.09	12.91	1	150	1000	1.1.2			
10388-AAA	QPSK Waveform, 10 MHz	X	1.77	65.04	13.47	1	150	±1.5%	±9.6%			
- concernery	and the second	Y	1.97	67.16	15.01	0.00	0 150	150	Energy.	1.1.1		
	the second se	Z	1.81	66.06	14.28		150	1				
10396-AAA	64-QAM Waveform, 100 kHz	X	2.27	67.24	17.73	1.00	150	±0.9%	±0.9%	±9.6%		
A POSTAN	Frankeren an fred also and	Y	2.50	69.43	19.12	3.01	150					
	and the second se	Z	2.22	67.67	18.11	2.0	150					
10414-AAA	WLAN CCDF, 64-QAM, 40MHz	X	4.59	65.39	15.13	1	150	50 ±3.7%	% ±9.6%			
around the	COLUMN STATEMENT STOLEN AND	Y	4.67	65.83	15.53	0.00	150		1.00			
		Z	4.55	65.64	15.34	I family a	150		1.000			

Note: For details on UID parameters see Appendix

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

<sup>A</sup> The uncertainties of Norm X, Y, Z do not affect the E<sup>2</sup> field uncertainty inside TSL (see Page 5).

 Internical linearization parameter: uncertainty not required.
 Uncertainty is determined using the max, deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

Certificate No:Z22-60223

Page 3 of 22







Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2117 E-mail: ettl@chinattl.com http://www.eaict.ac.cn

# DASY/EASY - Parameters of Probe: EX3DV4 - SN: 3677

## Sensor Model Parameters

	C1 fF	C2 fF	α V-1	T1 ms.V <sup>-2</sup>	T2 ms.V <sup>-1</sup>	T3 ms	T4 V-2	T5 V1	TG
X	31.29	236.58	35.88	18.80	0.00	4.90	0.00	0.26	1.02
Y	31.84	237.52	35.33	17.20	0.00	4.90	0.23	0.24	1.02
Z	27.77	207.22	35.23	19.61	0.00	4.90	0.18	0.18	1.02


## Other Probe Parameters

Sensor Arrangement	Triangular
Connector Angle (°)	117.3
Mechanical Surface Detection Mode	enabled
Optical Surface Detection Mode	disable
Probe Overall Length	337mm
Probe Body Diameter	10mm
Tip Length	9mm
Tip Diameter	2.5mm
Probe Tip to Sensor X Calibration Point	1mm
Probe Tip to Sensor Y Calibration Point	1mm
Probe Tip to Sensor Z Calibration Point	1mm
Recommended Measurement Distance from Surface	1.4mm

Certificate No:Z22-60223

Page 4 of 22





Report No.: R2209A0844-S1



Add: No.52 HuaYunnBei Road, Haldian District, Beijing, 100191, China Tel: +86-10-62304633-2117 E-mail: enl@chinatiLeom http://www.enlet.ac.en

# DASY/EASY – Parameters of Probe: EX3DV4 – SN:3677

f (MHz) <sup>C</sup>	Relative Permittivity <sup>F</sup>	Conductivity (S/m) <sup>F</sup>	ConvF X	ConvF Y	ConvF Z	Alpha <sup>G</sup>	Depth <sup>o</sup> (mm)	Unct. (k=2)
750	41.9	0.89	9.63	9.63	9.63	0.15	1.35	±12,1%
835	41.5	0.90	9.34	9.34	9.34	0.14	1.46	±12.1%
1750	40.1	1.37	8.25	8.25	8.25	0.26	1.06	±12.1%
1900	40.0	1.40	7.84	7.84	7.84	0.27	1.05	±12.1%
2000	40.0	1.40	7.92	7.92	7.92	0.21	1,27	±12.1%
2300	39.5	1.67	7.76	7.76	7.76	0.65	0.67	±12.1%
2450	39.2	1.80	7.46	7.46	7.46	0.64	0.70	±12.1%
2600	39.0	1.96	7.27	7.27	7.27	0.65	0.68	±12.1%
3300	38.2	2.71	7.02	7.02	7.02	0.45	0.92	±13.3%
3500	37.9	2.91	6.90	6.90	6.90	0.44	0.96	±13.3%
3700	37.7	3.12	6.64	6.64	6.64	0.44	1.01	±13.3%
3900	37.5	3.32	6,58	6.58	6.68	0.40	1.25	±13.3%
4100	37.2	3.53	6.60	6.60	6.60	0.40	1.15	±13.3%
4400	36.9	3.84	6.40	6.40	6.40	0.40	1.25	±13.3%
4600	36.7	4.04	6.31	6.31	6.31	0.45	1.25	±13.3%
4800	36.4	4.25	6.26	6.26	6.26	0.50	1.20	±13.3%
4950	36.3	4.40	6.03	6.03	6.03	0.45	1.30	±13.3%
5250	35.9	4.71	5.48	5.48	5.48	0.50	1.20	±13.3%
5600	35.5	5.07	4.97	4.97	4.97	0.50	1.30	±13.3%
5750	35.4	5.22	5.00	5.00	5.00	0.50	1.32	±13.3%

## Calibration Parameter Determined in Head Tissue Simulating Media


<sup>6</sup> Frequency validity above 300 MHz of ±100MHz only applies for DASY v4.4 and higher (Page 2), else it is restricted to ±50MHz. The uncertainty is the RSS of ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.

<sup>F</sup> At frequency below 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) can be relaxed to ±10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) is restricted to ±5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters. <sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for the frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

Certificate No:Z22-60223

Page 5 of 22










Add: No.52 HuaYuanBei Road, Haidian District, Heijing, 100191, China Tel: +86-10-62304633-2117 E-mail: ettl@chinattl.com http://www.calet.ac.en


# Frequency Response of E-Field (TEM-Cell: ifi110 EXX, Waveguide: R22)




Uncertainty of Frequency Response of E-field: ±7.4% (k=2)

Certificate No:Z22-60223

Page 6 of 22





270

Ý

- Z

Tot



Certificate No:Z22-60223


270


Y

- z

x

Page 7 of 22







# **Conversion Factor Assessment**

f=750 MHz,WGLS R9(H\_convF)

f=1750 MHz,WGLS R22(H\_convF)



# Deviation from Isotropy in Liquid



Uncertainty of Spherical Isotropy Assessment: ±3.2% (k=2)

Certificate No:Z22-60223

Page 9 of 22



17

.....

S



Add: No.52 HuaYuanBci Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2117 E-mail: cttl@chinattl.com http://www.caiet.ac.cn

а

## Appendix: Modulation Calibration Parameters

In Collaboration with

pe

CALIERATION LABORATORY

סוט	Rev	Communication System Name	Group	PAR (dB)	UncE (k=2)
0		CW	CW	0.00	±4.79
10010	CAA	SAR Validation (Square, 100ms, 10ms)	Test	10.00	±96%
10011	CAB	UMTS-FDD (WCDMA)	WCDMA	2.91	±96%
10012	CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps)	WLAN	1.87	± 9.6 9
10013	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 6 Mbps)	WLAN	9.46	±96%
10021	DAC	GSM-FDD (TDMA, GMSK)	GSM	9.39	± 9.6 %
10023	DAC	GPRS-FDD (TDMA, GMSK, TN 0)	GSM	9.57	±9.6 %
10024	DAC	GPRS-FDD (TDMA, GMSK, TN 0-1)	GSM	6.56	± 9.6 %
10025	DAC		GSM	12.62	±96%
10026	DAC	EDGE-FDD (TDMA, 8PSK, TN 0-1)	GSM	9.55	± 9.6 %
10027	DAC	GPRS-FDD (TDMA, GMSK, TN 0-1-2)	GSM	4.80	±96%
10028	DAC	GPRS-FDD (TDMA, GMSK, TN 0-1-2-3)	GSM	3.55	±9.6 %
10029	DAC		GSM	7,78	±9.6 %
10025		IEEE 802.15.1 Bluetooth (GFSK, DH1)	Bluetooth	5.30	± 9.6 9
10030	CAA	IEEE 802.15.1 Bluetooth (GFSK, DH3)	Bluetooth	1.87	±9.6 %
10032		IEEE 802.15.1 Bluetooth (GFSK, DH5)	Bluetooth	1.16	±9.6 9
10032	CAA	IEEE 802.15.1 Bluetooth (PI/4-DQPSK, DH1)	Bluetooth	7.74	±9.6*
10034	CAA	IEEE 802.15.1 Bluetooth (PI/4-DQPSK, DH3)	Bluetooth	4.53	±9.6 °
10034	CAA	IEEE 802.15.1 Bluetooth (PI/4-DQPSK, DH5)	Bluetooth	3.83	±9.6 °
	CAA	IEEE 802.15.1 Bluetooth (8-DPSK, DH1)	Bluetooth	8.01	±9.6
10036		IEEE 802.15.1 Bluetooth (8-DPSK, DH3)	Bluetooth	4.77	±9.6*
10037	CAA	IEEE 802.15.1 Bluetooth (8-DPSK, DH5)	Bluetooth	4.10	±9.6
10038	CAA	CDMA2000 (1xRTT, RC1)	CDMA2000	4.57	±9.6*
10039	CAB	IS-54 / IS-136 FDD (TDMA/FDM, PI/4-DQPSK, Halfrate)	AMPS	7.78	±9.64
10042	CAB		AMPS	0.00	± 9.6 °
10044	CAA	IS-91/EIA/TIA-553 FDD (FDMA, FM)	DECT	13.80	±9.6
10048	CAA	DECT (TDD, TDMA/FDM, GFSK, Full Slot, 24)	DECT	10.79	±9.6
10049	CAA	DECT (TDD, TDMA/FDM, GFSK, Double Slot, 12)	TD-SCDMA	11.01	±9.6
10056	CAA	UMTS-TDD (TD-SCDMA, 1.28 Mcps)	GSM	6.52	±9.64
10058	DAC	EDGE-FDD (TDMA, 8PSK, TN 0-1-2-3)	WLAN	2.12	±9.6
10059	CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 2 Mbps)	WLAN	2.83	±9.6
10060	CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 5.5 Mbps)	WLAN	3.60	±9.6
10061	CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 11 Mbps)	WLAN	8.68	±9.6
10062	CAD	IEEE 802.11a/h WiFi 5 GHz (OFDM, 6 Mbps)	WLAN	8.63	± 9.64
10063	CAD	IEEE 802.11a/h WiFi 5 GHz (OFDM, 9 Mbps)	WLAN	9.09	19.6
10064	CAD	IEEE 802 11a/n WiFi 5 GHz (OFDM, 12 Mbps)	WLAN	9.09	± 9.6
10065	CAD	IEEE 802.11a/h WiFi 5 GHz (OFDM, 18 Mbps)	WLAN	9.38	19.6
10066	CAD	IEEE 802 11a/h WiFi 5 GHz (OFDM, 24 Mbps)	WLAN		
10067	CAD	IEEE 802.11a/h WiFi 5 GHz (OFDM, 36 Mbps)		10.12	±9.6
10068	CAD	IEEE 802 11a/h WIFI 5 GHz (OFDM, 48 Mbps)	WLAN	10.24	±9.6 %
10069	CAD	IEEE 802.11a/h WiFi 5 GHz (OFDM, 54 Mbps)	WLAN	10.55	±9.6
10071	CAB	IEEE 802 11g WiFi 2.4 GHz (DSSS/OFDM, 9 Mbps)	WLAN	9.83	±9.6
10072	CAB	IEEE 802 11g WiFi 2.4 GHz (DSSS/OFDM, 12 Mbps)	WLAN	9.62	±96
10073	CAB	IEEE 802 11g WiFI 2.4 GHz (DSSS/OFDM, 18 Mbps)	WLAN	9.94	±96
10074	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 24 Mbps)	WLAN	10.30	±9.6
10075	CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 36 Mbps)	WLAN	10.77	±96'
10076	CAB	IEEE 802.11g WIFI 2.4 GHz (DSSS/OFDM, 48 Mbps)	WLAN	10.94	±96
10077	CAB	IEEE 802.11g WiFI 2.4 GHz (DSSS/OFDM, 54 Mbps)	WLAN	11.00	±96
10081	CAB	COMA2000 (1xRTL RC3)	CDMA2000	3.97	±9.6
10082	CAB	IS-54 / IS-136 FDD (TDMA/FDM, PI/4-DQPSK, Fullrate)	AMPS	4.77	±96
10090	DAC	GPRS-FDD (TDMA, GMSK, TN 0-4)	GSM	6.56	± 9.6
10097	CAC	UMTS-FDD (HSDPA)	WCDMA	3.98	± 9.6
10098	DAC	UMTS-FDD (HSUPA, Subtest 2)	WCDMA	3.98	±96
10099	CAC	EDGE-FDD (TDMA, 8PSK, TN 0-4)	GSM	9.55	196
10100	CAC	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	5.67	±96
10101	CAR	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM)	LTE-FDD	6.42	±9.6


Certificate No:Z22-60223

Page 10 of 22



.

In Collaboration with





Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2117 E-mail: ettl@chinattl.com http://www.caict.ac.cn

10102	CAB	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM)	LTE-FDD	6.60	± 9.6 %
10103	DAC	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-TDD	9.29	± 9.6 %
10104	CAE	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM)	LTE-TDD	9.97	± 9.6 %
10105	CAE	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM)	LTE-TOD	10.01	± 9.6 %
10108	CAE	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, QPSK)	LTE-FDD	5.80	± 9.6 %
10109	CAG	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM)	LTE-FDD	6.43	± 9.6 %
10110	CAG	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, QPSK)	LTE-FDD	5.75	± 9.6 %
10111	CAG	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, 16-QAM)	LTE-FDD	6.44	±9.6%
10112	CAG	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM)	LTE-FDD	6.59	± 9.6 %
10113	CAG	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, 64-QAM)	LTE-FDD	6.62	± 9.6 %
10114	CAG	IEEE 802.11n (HT Greenfield, 13.5 Mbps, BPSK)	WLAN	8.10	± 9.6 %
10115	CAG	IEEE 802.11n (HT Greenfield, 81 Mbps, 16-QAM)	WLAN	8.46	±9.6%
10116	CAG	IEEE 802.11n (HT Greenfield, 135 Mbps, 64-QAM)	WLAN	8.15	±9.6%
10117	CAG	IEEE 802.11n (HT Mixed, 13.5 Mbps, BPSK)	WLAN	8.07	±9.6%
10118	CAD	IEEE 802.11n (HT Mixed, 81 Mbps, 16-QAM)	WLAN	8.59	± 9.6 %
10119	CAD	IEEE 802.11n (HT Mixed, 135 Mbps, 64-QAM)	WLAN	8.13	± 9.6 %
10140	CAD	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM)	LTE-FDD	6.49	±9.6%
10141	CAD	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM)	LTE-FDD	6.53	± 9.6 %
10142	CAD	LTE-FDD (SC-FDMA, 100% RB, 3 MHz, QPSK)	LTE-FDD	5,73	± 9.6 %
10143		LTE-FDD (SC-FDMA, 100% RB, 3 MHz, 16-OAM)	LTE-FDD	6.35	±9.6 %
10144		LTE-FDD (SC-FDMA, 100% RB, 3 MHz, 64-QAM)	LTE-FDD	6.65	±9.6 %
10145		LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, QPSK)	LTE-FDD	5.76	±9.6%
10146	CAC	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM)	LTE-FDD	6.41	±9.6 %
10147	CAC		LTE-FDD	6.72	±9.6 %
10149	CAE	LTE-FDD (SC-FDMA, 50% RB, 20 MHz, 16-QAM)	LTE-FDD	6.42	± 9.6 %
10150		LTE-FDD (SC-FDMA, 50% RB, 20 MHz, 64-QAM)	LTE-FDD	6.60	±9.6 %
10151	CAE		LTE-TDD	9.28	± 9.6 %
10152		LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 16-QAM)	LTE-TDD	9.92	±9.6%
10153		LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 64-QAM)	LTE-TDD	10.05	±9.6 %
10154		LTE-FDD (SC-FDMA, 50% RB, 10 MHz, QPSK)	LTE-FDD	5.75	±9.6 %
10155	CAF	LTE-FDD (SC-FDMA, 50% RB, 10 MHz, 16-QAM)	LTE-FDD	6.43	±9.6 %
10156	CAF	LTE-FDD (SC-FDMA, 50% RB, 5 MHz, QPSK)	LTE-FDD	5.79	±9.6%
10157		LTE-FDD (SC-FDMA, 50% RB, 5 MHz, 16-QAM)	LTE-FDD	6.49	±9.6 %
10158		LTE-FDD (SC-FDMA, 50% RB, 10 MHz, 64-QAM)	LTE-FDD	6.62	±9.6%
10159		LTE-FDD (SC-FDMA, 50% RB, 5 MHz, 64-QAM)	LTE-FDD	6.56	±9.6%
10160		LTE-FDD (SC-FDMA, 50% RB, 15 MHz, QPSK)	LTE-FDD	5.82	± 9.6 %
10161		LTE-FDD (SC-FDMA, 50% RB, 15 MHz, 16-QAM)	LTE-FDD	6.43	±9.6%
10162		LTE-FDD (SC-FDMA, 50% RB, 15 MHz, 64-QAM)	LTE-FDD	6.58	±9.6%
10166		LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, QPSK)	LTE-FDD	5.46	±9.6%
10167		LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, 16-QAM)	LTE-FDD	6.21	± 9.6 %
10168		LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, 64-QAM)	LTE-FDD	6.79	± 9.6 %
10169		LTE-FDD (SC-FDMA, 1 RB, 20 MHz, QPSK) LTE-FDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM)	LTE-FDD	5.73	± 9.6 %
10170		LTE-FDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM)	LTE-FDD	6.52	±9.6%
10172	CAE	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 04-GAM)	LTE-FDD	6.49	± 9.6 %
10172	CAE		LTE-TDD	9.21	± 9.6 %
10173	CAF	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM)	LTE-TDD	9.48	± 9.6 %
10174	CAF	LTE-FDD (SC-FDMA, 1 RB, 10 MHz, OPSK)	LTE-TOD	10.25	± 9.6 %
10175	CAF	LTE-FDD (SC-FDMA, 1 RB, 10 MHz, 16-QAM)	LTE-FDD	5.72	±969
10176	CAF	LTE-FDD (SC-FDMA, 1 RB, 5 MHz, 10-QAM)	LTE-FDD	6.52	±9.6 %
10178	CAE	LTE-FDD (SC-FDMA, 1 RB, 5 MHz, 16-QAM)	LTE-FDD	5.73	±9.6 %
10178	AAE	LTE-FDD (SC-FDMA, 1 RB, 10 MHz, 64-QAM)	LTE-FDD	6.52	±9.6%
101/9	CAG		LTE-FDD	6.50	±9.6%
10180		LTE-FDD (SC-FDMA, 1 RB, 15 MHz, 04-0AM)	LTE-FDD	6,50	± 9,6 %
10181	CAG		LTE-FDD	5.72	± 9.6 %
10182		LTE-FDD (SC-FDMA, 1 RB, 15 MHz, 16-QAM)	LTE-FDD	6,52	±9.6%
10183	CAG		LTE-FDD	6.50	±9.6 %
10185	CAU	LTE-FDD (SC-FDMA, 1 RB, 3 MHz, 16-QAM)	LTE-FDD	5.73	± 9.6 %
10185	CAG	LTE-FDD (SC-FDMA, 1 RB, 3 MHz, 64-QAM)	LTE-FDD	6.51	± 9.6 %
10100	CAG	Tere too too tomin, Tho, 3 Minz, 04-0 Minj	LTE-FDD	6.50	±9.6


Certificate No:Z22-60223

Page 11 of 22

TA Technology (Shanghai) Co., Ltd.TA-MB-06-003SPage 46 of 79This report shall not be reproduced except in full, without the written approval of TA Technology (Shanghai) Co., Ltd.



In Collaboration with



CALIBRATION LABORATORY





 
 CAG
 LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz, QPSK)

 CAG
 LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM)

 CAE
 LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz, 64-QAM)
 LTE-FDD ±9.6 % 5.73 10187 LTE-FDD 6.52 ±9.6% 10188 LTE-FDD 6.50 ± 9.6 % 10189 WLAN ±9.6% IEEE 802.11n (HT Greenfield, 6.5 Mbps, BPSK) 8.09 10193 CAE ±9.6% AAD | IEEE 802.11n (HT Greenfield, 39 Mbps, 16-QAM) WLAN 8.12 10194 ±9.6% CAE | IEEE 802.11n (HT Greenfield, 65 Mbps, 64-QAM) WLAN 8.21 10195 WLAN 8.10 ±9.6% CAE IEEE 802.11n (HT Mixed, 6.5 Mbps, BPSK) 10196 ±9.6% WLAN 8.13 10197 AAE IEEE 802.11n (HT Mixed, 39 Mbps, 16-QAM) 8.27 ±9.6% CAF IEEE 802.11n (HT Mixed, 65 Mbps, 64-QAM) CAF IEEE 802.11n (HT Mixed, 7.2 Mbps, BPSK) WLAN 10198 ± 9.6 % WLAN 8.03 10219 ± 9.6 % WLAN 8.13 10220 AAF IEEE 802.11n (HT Mixed, 43.3 Mbps, 16-QAM WLAN 8.27 ±9.6% 10221 CAC | IEEE 802.11n (HT Mixed, 72.2 Mbps, 64-QAM) CAC IEEE 802.11n (HT Mixed, 15 Mbps, BPSK) CAD IEEE 802.11n (HT Mixed, 90 Mbps, 16-QAM) ±9.6% WLAN 8.06 10222 8.48 ± 9.6 % 10223 WLAN ± 9.6 % ± 9.6 % WLAN 8.08 CAD IEEE 802.11n (HT Mixed, 150 Mbps, 64-QAM) 10224 5.97 WCDMA 10225 CAD UMTS-FDD (HSPA+) 9.49 ± 9.6 % 10.26 ± 9.6 % CAD LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM) LTE-TDD 10226 LTE-TDD CAD LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 64-QAM) 10227 9.22 ± 9.6 % 9.48 ± 9.6 % CAD LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, QPSK) DAC LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 16-QAM LTE-TDD 10228 9.48 LTE-TDD 10229 16-QAM 10230 CAC LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 64-QAM) 10231 CAC LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 64-QAM) 10232 CAD LTE-TDD (SC-FDMA, 1 RB, 3 MHz, QPSK) 10.25 ± 9.6 % 9.19 ± 9.6 % LTE-TDD LTE-TDD 9.48 ± 9.6 % 10.25 ± 9.6 % LTE-TDD LTE-TDD 10233 CAD LTE-TDD (SC-FDMA, 1 RB, 5 MHz, 64-QAM) 9.21 ± 9.6 % 9.48 ± 9.6 % CAD LTE-TDD (SC-FDMA, 1 RB, 5 MHz, QPSK) CAD LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 16-QAM) LTE-TDD 10234 LTE-TDD 10235 64-QAM) LTE-TDD 10.25 ± 9.6 % 9.21 ± 9.6 % CAD LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 10235 LTE-TDD 10237 CAD LTE-TDD (SC-FDMA, 1 RB, 10 MHz, QPSK) 9.48 ± 9.6 % CAB LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 16-QAM) LTE-TDD 10238 10.25 ± 9.6 % ITE-TOD CAB LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 10239 64-QAM) LTE-TDD 9.21 ±9.6% 9.82 ±9.6% CAB LTE-TDD (SC-FDMA, 1 RB, 15 MHz, QPSK) 10240 CAB LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 16-QAM) LTE-TDD 10241 CAD LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 64-QAM) LTE-TDD 9.86 ±9.6 % 10242 LTE-TDD 9.46 ± 9.6 % CAD LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, QPSK) 10243 ± 9.6 % LTE-TDD 10.06 10244 CAD LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 16-QAM) ITE-TDD 10.06 ± 9.6 % 10245 CAG LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 64-QAM) 10246 CAG LTE-TDD (SC-FDMA, 50% RB, 3 MHz, OPSK) 10247 CAG LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 16-OAM) LTE-TDD 9.30 ± 9.6 % LTE-TDD 9.91 ± 9.6 % 10.09 ± 9.6 % 10248 CAG LTE-TOD (SC-FDMA, 50% RB, 5 MHz, 64-QAM) LTE-TDD LTE-TDD 9.29 ± 9.6 % 10249 CAG LTE-TDD (SC-FDMA, 50% RB, 5 MHz, QPSK) 10250 CAG LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 16-QAM) LTE-TDD 9.81 ± 9.6 % LTE-TDD 10.17 ± 9.6 % CAF LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 64-QAM) 10251 LTE-TDD LTE-TDD 10252 CAF LTE-TDD (SC-FDMA, 50% RB, 10 MHz, QPSK) 9.24 ±9.6% 16-QAM 10253 CAF LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 9.90 ±96% 10.14 ± 9.6 % 10254 CAB LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 64-0/ 10255 CAB LTE-TDD (SC-FDMA, 50% RB, 15 MHz, QPSK) 64-OAM) LTE-TDD LTE-TDD 9.20 ± 9.6 % 10256 CAB LTE-TOD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM) LTE-TDD 9.96 ± 9.6 % 10257 CAD LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM) LTE-TDD 10.08 ± 9.6 % 10258 CAD LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, OPSK) 10259 CAD LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 16-OAM) LTE-TDD 9.34 ± 9.6 % LTE-TDD 9.98 ±96% 10260 CAG LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 64-QAM) LTE-TDD 9.97 ±9.6% CAG LTE-TDD (SC-FDMA, 100% RB, 3 MHz, QPSK) LTE-TDD ±9.6% 9.24 10261 CAG LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 16-QAM) LTE-TDD ±9.6% 9.83 10262 CAG LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 64-QAM) CAG LTE-TDD (SC-FDMA, 100% RB, 5 MHz, QPSK) 10.16 ± 9.6 % LTE-TDD 10263 LTE-TDD 9.23 ±9.6% 10264 CAG LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 16-0AM LTE-TDD ± 9.6 % 9.92 10265 CAF LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 64-OAM) LTE-TDD 10.07 ± 9.6 % 10266 QPSK) CAF LTE-TOD (SC-FDMA, 100% RB, 10 MHz, ±9.6% LTE-TDD 9.30 10267 CAF LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM) LTE-TDD 10.06 ± 9.6 % 10268

Certificate No:Z22-60223

Page 12 of 22





Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2117 E-mail: cttl@chinattl.com http://www.caict.ac.cn

а

In Collaboration with

p e

CALIBRATION LABORATORY

5

10269	CAB	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM)	LTE-TDD	10.13	196%
10270	CAB	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, OPSK)	LTE-TDD	9.58	±9.6 %
0274	CAB	UMTS-FDD (HSUPA, Subtest 5, 3GPP Rel8.10)	WCDMA	4.87	±96%
0275	CAD	and the state state state state state of the	WCDMA	3.96	196%
0277		PHS (QPSK)	PHS	11.81	±96%
0278		PHS (QPSK, BW 884MHz, Rolloff 0.5)	PHS	11.81	±9.6%
10279	CAG	PHS (QPSK, 8W 884MHz, Rolloff 0.38)	PHS	12.18	± 9.6 %
10290	CAG		CDMA2000	3.91	±9.6%
10291	CAG	COMA2000, RC3, SO55, Full Rate	CDMA2000	3.46	±9.6%
10292		CDMA2000, RC3, SO32, Full Rate	CDMA2000	3.39	±96%
10293	CAG		CDMA2000	3,50	±96%
10295		CDMA2000, RC1, SO3, 1/8th Rate 25 fr.	CDMA2000	12.49	±9.6%
10297	CAF	LTE-FDD (SC-FDMA, 50% RB, 20 MHz, QPSK)	LTE-FDD	5.81	±9.6%
10298	CAF	LTE-FDD (SC-FDMA, 50% RB, 3 MHz, QPSK)	LTE-FOD	5.72	±9.6 %
10299		LTE-FDD (SC-FDMA, 50% RB, 3 MHz, 16-QAM)	LTE-FDD	6.39	±9.6 %
10300		LTE-FDD (SC-FDMA, 50% RB, 3 MHz, 64-QAM)	LTE-FDD	6.60	± 9.6 %
10301	CAC		WIMAX	12.03	±9.6%
10302	CAB	IEEE 802.16e WIMAX (29:18, 5ms, 10MHz, QPSK, PUSC, 3CTRL)	WIMAX	12.57	±9.6 %
10303	CAB		WIMAX	12.52	±96%
10304	CAA	IEEE 802.16e WIMAX (29:18, 5ms, 10MHz, 64QAM, PUSC)	WIMAX	11.86	±9.6%
10305	CAA	IEEE 802.16e WIMAX (31:15, 10ms, 10MHz, 64QAM, PUSC)	WIMAX	15.24	± 9.6 %
10305	CAA		WIMAX	14.67	±9.6 %
10307		IEEE 802.16e WIMAX (29:18, 10ms, 10MHz, QPSK, PUSC)	WIMAX	14.49	±9.6%
10308	AAB	IEEE 802.16e WIMAX (29:18, 10ms, 10MHz, 16QAM, PUSC)	WIMAX	14.46	±9.6%
10309	AAB	IEEE 802.16e WIMAX (29:18, 10ms, 10MHz, 16QAM,AMC 2x3)	WIMAX	14.58	±9.6%
10310	AAB	IEEE 802.16e WIMAX (29.18, 10ms, 10MHz, QPSK, AMC 2x3	WIMAX	14,57	± 9.6 %
10311	AAB	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, QPSK)	LTE-FDD	6.06	±9.6%
10313	AAD	IDEN 1:3	IDEN	10.51	± 9.6 %
10314	AAD	IDEN 1:6	IDEN	13.48	±9.6%
10315	AAD	IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps, 96pc dc)	WLAN	1.71	± 9.6 %
10316	AAD	IEEE 802.11g WiFi 2.4 GHz (ERP-OFDM. 6 Mbps, 96pc dc)	WLAN	8.36	± 9.6 %
10352	and the second second second	IEEE 802.11a WiFi 5 GHz (OFDM, 6 Mbps, 96pc dc)	WLAN	8.36	±9.6 %
10352	AAA	Pulse Waveform (200Hz, 10%)	Generic	10.00	±9.6 %
10354	AAA	Pulse Waveform (200Hz, 20%) Pulse Waveform (200Hz, 40%)	Generic	6.99	±9.69
10355	AAA		Generic	3.98	± 9.6 %
10356	AAA	Pulse Waveform (200Hz, 60%)	Generic	2.22	±9.6 %
10350	AAA	Pulse Waveform (200Hz, 80%)	Generic	0.97	±9.6 %
10388	AAA	OPSK Waveform, 1 MHz OPSK Waveform, 10 MHz	Generic	5.10	±9.69
10396	AAA	64-QAM Waveform, 100 kHz	Generic	5.22	19.69
10399	AAA		Generic	6.27	± 9.6 %
10399	AAD	64-QAM Waveform, 40 MHz	Generic	6.27	± 9.6 %
10400	AAA	IEEE 802.11ac WiFi (20MHz, 64-QAM, 99pc dc) IEEE 802.11ac WiFi (40MHz, 64-QAM, 99pc dc)	WLAN	8,37	19.6 %
10402	AAA	IEEE 802.11ac WiFi (80MHz, 64-QAM, 99pc dc)	WLAN	8.60	± 9.6 %
10402	AAB	CDMA2000 (1xEV-DO, Rev. 0)	WLAN	8.53	± 9.6.9
10403	AAB	CDMA2000 (1xEV-DO, Rev. 0)	CDMA2000	3.76	±9.6%
10404	AAD	CDMA2000 (1xEV-DO, Rev. A) CDMA2000, RC3, SO32, SCH0, Full Rate	CDMA2000	3.77	±9.6%
10400	AAA	LTE-TDD (SC-FOMA, 1 RB, 10 MHz, QPSK, UL Sub=2,3,4,7,8,9)	CDMA2000	5.22	± 9.6 %
10414	AAA	WLAN CCDF, 64-QAM, 40MHz	LTE-TOD	7.82	±9.6 %
10415	AAA	IEEE 802.11b WIFI 2.4 GHz (DSSS, 1 Mbps, 99pc dc)	Generic	8.54	± 9.6 %
10416	AAA	IEEE 802.11g WiFi 2.4 GHz (CSSS, 1 Mbps, 59pc dc)	WLAN	1.54	± 9.6 %
10417	AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 6 Mbps, 99pc dc)	WLAN	8.23	± 9.6 %
10418	AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 6 Mbps, 99pc dc)	WLAN	8.23	± 9.6 '
10419	AAA	IEEE 802 11a WELZ 4 CHY (DSSS OFOM & Mops, 99pc, Long)	WLAN	8.14	± 9.6
10419	AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 6 Mbps, 99pc, Short) IEEE 802.11n (HT Greenfield, 7.2 Mbps, BPSK)	WLAN	8.19	±96'
10423	AAA		WLAN	8.32	±96
10423	AAA	IEEE 802.11n (HT Greenfield, 43.3 Mbps, 16-QAM)	WLAN	8.47	± 9.6 %
10424		IEEE 802.11n (HT Greenfield, 72.2 Mbps, 64-QAM)	WLAN	8.40	±9.6 9
10425	AAE	IEEE 802.11n (HT Greenfield, 15 Mbps, BPSK)	WLAN	8.41	±9.6
10420	AAE	IEEE 802.11n (HT Greenfield, 90 Mbps, 16-QAM)	WLAN	8.45	± 9,6 °

Certificate No:Z22-60223


Page 13 of 22

TA Technology (Shanghai) Co., Ltd.TA-MB-06-003SPage 48 of 79This report shall not be reproduced except in full, without the written approval of TA Technology (Shanghai) Co., Ltd.



In Collaboration with

s p e a



CAICT

Add: No.52 HuaYuanBei Road, Haidian District, Beljing, 100191, China Tel: +86-10-62304633-2117 E-mail: ettl.@chinat0.com http://www.caict.ac.en

10427	AAB	IEEE 802.11n (HT Greenfield, 150 Mbps, 64-QAM)	WLAN	8.41	± 9.6 %
0430	AAB	LTE-FDD (OFDMA, 5 MHz, E-TM 3.1)	LTE-FDD	8.28	198%
0431	AAC	LTE-FDD (OFDMA, 10 MHz, E-TM 3.1)	LTE-FDD	8.38	19.6%
0432	AAB	LTE-FDD (OFDMA, 15 MHz, E-TM 3.1)	LTE-FDD	8.34	19.6 %
0433	AAC	LTE-FDD (OFDMA, 20 MHz, E-TM 3.1)	LTE-FDD	8.34	19.6%
0434	AAG	W-CDMA (BS Test Model 1, 64 DPCH)	WCDMA	8.60	19.6 %
0435	AAA	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK, UL Sub)	LTE-TDD	7.82	±9.6%
0447	AAA	LTE-FDD (OFDMA, 5 MHz, E-TM 3.1, Clipping 44%)	LTE-FDD	7.58	1989
0448	AAA	LTE-FDD (OFDMA, 10 MHz, E-TM 3 1, Clippin 44%)	LTE-FDD	7.53	1969
0449	AAC	LTE-FDD (OFDMA, 15 MHz, E-TM 3.1, Cliping 44%)	LTE-FDD	7.51	±98%
0450	AAA	LTE-FDD (OFDMA, 20 MHz, E-TM 3 1, Clipping 44%)	LTE-FDD	7.48	± 9.6 %
0451	AAA	W-CDMA (BS Test Model 1, 64 DPCH, Clipping 44%)	WCDMA	7.59	±9.0 %
10453	AAC	Validation (Square, 10ms, 1ms)	Test	10.00	±9.8%
10456	AAC	IEEE 802.11ac WiFi (160MHz, 64-QAM, 99pc dc)	WLAN	8.63	±969
0457	AAC	UMTS-FDD (DC-HSDPA)	WCDMA	6.62	± 9.6 %
10458	AAC	CDMA2000 (1xEV-DO, Rev. B, 2 carriers)	CDMA2000	6.55	±98%
10459	AAC	CDMA2000 (1xEV-DO, Rev. B, 3 carriers)	CDMA2000	8.25	± 9.6 %
10460	AAC	UMTS-FDD (WCDMA, AMR)	WCDMA	2.39	±96%
10461	AAC	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, QPSK, UL Sub)	LTE-TOD	7.82	± 9.6 %
0462	AAC	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM, UL Sub)	LTE-TOD	8.30	±9.6 %
0463	AAD	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 64-QAM, UL Sub)	LTE-TDD	8.56	± 9.6 %
10464	AAD	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, QPSK, UL Sub)	LTE-TOD	7.82	± 9.6 %
10465	AAC	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 16-QAM, UL Sub)	LTE-TDD	8.32	± 9.6 %
10466	AAC		LTE-TDD	8.57	± 9.6 %
10457	AAA	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, QPSK, UL Sub)	LTE-TDD	7.82	±9.6 %
0468		LTE-TDD (SC-FDMA, 1 RB, 5 MHz, 16-QAM, UL Sub)	LTE-TDD	8.32	± 9.6 9
0469	AAD	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, 64-QAM, UL Sub)	LTE-TDD	8.56	± 9.6 %
10470		LTE-TDD (SC-FDMA, 1 RB, 10 MHz, QPSK, UL Sub)	LTE-TDD	7.82	±9.6 %
10471		LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 16-QAM, UL Sub)	LTE-TOD	8.32	± 9.6 %
10472		LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 64-QAM, UL Sub)	LTE-TDD	8.57	±9.6%
10473		LTE-TDD (SC-FDMA, 1 RB, 15 MHz, QPSK, UL Sub)	LTE-TDD	7.82	±9.6%
10474	AAC	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 16-QAM, UL Sub)	LTE-TDD	8.32	± 9.6 %
10475	AAD		LTE-TDD	8.57	± 9.6 %
10477	AAC	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM, UL Sub)	LTE-TDD	8.32	± 9.6 %
10478	AAC	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 64-QAM, UL Sub)	LTE-TDD	8.57	±9.6 %
10479	AAC	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, QPSK, UL Sub)	LTE-TDD	7.74	± 9.6 %
10480	AAA	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 16-QAM, UL Sub)	LTE-TDD	8.18	±9.6 %
10481	AAA	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 64-QAM, UL Sub)	LTE-TOD	8.45	±9.6 %
10482	AAA	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, QPSK, UL Sub)	LTE-TDD	7.71	±9.6 %
10483	AAA	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 16-QAM, Sub)	LTE-TDD	8.39	± 9.6 %
10454	AAB	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 64-QAM, UL Sub)	LTE-TDD	8.47	±9.6 %
10485	AAB	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, QPSK, UL Sub)	LTE-TDD	7.59	± 9.6 %
10486	AAB	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 16-OAM, UL Sub)	LTE-TDD	8.38	±9.6 %
0487	AAC	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 64-QAM, UL Sub)	LTE-TDD	8,60	±969
0488	AAC	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, QPSK, UL Sub)	LTE-TDD	7.70	±96%
10489	AAC	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 16-OAM, UL Sub)	LTE-TDD	8,31	±9.6 %
10490	AAF	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 64-OAM, UL Sub)	LTE-TOD	8.54	±9.6 %
0491	AAF	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, QPSK, UL Sub)	LTE-TDD	7.74	± 9.6 %
0492	AAF	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 16-OAM, UL Sub)	LTE-TDD	8.41	±96 %
0493	AAF	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 64-QAM, UL Sub)	LTE-TDD	8.55	± 9.6 °
0494	AAF	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, QPSK, UL Sub)	LTE-TDD	7.74	±9.6 %
0495	AAF	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 16-QAM, UL Sub)	LTE-TOD	8.37	±9.6 %
0496	AAE	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 64-QAM, UL Sub)	LTE-TDD	8.54	19.6
0497	AAE	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, QPSK, UL Sub)	LTE-TDD	7.67	± 9.6 °
0498	AAE	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM, UL Sub)	LTE-TDD	8.40	±9.6 9
0499	AAC	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM, UL Sub)	LTE-TDD	8,68	± 9.6 %
0500	AAF	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, QPSK, UL Sub)	LTE-TDD	7.67	±9.6 %
0501	AAF	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM, UL Sub)	LTE-TDD	8.44	± 9.6 %
0502	AAB	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 64-QAM, UL Sub)	LTE-TOD	8.52	± 9.6 %

Certificate No:Z22-60223

Page 14 of 22



Report No.: R2209A0844-S1

.0 In Collaboration with

spe a q CALIBRATION LABORATORY

Add: No.52 HuaYuanBel Road, Haidian District, Belling, 100191, China Tel: +86-10-62304633-2117 E-mail: ettl@chinattl.com http://www.enict.ac.en

10503	AAB	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, QPSK, UL Sub)	LTE-TDD	7.72	±98%
0504	AAB	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 16 OAM, UL Sub)	LTE-TDD	8.31	±98%
10505	AAC	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 64-QAM, UL Sub)	LTE-TDD	8.54	±96%
10506	AAC	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, QPSK, UL Sub)	LTE-TDD	7.74	±96%
10507	AAC	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM, UL Sub)	LTE-TOD	8.30	±96%
10508	AAF	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM, UL Sub)	LTE-TDD	8.55	±9.6%
10509	AAF	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, QPSK, UL Sub)	LTE-TDD	7.99	±9.6%
10510	AAF	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM, UL Sub)	LTE-TDD	8.49	± 9.6 %
10511	AAF	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM, UL Sub)	LTE-TDD	8.51	±96%
10512	AAF	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, QPSK, UL Sub)	LTE-TDD		196%
10513	AAF	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM, UL Sub)	LTE-TDD	8.42	196%
10514	AAE	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM, UL Sub)	LTE-TDD	1.58	196%
10515	AAE	IEEE 802.11b WiFi 2.4 GHz (DSSS, 2 Mbps, 99pc dc)	WLAN	1.57	196%
10516	AAE	IEEE 802.11b WiFi 2.4 GHz (DSSS, 5.5 Mbps, 99pc dc)	WLAN	1.58	19.6%
10517	AAF	IEEE 802.11b WiFi 2.4 GHz (DSSS, 11 Mbps, 99pc dc)	WLAN	8.23	19.6%
10518	AAF	IEEE 802.11a/h WiFi 5 GHz (OFDM, 9 Mbps, 99pc dc)	WLAN	8.39	19.6%
10519	AAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 12 Mbps, 99pc dc) IEEE 802.11a/h WiFi 5 GHz (OFDM, 18 Mbps, 99pc dc)	WLAN	8.12	19.6%
	AAB		WLAN	7.97	±9.6 %
10521	AAB	IEEE 802 11a/h WiFi 5 GHz (OFDM, 24 Mbps, 99pc dc)	WLAN	8.45	196%
10522		IEEE 802 11a/h WiFi 5 GHz (OFDM, 36 Mbps, 99pc dc)	WLAN	8.08	±96%
10523	AAC	IEEE 802.11a/h WiFi 5 GHz (OFDM, 48 Mbps, 99pc dc) IEEE 802.11a/h WiFi 5 GHz (OFDM, 54 Mbps, 99pc dc)	WLAN	8.27	19.6%
10524	AAC	IEEE 802,11a/N WIFI 5 GHZ (OFDM, 54 MDps, 59pc dc)	WLAN	8.36	19.6%
10525	AAF	IEEE 802.11ac WiFi (20MHz, MCS0, 99pc dc)	WLAN	8.42	±9.6%
10520	AAF	IEEE 802.11ac WiFi (20MHz, MCS2, 99pc dc)	WLAN	8.21	±96%
10528	AAF	1EEE 802.11ac WiFI (20MHz, MCS2, 99pc dc)	WLAN	8.36	±9.6%
10528	AAF	IEEE 802.11ac WiFi (20MHz, MCS3, 350c dc)	WLAN	8.36	±9.6%
10529	AAF	IEEE 802, 11ac WiFi (20MHz, MCS6, 99pc dc)	WLAN	8.43	±9.6%
10532	AAF	IEEE 802.11ac WiFi (20MHz, MCS3, 99pc dc)	WLAN	8.29	± 9.6 %
10533	AAE	IEEE 802.11ac WiFi (20MHz, MCS8, 99pc dc)	WLAN	8.38	±9.6%
10534	AAE	IEEE 802.11ac WiFi (40MHz, MCS0, 99pc dc)	WLAN	8.45	±9.6 %
10535	AAE	IEEE 802.11ac WiFi (40MHz, MCS1, 99pc dc)	WLAN	8.45	±9.6%
10536	AAF	IEEE 802.11ac WiFi (40MHz, MCS2, 99pc dc)	WLAN	8.32	±9.6 %
10537	AAF	IEEE 802.11ac WiFi (40MHz, MCS3, 99pc dc)	WLAN	8.44	± 9.6 %
10538	AAF	IEEE 802 11ac WiFi (40MHz, MCS4, 99pc dc)	WLAN	8.54	± 9.6 %
10540	AAA	IEEE 802.11ac WiFi (40MHz, MCS6, 99pc dc)	WLAN	8.39	±9.6 %
10541	AAA	IEEE 802.11ac WiFi (40MHz, MCS7, 99pc dc)	WLAN	8.46	± 9.6 %
10542	AAA	IEEE 802.11ac WiFi (40MHz, MCS8, 99pc dc)	WLAN	8.65	±9.6%
10543	AAC	IEEE 802.11ac WiFi (40MHz, MCS9, 99pc dc)	WLAN	8.65	±9.6 %
10544	AAC	IEEE 802 11ac WiFI (80MHz, MCS0, 99pc dc)	WLAN	8.47	±96%
10545	AAC	IEEE 802.11ac WiFI (80MHz, MCS1, 99pc dc)	WLAN	8.55	± 9.6 %
10546	AAC	IEEE 802.11ac WiFi (80MHz, MCS2, 99pc dc)	WLAN	8.35	±96%
10547	AAC	IEEE 802.11ac WiFi (80MHz, MCS3, 99pc dc)	WLAN	8.49	±96%
10548	AAC	IEEE 802 11ac WiFI (BOMHz, MCS4, 99pc dc)	WLAN	8.37	±96%
10550	AAC	IEEE 802.11ac WiFI (80MHz, MCS6, 99pc dc)	WLAN	8.38	±9.6%
10551	AAC	IEEE 802.11ac WiFi (80MHz, MCS7, 99pc dc)	WLAN	8.50	±9.6 %
10552	AAC	IEEE 802.11ac WiFi (80MHz, MCS8, 99pc dc)	WLAN	8.42	±96%
10553	AAC	JEEE 802.11ac WiFi (80MHz, MCS9, 99pc dc)	WLAN	8.45	±9.6 %
10554	AAC	IEEE 802.11ac WiFi (160MHz, MCS0, 99pc dc)	WLAN	8.48	196%
10555	AAC	IEEE 802.11ac WiFI (160MHz, MCS1, 99pc dc)	WLAN	8.47	±96%
10556	AAC	IEEE 802.11ac WiFi (160MHz, MCS2, 99pc dc)	WLAN	8.50	±9.6%
10557	AAC	IEEE 802.11ac WiFI (160MHz, MCS3, 99pc dc)	WLAN	8.52	19.6%
10558	AAC	IEEE 802.11ac WiFi (160MHz, MCS4, 99pc dc)	WLAN	8.61	± 9.6 %
10560	AAC	IEEE 802.11ac WiFi (160MHz, MCS6, 99pc dc)	WLAN	8.73	19.6%
10561	AAC	IEEE 802.11ac WiFi (160MHz, MCS7, 99pc dc)	WLAN	8.56	± 9.6 %
10562	AAC	IEEE 802.11ac WiFi (160MHz, MCS8, 99pc dc)	WLAN	8.69	± 9.6 %
10563	AAC	IEEE 802.11ac WiFi (160MHz, MCS9, 99pc dc)	WLAN	8.77	±9.6%
10564	AAC	IEEE 802,11g WiFi 2.4 GHz (DSSS-OFDM, 9 Mbps, 99pc dc)	WLAN	8.25	19.6%
10565	AAC	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 12 Mbps, 99pc dc)	WLAN	8.45	±9.6%


Certificate No:Z22-60223

Page 15 of 22



Report No.: R2209A0844-S1

In Collaboration with





Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2117 E-mail: ettl@chinattl.com http://www.caicl.ac.cn

10566	AAC	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 18 Mbps, 99pc dc)	WLAN	6.13	±9.6%
10567	AAC	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 24 Mbps, 99pc dc)	WLAN	8.00	±9.6 %
0568	AAC	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 36 Mbps, 99pc dc)	WLAN	8.37	±96%
0569	AAC	IEEE 802.11g WIFi 2.4 GHz (DSSS-OFDM, 48 Mbps, 99pc dc)	WLAN	8.10	±9.6%
0570	AAC	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 54 Mbps, 99pc dc)	WLAN	8.30	±9.6%
10571	AAC	IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps, 90pc dc)	WLAN	1.99	±9.6%
10572	AAC	IEEE 802.11b WiFi 2.4 GHz (DSSS, 2 Mbps, 90pc dc)	WLAN	1.99	±9.6%
10573	AAC	IEEE 802.11b WiFi 2.4 GHz (DSSS, 5.5 Mbps, 90pc dc)	WLAN	1.98	±9.6%
10574	AAC	IEEE 802.11b WiFi 2.4 GHz (DSSS, 11 Mbps, 90pc dc)	WLAN	1.98	± 9.6 %
10575	AAC	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 6 Mbps, 90pc dc)	WLAN	8.59	±9.6%
10576	AAC	IEEE 802.11g WIFi 2.4 GHz (DSSS-OFDM, 9 Mbps, 90pc dc)	WLAN	8.60	±9.6%
10577	AAC	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 12 Mbps, 90pc dc)	WLAN	8.70	± 9.6 %
10578	AAD	IEEE 802.11g WIFI 2.4 GHz (DSSS-OFDM, 18 Mbps, 90pc dc)	WLAN	8.49	± 9.6 %
10579	AAD	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 24 Mbps, 90pc dc)	WLAN	8.36	±9.6 %
10580	AAD	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 36 Mbps, 90pc dc)	WLAN	8.76	±9.6 %
10581	AAD	IEEE 802.11g WIFI 2.4 GHz (DSSS-OFDM, 48 Mbps, 90pc dc)	WLAN	8.35	± 9.6 %
10582	AAD	IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 54 Mbps, 90pc dc)	WLAN	8.67	± 9.6 %
10583	AAD	IEEE 802.11a/h WiFi 5 GHz (OFDM, 6 Mbps, 90pc dc)	WLAN	8.59	± 9.6 %
10584	AAD	IEEE 802.11a/h WiFi 5 GHz (OFDM, 9 Mbps, 90pc dc)	WLAN	8.60	± 9.6 %
10585	AAD	IEEE 802.11a/h WiFi 5 GHz (OFDM, 12 Mbps, 90pc dc)	WLAN	8.70	± 9.6 %
10586	AAD	IEEE 802.11a/h WIFI 5 GHz (OFDM, 18 Mbps, 90pc dc)	WLAN	8.49	± 9.6 %
10587	AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 24 Mbps, 90pc dc)	WLAN	8.36	± 9.6 %
10588	AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 36 Mbps, 90pc dc)	WLAN	8.76	±9.6%
10589	AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 48 Mbps, 90pc dc)	WLAN	8.35	±9.6 %
10590	AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 54 Mbps, 90pc dc)	WLAN	8.67	±9.6%
10591	AAA	IEEE 802.11n (HT Mixed, 20MHz, MCS0, 90pc dc)	WLAN	8.63	± 9.6 %
10592	AAA	IEEE 802.11n (HT Mixed, 20MHz, MCS1, 90pc dc)	WLAN	8.79	±9.6%
10593	AAA	IEEE 802.11n (HT Mixed, 20MHz, MCS2, 90pc dc)	WLAN	8.64	± 9.6 %
10594	AAA	IEEE 802.11n (HT Mixed, 20MHz, MCS3, 90pc dc)	WLAN	8.74	±9.6%
10595	AAA	IEEE 802.11n (HT Mixed, 20MHz, MCS4, 90pc dc)	WLAN	8.74	± 9.6 %
10596	AAA	IEEE 802.11n (HT Mixed, 20MHz, MCS5, 90pc dc)	WLAN	8.71	± 9.6 %
10597	AAA	IEEE 802.11n (HT Mixed, 20MHz, MCS6, 90pc dc)	WLAN	8.72	±9.6%
10598	AAA	IEEE 802.11n (HT Mixed, 20MHz, MCS7, 90pc dc)	WLAN	8.50	±9.6%
10599	AAA	IEEE 802.11n (HT Mixed, 40MHz, MCS0, 90pc dc)	WLAN	8.79	±9.6%
10600	AAA	IEEE 802.11n (HT Mixed, 40MHz, MCS1, 90pc dc)	WLAN	8.88	± 9,6 %
10601	AAA	IEEE 802.11n (HT Mixed, 40MHz, MCS2, 90pc dc)	WLAN	8.82	±9.6%
10602	AAA	IEEE 802.11n (HT Mixed, 40MHz, MCS3, 90pc dc)	WLAN	8.94	±9.6%
10603	AAA	IEEE 802.11n (HT Mixed, 40MHz, MCS4, 90pc dc)	WLAN	9.03	± 9.6 %
10604	AAA	IEEE 802.11n (HT Mixed, 40MHz, MCS5, 90pc dc)	WLAN	8.76	±9.69
10605	AAA	IEEE 802.11n (HT Mixed, 40MHz, MCS6, 90pc dc)	WLAN	8.97	±9.6 %
10606	AAC	IEEE 802.11n (HT Mixed, 40MHz, MCS7, 90pc dc)	WLAN	8.82	±9.6 %
10607	AAC	IEEE 802.11ac WiFI (20MHz, MCS0, 90pc dc)	WLAN	8.64	±9.6 %
10608	AAC	IEEE 802.11ac WiFi (20MHz, MCS1, 90pc dc)	WLAN	8.77	± 9.6 %
10609	AAC	IEEE 802.11ac WiFi (20MHz, MCS2, 90pc dc)	WLAN	8.57	± 9.6 %
10610	AAC	IEEE 802.11ac WiFi (20MHz, MCS3, 90pc dc)	WLAN	8.78	±9.6 %
10611	AAC	IEEE 802,11ac WiFi (20MHz, MCS4, 90pc dc)	WLAN	8.70	± 9.6 %
10612	AAC	IEEE 802.11ac WiFi (20MHz, MCS5, 90pc dc)	WLAN	8,77	±9.6%
10613	AAC	IEEE 802.11ac WiFi (20MHz, MCS6, 90pc dc)	WLAN	8.94	±96%
10614	AAC	IEEE 802.11ac WiFI (20MHz, MCS7, 90pc dc)	WLAN	8.59	±9.6%
10615	AAC	IEEE 802.11ac WiFi (20MHz, MCS8, 90pc dc)	WLAN	8.82	±9.6%
10616	AAC	IEEE 802.11ac WiFi (40MHz, MCSO, 90pc dc)	WLAN	8.82	±9.6%
10617	AAC	IEEE 802.11ac WiFi (40MHz, MCS1, 90pc dc)	WLAN	8.81	±9.6 %
10618	AAC	IEEE 802.11ac WiFi (40MHz, MCS2, 90pc dc)	WLAN	8.58	±9.69
10619	AAC	IEEE 802 11ac WiFi (40MHz, MCS3, 90pc dc)	WLAN	8.86	±9.6 %
10620	AAC	IEEE 802.11ac WiFi (40MHz, MCS4, 90pc dc)	WLAN	8.87	±9.6 °
10621	AAC	IEEE 802.11ac WiFi (40MHz, MCS5, 90pc dc)	WLAN	8.77	±9.69
10622	AAC	IEEE 802.11ac WiFi (40MHz, MCS6, 90pc dc)	WLAN	8.68	±9.6 %
10623	AAC	IEEE 802.11ac WiFI (40MHz, MCS7, 90pc dc)	WLAN	8.82	± 9.6 %
10624	AAC	IEEE 802 11ac WIFi (40MHz, MCS8, 90pc dc)	WLAN	8.96	±9.6 %

Certificate No:Z22-60223

Page 16 of 22



L

Report No.: R2209A0844-S1



CALIBRATION LABORATORY Add: No.52 HuaYuanBei Read, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2117 E-mail: cttl@chinattl.com

a

a

http://www.caict.ac.en

e

In Collaboration with

p

s

10625	AAC	IEEE 802,11ac WiFi (40MHz, MCS9, 90pc dc)	WLAN	8.96	± 9.6 %
0626	AAC	IEEE 802.11ac WiFi (80MHz, MCS0, 90pc dc)	WLAN	8.83	± 9.6 %
0627	AAC	IEEE 802.11ac WiFi (80MHz, MCS1, 90pc dc)	WLAN	8.88	± 9.6 %
0628	AAC	IEEE 802.11ac WiFi (80MHz, MCS2, 90pc dc)	WLAN	8.71	±9.6 %
0629	AAC	IEEE 802.11ac WiFi (80MHz, MCS3, 90pc dc)	WLAN	8.85	± 9.6 %
0630	AAC	IEEE 802.11ac WiFi (80MHz, MCS4, 90pc dc)	WLAN	8.72	± 9.6 %
0631	AAC	IEEE 802.11ac WiFi (80MHz, MCS5, 90pc dc)	WLAN	8.81	± 9.6 %
0632	AAC	IEEE 802, 11ac WiFi (80MHz, MCS6, 90pc dc)	WLAN	8.74	±9.6 %
0633	AAC	IEEE 802,11ac WiFi (80MHz, MCS7, 90pc dc)	WLAN	8.83	±9.6%
10634	AAC	IEEE 802.11ac WiFi (80MHz, MCS8, 90pc dc)	WLAN	8.80	±9.6 %
0635	AAC	IEEE 802.11ac WiFi (80MHz, MCS9, 90pc dc)	WLAN	8.81	±9.6 %
0636	AAC	IEEE 802.11ac WiFi (160MHz, MCS0, 90pc dc)	WLAN	8.83	±9.6 %
0637	AAC	IEEE 802.11ac WiFi (160MHz, MCS1, 90pc dc)	WLAN	8.79	±9.6 %
0638	AAC	IEEE 802.11ac WiFi (160MHz, MCS2, 90pc dc)	WLAN	8.86	±9.6 %
0639	AAC	IEEE 802.11ac WiFi (160MHz, MCS3, 90pc dc)	WLAN	8.85	±9.6 %
0640	AAC	IEEE 802.11ac WiFi (160MHz, MCS4, 90pc dc)	WLAN	8.98	±9.6 %
0641	AAC	IEEE 802.11ac WIFI (160MHz, MCS5, 90pc dc)	WLAN	9.05	±9.6%
0642	AAC	IEEE 802.11ac WiFi (160MHz, MCS6, 90pc dc)	WLAN	9.06	± 9.6 %
0643	AAC	IEEE 802.11ac WiFi (160MHz, MCS7, 90pc dc)	WLAN	8.89	±9.6 %
0644	AAC	IEEE 802.11ac WiFi (160MHz, MCS8, 90pc dc)	WLAN	9.05	± 9.6 %
0645	AAC	IEEE 802.11ac WiFI (160MHz, MCS9, 90pc dc)	WLAN	9.11	±9.6 %
0646	AAC	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, QPSK, UL Sub=2,7)	LTE-TDD	11.96	±9.6
0647	AAC	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK, UL Sub=2,7)	LTE-TDD	11.96	± 9.6 %
0648	AAC	CDMA2000 (1x Advanced)	CDMA2000	3.45	± 9.6 %
0652	AAC	LTE-TDD (OFDMA, 5 MHz, E-TM 3.1, Clipping 44%)	LTE-TDD	6.91	± 9.6 5
0653	AAC	LTE-TDD (OFDMA, 10 MHz, E-TM 3.1, Clipping 44%)	LTE-TDD	7.42	±9.6 °
0654	AAC	LTE-TDD (OFDMA, 15 MHz, E-TM 3.1, Clipping 44%)	LTE-TDD	6.96	±9.6
0655	AAC	LTE-TDD (OFDMA, 20 MHz, E-TM 3.1, Clipping 44%)	LTE-TDD	7.21	± 9.6 9
0658	AAC	Pulse Waveform (200Hz, 10%)	Test	10.00	±9.6 9
0659	AAC	Pulse Waveform (200Hz, 20%)	Test	6.99	±9.6 °
0660	AAC	Pulse Waveform (200Hz, 40%)	Test	3.98	±9.6 9
0661	AAC	Pulse Waveform (200Hz, 60%)	Test	2.22	±9.6 %
0662	AAC	Pulse Waveform (200Hz, 80%)	Test	0.97	±9.6 %
0670	AAG	Bluetooth Low Energy	Bluetooth	2.19	±9.6
0671	AAD	IEEE 802.11ax (20MHz, MCS0, 90pc dc)	WLAN	9.09	± 9.6 °
0672	AAD	IEEE 802.11ax (20MHz, MCS1, 90pc dc)	WLAN	8.57	±9.6
0673	AAD	IEEE 802.11ax (20MHz, MCS2, 90pc dc)	WLAN	8.78	±9.6 °
0674	AAD	IEEE 802.11ax (20MHz, MCS3, 90pc dc)	WLAN	8.74	±9.6
0675	AAD	IEEE 802.11ax (20MHz, MCS4, 90pc dc)	WLAN	8.90	± 9.6 5
0676	AAD	IEEE 802.11ax (20MHz, MCS5, 90pc dc)	WLAN	8.77	±9.6
0677	AAD	IEEE 802.11ax (20MHz, MCS6, 90pc dc)	WLAN	8.73	± 9.6 9
0678	AAD	IEEE 802.11ax (20MHz, MCS7, 90pc dc)	WLAN	8.78	±9.6
0679	AAD	IEEE 802.11ax (20MHz, MCS8, 90pc dc)	WLAN	8.89	±9.6 5
0680	AAD	IEEE 802.11ax (20MHz, MCS9, 90pc dc)	WLAN	8.80	±9.6
0681	AAG	IEEE 802.11ax (20MHz, MCS10, 90pc dc)	WLAN	8.62	±9.6
0682	AAF	IEEE 802.11ax (20MHz, MCS11, 90pc dc)	WLAN	8.83	±9.6 9
0683	AAA	IEEE 802.11ax (20MHz, MCS0, 99pc dc)	WLAN	8.42	±9.6 9
0684	AAC	IEEE 802.11ax (20MHz, MCS1, 99pc dc)	WLAN	8.26	±9.6 5
0685	AAC	IEEE 802.11ax (20MHz, MCS2, 99pc dc)	WLAN	8.33	±9.6
0686	AAC	IEEE 802.11ax (20MHz, MCS3, 99pc dc)	WLAN	8.28	±9.6
0687	AAE	IEEE 802.11ax (20MHz, MCS4, 99pc dc)	WLAN	8.45	±9.6
8860	AAE	IEEE 802.11ax (20MHz, MCS5, 99pc dc)	WLAN	8.29	± 9.6 °
0689	AAD	IEEE 802.11ax (20MHz, MCS6, 99pc dc)	WLAN	8.55	±9.6
0690	AAE	IEEE 802.11ax (20MHz, MCS7, 99pc dc)	WLAN	8.29	± 9.6
0691	AAB	IEEE 802.11ax (20MHz, MCS8, 99pc dc)	WLAN	8.25	± 9.6
0692	AAA	IEEE 802.11ax (20MHz, MCS9, 99pc dc)	WLAN	8.29	± 9.6
0693	AAA	IEEE 802.11ax (20MHz, MCS10, 99pc dc)	WLAN	8.25	±9.6
	AAA	IEEE 802.11ax (20MHz, MCS11, 99pc dc)	WLAN	8.57	± 9.6
0694					

Certificate No:Z22-60223

Page 17 of 22



Report No.: R2209A0844-S1





Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2117 E-mail: ettl@chinattl.com http://www.caiet.oc.cn

10696	AAA	IEEE 802.11ax (40MHz, MCS1, 90pc dc)	WLAN	8.91	±9.6 %
10697	AAA	IEEE 802.11ax (40MHz, MCS2, 90pc dc)	WLAN	8.61	19.6%
10698	AAA	IEEE 802.11ax (40MHz, MCS3, 90pc dc)	WLAN	8.89	19.6%
10699	AAA	IEEE 802.11ax (40MHz, MCS4, 90pc dc)	WLAN	8.82	196%
10700	AAA	IEEE 802.11ax (40MHz, MCS5, 90pc dc)	WLAN	8.73	± 9.8 %
10701	AAA	IEEE 802.11ax (40MHz, MCS6, 90pc dc)	WLAN	8.86	±9.8%
10702	AAA	IEEE 802.11ax (40MHz, MCS7, 90pc dc)	WLAN	8.70	± 9.6 %
10703	AAA	IEEE 802.11ax (40MHz, MCS8, 90pc dc)	WLAN	8.82	±96%
10704	AAA	IEEE 802.11ax (40MHz, MCS9, 90pc dc)	WLAN	8.56	± 9.6 %
10705	AAA	IEEE 802.11ax (40MHz, MCS10, 90pc dc)	WLAN	8.69	±9.6 %
10706	AAC	IEEE 802.11ax (40MHz, MCS11, 90pc dc)	WLAN	8.66	± 9.6 %
10707	AAC	IEEE 802.11ax (40MHz, MCS0, 99pc dc)	WLAN	8.32	± 9.6 %
10708	AAC	IEEE 802 11ax (40MHz, MCS1, 99pc dc)	WLAN	8.55	±9.6 %
10709	AAC	IEEE 802.11ax (40MHz, MCS2, 99pc dc)	WLAN	8.33	± 9.6 %
10710	AAC	IEEE 802 11ax (40MHz, MCS3, 99pc dc)	WLAN	8.29	± 9.6 %
10711	AAC	IEEE 802.11ax (40MHz, MCS4, 99pc dc)	WLAN	8.39	± 9.6 %
10712	AAC	IEEE 802.11ax (40MHz, MCS5, 99pc dc)	WLAN	8.67	± 9.6 %
10713	AAC	IEEE 802.11ax (40MHz, MCS6, 99pc dc)	WLAN	8.33	± 9.6 %
10714	AAC	IEEE 802.11ax (40MHz, MCS7, 99pc dc)	WLAN	8.26	± 9.6 %
10715	AAC	IEEE 802.11ax (40MHz, MCS8, 99pc dc)	WLAN	8.45	±9.6 %
10716	AAC	IEEE 802.11ax (40MHz, MCS9, 99pc dc)	WLAN	8.30	± 9.6 %
10717	AAC	IEEE 802.11ax (40MHz, MCS10, 99pc dc)	WLAN	8.48	± 9.6 %
10718	AAC	IEEE 802.11ax (40MHz, MCS11, 99pc dc)	WLAN	8.24	± 9.6 %
10719	AAC	IEEE 802.11ax (80MHz, MCS0, 90pc dc)	WLAN	8.81	± 9.6 %
10720	AAC	IEEE 802.11ax (80MHz, MCS1, 90pc dc)	WLAN	8.87	± 9.6 %
10721	AAC	IEEE 802.11ax (80MHz, MCS2, 90pc dc)	WLAN	8.76	± 9.6 %
10722	AAC	IEEE 802.11ax (80MHz, MCS3, 90pc dc)	WLAN	8.55	± 9.6 %
10723	AAC	IEEE 802.11ax (80MHz, MCS4, 90pc dc)	WLAN	8.70	± 9.6 %
10724	AAC	IEEE 802.11ax (80MHz, MCS5, 90pc dc)	WLAN	8.90	± 9.6 %
10725	AAC	IEEE 802.11ax (80MHz, MCS6, 90pc dc)	WLAN	8.74	±9.6 %
10726	AAC	IEEE 802.11ax (80MHz, MCS7, 90pc dc)	WLAN	8.72	± 9.6 %
10727	AAC	IEEE 802.11ax (80MHz, MCS8, 90pc dc)	WLAN	8.66	± 9.6 %
10728	AAC	IEEE 802.11ax (80MHz, MCS9, 90pc dc)	WLAN	8.65	± 9.6 %
10729	AAC	IEEE 802.11ax (80MHz, MCS10, 90pc dc)	WLAN	8.64	± 9.6 %
10730	AAC	IEEE 802.11ax (80MHz, MCS11, 90pc dc)	WLAN	8.67	± 9.6 %
10731	AAC	IEEE 802.11ax (80MHz, MCS0, 99pc dc)	WLAN	8.42	± 9.6 %
10732	AAC	IEEE 802.11ax (80MHz, MCS1, 99pc dc)	WLAN	8,46	± 9.6 %
10733	AAC	IEEE 802.11ax (80MHz, MCS2, 99pc dc)	WLAN	8.40	± 9.6 %
10734	AAC	IEEE 802.11ax (80MHz, MCS3, 99pc dc)	WLAN	8.25	± 9.6 %
10735	AAC	IEEE 802.11ax (80MHz, MCS4, 99pc dc)	WLAN	8.33	± 9.6 %
10736	AAC	IEEE 802.11ax (80MHz, MCS5, 99pc dc)	WLAN	8.27	± 9.6 %
10737	AAC	IEEE 802.11ax (80MHz, MCS6, 99pc dc)	WLAN	8.36	± 9.6 %
10738	AAC	IEEE 802.11ax (80MHz, MCS7, 99pc dc)	WLAN	8.42	± 9.6 %
10739	AAC	IEEE 802.11ax (80MHz, MCS8, 99pc dc)	WLAN	8.29	± 9.6 %
10740	AAC	IEEE 802.11ax (80MHz, MCS9, 99pc dc)	WLAN	8.48	± 9.6 %
10741	AAC	IEEE 802.11ax (80MHz, MCS10, 99pc dc)	WLAN	8.40	± 9.6 %
10742	AAC	IEEE 802.11ax (80MHz, MCS11, 99pc dc)	WLAN	8.43	± 9.6 %
10743	AAC	IEEE 802.11ax (160MHz, MCS0, 90pc dc)	WLAN	8.94	± 9.6 %
10744	AAC	IEEE 802.11ax (160MHz, MCS1, 90pc dc)	WLAN	9.16	± 9.6 %
10745	AAC	IEEE 802.11ax (160MHz, MCS2, 90pc dc)	WLAN	8.93	± 9.6 %
10746	AAC	IEEE 802.11ax (160MHz, MCS3, 90pc dc)	WLAN	9.11	±9.6%
10747	AAC	IEEE 802.11ax (160MHz, MCS4, 90pc dc)	WLAN	9.04	± 9.6 %
10748	AAC	IEEE 802.11ax (160MHz, MCS5, 90pc dc)	WLAN	8.93	±9.6%
10749	AAC	IEEE 802.11ax (160MHz, MCS6, 90pc dc)	WLAN	8.90	± 9.6 %
10750	AAC	IEEE 802.11ax (160MHz, MCS7, 90pc dc)	WLAN	8.79	± 9.6 %
10751	AAC	IEEE 802.11ax (160MHz, MCS8, 90pc dc)	WLAN	8.82	±9.6%
10752	AAC	IEEE 802.11ax (160MHz, MCS9, 90pc dc)	WLAN	8.81	±9.6 %
10753	AAC	IEEE 802.11ax (160MHz, MCS10, 90pc dc)	WLAN	9.00	±9.6 %
10754	AAC	IEEE 802.11ax (160MHz, MCS11, 90pc dc)	WLAN	8.94	±9.6%

Certificate No:Z22-60223

Page 18 of 22



a

S



Add: No.52 HuaYuanBel Road, Haidian District, Heijlag, 100191, Chim Tel: +86-10-62304633-2117 E-mail: ettl@ebinattl.com http://www.esiet.ac.en http://www.caicl.nc.cn

а

q

In Collaboration with

pe

CALIBRATION LABORATORY

10755	AAC	IEEE 802.11ax (160MHz, MCS0, 99pc dc)	WLAN	8.64	10.0%
10756	AAC	IEEE 802.11ax (160MHz, MCS1, 99pc dc)	WLAN	8.77	19.6 %
10757	AAC	IEEE 802.11ax (160MHz, MCS2, 99pc dc)	WLAN	8.77	± 9.6 %
0758	AAC	IEEE 802.11ax (160MHz, MCS3, 99pc dc)	WLAN	8.69	19.6 %
0759	AAC	IEEE 802.11ax (160MHz, MCS4, 99pc dc)	WLAN	8.59	19.6%
0760	AAC	IEEE 802.11ax (160MHz, MCS5, 99pc dc)	WLAN	8.49	± 9.0 %
0761	AAC	IEEE 802.11ex (160MHz, MCS6, 99pc dc)	WLAN	8.58	± 9.6 %
10762	AAC	IEEE 802.11ax (160MHz, MCS7, 99pc dc)	WLAN	8.49	± 9.6 %
0763	AAC	IEEE 802.11ax (160MHz, MCS8, 99pc dc)	WLAN	8.53	± 0.6 %
10764	AAC	IEEE 802.11ax (160MHz, MCS9, 99pc dc)	WLAN	B.54	198%
0765	AAC	IEEE 802.11ax (160MHz, MCS10, 99pc dc)	WLAN	8.54	± 0.6 %
0766	AAC	IEEE 802.11ax (160MHz, MCS11, 99pc dc)	WLAN	8.51	19.8%
0767	AAC	5G NR (CP-OFDM, 1 RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	7.99	± 0.6 %
0768	AAC	5G NR (CP-OFDM, 1 RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.01	±9.8%
0769	AAC	5G NR (CP-OFDM, 1 RB, 15 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.01	± 8.6 %
0770	AAC	5G NR (CP-OFDM, 1 RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.02	± 9.6 %
0771	AAC	5G NR (CP-OFDM, 1 RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.02	± 8.6 %
0772	AAC	5G NR (CP-OFDM, 1 RB, 30 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.23	±9.6%
0773	AAC	5G NR (CP-OFDM, 1 RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.03	±9.6 %
0774	AAC	5G NR (CP-OFDM, 1 RB, 50 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.02	± 9.6 %
0775	AAC	5G NR (CP-OFDM, 50% RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.31	± 0.6 %
0776	AAC	5G NR (CP-OFDM, 50% RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.30	± 9.6 %
0777	AAC	5G NR (CP-OFDM, 50% RB, 15 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.30	± 9.6 %
0778	AAC	5G NR (CP-OFDM, 50% RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.34	± 9.8 4
0779	AAC	5G NR (CP-OFDM, 50% RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.42	± 9.6 4
0780	AAC	5G NR (CP-OFDM, 50% RB, 30 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.38	± 9.6 4
0781	AAC	5G NR (CP-OFDM, 50% RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.38	±9.6
0782	AAC	5G NR (CP-OFDM, 50% RB, 50 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.43	± 9.6 %
0783	AAC	5G NR (CP-OFDM, 100% RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.31	±9.6
0784	AAC	5G NR (CP-OFDM, 100% RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.29	± 9.6 9
0785	AAC	5G NR (CP-OFDM, 100% RB, 15 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.40	± 9.6 °
10786	AAC	5G NR (CP-OFDM, 100% RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.35	±9.6
0787	AAC	5G NR (CP-OFDM, 100% RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.44	±9.6 %
0788	AAC	5G NR (CP-OFDM, 100% RB, 30 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.39	±9.8
10789	AAC	5G NR (CP-OFDM, 100% RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.37	± 9.6 4
0790	AAC	5G NR (CP-OFDM, 100% RB, 50 MHz, QPSK, 15 kHz)	5G NR FR1 TDD	8.39	±9.8
10791	AAC	5G NR (CP-OFDM, 1 RB, 5 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.83	± 9.6
10792	AAC	5G NR (CP-OFDM, 1 RB, 10 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.92	± 9.6
10793	AAC	5G NR (CP-OFDM, 1 RB, 15 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.95	± 9.6
10794	AAC	5G NR (CP-OFDM, 1 RB, 20 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.82	± 9.6
10795	AAC	5G NR (CP-OFDM, 1 RB, 25 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.84	±9.6
10796	AAC	5G NR (CP-OFDM, 1 RB, 30 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.82	± 9.6
0797	AAC	5G NR (CP-OFDM, 1 RB, 40 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.01	± 9.6
0798	AAC	5G NR (CP-OFDM, 1 RB, 50 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.89	± 9.6
0799	AAC	5G NR (CP-OFDM, 1 RB, 60 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.93	± 9.8
10801	AAC	5G NR (CP-OFDM, 1 RB, 80 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.89	± 9.6
10802	AAC	5G NR (CP-OFDM, 1 RB, 90 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.87	±9.6
10803	AAE	5G NR (CP-OFDM, 1 RB, 100 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	7.93	±9.6
0805	AAD	5G NR (CP-OFDM, 50% RB, 10 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	6.34	± 9.6
0806	AAD	5G NR (CP-OFDM, 50% RB, 15 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.37	± 9.6
0809	AAD	5G NR (CP-OFDM, 50% RB, 30 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.34	± 9.6
0810	AAD	5G NR (CP-OFDM, 50% RB, 40 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.34	± 9.6
0812	AAD	5G NR (CP-OFDM, 50% RB, 60 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.35	19.6
0817	AAD	5G NR (CP-OFDM, 100% RB, 5 MHz, QPSK, 30 kHz)	5G NR FR1 TDD		
0818	AAD	5G NR (CP-OFDM, 100% RB, 10 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.35	±9.6
0819	AAD	5G NR (CP-OFDM, 100% RB, 15 MHz, QPSK, 30 kHz)		8.34	± 9.6
0820	AAD	5G NR (CP-OFDM, 100% RB, 20 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.33	± 9.6
0821	AAC	5G NR (CP-OFDM, 100% RB, 25 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.30	± 9.6
0822	AAD	5G NR (CP-OFDM, 100% RB, 30 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8,41	±9.6
OULL	1 AVID	So the for on DW, to with to, so with an one so with	5G NR FR1 TDD	8.41	± 9.6

Certificate No:Z22-60223

Page 19 of 22



Report No.: R2209A0844-S1

In Collaboration with s р e а



Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2117 E-mail: ctul@chinattl.com http://www.caict.ac.en

10823	AAC	5G NR (CP-OFDM, 100% RB, 40 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.36	1 19.6 9
10824	AAD	5G NR (CP-OFDM, 100% RB, 50 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.39	± 9.6 %
10825	AAD	5G NR (CP-OFDM, 100% RB, 60 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.41	±969
10827	AAD	5G NR (CP-OFDM, 100% RB, 60 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.42	± 9.6 3
0828	AAE	5G NR (CP-OFDM, 100% RB, 90 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.43	19.8 9
0829	AAD	5G NR (CP-OFDM, 100% RB, 100 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	8.40	19.69
10830	AAD	5G NR (CP-OFDM, 1 RB, 10 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.63	19.6 9
0831	AAD	5G NR (CP-OFDM, 1 RB, 15 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.73	±9.6 %
10832	AAD	5G NR (CP-OFDM, 1 R8, 20 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.74	±989
10833	AAD	5G NR (CP-OFDM, 1 RB, 25 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.70	±969
10834	AAD	5G NR (CP-OFDM, 1 RB, 30 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.75	±983
10835	AAD	5G NR (CP-OFDM, 1 RB, 40 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.70	±9.6 %
10836	AAE	5G NR (CP-OFDM, 1 RB, 50 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.68	±9.89
10837	AAD	5G NR (CP-OFDM, 1 RB, 60 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.68	± 9.6 9
10839	AAD	5G NR (CP-OFDM, 1 RB, 80 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.70	1989
10840	AAD	5G NR (CP-OFDM, 1 RB, 90 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.67	± 9.6 %
10841	AAD	5G NR (CP-OFDM, 1 RB, 100 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	7.71	29.69
10843	AAD	5G NR (CP-OFDM, 50% RB, 15 MHz, OPSK, 60 kHz)	5G NR FR1 TDD	8.49	±9.6 %
10844	AAD	5G NR (CP-OFDM, 50% RB, 20 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.34	1969
10846	AAD	5G NR (CP-OFDM, 50% RB, 30 MHz, QPSK, 60 KHz)	5G NR FR1 TDD	8.41	1969
10854	AAD	5G NR (CP-OFDM, 100% RB, 10 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.34	± 9.6 9
10855	AAD	5G NR (CP-OFDM, 100% RB, 15 MHz, QPSK, 60 KHz)	5G NR FR1 TDD	8.36	19.6 9
10856	AAD	5G NR (CP-OFDM, 100% RB, 20 MHz, 0PSK, 60 kHz)	5G NR FR1 TDD	8.37	±983
10857	AAD	5G NR (CP-OFDM, 100% RB, 25 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.35	±9.6 9
10858	AAD	5G NR (CP-OFDM, 100% RB, 30 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.36	±983
10859	AAD	5G NR (CP-OFDM, 100% RB, 40 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.34	± 9.6 %
10859	AAD	5G NR (CP-OFDM, 100% RB, 50 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.41	±9.69
10861	AAD	5G NR (CP-OFDM, 100% RB, 60 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.40	±9.6 %
10863	AAD	5G NR (CP-OFDM, 100% RB, 80 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.41	+965
10864	AAE	5G NR (CP-OFDM, 100% RB, 90 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.37	±9.6 %
10865	AAD	5G NR (CP-OFDM, 100% RB, 100 MHz, QPSK, 60 kHz)	5G NR FR1 TDD	8.41	19.6 %
10866	AAD	5G NR (DFT-s-OFDM, 1 RB, 100 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	19.69
10868	AAD	5G NR (DFT-s-OFDM, 100% RB, 100 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.89	±969
10869	AAD	5G NR (DFT-S-OFDM, 1 RB, 100 MHz, QPSK, 120 kHz)	5G NR FR2 TDD	5.75	±9.6 %
10870	AAD	5G NR (DFT-s-OFDM, 100% RB, 100 MHz, QPSK, 120 kHz)	5G NR FR2 TOD	5.86	±9.6 %
10871	AAD	5G NR (DFT-s-OFDM, 1 RB, 100 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD	5.75	1969
10872	AAD	5G NR (DFT-s-OFDM, 100% RB, 100 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD	8.52	1969
10873	AAD	5G NR (DFT-s-OFDM, 1 RB, 100 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	6.61	19.69
10874	AAD	5G NR (DFT-s-OFDM, 100% RB, 100 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	6.65	±9.6 %
10875	AAD	5G NR (CP-OFDM, 1 RB, 100 MHz, QPSK, 120 kHz)	5G NR FR2 TDD	7.78	±969
10876	AAD	5G NR (CP-OFDM, 100% RB, 100 MHz, QPSK, 120 kHz)	5G NR FR2 TDD	8.39	+9.6 9
10877	AAD	5G NR (CP-OFDM, 1 RB, 100 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD	7,95	1969
10878	AAD	5G NR (CP-OFDM, 100% RB, 100 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD	8.41	±9.6 %
10879	AAD	5G NR (CP-OFDM, 1 RB, 100 MHz, 64QAM, 120 kHz)	5G NR FR2 TOD	8.12	1989
10880	AAD	5G NR (CP-OFOM, 100% RB, 100 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	8.38	1969
10880	AAD	5G NR (DFT-s-OFDM, 1 RB, 50 MHz, QPSK, 120 kHz)	5G NR FR2 TDD	5.75	1969
10882	AAD	5G NR (DFT-s-OFDM, 100% RB, 50 MHz, QPSK, 120 kHz)	5G NR FR2 TDD	5.96	196%
10883	AAD	5G NR (DFT-s-OFDM, 1 RB, 50 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD	6.57	19.6 %
10884	AAD	5G NR (DFT-s-OFDM, 100% RB, 50 MHz, 180AM, 120 kHz)	5G NR FR2 TDD	6.53	1969
10885	AAD	5G NR (DFT-s-OFDM, 100% RB, 50 MHz, 64QAM, 120 KHz)	5G NR FR2 TDD	6.61	1969
10886	AAD	5G NR (DFT-s-OFDM, 100% RB, 50 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	6.65	±9.6 %
	AAD	5G NR (CP-OFDM, 1 RB, 50 MHz, OPSK, 120 KHz)	5G NR FR2 TDD	7.78	19.69
10887	AAD	5G NR (CP-OFDM, 100% RB, 50 MHz, GPSK, 120 kHz)	5G NR FR2 TDD	8.35	1969
10888	AAD	5G NR (CP-OFDM, 100% RB, 50 MHz, 16QAM, 120 KHz)	5G NR FR2 TDD	8.02	1969
10889	-	5G NR (CP-OFDM, 100% RB, 50 MHz, 16QAM, 120 kHz)	5G NR FR2 TDD		19.69
10890	AAD			8.40	
10891	AAD	5G NR (CP-OFDM, 1 RB, 50 MHz, 64QAM, 120 kHz)	5G NR FR2 TDD	8.13	±9.6 %
10892	AAD	5G NR (CP-OFDM, 100% RB, 50 MHz, 640AM, 120 kHz)	5G NR FR2 TDD	8.41	±96%
10897	AAD	5G NR (DFT-s-OFDM, 1 RB, 5 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.66	± 9.6 %
10898	AAD	5G NR (DFT-s-OFDM, 1 RB, 10 MHz, OPSK, 30 kHz)	5G NR FR1 TDD	5,67	±9.6%

Certificate No:Z22-60223



In Collaboration with

CALIBRATION LABORATORY



Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2117 E-mail: ettl@chinattl.com http://www.eaiet.ac.en

10899	AAD	5G NR (DFT-s-OFDM, 1 RB, 15 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.67	± 9.6 %
10900	AAD	5G NR (DFT-s-OFDM, 1 RB, 20 MHz, OPSK, 30 kHz)	5G NR FR1 TDD	5.68	±9.6%
10901	AAD	5G NR (DFT-s-OFDM, 1 RB, 25 MHz, OPSK, 30 kHz)	5G NR FR1 TDD	5.68	± 9.6 %
10902	AAD	5G NR (DFT-s-OFDM, 1 RB, 30 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	±9.6 %
10903	AAD	5G NR (DFT-s-OFDM, 1 RB, 40 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	± 9.6 %
10904	AAD	5G NR (DFT-s-OFDM, 1 RB, 50 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	±9.6 %
10905	AAD	5G NR (DFT-s-OFDM, 1 RB, 60 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	± 9.6 %
10906	AAD	5G NR (DFT-s-OFDM, 1 RB, 80 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.68	±9.6 %
10907	AAD	5G NR (DFT-s-OFDM, 50% RB, 5 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.78	± 9.6 %
10908	AAD	5G NR (DFT-s-OFDM, 50% RB, 10 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.93	± 9.6 %
10909	AAD	5G NR (DFT-s-OFDM, 50% RB, 15 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.96	± 9.6 %
10910	AAD	5G NR (DFT-s-OFDM, 50% RB, 20 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.83	± 9,6 %
10911	AAD	5G NR (DFT-s-OFDM, 50% RB, 25 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.93	± 9.6 %
10912	AAD	5G NR (DFT-s-OFDM, 50% RB, 30 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.84	±9.6 %
10913	AAD	5G NR (DFT-s-OFDM, 50% RB, 40 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.84	± 9.6 %
10914	AAD	5G NR (DFT-s-OFDM, 50% RB, 50 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.85	± 9.6 %
10915	AAD	5G NR (DFT-s-OFDM, 50% RB, 60 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.83	± 9.6 %
10916	AAD	5G NR (DFT-s-OFDM, 50% RB, 80 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.87	± 9.6 %
10917	AAD	5G NR (DFT-s-OFDM, 50% RB, 100 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.94	± 9.6 %
10918	AAD	5G NR (DFT-s-OFDM, 100% RB, 5 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.86	± 9.6 %
10919	AAD	5G NR (DFT-s-OFDM, 100% RB, 10 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.86	19.6 %
10920	AAD	5G NR (DFT-s-OFDM, 100% RB, 15 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.87	± 9.6 %
10921	AAD	5G NR (DFT-s-OFDM, 100% RB, 20 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.84	± 9.6 %
10922	AAD	5G NR (DFT-s-OFDM, 100% RB, 25 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.82	± 9.6 %
10923	AAD	5G NR (DFT-s-OFDM, 100% RB, 30 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.84	± 9.6 %
10924	AAD	5G NR (DFT-s-OFDM, 100% RB, 40 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.84	± 9.6 %
10925	AAD	5G NR (DFT-s-OFDM, 100% RB, 50 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.95	±9.6 %
10926	AAD	5G NR (DFT-s-OFDM, 100% RB, 60 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5,84	± 9.6 %
10927	AAD	5G NR (DFT-s-QFDM, 100% RB, 80 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	5.94	±9.6 %
10928	AAD	5G NR (DFT-s-OFDM, 1 RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.52	± 9.6 %
10929	AAD	5G NR (DFT-s-OFDM, 1 RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.52	± 9.6 %
10930	AAD	5G NR (DFT-s-OFDM, 1 RB, 15 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.52	± 9.6 %
10931	AAD	5G NR (DFT-s-OFDM, 1 RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.51	± 9.6 %
10932	AAB	5G NR (DFT-s-OFDM, 1 RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.51	±9.6 %
10933	AAA	5G NR (DFT-s-OFDM, 1 RB, 30 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.51	± 9.6 %
10934	AAA	5G NR (DFT-s-OFDM, 1 RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.51	± 9.6 %
10935	AAA	5G NR (DFT-s-OFDM, 1 R8, 50 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.51	±9.6 %
10936	AAC	5G NR (DFT-s-OFDM, 50% RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.90	± 9.6 %
10937	AAB	5G NR (DFT-s-OFDM, 50% RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.77	±9.6%
10938	AAB	5G NR (DFT-s-OFDM, 50% RB, 15 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.90	± 9.6 9
10939	AAB	5G NR (DFT-s-OFDM, 50% RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.82	±9.6%
10940	AAB	5G NR (DFT-s-OFDM, 50% RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.89	± 9.6 %
10941	AAB	5G NR (DFT-s-OFDM, 50% RB, 30 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.83	± 9.6 %
10942	AAB	5G NR (DFT-s-OFDM, 50% RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.85	19.6%
10943	AAB	5G NR (DFT-s-OFDM, 50% RB, 50 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.95	± 9.6 %
10944	AAB	5G NR (DFT-s-OFDM, 100% RB, 5 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.81	± 9.6 %
10945	AAB	5G NR (DFT-s-OFDM, 100% RB, 10 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.85	19.6%
10946	AAC	5G NR (DFT-s-OFDM, 100% RB, 15 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.83	±9.6%
10947	AAB	5G NR (DFT-s-OFDM, 100% RB, 20 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.87	±9.6%
10948	AAB	5G NR (DFT-s-OFDM, 100% RB, 25 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.94	±96%
10949	AAB	5G NR (DFT-s-OFDM, 100% RB, 30 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.87	±9.69
10950	AAB	5G NR (DFT-s-OFDM, 100% RB, 40 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.94	±9.6 %
10951	AAB	5G NR (DFT-s-OFDM, 100% RB, 50 MHz, QPSK, 15 kHz)	5G NR FR1 FDD	5.92	± 9.6 9
10952	AAB	5G NR DL (CP-OFDM, TM 3.1, 5 MHz, 64-QAM, 15 kHz)	5G NR FR1 FDD	8.25	± 9.6 9
10953	AAB	5G NR DL (CP-OFDM, TM 3.1, 10 MHz, 64-QAM, 15 kHz)	5G NR FR1 FDD	8.15	± 9.6 %
10954	AAB	5G NR DL (CP-OFDM, TM 3.1, 15 MHz, 64-QAM, 15 kHz)	5G NR FR1 FDD	8.23	±9.6 %
10955	AAB	5G NR DL (CP-OFDM, TM 3.1, 20 MHz, 64-QAM, 15 kHz)	5G NR FR1 FDD	8.42	± 9.6 9
10956	AAB	5G NR DL (CP-OFDM, TM 3.1, 5 MHz, 64-QAM, 30 kHz)	5G NR FR1 FDD	8.14	± 9.6 %
10957	AAC	5G NR DL (CP-OFDM, TM 3.1, 10 MHz, 64-QAM, 30 kHz)	5G NR FR1 FDD	8.31	19.6 2

Certificate No:Z22-60223

Page 21 of 22

TA Technology (Shanghai) Co., Ltd.TA-MB-06-003SPage 56 of 79This report shall not be reproduced except in full, without the written approval of TA Technology (Shanghai) Co., Ltd.



.

S

In Collaboration with

CAICT

Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2117 E-mail: ettl@chinattl.com http://www.caict.ac.en

a

a

pe

CALIBRATION LABORATORY

10958	AAB	5G NR DL (CP-OFDM, TM 3.1, 15 MHz, 64-QAM, 30 kHz)	5G NR FR1 FDD	8.61	± 9.6 %
10959	AAB	5G NR DL (CP-OFDM, TM 3.1, 20 MHz, 64-QAM, 30 kHz)	5G NR FR1 FDD	8.33	±9.6%
10960	AAB	5G NR DL (CP-OFDM, TM 3.1, 5 MHz, 64-QAM, 15 kHz)	5G NR FR1 TDD	9.32	±9.6%
10961	AAB	5G NR DL (CP-OFDM, TM 3.1, 10 MHz, 64-QAM, 15 kHz)	5G NR FR1 TDD	9.36	±96%
10962	AAB	5G NR DL (CP-OFDM, TM 3.1, 15 MHz, 64-QAM, 15 kHz)	5G NR FR1 TDD	9.40	±9.6 %
10963	AAB	5G NR DL (CP-OFDM, TM 3.1, 20 MHz, 64-QAM, 15 kHz)	5G NR FR1 TDD	9.55	±9.6 %
10964	AAB	5G NR DL (CP-OFDM, TM 3.1, 5 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.29	± 9.6 %
10965	AAB	5G NR DL (CP-OFDM, TM 3.1, 10 MHz, 64-OAM, 30 kHz)	5G NR FR1 TDD	9.37	±9.6 %
10966	AAB	5G NR DL (CP-OFDM, TM 3.1, 15 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.55	± 9.6 %
10967	AAB	5G NR DL (CP-OFDM, TM 3.1, 20 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.42	± 9.6 %
10968	AAB	5G NR DL (CP-OFDM, TM 3.1, 100 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.49	±9.6%
10972	AAB	5G NR (CP-OFDM, 1 RB, 20 MHz, OPSK, 15 kHz)	5G NR FR1 TDD	11.59	±9.6%
10973	AAB	5G NR (DFT-s-OFDM, 1 RB, 100 MHz, QPSK, 30 kHz)	5G NR FR1 TDD	9.06	±9.6%
10974	AAB	5G NR (CP-OFDM, 100% RB, 100 MHz, 256-QAM, 30 kHz)	5G NR FR1 TDD	10.28	±9.6%
10978	AAA	ULLABOR	ULLA	1.16	±9.6%
10979	AAA	ULLA HDR4	ULLA	8.58	±9.6%
10980	AAA	ULLA HDR8	ULLA	10.32	±9.6%
10981	AAA	ULLA HDRp4	ULLA	3.19	±9.6%
10982	AAA	ULLA HDRp8	ULLA	3.43	±9.6%
10983	AAA	5G NR DL (CP-OFDM, TM 3 1, 40 MHz, 64-QAM, 15 kHz)	5G NR FR1 TDD	9.31	±9.6%
10984	AAA	5G NR DL (CP-OFDM, TM 3.1, 50 MHz, 64-QAM, 15 kHz)	5G NR FR1 TDD	9.42	19.6%
10985	AAA	5G NR DL (CP-OFDM, TM 3 1, 40 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.54	± 9.6 %
10986	AAA	5G NR DL (CP-OFDM, TM 3.1, 50 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.50	19.6%
10987	AAA	5G NR DL (CP-OFDM, TM 3.1, 60 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.53	±9.6%
10988	AAA	5G NR DL (CP-OFDM, TM 3.1, 70 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.38	± 9.6 %
95901	AAA	5G NR DL (CP-OFDM, TM 3.1, 80 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.33	±9.6 %
10990	AAA	5G NR DL (CP-OFDM, TM 3.1, 90 MHz, 64-QAM, 30 kHz)	5G NR FR1 TDD	9.52	± 9.6 %

\* Uncertainty is determined using the max, deviation from linear response applying rectangular distribution and is expressed for the

square of the field value.

Certificate No:Z22-60223

Page 22 of 22

🛟 eurofins


# ANNEX E: D2450V2 Dipole Calibration Certificate

Tel: +86-10-623046 E-mail: cttl@chinatt Client TA(S		86-10-62304633-2504	CNAS L0570
	insom mupar		
	hanghai)		20-60298
CALIBRATION CE	State Services	E	
Object	D2450	/2 - SN: 786	
Calibration Procedure(s)			
valioration Procedure(s)	FF-Z11		
	Calibra	tion Procedures for dipole validation kits	
Calibration date:	August	27, 2020	
pages and are part of the ce	ertificate.		are given on the following
	conducted in	the closed laboratory facility: environmen or calibration)	
All calibrations have been numidity<70%. Calibration Equipment used	conducted in	or calibration)	
All calibrations have been numidity<70%. Calibration Equipment used	n conducted in		t temperature(22±3)°C and
All calibrations have been numidity<70%. Calibration Equipment used Primary Standards	I (M&TE critical fo	or calibration) Cal Date(Calibrated by, Certificate No.)	t temperature(22±3)°C and Scheduled Calibration
All calibrations have been numidity<70%. Calibration Equipment used Primary Standards Power Meter NRP2	I Conducted in (M&TE critical for ID # 106276 101369	Cal Date(Calibrated by, Certificate No.) 12-May-20 (CTTL, No.J20X02965) 12-May-20 (CTTL, No.J20X02965) 30-Jan-20(SPEAG,No.EX3-3617_Jan20)	t temperature(22±3)°C and Scheduled Calibration May-21 May-21 Jan-21
All calibrations have been numidity<70%. Calibration Equipment used Primary Standards Power Meter NRP2 Power sensor NRP6A	I Conducted in (M&TE critical for ID # 106276 101369	Cal Date(Calibrated by, Certificate No.) 12-May-20 (CTTL, No.J20X02965) 12-May-20 (CTTL, No.J20X02965)	t temperature(22±3)°C and Scheduled Calibration May-21 May-21
All calibrations have been numidity<70%. Calibration Equipment used Primary Standards Power Meter NRP2 Power sensor NRP6A Reference Probe EX3DV4 DAE4	I conducted in (M&TE critical for ID # 106276 101369 SN 3617	Cal Date(Calibrated by, Certificate No.) 12-May-20 (CTTL, No.J20X02965) 12-May-20 (CTTL, No.J20X02965) 30-Jan-20(SPEAG,No.EX3-3617_Jan20) 10-Feb-20(CTTL-SPEAG,No.Z20-60017)	t temperature(22±3)°C and Scheduled Calibration May-21 May-21 Jan-21
All calibrations have been numidity<70%. Calibration Equipment used Primary Standards Power Meter NRP2 Power sensor NRP6A Reference Probe EX3DV4	ID # 106276 101369 SN 3617 SN 771	Cal Date(Calibrated by, Certificate No.) 12-May-20 (CTTL, No.J20X02965) 12-May-20 (CTTL, No.J20X02965) 30-Jan-20(SPEAG,No.EX3-3617_Jan20)	t temperature(22±3) <sup>v</sup> C and Scheduled Calibration May-21 May-21 Jan-21 Feb-21
All calibrations have been numidity<70%. Calibration Equipment used Primary Standards Power Meter NRP2 Power sensor NRP6A Reference Probe EX3DV4 DAE4 Secondary Standards	I conducted in (M&TE critical for 106276 101369 SN 3617 SN 771 ID # ID # MY49071430	Cal Date(Calibrated by, Certificate No.) 12-May-20 (CTTL, No.J20X02965) 12-May-20 (CTTL, No.J20X02965) 30-Jan-20(SPEAG,No.EX3-3617_Jan20) 10-Feb-20(CTTL-SPEAG,No.Z20-60017) Cal Date(Calibrated by, Certificate No.)	t temperature(22±3)°C and Scheduled Calibration May-21 May-21 Jan-21 Feb-21 Scheduled Calibration
All calibrations have been numidity<70%. Calibration Equipment used Primary Standards Power Meter NRP2 Power sensor NRP6A Reference Probe EX3DV4 DAE4 Secondary Standards Signal Generator E4438C	I conducted in (M&TE critical for 106276 101369 SN 3617 SN 771 ID # ID # MY49071430	Cal Date(Calibrated by, Certificate No.) 12-May-20 (CTTL, No.J20X02965) 12-May-20 (CTTL, No.J20X02965) 30-Jan-20(SPEAG,No.EX3-3617_Jan20) 10-Feb-20(CTTL-SPEAG,No.Z20-60017) Cal Date(Calibrated by, Certificate No.) 25-Feb-20 (CTTL, No.J20X00516)	t temperature(22±3)°C and Scheduled Calibration May-21 May-21 Jan-21 Feb-21 Scheduled Calibration Feb-21
All calibrations have been numidity<70%. Calibration Equipment used Primary Standards Power Meter NRP2 Power sensor NRP6A Reference Probe EX3DV4 DAE4 Secondary Standards Signal Generator E4438C NetworkAnalyzer E5071C	ID# 106276 101369 SN 3617 SN 771 ID# MY49071430 MY46107873	Cal Date(Calibrated by, Certificate No.) 12-May-20 (CTTL, No.J20X02965) 12-May-20 (CTTL, No.J20X02965) 30-Jan-20(SPEAG,No.EX3-3617_Jan20) 10-Feb-20(CTTL-SPEAG,No.Z20-60017) Cal Date(Calibrated by, Certificate No.) 25-Feb-20 (CTTL, No.J20X00516) 10-Feb-20 (CTTL, No.J20X00515)	t temperature(22±3)°C and Scheduled Calibration May-21 May-21 Jan-21 Feb-21 Scheduled Calibration Feb-21 Feb-21
All calibrations have been numidity<70%. Calibration Equipment used Primary Standards Power Meter NRP2 Power sensor NRP6A Reference Probe EX3DV4 DAE4 Secondary Standards Signal Generator E4438C	a conducted in (M&TE critical for 10 # 106276 101369 SN 3617 SN 771 ID # MY49071430 MY46107873 Name	Cal Date(Calibrated by, Certificate No.) 12-May-20 (CTTL, No.J20X02965) 12-May-20 (CTTL, No.J20X02965) 30-Jan-20(SPEAG,No.EX3-3617_Jan20) 10-Feb-20(CTTL-SPEAG,No.Z20-60017) Cal Date(Calibrated by, Certificate No.) 25-Feb-20 (CTTL, No.J20X00516) 10-Feb-20 (CTTL, No.J20X00515) Function	t temperature(22±3)°C and Scheduled Calibration May-21 May-21 Jan-21 Feb-21 Scheduled Calibration Feb-21 Feb-21

Certificate No: Z20-60298

Page 1 of 8





Tel: +86-10-62304633-2079 E-mail: cttl@chinattl.com

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China Fax: +86-10-62304633-2504 http://www.chinattl.cn

Glossary:	
TSL	tissue simulating liquid
ConvF	sensitivity in TSL / NORMx,y,z
N/A	not applicable or not measured

#### Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", September 2013
- b) IEC 62209-1, "Measurement procedure for assessment of specific absorption rate of human exposure to radio frequency fields from hand-held and body-mounted wireless communication devices- Part 1: Device used next to the ear (Frequency range of 300MHz to 6GHz)", July 2016
- c) IEC 62209-2, "Procedure to measure the Specific Absorption Rate (SAR) For wireless communication devices used in close proximity to the human body (frequency range of 30MHz to 6GHz)", March 2010
- d) KDB865664, SAR Measurement Requirements for 100 MHz to 6 GHz

#### Additional Documentation:

e) DASY4/5 System Handbook

#### Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z20-60298

Page 2 of 8





E-mail: cttl@chinattl.com

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504 http://www.chinattl.cn

#### **Measurement Conditions**

DASY system configuration, as far as not given on page 1.

DASY Version	DASY52	V52.10.4
Extrapolation	Advanced Extrapolation	
Phantom	Triple Flat Phantom 5.1C	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	2450 MHz ± 1 MHz	

g

#### Head TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	39.2	1.80 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	39.5 ± 6 %	1.79 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	-	

#### SAR result with Head TSL

SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	13.0 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	52.3 W/kg ± 18.8 % (k=2)
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	Condition	
SAR measured	250 mW input power	5.99 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	24.0 W/kg ± 18.7 % (k=2)

Body TSL parameters The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	52.7	1.95 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	52.1 ± 6 %	1.94 mho/m ± 6 %
Body TSL temperature change during test	<1.0 °C	1 Sec.	

#### SAR result with Body TSL

SAR averaged over 1 cm <sup>3</sup> (1 g) of Body TSL	Condition	
SAR measured	250 mW input power	13.1 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	52.4 W/kg ± 18.8 % (k=2)
SAR averaged over 10 cm <sup>3</sup> (10 g) of Body TSL	Condition	
SAR measured	250 mW input power	6.08 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	24.3 W/kg ± 18.7 % (k=2)

Certificate No: Z20-60298

Page 3 of 8





http://www.chinattl.cn Appendix (Additional assessments outside the scope of CNAS L0570)

#### Antenna Parameters with Head TSL

E-mail: cttl@chinattl.com

Impedance, transformed to feed point	54.5Ω+ 1.44 jΩ
Return Loss	- 26.9dB

#### Antenna Parameters with Body TSL

Impedance, transformed to feed point	50.9Ω+ 5.09 jΩ	
Return Loss	- 25.8dB	

#### General Antenna Parameters and Design

Electrical Delay (one direction)	1.018 ns

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.


#### Additional EUT Data

Manufactured by	SPEAG

Certificate No: Z20-60298

Page 4 of 8






0 dB = 22.0 W/kg = 13.42 dBW/kg

Certificate No: Z20-60298

Page 5 of 8






E-mail: cttl@chinattl.com

Fax: +86-10-62304633-2504 http://www.chinattl.cn


Impedance Measurement Plot for Head TSL




Certificate No: Z20-60298

Page 6 of 8





dy=5mm, dz=5mm Reference Value = 102.9 V/m; Power Drift = -0.03 dB Peak SAR (extrapolated) = 26.9 W/kg SAR(1 g) = 13.1 W/kg; SAR(10 g) = 6.08 W/kg Smallest distance from peaks to all points 3 dB below = 8.5 mm Ratio of SAR at M2 to SAR at M1 = 49.9% Maximum value of SAR (measured) = 21.8 W/kg



0 dB = 21.8 W/kg = 13.38 dBW/kg

Certificate No: Z20-60298


Page 7 of 8





Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504 E-mail: ettl@chinattl.com http://www.chinattl.cn

Impedance Measurement Plot for Body TSL



Certificate No: Z20-60298

Page 8 of 8



Report No.: R2209A0844-S1

# ANNEX F: D5GHzV2 Dipole Calibration Certificate

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Schweizerischer Kalibrierdienst Service suisse d'étalonnage Servizio svizzero di taratura

Swiss Calibration Service

Accreditation No.: SCS 0108

S

С

S

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client Auden

Certificate No: D5GHzV2-1203\_Dec22

Dbject	D5GHzV2 - SN:1	203	
Calibration procedure(s)	QA CAL-22.v7 Calibration Proce	dure for SAR Validation Sources	between 3-10 GHz
Calibration date;	December 09, 20	22	
		onal standards, which realize the physical unit	
		y facility: environment temperature $(22 \pm 3)^{\circ}$ C	
Calibration Equipment used (M&TE			
Primary Standards	LID #	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NRP	SN: 104778	04-Apr-22 (No. 217-03525/03524)	Apr-23
Power sensor NRP-Z91	SN: 103244	04-Apr-22 (No. 217-03524)	Apr-23
Power sensor NRP-Z91	SN: 103245	04-Apr-22 (No. 217-03525)	Apr-23
	SN: BH9394 (20k)	04-Apr-22 (No. 217-03527)	Apr-23
Heterence 20 dB Attenuator	Contraction of the second s		Apr-23
	SN: 310982 / 06327	04-ADF-22 (NO. 21/-03528)	
Type-N mismatch combination	SN: 310982 / 06327 SN: 3503	04-Apr-22 (No. 217-03528) 08-Mar-22 (No. EX3-3503 Mar22)	
Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4		04-Apr-22 (No. 217-03528) 08-Mar-22 (No. EX3-3503_Mar22) 31-Aug-22 (No. DAE4-601_Aug22)	Mar-23 Aug-23
Type-N mismatch combination Reference Probe EX3DV4	SN: 3503	08-Mar-22 (No. EX3-3503_Mar22)	Mar-23
Type-N mismatch combination Reference Probe EX3DV4 DAE4	SN: 3503 SN: 601	08-Mar-22 (No. EX3-3503_Mar22) 31-Aug-22 (No. DAE4-601_Aug22)	Mar-23 Aug-23
Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards	SN: 3503 SN: 601 ID #	08-Mar-22 (No. EX3-3503_Mar22) 31-Aug-22 (No. DAE4-601_Aug22) Check Date (in house)	Mar-23 Aug-23 Scheduled Check
Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B	SN: 3503 SN: 601 ID # SN: GB39512475	08-Mar-22 (No. EX3-3503_Mar22) 31-Aug-22 (No. DAE4-601_Aug22) Check Date (in house) 30-Oct-14 (in house check Oct-22)	Mar-23 Aug-23 Scheduled Check In house check: Oct-24
Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A	SN: 3503 SN: 601 ID # SN: GB39512475 SN: US37292783	08-Mar-22 (No, EX3-3503_Mar22) 31-Aug-22 (No. DAE4-601_Aug22) Check Date (in house) 30-Oct-14 (in house check Oct-22) 07-Oct-15 (in house check Oct-22)	Mar-23 Aug-23 Scheduled Check In house check: Oct-24 In house check: Oct-24
Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A	SN: 3503 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: MY41093315	08-Mar-22 (No, EX3-3503_Mar22) 31-Aug-22 (No, DAE4-601_Aug22) Check Date (in house) 30-Oct-14 (in house check Oct-22) 07-Oct-15 (in house check Oct-22) 07-Oct-15 (in house check Oct-22)	Mar-23 Aug-23 Scheduled Check In house check: Oct-24 In house check: Oct-24 In house check: Oct-24
Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06	SN: 3503 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: MY41093315 SN: 100972	08-Mar-22 (No, EX3-3503_Mar22) 31-Aug-22 (No, DAE4-601_Aug22) Check Date (in house) 30-Oct-14 (in house check Oct-22) 07-Oct-15 (in house check Oct-22) 07-Oct-15 (in house check Oct-22) 15-Jun-15 (in house check Oct-22)	Mar-23 Aug-23 Scheduled Check In house check: Oct-24 In house check: Oct-24 In house check: Oct-24 In house check: Oct-24 In house check: Oct-24 Signature
Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06	SN: 3503 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: MY41093315 SN: 100972 SN: US41080477	08-Mar-22 (No, EX3-3503_Mar22) 31-Aug-22 (No, DAE4-601_Aug22) Check Date (in house) 30-Oct-14 (in house check Oct-22) 07-Oct-15 (in house check Oct-22) 07-Oct-15 (in house check Oct-22) 15-Jun-15 (in house check Oct-22) 31-Mar-14 (in house check Oct-22)	Mar-23 Aug-23 Scheduled Check In house check: Oct-24 In house check: Oct-24 In house check: Oct-24 In house check: Oct-24 In house check: Oct-24 Signature
Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent E8358A	SN: 3503 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: MY41093315 SN: 100972 SN: US41080477 Name	08-Mar-22 (No, EX3-3503_Mar22) 31-Aug-22 (No, DAE4-601_Aug22) Check Date (in house) 30-Oct-14 (in house check Oct-22) 07-Oct-15 (in house check Oct-22) 07-Oct-15 (in house check Oct-22) 15-Jun-15 (in house check Oct-22) 31-Mar-14 (in house check Oct-22) Function	Mar-23 Aug-23 Scheduled Check In house check: Oct-24 In house check: Oct-24 In house check: Oct-24 In house check: Oct-24 In house check: Oct-24



Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





S Schweizerischer Kalibrierdienst C Service suisse d'étalonnage

Servizio svizzero di taratura

S Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

#### Glossary:

TSL	tissue simulating liquid
ConvF	sensitivity in TSL / NORM x,y,z
N/A	not applicable or not measured

#### Calibration is Performed According to the Following Standards:

- a) IEC/IEEE 62209-1528, "Measurement Procedure For The Assessment Of Specific Absorption Rate Of Human Exposure To Radio Frequency Fields From Hand-Held And Body-Worn Wireless Communication Devices - Part 1528: Human Models, Instrumentation And Procedures (Frequency Range of 4 MHz to 10 GHz)", October 2020.
- b) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

#### Additional Documentation:

c) DASY System Handbook

#### Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end
  of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The source is mounted in a touch configuration below the center marking of the flat phantom.
- Return Loss: This parameter is measured with the source positioned under the liquid filled phantom (as described in the measurement condition clause). The Return Loss ensures low reflected power. No uncertainty required.
- · SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: D5GHzV2-1203\_Dec22

Page 2 of 9



#### **Measurement Conditions**

DASY system configuration, as far as not given on page 1.

DASY Version	DASY52	V52.10.4
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom V5.0	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy = 4.0 mm, dz = 1.4 mm	Graded Ratio = 1.4 (Z direction)
Frequency	5250 MHz ± 1 MHz 5600 MHz ± 1 MHz 5750 MHz ± 1 MHz 5850 MHz ± 1 MHz	

## Head TSL parameters at 5250 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.9	4.71 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	36.4 ± 6 %	4.61 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		

#### SAR result with Head TSL at 5250 MHz

SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	7.76 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	77.7 W/kg ± 19.9 % (k=2)
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	condition	
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL SAR measured	condition 100 mW input power	2.24 W/kg

Certificate No: D5GHzV2-1203\_Dec22

Page 3 of 9

#### Head TSL parameters at 5600 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.5	5.07 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.9 ± 6 %	4.98 mho/m ± 6 %
Head TSL temperature change during test	< 0,5 °C	1.000	

#### SAR result with Head TSL at 5600 MHz

SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	8.02 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	80.3 W/kg ± 19.9 % (k=2)
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	condition	
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL SAR measured	condition 100 mW input power	2.30 W/kg

#### Head TSL parameters at 5750 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.4	5.22 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.7 ± 6 %	5.14 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		

#### SAR result with Head TSL at 5750 MHz

SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	7.68 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	76.8 W/kg ± 19.9 % (k=2)

SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	condition	
SAR measured	100 mW input power	2.19 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	22.0 W/kg ± 19.5 % (k=2)

Certificate No: D5GHzV2-1203\_Dec22

Page 4 of 9

#### Head TSL parameters at 5850 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.2	5.32 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.6 ± 6 %	5.24 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		

### SAR result with Head TSL at 5850 MHz

SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	7.90 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	79.1 W/kg ± 19.9 % (k=2)
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	condition	
SAR averaged over 10 cm <sup>a</sup> (10 g) of Head TSL SAR measured	condition 100 mW input power	2.25 W/kg

Certificate No: D5GHzV2-1203\_Dec22

Page 5 of 9

## Appendix (Additional assessments outside the scope of SCS 0108)

#### Antenna Parameters with Head TSL at 5250 MHz

Impedance, transformed to feed point	48.5 Ω - 3.2 jΩ	
Return Loss	- 29.0 dB	

#### Antenna Parameters with Head TSL at 5600 MHz

Impedance, transformed to feed point	51.7 Ω + 2.6 <b>]</b> Ω
Return Loss	- 30.4 dB

#### Antenna Parameters with Head TSL at 5750 MHz

Impedance, transformed to feed point	53.6 Ω + 4.3 ϳΩ
Return Loss	- 25.3 dB

#### Antenna Parameters with Head TSL at 5850 MHz

Impedance, transformed to feed point	52.4 Ω + 4.2 jΩ
Return Loss	- 26.5 dB

#### **General Antenna Parameters and Design**

Electrical Delay (one direction)	1.191 ns	
----------------------------------	----------	--

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

#### Additional EUT Data

Manufactured by	SPEAG
-----------------	-------

Certificate No: D5GHzV2-1203\_Dec22

Page 6 of 9



#### DASY5 Validation Report for Head TSL

Date: 09.12.2022

Test Laboratory: SPEAG, Zurich, Switzerland

#### DUT: Dipole D5GHzV2; Type: D5GHzV2; Serial: D5GHzV2 - SN:1203

Communication System; UID 0 - CW; Frequency; 5250 MHz, Frequency; 5600 MHz, Frequency; 5750 MHz, Frequency; 5850 MHz Medium parameters used: f = 5250 MHz;  $\sigma = 4.61$  S/m;  $\epsilon_r = 36.4$ ;  $\rho = 1000$  kg/m<sup>3</sup>. Medium parameters used: f = 5600 MHz;  $\sigma = 4.98$  S/m;  $\epsilon_r = 35.9$ ;  $\rho = 1000$  kg/m<sup>3</sup>. Medium parameters used: f = 5750 MHz;  $\sigma = 5.14$  S/m;  $\epsilon_r = 35.7$ ;  $\rho = 1000$  kg/m<sup>3</sup>. Medium parameters used: f = 5850 MHz;  $\sigma = 5.24$  S/m;  $\epsilon_r = 35.6$ ;  $\rho = 1000$  kg/m<sup>3</sup>. Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

**DASY52** Configuration:

- Probe: EX3DV4 SN3503; ConvF(5.5, 5.5, 5.5) @ 5250 MHz, ConvF(5.1, 5.1, 5.1) @ 5600 MHz, ConvF(5.08, 5.08, 5.08) @ 5750 MHz, ConvF(4.99, 4.99, 4.99) @ 5850 MHz; Calibrated: 08.03.2022
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 31.08.2022
- Phantom: Flat Phantom 5.0 (front); Type: QD 000 P50 AA; Serial: 1001
- DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501)

Dipole Calibration for Head Tissue/Pin=100mW, dist=10mm, f=5250 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 73.31 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 26.3 W/kg SAR(1 g) = 7.76 W/kg; SAR(10 g) = 2.24 W/kg Smallest distance from peaks to all points 3 dB below = 7.5 mm Ratio of SAR at M2 to SAR at M1 = 70.6% Maximum value of SAR (measured) = 17.6 W/kg

Dipole Calibration for Head Tissue/Pin=100mW, dist=10mm, f=5600 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 72.76 V/m; Power Drift = -0.07 dB Peak SAR (extrapolated) = 29.4 W/kg SAR(1 g) = 8.02 W/kg; SAR(10 g) = 2.30 W/kg Smallest distance from peaks to all points 3 dB below = 7.5 mm Ratio of SAR at M2 to SAR at M1 = 67.9% Maximum value of SAR (measured) = 18.9 W/kg

Certificate No: D5GHzV2-1203\_Dec22

Page 7 of 9



## Dipole Calibration for Head Tissue/Pin=100mW, dist=10mm, f=5750 MHz/Zoom Scan,

dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 70.15 V/m; Power Drift = -0.02 dB Peak SAR (extrapolated) = 29.7 W/kg SAR(1 g) = 7.68 W/kg; SAR(10 g) = 2.19 W/kg Smallest distance from peaks to all points 3 dB below = 7.5 mm Ratio of SAR at M2 to SAR at M1 = 66.2% Maximum value of SAR (measured) = 18.3 W/kg

#### Dipole Calibration for Head Tissue/Pin=100mW, dist=10mm, f=5850 MHz/Zoom Scan,

dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 70.55 V/m; Power Drift = -0.03 dB Peak SAR (extrapolated) = 31.6 W/kg SAR(1 g) = 7.90 W/kg; SAR(10 g) = 2.25 W/kg Smallest distance from peaks to all points 3 dB below = 7.5 mm Ratio of SAR at M2 to SAR at M1 = 65.2% Maximum value of SAR (measured) = 19.2 W/kg




0 dB = 19.2 W/kg = 12.84 dBW/kg

Certificate No: D5GHzV2-1203\_Dec22

Page 8 of 9



## Impedance Measurement Plot for Head TSL



Certificate No: D5GHzV2-1203\_Dec22

Page 9 of 9


# ANNEX G: DAE4 Calibration Certificate (SN: 1692)

E-mail: emf@caict.ac.en Client : TA	http://www.caict.ac. (Shanghal)		No: Z22-60518
CALIBRATION	CERTIFICA	TE	
Object	DAE4	- SN: 1692	
Calibration Procedure(s)	FF-Z1	1-002-01 ration Procedure for the Data Acquis x)	ition Electronics
Calibration date:	Nover	mber 18, 2022	
pages and are part of the	e certificate.	d the uncertainties with confidence probative the closed laboratory facility: enviror	
pages and are part of the	e certificate. een conducted in sed (M&TE critical	the closed laboratory facility: enviror	
pages and are part of the All calibrations have be humidity<70%. Calibration Equipment u	e certificate. een conducted in sed (M&TE critical	the closed laboratory facility: enviror for calibration)	nment temperature(22±3)°C ar
pages and are part of the All calibrations have be humidity<70%. Calibration Equipment u Primary Standards	e certificate. een conducted in sed (M&TE critical ID # Ci	the closed laboratory facility: enviror for calibration) al Date(Calibrated by, Certificate No.)	nment temperature(22±3)°C an Scheduled Calibration Jun-23
pages and are part of the All calibrations have be humidity<70%. Calibration Equipment u Primary Standards Process Calibrator 753	e certificate. een conducted in sed (M&TE critical ID # C: 1971018	the closed laboratory facility: enviror for calibration) al Date(Calibrated by, Certificate No.) 14-Jun-22 (CTTL, No.J22X04180)	nment temperature(22±3)°C an
pages and are part of the All calibrations have be humidity<70%. Calibration Equipment u Primary Standards	e certificate. een conducted in sed (M&TE critical ID # C: 1971018 Name	the closed laboratory facility: enviror for calibration) al Date(Calibrated by, Certificate No.) 14-Jun-22 (CTTL, No.J22X04180) Function	nment temperature(22±3)°C an Scheduled Calibration Jun-23

Certificate No: Z22-60518

Page 1 of 3









Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2117 E-mail: emf@caict.ac.cn http://www.caict.ac.cn

Glossary: DAE Connector angle


data acquisition electronics information used in DASY system to align probe sensor X to the robot coordinate system.

## Methods Applied and Interpretation of Parameters:

- DC Voltage Measurement: Calibration Factor assessed for use in DASY system by comparison with a calibrated instrument traceable to national standards. The figure given corresponds to the full scale range of the voltmeter in the respective range.
- Connector angle: The angle of the connector is assessed measuring the angle mechanically by a tool inserted. Uncertainty is not required.
- The report provide only calibration results for DAE, it does not contain other performance test results.

Certificate No: Z22-60518









Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2117 E-mail: emf@caict.ac.cn http://www.caict.ac.cn

#### DC Voltage Measurement

 A/D - Converter Resolution nominal

 High Range:
 1LSB =
 6.1μV ,
 full range =
 -100...+300 mV

 Low Range:
 1LSB =
 61nV ,
 full range =
 -1.....+3mV

 DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

Calibration Factors	x	Y	Z
High Range	404.475±0.15% (k=2)	404.550 ± 0.15% (k=2)	404.407 ± 0.15% (k=2)
Low Range	3.95073 ± 0.7% (k=2)	$4.00277 \pm 0.7\%$ (k=2)	3.97904 ± 0.7% (k=2)

#### **Connector Angle**

Connector Angle to be used in DASY system	335°±1°
New York, and the second s	

Certificate No: Z22-60518

Page 3 of 3



# **ANNEX H: The EUT Appearance**

The EUT Appearance are submitted separately.



# **ANNEX I: Test Setup Photos**

The Test Setup Photos are submitted separately.