



Date: 18 May 2022

DECLARATION OF EQUIVALENCE

This document declares that the following designated products are equivalent to the unit under test **39428**.

For USA:

| Model Name / Product Code | Marketing Name |
|----------------------------------|-----------------------|
| 39491 | AIRCURVE 11 ASV USA |
| 39492 | AIRCURVE 11 S USA |
| 39493 | AIRCURVE 11 ST USA |
| 39494 | AIRCURVE 11 VAUTO USA |

For Canada:

| Model Name / Product Code | Marketing Name |
|----------------------------------|-----------------------|
| 39495 | AIRCURVE 11 VAUTO CAN |
| 39496 | AIRCURVE 11 S CAN |
| 39497 | AIRCURVE 11 ST CAN |
| 39498 | AIRCURVE 11 ASV CAN |

For USA & Canada:


| Model Name / Product Code | Marketing Name |
|----------------------------------|------------------------------|
| 51400 | AIRCURVE 11 ST NORTH AMERICA |

All the above stated products have the same hardware, cellular firmware and Bluetooth firmware.

Different models are derived for marketing purposes. The differences between S, ST, VAUTO and ASV models are the therapy modes and customer/region centric features supported by the device. These settings are configured at manufacturing via a configuration file and is independent of the cellular, Bluetooth and FG software.

In summary, the differences between the device models mentioned herein are:

- Therapy modes
- Data settings
- UI language
- Name plaque
- Data connection parameters.



ResMed

*Changing lives
with every breath*

Applicant:

Company Name: ResMed Pty Ltd
Address: 1 Elizabeth Macarthur Drive,
Bella Vista NSW 2153
Australia

By,

Christopher Jenkins

Christopher Jenkins

Title: Associate Manager – Systems Engineering
Company: ResMed Pty Ltd
Telephone: +61 2 8884 1517
e-mail: Christopher.jenkins@resmed.com.au