



ResMed

Date: May 18, 2022

Re: Declaration about validity of test results from testing/certification of the previous model

We,

ResMed Pty Ltd,
1 Elizabeth Macarthur Dr,
Bella Vista NSW 2153,
Australia

Related to product:

Type of Equipment: Bilevel device with integrated cellular and Bluetooth connectivity
Brand name: ResMed
Model name: 39491, 39492, 39493, 39494, 39495, 39496, 39497, 39498, 51400
Commercial name: AIRCURVE 11 VAUTO, AIRCURVE 11 S, AIRCURVE 11 ST, AIRCURVE 11 ASV
FCC ID: 2ACHL-AIR11M1B
IC: 9103A-AIR11M1B

To whom it may concern,

We declare that the model 39001 with the FCC ID: 2ACHL-AIR11M1 and IC: 9103A-AIR11M1 and the model 39428 support the same module BLE and the same cellular module as it can be confirmed with the electrical diagram and list of materials sent with along this declaration.

Then, the results from conducted test according to FCC and ISED standards remain applicable, valid and representative for the certification of the host device, AirCurve 11

Test report Number	Test Standards	Date of Issue
70437RRF.002A1	USA FCC Part 15.247, 15.209 CANADA RSS-247, RSS-Gen	2022-05-06
70437RRF.004	USA FCC Part 22 & Part 90 CANADA RSS-132, RSS-Gen	2022-05-06

Sincerely,

Christopher Jenkins

By: Christopher Jenkins
Title: Associate Manager – Systems Engineering
Company: ResMed Pty Ltd
Telephone: +61 2 8884 1517
e-mail: Christopher.Jenkins@resmed.com.au