

Bickland Industrial Park, Falmouth, Cornwall, TR11 4TA, UK
Tel: 01326 372753 [International +44 1326 372 753]
Fax: 01326 378783 [International +44 1326 378 783]
e-mail: sales@research-instruments.com
www.research-instruments.com

Date: 10 May 2018

Subject: Request for permissive change for FCC ID 2ACOO-670807 & IC ID 12146A-670807

Research Instruments is submitting an application to request permissive change on the product Embryology Heated Plate. We confirm the following changes:

- The antenna matching circuit is no longer configurable, it is now fixed by design.
- The circuit no longer measures SWR, it now measures forward power.
- Reed relays used to switch between antennas Meder CRR05-1A have been replaced by Coto 9092-05-1x.
- The transmitter module Tagsys Medio P233 has not been modified.

Sincerely,

By:

Will Thallien
ELECTRONICS DESIGN MANAGER
(Signature/Title)

WILL THALLIENS
(Print name)