



America

5945 Cabot Parkway, Suite 100
Alpharetta, GA 30005

February 21, 2018

TÜV SÜD TCB
5945 Cabot Parkway, Suite 100
Alpharetta, GA 30005

Re: FCC ID: 2AD2U-MI001

To Whom It May Concern:

The following application is submitted on behalf of our client, DJO, LLC, for evaluation of their models 11-4000-0-06000 and 11-4001-0-06000 for certification under FCC Part 15.247.

The manufacturer has declared that the models 11-4000-0-06000 and 11-4001-0-06000 mechanically/electrically are equivalent products and only differ in orientation based on the leg for which the X4 knee brace is installed. Therefore, only the 11-4000-0-06000 was evaluated.

The 11-4000-0-06000 knee brace is a body worn portable device integrating a Bluetooth Low energy transceiver.

The 11-4000-0-06000 was tested in full to the requirements of the aforementioned rules and was found to be in compliance.

Sincerely,

A handwritten signature in black ink that reads "R. Sam Wismer".

Sam Wismer
Southeast Regional Manager - EMC
TÜV SÜD America Inc.