



# Compliance Testing, LLC

Previously Flom Test Lab

EMI, EMC, RF Testing Experts Since 1963

toll-free: (866) 311-3268

fax: (480) 926-3598

<http://www.ComplianceTesting.com>

[info@ComplianceTesting.com](mailto:info@ComplianceTesting.com)

## Test Report

Prepared for: Emotiv Inc

Model: EPOC Flex

Description: EEG Headset

Serial Number: FLEX01

FCC ID: 2ADIH-FLEX01

To

FCC Part 1.1310

Date of Issue: September 13, 2018

On the behalf of the applicant:

Emotiv Research Pty Ltd  
490 Post Street  
Suite 824  
San Francisco, CA 94102

Attention of:

Dr. Geoffrey Mackellar, Chief Technical Officer  
Ph: 6 141 453-4561  
Email: [geoff@emotiv.co](mailto:geoff@emotiv.co)

Prepared By  
Compliance Testing, LLC  
1724 S. Nevada Way  
Mesa, AZ 85204  
(480) 926-3100 phone / (480) 926-3598 fax  
[www.compliancetesting.com](http://www.compliancetesting.com)  
Project No: p1860040

Poona Saber  
Project Test Engineer

This report may not be reproduced, except in full, without written permission from Compliance Testing  
All results contained herein relate only to the sample tested



### Test Report Revision History

Revision	Date	Revised By	Reason for Revision
1.0	August 15, 2018	Poona Saber	Original Document
2.0	September 11, 2018	Poona Saber	Revised SAR exclusion calculation

## ILAC / A2LA

Compliance Testing, LLC, has been accredited in accordance with the recognized International Standard ISO/IEC 17025:2005. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer joint ISO-ILAC-IAF Communiqué dated January 2009)

The tests results contained within this test report all fall within our scope of accreditation, unless below

Please refer to <http://www.compliancetesting.com/labscope.html> for current scope of accreditation.

Testing Certificate Number: **2152.01**



**FCC Site Reg. #349717**

**IC Site Reg. #2044A-2**

**Non-accredited tests contained in this report:**

**N/A**

### **EUT Description**

**Model:** Epoc Flex

**Description:** wireless EEG system

**Firmware:** NA

**Software:** emotive pro

**Serial Number:** NA

**Additional Information:** The EPOC Flex is a 32-channel flexible EEG system consist of a controller, universal USB receiver, cap with electrode sensors on. The controller has a Bluetooth low energy radio with mini chip antenna. Frequency of operation is 2400-2483.5 MHz and unit is a battery operated with recharging capability through mini USB cable

Below is Calculation for SAR exclusion per KDB 447498.

1) The 1-g and 10-g SAR test exclusion thresholds for 100 MHz to 6 GHz at *test separation distances*  $\leq 50$  mm are determined by:

$[(\text{max. power of channel, including tune-up tolerance, mW})/(\text{min. test separation distance, mm})] \cdot [\sqrt{f(\text{GHz})}] \leq 3.0$  for 1-g SAR and  $\leq 7.5$  for 10-g extremity SAR,25 where

- f(GHz) is the RF channel transmit frequency in GHz
- Power and distance are rounded to the nearest mW and mm before calculation<sup>26</sup>
- The result is rounded to one decimal place for comparison
- 3.0 and 7.5 are referred to as the numeric thresholds in the step 2 below

The test exclusions are applicable only when the minimum *test separation distance* is  $\leq 50$  mm and for transmission frequencies between 100 MHz and 6 GHz. When the minimum *test separation distance* is  $< 5$  mm, a distance of 5 mm according to 5) in section 4.1 in KDB 447498 is applied to determine SAR test exclusion.

The Maximum output power per manufacturer declaration is 4 dBm.

$$\frac{2.5 \text{ mW}}{5 \text{ mm}} \cdot \sqrt{2.402 \text{ GHz}} = 1.54$$

Since the above number is below 1-g SAR limit this device is excluded for SAR measurements

END OF TEST REPORT