



EMC Test Data

Client:	iRhythm	PR Number:	PR163245
Model:	Zio MCT ECG Monitor	T-Log Number:	TL163245-RANA-Patch
Contact:	Jason Van Vliet	Project Manager:	Deepa Shetty
Standard:	FCC Part 15.247, RSS-247	Project Engineer:	David Bare
		Class:	N/A

SAR Exclusion

Specific Details

Objective: Evaluate the RF Exposure requirements per FCC 1.1310, 2.1093.

Date of Test: 3/3/2023

Test Engineer: David Bare

Summary of Results

Device complies with SAR exclusion as the output power is less than 1mW