



October 9, 2015

To Whom It May Concern:

We, iRhythm Technologies, Inc., located at 650 Townsend Street, Suite 380, San Francisco, CA 94103 U.S.A., confirm that the results from conducted tests acc. to the relevant FCC rule part and as applied to the radio modules:

Telit CE910-DUAL, FCC-ID: RI7CE910-DUAL (CDMA dual band)
LS Research, LLC, FCC-ID: TFB-BT2 (Bluetooth LE)

which are incorporated into our host equipment

ZIO SR ECG GATEWAY, FCC-ID: 2AFBP-SR15G

and as reported in the associate module test reports

#NK-12-R-098, 07-27-2012, Nemko Korea Co., Ltd.
#313136A, 20-08-2013, LS Research, LLC, Wisconsin

accurately represent the test results under the new conditions when the identified radio module is integrated in the identified host equipment.

By integration of the said radio module into the identified host equipment no changes were made to it or by external adaptation circuitry that impacts the validity of the results of the original module's test reports.

Sincerely,



10/9/2015

Richard Laguna
Director of Regulatory Affairs