

RE051-18-100576-1-A Ed. 2

This report cancels and replaces the test report N°RE051-18-100576-1 Edition 1

SAR TEST REPORT	
According to the standards: FCC 47 CFR § 2.1093	
Equipments under test: Wireless intercom APPI-COM LR	
FCC ID: 2AG7HBSAPC2U01 IC: 21024-BSAPC2U01	
Company: APPI-TECHNOLOGY SAS	

DISTRIBUTION: Mr. COULON

Company: APPI-TECHNOLOGY SAS

Number of pages: 56 with 3 annexes

Ed.	Date	Modified page(s)	Technical Verification Name-function Visa	Quality Approval Name-function Visa
2	June 14, 2018	Identified by a vertical line	Gilles Hyaumet – SAR and EMC lab manager	Eric Coeuret –Technical director

Duplication of this test report is only permitted for an integral photographic facsimile. It includes the number of pages referenced here above.

This document is the result of testing a specimen or a sample of the product submitted. It does not imply an assessment of the conformity of the whole production of the tested sample.







EQUIPMENTS UNDER TEST:	Wireless intercom
Model:	APPI-COM LR
Serial number (S/N):	Not identified
Hardware version:	R10D (according to the applicant)
Software version:	2.3 (according to the applicant)
MANUFACTURER:	APPI-TECHNOLOGY SAS
APPLICANT:	
Company:	APPI-TECHNOLOGY SAS
Address:	443 AVENUE JEAN PROUVÉ LE MINOTAURE 30900 NIMES FRANCE
Contact person:	Mr. J. CHAIX
Person(s) present(s) during the test:	-
DATE(S) OF TEST(S):	April 23 and 24, 2018
TEST SITE:	EMITECH ANGERS – MRA : FR00009 - IC : 4452A PA de Lanserre 21 rue de la Fuye 49610 Juigné-sur-Loire France
TESTS OPERATORS:	Antoine BOUCLEY and Gilles HYAUMET
TESTS TUTOR	Gilles HYAUMET



SUMMARY

1.		4
2.	REFERENCE DOCUMENTS	4
3.	PRESENTATION OF EQUIPMENT FOR TESTING PURPOSES	5
4.	TESTS RESULTS SUMMARY	11
5.	ENVIRONNEMENTAL CONDITIONS	12
6.	EQUIPMENT USED FOR THE TESTING	12
7.	MEASUREMENT RESULTS	14
8.	GRAPHICAL REPRESENTATIONS OF THE COARSE SCAN	16
9.	PHOTOGRAPHS OF THE EQUIPMENT UNDER TEST	25
<i>10</i> .	MEASUREMENT UNCERTAINTY	28
<i>11</i> .	SPATIAL PEAK SAR EVALUATION	30
<i>12</i> .	EQUIPMENT TEST CONDITIONS	31
<i>13</i> .	MEASUREMENT SYSTEM DESCRIPTION	36
14.	LIQUID MEASUREMENT: TEST CONDITIONS & RESULTS	37
15.	SYSTEM VALIDATION: TEST CONDITIONS & RESULTS	38
ANI	NEX 1: DAE3 CALIBRATION CERTIFICATE	40
ANI	NEX 2: E-FIELD PROBE CALIBRATION CERTIFICATE	45
ANI	NEX 3: D900V2 CALIBRATION CERTIFICATE	56



1. INTRODUCTION

In this test report, Specific Absorption Rate (SAR) measurements for the wireless communication device APPI-COM LR are presented.

EUT, referenced BS-APC2U-0x/Bx, has been already tested according to the reference measurement with the integrated and external antennas.

The tested antennas and the results are listed in the test report N° RE051-16-103329-1-A Ed. 0 delivered by EMITECH

According to the applicant, 2 new references of external antenna have been added, with the increasing of the duration of the RF transmitting and FM deviation value.

2. REFERENCE DOCUMENTS

The reference documents referred throughout this report are listed below.

These reference documents are applicable to the entire report, although extensions (version, date and amendment) are not repeated.

Reference	Document title	Date
FCC 47 CFR	§ 2.1093 Radiofrequency radiation exposure evaluation: portable devices	2018
KDB 447498	D01 General RF Exposure Guidance v06	2015
KDB 447498	D03 Supplement C Cross-Reference v01	2014
IEEE Std C95.1a	IEEE Standard for Safety Levels with Respect to Human Exposure to Radio Frequency Electromagnetic Fields, 3 kHz to 300 GHz. Amendment 1: Specifes Ceiling Limits for Induced and Contact Current, Clarifes Distinctions between Localized Exposure and Spatial Peak Power Density	2010
IEEE Std 1528	IEEE Recommended Practice for Determining the Peak Spatial-Average Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Technique	2013



3. PRESENTATION OF EQUIPMENT FOR TESTING PURPOSES

The wireless communication device uses the frequency band 915MHz and is shown in Fig. 1a to Fig. 1c.

Antenna	Half wave	Half wave	
Model	DELTA 6A	463 MODEL	
Manufacturer	SIRETTA	NEARSON	
Gain	4.05 dBi	2.0 dBi	
See Fig.	1.b	1.c	
APPi-COM type	BS-APC2U-X4	BS-APC2U-X5	

EUT is intended to be inserted into a protective cover (rubber cover or an attachment clip allowing separation distance of 5mm to the body) to be used with the neck strap so that its rear side is facing the body of the user as defined by the applicant (using based on the test report N° RE051-16-103329-1-A Ed. 0)

As specified in the user guide (Version 2.0), APPI-COM must be used with the provided Clip or Rubber cover, allowing a minimum distance of 5 mm with the body. (Using based on the test report N° RE051-16-103329-1-A Ed. 0)

For the present measurements, Clip or Rubber cover were not used.

Note: the external antennas are non-detachable as defined by the applicant for the end-users (using based on the test report N° RE051-16-103329-1-A Ed. 0). For this testing purpose, the 2 external antennas were removable and tested on the same sample.



Front and rear sides

Internal views





BiBa

E Cool

=





Fig. 1a: APPI-COM without antenna





Front and rear sides

Fig. 1b: APPI-COM with SIRETTA antenna





Front and rear sides

Fig. 1c: APPI-COM with NEARSON antenna



Specifications of the tested antenna, provided by the applicant:

Antenna	Half wave	Half wave
Model	s463xxx-915	DELTA 6A
Manufac	NEARSON	SIRETTA
Gain	2	4,05
see fig		



4. TESTS RESULTS SUMMARY

Object	Respected Standard ? ≤ 1.6W/kg in 1 g ≤ 2W/kg in 10 g		Maximum SAR value measured) in 1g / in 10g	
	Yes	No	(W/kg	
EUT with SIRETTA antenna SAR measurements at 0 mm from the body	х		0.606 / 0.372	
EUT with SIRETTA antenna SAR measurements at 5 mm from the body	х		0.421 / 0.279	
EUT with NEARSON antenna SAR measurements at 0 mm from the body	x		0.502 / 0.308	

Conclusion:

The sample APPI-COM, fitted with the referenced antennas submitted to test when worn at 0 mm from the body, is in conformity with the FCC Guidelines, for general population/uncontrolled exposure, according to FCC 47 CFR § 2.1093

To declare, or not, the compliance with the specifications, it was not explicitly taken into account of uncertainty associated with the result(s).



5. ENVIRONNEMENTAL CONDITIONS

Condition	Measured Value
Liquid Temperature	See Graphical Representations and §14
Ambient Temperature	See Graphical Representations and §14

6. EQUIPMENT USED FOR THE TESTING

Plateform ID Emitech N°		Category	Brand	Туре	Last	Next
					calibration	calibration
1 – RF	8523	Spectrum analyzer	Rohde-Schwarz	FSEM30	Aug 25, 2016	Aug 25, 2018
monitoring					-	-
	8524	Spectrum analyzer	Hewlett-Packard	8591EM	Apr 27, 2016	Apr 27, 2018
2 DASY4	7321	Software	Speag	DASY4	-	-
	9485	E-Field Probe	Speag	ES3DV3	Aug. 23, 16	Aug. 23, 18
	7192	Data acquisition	Speag	DAE3	Aug. 9, 16	Aug. 9, 18
	7194	Dipole 900MHz	Speag	D900V2	Oct 17, 16	Oct 17, 18
	7324	Phantom	Speag	ELI4	-	-
3 Liquid	-	Software	Hewlett-Packard	HP85070C	-	-
Measure						
	1402	Network analyzer	Hewlett-Packard	HP8753C	Jan 03, 18	Jan 03, 19
	9777	S-Parameter	Hewlett-Packard	HP85047A	Jan 02, 18	Jan 02, 19
7218 Dielectric probe		Hewlett-Packard	HP85070C	-	-	
6980 Thermometer		Testo	922	Mar. 11, 16	Mar.11, 18 (1)	
4 System 7215 Signal generator		Marconi	2024	Feb. 11, 16	Feb.11, 18 (2)	
Validation						
	7209	Amplifier	Mini-circuits	ZHL42	-	-
	7212	Power meter	Rohde-Schwarz	NRVS	Mar 02, 17	Mar 02, 19
	7211	Probe power meter	Rohde-Schwarz	NRV-Z31	Mar 02, 17	Mar 02, 19
	7035	Power meter	Rohde-Schwarz	NRVD	Jan 09, 17	Jan 09, 19
7034 Probe power meter		Rohde-Schwarz	NRV-Z1	Jan 09, 17	Jan 09, 19	
	7210	Coupler	MEB	RK100	Feb. 11, 16	Feb.11, 18 (2)
	7315	Attenuator	Radiall	R411810124	Feb. 11, 16	Feb. 11, 18 (2)
				R411806124		
9161 50 ohms load		Diconex	17-0193	Feb. 11, 16	Feb. 11, 18 (2)	

(1) validity date extended to 2 months

(2) under the derogation N° EQSDER000S5100085. Calibrated on Apr. 26, 2018.



Measuring equipment specifications:

ES3DV3 Isotropic E-Field Probe Overview:

ELI4 Elliptical phantom Overview:

DimensionsLength 600 mm \pm 5 mm and width 400 mm \pm 5 mmShape Ellipse2.0 mm with a tolerance of \pm 0.2 mmLiquid depth150 mm

System Validation Kit Overview:

 Construction
 Symmetrical dipole with I/4 balun

 Enables measurement of feedpoint impedance with NWA

 Matched for use near flat phantoms filled with head/body simulating solutions

 Return Loss
 > 20 dB at specified validation position

 Dimensions
 D2450V3 dipole length: 51.8 mm

Mounting Device for Transmitters Overview:

ConstructionEnables the position of the mounted transmitter device according to IEEE and IEC specificationsMaterialPOM



7. <u>MEASUREMENT RESULTS</u>

The wireless communication device antennas and batteries were provided by the manufacturer. The battery is fully charged before each measurement.

The output power and frequency are controlled using a spectrum analyzer. The wireless communication device was set by the applicant to transmit at its highest output peak power level.

The wireless communication was placed against the flat phantom at 0 mm with its rear side. The SAR test was performed for each antenna at the Low, Middle and High frequencies in order to find the worst case.

Then the measurements were repeated at 5 mm as specified in the user guide and in the KDB 447498 for the devices that are designed to operate on the body of users using a conservative minimum test separation distance \leq 5 mm

Measurement results (SAR values averaged over a mass of 1g):

		SAR 1g			
Configuration	Test Position	Low channel	Middle channel	High channel	Meas. date
		903 MHz	915 MHz	927 MHz	
SIRETTA antenna 4.05dBi Short half wave	Rear side at 0mm	0.606	0.574	0.507	April 23, 2018
NEARSON antenna 2dBi Short half wave	Rear side at 0mm	0.397	0.489	0.502	April 24, 2018
SIRETTA antenna 4.05dBi Short half wave	Rear side at 5mm	0.421	0.343	0.251	April 24, 2018

Measurement results (SAR values averaged over a mass of 10g):

		SAR 10	SAR 10g (W/kg) - Limit = 2W/kg			
Configuration	Test Position	Low channel 903 MHz	Middle channel 915 MHz	High channel 927 MHz	Meas. date	
SIRETTA antenna 4.05dBi Short half wave	Rear side at 0mm	0.372	0.351	0.309	April 23, 2018	
NEARSON antenna 2dBi Short half wave	Rear side at 0mm	0.247	0.303	0.308	April 24, 2018	
SIRETTA antenna 4.05dBi Short half wave	Rear side at 5mm	0.279	0.225	0.166	April 24, 2018	



Simultaneous transmission

The equipment under test is intended to operate simultaneously with 915MHz and Bluetooth.

As the present highest measured SAR value is significantly lower than the one listed in the test report N° RE051-16-103329-1-A Ed. 0, evaluation has not been done again for simultaneous RF transmissions.

The sum of 1g and 10g SAR of all simultaneously transmission mode (ISM 915 MHz and Bluetooth) is assessed within the SAR limit (\leq 1.6 W/kg – 1g, \leq 2 W/kg - 10g).



8. GRAPHICAL REPRESENTATIONS OF THE COARSE SCAN

The graphical representations of the coarse scan for the worst case are shown in Fig. 2 to Fig. 10.

DUT: APPI-COM_SIRETTA

Communication System: APPI-COM; Frequency: 903 MHz; Duty Cycle: 1:8.33 Medium parameters used: σ = 1.03001 mho/m, ϵ_r = 52.9798; ρ = 1000 kg/m³ Phantom section: Flat Section

Program Notes: Ambient temperature: 23.6 °C, Liquid temperature: 22.0 °C

DASY4 Configuration:

- Probe: ES3DV3 SN3303; ConvF(5.92, 5.92, 5.92)
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE3 Sn402
- Phantom: ELI 4.0; Type: QDOVA001BA; Serial: xxxx
- Measurement SW: DASY4, V4.5 Build 19; Postprocessing SW: SEMCAD, V1.8 Build 145

Position 0cm, Low channel/Area Scan (41x201x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.692 mW/g

Position 0cm, Low channel/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 20.4 V/m; Power Drift = 0.024 dB Peak SAR (extrapolated) = 0.978 W/kg SAR(1 g) = 0.606 mW/g; SAR(10 g) = 0.372 mW/g Maximum value of SAR (measured) = 0.724 mW/g



Fig. 2: SAR distribution for SIRRETA antenna, Low channel (903 MHz), rear side at 0 mm



Communication System: APPI-COM; Frequency: 915 MHz; Duty Cycle: 1:8.33 Medium parameters used: σ = 1.0401 mho/m, ϵ_r = 52.8573; ρ = 1000 kg/m³ Phantom section: Flat Section

Program Notes: Ambient temperature: 23.7 °C, Liquid temperature: 22 °C

DASY4 Configuration:

- Probe: ES3DV3 SN3303; ConvF(5.92, 5.92, 5.92)
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE3 Sn402
- Phantom: ELI 4.0; Type: QDOVA001BA; Serial: xxxx
- Measurement SW: DASY4, V4.5 Build 19; Postprocessing SW: SEMCAD, V1.8 Build 145

Position 0cm, Middle channel/Area Scan (41x201x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 0.654 mW/g

Position 0cm, Middle channel/Zoom Scan (7x7x7) (7x8x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 20.4 V/m; Power Drift = 0.052 dB Peak SAR (extrapolated) = 0.935 W/kg SAR(1 g) = 0.574 mW/g; SAR(10 g) = 0.351 mW/g Maximum value of SAR (measured) = 0.692 mW/g



Fig. 3: SAR distribution for SIRRETA antenna, Middle channel (915 MHz), rear side at 0 mm



Communication System: APPI-COM; Frequency: 927 MHz; Duty Cycle: 1:8.33 Medium parameters used: σ = 1.04865 mho/m, ϵ_r = 52.7221; ρ = 1000 kg/m³ Phantom section: Flat Section

Program Notes: Ambient temperature: 21.2 °C, Liquid temperature: 21.3 °C

DASY4 Configuration:

- Probe: ES3DV3 SN3303; ConvF(5.92, 5.92, 5.92)
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE3 Sn402
- Phantom: ELI 4.0; Type: QDOVA001BA; Serial: xxxx
- Measurement SW: DASY4, V4.5 Build 19; Postprocessing SW: SEMCAD, V1.8 Build 145

Position 0cm, High channel/Area Scan (41x201x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.583 mW/g

Position 0cm, High channel/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 19.8 V/m; Power Drift = 0.018 dB Peak SAR (extrapolated) = 0.839 W/kg SAR(1 g) = 0.507 mW/g; SAR(10 g) = 0.309 mW/g Maximum value of SAR (measured) = 0.619 mW/g







DUT: APPI-COM_NEARSON

Communication System: APPI-COM; Frequency: 903 MHz; Duty Cycle: 1:8.33 Medium parameters used: σ = 1.02157 mho/m, ϵ_r = 52.532; ρ = 1000 kg/m³ Phantom section: Flat Section

Program Notes: Ambient temperature: 23.0 °C, Liquid temperature: 21.1 °C

DASY4 Configuration:

- Probe: ES3DV3 SN3303; ConvF(5.92, 5.92, 5.92)
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE3 Sn402
- Phantom: ELI 4.0; Type: QDOVA001BA; Serial: xxxx
- Measurement SW: DASY4, V4.5 Build 19; Postprocessing SW: SEMCAD, V1.8 Build 145

Position 0cm, Low channel/Area Scan (41x201x1): Measurement grid: dx=15mm, dy=15mm

Maximum value of SAR (interpolated) = 0.455 mW/g

Position 0cm, Low channel/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 17.3 V/m; Power Drift = -0.026 dB Peak SAR (extrapolated) = 0.635 W/kg SAR(1 g) = 0.397 mW/g; SAR(10 g) = 0.247 mW/g Maximum value of SAR (measured) = 0.477 mW/g



Fig. 5: SAR distribution for NEARSON antenna, Low channel (903 MHz), rear side at 0 mm



DUT: APPI-COM_NEARSON

Communication System: APPI-COM; Frequency: 915 MHz; Duty Cycle: 1:8.33 Medium parameters used: σ = 1.03373 mho/m, ϵ_r = 52.4694; ρ = 1000 kg/m³ Phantom section: Flat Section

Program Notes: Ambient temperature: 22.1 °C, Liquid temperature: 21.1 °C

DASY4 Configuration:

- Probe: ES3DV3 SN3303; ConvF(5.92, 5.92, 5.92)
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE3 Sn402
- Phantom: ELI 4.0; Type: QDOVA001BA; Serial: xxxx
- Measurement SW: DASY4, V4.5 Build 19; Postprocessing SW: SEMCAD, V1.8 Build 145

Position 0cm, Middle channel/Area Scan (41x201x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 0.565 mW/g

Position 0cm, Middle channel/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid: dx=5mm,

dy=5mm, dz=5mm Reference Value = 18.6 V/m; Power Drift = 0.00 dB Peak SAR (extrapolated) = 0.780 W/kg SAR(1 g) = 0.489 mW/g; SAR(10 g) = 0.303 mW/g Maximum value of SAR (measured) = 0.588 mW/g







DUT: APPI-COM_NEARSON

Communication System: APPI-COM; Frequency: 927 MHz; Duty Cycle: 1:8.33 Medium parameters used: $\sigma = 1.04425$ mho/m, $\varepsilon_r = 52.4234$; $\rho = 1000$ kg/m³ Phantom section: Flat Section

Program Notes: Ambient temperature: 21.6 °C, Liquid temperature: 21.1 °C

DASY4 Configuration:

- Probe: ES3DV3 SN3303; ConvF(5.92, 5.92, 5.92); Calibrated: 8/21/2015
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE3 Sn402; Calibrated: 8/13/2015
- Phantom: ELI 4.0; Type: QDOVA001BA; Serial: xxxx
- Measurement SW: DASY4, V4.5 Build 19; Postprocessing SW: SEMCAD, V1.8 Build 145

Position 0cm, High channel/Area Scan (41x201x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 0.591 mW/g

Position 0cm, High channel/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 18.8 V/m; Power Drift = -0.189 dB Peak SAR (extrapolated) = 0.817 W/kg SAR(1 g) = 0.502 mW/g; SAR(10 g) = 0.308 mW/g Maximum value of SAR (measured) = 0.606 mW/g



Fig. 7: SAR distribution for NEARSON antenna, High channel (927 MHz), rear side at 0 mm



Communication System: APPI-COM; Frequency: 903 MHz; Duty Cycle: 1:8.33 Medium parameters used: σ = 1.02157 mho/m, ϵ_r = 52.532; ρ = 1000 kg/m³ Phantom section: Flat Section

Program Notes: Ambient temperature: 22.2 °C, Liquid temperature: 21.3 °C

DASY4 Configuration:

- Probe: ES3DV3 SN3303; ConvF(5.92, 5.92, 5.92)
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE3 Sn402
- Phantom: ELI 4.0; Type: QDOVA001BA; Serial: xxxx
- Measurement SW: DASY4, V4.5 Build 19; Postprocessing SW: SEMCAD, V1.8 Build 145

Position 5mm, Low channel/Area Scan (41x201x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 0.482 mW/g

Position 5mm, Low channel/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 18.5 V/m; Power Drift = -0.025 dB Peak SAR (extrapolated) = 0.614 W/kg SAR(1 g) = 0.421 mW/g; SAR(10 g) = 0.279 mW/g Maximum value of SAR (measured) = 0.489 mW/g



Fig. 8: SAR distribution for SIRRETA antenna, Low channel (903 MHz), rear side at 5 mm



Communication System: APPI-COM; Frequency: 915 MHz; Duty Cycle: 1:8.33 Medium parameters used: σ = 1.03373 mho/m, ϵ_r = 52.4694; ρ = 1000 kg/m³ Phantom section: Flat Section

Program Notes: Ambient temperature: 22.9 °C, Liquid temperature: 21.2°C

DASY4 Configuration:

- Probe: ES3DV3 SN3303; ConvF(5.92, 5.92, 5.92); Calibrated: 8/21/2015
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE3 Sn402; Calibrated: 8/13/2015
- Phantom: ELI 4.0; Type: QDOVA001BA; Serial: xxxx
- Measurement SW: DASY4, V4.5 Build 19; Postprocessing SW: SEMCAD, V1.8 Build 145

Position 5mm, Middle channel/Area Scan (41x201x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 0.390 mW/g

Position 5mm, Middle channel/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid: dx=5mm,

dy=5mm, dz=5mm Reference Value = 15.7 V/m; Power Drift = -0.028 dB Peak SAR (extrapolated) = 0.513 W/kg SAR(1 g) = 0.343 mW/g; SAR(10 g) = 0.225 mW/g Maximum value of SAR (measured) = 0.404 mW/g







Communication System: APPI-COM; Frequency: 927 MHz; Duty Cycle: 1:8.33 Medium parameters used: σ = 1.04425 mho/m, ϵ_r = 52.4234; ρ = 1000 kg/m³ Phantom section: Flat Section

Program Notes: Ambient temperature: 23.1 °C, Liquid temperature: 21.3 °C

DASY4 Configuration:

- Probe: ES3DV3 SN3303; ConvF(5.92, 5.92, 5.92)
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE3 Sn402
- Phantom: ELI 4.0; Type: QDOVA001BA; Serial: xxxx
- Measurement SW: DASY4, V4.5 Build 19; Postprocessing SW: SEMCAD, V1.8 Build 145

Position 5mm, Middle channel/Area Scan (41x201x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 0.291 mW/g

Position 5mm, Middle channel/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid: dx=5mm,

dy=5mm, dz=5mm Reference Value = 14.1 V/m; Power Drift = -0.01 dB Peak SAR (extrapolated) = 0.373 W/kg SAR(1 g) = 0.251 mW/g; SAR(10 g) = 0.166 mW/g Maximum value of SAR (measured) = 0.294 mW/g







9. PHOTOGRAPHS OF THE EQUIPMENT UNDER TEST

The photographs of the equipment under test are shown in Fig. 11 to Fig. 16.



Fig. 11: SIRETTA antenna, rear side at 0 mm from the phantom



Fig. 12: SIRETTA antenna, rear side at 0 mm from the phantom





Fig. 13: NEARSON antenna, rear side at 0 mm from the phantom



Fig. 14: NEARSON antenna, rear side at 0 mm from the phantom





Fig. 15: SIRETTA antenna, rear side at 5 mm from the phantom



Fig. 16: SIRETTA antenna, rear side at 5 mm from the phantom



10. MEASUREMENT UNCERTAINTY

- Measurement uncertainty of SAR evaluations

The uncertainty of the measurements was evaluated according to the IEEE Std 1528, Tableau 9 (Handset). The expanded uncertainty is \pm 26.4 % in 1g.

Uncertainty component	Uncertainty Value (%)	Probability Distribution	Divisor	Ci (1a)	Ci (10g)	1g //:(%)	10g //: (%)
Measurement system		Distribution		(19)	(109)	un (10)	un (70)
Probe calibration	± 6.0	Normal	1	1	1	± 6.0	± 6.0
Probe calibration drift	± 7.0	Rectangular	2√3	1	1	+ 2.0	+ 2.0
Axial isotropy	± 4.7	Rectangular	√3	0.7	0.7	± 1.9	± 1.9
Hemispherical isotropy	± 9.6	Rectangular	√3	0.7	0.7	± 3.9	± 3.9
Boundary effect	± 1.0	Rectangular	√3	1	1	± 0.6	± 0.6
Linearity	± 4.7	Rectangular	√3	1	1	± 2.7	± 2.7
System detection limits	± 1.0	Rectangular	√3	1	1	± 0.6	± 0.6
Modulation response	± 2.4	Rectangular	√3	1	1	± 1.4	± 1.4
Readout electronics	± 0.3	Normal	1	1	1	± 0.3	± 0.3
Response time	± 0.8	Rectangular	√3	1	1	± 0.5	± 0.5
Integration time	± 2.6	Rectangular	√3	1	1	± 1.5	± 1.5
RF ambient conditions - noise	± 3.0	Rectangular	√3	1	1	± 1.7	± 1.7
RF ambient conditions - reflections	± 5.0	Rectangular	√3	1	1	± 2.9	± 2.9
Probe positioner mechanical tolerance	± 0.4	Rectangular	√3	1	1	± 0.2	± 0.2
Probe positioning with respect to phantom shell	± 2.9	Rectangular	√3	1	1	± 1.7	± 1.7
Extrapolation, interpolation and integration algorithms for max. SAR evaluation	± 2.0	Rectangular	√3	1	1	± 1.2	± 1.2
Test sample related							
Test sample positioning	± 2.9	Normal	1	1	1	±2.9	± 2.9
Device holder uncertainty	\pm 3.6	Normal	1	1	1	± 3.6	\pm 3.6
Output power variation – SAR drift	± 5.0	Rectangular	√3	1	1	± 2.9	± 2.9
measurement			1-				
SAR scaling	± 0.0	Rectangular	√3	1	1	±0.0	± 0.0
Phantom and tissue parameters			10				
Phantom shell uncertainty – shape, thickness and permittivity	± 6.1	Rectangular	√3	1	1	± 3.5	± 3.5
Uncertainty in SAR correction for deviations in permittivity and conductivity	± 1.9	Normal	1	1	0.84	± 1.9	± 1.6
Liquid conductivity measurement	± 2.5	Normal	1	0.78	0.71	± 2.0	± 1.8
Liquid permittivity measurement	± 2.5	Normal	1	0.23	0.26	± 0.6	± 0.7
Liquid conductivity – temperature uncertainty	± 2.8	Rectangular	√3	0.78	0.71	± 1.3	± 1.2
Liquid permittivity – temperature uncertainty	± 1.9	Rectangular	√3	0.23	0.26	± 0.3	± 0.3
Combined standard uncertainty						± 13.2	± 13.2
Expanded uncertainty (95% confidence interval)						± 26.4	± 26.4



- Uncertainty of SAR system validation

The uncertainty of the system validation was evaluated according to the IEEE Std 1528 Tableau 10 (System validation). The expanded uncertainty is \pm 24 % in 1g.

Uncertainty component	Uncertainty Value (%)	Probability Distribution	Divisor	сі (1g)	сі (10g)	1g <i>u_i</i> (%)	10g <i>u</i> i (%)
Measurement system				. •			
Probe calibration	± 6.0	Normal	1	1	1	± 6.0	± 6.0
Probe calibration drift	± 7.0	Rectangular	2√3	1	1	± 2.0	± 2.0
Axial isotropy	± 4.7	Rectangular	√3	1	1	± 2.7	± 2.7
Hemispherical isotropy	± 9.6	Rectangular	√3	0	0	± 0.0	± 0.0
Boundary effect	± 1.0	Rectangular	√3	1	1	± 0.6	± 0.6
Linearity	± 4.7	Rectangular	√3	1	1	± 2.7	± 2.7
System detection limits	± 1.0	Rectangular	√3	1	1	± 0.6	± 0.6
Modulation response	± 0.0	Rectangular	√3	1	1	± 0.0	± 0.0
Readout electronics	± 0.3	Normal	1	1	1	± 0.3	± 0.3
Response time (CW)	± 0.0	Rectangular	√3	1	1	± 0.0	± 0.0
Integration time	± 0.0	Rectangular	√3	1	1	± 0.0	± 0.0
RF ambient conditions - noise	± 2.0	Rectangular	√3	1	1	± 1.2	± 1.2
RF ambient conditions - reflections	± 2.0	Rectangular	√3	1	1	± 1.2	± 1.2
Probe positioner mechanical tolerance	± 0.4	Rectangular	√3	1	1	± 0.2	± 0.2
Probe positioning with respect to phantom shell	± 2.9	Rectangular	√3	1	1	± 1.7	± 1.7
Extrapolation, interpolation and integration algorithms for max. SAR evaluation	± 2.0	Rectangular	√3	1	1	± 1.2	± 1.2
System validation source							
Deviation of experimental source from numerical source	± 4.0	Normal	1	1	1	± 4.0	± 4.0
Input power and SAR drift measurement	± 3.4	Rectangular	√3	1	1	± 2.0	± 2.0
Other source contributions	± 2.0	Normal	1	1	1	± 2.0	± 2.0
Phantom and tissue parameters							
Phantom shell uncertainty – shape, thickness and permittivity	± 6.1	Rectangular	√3	1	1	± 3.5	\pm 3.5
Uncertainty in SAR correction for deviations in permittivity and conductivity	± 1.9	Normal	1	1	0.84	± 1.9	± 1.6
Liquid conductivity measurement	± 2.5	Normal	1	0.78	0.71	± 2.0	± 1.8
Liquid permittivity measurement	± 2.5	Normal	1	0.23	0.26	±0.6	± 0.7
Liquid conductivity – temperature uncertainty	± 2.8	Rectangular	√3	0.78	0.71	± 1.3	± 1.2
Liquid permittivity – temperature uncertainty	± 1.9	Rectangular	√3	0.23	0.26	± 0.3	± 0.3
Combined standard uncertainty						± 12.0	± 12.0
Expanded uncertainty (95% confidence interval)						± 24.0	± 24.0



11. SPATIAL PEAK SAR EVALUATION

From Schmid & Partner Engineering AG, [DASY4 Manual, March 2003, Application Note: Spatial Peak SAR Evaluation].

Spatial Peak SAR

The DASY4 software includes all numerical procedures necessary to evaluate the spatial peak SAR values.

The spatial-peak SAR can be computed over any required mass.

The base for the evaluation is a "cube" measurement in a volume of 30mm³ (7x7x7 points). The measured volume includes the 1g and 10g cubes with the highest averaged SAR values. For that purpose, the center of the measured volume is aligned to the interpolated peak SAR value of a previously performed area scan. The entire evaluation of the spatial peak values is performed within the post-processing engine (SEMCAD). The system always gives the maximum values for the 1g and 10g cubes. The algorithm to find the cube with highest averaged SAR is divided into the following stages:

1. Extraction of the measured data (grid and values) from the Zoom Scan,

2. Calculation of the SAR value at every measurement point based on all stored data (A/D values and measurement parameters),

- 3. Generation of a high-resolution mesh within the measured volume,
- 4. Interpolation of all measured values from the measurement grid to the high-resolution grid,
- 5. Extrapolation of the entire 3-D field distribution to the phantom surface over the distance from sensor to surface,
- 6. Calculation of the averaged SAR within masses of 1g and 10g.

Interpolation, Extrapolation and Detection of Maxima

The probe is calibrated at the center of the dipole sensors which is located 2 mm away from the probe tip. During measurements, the dipole sensors are 3 mm above the phantom surface. Both distances are included as parameters in the probe configuration file. The software always knows exactly how far away the measured point is from the surface. As the probe cannot directly measure at the surface, the values between the deepest measured point and the surface must be extrapolated.

The interpolation, extrapolation and maximum search routines are all based on the modified Quadratic Shepard's method [Robert J. Renka, "Multivariate Interpolation Of Large Sets Of Scattered Data", University of North Texas ACM Transactions on Mathematical Software, vol. 14, no. 2, June 1988, pp. 139-148.].

Averaging and Determination of Spatial Peak SAR

The interpolated data is used to average the SAR over the 1g and 10g cubes by spatially discretizing the entire measured volume. The resolution of this spatial grid used to calculate the averaged SAR is 1mm or about 42875 interpolated points. The resulting volumes are defined as cubical volumes containing the appropriate tissue parameters that are centered at the location. The location is defined as the center of the incremental volume (voxel).



12. EQUIPMENT TEST CONDITIONS

The following test conditions are given for information; the maximum output powers were measured with a spectrum analyser

Frequency range:902MHz – 928MHz, FHSS, according to the test report N°RE051-16-103329-1-A
Ed.0Modulation:FM, according to the test report N°RE051-16-103329-1-A Ed.0Test program:supplied by the applicant

Measured output power: RBW 1MHz

Frequency (MHz)	Conducted maximum peak power (dBm)
903 (Low)	26.2
915 (Middle)	26.7
927 (High)	26.6







Measured Crest factor:

8.33 (total Tx duration is 57.6ms within a frame length of 480ms)

903MHz:





RE051-18-100576-1-A Ed. 2









Secondary transmitter:Bluetooth Module : see the test report N° °RE051-16-103329-1-A Ed.0



13. MEASUREMENT SYSTEM DESCRIPTION

The automated near-field scanning system Dosimetric Assessment System DASY4 from Schmid & Partner Engineering AG was used. The measurement is performed using platform n° 2 (DASY4) referenced in paragraph 6 of this test report. The system consists of a computer controlled, high precision robotics system, robot controller, extreme near-field probes and the phantom containing the liquid. The six axis robot precisely positions the probe at the points of maximum electromagnetic field. A device holder made of low-loss dielectric material is used to maintain the test position of the equipment under test against the phantom. Fig. 17 shows the system.



Fig. 17: The measurement setup



14.LIQUID MEASUREMENT: TEST CONDITIONS & RESULTS

The liquid measurement is performed using platform n° 3 (Liquid measure) referenced in paragraph 6 of this test report. The following ingredients (in % by weight) are theoretical and given for information.

900 MHz liquid: Sugar 41.76 % De-ionised water 56.0 % NaCl salt 0.76 % HEC 1.21 % Bactericide 0.27 %

The dielectric parameters of the liquid were controlled prior to assessment (contact probe method).

Dielectric properties measured:

Date	Frequency (MHz)	ε _r (F/m) Targeted value	ε _r (F/m) Measured value	σ (S/m) ^{Targeted} value	σ (S/m) Measured value	Liquid temperature (°C)	Ambient temperature (°C)
	900	$55.0\pm5~\%$	53.1	$1.05\pm5~\%$	1.02	21.8	
April 23	905	$55.0\pm5~\%$	53	1.05 ± 5 %	1.03		21.7
	915	$55.0\pm5~\%$	52.9	1.06 ± 5 %	1.04		
	925	$55.0\pm5~\%$	52.7	1.06 ± 5 %	1.05		
	900	$55.0\pm5~\%$	52.6	1.05 ± 5 %	1.02		
A	905	$55.0\pm5~\%$	52.5	1.05 ± 5 %	1.03	01.0	<u>оо г</u>
April 24	915	$55.0\pm5~\%$	52.5	1.06 ± 5 %	1.03	21.0	22.0
	925	$55.0\pm5~\%$	52.4	$1.06 \pm 5 \%$	1.04		

The phantom shall be filled with tissue-equivalent liquid to a depth of at least 15cm.





15. SYSTEM VALIDATION: TEST CONDITIONS & RESULTS

The system validation is performed using platform n° 4 (System validation) referenced in paragraph 6 of this test report.

Measurement conditions:

The measurements were performed with the ELI4 phantom filled with liquid. The validation dipole input power was 250 mW. Prior to the assessment, the validation dipole were used to check whether the system was operating within its specification of \pm 10 %.

Measurement results:

The results is below and shown in Fig. 18

Meas. date	Frequency (MHz)	SAR 1g (W/kg) Targeted value	SAR 1g (W/kg) Measured value
Apr. 23	900	$2.775\pm10\%$	2.700



DUT: Dipole 900 MHz

Communication System: CW; Frequency: 900 MHz; Duty Cycle: 1:1 Medium parameters used: $\sigma = 1.02483$ mho/m, $\epsilon_r = 53.0886$; $\rho = 1000$ kg/m³ Phantom section: Flat Section

Program Notes: Ambient temperature: 21.7 °C, Liquid temperature: 21.8°C

DASY4 Configuration:

- Probe: ES3DV3 SN3303; ConvF(6.29, 6.29, 6.29)
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE3 Sn402
- Phantom: ELI 4.0; Type: QDOVA001BA; Serial: xxxx
- Measurement SW: DASY4, V4.5 Build 19; Postprocessing SW: SEMCAD, V1.8 Build 145

d=15mm, Pin=250mW/Area Scan (61x81x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 3.16 mW/g

d=15mm, Pin=250mW/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid: dx=5mm,

dy=5mm, dz=5mm Reference Value = 57.4 V/m; Power Drift = -0.040 dB Peak SAR (extrapolated) = 4.01 W/kg SAR(1 g) = 2.7 mW/g; SAR(10 g) = 1.76 mW/g Maximum value of SAR (measured) = 3.17 mW/g





□□□ End of report, 3 annexes to be forwarded □□□



Γ

ANNEX 1: DAE3 CALIBRATION CERTIFICATE

Engliteening AG Zeughausstrasse 43, 8004 Zurid	ch, Switzerland		Schweizenscher Kaupherdien: Service suisse d'étalonnage Servizio svizzero di taratura Swiss Calibration Service
Accredited by the Swiss Accredit The Swiss Accreditation Servic Multilateral Agreement for the	tation Service (SAS) ce is one of the signatories recognition of calibration	Accreditation s to the EA certificates	No.: SCS 0108
Client EMITECH Le I	Mans	Certificate No	: DAE3-402_Aug16
Object	DAE3 - SD 000 D	03 AA - SN: 402	
Calibration procedure(s)	QA CAL-06.v29 Calibration proced	dure for the data acquisition elect	ronics (DAE)
Calibration date:	August 09, 2016		
This calibration certificate docum The measurements and the unce All calibrations have been condu	nents the traceability to natio entainties with confidence pro- cted in the closed laboratory	mal standards, which realize the physical unit obability are given on the following pages and r facility: environment temperature $(22 \pm 3)^{\circ}$ C	s of measurements (SI). I are part of the certificate. and humidity < 70%.
This calibration certificate docum The measurements and the unce All calibrations have been condu Calibration Equipment used (M& Primary Standards Keithley Multimeter Type 2001	nents the traceability to natio entainties with confidence pro- cted in the closed laboratory TE critical for calibration) LD # SN: 0810278	mal standards, which realize the physical unit obability are given on the following pages and r facility: environment temperature (22 ± 3)°C Cal Date (Certificate No.) 09-Sep-15 (No:17153)	s of measurements (SI). I are part of the certificate. and humidity < 70%. Scheduled Calibration Sep-16
This calibration certificate docum The measurements and the unce All calibrations have been condu Calibration Equipment used (M& <u>Primary Standards</u> Keithley Multimeter Type 2001 Secondary Standards	nents the traceability to natio entainties with confidence pro- icted in the closed laboratory TE critical for calibration) 1D # SN: 0810278	onal standards, which realize the physical unit obability are given on the following pages and r facility: environment temperature (22 ± 3)°C <u>Cal Date (Certificate No.)</u> 09-Sep-15 (No:17153) Check Date (in bouse)	s of measurements (SI). I are part of the certificate. and humidity < 70%. Scheduled Calibration Sep-16 Schertuled Check
This calibration certificate docum The measurements and the unce All calibrations have been condu Calibration Equipment used (M& Primary Standards Keithley Multimeter Type 2001 Secondary Standards Auto DAE Calibration Unit Calibrator Box V2.1	ents the traceability to natio entainties with confidence pro- cted in the closed laboratory TE critical for calibration) ID # SE UWS 053 AA 1001 SE UWS 056 AA 1002	onal standards, which realize the physical unit obability are given on the following pages and r facility: environment temperature (22 ± 3)°C Cal Date (Certificate No.) 09-Sep-15 (No:17153) Check Date (in house) 05-Jan-16 (in house check) 05-Jan-16 (in house check)	s of measurements (SI). I are part of the certificate. and humidity < 70%. Scheduled Calibration Sep-16 Scheduled Check In house check: Jan-17 In house check: Jan-17
This calibration certificate docum The measurements and the unce All calibrations have been condu Calibration Equipment used (M& <u>Primary Standards</u> Keithley Multimeter Type 2001 <u>Secondary Standards</u> Auto DAE Calibration Unit Calibrator Box V2.1 Calibrated by:	hents the traceability to natio entainties with confidence pro- cted in the closed laboratory TE: critical for calibration) ID # ID # ID # SE UWS 053 AA 1001 SE UWS 056 AA 1002 Name Dominique Steffen	rnal standards, which realize the physical unit obability are given on the following pages and r facility: environment temperature (22 ± 3)°C <u>Cal Date (Certificate No.)</u> 09-Sep-15 (No:17153) <u>Check Date (in house)</u> 05-Jan-16 (in house check) 05-Jan-16 (in house check) 05-Jan-16 (in house check) Function Technician	s of measurements (SI). I are part of the certificate. and humidity < 70%. Scheduled Calibration Sep-16 Scheduled Check In house check: Jan-17 In house check: Jan-17
This calibration certificate docum The measurements and the unce All calibrations have been condu Calibration Equipment used (M& <u>Primary Standards</u> Keithley Multimeter Type 2001 <u>Secondary Standards</u> Auto DAE Calibration Unit Calibrator Box V2.1 Calibrated by: Approved by:	hents the traceability to natio entainties with confidence pro- cted in the closed laboratory TE critical for calibration) ID # ID # ID # SE UWS 053 AA 1001 SE UWS 053 AA 1001 SE UMS 006 AA 1002 Name Dominique Steffen Fin Bomholt	nal standards, which realize the physical unit obability are given on the following pages and r facility: environment temperature (22 ± 3)°C <u>Cal Date (Certificate No.)</u> 09-Sep-15 (No:17153) <u>Check Date (in house)</u> 05-Jan-16 (in house check) 05-Jan-16 (in house check) 05-Jan-16 (in house check) 05-Jan-16 (in house check) Function Technician Deputy Technical Manager	s of measurements (SI). I are part of the certificate. and humidity < 70%. Scheduled Calibration Sep-16 Scheduled Check In house check: Jan-17 In house check: Jan-17 Signature



Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





Schweizerischer Kalibrierdienst

Service suisse d'étalonnage Servizio svizzero di taratura

Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multifateral Agreement for the recognition of calibration certificates

Accreditation No.: SCS 0108

Glossary

EMITECH

DAE Connector angle

data acquisition electronics information used in DASY system to align probe sensor X to the robot coordinate system.

Methods Applied and Interpretation of Parameters

- *DC Voltage Measurement:* Calibration Factor assessed for use in DASY system by comparison with a calibrated instrument traceable to national standards. The figure given corresponds to the full scale range of the voltmeter in the respective range.
- Connector angle: The angle of the connector is assessed measuring the angle mechanically by a tool inserted. Uncertainty is not required.
- The following parameters as documented in the Appendix contain technical information as a result from the performance test and require no uncertainty.
 - DC Voltage Measurement Linearity: Verification of the Linearity at +10% and -10% of the nominal calibration voltage. Influence of offset voltage is included in this measurement.
 - Common mode sensitivity: Influence of a positive or negative common mode voltage on the differential measurement.
 - Channel separation: Influence of a voltage on the neighbor channels not subject to an input voltage.
 - AD Converter Values with inputs shorted: Values on the internal AD converter corresponding to zero input voltage
 - Input Offset Measurement: Output voltage and statistical results over a large number of zero voltage measurements.
 - Input Offset Current: Typical value for information; Maximum channel input offset current, not considering the input resistance.
 - Input resistance: Typical value for information: DAE input resistance at the connector, during internal auto-zeroing and during measurement.
 - Low Battery Alarm Voltage: Typical value for information. Below this voltage, a battery alarm signal is generated.
 - Power consumption: Typical value for information. Supply currents in various operating modes.

Certificate No: DAE3-402_Aug16

Page 2 of 5



DC Voltage Measurement

A/D - Converter Reso	lution nominal			
High Range:	1LSB =	6.1µV,	full range =	-100+300 mV
Low Range:	1LSB =	61nV,	full range =	-1+3mV
DASY measurement	parameters: Aut	to Zero Time: 3	sec; Measuring	time: 3 sec

Calibration Factors	X	Ŷ	Z
High Range	403.396 ± 0.02% (k=2)	403.316 ± 0.02% (k=2)	403.886 ± 0.02% (k=2)
Low Range	3.93593 ± 1.50% (k=2)	3.96106 ± 1.50% (k=2)	3.96586 ± 1.50% (k=2)

Connector Angle

Connector Angle to be used in DASY system	241.0 ° ± 1 °

Certificate No: DAE3-402_Aug16

Page 3 of 5



Appendix (Additional assessments outside the scope of SCS0108)

1. DC Voltage Linearity

High Range	Reading (µV)	Difference (µV)	Error (%)
Channel X + Inpu	t 200037.48	2.13	0.00
Channel X + Inpu	20013.38	8.54	0.04
Channel X - Input	-20000.08	5.30	-0.03
Channel Y + Input	200037.31	1.89	0.00
Channel Y + Input	20011.20	6.51	0.03
Channel Y - Input	-20005.18	0.25	-0.00
Channel Z + Input	200038.65	3.05	0.00
Channel Z + Input	20003.69	-0.95	-0.00
Channel Z - Input	-20014.52	-9,04	0.05

Low Range		Reading (µV)	Difference (µV)	Error (%)
Channel X	+ Input	2001.18	0.15	0.01
Channel X	+ Input	201.28	0.22	0.11
Channel X	- Input	-198.57	0.24	-0.12
Channel Y	+ Input	2000.82	-0.12	-0.01
Channel Y	+ Input	200.60	-0.39	-0.20
Channel Y	- Input	-199,38	-0.36	0.18
Channel Z	+ Input	2000.71	-0.07	-0.00
Channel Z	+ Input	199.90	-0.95	-0.47
Channel Z	- Input	-200.15	-1.04	0.52

2. Common mode sensitivity

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

-	Common mode Input Voltage (mV)	High Range Average Reading (µV)	Low Range Average Reading (μV)
Channel X	200	7.24	5.36
	- 200	-4.30	-5.84
Channel Y	200	-1.36	-1.47
	- 200	0.98	0,39
Channel Z	200	0.55	0.71
	- 200	-2.54	-2,93

3. Channel separation

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	Input Voltage (mV)	Channel X (μV)	Channel Y (µV)	Channel Z (µV)
Channel X	200	u	2.78	-2.08
Channel Y	200	8.00	-	3.96
Channel Z	200	8.11	5,41	*

Certificate No: DAE3-402_Aug16

Page 4 of 5



4. AD-Converter Values with inputs shorted

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	High Range (LSB)	Low Range (LSB)
Channel X	16283	15573
Channel Y	15888	17336
Channel Z	16453	17177

5. Input Offset Measurement

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec Input 10M Ω

	Average (μV)	min. Offset (µV)	max. Offset (μV)	Std. Deviation (µV)
Channel X	1.28	0.48	2.74	0.36
Channel Y	0.43	-0.70	1.50	0.33
Channel Z	-0.54	-1.47	0.39	0.36

6. Input Offset Current

Nominal Input circuitry offset current on all channels: <25fA

7. Input Resistance (Typical values for information)

	Zeroing (kOhm)	Measuring (MOhm)
Channel X	200	200
Channel Y	200	200
Channel Z	200	200

8. Low Battery Alarm Voltage (Typical values for information)

Typical values	Alarm Level (VDC)
Supply (+ Vcc)	+7.9
Supply (- Vcc)	-7.6

9. Power Consumption (Typical values for information)

Typical values	Switched off (mA)	Stand by (mA)	Transmitting (mA)	
Supply (+ Vcc)	+0.01	+6	+14	
Supply (- Vcc)	0.01	~8	-9	



ANNEX 2: E-FIELD PROBE CALIBRATION CERTIFICATE

Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zu	t ory of urich, Switzerland	BC MEA	Schweizerischer Kalibrierdiens Service suisse d'étalonnage Servizio svizzero di taratura Swiss Calibration Service
Accredited by the Swiss Accredited by the Swiss Accreditation Sen	ditation Service (SAS) vice is one of the signatorie	es to the EA	ccreditation No.: SCS 0108
Multilateral Agreement for the Client Emitech Le I	e recognition of calibration	certificates	» ES3-3303 Auo16
CALIBRATION	CERTIFICAT		
Object	ES3DV3 - SN:33	03	
Calibration procedure(s)	QA CAL-01.v9, C Calibration proce	DA CAL-12.v9, QA CAL-23.v5, QA dure for dosimetric E-field probes	A CAL-25.v6
Calibration date:	August 23, 2016		
All calibrations have been cond Calibration Equipment used (M	certainties with confidence pr incled in the closed laborator &TE critical for calibration)	obability are given on the following pages and y facility: environment temperature $(22 \pm 3)^{\circ}$ C	and humidity < 70%.
All calibrations have been cond Calibration Equipment used (M Primary Standards	certainties with confidence pr iucted in the closed laborator &TE critical for calibration)	obability are given on the following pages and y facility: environment temperature (22 ± 3)°C	and humidity < 70%.
All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP	ertainties with confidence pr lucted in the closed laborator &TE critical for calibration) 	Obability are given on the following pages and y facility: environment temperature (22 ± 3)°C Cal Date (Certificate No.) 06-Apr-16 (No. 217-02288/02289)	and humidity < 70%.
All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91	ertainties with confidence pr lucted in the closed laborator &TE critical for calibration) ID SN: 104778 SN: 103244	Cal Date (Certificate No.) 06-Apr-16 (No. 217-02288)	and humidity < 70%. <table> Scheduled Calibration Apr-17 Apr-17</table>
All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenueter	ertainties with confidence pr incled in the closed laborator &TE critical for calibration) ID SN: 104776 SN: 103244 SN: 103245	Cal Date (Certificate No.) 06-Apr-16 (No. 217-02288) 06-Apr-16 (No. 217-02288) 06-Apr-16 (No. 217-02288) 06-Apr-16 (No. 217-02288)	and humidity < 70%. Scheduled Calibration Apr-17 Apr-17 Apr-17
All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Reference Probe ES3DV2	Entrainties with confidence pr iucted in the closed laborator &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: 55277 (20x) SN: 3013	Cal Date (Certificate No.) Cal Date (Certificate No.) 06-Apr-16 (No. 217-02288/02289) 06-Apr-16 (No. 217-02288) 06-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02283) 21 Opo 16 (No. 217-02283)	and humidity < 70%. Scheduled Calibration Apr-17 Apr-17 Apr-17
All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Reference Probe ES3DV2 DAE4	Ertainties with confidence pr iucted in the closed laborator &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: 55277 (20x) SN: 3013 SN: 660	Obability are given on the following pages and y facility: environment temperature (22 ± 3)°C Cal Date (Certificate No.) 06-Apr-16 (No. 217-02288/02289) 06-Apr-16 (No. 217-02288) 06-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02289) 31-Dec-15 (No. ES3-3013, Dec15) 23-Dec-15 (No. DAE4-660 Dec15)	and humidity < 70%. Scheduled Calibration Apr-17 Apr-17 Apr-17 Apr-17 Dec-16 Dec-16
All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Reference Probe ES3DV2 DAE4	Ertainties with confidence pr iucted in the closed laborator &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: 103245 SN: 55277 (20x) SN: 3013 SN: 860	Cal Date (Certificate No.) Cal Date (Certificate No.) 06-Apr-16 (No. 217-02288) 06-Apr-16 (No. 217-02288) 06-Apr-16 (No. 217-02288) 06-Apr-16 (No. 217-0228) 05-Apr-16 (No. 217-0228) 05-Apr-16 (No. 217-0228) 05-Apr-16 (No. 217-0228) 05-Apr-16 (No. 217-0228) 23-Dec-15 (No. DAE4-660_Dec15)	and humidity < 70%. Scheduled Calibration Apr-17 Apr-17 Apr-17 Apr-17 Dec-16 Dec-16
All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Reference Probe ES3DV2 DAE4 Secondary Standards Power meter E44109	Ertainties with confidence pr iucted in the closed laborator &TE critical for calibration) ID SN: 104776 SN: 103244 SN: 103245 SN: 3013 SN: 3013 SN: 660 ID	Cal Date (Certificate No.) Cal Date (Certificate No.) 06-Apr-16 (No. 217-02288/02289) 06-Apr-16 (No. 217-02288) 06-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02289)	and humidity < 70%. Scheduled Calibration Apr-17 Apr-17 Apr-17 Apr-17 Dec-16 Dec-16 Scheduled Check
All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP- Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Reference Probe ES3DV2 DAE4 Secondary Standards Power meter E4419B Power sensor E4412A	Ertainties with confidence pr iucted in the closed laborator &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: 3013 SN: 3013 SN: 60 ID SN: GB41293874 SN: WY4140907	Obability are given on the following pages and y facility: environment temperature (22 ± 3)°C Cal Date (Certificate No.) 06-Apr-16 (No. 217-02288/02289) 06-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02289) 06-Apr-16 (No. 217-02289) 06-Apr-16 (In house) 06-Apr-16 (In house) 06-Apr-16 (In house check Jun-16)	and humidity < 70%. Scheduled Calibration Apr-17 Apr-17 Apr-17 Apr-17 Dec-16 Dec-16 Scheduled Check In house check: Jun-18
All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Reference 20 dB Attenuator Reference Probe ES3DV2 DAE4 Secondary Standards Power meter E4419B Power sensor E4412A Power sensor E4412A	Ertainties with confidence pr iucted in the closed laborator &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: 55277 (20x) SN: 3013 SN: 660 ID SN: GB41293674 SN: MY41498087 SN: 000110210	Obability are given on the following pages and y facility: environment temperature (22 ± 3)°C Cal Date (Certificate No.) 06-Apr-16 (No. 217-02288/02289) 06-Apr-16 (No. 217-02288) 06-Apr-16 (No. 217-02283) 05-Apr-16 (No. 217-02283) 06-Apr-16 (No. 217-02283)	and humidity < 70%. Scheduled Calibration Apr-17 Apr-17 Apr-17 Apr-17 Dec-16 Dec-16 Scheduled Check In house check: Jun-18 In house check: Jun-18 In house check: Jun-18
All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Reference Probe ES3DV2 DAE4 Secondary Standards Power meter E4419B Power sensor E4412A Power sensor E4412A RF generator HP 8648C	Ertainties with confidence pr iucted in the closed laborator &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: 55277 (20x) SN: 3013 SN: 660 ID SN: GB41293674 SN: 60110210 SN: 000110210 SN: US3642U01700	Obability are given on the following pages and y facility: environment temperature (22 ± 3)°C Cal Date (Certificate No.) 06-Apr-16 (No. 217-02288/02288) 06-Apr-16 (No. 217-02288) 06-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02283) 31-Dec-15 (No. ES3-3013, Dec15) 23-Dec-15 (No. DAE4-660_Dec15) Check Date (in house) 06-Apr-16 (in house check Jun-16) 06-Apr-16 (in house check Jun-16) 06-Apr-16 (in house check Jun-16)	and humidity < 70%. Scheduled Calibration Apr-17 Apr-17 Apr-17 Apr-17 Dec-16 Dec-16 Scheduled Check In house check: Jun-18 In house check: Jun-18
The measurements and the un All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Reference Probe ES3DV2 DAE4 Secondary Standards Power meter E4419B Power sensor E4419A Power sensor E4412A Power sensor E4412A Power sensor E4412A RF generator HP 8648C Network Analyzer HP 8753E	Ertainties with confidence pr iucted in the closed laborator &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: 3013 SN: 3013 SN: 660 ID SN: GB41293674 SN: MY41498087 SN: 000110210 SN: US3642U01700 SN: US37390585	Cal Date (Certificate No.) Cal Date (Certificate No.) 06-Apr-16 (No. 217-02288/02289) 06-Apr-16 (No. 217-02288) 06-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02289) 06-Apr-16 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) 06-Apr-19 (In house check Jun-16) 04-Aug-99 (In house check Jun-16) 18-Oct-01 (In house check Oct-15)	and humidity < 70%. Scheduled Calibration Apr-17 Apr-17 Apr-17 Apr-17 Dec-16 Dec-16 Dec-16 Scheduled Check In house check: Jun-18 In house check: Jun-18
All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Reference Probe ES3DV2 DAE4 Secondary Standards Power meter E4419B Power sensor E4412A Power sensor E4412A RF generator HP 8648C Network Analyzer HP 8753E	certainties with confidence prediction iucled in the closed laborator &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: 103245 SN: 03244 SN: 0313 SN: 660 ID SN: GB41293674 SN: 000110210 SN: US3642U01700 SN: US37390585 Name	Obability are given on the following pages and y facility: environment temperature (22 ± 3)°C Cal Date (Certificate No.) 06-Apr-16 (No. 217-02288/02289) 06-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02293) 31-Dec-15 (No. ES3-3013, Dec15) 23-Dec-15 (No. DAE4-660_Dec15) Check Date (in house) 06-Apr-16 (in house check Jun-16) 07-Apr-16 (in house check Jun-16) 08-Apr-16 (in house check Jun-16) 08-Apr-16 (in house check Jun-16) 08-Apr-16 (in house check Jun-16) <td>and humidity < 70%. Scheduled Calibration Apr-17 Apr-17 Apr-17 Apr-17 Dec-16 Dec-16 Dec-16 Scheduled Check In house check: Jun-18 In house check: Jun-18</td>	and humidity < 70%. Scheduled Calibration Apr-17 Apr-17 Apr-17 Apr-17 Dec-16 Dec-16 Dec-16 Scheduled Check In house check: Jun-18 In house check: Jun-18
All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Reference Probe ES3DV2 DAE4 Secondary Standards Power meter E4419B Power sensor E4412A Power sensor E4412A Power sensor E4412A RF generator HP 8648C Network Analyzer HP 8763E	Ertainties with confidence pr iucted in the closed laborator &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: 55277 (20x) SN: 3013 SN: 660 ID SN: GB41293674 SN: 60110210 SN: 000110210 SN: US3642U01700 SN: US37390585 Name Jeton Kastrati	Obability are given on the following pages and y facility: environment temperature (22 ± 3)°C Cal Date (Certificate No.) 06-Apr-16 (No. 217-02288/02288) 06-Apr-16 (No. 217-02288) 06-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02293) 31-Dec-15 (No. ES3-3013, Dec15) 23-Dec-15 (No. DAE4-660_Dec15) Check Date (In house) 06-Apr-16 (in house check Jun-16) 04-Aug-99 (in house check Jun-16) 18-Oct-01 (in house check Jun-16) Function Laboratory Technician	and humidity < 70%. Scheduled Calibration Apr-17 Apr-17 Apr-17 Apr-17 Dec-16 Dec-16 Scheduled Check In house check: Jun-18 In house check: Jun-18
The measurements and the un All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Reference Probe ES3DV2 DAE4 Secondary Standards Power sensor E4419B Power sensor E4419B Power sensor E4412A Power sensor E4412A RF generator HP 8648C Network Analyzer HP 8763E Calibrated by:	Ertainties with confidence pr iucted in the closed laborator &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: 3013 SN: 860 ID SN: GB41293674 SN: 000110210 SN: US3642U01700 SN: US3642U01700 SN: US37390585 Name Jeton Kastrati Katja Pokovic	obability are given on the following pages and y facility: environment temperature (22 ± 3)°C Cal Date (Certificate No.) 06-Apr-16 (No. 217-02288/02289) 06-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02293) 31-Dec-15 (No. ES3-3013, Dec15) 23-Dec-15 (No. DAE4-660_Dec15) Check Date (in house) 06-Apr-16 (in house check Jun-16) 05-Apr-16 (in house check Jun-16) 18-Oct-01 (in house check Oct-15) Function Laboratory Technician Technical Manager	and humidity < 70%. Scheduled Calibration Apr-17 Apr-17 Apr-17 Apr-17 Apr-17 Dec-16 Dec-16 Dec-16 Scheduled Check In house check: Jun-18 In house check: Jun-18 I
All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Reference Probe ES3DV2 DAE4 Secondary Standards Power sensor E4412A Power sensor E4412A RF generator HP 8648C Network Analyzer HP 8753E Calibrated by: Opproved by:	Ertainties with confidence pr iucted in the closed laborator &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: 3013 SN: 660 ID SN: GB41293674 SN: 000110210 SN: US3642U01700 SN: US3642U01700 SN: US37390585 Name Jeton Kastrati Katja Pokovic not be reproduced except in f	Obability are given on the following pages and y facility: environment temperature (22 ± 3)°C Cal Date (Certificate No.) 06-Apr-16 (No. 217-02288) 06-Apr-16 (No. 217-02289) 06-Apr-16 (No. 217-02289) 06-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02289) 05-Apr-16 (No. 217-02293) 31-Dec-15 (No. DAE4-660_Dec15) 23-Dec-15 (No. DAE4-660_Dec15) Check Date (in house) 06-Apr-16 (in house check Jun-16) 18-Oct-01 (in house check Jun-16) Isoratory Technician Technical Manager ull without written approval of the laboratory	and humidity < 70%. Scheduled Calibration Apr-17 Apr-17 Apr-17 Apr-17 Dec-16 Dec-16 Scheduled Check In house check: Jun-18 In house check: J





Calibration Laboratory of Schmid & Partner

Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



- S Schweizerischer Kalibrierdienst
- C Service suisse d'étalonnage
- S Servizio svizzero di taratura

Accreditation No.: SCS 0108

Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Glossary:

TSL	tissue simulating liquid
NORMx,y,z	sensitivity in free space
ConvF	sensitivity in TSL / NORMx.v.z
DCP	diode compression point
)F	crest factor (1/duty, cycle) of the RF signal
A, B, C, D	modulation dependent linearization parameters
Polarization g	@ rotation around probe axis
olarization 8	9 rotation around an axis that is in the plane normal to probe axis (at measurement center),
Dommentary Avenue	i.e., a = 0 is normal to probe axis

Connector Angle information used in DASY system to align probe sensor X to the robot coordinate system

Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- b) IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005
- c) IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- d) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Methods Applied and Interpretation of Parameters:

- NORMx, y,z: Assessed for E-field polarization 9 = 0 (f ≤ 900 MHz in TEM-cell; f > 1800 MHz: R22 waveguide). NORMx, y,z are only intermediate values, i.e., the uncertainties of NORMx, y,z does not affect the E²-field uncertainty inside TSL (see below *ConvF*).
- NORM(f)x,y,z = NORMx,y,z * frequency_response (see Frequency Response Chart). This linearization is
 implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included
 in the stated uncertainty of ConvF.
- DCPx,y,z: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- PAR: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- Ax,y,z; Bx,y,z; Cx,y,z; Dx,y,z; VRx,y,z; A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters: Assessed in flat phantom using E-field (or Temperature Transfer Standard for f ≤ 800 MHz) and inside waveguide using analytical field distributions based on power measurements for f > 800 MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORMx,y,z * ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from ± 50 MHz to ± 100 MHz.
- Spherical isotropy (3D deviation from isotropy): in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle: The angle is assessed using the information gained by determining the NORMx (no uncertainty required).

Certificate No: ES3-3303_Aug16

Page 2 of 11



ES3DV3 - SN:3303

August 23, 2016

Probe ES3DV3

SN:3303

Manufactured: August 27, 2010 Calibrated: August 23, 2016

Calibrated for DASY/EASY Systems (Note: non-compatible with DASY2 system!)

Certificate No: ES3-3303_Aug16

Page 3 of 11





ES3DV3-SN:3303

August 23, 2016

DASY/EASY - Parameters of Probe: ES3DV3 - SN:3303

Basic Calibration Parameters

	Sensor X	Sensor Y	Sensor Z	Unc (k=2)
Norm $(\mu V/(V/m)^2)^A$	1.32	1.35	1.36	±10.1 %
DCP (mV) ^b	102.5	101.5	102.1	

Modulation Calibration Parameters

UID	Communication System Name		A dB	B dBõV	С	D dB	VR mV	Unc [⊨] (k≕2)
0	CW	X	0.0	0.0	1.0	0.00	229.1	±3.5 %
		Y	0.0	0.0	1.0		233.9	
		Z	0.0	0.0	1.0		223.8	

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

^A The uncertainties of Norm X,Y,Z do not affect the E²-field uncertainty inside TSL (see Pages 5 and 6).
 ^B Numerical linearization parameter: uncertainty not required.
 ^E Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

Certificate No: ES3-3303_Aug16

Page 4 of 11

ES3DV3-- SN:3303

EMITECH

August 23, 2016

DASY/EASY - Parameters of Probe: ES3DV3 - SN:3303

f (MHz) ^C	Relative Permittivity ^F	Conductivity (S/m) ^F	ConvF X	ConvF Y	ConvF Z	Alpha ^G	Depth ^G (mm)	Unc (k=2)
450	43.5	0.87	6.93	6.93	6.93	0.20	2.20	± 13.3 %
750	41.9	0.89	6.79	6.79	6.79	0.53	1.45	± 12.0 %
900	41.5	0.97	6.37	6.37	6.37	0.80	1.14	± 12.0 %
1810	40.0	1.40	5.35	5.35	5.35	0.77	1.20	± 12.0 %
1950	40.0	1.40	5.14	5.14	5.14	0.72	1.27	± 12.0 %
2150	39.7	1.53	5.10	5.10	5.10	0.80	1.16	± 12.0 %
2300	39.5	1.67	4.96	4.96	4.96	0.80	1.22	± 12.0 %
2450	39.2	1.80	4.65	4.65	4.65	0.65	1.40	± 12.0 %
2600	39.0	1.96	4.55	4.55	4.55	0.76	1.35	± 12.0 %

Calibration	Parameter	Determined i	n Head Tissue	Simulating Modia

^C Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency uncertainty is the RSS of the ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively.

below 300 MHz is \pm 10, 25, 40, 50 and 70 MHz for Conver assessments at 30, 04, 120, 130 and 20 MHz responsely. Access 6 one inspectively, Access 6 one in

Certificate No: ES3-3303_Aug16

Page 5 of 11

ES3DV3-- SN:3303

EMITECH

August 23, 2016

DASY/EASY - Parameters of Probe: ES3DV3 - SN:3303

f (MHz) ^c	Relative Permittivity [#]	Conductivity (S/m) ^F	ConvF X	ConvF Y	ConvF Z	Aipha ^G	Depth ^G (mm)	Unc (k=2)
750	55.5	0.96	6.24	6.24	6.24	0.80	1.11	± 12.0 %
900	55.0	1.05	6,29	6.29	6.29	0.80	1.19	± 12.0 %
1810	53.3	1.52	5.01	5.01	5.01	0.57	1.45	± 12.0 %
2000	53.3	1.52	5.01	5.01	5.01	0.75	1.28	± 12.0 %
2450	52.7	1.95	4,52	4.52	4.52	0.80	1.10	± 12.0 %

Calibration Parameter Determined in Body Tissue Simulating Media

^c Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity validity can be extended to ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity validity can be extended to ± 110 MHz.
^c At frequencies below 3 GHz, the validity of tissue parameters (*x* and *σ*) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters (*x* and *σ*) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.
^c Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

diameter from the boundary.

Certificate No: ES3-3303_Aug16

Page 6 of 11















Page 53 of 56





Page 54 of 56

ES3DV3-- SN:3303

EMITECH

August 23, 2016

DASY/EASY - Parameters of Probe: ES3DV3 - SN:3303

Other Probe Parameters

Sensor Arrangement	· · · · · · · · · · · · · · · · · · ·
Connector Angle (°)	
Mechanical Surface Detection Mode	44.7
Optical Surface Detection Mode	enabled
Probe Overall Length	disabled
Probe Body Diameter	337 mm
Tip Length	10 mm
Tip Diameter	10 mm
Probe Tin to Sensor X Colibration Date	4 mm
Probe Tip to Sensor X Collibration Politi	2 mm
Probe Tip to Sensor 7 Calibration Point	2 mm
Possessed and Management And Andrews Andre	2 mm
Neconmenueu weasurement Distance from Surface	3 mm

Certificate No: ES3-3303_Aug16

Page 11 of 11



ANNEX 3: D900V2 CALIBRATION CERTIFICATE

CALIBRATION CERTIFICATE

Equipment under calibration: Designation: 900MHz dipole Brand: Schmid & Partner Engineering AG Type: D900V2 Serial number: 086 Emitech number: 7194

Calibration date: October 17, 2016

Operator: Emmanuel TOUTAIN

Calibration procedure: PRTFIC000MET00050

Environnemental conditions : Ambient Temperature: 21.3°C Liquid Temperature: 20.1°C Hygrometry: 38% HR

Equipment used:

EMITECH N°	DESIGNATION	BRAND	TYPE	CALIBRATION	CALIBRATION
				PERIODICITY	DATE
1402	Network Analyser	Hewlett Packard	8753C	12 months	October 04, 2016
7217	Calibration kit	Hewlett Packard	85033D	12 months	October 04, 2016

Liquid measurements:

Frequency	Liquid :	Head (1)	Liquid : Body		
(MHz)	Sigma	Epsilon	Sigma	Epsilon	
900	0.95	40.8	-	-	

Note (1): dielectric properties according to EN62209-1: 2006 and EN62209-2: 2010

Return loss measurements:

	Frequency (MHz)	Liquid	: Head	Liquid : Body		
Dipole		Return loss (dB)	Verdict (2) \leq -20dB	Return loss (dB)	Verdict (2) \leq -20dB	
D900V2	900	-23.3	PASS	-	-	

Note (2): The reference dipole shall have a return loss better than -20 dB.

Conclusion: In Tolerance

(in head/body tissue at 900MHz according to EN62209-1 and EN62209-2)

<u>Visa:</u>

Intais