



Date (04 October 2018)

Federal Communications Commission  
Authorization and Evaluation Division  
7435 Oakland Mills Road  
Columbia, Maryland 21046

Re: FCC Application  
FCC ID: 2AGO22740TM

To whom it may concern,

We, Inspire Medical Systems Inc., hereby apply for original certification of the Telemetry Module radio included in the Model 2740 Physician Programmer system. This intentional radiator operates at 175kHz pursuant to 47 CFR §15.209 and is attached to a pre-certified radio (FCC ID: T9JRN41-3).

Sincerely,

A handwritten signature in blue ink, which appears to be "Dave Dieken", followed by the date "4 Oct 2018" written in the same ink.

Company Officer: Dave Dieken

Telephone Number: 763-235-6723

Email: [davedieken@inspiresleep.com](mailto:davedieken@inspiresleep.com)

Inspire Medical Systems, Inc.  
9700 63<sup>rd</sup> Avenue North, Suite 200 | Maple Grove, MN 55369  
Telephone: 763-205-7970 | Fax: 763-537-4310 | [www.inspiresleep.com](http://www.inspiresleep.com)