

Class II Permissive Change (C2PC)

2021-05-25

Federal Communication Commission

Equipment Authorization Division, Application Processing Branch
7435 Oakland Mills Road
Columbia, MD 21048

Certification and Engineering Bureau

Industry Canada
Spectrum Engineering Branch
3701 Carling Avenue, Building 94
Ottawa, Ontario K2H 8S2

TO WHOM IT MAY CONCERN

Pursuant to the applicable paragraphs in ISED's RSP-100 §7 and CFR 47 FCC part §2.1043(b)(2) we hereby declare the following changes under Class II Permissive Change to the product:

Hearing Instrument: Ponto 4 45
FCC ID: 2AHJWPONTO4
IC: 21199-PONTO4
Original grant: June 25th, 2019

We, Oticon Medical AB hereby declare the following Class 2 Permissive Change to the radio model for hearing instruments related to the FCC part 15.247 / RSS-247 DTS radio:

- The change of Bluetooth BAW filter and ESD protection diode due to "End-of-Life" components changes has been performed with a minor PCB layout change handled in correlation with KDB 178919 D01, section III, subclause D + Page 3, Note 5, bullet 2.
- Only passive components have been changed.
- No mechanical changes has been applied to the product for this C2PC.
- The radio model for hearing instruments performance is equal to the previously certified product performance.

We, Oticon Medical AB declare that the radio model for hearing instruments changes as documented in the test report will have a similar product performance as established in the FCC/ISED original certification.



Birgitta Bolander, Head of RA
Oticon Medical AB
Datavägen 37B
Sweden
Phone: +46 31 748 61 00
Fax: +46 31 68 77 56
www.oticonmedical.com