

Manufacturer's Statement

FCC ID location justification

FCC ID: 2AHJWSENTIO1

TO WHOM IT MAY CONCERN

We, the undersigned, hereby submit the justification for placing the FCC ID in the Associated User Manual for the Bone Anchor Hearing System having the above-mentioned FCC IDs.

According to KDB Publication 784748 D01, when the device is so small, or for such use that it is impracticable to label with a font size that is four points or larger (and the device does not utilize electronic labelling), then the FCC ID shall be placed in the user manual.

The Sound Processor part of the Bone Anchor Hearing System having the above-mentioned FCC IDs are so small and considering the dimensions of the equipment and the label area, it is impracticable to label FCC ID with a font size required on the devices.

Therefore, the FCC ID label is placed in the User Manual.

Yours sincerely,



Per Klaus Nielsen
Principal certification Specialist, Regulatory Affairs
Oticon Medical

