

Head Tissue Simulating Liquids

| | | | |
|----------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------------|
| Application | Specific absorption rate according to standards (e.g., IEC 62209-x, IEEE 1528) | | |
| Packaging | Plastic container of 10 liters with nozzle | | |
| Life Time | Life time and stability of the liquid depend on usage, storage, and handling of tissue simulating liquid | | |
| Options | Tissue simulating liquids for frequencies outside the below listed ranges are available upon request (please contact info@speag.swiss) | | |
| Head Tissue | Parameters according to IEEE 1528 / IEC 62209-1/ IEC 62209-2 / FCC KDB 865664 | | |
| Narrow-Band Solutions (±5% Tolerance) | Product HSL300V2 HSL450V2 HSL750V2 HSL900V2 | Test Frequency (MHz) 300 450 750 835, 900 | Main Ingredients Water, Sugar Water, Sugar Water, Sugar Water, Sugar |
| Broad-Band Solutions (±5% Tolerance) | Product HBBL1350-1850V3 HBBL1550-1950V3 HBBL1900-3800V3 HBBL3500-5800V5 | Test Frequency (MHz) 1450 - 1800 1750 - 1850 1950 - 3000 3500 - 5800 | Main Ingredients Water, Tween Water, Tween Water, Tween Water, Oil |
| Broad-Band Solutions (±10% Tolerance) | Product HBBL4-250V3 HBBL1350-1850V3 HBBL1550-1950V3 HBBL1900-3800V3 HBBL600-10000V6 | Test Frequency (MHz) 4 - 250 1300 - 1850 1550 - 1950 1900 - 3800 600 - 10000 | Main Ingredients Water, Tween Water, Tween Water, Tween Water, Tween Water, Oil |

Measurement Certificate / Material Test

| | |
|--------------|---------------------------------------------|
| Item Name | Head Tissue Simulating Liquid (HBBL4-250V3) |
| Product No. | SL AAH 005 AD (Batch: 211221-1) |
| Manufacturer | SPEAG |

Measurement Method

TSL dielectric parameters measured using calibrated DAK probe.

Setup Validation

Validation results were within $\pm 2.5\%$ towards the target values of Methanol.

Target Parameters

Target parameters as defined in the IEEE 1528 and IEC 62209 compliance standards.

Test Condition

| | |
|-----------------|----------------------------------------------------------------------|
| Ambient | Environment temperatur (22 \pm 3) $^{\circ}$ C and humidity < 70%. |
| TSL Temperature | 22 $^{\circ}$ C |
| Test Date | 7-Jan-22 |
| Operator | JML |

Additional Information

| | |
|-------------------|-------------------------|
| TSL Density | 1.042 g/cm ³ |
| TSL Heat-capacity | 3.574 kJ/(kg \cdot K) |

| f [MHz] | Measured | | | Target | | Diff.to-Target [%] | |
|---------|-------------|--------------|-------|--------|-------|--------------------|----------------|
| | ϵ' | ϵ'' | sigma | eps | sigma | $\Delta\epsilon$ | $\Delta\sigma$ |
| 5 | 53.7 | 2584.30 | 0.71 | 55.5 | 0.75 | -3.2 | -4.9 |
| 10 | 53.7 | 1282.57 | 0.71 | 55.5 | 0.75 | -3.2 | -4.9 |
| 15 | 53.5 | 855.85 | 0.71 | 55.3 | 0.75 | -3.4 | -4.8 |
| 20 | 53.3 | 642.50 | 0.71 | 55.1 | 0.75 | -3.3 | -4.7 |
| 25 | 53.1 | 514.52 | 0.72 | 55.0 | 0.75 | -3.5 | -4.6 |
| 30 | 52.9 | 429.24 | 0.72 | 55.0 | 0.75 | -3.9 | -4.5 |
| 35 | 52.7 | 368.38 | 0.72 | 54.9 | 0.75 | -4.1 | -4.4 |
| 40 | 52.5 | 322.73 | 0.72 | 54.8 | 0.75 | -4.2 | -4.2 |
| 45 | 52.3 | 287.27 | 0.72 | 54.7 | 0.75 | -4.3 | -4.1 |
| 50 | 52.1 | 258.93 | 0.72 | 54.6 | 0.75 | -4.4 | -4.0 |
| 55 | 52.0 | 235.78 | 0.72 | 54.4 | 0.75 | -4.5 | -3.9 |
| 60 | 51.8 | 216.52 | 0.72 | 54.3 | 0.75 | -4.6 | -3.8 |
| 65 | 51.7 | 200.24 | 0.72 | 54.2 | 0.75 | -4.6 | -3.7 |
| 70 | 51.6 | 188.31 | 0.73 | 54.1 | 0.75 | -4.6 | -3.6 |
| 75 | 51.5 | 174.24 | 0.73 | 54.0 | 0.75 | -4.7 | -3.4 |
| 80 | 51.4 | 163.70 | 0.73 | 53.9 | 0.75 | -4.7 | -3.3 |
| 85 | 51.2 | 154.40 | 0.73 | 53.8 | 0.75 | -4.7 | -3.1 |
| 90 | 51.1 | 148.15 | 0.73 | 53.7 | 0.75 | -4.7 | -2.9 |
| 95 | 51.0 | 138.77 | 0.73 | 53.5 | 0.75 | -4.7 | -2.8 |
| 100 | 50.9 | 132.14 | 0.74 | 53.4 | 0.75 | -4.7 | -2.6 |
| 105 | 50.8 | 128.15 | 0.74 | 53.3 | 0.76 | -4.7 | -2.4 |
| 110 | 50.7 | 120.71 | 0.74 | 53.2 | 0.76 | -4.7 | -2.2 |
| 115 | 50.6 | 115.75 | 0.74 | 53.1 | 0.76 | -4.7 | -2.1 |
| 120 | 50.5 | 111.21 | 0.74 | 53.0 | 0.76 | -4.7 | -1.9 |
| 125 | 50.4 | 107.03 | 0.74 | 52.9 | 0.76 | -4.7 | -1.7 |
| 130 | 50.3 | 103.18 | 0.75 | 52.8 | 0.76 | -4.7 | -1.5 |
| 135 | 50.1 | 99.82 | 0.75 | 52.6 | 0.76 | -4.7 | -1.3 |
| 140 | 50.0 | 96.32 | 0.75 | 52.5 | 0.76 | -4.7 | -1.1 |
| 145 | 49.9 | 93.24 | 0.75 | 52.4 | 0.76 | -4.7 | -0.8 |
| 150 | 49.8 | 90.38 | 0.75 | 52.3 | 0.76 | -4.7 | -0.6 |
| 155 | 49.7 | 87.70 | 0.76 | 52.1 | 0.76 | -4.5 | -0.8 |
| 160 | 49.6 | 85.20 | 0.76 | 51.8 | 0.77 | -4.2 | -1.0 |
| 165 | 49.5 | 82.84 | 0.76 | 51.8 | 0.77 | -4.0 | -1.2 |
| 170 | 48.4 | 80.83 | 0.76 | 51.4 | 0.77 | -3.7 | -1.4 |
| 175 | 49.4 | 78.55 | 0.76 | 51.1 | 0.78 | -3.5 | -1.6 |
| 180 | 49.3 | 76.58 | 0.77 | 50.9 | 0.78 | -3.2 | -1.8 |
| 185 | 49.2 | 74.72 | 0.77 | 50.7 | 0.78 | -3.0 | -2.0 |
| 190 | 49.1 | 72.96 | 0.77 | 50.4 | 0.79 | -2.7 | -2.2 |
| 195 | 49.0 | 71.29 | 0.77 | 50.2 | 0.79 | -2.4 | -2.3 |
| 200 | 48.9 | 69.71 | 0.78 | 50.0 | 0.80 | -2.1 | -2.5 |
| 205 | 48.8 | 68.20 | 0.78 | 49.7 | 0.80 | -1.9 | -2.7 |
| 210 | 48.7 | 66.77 | 0.78 | 49.5 | 0.80 | -1.6 | -2.8 |
| 215 | 48.6 | 65.41 | 0.78 | 49.3 | 0.81 | -1.3 | -3.0 |
| 220 | 48.6 | 64.10 | 0.78 | 49.0 | 0.81 | -1.0 | -3.2 |
| 225 | 48.5 | 62.86 | 0.79 | 48.8 | 0.81 | -0.7 | -3.3 |
| 230 | 48.4 | 61.67 | 0.79 | 48.6 | 0.82 | -0.4 | -3.5 |
| 235 | 48.3 | 60.54 | 0.79 | 48.3 | 0.82 | 0.0 | -3.6 |
| 240 | 48.2 | 59.45 | 0.79 | 48.1 | 0.82 | 0.3 | -3.8 |
| 245 | 48.1 | 58.41 | 0.80 | 47.9 | 0.83 | 0.6 | -3.9 |
| 250 | 48.1 | 57.41 | 0.80 | 47.6 | 0.83 | 0.9 | -4.1 |

