



TEST REPORT

No.I21N00886-HAC RF

For

HMD Global Oy

Smart Phone

Model Name: TA-1357

With

Hardware Version: V01A

Software Version: 00WW_0_010

FCC ID: 2AJOTTA-1357

Results Summary: M Category = M4

Issued Date: 2021-05-26

Designation Number: CN1210

Note:

The test results in this test report relate only to the devices specified in this report. This report shall not be reproduced except in full without the written approval of SAICT.

Test Laboratory:

SAICT, Shenzhen Academy of Information and Communications Technology

Building G, Shenzhen International Innovation Center, No.1006 Shennan Road, Futian District, Shenzhen, Guangdong, P. R. China 518000.

Tel:+86(0)755-33322000, Fax:+86(0)755-33322001

Email: yewu@caict.ac.cn. www.saict.ac.cn



REPORT HISTORY

Report Number	Revision	Description	Issue Date
I21N00886-HAC RF	Rev.0	1st edition	2021-05-26



No. I21N00886-HAC RF

CONTENTS

1. SUMMARY OF TEST REPORT	5
1.1. Тезт Ітемз	5
1.2. Test Standards	5
1.3. TEST RESULT	5
1.4. TESTING LOCATION	5
1.5. PROJECT DATA	5
1.6. SIGNATURE	5
2. CLIENT INFORMATION	6
2.1. APPLICANT INFORMATION	6
2.2. MANUFACTURER INFORMATION	6
3. EQUIPMENT UNDER TEST (EUT) AND ANCILLARY EQUIPMENT (AE)	7
3.1. ABOUT EUT	7
3.2. INTERNAL IDENTIFICATION OF EUT USED DURING THE TEST	7
3.3. INTERNAL IDENTIFICATION OF AE USED DURING THE TEST	7
3.4. AIR INTERFACES / BANDS INDICATING OPERATING MODES	7
4. REFERENCE DOCUMENTS	8
5. OPERATIONAL CONDITIONS DURING TEST	9
5.1. HAC MEASUREMENT SET-UP	9
5.2. PROBE SPECIFICATION	
5.3. TEST ARCH PHANTOM & PHONE POSITIONER	
5.4. ROBOTIC SYSTEM SPECIFICATIONS	11
6. EUT ARRANGEMENT	12
6.1. WD RF EMISSION MEASUREMENTS REFERENCE AND PLANE	
7. SYSTEM VALIDATION	13
7.1. VALIDATION PROCEDURE	13
7.2. VALIDATION RESULT	13
8. MODULATION INTERFERENCE FACTOR (MIF)	14
9. EVALUATION FOR LOW-POWER EXEMPTION	16
9.1. PRODUCT TESTING THRESHOLD	16
9.2. Conducted power	16
10. RF TEST PROCEDURES	17
11. MEASUREMENT RESULTS (E-FIELD)	18
12. ANSI C 63.19-2011 LIMITS	18
13. MEASUREMENT UNCERTAINTY	19



14. MAIN TEST INSTRUMENTS	20
ANNEX A: RF EMISSION TEST PLOT	21
ANNEX B: SYSTEM VALIDATION RESULT	27
ANNEX C: PROBE CALIBRATION CERTIFICATE	29
ANNEX D: DAE CALIBRATION CERTIFICATE	
ANNEX E: DIPOLE CALIBRATION CERTIFICATE	41
ANNEX F: UID SPECIFICATION	51



No. I21N00886-HAC RF

1. Summary of Test Report

1.1. Test Items

Description:	Smart Phone
Model Name:	TA-1357
Applicant's name:	HMD Global Oy
Manufacturer's Name:	HMD Global Oy

1.2. Test Standards

ANSI C63.19-2011

1.3. Test Result

Pass

1.4. Testing Location

Address: Building G, Shenzhen International Innovation Center, No.1006 Shennan Road, Futian District, Shenzhen, Guangdong, P. R. China

1.5. Project Data

Testing Start Date: 2021-04-18

Testing End Date: 2021-04-18

1.6. Signature

本的高

Li Yongfu (Prepared this test report)

Cao Junifei (Approved this test report)

[\]Zhang Yunzhuan (Reviewed this test report)



2. Client Information

2.1. Applicant Information

Company Name:	HMD Global Oy
Address:	Bertel Jungin aukio 9, 02600 Espoo, Finland
City:	/
Country:	/
Telephone:	+393 31 6272922

2.2. Manufacturer Information

Company Name:	HMD Global Oy
Address:	Bertel Jungin aukio 9, 02600 Espoo, Finland
City:	1
Country:	1
Telephone:	+393 31 6272922



3. Equipment under Test (EUT) and Ancillary Equipment (AE)

3.1. About EUT

Description:	Smart Phone
Mode Name:	TA-1357
Condition of EUT as received:	No obvious damage in appearance
Operating mode(a) :	GSM 850/1900, WCDMA Band 2/4/5
Operating mode(s) :	LTE Band 2/4/5/7/12/13/17/28/66, Bluetooth, WLAN 2.4G

3.2. Internal Identification of EUT used during the test

EUT ID*	IMEI	HW Version	SW Version	Receipt Date
UT03aa	350872080007246	V01A	00WW_0_010	2021-04-10

*EUT ID: is used to identify the test sample in the lab internally.

Note: It is performed to test HAC with the UT03aa.

3.3. Internal Identification of AE used during the test

AE ID*	Description	Model	Manufacturer
AE1	Battery	SE681	Shenzhen Aerospac Electronic CO.,Ltd.

*AE ID: is used to identify the test sample in the lab internally.

3.4. Air Interfaces / Bands Indicating Operating Modes

Air-interface	Band(MHz)	Туре	C63.19 /	Simultaneous	Name of Voice	Power
		.,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,	tested	Transmissions	Service	Reduction
GSM	850 / 1900	VO	Yes	BT,WLAN	CMRS Voice	No
EDGE	850 / 1900	DT	No	BT,WLAN	NA	NO
WCDMA	B2 / B4 / B5	VO	Yes	BT,WLAN	CMRS Voice	No
	HSPA	DT	No	BT,WLAN	NA	
LTE (FDD)	2/4/5/7/12/13/17/28/66	VD	Yes	BT,WLAN	VoLTE	No
WLAN	2.4G	VD	Yes	WWAN	VoWIFI	No
Bluetooth	2.4G	DT	No	WWAN	NA	No

VO: Voice CMRS/PSTN Service Only

VD: Voice CMRS/PSTN and Data Service

DT: Digital Transport

* HAC Rating was not based on concurrent voice and data modes; Non-current mode was found to represent worst case rating for both M and T rating



4. Reference Documents

The following document listed in this section is referred for testing.

Reference	Title	Version
ANSI C63.19-2011	American National Standard for Methods of Measurement of Compatibility between Wireless Communication Devices and Hearing Aids	2011
KDB 285076 D01	Equipment Authorization Guidance for Hearing Aid Compatibility	v05
KDB 285076 D02	Guidance for performing T-Coil tests for air interfaces supporting voice over IP (e.g., LTE and WiFi) to support CMRS based telephone services	v03



5. Operational Conditions During Test

5.1. HAC Measurement Set-up

These measurements are performed using the DASY5 NEO automated dosimetric assessment system. It is made by Schmid & Partner Engineering AG (SPEAG) in Zurich, Switzerland. It consists of high precision robotics system (Stäubli), robot controller, Intel Core2 computer, near-field probe, probe alignment sensor. The robot is a six-axis industrial robot performing precise movements. A cell controller system contains the power supply, robot controller, teach pendant (Joystick), and remote control, is used to drive the robot motors. The PC consists of the HP Intel Core2 1.86 GHz computer with Windows XP system and HAC Measurement Software DASY5 NEO, A/D interface card, monitor, mouse, and keyboard. The Stäubli Robot is connected to the cell controller to allow software manipulation of the robot. A data acquisition electronic (DAE) circuit performs the signal amplification, signal multiplexing, AD-conversion, offset measurements, mechanical surface detection, collision detection, etc. is connected to the Electro-optical coupler (EOC). The EOC performs the conversion from the optical into digital electric signal of the DAE and transfers data to the PC plug-in card.




Fig. 1 HAC Test Measurement Set-up

The DAE4 consists of a highly sensitive electrometer-grade preamplifier with auto-zeroing, a channel and gain-switching multiplexer, a fast 16 bit AD-converter and a command decoder and control logic unit. Transmission to the PC-card is accomplished through an optical downlink for data and status information and an optical uplink for commands and clock lines. The mechanical probe mounting device includes two different sensor systems for frontal and sidewise probe contacts. They are also used for mechanical surface detection and probe collision detection. The robot uses its own controller with a built in VME-bus computer.



5.2. Probe Specification

E-Field Probe Description

Construction	One dipole parallel, two dipoles normal to probe axis	
Construction	Built-in shielding against static charges	
	PEEK enclosure material	
	FER enclosure material	
Calibration	In air from 100 MHz to 3.0 GHz (absolute accuracy ±6.0%,	
	k=2)	
		[ER3DV6]
Frequency	40 MHz to > 6 GHz (can be extended to < 20 MHz)	
	Linearity: ± 0.2 dB (100 MHz to 3 GHz)	
Directivity	± 0.2 dB in air (rotation around probe axis)	
	± 0.4 dB in air (rotation normal to probe axis)	
Dynamic Range	2 V/m to > 1000 V/m; Linearity: ± 0.2 dB	
Dynamie Kange		
Dimensions	Overall length: 330 mm (Tip: 16 mm)	
Dimensions	Tip diameter: 8 mm (Body: 12 mm)	
	Distance from probe tip to dipole centers: 2.5 mm	
Application	General near-field measurements up to 6 GHz	
	Field component measurements	
	Fast automatic scanning in phantoms	



5.3. Test Arch Phantom & Phone Positioner

The Test Arch phantom should be positioned horizontally on a stable surface. Reference markings on the Phantom allow the complete setup of all predefined phantom positions and measurement grids by manually teaching three points in the robot. It enables easy and well defined positioning of the phone and validation dipoles as well as simple teaching of the robot (Dimensions: 370 x 370 x 370 mm).

The Phone Positioner supports accurate and reliable positioning of any phone with effect on near field $<\pm 0.5$ dB.



Fig. 2 HAC Phantom & Device Holder

5.4. Robotic System Specifications

Specifications

Positioner: Stäubli Unimation Corp. Robot Model: RX160XL Repeatability: ±0.02 mm No. of Axis: 6 Data Acquisition Electronic (DAE) System Cell Controller Processor: Intel Core2 Clock Speed: 1.86 GHz Operating System: Windows XP Data Converter Features:Signal Amplifier, multiplexer, A/D converter, and control logic Software: DASY5 software Connecting Lines: Optical downlink for data and status info. Optical uplink for commands and clock



6. EUT Arrangement

6.1. WD RF Emission Measurements Reference and Plane

Figure 4 illustrates the references and reference plane that shall be used in the WD emissions measurement.

- The grid is 5 cm by 5 cm area that is divided into 9 evenly sized blocks or sub-grids.
- The grid is centered on the audio frequency output transducer of the WD (speaker or T-coil).

• The grid is located by reference to a reference plane. This reference plane is the planar area that contains the highest point in the area of the WD that normally rests against the user's ear

• The measurement plane is located parallel to the reference plane and 15 mm from it, out from the phone. The grid is located in the measurement plane.




Fig. 3 WD reference and plane for RF emission measurements



7. System Validation

7.1. Validation Procedure

Place a dipole antenna meeting the requirements given in ANSI C63.19 in the position normally occupied by the WD. The dipole antenna serves as a known source for an electrical output. Position the E-field probes so that:

• The probes and their cables are parallel to the coaxial feed of the dipole antenna

• The probe cables and the coaxial feed of the dipole antenna approach the measurement area from opposite directions

• The center point of the probe element(s) are 15 mm from the closest surface of the dipole elements.




Fig. 4 Dipole Validation Setup

7.2. Validation Result

	E-Field Scan					
Mode	Frequency Input Power Measured ¹ Target ² Deviation ³ Limit ⁴					
wode	(MHz)	(mW)	Value(dBV/m)	Value(dBV/m)	(%)	(%)
CW	835	100	43.27	40.72	6.3	±25
CW	1880	100	39.93	39.06	2.2	±25

Notes:

1. Please refer to the attachment for detailed measurement data and plot.

2. Target value is provided by SPEAD in the calibration certificate of specific dipoles.

3. Deviation (%) = 100 * (Measured value minus Target value) divided by Target value.

4. ANSI C63.19 requires values within \pm 25% are acceptable, of which 12% is deviation and 13% is measurement uncertainty. Values independently validated for the dipole actually used in the measurements should be used, when available.



8. Modulation Interference Factor (MIF)

The HAC Standard ANSI C63.19-2011 defines a new scaling using the Modulation Interference Factor (MIF) which replaces the need for the Articulation Weighting Factor (AWF) during the evaluation and is applicable to any modulation scheme.

The Modulation Interference factor (MIF, in dB) is added to the measured average E-field (in dBV/m) and converts it to the RF Audio Interference level (in dBV/m). This level considers the audible amplitude modulation components in the RF E-field. CW fields without amplitude modulation are assumed to not interfere with the hearing aid electronics. Modulations without time slots and low fluctuations at low frequencies have low MIF values, TDMA modulations with narrow transmission and repetition rates of few 100 Hz have high MIF values and give similar classifications as ANSI C63-2007.

Definitions

ER3D, E-field probes have a bandwidth <10 kHz and can therefore not evaluate the RF envelope in the full audio band. DASY52 is therefore using the "indirect" measurement method according to ANSI C63.19-2011 which is the primary method. These near field probes read the averaged E-field measurement. Especially for the new high peak-to-average (PAR) signal types, the probes shall be linearized by probe modulation response (PMR) calibration in order to not overestimate the field reading.

The evaluation method or the MIF is defined in ANSI C63.19-2011 section D.7. An RMS demodulated RF signal is fed to a spectral filter (similar to an A weighting filter) and forwarded to a temporal filter acting as a quasi-peak detector. The averaged output of these filtering is called to a 1 kHz 80% AM signal as reference. MIF measurement requires additional instrumentation and is not well suited for evaluation by the end user with reasonable uncertainty It may alternatively be determined through analysis and simulation, because it is constraint and characteristic for a communication signal. DASY52 uses well defined signals for PMR calibration. The MIF of these signals has been determined by simulation and is automatically applied.

MIF values were not tested by a probe or as specified in the standards but are based on analysis provided by SPEAG for all the air interfaces (GSM, WCDMA, CDMA, LTE). The data included in this report are for the worst case operating modes. The UIDs used are listed below:

UID	Communication System Name	MIF (dB)
10021	GSM-FDD (TDMA, GMSK)	3.63
10011	UMTS-FDD (WCDMA)	-27.23
10170	LTE-FDD(SC-FDMA, 1RB, 20MHz, 16-QAM)	-9.76
10176	LTE-FDD(SC-FDMA, 1RB, 10MHz, 16-QAM)	-9.76
10061	IEEE 802.11b WiFi 2.4 GHz (DSSS, 11 Mbps)	-2.02



No. I21N00886-HAC RF

A PMR calibrated probe is linearized for the selected waveform over the full dynamic range within the uncertainty specified in its calibration certificate. ER3D, EF3D and EU2D E-field probes have a bandwidth <10 kHz and can therefore not evaluate the RF envelope in the full audio band. DASY52 is therefore using the \indirect" measurement method according to ANSI C63.19-2011 which is the primary method. These near field probes read the averaged E-field measurement. Especially for the new high peak-to-average (PAR) signal types, the probes shall be linearized by PMR calibration in order to not overestimate the field reading.

The MIF measurement uncertainty is estimated as follows, for modulation frequencies from slotted waveforms with fundamental frequency and at least 2 harmonics within 10 kHz:

0.2 dB for MIF -7 to +5 dB, 0.5 dB for MIF -13 to +11 dB 1 dB for MIF > -20 dB



9. Evaluation for low-power exemption

9.1. Product testing threshold

There are two methods for exempting an RF air interface technology from testing. The first method requires evaluation of the MIF for the worst-case operating mode. An RF air interface technology of a device is exempt from testing when its average antenna input power plus its MIF is \leq 17 dBm for any of its operating modes. The second method does not require determination of the MIF. The RF emissions testing exemption shall be applied to an RF air interface technology in a device whose peak antenna input power, averaged over intervals \leq 50 µs20, is \leq 23 dBm. An RF air interface technology that is exempted from testing by either method shall be rated as M4.

The first method is used to be exempt from testing for the RF air interface technology in this report.

Band	power (dBm)	MIF (dB)	Sum (dBm)	HAC Test
GSM 850	33.0	3.63	36.63	Yes
GSM 1900	30.5	3.63	34.13	Yes
WCDMA B2	24.5	-27.23	-2.73	No
WCDMA B4	24.0	-27.23	-3.23	No
WCDMA B5	24.5	-27.23	-2.73	No
LTE Band 2	24.5	-9.76	14.74	No
LTE Band 4	24.5	-9.76	14.74	No
LTE Band 5	24.0	-9.76	14.24	No
LTE Band 7	24.0	-9.76	14.24	No
LTE Band 12	24.0	-9.76	14.24	No
LTE Band 13	24.0	-9.76	14.24	No
LTE Band 17	24.0	-9.76	14.24	No
LTE Band 28	24.5	-9.76	14.74	No
LTE Band 66	24.5	-9.76	14.74	No
WLAN 2.4G	13.5	-2.02	11.48	No

9.2. Conducted power

Note:

1. Power = Max tune-up limit



10. RF Test Procedures

The evaluation was performed with the following procedure:

- 1) Confirm proper operation of the field probe, probe measurement system and other instrumentation and the positioning system.
- 2) Position the WD in its intended test position. The gauge block can simplify this positioning.
- 3) Configure the WD normal operation for maximum rated RF output power, at the desired channel and other operating parameters (e.g., test mode), as intended for the test.
- 4) The center sub-grid shall centered on the center of the T-Coil mode axial measurement point or the acoustic output, as appropriate. Locate the field probe at the initial test position in the 50 mm by 50 mm grid, which is contained in the measurement plane. If the field alignment method is used, align the probe for maximum field reception.
- 5) Record the reading.
- 6) Scan the entire 50 mm by 50 mm region in equally spaced increments and record the reading at each measurement point. The distance between measurement points shall be sufficient to assure the identification of the maximum reading.
- 7) Identify the five contiguous sub-grids around the center sub-grid whose maximum reading is the lowest of all available choices. This eliminates the three sub-grids with the maximum readings. Thus, the six areas to be used to determine the WD's highest emissions are identified.
- 8) Identify the maximum field reading within the non-excluded sub-grids identified in Step 7)
- 9) Evaluate the MIF and add to the maximum steady-state rms field-strength reading to obtain the RF audio interference level..
- 10) Compare this RF audio interference level with the categories and record the resulting WD category rating.



Frequ	iency	Measured Value	Power Drift	Category
MHz	Channel	(dBV/m)	(dBV/m) (dB) C	
		GSM 85	50	
848.8	251	37.13	-0.07	M4 (see Fig A.1)
836.6	190	37.58	-0.04	M4 (see Fig A.2)
824.2	128	37.35	-0.02	M4 (see Fig A.3)
		GSM 19	00	
1909.8	810	27.71	0.01	M4 (see Fig A.4)
1880.0	661	28.07	0.03	M4 (see Fig A.5)
1850.2	512	28.53	0.02	M4 (see Fig A.6)

11. Measurement Results (E-Field)

12. ANSI C 63.19-2011 Limits

WD RF audio interference level categories in logarithmic units

Emission categories	< 960 MHz		
/	E-field e	emissions	
Category M1	50 to 55	dB (V/m)	
Category M2	45 to 50	dB (V/m)	
Category M3	40 to 45	dB (V/m)	
Category M4	< 40	dB (V/m)	
Emission categories	> 96	0 MHz	
/	E-field e	emissions	
Category M1	40 to 45	dB (V/m)	
Category M2	35 to 40	dB (V/m)	
Category M3	30 to 35 dB (V/m)		
Category M4	< 30 dB (V/m)		



13. Measurement Uncertainty

No.	Error source	Туре	Uncert ainty Value (%)	Prob. Dist.	k	C _i E	Standard Uncertainty (%) u_i^{+} (%) E	Degree of freedom V _{eff} or v _i	source
1	System repeatability	А	0.24	Ν	1	1	0.24	9	Measurement
Meas	surement System								
2	Probe Calibration	В	10.1	Ν	1	1	10.1	8	Manufacturer
3	Axial Isotropy	В	0.5	R	$\sqrt{3}$	1	0.5	∞	Cal report
4	Sensor Displacement	В	16.5	R	$\sqrt{3}$	1	9.5	8	Manufacturer
5	Boundary Effects	В	2.4	R	$\sqrt{3}$	1	1.4	8	Manufacturer
6	Linearity	В	0.6	R	$\sqrt{3}$	1	0.35	8	Cal report
7	Scaling to Peak Envolope Power	В	2.0	R	$\sqrt{3}$	1	1.2	∞	Standard
8	System Detection Limit	В	1.0	R	$\sqrt{3}$	1	0.6	∞	Manufacturer
9	Readout Electronics	В	0.3	N	1	1	0.3	∞	Manufacturer
10	Response Time	В	0.8	R	$\sqrt{3}$	1	0.5	∞	Manufacturer
11	Integration Time	В	2.6	R	$\sqrt{3}$	1	1.5	∞	Manufacturer
12	RF Ambient Conditions	В	3.0	R	$\sqrt{3}$	1	1.7	∞	Measurement
13	RF Reflections	В	12.0	R	$\sqrt{3}$	1	6.9	∞	Measurement
14	Probe Positioner	А	1.2	R	$\sqrt{3}$	1	0.7	∞	Manufacturer
15	Probe Positioning	А	4.7	R	$\sqrt{3}$	1	2.7	∞	Manufacturer
16	Extra. And Interpolation	В	1.0	R	$\sqrt{3}$	1	0.6	∞	Manufacturer
Test	Sample Related								
17	Device Positioning Vertical	В	4.7	R	$\sqrt{3}$	1	2.7	8	Manufacturer
18	Device Positioning Lateral	В	1.0	R	$\sqrt{3}$	1	0.6	8	Manufacturer
19	Device Holder and Phantom	В	2.4	R	$\sqrt{3}$	1	1.4	8	Manufacturer
20	Power Drift	В	5.0	R	$\sqrt{3}$	1	2.9	8	Measurement
Phar	Phantom and Setup related								
21	Phantom Thickness	В	2.4	R	$\sqrt{3}$	1	1.4	∞	Manufacturer
PMF	related								
22	Monitor amplitude	В	3.5	R	$\sqrt{3}$	1	2.02	8	Manufacturer
23	Setup repeatability	А	2.3	Ν	1	1	2.3	9	Manufacturer
24	Sensor amplitude	В	12	R	$\sqrt{3}$	1	6.93	∞	Manufacturer
	Combined standard uncertainty(%)					18.3			
	Expanded uncertainty (confidence interval of 95 %)	<i>U</i> _e	$= 2u_c$	Ν	k=	=2	36.6		



14. Main Test Instruments

No.	Name	Туре	Serial Number	Calibration Date	Valid Period
01	Signal Generator	E8257D	MY47461211	2021-01-15	One year
02	Power meter	E4418B	MY50000366	2020-12-13	Opener
03	Power sensor	E9304A	MY50000188	2020-12-13	One year
04	Amplifier	VTL5400	0404	/	
05	HAC Test Arch	N/A	1150	/	
06	DAE	DAE4	1527	2020-11-06	One year
07	E-Field Probe	ER3DV6	2424	2021-03-04	Three year
08	HAC Dipole	CD835V3	1165	2018-07-19	Three year
09	HAC Dipole	CD1880V3	1149	2018-07-19	Three year
10	BTS	CMW500	152499	2020-07-17	One year
11	Software	DASY5	52.8.8.1222	/	/

Table 14-1: List of Main Instruments

No. I21N00886-HAC RF



ANNEX A: RF Emission Test Plot


HAC RF E-Field GSM 850 High

Date: 2021-4-18 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: σ = 0 S/m, ϵ_r = 1; ρ = 1000 kg/m³ Communication System: UID 0, GSM Frequency: 848.8 MHz Duty Cycle: 1:8.3 Probe: ER3DV6 – SN2424 ConvF (1, 1, 1);

E Scan - ER3DV6 - 2011: 15 mm from Probe Center to the Device /Hearing Aid Compatibility

Test (101x101x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 60.64 V/m; Power Drift = -0.07 dB Applied MIF = 3.63 dB RF audio interference level = 37.13 dBV/m **Emission category: M4**

Grid 1 M4	Grid 2 M4	Grid 3 M4
35.93 dBV/m	36.63 dBV/m	36.27 dBV/m
Grid 4 M4	Grid 5 M4	Grid 6 M4
36.07 dBV/m	37.13 dBV/m	36.99 dBV/m
Grid 7 M4	Grid 8 M4	Grid 9 M4
36 dBV/m	37.22 dBV/m	37.04 dBV/m








HAC RF E-Field GSM 850 Middle

Date: 2021-4-18 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\epsilon_r = 1$; $\rho = 1000$ kg/m³ Communication System: UID 0, GSM Frequency: 836.6 MHz Duty Cycle: 1:8.3 Probe: ER3DV6 – SN2424 ConvF (1, 1, 1);

E Scan - ER3DV6 - 2011: 15 mm from Probe Center to the Device /Hearing Aid Compatibility

Test (101x101x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 61.39 V/m; Power Drift = -0.04 dB Applied MIF = 3.63 dB RF audio interference level = 37.58 dBV/m Emission category: M4

Grid 1 M4	Grid 2 M4	Grid 3 M4
35.72 dBV/m	36.67 dBV/m	36.45 dBV/m
Grid 4 M4	Grid 5 M4	Grid 6 M4
36.04 dBV/m	37.58 dBV/m	37.5 dBV/m
Grid 7 M4	Grid 8 M4	Grid 9 M4
36.25 dBV/m	37.84 dBV/m	37.73 dBV/m








HAC RF E-Field GSM 850 Low

Date: 2021-4-18 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\epsilon_r = 1$; $\rho = 1000$ kg/m³ Communication System: UID 0, GSM Frequency: 824.2 MHz Duty Cycle: 1:8.3 Probe: ER3DV6 – SN2424 ConvF (1, 1, 1);

E Scan - ER3DV6 - 2011: 15 mm from Probe Center to the Device /Hearing Aid Compatibility

Test (101x101x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 59.38 V/m; Power Drift = -0.02 dB Applied MIF = 3.63 dB RF audio interference level = 37.35 dBV/m Emission category: M4

Grid 1 M4	Grid 2 M4	Grid 3 M4
35.45 dBV/m	36.41 dBV/m	36.2 dBV/m
Grid 4 M4	Grid 5 M4	Grid 6 M4
35.81 dBV/m	37.35 dBV/m	37.26 dBV/m
Grid 7 M4	Grid 8 M4	Grid 9 M4
36.04 dBV/m	37.55 dBV/m	37.42 dBV/m







HAC RF E-Field GSM 1900 High

Date: 2021-4-18 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\epsilon_r = 1$; $\rho = 1000$ kg/m³ Communication System: UID 0, GSM Frequency: 1909.8 MHz Duty Cycle: 1:8.3 Probe: ER3DV6 – SN2424 ConvF (1, 1, 1);

E Scan - ER3DV6 - 2011: 15 mm from Probe Center to the Device /Hearing Aid Compatibility

Test (101x101x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 19.37 V/m; Power Drift = 0.01 dB Applied MIF = 3.63 dB RF audio interference level = 27.71 dBV/m Emission category: M4

 MIF scaled E-field

 Grid 1 M4
 Grid 2 M4
 Grid 3 M4


 26.93 dBV/m
 28.07 dBV/m
 27.59 dBV/m

 Grid 4 M4
 Grid 5 M4
 Grid 6 M4

 26 dBV/m
 27.71 dBV/m
 27.52 dBV/m

 Grid 7 M4
 Grid 8 M4
 Grid 9 M4

26.75 dBV/m



24.39 dBV/m 26.8 dBV/m






HAC RF E-Field GSM 1900 Middle

Date: 2021-4-18 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\epsilon_r = 1$; $\rho = 1000$ kg/m³ Communication System: UID 0, GSM Frequency: 1880 MHz Duty Cycle: 1:8.3 Probe: ER3DV6 – SN2424 ConvF (1, 1, 1);

E Scan - ER3DV6 - 2011: 15 mm from Probe Center to the Device /Hearing Aid Compatibility

Test (101x101x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 19.61 V/m; Power Drift = 0.03 dB Applied MIF = 3.63 dB RF audio interference level = 28.07 dBV/m **Emission category: M4**

MIF scaled E-field				
Grid 1 M4	Grid 2 M4	Grid 3 M4		
27.97 dBV/m	28.8 dBV/m	27.86 dBV/m		
Grid 4 M4	Grid 5 M4	Grid 6 M4		
26.74 dBV/m	28.07 dBV/m	27.67 dBV/m		
Grid 7 M4	Grid 8 M4	Grid 9 M4		
24.85 dBV/m	26.71 dBV/m	26.61 dBV/m		








HAC RF E-Field GSM 1900 Low

Date: 2021-4-18 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\epsilon_r = 1$; $\rho = 1000$ kg/m³ Communication System: UID 0, GSM Frequency: 1850.2 MHz Duty Cycle: 1:8.3 Probe: ER3DV6 – SN2424 ConvF (1, 1, 1);

E Scan - ER3DV6 - 2011: 15 mm from Probe Center to the Device /Hearing Aid Compatibility

Test (101x101x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 21.13 V/m; Power Drift = 0.02 dB Applied MIF = 3.63 dB RF audio interference level = 28.53 dBV/m Emission category: M4

Grid 1 M4	Grid 2 M4	Grid 3 M4
28.15 dBV/m	28.94 dBV/m	28.16 dBV/m
Grid 4 M4	Grid 5 M4	Grid 6 M4
27.23 dBV/m	28.53 dBV/m	28.16 dBV/m
Grid 7 M4	Grid 8 M4	Grid 9 M4
25.78 dBV/m	27.54 dBV/m	27.36 dBV/m







ANNEX B: System Validation Result


835 MHz

Date: 2021-4-18 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\varepsilon_r = 1$; $\rho = 1000$ kg/m³ Communication System: CW; Frequency: 835 MHz; Duty Cycle: 1:1 Probe: ER3DV6 – SN2424 ConvF (1, 1, 1);

E Scan - measurement distance from the probe sensor center to CD835 Dipole = 15mm / Hearing Aid Compatibility Test (41x361x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm

Device Reference Point: 0, 0, -6.3 mm Reference Value = 118.4 V/m; Power Drift = 0.06 dB Applied MIF = 0.00 dB RF audio interference level = 43.27 dBV/m **Emission category: M3**

Grid 1 M3	Grid 2 M3	Grid 3 M3
42.62 dBV/m	43.15 dBV/m	43.02 dBV/m
Grid 4 M4	Grid 5 M4	Grid 6 M4
38.16 dBV/m	38.55 dBV/m	38.47 dBV/m
Grid 7 M3	Grid 8 M3	Grid 9 M3
42.68 dBV/m	43.27 dBV/m	43.09 dBV/m




0 dB = 43.27 dBV/m




1880 MHz Date: 2021-4-18 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\varepsilon_r = 1$; $\rho = 1000$ kg/m³ Communication System: CW; Frequency: 1880 MHz; Duty Cycle: 1:1 Probe: ER3DV6 – SN2424 ConvF (1, 1, 1);

E Scan - measurement distance from the probe sensor center to CD1880 Dipole = 15mm / Hearing Aid Compatibility Test (41x181x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 139.2 V/m; Power Drift = 0.09 dB Applied MIF = 0.00 dB RF audio interference level = 39.93 dBV/m Emission category: M2

Grid 1 M2	Grid 2 M2	Grid 3 M2
39.48 dBV/m	39.93 dBV/m	39.81 dBV/m
Grid 4 M2	Grid 5 M2	Grid 6 M2
37.86 dBV/m	37.99 dBV/m	37.93 dBV/m
Grid 7 M2	Grid 8 M2	Grid 9 M2
39.42 dBV/m	39.88 dB V/m	39.76 dBV/m







ANNEX C: Probe Calibration Certificate

Schmid & Partner Engineering AG		ilac-MRA	Service suisse d'étalonnage
Coughausstrasse 43, 8004 Zu	rich, Switzerland	🐨 💓 s	Servizio svizzero di taratura Swiss Calibration Service
Accredited by the Swiss Accred	titation Service (SAS)		creditation No.: SCS 0108
he Swiss Accreditation Serv			creditation No.: 3CS 0108
Aultilateral Agreement for the			
Client TMC-SZ (Auc			: ER3-2424_Mar21
CALIBRATION	CERTIFICATE		
Object	ER3DV6- SN:242	4	Contract of the Party of
Calibration procedure(s)	QA CAL-02.v9, Q	A.CAI -25 v7	
		fure for E-field probes optimized t	for close near field
	evaluations in air	and for a mana probles optimized i	ion close near nea
	Contraction of the		and the second se
Calibration date:	March 4, 2021		
and the deliver state.	Warut 4, 2021		
The Device a line start is an internal two in the start of the			
This calibration ceraticate docu	iments the traceability to nation	nal standards, which realize the physical units bability are given on the following pages and	of measurements (SI)
The measurements and the un	certainties with confidence pro	bability are given on the following pages and	are part of the certificate.
		second as a surger manual bedres and	
		facility: environment temperature (22 ± 3) °C a	
All calibrations have been cond	fucted in the closed laboratory		
All calibrations have been cond	fucted in the closed laboratory		
All calibrations have been cond Calibration Equipment used (M	ducted in the closed laboratory		
All calibrations have been cond Calibration Equipment used (M Primary Standards	fucted in the closed laboratory		
All calibrations have been conc Calibration Equipment used (M Primary Standards Power meter NRP	International Internation Internation Internation	facility: environment temperature (22 ± 3) °C a Cal Date (Certificate No.) 01-Apr-20 (No. 217-03100/03101)	and humidily < 70%.
All calibrations have been conc Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-291	Increase of the closed laboratory	facility: environment temperature (22 ± 3) °C a Cal Date (Certificate No.)	and humidity < 70%.
All calibrations have been conc Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291	ID SN: 104778 SN: 103246	facility: environment temperature (22 ± 3) °C a Cal Date (Certificate No.) 01-Apr-20 (No. 217-03100/03101)	and humidity < 70%. Scheduled Calibration Apr-21
All calibrations have been cond Calibration Equipment used (M Primary Standards Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator	ID SN: 104778 SN: 104778 SN: 103244 SN: 003245 SN: CC2552 (20x)	facility: environment temperature (22 ± 3) °C a Cal Date (Certificate No.) 01-Apr-20 (No. 217-03100/03101) 01-Apr-20 (No. 217-03100) 01-Apr-20 (No. 217-03101) 31-Mar-20 (No. 217-03106)	and humidity < 70%. Scheduled Calibration Apr-21 Apr-21
All calibrations have been cond Calibration Equipment used (M Primary Standards Power sensor NRP- Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator DAE4	ID ID ID SN: 104778 SN: 103244 SN: 103246 SN: CC2552 (20x) SN: 789	facility: environment temperature (22 ± 3) °C a Cal Date (Certificate No.) 01-Apr-20 (No. 217-03100/03101) 01-Apr-20 (No. 217-03100) 01-Apr-20 (No. 217-03101)	and humidily < 70%. Scheduled Calibration Apr-21 Apr-21 Apr-21
All calibrations have been cond Calibration Equipment used (M Primary Standards Power sensor NRP- Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator DAE4	ID SN: 104778 SN: 104778 SN: 103244 SN: 003245 SN: CC2552 (20x)	facility: environment temperature (22 ± 3) °C a Cal Date (Certificate No.) 01-Apr-20 (No. 217-03100/03101) 01-Apr-20 (No. 217-03100) 01-Apr-20 (No. 217-03101) 31-Mar-20 (No. 217-03106)	Scheduled Calibration Apr-21 Apr-21 Apr-21 Apr-21
All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator DAE4 Reference Probe ER3DV6	ID ID SN: 104778 SN: 103244 SN: 02252 (20x) SN: 2328 SN: 2328	facility: environment temperature (22 ± 3) °C a Cal Date (Certificate No.) 01-Apr-20 (No. 217-03100/03101) 01-Apr-20 (No. 217-03100) 01-Apr-20 (No. 217-03101) 31-Mar-20 (No. 217-03106) 23-Dec-20 (No. DAE4-789_Dec20)	Apr-21 Apr-21 Apr-21 Apr-21 Apr-21 Dec-21
All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator DAE4 Reference Probe ER3DV6 Secondary Standards	ID ID ID SN: 104778 SN: 103244 SN: 103246 SN: CC2552 (20x) SN: 789	facility: environment temperature (22 ± 3) °C a Cal Date (Certificate No.) 01-Apr-20 (No. 217-03100/03101) 01-Apr-20 (No. 217-03100) 01-Apr-20 (No. 217-03101) 31-Mar-20 (No. 217-03106) 23-Dec-20 (No. DAE4-789_Dec20)	Apr-21 Apr-21 Apr-21 Apr-21 Apr-21 Dec-21
All calibrations have been cond Calibration Equipment used (M Primary Standards Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator DAE4 Reference Probe ER3DV6 Secondary Standards Power meter E4419B	Ucted in the closed laboratory IID SN: 104778 SN: 104778 SN: 103244 SN: 103244 SN: 002552 (20x) SN: CC2552 (20x) SN: 2328 ID SN: GB41293874	facility: environment temperature (22 ± 3) °C ± Cal Date (Certificate No.) 01-Apr-20 (No. 217-03100/03101) 01-Apr-20 (No. 217-03100) 01-Apr-20 (No. 217-03106) 23-Disc-20 (No. 217-03106) 23-Disc-20 (No. DAE4-789 Deb20) 05-Oct-20 (No. ER3-2328_Oct20)	and humidity < 70%. Scheduled Calibration Apr-21 Apr-21 Apr-21 Apr-21 Dec-21 Dec-21 Oct-21
All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator DAE4 Reference Probe ER3DV6 Secondary Standards Power meter E4419B	Ucted in the closed laboratory IBTE critical for calibration) ID SN: 104778 SN: 103244 SN: 103244 SN: 103245 SN: CC2552 (20x) SN: 789 SN: 2328 ID	facility: environment temperature (22 ± 3) °C ± Cal Date (Certificate No.) 01-Apr-20 (No. 217-03100/03101) 01-Apr-20 (No. 217-03100) 01-Apr-20 (No. 217-03100) 23-Dac-20 (No. 217-03106) 23-Dac-20 (No. 217-03106) 23-Dac-20 (No. DAE4-788_Dec20) 05-Oct-20 (No. ER3-2328_Oct20) Check Date (in house)	and humidity < 70%. Scheduled Calibration Apr-21 Apr-21 Apr-21 Dec-21 Oct-21 Oct-21 Scheduled Check
All calibrations have been cond Calibration Equipment used (M Primary Standards Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator DAE4 Reference Probe ER3DV6 Secondary Standards Power meter E44198 Power sensor E4412A	Ucted in the closed laboratory IID SN: 104778 SN: 104778 SN: 103244 SN: 103244 SN: 002552 (20x) SN: CC2552 (20x) SN: 2328 ID SN: GB41293874	facility: environment temperature (22 ± 3) °C a Cal Date (Certificate No.) 01-Apr-20 (No. 217-03100/03101) 01-Apr-20 (No. 217-03100) 01-Apr-20 (No. 217-03101) 31-Mar-20 (No. 217-03106) 23-Dec-20 (No. DAE4-769_Dec20) 05-Oct-20 (No. EE3-2328_Oct20) Check Date (in house) 06-Apr-16 (in house)	and humidily < 70%. Scheduled Calibration Apr-21 Apr-21 Apr-21 Apr-21 Dec-21 Oct-21 Oct-21 Scheduled Check In house check: Jun-22
All calibrations have been cond Calibration Equipment used (M Primary Standards Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator DAE4 Reference Probe ER3DV6 Secondary Standards Power meter E44198 Power sensor E4412A	Aucted in the closed laboratory INTE oritical for calibration) ID SN: 104778 SN: 103244 SN: 002456 SN: CC2552 (20x) SN: 789 SN: 2328 ID SN: GB41293674 SN: MY41498087	facility: environment temperature (22 ± 3) °C a Cal Date (Certificate No.) 01-Apr-20 (No. 217-03100/03101) 01-Apr-20 (No. 217-03100) 01-Apr-20 (No. 217-03101) 31-Mar-20 (No. 217-03106) 23-Dec-20 (No. DAE4-789_Dec20) 05-Oct-20 (No. ER3-2328_Oct20) Check Date (in house) 06-Apr-16 (in house check Jun-20) 06-Apr-16 (in house check Jun-20)	and humidily < 70%. Scheduled Calibration Apr-21 Apr-21 Apr-21 Dec-21 Dec-21 Dec-21 Oct-21 Scheduled Check In house check: Jun-22 In house check: Jun-22
All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator DAE4 Reference Probe ER3DV6 Secondary Standards Power meter E44198 Power sensor E4412A Power sensor E4412A RF generator HP 8648C	Aucted in the closed laboratory ID SN: 104778 SN: 103244 SN: 103245 SN: 202552 (20x) SN: 789 SN: 2328 ID SN: GB41293674 SN: MY41498087 SN: 000110210	facility: environment temperature (22 ± 3) °C a Cal Date (Certificate No.) 01-Apr-20 (No. 217-03100/03101) 01-Apr-20 (No. 217-03100) 01-Apr-20 (No. 217-03101) 31-Mar-20 (No. 217-03106) 23-Dec-20 (No. DAE4-789_Dec20) 05-Oct-20 (No. ER3-2328_Oct20) Check Date (in house) 06-Apr-16 (in house check Jun-20) 06-Apr-16 (in house check Jun-20)	and humidity < 70%. Scheduled Calibration Apr-21 Apr-21 Apr-21 Dec-21 Oct-21 Oct-21 Scheduled Check In house check: Jun-22 In house check: Jun-22
All calibrations have been cond Calibration Equipment used (M Primary Standards Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator DAE4 Reference Probe ER3DV6 Secondary Standards Power meter E44198 Power sensor E4412A Referenca E4412A RF generator HP 8648C	ID ID SN: 104778 IO SN: 103244 IO SN: 103244 IO SN: 103245 IO SN: 103246 IO SN: 103247 IO SN: 103248 IO SN: 002552 (2Dx) IO SN: 2328 ID SN: GB41293674 SN: MY41498087 SN: US3642U01700 SN: US3642U01700 SN: US41060477 IO	facility: environment temperature (22 ± 3) °C ± Cal Date (Certificate No.) 01-Apr-20 (No. 217-03100/03101) 01-Apr-20 (No. 217-03100) 01-Apr-20 (No. 217-03100) 01-Apr-20 (No. 217-03106) 23-Dec-20 (No. 217-03106) 23-Dec-20 (No. 217-03106) 23-Dec-20 (No. 217-03106) 23-Dec-20 (No. 217-03106) 23-Dec-20 (No. ER3-2328_Oct20) 06-Apr-16 (in house) 06-Apr-16 (in house check Jun-20) 06-Apr-16 (in house check Jun-20) 04-Aug-99 (in house check Jun-20) 04-Aug-99 (in house check Jun-20) 31-Mar-14 (in house check Oct-20)	and humidily < 70%. Scheduled Calibration Apr-21 Apr-21 Apr-21 Dec-21 Oct-21 Oct-21 Scheduled Check In house check: Jun-22 In house check: Jun-22
All calibrations have been cond Calibration Equipment used (M Primary Standards Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator DAE4 Reference Probe ER3DV6 Secondary Standards Power sensor E44198 Power sensor E44198 Power sensor E4412A RF generator HP 8648C Network Analyzer E8358A	Aucted in the closed laboratory ID SN: 104778 SN: 103244 SN: 103245 SN: 202552 (20x) SN: 789 SN: 2328 ID SN: 6841293674 SN: MY41498087 SN: 000110210 SN: US3642001700 SN: US3642001700 SN: US361080477 Name	facility: environment temperature (22 ± 3) °C a Cal Date (Certificate No.) 01-Apr-20 (No. 217-03100/03101) 01-Apr-20 (No. 217-03100) 01-Apr-20 (No. 217-03101) 31-Mar-20 (No. 217-03106) 23-Dec-20 (No. DAE4-789_Dec20) 05-Oct-20 (No. ER3-2328_Oct20) 05-Oct-20 (No. ER3-2328_Oct20) 06-Apr-16 (in house) 06-Apr-16 (in house check Jun-20) 06-Apr-16 (in house check Jun-20) 06-Apr-16 (in house check Jun-20) 04-Aug-99 (in house check Jun-20) 04-Aug-99 (in house check Jun-20) 31-Mar-14 (in house check Jun-20) Function	and humidily < 70%. Scheduled Calibration Apr-21 Apr-21 Apr-21 Dec-21 Dec-21 Oct-21 Scheduled Check In house check: Jun-22 In house check: Jun-22 In house check: Jun-22 In house check: Jun-22
All calibrations have been cond Calibration Equipment used (M Primary Standards Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator DAE4 Reference Probe ER3DV6 Secondary Standards Power meter E44198 Power sensor E4412A Power sensor E4412A RF generator HP 8648C Network Analyzer E8358A	ID ID SN: 104778 IO SN: 103244 IO SN: 103244 IO SN: 103245 IO SN: 103246 IO SN: 103247 IO SN: 103248 IO SN: 002552 (2Dx) IO SN: 2328 ID SN: GB41293674 SN: MY41498087 SN: US3642U01700 SN: US3642U01700 SN: US41060477 IO	facility: environment temperature (22 ± 3) °C ± Cal Date (Certificate No.) 01-Apr-20 (No. 217-03100/03101) 01-Apr-20 (No. 217-03100) 01-Apr-20 (No. 217-03100) 01-Apr-20 (No. 217-03106) 23-Dec-20 (No. 217-03106) 23-Dec-20 (No. 217-03106) 23-Dec-20 (No. 217-03106) 23-Dec-20 (No. 217-03106) 23-Dec-20 (No. ER3-2328_Oct20) 06-Apr-16 (in house) 06-Apr-16 (in house check Jun-20) 06-Apr-16 (in house check Jun-20) 04-Aug-99 (in house check Jun-20) 04-Aug-99 (in house check Jun-20) 31-Mar-14 (in house check Oct-20)	and humidily < 70%. Scheduled Calibration Apr-21 Apr-21 Apr-21 Apr-21 Dec-21 Oct-21 Scheduled Check In house check: Jun-22 In house check: Jun-22
All calibrations have been cond Calibration Equipment used (M Primary Standards Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator DAE4 Reference Probe ER3DV6 Secondary Standards Power meter E44198 Power sensor E4412A Power sensor E4412A RF generator HP 8648C Network Analyzer E8358A	Aucted in the closed laboratory ID SN: 104778 SN: 103244 SN: 103245 SN: 202552 (20x) SN: 789 SN: 2328 ID SN: 6841293674 SN: MY41498087 SN: 000110210 SN: US3642001700 SN: US3642001700 SN: US361080477 Name	facility: environment temperature (22 ± 3) °C a Cal Date (Certificate No.) 01-Apr-20 (No. 217-03100/03101) 01-Apr-20 (No. 217-03100) 01-Apr-20 (No. 217-03101) 31-Mar-20 (No. 217-03106) 23-Dec-20 (No. DAE4-789_Dec20) 05-Oct-20 (No. ER3-2328_Oct20) 05-Oct-20 (No. ER3-2328_Oct20) 06-Apr-16 (in house) 06-Apr-16 (in house check Jun-20) 06-Apr-16 (in house check Jun-20) 06-Apr-16 (in house check Jun-20) 04-Aug-99 (in house check Jun-20) 04-Aug-99 (in house check Jun-20) 31-Mar-14 (in house check Jun-20) Function	and humidily < 70%. Scheduled Calibration Apr-21 Apr-21 Apr-21 Apr-21 Dec-21 Oct-21 Scheduled Check In house check: Jun-22 In house check: Jun-22
All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator DAE4 Reference Probe ER3DV6 Secondary Standards Power meter E44198 Power sensor E4412A RF generator HP 8648C Network Analyzer E8358A Calibrated by:	Aucted in the closed laboratory IBTE critical for calibration) ID SN: 104778 SN: 103244 SN: 103244 SN: 103244 SN: 022552 (20x) SN: 2328 ID SN: 2328 ID SN: 0841293874 SN: MY41498087 SN: 000110210 SN: US36421001700 SN: US36421001700 SN: US36421001700 SN: US36421001700	facility: environment temperature (22 ± 3) C ± Cal Date (Certificate No.) 01-Apr-20 (No. 217-03100/03101) 01-Apr-20 (No. 217-03100) 01-Apr-20 (No. 217-03100) 31-Mar-20 (No. 217-03106) 23-Dec-20 (No. 217-03106) 06-Apr-16 (in house check Jun-20) 06-Apr-16 (in house check Jun-20) 06-Apr-16 (in house check Jun-20) 04-Aug-99 (in house check Jun-20) 31-Mar-14 (in house check Oct-20) Function Laboratory Technician	and humidily < 70%. Scheduled Calibration Apr-21 Apr-21 Apr-21 Apr-21 Dec-21 Oct-21 Scheduled Check In house check: Jun-22 In house check: Jun-22
All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator DAE4 Reference Probe ER3DV6 Secondary Standards Power meter E44198 Power sensor E4412A RF generator HP 8648C Network Analyzer E8358A Calibrated by:	Aucted in the closed laboratory ID SN: 104778 SN: 103244 SN: 103245 SN: 202552 (20x) SN: 789 SN: 2328 ID SN: 6841293674 SN: MY41498087 SN: 000110210 SN: US3642001700 SN: US3642001700 SN: US361080477 Name	facility: environment temperature (22 ± 3) °C a Cal Date (Certificate No.) 01-Apr-20 (No. 217-03100/03101) 01-Apr-20 (No. 217-03100) 01-Apr-20 (No. 217-03101) 31-Mar-20 (No. 217-03106) 23-Dec-20 (No. DAE4-789_Dec20) 05-Oct-20 (No. ER3-2328_Oct20) 05-Oct-20 (No. ER3-2328_Oct20) 06-Apr-16 (in house) 06-Apr-16 (in house check Jun-20) 06-Apr-16 (in house check Jun-20) 06-Apr-16 (in house check Jun-20) 04-Aug-99 (in house check Jun-20) 04-Aug-99 (in house check Jun-20) 31-Mar-14 (in house check Jun-20) Function	and humidily < 70%. Scheduled Calibration Apr-21 Apr-21 Apr-21 Apr-21 Dec-21 Oct-21 Scheduled Check In house check: Jun-22 In house check: Jun-22
All calibrations have been cond Calibration Equipment used (M Primary Standards Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator DAE4 Reference Probe ER3DV6 Secondary Standards Power meter E44198 Power sensor E4412A Referencator HP 8648C Network Analyzer E8358A	Aucted in the closed laboratory IBTE critical for calibration) ID SN: 104778 SN: 103244 SN: 103244 SN: 103244 SN: 022552 (20x) SN: 2328 ID SN: 2328 ID SN: 0841293874 SN: MY41498087 SN: 000110210 SN: US36421001700 SN: US36421001700 SN: US36421001700 SN: US36421001700	facility: environment temperature (22 ± 3) C ± Cal Date (Certificate No.) 01-Apr-20 (No. 217-03100/03101) 01-Apr-20 (No. 217-03100) 01-Apr-20 (No. 217-03100) 31-Mar-20 (No. 217-03106) 23-Dec-20 (No. 217-03106) 06-Apr-16 (in house check Jun-20) 06-Apr-16 (in house check Jun-20) 06-Apr-16 (in house check Jun-20) 04-Aug-99 (in house check Jun-20) 31-Mar-14 (in house check Oct-20) Function Laboratory Technician	and humidily < 70%. Scheduled Calibration Apr-21 Apr-21 Apr-21 Apr-21 Dec-21 Oct-21 Scheduled Check In house check: Jun-22 In house check: Jun-22
All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator DAE4 Reference Probe ER3DV6 Secondary Standards Power sensor E44198 Power sensor E44198 Power sensor E4412A RF generator HP 8648C Network Analyzer E8358A Calibrated by:	Aucted in the closed laboratory IBTE critical for calibration) ID SN: 104778 SN: 103244 SN: 103244 SN: 103244 SN: 022552 (20x) SN: 2328 ID SN: 2328 ID SN: 0841293874 SN: MY41498087 SN: 000110210 SN: US36421001700 SN: US36421001700 SN: US36421001700 SN: US36421001700	facility: environment temperature (22 ± 3) C ± Cal Date (Certificate No.) 01-Apr-20 (No. 217-03100/03101) 01-Apr-20 (No. 217-03100) 01-Apr-20 (No. 217-03100) 31-Mar-20 (No. 217-03106) 23-Dec-20 (No. 217-03106) 06-Apr-16 (in house check Jun-20) 06-Apr-16 (in house check Jun-20) 06-Apr-16 (in house check Jun-20) 04-Aug-99 (in house check Jun-20) 31-Mar-14 (in house check Oct-20) Function Laboratory Technician	and humidily < 70%. Scheduled Calibration Apr-21 Apr-21 Apr-21 Apr-21 Dec-21 Oct-21 Scheduled Check In house check: Jun-22 In house check: Jun-22
All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator DAE4 Reference Probe ER3DV6 Secondary Standards Power sensor E44198 Power sensor E44198 Power sensor E4412A RF generator HP 8648C Network Analyzer E8358A Calibrated by:	Aucted in the closed laboratory IBTE critical for calibration) ID SN: 104778 SN: 103244 SN: 103244 SN: 103244 SN: 022552 (20x) SN: 2328 ID SN: 2328 ID SN: 0841293874 SN: MY41498087 SN: 000110210 SN: US36421001700 SN: US36421001700 SN: US36421001700 SN: US36421001700	facility: environment temperature (22 ± 3) C ± Cal Date (Certificate No.) 01-Apr-20 (No. 217-03100/03101) 01-Apr-20 (No. 217-03100) 01-Apr-20 (No. 217-03100) 31-Mar-20 (No. 217-03106) 23-Dec-20 (No. 217-03106) 06-Apr-16 (in house check Jun-20) 06-Apr-16 (in house check Jun-20) 06-Apr-16 (in house check Jun-20) 04-Aug-99 (in house check Jun-20) 31-Mar-14 (in house check Oct-20) Function Laboratory Technician	and humidily < 70%. Scheduled Calibration Apr-21 Apr-21 Apr-21 Apr-21 Dec-21 Oct-21 Scheduled Check In house check: Jun-22 In house check: Jun-22

Certificate No: ER3-2424_Mar21

Page 1 of 9



No. I21N00886-HAC RF

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Schweizerischer Kalibrierdienst Service suisse d'étalonnage Servizio svizzero di taratura Swiss Calibration Service

Accreditation No.: SCS 0108

S

С

S

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Glossary:

NORMx.y.z DCP CF A, B, C, D En Ep Polarization ϕ	sensitivity in free space diode compression point crest factor (1/duty_cycle) of the RF signal modulation dependent linearization parameters incident E-field orientation normal to probe axis incident E-field orientation parallel to probe axis or otation around probe axis
Polarization 3	9 rotation around an axis that is in the plane normal to probe axis (at measurement center),
Connector Angle	i.e., 9 = 0 is normal to probe axis information used in DASY system to align probe sensor X to the robot coordinate system

Calibration is Performed According to the Following Standards:

- a) IEEE Std 1309-2005, * IEEE Standard for calibration of electromagnetic field sensors and probes, excluding
- antennas, from 9 kHz to 40 GHz", December 2005
 b) CTIA Test Plan for Hearing Aid Compatibility, Rev 3.1.1, May 2017

- Methods Applied and Interpretation of Parameters:
 NORMx.y.z: Assessed for E-field polarization 9 = 0 for XY sensors and 9 = 90 for Z sensor (f ≤ 900 MHz in TEM-cell; f > 1800 MHz: R22 waveguide).
 - NORM(f)x,y,z = NORMx,y,z * frequency_response (see Frequency Response Chart).
 - DCPx,y,z: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
 - PAR: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
 - Ax,y,z; Bx,y,z; Cx,y,z; Dx,y,z; VRx,y,z; A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
 - Spherical isotropy (3D deviation from isotropy): in a locally homogeneous field realized using an open waveguide setup.
 - Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
 - Connector Angle: The angle is assessed using the information gained by determining the NORMx (no uncertainty required).

Certificate No: ER3-2424_Mar21

Page 2 of 9



March 4, 2021

DASY/EASY - Parameters of Probe: ER3DV6 - SN:2424

Basic Calibration Parameters

	Sensor X	Sensor Y	Sensor Z	Unc (k=2)
Norm (µV/(V/m) ²)	1.53	1.55	1.83	± 10.1 %
DCP (mV) ^B	99.3	99.8	101.3	

Calibration results for Frequency Response (30 MHz - 3 GHz)

Frequency MHz	Target E-Field V/m	Measured E-field (En) V/m	Deviation E-normal in %	Measured E-field (Ep) V/m	Deviation E-normal in %	Unc (k=2)
30	77.1	76.6	-0.7%	77.4	0.4%	± 5.1%
100	77.2	78.5	1.8%	77.9	0.9%	± 5.1%
450	77.2	78.6	1.9%	77.8	0.8%	± 5.1 %
600	77.0	78.2	1.5%	77.5	0.6%	± 5.1 %
750	77.0	78.1	1.5%	77.5	0.7%	± 5.1 %
1800	143.0	141.7	-0.9%	141.1	-1.3%	± 5.1 %
2000	135.1	134.4	-0.5%	133.5	-1.2%	± 5.1 %
2200	127.7	126.2	-1.2%	127.5	-0.1%	± 5.1 %
2500	125.5	126.0	0.4%	126.8	1.1%	± 5.1 %
3000	79.4	78.2	-1.6%	81.3	2.4%	± 5.1 %

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

⁶ Numerical linearization parameter: uncertainty not required.
⁶ Uncertainty is determined using the max, deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

Certificate No: ER3-2424_Mar21

Page 3 of 9



March 4, 2021

DASY/EASY - Parameters of Probe: ER3DV6 - SN:2424

Calibration Results for Modulation Response

UID	Communication System Name		A dB	B dBõV	C	D dB	VR mV	Max dev.	Unc ^L (k=2)
0	CW	X	0.0	0.0	1.0	0.00	207.1	±3.5 %	± 4.7 %
		Y	0.0	0.0	1.0		194.8	0000000000	The sector of
		Z	0.0	0.0	1.0		208.5		
10021- DAC	GSM-FDD (TDMA, GMSK)	X	13.38	91.7	25.7	9.39	127.8	±3.0 %	±.4.7 %
		Y	20.31	99.9	28.1		115.1		
		Z	25.39	99.9	28.1		145.9		
10061- CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 11 Mbps)	×	4,95	75.3	21.8	3.60	114.8	±2.2 %	± 4.7 %
		Y	4.11	72.3	20.5		106.0		
_		Z	5.66	76.6	21.8		117.0		
10077- IEEE 80 CAB (DSSS/0	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 54 Mbps)	X	12.21	74.8	27.6	11.00	104.9	±2.2 %	±4.7 %
		Y.	13,33	78.3	29.7		144.6		
Charles and a second second		Z	12.02	73.8	26.5		107.7		
10172- CAG	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK)	X	10.38	84.9	32.0	9.21	140,1	±2.5 %	± 4.7 %
		Y	8.50	78.8	28.9		126.9		
		Z	11.14	85.0	31.1		148.0		
	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM)	×	10.60	84.6	31.9	9.48	139.5	±2.5 %	±4.7 %
		Y	9.11	80.2	29.6		127.0		
		Z	12.00	86,6	31.9		148.3		
10295- AAB	CDMA2000, RC1, SO3, 1/8th Rate 25 fr.	×	16,51	99.7	40.5	12.49	113.2	±3.5 %	± 4.7 %
		Y	15.91	100.0	40.9		101.3		
		Z	18.42	100.0	39.2		126.2		

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

⁸ Numerical linearization parameter: uncortaintly not required.
⁹ Uncortaintly is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

Certificate No: ER3-2424_Mar21

Page 4 of 9



March 4, 2021

DASY/EASY - Parameters of Probe: ER3DV6 - SN:2424

Sensor Frequency Model Parameters

	Sensor X	Sensor Y	Sensor Z
Frequency Corr. (LF)	-1.78	-1.32	0.22
Frequency Corr. (HF)	0.00	0.00	0.00

Other Probe Parameters

Sensor Arrangement	Rectangular
Connector Angle (*)	165.5
Mechanical Surface Detection Mode	enabled
Optical Surface Detection Mode	disabled
Probe Overall Length	337 mm
Probe Body Diameter	10 mm
Tip Length	10 mm
Tip Diameter	8 mm
Probe Tip to Sensor X Calibration Point	2.5 mm
Probe Tip to Sensor Y Calibration Point	2.5 mm
Probe Tip to Sensor Z Calibration Point	2.5 mm

Certificate No: ER3-2424_Mar21

Page 5 of 9




March 4, 2021



Receiving Pattern (ϕ), $\vartheta = 0^{\circ}$

Receiving Pattern (ϕ), ϑ = 90°




Certificate No: ER3-2424_Mar21

Page 6 of 9




March 4, 2021



Uncertainty of Axial Isotropy Assessment: ± 0.5% (k=2)

Receiving Pattern (ϕ), $\vartheta = 90^{\circ}$



Uncertainty of Axial Isotropy Assessment: ± 0.5% (k=2)

Certificate No: ER3-2424_Mar21


Page 7 of 9



No. I21N00886-HAC RF


ER3DV6 - SN:2424

March 4, 2021



Certificate No: ER3-2424_Mar21


Page 8 of 9





ER3DV6 - SN:2424

March 4, 2021



Uncertainty of Spherical Isotropy Assessment: ± 2.6% (k=2)

Certificate No: ER3-2424_Mar21

Page 9 of 9





ANNEX D: DAE Calibration Certificate

Tel: +86-10-623 E-mail: cttl <i>a</i> ch	eyuan Road, Haidian Dis 04633-2512 Fax: -	e a g TON LABORATORY strict, Beijing, 100191, China +86-10-62304633-2504 //www.chinattl.cn	Certificate	No: Z20-60433	中国认可 国际互认 校准 CALIBRATIO CNAS L057(
CALIBRATION	CERTIFICAT	ſE			
Object	DAE4	- SN: 1527			
Calibration Procedure(s)	FF-Z1	1-002-01 ation Procedure for the)	e Data Acquisi	tion Electronics	-
Calibration date:	Noven	nber 06, 2020			
pages and are part of the All calibrations have be humidity<70%. Calibration Equipment us Primary Standards	een conducted in sed (M&TE critical			iment temperatur	
Process Calibrator 753	1971018	16-Jun-20 (CTTL, No.J		Jun-2	1
	Name	Function		Signature	
Calibrated by:	Yu Zongying	SAR Test Enginee	er	1-17	D
Reviewed by:	Lin Hao	SAR Test Enginee	er	林光	2
Approved by:	Qi Dianyuan	SAR Project Lead	er	ton	
This calibration certificate	e shall not be repro	oduced except in full with		ssued: November roval of the labora	

Certificate No: Z20-60433

Page 1 of 3





Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2512 Fax: +86-10-62304633-2504 E-mail: ettl@chinattl.com Http://www.chinattl.en

Glossary:

DAE Connector angle data acquisition electronics information used in DASY system to align probe sensor X to the robot coordinate system.

Methods Applied and Interpretation of Parameters:

- DC Voltage Measurement: Calibration Factor assessed for use in DASY system by comparison with a calibrated instrument traceable to national standards. The figure given corresponds to the full scale range of the voltmeter in the respective range.
- *Connector angle*: The angle of the connector is assessed measuring the angle mechanically by a tool inserted. Uncertainty is not required.
- The report provide only calibration results for DAE, it does not contain other performance test results.

Certificate No: Z20-60433

Page 2 of 3





 Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China

 Tel: +86-10-62304633-2512
 Fax: +86-10-62304633-2504

 E-mail: cttl@chinattl.com
 Http://www.chinattl.cn

DC Voltage Measurement

A/D - Converter Res	solution nomi	nal		
High Range:	1LSB =	6.1µV,	full range =	-100+300 mV
Low Range:	1LSB =	61nV,	full range =	-1+3mV
DASY measuremen	t parameters	: Auto Zero 7	Time: 3 sec; Meas	uring time: 3 sec

Calibration Factors	х	Y	Z
High Range	403.863 ± 0.15% (k=2)	403.582 ± 0.15% (k=2)	403.801 ± 0.15% (k=2)
Low Range	3.95875 ± 0.7% (k=2)	3.98892 ± 0.7% (k=2)	3.96720 ± 0.7% (k=2)

Connector Angle

Connector Angle to be used in DASY system	223.5°±1°
---	-----------

Certificate No: Z20-60433

Page 3 of 3





ANNEX E: Dipole Calibration Certificate

Dipole 835 MHz

Engineering AG Ighausstrasse 43, 8004 Zurich, S	witzerland	C S	Service suisse d'étalonnage Servizio svizzero di taratura Swiss Calibration Service
credited by the Swiss Accreditation e Swiss Accreditation Service is			creditation No.: SCS 0108
ultilateral Agreement for the reco	gnition of calibration c	ertificates	
ent CTTL (Auden)		Certificate No:	CD835V3-1165_Jul18
CALIBRATION CI	ERTIFICATE		
Dbject	CD835V3 - SN: 1	165	
Calibration procedure(s)	QA CAL-20.v6 Calibration proces	dure for dipoles in air	
Calibration date:	July 19, 2018		
The measurements and the uncerta	ainties with confidence pr ed in the closed laborator	onal standards, which realize the physical unirobability are given on the following pages an ry facility: environment temperature $(22 \pm 3)^{\circ}$	id are part of the certificate.
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE	ainties with confidence pr ed in the closed laborator E critical for calibration)	robability are given on the following pages an ry facility: environment temperature (22 \pm 3)°(id are part of the certificate.
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards	ainties with confidence pr ed in the closed laborator	robability are given on the following pages an	id are part of the certificate. C and humidity < 70%.
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power meter NRP	ainties with confidence pr ed in the closed laborator E critical for calibration)	robability are given on the following pages an ry facility: environment temperature (22 ± 3)°0 Cal Date (Certificate No.)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-Z91	ainties with confidence pr ed in the closed laborator E critical for calibration) ID # SN: 104778	robability are given on the following pages an ny facility: environment temperature (22 ± 3)°C Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-19
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91	ainties with confidence pr ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244	Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator	ainties with confidence pr ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245	robability are given on the following pages an ry facility: environment temperature (22 ± 3)°(Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Apr-19
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination	ainties with confidence pr ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 5058 (20k) SN: 5047.2 / 06327 SN: 4013	robability are given on the following pages an ry facility: environment temperature (22 ± 3)°(Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. EF3-4013_Mar18)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Mar-19
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Probe EF3DV3 Probe H3DV6	ainties with confidence pr ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 5058 (20k) SN: 5058 (20k) SN: 5047.2 / 06327 SN: 4013 SN: 6065	robability are given on the following pages an ry facility: environment temperature (22 ± 3)°(<u>Cal Date (Certificate No.)</u> 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. EF3-4013_Mar18) 30-Dec-17 (No. H3-6065_Dec17)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Mar-19 Dec-18
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Probe EF3DV3 Probe H3DV6	ainties with confidence pr ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 5058 (20k) SN: 5047.2 / 06327 SN: 4013	robability are given on the following pages an ry facility: environment temperature (22 ± 3)°(Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. EF3-4013_Mar18)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Mar-19
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Probe EF3DV3 Probe H3DV6 DAE4	ainties with confidence pr ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 5058 (20k) SN: 5058 (20k) SN: 5047.2 / 06327 SN: 4013 SN: 6065	robability are given on the following pages an ry facility: environment temperature (22 ± 3)°(<u>Cal Date (Certificate No.)</u> 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. EF3-4013_Mar18) 30-Dec-17 (No. H3-6065_Dec17)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Mar-19 Dec-18
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Probe EF3DV3 Probe H3DV6 DAE4 Secondary Standards	ainties with confidence pr ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 6065 SN: 781	Cal Date (Certificate No.) O4-Apr-18 (No. 217-02672/02673) O4-Apr-18 (No. 217-02672/02673) O4-Apr-18 (No. 217-02672) O4-Apr-18 (No. 217-02673) O4-Apr-18 (No. 217-02683) O5-Mar-18 (No. EF3-4013_Mar18) 30-Dec-17 (No. H3-6065_Dec17) 17-Jan-18 (No. DAE4-781_Jan18)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Mar-19 Dec-18 Jan-19
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Probe EF3DV3 Probe H3DV6 DAE4 Secondary Standards Power meter Agilent 4419B	ainties with confidence pre- ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 4013 SN: 6065 SN: 781 ID #	Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. DAE4-781_Jan18) Check Date (in house) 09-Oct-09 (in house check Oct-17) 05-Jan-10 (in house check Oct-17)	Id are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Mar-19 Dec-18 Jan-19 Scheduled Check In house check: Oct-20 In house check: Oct-20
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Probe EF3DV3 Probe H3DV6 DAE4 Secondary Standards Power meter Agilent 4419B Power sensor HP E4412A	ainties with confidence pre- ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103245 SN: 5058 (20k) SN: 5058 (20k) SN: 5047.2 / 06327 SN: 4013 SN: 6065 SN: 781 ID # SN: GB42420191 SN: US38485102 SN: US37295597	Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. DAE4-781_Jan18) Check Date (in house) 09-Oct-09 (in house check Oct-17) 05-Jan-10 (in house check Oct-17) 09-Oct-09 (in house check Oct-17)	Id are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Mar-19 Dec-18 Jan-19 Scheduled Check In house check: Oct-20 In house check: Oct-20 In house check: Oct-20
The measurements and the uncertained in the uncerta	ainties with confidence pre- ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103245 SN: 5058 (20k) SN: 5058 (20k) SN: 5047.2 / 06327 SN: 4013 SN: 6065 SN: 781 ID # SN: GB42420191 SN: US38485102 SN: US37295597 SN: 832283/011	Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. 217-02672) 04-Apr-18 (No. 217-02683) 05-Jan-10 (No. H3-6065 Dec17) 17-Jan-18 (No. DAE4-781_Jan18) Check Date (in house) 09-Oct-09 (in house check Oct-17) 05-Jan-10 (in house check Oct-17) 09-Oct-09 (in house check Oct-17) 27-Aug-12 (in house check Oct-17)	Id are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Mar-19 Dec-18 Jan-19 Scheduled Check In house check: Oct-20 In house check: Oct-20
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Probe EF3DV3 Probe H3DV6 DAE4 Secondary Standards Power meter Agilent 4419B Power sensor HP E4412A Power sensor HP 8482A RF generator R&S SMT-06	ainties with confidence pre- ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103245 SN: 5058 (20k) SN: 5058 (20k) SN: 5047.2 / 06327 SN: 4013 SN: 6065 SN: 781 ID # SN: GB42420191 SN: US38485102 SN: US37295597 SN: 832283/011	Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. DAE4-781_Jan18) Check Date (in house) 09-Oct-09 (in house check Oct-17) 05-Jan-10 (in house check Oct-17) 09-Oct-09 (in house check Oct-17)	Id are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Mar-19 Dec-18 Jan-19 Scheduled Check In house check: Oct-20 In house check: Oct-20 In house check: Oct-20
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Probe EF3DV3 Probe H3DV6 DAE4 Secondary Standards Power meter Agilent 4419B Power sensor HP E4412A Power sensor HP E4412A Power sensor HP 8482A RF generator R&S SMT-06 Network Analyzer Agilent E8358A	ainties with confidence pre- ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 4013 SN: 6065 SN: 781 ID # SN: GB42420191 SN: US38485102 SN: US37295597 SN: 832283/011 SN: US41080477 Name	Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. 217-02683) 05-Mar-18 (No. EF3-4013_Mar18) 30-Dec-17 (No. H3-6065_Dec17) 17-Jan-18 (No. DAE4-781_Jan18) Check Date (in house) 09-Oct-09 (in house check Oct-17) 05-Jan-10 (in house check Oct-17) 27-Aug-12 (in house check Oct-17) 31-Mar-14 (in house check Oct-17) Function	Id are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Mar-19 Dec-18 Jan-19 Scheduled Check In house check: Oct-20 In house check: Oct-20
The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Probe EF3DV3 Probe H3DV6 DAE4 Secondary Standards Power meter Agilent 4419B Power sensor HP E4412A Power sensor HP E4412A Power sensor HP 8482A RF generator R&S SMT-06	ainties with confidence pre- ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 4013 SN: 6065 SN: 781 ID # SN: GB42420191 SN: US38485102 SN: US3725597 SN: 832283/011 SN: US41080477	Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. 217-02673) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. 217-02683) 05-Mar-18 (No. 217-02683) 05-Mar-18 (No. DAE4-781_Jan18) Check Date (in house) 09-Oct-09 (in house check Oct-17) 05-Jan-10 (in house check Oct-17) 07-Aug-12 (in house check Oct-17) 27-Aug-12 (in house check Oct-17) 31-Mar-14 (in house check Oct-17)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Mar-19 Dec-18 Jan-19 Scheduled Check In house check: Oct-20 In house check: Oct-20

Certificate No: CD835V3-1165_Jul18

Page 1 of 5



Calibration Laboratory of Schmid & Partner

Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizeris C Service suis Servizio svi Swiss Calib

Schweizerischer Kalibrierdienst Service suisse d'étalonnage Servizio svizzero di taratura Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

References

- [1] ANSI-C63.19-2011
 - American National Standard, Methods of Measurement of Compatibility between Wireless Communications Devices and Hearing Aids.

Methods Applied and Interpretation of Parameters:

- Coordinate System: y-axis is in the direction of the dipole arms. z-axis is from the basis of the antenna (mounted on the table) towards its feed point between the two dipole arms. x-axis is normal to the other axes. In coincidence with the standards [1], the measurement planes (probe sensor center) are selected to be at a distance of 15 mm above the top metal edge of the dipole arms.
- Measurement Conditions: Further details are available from the hardcopies at the end of the certificate. All
 figures stated in the certificate are valid at the frequency indicated. The forward power to the dipole connector
 is set with a calibrated power meter connected and monitored with an auxiliary power meter connected to a
 directional coupler. While the dipole under test is connected, the forward power is adjusted to the same level.
- Antenna Positioning: The dipole is mounted on a HAC Test Arch phantom using the matching dipole positioner with the arms horizontal and the feeding cable coming from the floor. The measurements are performed in a shielded room with absorbers around the setup to reduce the reflections. It is verified before the mounting of the dipole under the Test Arch phantom, that its arms are perfectly in a line. It is installed on the HAC dipole positioner with its arms parallel below the dielectric reference wire and able to move elastically in vertical direction without changing its relative position to the top center of the Test Arch phantom. The vertical distance to the probe is adjusted after dipole mounting with a DASY5 Surface Check job. Before the measurement, the distance between phantom surface and probe tip is verified. The proper measurement distance is selected by choosing the matching section of the HAC Test Arch phantom with the proper device reference point (upper surface of the dipole) and the matching grid reference point (tip of the probe) considering the probe sensor offset. The vertical distance to the probe is essential for the accuracy.
- Feed Point Impedance and Return Loss: These parameters are measured using a HP 8753E Vector Network Analyzer. The impedance is specified at the SMA connector of the dipole. The influence of reflections was eliminating by applying the averaging function while moving the dipole in the air, at least 70cm away from any obstacles.
- E-field distribution: E field is measured in the x-y-plane with an isotropic ER3D-field probe with 100 mW forward power to the antenna feed point. In accordance with [1], the scan area is 20mm wide, its length exceeds the dipole arm length (180 or 90mm). The sensor center is 15 mm (in z) above the metal top of the dipole arms. Two 3D maxima are available near the end of the dipole arms. Assuming the dipole arms are perfectly in one line, the average of these two maxima (in subgrid 2 and subgrid 8) is determined to compensate for any non-parallelity to the measurement plane as well as the sensor displacement. The E-field value stated as calibration value represents the maximum of the interpolated 3D-E-field, in the plane above the dipole surface.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: CD835V3-1165_Jul18

Page 2 of 5



Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY5	V52.10.1
HAC Test Arch	
15 mm	
dx, dy = 5 mm	
835 MHz ± 1 MHz	
< 0.05 dB	
	HAC Test Arch 15 mm dx, dy = 5 mm 835 MHz ± 1 MHz

Maximum Field values at 835 MHz

E-field 15 mm above dipole surface	condition	Interpolated maximum
Maximum measured above high end	100 mW input power	108.7 V/m = 40.72 dBV/m
Maximum measured above low end	100 mW input power	108.6 V/m = 40.72 dBV/m
Averaged maximum above arm	100 mW input power	108.7 V/m ± 12.8 % (k=2)

Appendix (Additional assessments outside the scope of SCS 0108)

Antenna Parameters

Frequency	Return Loss	Impedance
800 MHz	16.4 dB	40.0 Ω - 9.2 jΩ
835 MHz	25.5 dB	53.7 Ω + 4.0 jΩ
880 MHz	17.8 dB	60.3 Ω - 9.8 jΩ
900 MHz	16.5 dB	51.6 Ω - 15.3 jΩ
945 MHz	21.7 dB	43.9 Ω + 4.8 jΩ

3.2 Antenna Design and Handling

The calibration dipole has a symmetric geometry with a built-in two stub matching network, which leads to the enhanced bandwidth.

The dipole is built of standard semirigid coaxial cable. The internal matching line is open ended. The antenna is therefore open for DC signals.

Do not apply force to dipole arms, as they are liable to bend. The soldered connections near the feedpoint may be damaged. After excessive mechanical stress or overheating, check the impedance characteristics to ensure that the internal matching network is not affected.

After long term use with 40W radiated power, only a slight warming of the dipole near the feedpoint can be measured.


Certificate No: CD835V3-1165_Jul18

Page 3 of 5





Impedance Measurement Plot



Certificate No: CD835V3-1165_Jul18

Page 4 of 5



DASY5 E-field Result

Date: 19.07.2018

Test Laboratory: SPEAG Lab2

DUT: HAC-Dipole 835 MHz; Type: CD835V3; Serial: CD835V3 - SN: 1165

Communication System: UID 0 - CW ; Frequency: 835 MHz Medium parameters used: $\sigma = 0$ S/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³ Phantom section: RF Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:


- Probe: EF3DV3 SN4013; ConvF(1, 1, 1) @ 835 MHz; Calibrated: 05.03.2018
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn781; Calibrated: 17.01.2018
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA; Serial: 1070
- DASY52 52.10.1(1476); SEMCAD X 14.6.11(7439)

Dipole E-Field measurement @ 835MHz/E-Scan - 835MHz d=15mm/Hearing Aid Compatibility Test (41x361x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm

Reference Value = 130.9 V/m; Power Drift = 0.02 dB Applied MIF = 0.00 dB RF audio interference level = 40.73 dBV/m Emission category: M3

MIF scaled E-field

Grid 1 M3 40.28 dBV/m		Grid 3 M3 40.67 dBV/m
	Grid 5 M4 35.96 dBV/m	Grid 6 M4 35.94 dBV/m
	Grid 8 M3 40.73 dBV/m	Grid 9 M3 40.67 dBV/m



Certificate No: CD835V3-1165_Jul18

Page 5 of 5



Dipole 1880 MHz

Calibration Laboratory of
Schmid & Partner
Engineering AG
Zeughausstrasse 43, 8004 Zurich, Switzerland



Schweizerischer Kalibrierdienst Service suisse d'étalonnage Servizio svizzero di taratura Swiss Calibration Service

S

С

S

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client CTTL (Auden)

Certificate No: CD1880V3-1149_Jul18

Accreditation No.: SCS 0108

CALIBRATION CI	ERTIFICATE			
Object	CD1880V3 - SN: 1149			
Calibration procedure(s)	QA CAL-20.v6 Calibration proce	dure for dipoles in air		
Calibration date:	July 19, 2018			
The measurements and the uncerta	ainties with confidence pr ed in the closed laborator	onal standards, which realize the physical un obability are given on the following pages an y facility: environment temperature $(22 \pm 3)^{\circ}($	d are part of the certificate.	
Primary Standards		Cal Date (Certificate No.)	Scheduled Calibration	
Power meter NRP	SN: 104778	04-Apr-18 (No. 217-02672/02673)	Apr-19	
Power sensor NRP-Z91	SN: 103244	04-Apr-18 (No. 217-02672)	Apr-19	
Power sensor NRP-Z91	SN: 103245	04-Apr-18 (No. 217-02673)	Apr-19	
Reference 20 dB Attenuator	SN: 5058 (20k)	04-Apr-18 (No. 217-02682)	Apr-19	
Type-N mismatch combination	SN: 5047.2 / 06327	04-Apr-18 (No. 217-02683)	Apr-19	
Probe EF3DV3	SN: 4013	05-Mar-18 (No. EF3-4013_Mar18)	Mar-19	
Probe H3DV6	SN: 6065	30-Dec-17 (No. H3-6065_Dec17)	Dec-18	
DAE4	SN: 781	17-Jan-18 (No. DAE4-781_Jan18)	Jan-19	
Secondary Standards	ID #	Check Date (in house)	Scheduled Check	
Power meter Agilent 4419B	SN: GB42420191	09-Oct-09 (in house check Oct-17)	In house check: Oct-20	
Power sensor HP E4412A	SN: US38485102	05-Jan-10 (in house check Oct-17)	In house check: Oct-20	
Power sensor HP 8482A	SN: US37295597	09-Oct-09 (in house check Oct-17)	In house check: Oct-20	
RF generator R&S SMT-06	SN: 832283/011	27-Aug-12 (in house check Oct-17)	In house check: Oct-20	
Network Analyzer Agilent E8358A	SN: US41080477	31-Mar-14 (in house check Oct-17)	In house check: Oct-18	
	Name	Function	Signature	
Calibrated by:	Leif Klysner	Laboratory Technician	Sef Thepen	
Approved by:	Katja Pokovic	Technical Manager	00101	
, pprotos ay.			6605	
			Issued: July 19, 2018	
This calibration certificate shall not	be reproduced except in	n full without written approval of the laborator	у.	

Certificate No: CD1880V3-1149_Jul18

Page 1 of 5



Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



S C S

Schweizerischer Kalibrierdienst Service suisse d'étalonnage Servizio svizzero di taratura Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

References

- [1] ANSI-C63.19-2011
 - American National Standard, Methods of Measurement of Compatibility between Wireless Communications Devices and Hearing Aids.

Methods Applied and Interpretation of Parameters:

- Coordinate System: y-axis is in the direction of the dipole arms. z-axis is from the basis of the antenna (mounted on the table) towards its feed point between the two dipole arms. x-axis is normal to the other axes. In coincidence with the standards [1], the measurement planes (probe sensor center) are selected to be at a distance of 15 mm above the top metal edge of the dipole arms.
- Measurement Conditions: Further details are available from the hardcopies at the end of the certificate. All
 figures stated in the certificate are valid at the frequency indicated. The forward power to the dipole connector
 is set with a calibrated power meter connected and monitored with an auxiliary power meter connected to a
 directional coupler. While the dipole under test is connected, the forward power is adjusted to the same level.
- Antenna Positioning: The dipole is mounted on a HAC Test Arch phantom using the matching dipole positioner with the arms horizontal and the feeding cable coming from the floor. The measurements are performed in a shielded room with absorbers around the setup to reduce the reflections. It is verified before the mounting of the dipole under the Test Arch phantom, that its arms are perfectly in a line. It is installed on the HAC dipole positioner with its arms parallel below the dielectric reference wire and able to move elastically in vertical direction without changing its relative position to the top center of the Test Arch phantom. The vertical distance to the probe is adjusted after dipole mounting with a DASY5 Surface Check job. Before the measurement, the distance between phantom surface and probe tip is verified. The proper measurement distance is selected by choosing the matching section of the HAC Test Arch phantom with the proper device reference point (upper surface of the dipole) and the matching grid reference point (tip of the probe) considering the probe sensor offset. The vertical distance to the probe is essential for the accuracy.
- Feed Point Impedance and Return Loss: These parameters are measured using a HP 8753E Vector Network Analyzer. The impedance is specified at the SMA connector of the dipole. The influence of reflections was eliminating by applying the averaging function while moving the dipole in the air, at least 70cm away from any obstacles.
- E-field distribution: E field is measured in the x-y-plane with an isotropic ER3D-field probe with 100 mW forward power to the antenna feed point. In accordance with [1], the scan area is 20mm wide, its length exceeds the dipole arm length (180 or 90mm). The sensor center is 15 mm (in z) above the metal top of the dipole arms. Two 3D maxima are available near the end of the dipole arms. Assuming the dipole arms are perfectly in one line, the average of these two maxima (in subgrid 2 and subgrid 8) is determined to compensate for any non-parallelity to the measurement plane as well as the sensor displacement. The E-field value stated as calibration value represents the maximum of the interpolated 3D-E-field, in the plane above the dipole surface.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: CD1880V3-1149_Jul18

Page 2 of 5



Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V52.10.1
Phantom	HAC Test Arch	
Distance Dipole Top - Probe Center	15 mm	
Scan resolution	dx, dy = 5 mm	
Frequency	1880 MHz ± 1 MHz	
Input power drift	< 0.05 dB	

Maximum Field values at 1880 MHz

E-field 15 mm above dipole surface	condition	Interpolated maximum
Maximum measured above high end	100 mW input power	89.8 V/m = 39.06 dBV/m
Maximum measured above low end	100 mW input power	89.3 V/m = 39.02 dBV/m
Averaged maximum above arm	100 mW input power	89.5 V/m ± 12.8 % (k=2)

Appendix (Additional assessments outside the scope of SCS 0108)

Antenna Parameters

Frequency	Return Loss	Impedance
1730 MHz	23.9 dB	53.9 Ω + 5.4 jΩ
1880 MHz	22.5 dB	54.7 Ω + 6.3 jΩ
1900 MHz	23.4 dB	55.6 Ω + 4.5 jΩ
1950 MHz	30.3 dB	52.9 Ω - 1.3 jΩ
2000 MHz	21.3 dB	44.2 Ω + 5.7 jΩ

3.2 Antenna Design and Handling

The calibration dipole has a symmetric geometry with a built-in two stub matching network, which leads to the enhanced bandwidth.


The dipole is built of standard semirigid coaxial cable. The internal matching line is open ended. The antenna is therefore open for DC signals.

Do not apply force to dipole arms, as they are liable to bend. The soldered connections near the feedpoint may be damaged. After excessive mechanical stress or overheating, check the impedance characteristics to ensure that the internal matching network is not affected.

After long term use with 40W radiated power, only a slight warming of the dipole near the feedpoint can be measured.


Certificate No: CD1880V3-1149_Jul18

Page 3 of 5





Impedance Measurement Plot



Certificate No: CD1880V3-1149_Jul18

Page 4 of 5



DASY5 E-field Result

Date: 19.07.2018

Test Laboratory: SPEAG Lab2

DUT: HAC Dipole 1880 MHz; Type: CD1880V3; Serial: CD1880V3 - SN: 1149

 $\begin{array}{l} \mbox{Communication System: UID 0 - CW ; Frequency: 1880 MHz \\ \mbox{Medium parameters used: } \sigma = 0 \mbox{S/m}, \mbox{ϵ_r} = 1; \mbox{ρ} = 0 \mbox{kg/m^3} \\ \mbox{Phantom section: } RF \mbox{Section} \\ \mbox{Measurement Standard: } DASY5 (IEEE/IEC/ANSI C63.19-2011) \\ \end{array}$


DASY52 Configuration:

- Probe: EF3DV3 SN4013; ConvF(1, 1, 1) @ 1880 MHz; Calibrated: 05.03.2018
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn781; Calibrated: 17.01.2018
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA; Serial: 1070
- DASY52 52.10.1(1476); SEMCAD X 14.6.11(7439)

Dipole E-Field measurement @ 1880MHz/E-Scan - 1880MHz d=15mm/Hearing Aid Compatibility Test (41x181x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Volue = 160 1 V(m; Power Drift = -0.04 dB

Reference Value = 160.1 V/m; Power Drift = -0.04 dB Applied MIF = 0.00 dB RF audio interference level = 39.06 dBV/m Emission category: M2

 Grid 2 M2 39.06 dBV/m	Grid 3 M2 39.01 dBV/m
 Grid 5 M2 36.15 dBV/m	Grid 6 M2 36.1 dBV/m
 Grid 8 M2 39.02 dBV/m	Grid 9 M2 38.91 dBV/m



Certificate No: CD1880V3-1149_Jul18



ANNEX F: UID Specification

Calibration Laboratory of

Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland

Name:	GSM-FDD (TDMA, GMSK)	
Group:	GSM	
UID:	10021-DAC	
PAR: 1	9.39 dB	
MIF: 2	3.63 dB	
Standard Reference:	ETSI TS 100 909 V8.9.0 (2005-01)	
	FCC OET KDB 941225, D03 and D04	
Category:	Periodic pulsed modulation	
Modulation:	GMSK	
Frequency Band:	GSM 450 (450.4 - 457.6 MHz)	
	GSM 480 (478.8 - 486.0 MHz)	
	GSM 710 (698.0 - 716.0 MHz)	
	GSM 750 (747.0 - 763.0 MHz)	
	GSM 850 (824.0 - 849.0 MHz)	
	P-GSM 900 (890.0 - 915.0 MHz)	
	E-GSM 900 (880.0 - 915.0 MHz)	
	R-GSM 900 (876.0 - 915.0 MHz)	
	DCS 1800 (1710.0 - 1785.0 MHz)	
	PCS 1900 (1850.0 - 1910.0 MHz)	
	ER-GSM 900 (873.0 - 915.0 MHz)	
	Validation band (0.0 - 6000.0 MHz)	
Detailed Specification:	Active Slot: TN0	
	Data: PN9 continuous	
	Frame: composed out of 8 Slots	
	Multiframe: 26th (IDLE) Frame set blank	
	Slottype & -timing: Normal burst for GMSK	
Bandwidth:	0.2 MHz	
Integration Time:	120.0 ms	

PAR (0.1%) in accordance with FCC KDB 971168, Section 6.0 "Measurement of the Peak-to-Average Power Ratio (PAPR)"
 Modulation Interference Factor (MIF) value valid only in conjunction with advanced probe response linearization calibration for the same communication system (same UID and version).

UID Specification Sheet


UID 10021-DAC page 1/2

16.11.2016




Calibration Laboratory of Schmid & Partner

Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Complementary Cumulative Distribution Function (CCDF)



UID Specification Sheet

UID 10021-DAC page 2/2

16.11.2016



Calibration Laboratory of Schmid & Partner

Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland

Name:	UMTS-FDD (WCDMA)
Group: UID:	WCDMA 10011-CAB
PAR: ¹ MIF: ²	2.91 dB -27.23 dB
WIF.	-27.23 dB
Standard Reference:	3GPP TS 25.141 Annex A
	FCC OET KDB 941225 D01 SAR test for 3G devices v02
Category:	Random amplitude modulation
Modulation: Frequency Band:	QPSK Band 1, UTRA/FDD (1920.0-1980.0 MHz, 20000)
riequency band.	Band 2, UTRA/FDD (1850.0-1910.0 MHz, 2000)
	Band 3, UTRA/FDD (1710.0-1785.0 MHz, 20002)
	Band 4, UTRA/FDD (1710.0-1755.0 MHz, 20003)
	Band 5, UTRA/FDD (824.0-849.0 MHz, 20004)
	Band 6, UTRA/FDD (830.0-840.0 MHz, 20005)
	Band 7, UTRA/FDD (2500.0-2570.0 MHz, 20006)
	Band 8, UTRA/FDD (880.0-915.0 MHz, 20007)
	Band 9, UTRA/FDD (1749.9-1784.9 MHz, 20008)
	Band 10, UTRA/FDD (1710.0-1770.0 MHz, 20009)
	Band 11, UTRA/FDD (1427.9-1452.9 MHz, 20010)
	Band 12, UTRA/FDD (698.0-716.0 MHz, 20011)
	Band 13, UTRA/FDD (777.0-787.0 MHz, 20012)
	Band 14, UTRA/FDD (788.0-798.0 MHz, 20013)
	Band 19, UTRA/FDD (830.0-845.0 MHz, 20130)
	Band 20, UTRA/FDD (832.0-862.0 MHz, 20131) Band 21, UTRA/FDD (1447.9-1462.9 MHz, 20132)
	Band 21, UTRA/FDD (1447.3-1462.3 MHz, 20132) Band 22, UTRA/FDD (3410.0-3490.0 MHz, 20217)
	Band 25, UTRA/FDD (1850.0-1915.0 MHz, 20218)
	Band 26, UTRA/FDD (814.0-849.0 MHz, 20219)
Detailed Specification:	Dedicated Channel Type: RMC
	Bitrate: 12.2 kbps
	DPDCH: 60 kbps
	DPCCH: 15 kbps
5	DPCCH/DPDCH power ratio: -5.46 dB
Bandwidth:	5.0 MHz
Integration Time:	100.0 ms

 PAR (0.1%) in accordance with FCC KDB 971168, Section 6.0 "Measurement of the Peak-to-Average Power Ratio (PAPR)"
 Modulation Interference Factor (MIF) value valid only in conjunction with advanced probe response linearization calibration for the same communication system (same UID and version).

UID Specification Sheet


UID 10011-CAB page 1/2

16.01.2014




Calibration Laboratory of

Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Complementary Cumulative Distribution Function (CCDF)



Time Domain

UID 10011-CAB page 2/2

16.01.2014

UID Specification Sheet



Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland

Name:	LTE-FDD (SC-FDMA, 1 RB, 10 MHz, 16-QAM)	
Group:	LTE-FDD	
UID:	10176-CAE	
PAR: 1	6.52 dB	
MIF: 2	-9.76 dB	
Standard Reference:	3GPP / ETSI TS 136.101 V8.4.0	
	3GPP / ETSI TS 136.213 V8.4.0	
Category:	FCC OET KDB 941225 D05 SAR for LTE Devices v01 Random amplitude modulation	
Modulation:	16-QAM	
Frequency Band:	Band 1, E-UTRA/FDD (1920.0 - 1980.0 MHz)	
	Band 2, E-UTRA/FDD (1850.0 - 1910.0 MHz)	
	Band 3, E-UTRA/FDD (1710.0 - 1785.0 MHz) Band 4, E-UTRA/FDD (1710.0 - 1755.0 MHz)	
	Band 5, E-UTRA/FDD (824.0 - 849.0 MHz)	
	Band 6, E-UTRA/FDD (830.0 - 840.0 MHz)	
	Band 7, E-UTRA/FDD (2500.0 - 2570.0 MHz)	
	Band 8, E-UTRA/FDD (880.0 - 915.0 MHz)	
	Band 9, E-UTRA/FDD (1749.9 - 1784.9 MHz)	
	Band 10, E-UTRA/FDD (1710.0 - 1770.0 MHz) Band 11, E-UTRA/FDD (1427.9 - 1447.9 MHz)	
	Band 12, E-UTRA/FDD (699.0 - 716.0 MHz)	
	Band 13, E-UTRA/FDD (777.0 - 787.0 MHz)	
	Band 14, E-UTRA/FDD (788.0 - 798.0 MHz)	
	Band 17, E-UTRA/FDD (704.0 - 716.0 MHz)	
	Band 18, E-UTRA/FDD (815.0 - 830.0 MHz)	
	Band 19, E-UTRA/FDD (830.0 - 845.0 MHz)	
	Band 20, E-UTRA/FDD (832.0 - 862.0 MHz) Band 21, E-UTRA/FDD (1447.9 - 1462.9 MHz)	
	Band 22, E-UTRA/FDD (3410.0 - 3490.0 MHz)	
	Band 23, E-UTRA/FDD (2000.0 - 2020.0 MHz)	
	Band 24, E-UTRA/FDD (1626.5 - 1660.5 MHz)	
	Band 25, E-UTRA/FDD (1850.0 - 1915.0 MHz)	
	Band 26 E-UTRA/FDD (814.0 - 849.0 MHz)	
	Band 27 E-UTRA/FDD (807.0 - 824.0 MHz) Band 28 E-UTRA/FDD (703.0 - 748.0 MHz)	
	Band 30, E-UTRA/FDD (2305.0 - 2315.0 MHz)	
	Band 65, E-UTRA/FDD (1920.0 - 2010.0 MHz)	
	Band 66, E-UTRA/FDD (1710.0 - 1780.0 MHz)	
	Band 68, E-UTRA/FDD (698.0 - 728.0 MHz)	
	Band 70, E-UTRA/FDD (1695.0 - 1710.0 MHz)	
	Band 71, E-UTRA/FDD (663.0 - 698.0 MHz) Validation band (0.0 - 6000.0 MHz)	
Detailed Specification:	Modulation Scheme; SC-FDMA	
botanoa opoonioation.	Number of PUSCHs: 1	
	Settings for Subframe #0 to #9:	
	Modulation Scheme: QPSK	
	Data Type: UL-SCH	
	Number RB: 1	
	Transport Block Size: 256 TBS Index: 14	
	MCS Index: 15	
	Data Type: PN9	
Bandwidth:	10.0 MHz	
Integration Time:	10.0 ms	

 PAR (0.1%) in accordance with FCC KDB 971168, Section 6.0 "Measurement of the Peak-to-Average Power Ratio (PAPR)"
 Modulation Interference Factor (MIF) value valid only in conjunction with advanced probe response linearization calibration for the same communication system (same UID and version).


UID Specification Sheet

UID 10176-CAE page 1/2




Calibration Laboratory of

Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Complementary Cumulative Distribution Function (CCDF)



UID Specification Sheet

UID 10176-CAE page 2/2



Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland

Name:	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM)
Group: UID:	LTE-FDD 10170-CAD
PAR: 1 MIF: 2	6.52 dB -9.76 dB
Standard Reference:	3GPP / ETSI TS 136.101 V8.4.0 3GPP / ETSI TS 136.213 V8.4.0
Category:	FCC OET KDB 941225 D05 SAR for LTE Devices v01 Random amplitude modulation
Modulation:	16-QAM
Frequency Band:	Band 1, E-UTRA/FDD (1920.0 - 1980.0 MHz) Band 2, E-UTRA/FDD (1710.0 - 1785.0 MHz) Band 3, E-UTRA/FDD (1710.0 - 1755.0 MHz) Band 7, E-UTRA/FDD (1710.0 - 1755.0 MHz) Band 7, E-UTRA/FDD (1710.0 - 1770.0 MHz) Band 10, E-UTRA/FDD (1710.0 - 1770.0 MHz) Band 20, E-UTRA/FDD (1710.0 - 1770.0 MHz) Band 22, E-UTRA/FDD (1820.0 - 862.0 MHz) Band 23, E-UTRA/FDD (1820.0 - 849.0 MHz) Band 25, E-UTRA/FDD (1850.0 - 1915.0 MHz) Band 26, E-UTRA/FDD (1850.0 - 1915.0 MHz) Band 26, E-UTRA/FDD (1950.0 - 1915.0 MHz) Band 66, E-UTRA/FDD (1920.0 - 2010.0 MHz) Band 66, E-UTRA/FDD (1920.0 - 2010.0 MHz) Band 70, E-UTRA/FDD (1995.0 - 1710.0 MHz) Band 70, E-UTRA/FDD (1995.0 - 1710.0 MHz) Band 71, E-UTRA/FDD (1995.0 - 1710.0 MHz) Band 70, E-UTRA/FDD (1995.0 - 1710.0 MHz) Band 70, E-UTRA/FDD (1995.0 - 1710.0 MHz) Band 71, E-UTRA/FDD (1995.0 - 1710.0 MHz) Band 71, E-UTRA/FDD (1995.0 - 1710.0 MHz) Band 71, E-UTRA/FDD (1995.0 - 1710.0 MHz)
Detailed Specification:	Modulation Scheme: SC-FDMA Number of PUSCHs: 1 Settings for Subframe #0 to #9: Modulation Scheme: 16QAM Data Type: UL-SCH Number RB: 1 Transport Block Size: 256 TBS Index: 14 MCS Index: 15 Data Type: PN9
Bandwidth:	20.0 MHz
Integration Time:	10.0 ms

PAR (0.1%) in accordance with FCC KDB 971168, Section 6.0 "Measurement of the Peak-to-Average Power Ratio (PAPR)" Modulation Interference Factor (MIF) value valid only in conjunction with advanced probe response linearization calibration for the same communication system (same UID and version). 2


UID Specification Sheet

UID 10170-CAD page 1/2




Calibration Laboratory of

Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Complementary Cumulative Distribution Function (CCDF)



UID Specification Sheet

UID 10170-CAD page 2/2



Calibration Laboratory of Schmid & Partner

Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland

Name:	IEEE 802.11b WiFi 2.4 GHz (DSSS, 11 Mbps)
Group: UID:	WLAN 10061-CAB
PAR: ¹ MIF: ²	3.60 dB -2.02 dB
Standard Reference:	IEEE 802.11b-1999 , Part 11, FCC SAR meas for 802 11 a b g v01r02 (248227 D01)
Category:	Random amplitude modulation
Modulation:	DQPSK
Frequency Band:	WLAN 2.4GHz (2412.0-2484.0 MHz, 20230)
Detailed Specification:	Data Rate: 11 Mbps
	Spreading, Coding: CCK
	PPDU format: Long Preamble & Heading
	PSDU Length: 1024
Bandwidth:	PSDU Data: PN9 20.0 MHz
Integration Time:	1.5 ms

 PAR (0.1%) in accordance with FCC KDB 971168, Section 6.0 "Measurement of the Peak-to-Average Power Ratio (PAPR)"
 Modulation Interference Factor (MIF) value valid only in conjunction with advanced probe response linearization calibration for the same communication system (same UID and version).

UID Specification Sheet


UID 10061-CAB page 1/2

26.11.2014




Calibration Laboratory of

Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Complementary Cumulative Distribution Function (CCDF)



UID Specification Sheet

UID 10061-CAB page 2/2

26.11.2014

END OF REPORT