

Zhang, Apple-MP (Shenzhen)

From: oetech@fcc.gov
Sent: 2018年5月25日星期五 20:45
To: Zhang, Apple-MP (Shenzhen)
Subject: Response to Inquiry to FCC (Tracking Number 402348)

Inquiry on 05/24/2018 :

Inquiry:

Dear Sir/ Madam,

Our client Shenzhen Dongdixin Technology Co.,Ltd. is going to apply the FCC certificate. The device is a 3-in-1 Wireless Electrotherapy Devices, Photo shows as the attachment. There's not enough space to place the FCC label statement. Please advise if it is acceptable to put the statement to user manual and package instead.

Thank you.

FCC response on 05/25/2018

Yes this is acceptable.

The following information can be obtained from our website at <https://apps.fcc.gov/oetcf/kdb/forms/FTSSearchResultPage.cfm?switch=P&id=27980>.

Also, labelling guidelines can be found in Knowledge DataBase (KDB) Publication [#784748](#).

Attachment Details:

[Photo](#)

[Photo](#)

Do not reply to this message. Please select the [Reply to an Inquiry Response](#) link from the OET Inquiry System to add any additional information pertaining to this inquiry.