

System Check_H5250

DUT: Dipole 5GHzV2;Type:D5GHzV2; SN:1169

Communication System: CW; Frequency: 5250 MHz;Duty Cycle: 1:1

Medium: H5G Medium parameters used: $f = 5250$ MHz; $\sigma = 4.74$ S/m; $\epsilon_r = 36.79$; $\rho = 1000$ kg/m³

Ambient Temperature : 22.3 °C; Liquid Temperature : 22.1 °C

DASY5 Configuration:

- Probe: EX3DV4 - SN3970; Calibrated: 2024/6/25;

- Electronics: DAE4 Sn1418; Calibrated: 2024/5/17

- Phantom: ELI v5.0; Type: QDOVA002AA; Serial: TP:1231

- Measurement SW: DASY52, Version 52.8 (7); SEMCAD X Version 14.6.10 (7164)

Pin=100mW/Area Scan (91x91x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm

Maximum value of SAR (interpolated) = 22.1 W/kg

Pin=100mW/Zoom Scan (4x4x2mm) (7x7x12)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=2mm

Reference Value = 57.180 V/m; Power Drift = 0.14 dB

Peak SAR (extrapolated) = 37.4 W/kg

SAR(1 g) = 7.93 W/kg; SAR(10 g) = 2.14 W/kg

Maximum value of SAR (measured) = 23.3 W/kg

