

*SELVAS Healthcare, Inc.*  
*155, Sinseong-ro, Yuseong-gu, Daejeon, South Korea*

---

Date: 2022-11-22

FCC ID: 2AL4D-T90

Model Number: T90ET, T90EZ


To:  
Federal Communications Commission  
Authorization and Evaluation Division  
7435 Oakland Mills Road  
Columbia, MD

To Whom It May Concern,

We, SELVAS Healthcare, Inc. declare that this EUT (FCC ID: 2AL4D-T90) doesn't have Ad Hoc mode function on "on-US/Canada frequencies". In addition. The frequency selection feature is disabled by firmware for devices marked to the US/Canada. For DTS part of this device, only channels 1-11 will be used in North America, Country code selection is disabled.

Should you have any questions or comments regarding this matter, please have my best attention.

Sincerely,

(Signed) 

Name / Title: ILHYUNG KIM/Director  
Company: SELVAS Healthcare, Inc.  
Address: 155, Sinseong-ro, Yuseong-gu, Daejeon, South Korea  
Tel: +82-42-879-3000  
E-Mail: jeffrey.i.kim@selvashc.com