

CAM Pressure Mote Gen 2

Model # CAM-PR2

Airgas USA, LLC 184 Sandbank Rd, Cheshire, CT 06410, USA

Conforms to UL STD 60079-0, 60079-11 & 61010-1.

Certified to CSA STD C22.2#60079-0, 60079-11 & 61010-1-12.

CSA Cert. No. ETL21CA104530669X

UKCA Cert. No. ITS21UKEX0033X

IECEX certificate number: IECEX ETL 21.0002X

ATEX certificate number ITS-I21ATEX28699X

Class I Division 1 Groups A-D T4 IP55

Class I Zone 0 AEx ia IIC T4 Ga IP55

Ex ia IIC T4 Ga

 II I G Ex ia IIC T4 Ga

-20°C ≤ Ta ≤ 60°C

FCC ID: 2ALBX-CAMP RMR02


IC: 22533-CAMP RMR02



This device complies with Part 15 of the FCC rules.

1. This device may not cause harmful interference.

2. This device must accept any interference received, including interference that may cause undesired operations.

Battery Supply:  2 x AA 3.6VDC Primary

 CAUTION: Note polarity of batteries ++

ATTENTION: Verifiez la polarite devs pilzes ++

 See instruction manual

Voir le manuel d'instructions

