

## Plot 9


### A. Experimental conditions.

|                 |                            |
|-----------------|----------------------------|
| Probe           | SN 25/22 EPG0373           |
| ConvF           | 3.01                       |
| Area Scan       | surf_sam_plan.txt          |
| Zoom Scan       | 7x7x8,dx=5mm dy=5mm dz=5mm |
| Phantom         | Validation plane           |
| Device Position | Body                       |
| Band            | LTE band 5                 |
| Channels        | Middle                     |
| Signal          | LTE (Crest factor: 1.0)    |

### B. Permittivity

|                                        |         |
|----------------------------------------|---------|
| Frequency (MHz)                        | 836.500 |
| Relative permittivity (real part)      | 42.277  |
| Relative permittivity (imaginary part) | 19.400  |
| Conductivity (S/m)                     | 0.938   |

### C. SAR Surface and Volume




Maximum location: X=-10.00, Y=1.00 ; SAR Peak: 0.49 W/kg

### D. SAR 1g & 10g



|                                                       |        |
|-------------------------------------------------------|--------|
| SAR 10g (W/Kg)                                        | 0.157  |
| SAR 1g (W/Kg)                                         | 0.282  |
| Variation (%)                                         | -0.510 |
| Horizontal validation criteria: minimum distance (mm) | --     |
| Vertical validation criteria: SAR ratio M2/M1 (%)     | --     |

### E. Z Axis Scan

|            |       |       |       |       |       |
|------------|-------|-------|-------|-------|-------|
| Z (mm)     | 0.00  | 4.00  | 8.00  | 12.00 | 16.00 |
| SAR (W/Kg) | 0.488 | 0.301 | 0.185 | 0.117 | 0.079 |



### F. 3D Image

| 3D screen shot                                                                     | Hot spot position                                                                    |
|------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------|
|  |  |

## Plot 10


### A. Experimental conditions.

|                 |                            |
|-----------------|----------------------------|
| Probe           | SN 25/22 EPG0373           |
| ConvF           | 3.01                       |
| Area Scan       | surf_sam_plan.txt          |
| Zoom Scan       | 7x7x8,dx=5mm dy=5mm dz=5mm |
| Phantom         | Validation plane           |
| Device Position | Body                       |
| Band            | LTE band 5                 |
| Channels        | Middle                     |
| Signal          | LTE (Crest factor: 1.0)    |

### B. Permittivity

|                                        |         |
|----------------------------------------|---------|
| Frequency (MHz)                        | 836.500 |
| Relative permittivity (real part)      | 42.277  |
| Relative permittivity (imaginary part) | 19.400  |
| Conductivity (S/m)                     | 0.938   |

### C. SAR Surface and Volume




Maximum location: X=8.00, Y=-10.00 ; SAR Peak: 0.54 W/kg

### D. SAR 1g & 10g


|                                                       |        |
|-------------------------------------------------------|--------|
| SAR 10g (W/Kg)                                        | 0.201  |
| SAR 1g (W/Kg)                                         | 0.340  |
| Variation (%)                                         | -5.280 |
| Horizontal validation criteria: minimum distance (mm) | --     |
| Vertical validation criteria: SAR ratio M2/M1 (%)     | --     |

### E. Z Axis Scan

|            |       |       |       |       |       |
|------------|-------|-------|-------|-------|-------|
| Z (mm)     | 0.00  | 4.00  | 8.00  | 12.00 | 16.00 |
| SAR (W/Kg) | 0.535 | 0.368 | 0.252 | 0.175 | 0.123 |



### F. 3D Image



## Plot 11


### A. Experimental conditions.

|                 |                            |
|-----------------|----------------------------|
| Probe           | SN 25/22 EPG0373           |
| ConvF           | 2.96                       |
| Area Scan       | surf_sam_plan.txt          |
| Zoom Scan       | 7x7x8,dx=5mm dy=5mm dz=5mm |
| Phantom         | Validation plane           |
| Device Position | Body                       |
| Band            | LTE band 12                |
| Channels        | High                       |
| Signal          | LTE (Crest factor: 1.0)    |

### B. Permittivity

|                                        |         |
|----------------------------------------|---------|
| Frequency (MHz)                        | 711.000 |
| Relative permittivity (real part)      | 42.748  |
| Relative permittivity (imaginary part) | 23.152  |
| Conductivity (S/m)                     | 0.891   |

### C. SAR Surface and Volume




Maximum location: X=24.00, Y=-16.00 ; SAR Peak: 0.07 W/kg

### D. SAR 1g & 10g



|                                                       |        |
|-------------------------------------------------------|--------|
| SAR 10g (W/Kg)                                        | 0.027  |
| SAR 1g (W/Kg)                                         | 0.045  |
| Variation (%)                                         | -4.150 |
| Horizontal validation criteria: minimum distance (mm) | --     |
| Vertical validation criteria: SAR ratio M2/M1 (%)     | --     |

### E. Z Axis Scan

|            |       |       |       |       |       |
|------------|-------|-------|-------|-------|-------|
| Z (mm)     | 0.00  | 4.00  | 8.00  | 12.00 | 16.00 |
| SAR (W/Kg) | 0.074 | 0.049 | 0.032 | 0.021 | 0.015 |



### F. 3D Image

| 3D screen shot                                                                     | Hot spot position                                                                    |
|------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------|
|  |  |

## Plot 12


### A. Experimental conditions.

|                 |                            |
|-----------------|----------------------------|
| Probe           | SN 25/22 EPG0373           |
| ConvF           | 2.96                       |
| Area Scan       | surf_sam_plan.txt          |
| Zoom Scan       | 7x7x8,dx=5mm dy=5mm dz=5mm |
| Phantom         | Validation plane           |
| Device Position | Body                       |
| Band            | LTE band 12                |
| Channels        | High                       |
| Signal          | LTE (Crest factor: 1.0)    |

### B. Permittivity

|                                        |         |
|----------------------------------------|---------|
| Frequency (MHz)                        | 711.000 |
| Relative permittivity (real part)      | 42.748  |
| Relative permittivity (imaginary part) | 23.152  |
| Conductivity (S/m)                     | 0.891   |

### C. SAR Surface and Volume




### D. SAR 1g & 10g



|                                                       |        |
|-------------------------------------------------------|--------|
| SAR 10g (W/Kg)                                        | 0.061  |
| SAR 1g (W/Kg)                                         | 0.106  |
| Variation (%)                                         | -4.690 |
| Horizontal validation criteria: minimum distance (mm) | --     |
| Vertical validation criteria: SAR ratio M2/M1 (%)     | --     |

### E. Z Axis Scan

|            |       |       |       |       |       |
|------------|-------|-------|-------|-------|-------|
| Z (mm)     | 0.00  | 4.00  | 8.00  | 12.00 | 16.00 |
| SAR (W/Kg) | 0.164 | 0.113 | 0.077 | 0.054 | 0.038 |



### F. 3D Image

| 3D screen shot                                                                     | Hot spot position                                                                    |
|------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------|
|  |  |

## Plot 13


### A. Experimental conditions.

|                 |                            |
|-----------------|----------------------------|
| Probe           | SN 25/22 EPG0373           |
| ConvF           | 3.63                       |
| Area Scan       | surf_sam_plan.txt          |
| Zoom Scan       | 7x7x8,dx=5mm dy=5mm dz=5mm |
| Phantom         | Validation plane           |
| Device Position | Body                       |
| Band            | LTE band 41                |
| Channels        | Middle                     |
| Signal          | LTE (Crest factor: 1.0)    |

### B. Permittivity

|                                        |          |
|----------------------------------------|----------|
| Frequency (MHz)                        | 2605.000 |
| Relative permittivity (real part)      | 39.847   |
| Relative permittivity (imaginary part) | 13.318   |
| Conductivity (S/m)                     | 2.025    |

### C. SAR Surface and Volume




Maximum location: X=0.00, Y=7.00 ; SAR Peak: 0.15 W/kg

### D. SAR 1g & 10g



|                                                       |        |
|-------------------------------------------------------|--------|
| SAR 10g (W/Kg)                                        | 0.028  |
| SAR 1g (W/Kg)                                         | 0.062  |
| Variation (%)                                         | -0.070 |
| Horizontal validation criteria: minimum distance (mm) | --     |
| Vertical validation criteria: SAR ratio M2/M1 (%)     | --     |

### E. Z Axis Scan

|            |       |       |       |       |       |
|------------|-------|-------|-------|-------|-------|
| Z (mm)     | 0.00  | 4.00  | 8.00  | 12.00 | 16.00 |
| SAR (W/Kg) | 0.164 | 0.064 | 0.023 | 0.009 | 0.005 |



### F. 3D Image

| 3D screen shot                                                                     | Hot spot position                                                                    |
|------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------|
|  |  |

## Plot 14


### A. Experimental conditions.

|                 |                            |
|-----------------|----------------------------|
| Probe           | SN 25/22 EPG0373           |
| ConvF           | 3.63                       |
| Area Scan       | surf_sam_plan.txt          |
| Zoom Scan       | 7x7x8,dx=5mm dy=5mm dz=5mm |
| Phantom         | Validation plane           |
| Device Position | Body                       |
| Band            | LTE band 41                |
| Channels        | Middle                     |
| Signal          | LTE (Crest factor: 1.0)    |

### B. Permittivity

|                                        |          |
|----------------------------------------|----------|
| Frequency (MHz)                        | 2605.000 |
| Relative permittivity (real part)      | 39.847   |
| Relative permittivity (imaginary part) | 13.318   |
| Conductivity (S/m)                     | 2.025    |

### C. SAR Surface and Volume




Maximum location: X=-1.00, Y=1.00 ; SAR Peak: 0.87 W/kg

### D. SAR 1g & 10g



|                                                       |        |
|-------------------------------------------------------|--------|
| SAR 10g (W/Kg)                                        | 0.139  |
| SAR 1g (W/Kg)                                         | 0.344  |
| Variation (%)                                         | -0.460 |
| Horizontal validation criteria: minimum distance (mm) | --     |
| Vertical validation criteria: SAR ratio M2/M1 (%)     | --     |

### E. Z Axis Scan

|            |       |       |       |       |       |
|------------|-------|-------|-------|-------|-------|
| Z (mm)     | 0.00  | 4.00  | 8.00  | 12.00 | 16.00 |
| SAR (W/Kg) | 0.996 | 0.361 | 0.114 | 0.040 | 0.015 |



### F. 3D Image

| 3D screen shot                                                                     | Hot spot position                                                                    |
|------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------|
|  |  |

## Plot 15


### A. Experimental conditions.

|                 |                               |
|-----------------|-------------------------------|
| Probe           | SN 25/22 EPG0373              |
| ConvF           | 3.96                          |
| Area Scan       | surf_sam_plan.txt             |
| Zoom Scan       | 7x7x8,dx=5mm dy=5mm dz=5mm    |
| Phantom         | Validation plane              |
| Device Position | Body                          |
| Band            | IEEE 802.11b ISM              |
| Channels        | Middle                        |
| Signal          | IEEE802.b (Crest factor: 1.0) |

### B. Permittivity

|                                        |          |
|----------------------------------------|----------|
| Frequency (MHz)                        | 2437.000 |
| Relative permittivity (real part)      | 40.082   |
| Relative permittivity (imaginary part) | 13.212   |
| Conductivity (S/m)                     | 1.819    |

### C. SAR Surface and Volume




Maximum location: X=-16.00, Y=-24.00 ; SAR Peak: 0.14 W/kg

### D. SAR 1g & 10g



|                                                       |       |
|-------------------------------------------------------|-------|
| SAR 10g (W/Kg)                                        | 0.030 |
| SAR 1g (W/Kg)                                         | 0.062 |
| Variation (%)                                         | 0.420 |
| Horizontal validation criteria: minimum distance (mm) | --    |
| Vertical validation criteria: SAR ratio M2/M1 (%)     | --    |

### E. Z Axis Scan

|            |       |       |       |       |       |
|------------|-------|-------|-------|-------|-------|
| Z (mm)     | 0.00  | 4.00  | 8.00  | 12.00 | 16.00 |
| SAR (W/Kg) | 0.139 | 0.066 | 0.029 | 0.013 | 0.008 |



### F. 3D Image

| 3D screen shot                                                                     | Hot spot position                                                                    |
|------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------|
|  |  |

## Plot 16


### A. Experimental conditions.

|                 |                               |
|-----------------|-------------------------------|
| Probe           | SN 25/22 EPG0373              |
| ConvF           | 3.96                          |
| Area Scan       | surf_sam_plan.txt             |
| Zoom Scan       | 7x7x8,dx=5mm dy=5mm dz=5mm    |
| Phantom         | Validation plane              |
| Device Position | Body                          |
| Band            | IEEE 802.11b ISM              |
| Channels        | Middle                        |
| Signal          | IEEE802.b (Crest factor: 1.0) |

### B. Permittivity

|                                        |          |
|----------------------------------------|----------|
| Frequency (MHz)                        | 2437.000 |
| Relative permittivity (real part)      | 40.082   |
| Relative permittivity (imaginary part) | 13.212   |
| Conductivity (S/m)                     | 1.819    |

### C. SAR Surface and Volume




Maximum location: X=9.00, Y=-16.00 ; SAR Peak: 0.64 W/kg

### D. SAR 1g & 10g



|                                                       |       |
|-------------------------------------------------------|-------|
| SAR 10g (W/Kg)                                        | 0.109 |
| SAR 1g (W/Kg)                                         | 0.260 |
| Variation (%)                                         | 0.790 |
| Horizontal validation criteria: minimum distance (mm) | --    |
| Vertical validation criteria: SAR ratio M2/M1 (%)     | --    |

### E. Z Axis Scan

|            |       |       |       |       |       |
|------------|-------|-------|-------|-------|-------|
| Z (mm)     | 0.00  | 4.00  | 8.00  | 12.00 | 16.00 |
| SAR (W/Kg) | 0.640 | 0.270 | 0.099 | 0.035 | 0.015 |




### F. 3D Image

| 3D screen shot                                                                     | Hot spot position                                                                    |
|------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------|
|  |  |

## 16. CALIBRATION CERTIFICATES

**Probe-EPGO373 Calibration Certificate**  
**SID750Dipole Calibration Certificate**  
**SID835Dipole Calibration Certificate**  
**SID1800Dipole Calibration Certificate**  
**SID1900Dipole Calibration Certificate**  
**SID2450Dipole Calibration Certificate**  
**SID2600Dipole Calibration Certificate**





## COMOSAR E-Field Probe Calibration Report

Ref : ACR.180.5.22.BES.A

**SHENZHEN BCTC TECHNOLOGY CO., LTD.**  
1 ~2/ F, NO. B FACTORY BUILDING, PENGZHOU INDUSTRIAL  
PARK, FUYUAN 1ST ROAD,  
TANGWEI COMMUNITY, FUHAI STREET, BAO'AN DISTRICT,  
SHENZHEN, GUANGDONG, CHINA  
**MVG COMOSAR DOSIMETRIC E-FIELD PROBE**  
SERIAL NO.: SN 25/22 EPG0373

**Calibrated at MVG**  
**Z.I. de la pointe du diable**  
Technopôle Brest Iroise – 295 avenue Alexis de Rochon  
29280 PLOUZANE - FRANCE

**Calibration date: 06/29/2022**



Accreditations #2-6789  
Scope available on [www.cofrac.fr](http://www.cofrac.fr)

**The use of the Cofrac brand and the accreditation references is prohibited from any reproduction.**

### *Summary:*

This document presents the method and results from an accredited COMOSAR Dosimetric E-Field Probe calibration performed at MVG, using the CALIPROBE test bench, for use with a MVG COMOSAR system only. The test results covered by accreditation are traceable to the International System of Units (SI).

Page: 1/11


**COMOSAR E-FIELD PROBE CALIBRATION REPORT**

Ref: ACR.180.5.22.BES.A

|                                   | <i>Name</i>    | <i>Function</i>         | <i>Date</i> | <i>Signature</i>                                                                    |
|-----------------------------------|----------------|-------------------------|-------------|-------------------------------------------------------------------------------------|
| <i>Prepared by :</i>              | Jérôme Le Gall | Measurement Responsible | 6/30/2022   |  |
| <i>Checked &amp; approved by:</i> | Jérôme Luc     | Technical Manager       | 6/30/2022   |  |
| <i>Authorized by:</i>             | Yann Toutain   | Laboratory Director     | 6/30/2022   |  |

2022.06.30  
13:38:42 +02'00'

|                       | <i>Customer Name</i>                     |
|-----------------------|------------------------------------------|
| <i>Distribution :</i> | Shenzhen BCTC<br>Technology Co.,<br>Ltd. |

| <i>Issue</i> | <i>Name</i>    | <i>Date</i> | <i>Modifications</i> |
|--------------|----------------|-------------|----------------------|
| A            | Jérôme Le Gall | 6/30/2022   | Initial release      |
|              |                |             |                      |
|              |                |             |                      |
|              |                |             |                      |

Page: 2/11

**Template: ACR.DDD.N.YY.MVGB.ISSUE\_COMOSAR\_Probe\_vK**

*This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.*



## TABLE OF CONTENTS

|     |                                       |    |
|-----|---------------------------------------|----|
| 1   | Device Under Test .....               | 4  |
| 2   | Product Description .....             | 4  |
| 2.1 | General Information .....             | 4  |
| 3   | Measurement Method .....              | 4  |
| 3.1 | Linearity .....                       | 4  |
| 3.2 | Sensitivity .....                     | 4  |
| 3.3 | Lower Detection Limit .....           | 5  |
| 3.4 | Isotropy .....                        | 5  |
| 3.1 | Boundary Effect .....                 | 5  |
| 4   | Measurement Uncertainty .....         | 6  |
| 5   | Calibration Measurement Results ..... | 6  |
| 5.1 | Sensitivity in air .....              | 6  |
| 5.2 | Linearity .....                       | 7  |
| 5.3 | Sensitivity in liquid .....           | 8  |
| 5.4 | Isotropy .....                        | 9  |
| 6   | List of Equipment .....               | 10 |

Page: 3/11

**Template: ACR.DDD.N.YY.MVGB.ISSUE\_COMOSAR\_Probe vK**

*This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.*



## 1 DEVICE UNDER TEST

| Device Under Test                        |                                                                         |
|------------------------------------------|-------------------------------------------------------------------------|
| Device Type                              | COMOSAR DOSIMETRIC E FIELD PROBE                                        |
| Manufacturer                             | MVG                                                                     |
| Model                                    | SSE2                                                                    |
| Serial Number                            | SN 25/22 EPG0373                                                        |
| Product Condition (new / used)           | New                                                                     |
| Frequency Range of Probe                 | 0.15 GHz-6GHz                                                           |
| Resistance of Three Dipoles at Connector | Dipole 1: R1=0.234 MΩ<br>Dipole 2: R2=0.195 MΩ<br>Dipole 3: R3=0.250 MΩ |

## 2 PRODUCT DESCRIPTION

### 2.1 GENERAL INFORMATION

MVG's COMOSAR E field Probes are built in accordance to the IEC/IEEE 62209-1528 and FCC KDB865664 D01 standards.



**Figure 1 – MVG COMOSAR Dosimetric E field Probe**

|                                            |        |
|--------------------------------------------|--------|
| Probe Length                               | 330 mm |
| Length of Individual Dipoles               | 2 mm   |
| Maximum external diameter                  | 8 mm   |
| Probe Tip External Diameter                | 2.5 mm |
| Distance between dipoles / probe extremity | 1 mm   |

## 3 MEASUREMENT METHOD

The IEC/IEEE 62209-1528 and FCC KDB865664 D01 standards provide recommended practices for the probe calibrations, including the performance characteristics of interest and methods by which to assess their affect. All calibrations / measurements performed meet the fore mentioned standards.

### 3.1 LINEARITY

The evaluation of the linearity was done in free space using the waveguide, performing a power sweep to cover the SAR range 0.01W/kg to 100W/kg.

### 3.2 SENSITIVITY

The sensitivity factors of the three dipoles were determined using a two step calibration method (air and tissue simulating liquid) using waveguides as outlined in the standards.

Page: 4/11

*Template\_ACR.DDD.N.YY.MVGB.ISSUE\_COMOSAR\_Probe\_vK*

*This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.*



### 3.3 LOWER DETECTION LIMIT

The lower detection limit was assessed using the same measurement set up as used for the linearity measurement. The required lower detection limit is 10 mW/kg.

### 3.4 ISOTROPY

The axial isotropy was evaluated by exposing the probe to a reference wave from a standard dipole with the dipole mounted under the flat phantom in the test configuration suggested for system validations and checks. The probe was rotated along its main axis from 0 to 360 degrees in 15-degree steps. The hemispherical isotropy is determined by inserting the probe in a thin plastic box filled with tissue-equivalent liquid, with the plastic box illuminated with the fields from a half wave dipole. The dipole is rotated about its axis (0°–180°) in 15° increments. At each step the probe is rotated about its axis (0°–360°).

### 3.1 BOUNDARY EFFECT

The boundary effect is defined as the deviation between the SAR measured data and the expected exponential decay in the liquid when the probe is oriented normal to the interface. To evaluate this effect, the liquid filled flat phantom is exposed to fields from either a reference dipole or waveguide. With the probe normal to the phantom surface, the peak spatial average SAR is measured and compared to the analytical value at the surface.

The boundary effect uncertainty can be estimated according to the following uncertainty approximation formula based on linear and exponential extrapolations between the surface and  $d_{be} + d_{step}$  along lines that are approximately normal to the surface:

$$SAR_{uncertainty} [\%] = \Delta SAR_{be} \frac{(d_{be} + d_{step})^2}{2d_{step}} \frac{(e^{-d_{be}/\delta} - e^{-(d_{be} + d_{step})/\delta})}{\delta/2} \quad \text{for } (d_{be} + d_{step}) < 10 \text{ mm}$$

where

|                     |                                                                                                                                                                                                      |
|---------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| $SAR_{uncertainty}$ | is the uncertainty in percent of the probe boundary effect                                                                                                                                           |
| $d_{be}$            | is the distance between the surface and the closest <i>zoom-scan</i> measurement point, in millimetre                                                                                                |
| $\Delta_{step}$     | is the separation distance between the first and second measurement points that are closest to the phantom surface, in millimetre, assuming the boundary effect at the second location is negligible |
| $\delta$            | is the minimum penetration depth in millimetres of the head tissue-equivalent liquids defined in this standard, i.e., $\delta \approx 14 \text{ mm}$ at 3 GHz;                                       |
| $\Delta SAR_{be}$   | in percent of SAR is the deviation between the measured SAR value, at the distance $d_{be}$ from the boundary, and the analytical SAR value.                                                         |

The measured worst case boundary effect SAR uncertainty[%] for scanning distances larger than 4mm is 1.0% Limit ,2%).



#### 4 MEASUREMENT UNCERTAINTY

The guidelines outlined in the IEC/IEEE 62209-1528 and FCC KDB865664 D01 standards were followed to generate the measurement uncertainty associated with an E-field probe calibration using the waveguide technique. All uncertainties listed below represent an expanded uncertainty expressed at approximately the 95% confidence level using a coverage factor of  $k=2$ , traceable to the Internationally Accepted Guides to Measurement Uncertainty.

| Uncertainty analysis of the probe calibration in waveguide |                       |                          |         |    |                          |
|------------------------------------------------------------|-----------------------|--------------------------|---------|----|--------------------------|
| ERROR SOURCES                                              | Uncertainty value (%) | Probability Distribution | Divisor | ci | Standard Uncertainty (%) |
| Expanded uncertainty<br>95 % confidence level $k = 2$      |                       |                          |         |    | 14 %                     |

#### 5 CALIBRATION MEASUREMENT RESULTS

| Calibration Parameters |             |
|------------------------|-------------|
| Liquid Temperature     | 20 +/- 1 °C |
| Lab Temperature        | 20 +/- 1 °C |
| Lab Humidity           | 30-70 %     |


##### 5.1 SENSITIVITY IN AIR

|                                                           |                                                           |                                                           |
|-----------------------------------------------------------|-----------------------------------------------------------|-----------------------------------------------------------|
| Normx dipole<br>1 ( $\mu\text{V}/(\text{V}/\text{m})^2$ ) | Normy dipole<br>2 ( $\mu\text{V}/(\text{V}/\text{m})^2$ ) | Normz dipole<br>3 ( $\mu\text{V}/(\text{V}/\text{m})^2$ ) |
| 1.19                                                      | 0.77                                                      | 1.05                                                      |


|                      |                      |                      |
|----------------------|----------------------|----------------------|
| DCP dipole 1<br>(mV) | DCP dipole 2<br>(mV) | DCP dipole 3<br>(mV) |
| 108                  | 109                  | 110                  |

Calibration curves  $e_i=f(V)$  ( $i=1,2,3$ ) allow to obtain E-field value using the formula:

$$E = \sqrt{E_1^2 + E_2^2 + E_3^2}$$



## 5.2 LINEARITY



**Linearity:  $\pm 1.77\%$  ( $\pm 0.08\text{dB}$ )**

Page: 7/11

*Template: ACR.DDD.N.YY.MVGB.ISSUE\_COMOSAR\_Probe\_vK*

*This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.*



### 5.3 SENSITIVITY IN LIQUID

| Liquid | Frequency<br>(MHz +/-<br>100MHz) | ConvF |
|--------|----------------------------------|-------|
| HL450* | 450*                             | 3.00* |
| BL450* | 450*                             | 2.83* |
| HL750  | 750                              | 2.96  |
| BL750  | 750                              | 3.07  |
| HL850  | 835                              | 3.01  |
| BL850  | 835                              | 3.13  |
| HL900  | 900                              | 3.08  |
| BL900  | 900                              | 3.18  |
| HL1800 | 1800                             | 3.35  |
| BL1800 | 1800                             | 3.42  |
| HL1900 | 1900                             | 3.27  |
| BL1900 | 1900                             | 3.55  |
| HL2100 | 2100                             | 3.77  |
| BL2100 | 2100                             | 3.92  |
| HL2300 | 2300                             | 3.77  |
| BL2300 | 2300                             | 3.94  |
| HL2450 | 2450                             | 3.96  |
| BL2450 | 2450                             | 4.13  |
| HL2600 | 2600                             | 3.63  |
| BL2600 | 2600                             | 3.79  |
| HL5200 | 5200                             | 2.72  |
| BL5200 | 5200                             | 2.45  |
| HL5400 | 5400                             | 2.92  |
| BL5400 | 5400                             | 2.74  |
| HL5600 | 5600                             | 3.09  |
| BL5600 | 5600                             | 2.90  |
| HL5800 | 5800                             | 2.86  |
| BL5800 | 5800                             | 2.72  |


\* Frequency not cover by COFRAC scope, calibration not accredited

LOWER DETECTION LIMIT: 7mW/kg



#### 5.4 ISOTROPY

##### HL1800 MHz



Page: 9/11

**Template: ACR.DDD.N.YY.MVGB.ISSUE\_COMOSAR\_Probe\_vK**

*This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.*



## 6 LIST OF EQUIPMENT

| Equipment Summary Sheet            |                      |                         |                                               |                                               |
|------------------------------------|----------------------|-------------------------|-----------------------------------------------|-----------------------------------------------|
| Equipment Description              | Manufacturer / Model | Identification No.      | Current Calibration Date                      | Next Calibration Date                         |
| CALIPROBE Test Bench               | Version 2            | NA                      | Validated. No cal required.                   | Validated. No cal required.                   |
| Network Analyzer                   | Rohde & Schwarz ZVM  | 100203                  | 08/2021                                       | 08/2024                                       |
| Network Analyzer                   | Agilent 8753ES       | MY40003210              | 10/2019                                       | 10/2022                                       |
| Network Analyzer – Calibration kit | HP 85033D            | 3423A08186              | 06/2021                                       | 06/2027                                       |
| Multimeter                         | Keithley 2000        | 1160271                 | 02/2020                                       | 02/2023                                       |
| Signal Generator                   | Rohde & Schwarz SMB  | 106589                  | 03/2022                                       | 03/2025                                       |
| Amplifier                          | MVG                  | MODU-023-C-0002         | Characterized prior to test. No cal required. | Characterized prior to test. No cal required. |
| Power Meter                        | NI-USB 5680          | 170100013               | 06/2021                                       | 06/2024                                       |
| Power Meter                        | Rohde & Schwarz NRVD | 832839-056              | 11/2019                                       | 11/2022                                       |
| Directional Coupler                | Krytar 158020        | 131467                  | Characterized prior to test. No cal required. | Characterized prior to test. No cal required. |
| Waveguide                          | MVG                  | SN 32/16 WG4_1          | Validated. No cal required.                   | Validated. No cal required.                   |
| Liquid transition                  | MVG                  | SN 32/16 WGLIQ_0G900_1  | Validated. No cal required.                   | Validated. No cal required.                   |
| Waveguide                          | MVG                  | SN 32/16 WG6_1          | Validated. No cal required.                   | Validated. No cal required.                   |
| Liquid transition                  | MVG                  | SN 32/16 WGLIQ_1G500_1  | Validated. No cal required.                   | Validated. No cal required.                   |
| Waveguide                          | MVG                  | SN 32/16 WG8_1          | Validated. No cal required.                   | Validated. No cal required.                   |
| Liquid transition                  | MVG                  | SN 32/16 WGLIQ_1G800B_1 | Validated. No cal required.                   | Validated. No cal required.                   |
| Liquid transition                  | MVG                  | SN 32/16 WGLIQ_1G800H_1 | Validated. No cal required.                   | Validated. No cal required.                   |
| Waveguide                          | MVG                  | SN 32/16 WG10_1         | Validated. No cal required.                   | Validated. No cal required.                   |
| Liquid transition                  | MVG                  | SN 32/16 WGLIQ_3G500_1  | Validated. No cal required.                   | Validated. No cal required.                   |
| Waveguide                          | MVG                  | SN 32/16 WG12_1         | Validated. No cal required.                   | Validated. No cal required.                   |

Page: 10/11

Template: ACR.DDD.N.YY.MVGB.ISSUE\_COMOSAR\_Probe\_vK

This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.

**COMOSAR E-FIELD PROBE CALIBRATION REPORT**

Ref: ACR.180.5.22.BES.A

|                                  |              |                           |                                |                                |
|----------------------------------|--------------|---------------------------|--------------------------------|--------------------------------|
| Liquid transition                | MVG          | SN 32/16<br>WGLIQ_5G000_1 | Validated. No cal<br>required. | Validated. No cal<br>required. |
| Temperature / Humidity<br>Sensor | Testo 184 H1 | 44225320                  | 06/2021                        | 06/2024                        |

*Page: 11/11****Template\_ACR.DDD.N.YY.MVGB.ISSUE\_COMOSAR\_Probe vK***

*This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.*



## SAR Reference Dipole Calibration Report

Ref : ACR.329.8.21.BES.A

### **SHENZHEN BCTC TECHNOLOGY CO., LTD.**

**1 ~2/ F, NO. B FACTORY BUILDING, PENGZHOU  
INDUSTRIAL PARK, FUYUAN 1ST ROAD,  
TANGWEI COMMUNITY, FUHAI STREET, BAO'AN  
DISTRICT, SHENZHEN, GUANGDONG, CHINA  
MVG COMOSAR REFERENCE DIPOLE**

**FREQUENCY: 750 MHZ**

**SERIAL NO.: SN 47/21 DIP 0G750-620**

**Calibrated at MVG**

**Z.I. de la pointe du diable**

**Technopôle Brest Iroise – 295 avenue Alexis de Rochon  
29280 PLOUZANE - FRANCE**

**Calibration date: 11/25/2021**



Accreditations #2-6789 and #2-6814  
Scope available on [www.cofrac.fr](http://www.cofrac.fr)

**The use of the Cofrac brand and the accreditation references is prohibited from any reproduction.**


#### *Summary:*

This document presents the method and results from an accredited SAR reference dipole calibration performed in MVG using the COMOSAR test bench. All calibration results are traceable to national metrology institutions.

Page: 1/13


**SAR REFERENCE DIPOLE CALIBRATION REPORT**

Ref: ACR.329.8.21.BES.A

|                      | <i>Name</i>  | <i>Function</i>     | <i>Date</i> | <i>Signature</i>                                                                                                      |
|----------------------|--------------|---------------------|-------------|-----------------------------------------------------------------------------------------------------------------------|
| <i>Prepared by :</i> | Jérôme Luc   | Technical Manager   | 11/25/2021  |                                    |
| <i>Checked by :</i>  | Jérôme Luc   | Technical Manager   | 11/25/2021  |                                    |
| <i>Approved by :</i> | Yann Toutain | Laboratory Director | 11/25/2021  | <br>2021.11.25<br>11:51:55 +01'00' |

|                       | <i>Customer Name</i>                     |
|-----------------------|------------------------------------------|
| <i>Distribution :</i> | Shenzhen BCTC<br>Technology Co.,<br>Ltd. |

| <i>Issue</i> | <i>Name</i> | <i>Date</i> | <i>Modifications</i> |
|--------------|-------------|-------------|----------------------|
| A            | Jérôme Luc  | 11/25/2021  | Initial release      |
|              |             |             |                      |
|              |             |             |                      |
|              |             |             |                      |

Page: 2/13

**Template: ACR.DDD.N.YY.MVGB.ISSUE SAR Reference Dipole vJ**

*This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.*



## TABLE OF CONTENTS

|     |                                                |    |
|-----|------------------------------------------------|----|
| 1   | Introduction.....                              | 4  |
| 2   | Device Under Test .....                        | 4  |
| 3   | Product Description .....                      | 4  |
| 3.1 | General Information .....                      | 4  |
| 4   | Measurement Method .....                       | 5  |
| 4.1 | Return Loss Requirements .....                 | 5  |
| 4.2 | Mechanical Requirements .....                  | 5  |
| 5   | Measurement Uncertainty .....                  | 5  |
| 5.1 | Return Loss .....                              | 5  |
| 5.2 | Dimension Measurement .....                    | 5  |
| 5.3 | Validation Measurement .....                   | 5  |
| 6   | Calibration Measurement Results .....          | 6  |
| 6.1 | Return Loss and Impedance In Head Liquid ..... | 6  |
| 6.2 | Return Loss and Impedance In Body Liquid ..... | 6  |
| 6.3 | Mechanical Dimensions .....                    | 7  |
| 7   | Validation measurement .....                   | 7  |
| 7.1 | Head Liquid Measurement .....                  | 8  |
| 7.2 | SAR Measurement Result With Head Liquid .....  | 8  |
| 7.3 | Body Liquid Measurement .....                  | 11 |
| 7.4 | SAR Measurement Result With Body Liquid .....  | 12 |
| 8   | List of Equipment .....                        | 13 |



## 1 INTRODUCTION

This document contains a summary of the requirements set forth by the IEC/IEEE 62209-1528 and FCC KDB865664 D01 standards for reference dipoles used for SAR measurement system validations and the measurements that were performed to verify that the product complies with the fore mentioned standards.

## 2 DEVICE UNDER TEST

| Device Under Test              |                                  |
|--------------------------------|----------------------------------|
| Device Type                    | COMOSAR 750 MHz REFERENCE DIPOLE |
| Manufacturer                   | MVG                              |
| Model                          | SID750                           |
| Serial Number                  | SN 47/21 DIP 0G750-620           |
| Product Condition (new / used) | New                              |

## 3 PRODUCT DESCRIPTION

### 3.1 GENERAL INFORMATION

MVG's COMOSAR Validation Dipoles are built in accordance to the IEC/IEEE 62209-1528 and FCC KDB865664 D01 standards. The product is designed for use with the COMOSAR test bench only.



**Figure 1 – MVG COMOSAR Validation Dipole**



#### 4 MEASUREMENT METHOD

The IEC/IEEE 62209-1528 and FCC KDB865664 D01 standards provide requirements for reference dipoles used for system validation measurements. The following measurements were performed to verify that the product complies with the fore mentioned standards.

##### 4.1 RETURN LOSS REQUIREMENTS

The dipole used for SAR system validation measurements and checks must have a return loss of -20 dB or better. The return loss measurement shall be performed against a liquid filled flat phantom, with the phantom constructed as outlined in the fore mentioned standards. A direct method is used with a network analyser and its calibration kit, both with a valid ISO17025 calibration.

##### 4.2 MECHANICAL REQUIREMENTS

The IEC/IEEE 62209-1528 and FCC KDB865664 D01 standards specify the mechanical components and dimensions of the validation dipoles, with the dimension's frequency and phantom shell thickness dependent. The COMOSAR test bench employs a 2 mm phantom shell thickness therefore the dipoles sold for use with the COMOSAR test bench comply with the requirements set forth for a 2 mm phantom shell thickness. A direct method is used with a ISO17025 calibrated caliper.

#### 5 MEASUREMENT UNCERTAINTY

All uncertainties listed below represent an expanded uncertainty expressed at approximately the 95% confidence level using a coverage factor of  $k=2$ , traceable to the Internationally Accepted Guides to Measurement Uncertainty.

##### 5.1 RETURN LOSS

The following uncertainties apply to the return loss measurement:

| Frequency band | Expanded Uncertainty on Return Loss |
|----------------|-------------------------------------|
| 400-6000MHz    | 0.08 LIN                            |

##### 5.2 DIMENSION MEASUREMENT

The following uncertainties apply to the dimension measurements:

| Length (mm) | Expanded Uncertainty on Length |
|-------------|--------------------------------|
| 0 - 300     | 0.20 mm                        |
| 300 - 450   | 0.44 mm                        |

##### 5.3 VALIDATION MEASUREMENT

The guidelines outlined in the IEC/IEEE 62209-1528 and FCC KDB865664 D01 standards were followed to generate the measurement uncertainty for validation measurements.



**SAR REFERENCE DIPOLE CALIBRATION REPORT**

Ref: ACR.329.8.21.BES.A

| Scan Volume | Expanded Uncertainty |
|-------------|----------------------|
| 1 g         | 19 % (SAR)           |
| 10 g        | 19 % (SAR)           |


## 6 CALIBRATION MEASUREMENT RESULTS

### 6.1 RETURN LOSS AND IMPEDANCE IN HEAD LIQUID



| Frequency (MHz) | Return Loss (dB) | Requirement (dB) | Impedance                   |
|-----------------|------------------|------------------|-----------------------------|
| 750             | -24.50           | -20              | $55.7 \Omega - 1.7 j\Omega$ |

### 6.2 RETURN LOSS AND IMPEDANCE IN BODY LIQUID



| Frequency (MHz) | Return Loss (dB) | Requirement (dB) | Impedance                   |
|-----------------|------------------|------------------|-----------------------------|
| 750             | -27.04           | -20              | $53.8 \Omega + 2.3 j\Omega$ |

Page: 6/13

Template: ACR.DDD.N.YY.MVGB.ISSUE SAR Reference Dipole vJ

This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.