

D750V3 Sn:1101 (2/2)


In Collaboration with
TTL S P E A G
CALIBRATION LABORATORY

Add: No.31 Xuyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2304
E-mail: csl@chinafl.com http://www.chinafl.com

DASY5 Validation Report for Head TSL Date: 09.13.2017
Test Laboratory: CTTL, Beijing, China
DUT: Dipole 750 MHz; Type: D750V3; Serial: D750V3 - SN: 1101
Communication System: UID 0, CW; Frequency: 750 MHz; Duty Cycle: 1:1
Medium parameters used: $f = 750$ MHz; $\sigma = 0.879$ S/m; $\epsilon_r = 41.54$; $\rho = 1000$ kg/m³
Phantom section: Left Section
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)
DASY5 Configuration:

- Probe: EX3DV4 - SN7433; ConvF(10.01, 10.01, 10.01); Calibrated: 9/26/2016;
- Sensor-Surface: 1.8mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1331; Calibrated: 1/19/2017
- Phantom: Triple Flat Phantom 5.1C; Type: QD 000 P51 CA; Serial: 1161/1
- Measurement SW: DASY52, Version 52.10 (0); SEMCAD X Version 14.6.10 (7417)

Dipole Calibration/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm
Reference Value = 53.10 V/m; Power Drift = -0.05 dB
Peak SAR (extrapolated) = 3.17 W/kg
SAR(1 g) = 2.05 W/kg; SAR(10 g) = 1.34 W/kg
Maximum value of SAR (measured) = 2.77 W/kg




Certificate No: Z17-97134 Page 5 of 8

In Collaboration with
TTL S P E A G
CALIBRATION LABORATORY

Add: No.31 Xuyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2304
E-mail: csl@chinafl.com http://www.chinafl.com

Impedance Measurement Plot for Head TSL



Certificate No: Z17-97134 Page 6 of 8


In Collaboration with
TTL S P E A G
CALIBRATION LABORATORY

Add: No.31 Xuyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2304
E-mail: csl@chinafl.com http://www.chinafl.com

DASY5 Validation Report for Body TSL Date: 09.13.2017
Test Laboratory: CTTL, Beijing, China
DUT: Dipole 750 MHz; Type: D750V3; Serial: D750V3 - SN: 1101
Communication System: UID 0, CW; Frequency: 750 MHz; Duty Cycle: 1:1
Medium parameters used: $f = 750$ MHz; $\sigma = 0.946$ S/m; $\epsilon_r = 55.41$; $\rho = 1000$ kg/m³
Phantom section: Center Section
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)
DASY5 Configuration:

- Probe: EX3DV4 - SN7433; ConvF(9.83, 9.83, 9.83); Calibrated: 9/26/2016;
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1331; Calibrated: 1/19/2017
- Phantom: Triple Flat Phantom 5.1C; Type: QD 000 P51 CA; Serial: 1161/1
- Measurement SW: DASY52, Version 52.10 (0); SEMCAD X Version 14.6.10 (7417)

Dipole Calibration/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm
Reference Value = 53.35 V/m; Power Drift = -0.02 dB
Peak SAR (extrapolated) = 3.27 W/kg
SAR(1 g) = 2.15 W/kg; SAR(10 g) = 1.42 W/kg
Maximum value of SAR (measured) = 2.88 W/kg




Certificate No: Z17-97134 Page 7 of 8

In Collaboration with
TTL S P E A G
CALIBRATION LABORATORY

Add: No.31 Xuyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2304
E-mail: csl@chinafl.com http://www.chinafl.com

Impedance Measurement Plot for Body TSL



Certificate No: Z17-97134 Page 8 of 8

D835V2 Sn:4d023

In Collaboration with
TTL CALIBRATION LABORATORY
SPEAG
CNAS 中国认可 国际互认 校准 CALIBRATION CNAS L5570

Address: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62306633-2079 Fax: +86-10-62306633-2504
E-mail: cti@chinaatl.com http://www.chinaatl.com

Client: **SRTC** Certificate No: **Z17-97135**

CALIBRATION CERTIFICATE

Object: D835V2 - SN: 4d023

Calibration Procedure(s): FF-Z11-003-01
Calibration Procedures for dipole validation kits

Calibration date: September 13, 2017

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements(SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature(23.3)°C and humidity<70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration
Power Meter NRV0	102196	02-Mar-17 (CTTL, No.J17X01254)	Mar-18
Power sensor NRV-Z5	100596	02-Mar-17 (CTTL, No.J17X01254)	Mar-18
Reference Probe EX3D/4	SN 7433	28-Sep-16(SPEAG, No EX3-7433_Sep16)	Sep-17
DAE4	SN 1331	19-Jan-17(CTTL-SPEAG, No Z17-87015)	Jan-18
Secondary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	13-Jan-17 (CTTL, No.J17X00286)	Jan-18
Network Analyzer E5071C	MY46110673	13-Jan-17 (CTTL, No.J17X00286)	Jan-18

Calibrated by: Zhao Jing, SAR Test Engineer, Signature: [Signature]

Reviewed by: Yu Zongying, SAR Test Engineer, Signature: [Signature]

Approved by: Qi Dianyuan, SAR Project Leader, Signature: [Signature]

Issued: September 16, 2017

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z17-97135 Page 1 of 8

In Collaboration with
TTL CALIBRATION LABORATORY
SPEAG
CNAS 中国认可 国际互认 校准 CALIBRATION CNAS L5570

Address: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62306633-2079 Fax: +86-10-62306633-2504
E-mail: cti@chinaatl.com http://www.chinaatl.com

Client: **SRTC** Certificate No: **Z17-97135**

Glossary:

TSL: tissue simulating liquid
ConvF: sensitivity in TSL / NDRMx,y,z
N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Measurement procedure for assessment of specific absorption rate of human exposure to radio frequency fields from hand-held and body-mounted wireless communication devices- Part 1: Device used next to the ear (Frequency range of 300MHz to 6GHz)", July 2016
- IEC 62209-2, "Procedure to measure the Specific Absorption Rate (SAR) For wireless communication devices used in close proximity to the human body (frequency range of 30MHz to 6GHz)", March 2010
- KDB855664, SAR Measurement Requirements for 100 MHz to 6 GHz

Additional Documentation:

- DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is oriented at the frequency indicated. The dipole is positioned exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: Z17-97135 Page 2 of 8

In Collaboration with
TTL CALIBRATION LABORATORY
SPEAG
CNAS 中国认可 国际互认 校准 CALIBRATION CNAS L5570

Address: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62306633-2079 Fax: +86-10-62306633-2504
E-mail: cti@chinaatl.com http://www.chinaatl.com

Client: **SRTC** Certificate No: **Z17-97135**

Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY92	52.10.0.1446
Extrapolation	Advanced Extrapolation	
Phantom	Triple Flat Phantom 5.1C	
Distance Dipole Center - TSL	15 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	835 MHz ± 1 MHz	

Head TSL parameters

The following parameters and calculations were applied.

Nominal Head TSL parameters	Temperature	Permittivity	Conductivity
	22.0 °C	41.5	0.90 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	41.3 ± 6 %	0.90 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	2.35 mW / g
SAR measured	250 mW input power	2.35 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	9.37 mW / g ± 18.8 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	1.52 mW / g
SAR measured	250 mW input power	1.52 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	6.06 mW / g ± 18.7 % (k=2)

Body TSL parameters

The following parameters and calculations were applied.

Nominal Body TSL parameters	Temperature	Permittivity	Conductivity
	22.0 °C	55.2	0.97 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	55.7 ± 6 %	0.96 mho/m ± 6 %
Body TSL temperature change during test	<1.0 °C	---	---

SAR result with Body TSL

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	2.34 mW / g
SAR measured	250 mW input power	2.34 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	9.47 mW / g ± 18.8 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Body TSL	Condition	1.53 mW / g
SAR measured	250 mW input power	1.53 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	6.17 mW / g ± 18.7 % (k=2)

Certificate No: Z17-97135 Page 3 of 8

In Collaboration with
TTL CALIBRATION LABORATORY
SPEAG
CNAS 中国认可 国际互认 校准 CALIBRATION CNAS L5570

Address: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62306633-2079 Fax: +86-10-62306633-2504
E-mail: cti@chinaatl.com http://www.chinaatl.com

Client: **SRTC** Certificate No: **Z17-97135**

Appendix (Additional assessments outside the scope of CNAS L5070)

Antenna Parameters with Head TSL

Impedance, transformed to feed point	51.0Ω - 2.79jΩ
Return Loss	- 30.7dB

Antenna Parameters with Body TSL

Impedance, transformed to feed point	46.6Ω - 3.61jΩ
Return Loss	- 25.8dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.495 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
-----------------	-------

Certificate No: Z17-97135 Page 4 of 8

D835V2 Sn:4d023




In Collaboration with
TTL Calibration Laboratory
Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504
E-mail: cti@chinaatl.com http://www.chinaatl.cn

DASY5 Validation Report for Head TSL Date: 09.13.2017
Test Laboratory: CTTL, Beijing, China
DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN: 4d023
Communication System: UTD 0, CW; Frequency: 835 MHz; Duty Cycle: 1:1
Medium parameters used: $f = 835$ MHz; $\sigma = 0.903$ S/m; $\epsilon_r = 41.34$; $\rho = 1000$ kg/m³
Phantom section: Left Section
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY5 Configuration:


- Probe: EX3DV4 - SN7433; ConvF(9.82, 9.82); Calibrated: 9/26/2016;
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1331; Calibrated: 1/19/2017
- Phantom: Triple Flat Phantom 5.1C; Type: QD 000 P51 CA; Serial: 1161/1
- Measurement SW: DASY52, Version 52.10 (0); SEMCAD X Version 14.6.10 (7417)

Dipole Calibration/Zoom Scan (7x7x7) (7x7x7)/Cube 0; Measurement grid: dx=5mm, dy=5mm, dz=5mm
Reference Value = 56.28V/m; Power Drift = -0.02 dB
Peak SAR (extrapolated) = 3.66 W/kg
SAR(1 g) = 2.35 W/kg; SAR(10 g) = 1.52 W/kg
Maximum value of SAR (measured) = 3.21 W/kg



In Collaboration with
TTL Calibration Laboratory
Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504
E-mail: cti@chinaatl.com http://www.chinaatl.cn

Impedance Measurement Plot for Head TSL




In Collaboration with
TTL Calibration Laboratory
Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504
E-mail: cti@chinaatl.com http://www.chinaatl.cn

DASY5 Validation Report for Body TSL Date: 09.13.2017
Test Laboratory: CTTL, Beijing, China
DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN: 4d023
Communication System: UTD 0, CW; Frequency: 835 MHz; Duty Cycle: 1:1
Medium parameters used: $f = 835$ MHz; $\sigma = 0.958$ S/m; $\epsilon_r = 55.68$; $\rho = 1000$ kg/m³
Phantom section: Center Section
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY5 Configuration:


- Probe: EX3DV4 - SN7433; ConvF(9.5, 9.5); Calibrated: 9/26/2016;
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1331; Calibrated: 1/19/2017
- Phantom: Triple Flat Phantom 5.1C; Type: QD 000 P51 CA; Serial: 1161/1
- Measurement SW: DASY52, Version 52.10 (0); SEMCAD X Version 14.6.10 (7417)

Dipole Calibration/Zoom Scan (7x7x7) (7x7x7)/Cube 0; Measurement grid: dx=5mm, dy=5mm, dz=5mm
Reference Value = 56.17 V/m; Power Drift = -0.01 dB
Peak SAR (extrapolated) = 3.57 W/kg
SAR(1 g) = 2.34 W/kg; SAR(10 g) = 1.53 W/kg
Maximum value of SAR (measured) = 3.15 W/kg



In Collaboration with
TTL Calibration Laboratory
Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504
E-mail: cti@chinaatl.com http://www.chinaatl.cn

Impedance Measurement Plot for Body TSL



D1800V2 Sn:2d084



Address: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42334633-2079 Fax: +86-10-42334633-2504
E-mail: cti@chinaatl.com http://www.chinaatl.cn

Client: SRTC Certificate No: Z17-97138

CALIBRATION CERTIFICATE

Object: D1800V2 - SN: 2d084

Calibration Procedure(s): FF-Z11-003-01
Calibration Procedures for dipole validation kits

Calibration date: September 15, 2017

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements(SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature(2±3) °C and humidity<70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration
Power Meter NRP2	102196	02-Mar-17 (CTTL No.J17X01254)	Mar-18
Power sensor NRP-Z91	100596	02-Mar-17 (CTTL No.J17X01254)	Mar-18
Reference Probe EX3DV4	SN 7433	26-Sep-16(SPEAG No EX3-7433_Sep16)	Sep-17
D4E4	SN 1331	19-Jan-17(CTTL-SPEAG No.Z17-97015)	Jan-18
Secondary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	13-Jan-17 (CTTL No.J17X00286)	Jan-18
Network Analyzer E5071C	MY46110673	13-Jan-17 (CTTL No.J17X00286)	Jan-18

Calibrated by:	Name	Function	Signature
	Zhao Jing	SAR Test Engineer	
Reviewed by:	Name	Function	Signature
	Yu Zongying	SAR Test Engineer	
Approved by:	Name	Function	Signature
	Qi Dianyuan	SAR Project Leader	

Issued: September 18, 2017

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z17-97138 Page 1 of 8



Address: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42334633-2079 Fax: +86-10-42334633-2504
E-mail: cti@chinaatl.com http://www.chinaatl.cn

Glossary:
TSL: tissue simulating liquid
ConvF: sensitivity in TSL / NORMx,y,z
N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) For hand-held devices used in close proximity to the ear (frequency range of 300MHz to 3GHz)", February 2005
- IEC 62209-2, "Procedure to measure the Specific Absorption Rate (SAR) For wireless communication devices used in close proximity to the human body (frequency range of 30MHz to 6GHz)", March 2010
- KDB85664, SAR Measurement Requirements for 100 MHz to 6 GHz

Additional Documentation:
e) DASy4/5 System Handbook

- Methods Applied and Interpretation of Parameters:**
- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
 - Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
 - Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
 - Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
 - SAR measured:** SAR measured at the stated antenna input power.
 - SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
 - SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z17-97138 Page 2 of 8



Address: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42334633-2079 Fax: +86-10-42334633-2504
E-mail: cti@chinaatl.com http://www.chinaatl.cn

Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY52	52.10.0.1446
Extrapolation	Advanced Extrapolation	
Phantom	Triple Flat Phantom 5.1C	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	1800 MHz ± 1 MHz	

Head TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	40.0	1.40 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	40.4 ± 6 %	1.42 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	9.79 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	38.9 mW / g ± 18.8 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	250 mW input power	5.12 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	20.4 mW / g ± 18.7 % (k=2)

Body TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	53.3	1.52 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	53.8 ± 6 %	1.50 mho/m ± 6 %
Body TSL temperature change during test	<1.0 °C	---	---

SAR result with Body TSL

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	250 mW input power	9.84 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	39.7 mW / g ± 18.8 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Body TSL	Condition	
SAR measured	250 mW input power	5.18 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	20.8 mW / g ± 18.7 % (k=2)

Certificate No: Z17-97138 Page 3 of 8



Address: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42334633-2079 Fax: +86-10-42334633-2504
E-mail: cti@chinaatl.com http://www.chinaatl.cn

Appendix (Additional assessments outside the scope of CNAS L0570)

Antenna Parameters with Head TSL

Impedance, transformed to feed point	49.3Ω- 1.55Ω
Return Loss	- 35.4dB

Antenna Parameters with Body TSL

Impedance, transformed to feed point	48.0Ω- 1.32Ω
Return Loss	- 27.1dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.318 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.


The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data


Manufactured by	SPEAG
-----------------	-------

Certificate No: Z17-97138 Page 4 of 8


D1800V2 Sn:2d084




Certificate No: Z17-97138 Page 5 of 8



Certificate No: Z17-97138 Page 6 of 8



Certificate No: Z17-97138 Page 7 of 8



Certificate No: Z17-97138 Page 8 of 8

D2000V2 Sn:1009



Client: SRTC Certificate No: Z18-97021

CALIBRATION CERTIFICATE			
Object	D2000V2 - SN: 1009		
Calibration Procedure(s)	FF-Z11-003-01 Calibration Procedures for dipole validation kits		
Calibration date:	February 1, 2018		
This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements(SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.			
All calibrations have been conducted in the closed laboratory facility: environment temperature:(22±3)°C and humidity:70%.			
Calibration Equipment used (MATE critical for calibration)			
Primary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration
Power Meter NRVD	102196	02-Mar-17 (CTTL, No.J17X01254)	Mar-18
Power sensor NRV-Z5	100696	02-Mar-17 (CTTL, No.J17X01254)	Mar-18
Reference Probe EX3DV4 DAE4	SN 7464 SN 1525	12-Sep-17 (SPEAG, No.EX3-7464_Sep17) 02-Oct-17 (SPEAG, No.DAE4-1525_Oct17)	Sep-18 Oct-18
Secondary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY46071430	23-Jan-18 (CTTL, No.J18X00560)	Jan-19
Network Analyzer E6071C	MY46110673	24-Jan-18 (CTTL, No.J18X00561)	Jan-19
Calibrated by:	Name	Function	Signature
	Zhao Jing	SAR Test Engineer	
Reviewed by:	Lin Hao	SAR Test Engineer	
Approved by:	Qi Dianyuan	SAR Project Leader	
Issued: February 4, 2018			
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.			

Certificate No: Z18-97021 Page 1 of 4



Glossary:
TSL tissue simulating liquid
ConvF sensitivity in TSL / NORM_{x,y,z}
N/A not applicable or not measured

Calibration is Performed According to the Following Standards:
a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
b) IEC 62209-1, "Measurement procedure for assessment of specific absorption rate of human exposure to radio frequency fields from hand-held and body-mounted wireless communication devices- Part 1: Device used next to the ear (Frequency range of 300MHz to 6GHz)", July 2016
c) IEC 62209-2, "Procedure to measure the Specific Absorption Rate (SAR) For wireless communication devices used in close proximity to the human body (frequency range of 30MHz to 6GHz)", March 2010
d) KDB658564, SAR Measurement Requirements for 100 MHz to 6 GHz

Additional Documentation:
e) DASy4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z18-97021 Page 2 of 4



Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY52	52.10.0.1446
Extrapolation	Advanced Extrapolation	
Phantom	Triple Flat Phantom 5.1C	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	2000 MHz ± 1 MHz	

Head TSL parameters

The following parameters and calculations were applied

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	40.0	1.40 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	38.9 ± 6 %	1.42 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	10.2 mW/g
SAR for nominal Head TSL parameters	normalized to 1W	40.3 mW/g ± 18.8 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	250 mW input power	5.17 mW/g
SAR for nominal Head TSL parameters	normalized to 1W	20.6 mW/g ± 18.7 % (k=2)

Body TSL parameters

The following parameters and calculations were applied

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	53.3	1.52 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	51.8 ± 6 %	1.56 mho/m ± 6 %
Body TSL temperature change during test	<1.0 °C	---	---

SAR result with Body TSL

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	250 mW input power	10.3 mW/g
SAR for nominal Body TSL parameters	normalized to 1W	40.3 mW/g ± 18.8 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Body TSL	Condition	
SAR measured	250 mW input power	5.18 mW/g
SAR for nominal Body TSL parameters	normalized to 1W	20.4 mW/g ± 18.7 % (k=2)

Certificate No: Z18-97021 Page 3 of 4



Appendix (Additional assessments outside the scope of CNAS L0570)

Antenna Parameters with Head TSL

Impedance, transformed to feed point	48.8C- 2.06jΩ
Return Loss	- 33.6dB

Antenna Parameters with Body TSL

Impedance, transformed to feed point	46.3C- 1.63jΩ
Return Loss	- 27.9dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.047 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
-----------------	-------

Certificate No: Z18-97021 Page 4 of 4


D2000V2 Sn:1009




DASY5 Validation Report for Head TSL Date: 02.01.2018
Test Laboratory: CTTL, Beijing, China
DUT: Dipole 2000 MHz; Type: D2000V2; Serial: D2000V2 - SN: 1009
Communication System: UID 0, CW; Frequency: 2000 MHz; Duty Cycle: 1:1
Medium parameters used: $f = 2000$ MHz; $\sigma = 1.416$ S/m; $\epsilon_r = 38.89$; $\rho = 1000$ kg/m³
Phantom section: Left Section
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)
DASY5 Configuration:

- Probe: EX3DV4 - SN7464; ConvF(8.39, 8.39, 8.39); Calibrated: 9/12/2017,
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DA64 Sn1525; Calibrated: 10/2/2017
- Phantom: Triple Flat Phantom 5.1C; Type: QD 000 P51 CA; Serial: 1161/1
- Measurement SW: DASY52, Version 52.10 (0); SEMCAD X Version 14.6.10 (7417)

System Performance Check/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid:
 $d_x = 5$ mm, $d_y = 5$ mm, $d_z = 5$ mm
Reference Value = 95.98 V/m; Power Drift = -0.03 dB
Peak SAR (extrapolated) = 19.7 W/kg
SAR(1 g) = 10.2 W/kg; SAR(10 g) = 5.17 W/kg
Maximum value of SAR (measured) = 16.2 W/kg




Impedance Measurement Plot for Head TSL




DASY5 Validation Report for Body TSL Date: 02.01.2018
Test Laboratory: CTTL, Beijing, China
DUT: Dipole 2000 MHz; Type: D2000V2; Serial: D2000V2 - SN: 1009
Communication System: UID 0, CW; Frequency: 2000 MHz; Duty Cycle: 1:1
Medium parameters used: $f = 2000$ MHz; $\sigma = 1.564$ S/m; $\epsilon_r = 51.83$; $\rho = 1000$ kg/m³
Phantom section: Center Section
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)
DASY5 Configuration:

- Probe: EX3DV4 - SN7464; ConvF(8.24,8.24,8.24); Calibrated: 9/12/2017,
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DA64 Sn1525; Calibrated: 10/2/2017
- Phantom: Triple Flat Phantom 5.1C; Type: QD 000 P51 CA; Serial: 1161/1
- Measurement SW: DASY52, Version 52.10 (0); SEMCAD X Version 14.6.10 (7417)

System Performance Check/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid:
 $d_x = 5$ mm, $d_y = 5$ mm, $d_z = 5$ mm
Reference Value = 93.84 V/m; Power Drift = 0.02 dB
Peak SAR (extrapolated) = 19.7 W/kg
SAR(1 g) = 10.3 W/kg; SAR(10 g) = 5.18 W/kg
Maximum value of SAR (measured) = 16.3 W/kg



Impedance Measurement Plot for Body TSL



D2450V2 Sn:738



Client: **SRTC** Certificate No: **Z17-97140**

CALIBRATION CERTIFICATE

Object: D2450V2 - SN: 738

Calibration Procedure(s): FF-Z11-003-01
Calibration Procedures for dipole validation kits

Calibration date: September 18, 2017

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements(SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature(22±3)°C and humidity<70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration
Power Meter NRVD	102196	02-Mar-17 (CTTL No.J17X01254)	Mar-18
Power sensor NR-V25	100596	02-Mar-17 (CTTL No.J17X01254)	Mar-18
Reference Probe EX3DV4	SN 7433	26-Sep-16(SPEAG.No.EX3-7433_Sep16)	Sep-17
D4E4	SN 1331	19-Jan-17(CTTL-SPEAG.No.Z17-97015)	Jan-18

Secondary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY46071430	13-Jan-17 (CTTL No.J17X00285)	Jan-18
Network Analyzer E5071C	MY46110673	13-Jan-17 (CTTL No.J17X00285)	Jan-18

Calibrated by: Zhao Jing SAR Test Engineer

Reviewed by: Yu Zongying SAR Test Engineer

Approved by: Qi Dianyuan SAR Project Leader

Issued: September 21, 2017

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.



Glossary:
TSL tissue simulating liquid
ConvF sensitivity in TSL / NORMx,y,z
N/A not applicable or not measured

Calibration is Performed According to the Following Standards:
a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
b) IEC 62209-1, "Measurement procedure for assessment of specific absorption rate of human exposure to radio frequency fields from hand-held and body-mounted wireless communication devices- Part 1: Device used next to the ear (F frequency range of 300MHz to 6GHz)", July 2016
c) IEC 62209-2, "Procedure to measure the Specific Absorption Rate (SAR) For wireless communication devices used in close proximity to the human body (frequency range of 30MHz to 6GHz)", March 2010
d) KDB855664, SAR Measurement Requirements for 100 MHz to 6 GHz

Additional Documentation:
e) DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.



Measurement Conditions
DASY system configuration, as far as not given on page 1.

DASY Version	DASY52	52.10.0.1446
Extrapolation	Advanced Extrapolation	
Phantom	Triple Flat Phantom 5.1C	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	2450 MHz ± 1 MHz	

Head TSL parameters
The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	39.2	1.80 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	38.7 ± 6 %	1.79 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	13.1 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	52.4 mW / g ± 18.8 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	250 mW input power	6.10 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	24.4 mW / g ± 18.7 % (k=2)

Body TSL parameters
The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	52.7	1.95 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	52.5 ± 6 %	1.98 mho/m ± 6 %
Body TSL temperature change during test	<1.0 °C	---	---

SAR result with Body TSL

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	250 mW input power	13.2 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	52.3 mW / g ± 18.8 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Body TSL	Condition	
SAR measured	250 mW input power	6.10 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	24.3 mW / g ± 18.7 % (k=2)



Appendix (Additional assessments outside the scope of CNAS L0570)

Antenna Parameters with Head TSL

Impedance, transformed to feed point	51.30+ 5.92jΩ
Return Loss	- 24.5dB

Antenna Parameters with Body TSL

Impedance, transformed to feed point	47.60+ 6.38jΩ
Return Loss	- 23.1dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.268 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connectors near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
-----------------	-------


D2450V2 Sn:738




DASY5 Validation Report for Head TSL
Test Laboratory: CTTL, Beijing, China
DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN: 738
Communication System: UID 0, CW; Frequency: 2450 MHz; Duty Cycle: 1:1
Medium parameters used: $f = 2450$ MHz; $\sigma = 1.788$ S/m; $\epsilon_r = 38.67$; $\rho = 1000$ kg/m³
Phantom section: Left Section
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

- DASY5 Configuration:
- Probe: EX3DV4 - SN7433; ConvF(7.45, 7.45, 7.45); Calibrated: 9/26/2016;
 - Sensor-Surface: 1.4mm (Mechanical Surface Detection)
 - Electronics: DAE4 Sn1331; Calibrated: 1/19/2017
 - Phantom: Triple Flat Phantom 5.1C; Type: QD 000 P51 CA; Serial: 1161/1
 - Measurement SW: DASY52, Version 52.10 (0); SEMCAD X Version 14.6.10 (7417)

Dipole Calibration/Zoom Scan (7x7x7) (7x7x7)/Cube 0; Measurement grid: dx=5mm, dy=5mm, dz=5mm
Reference Value = 102.1 V/m; Power Drift = -0.01 dB
Peak SAR (extrapolated) = 27.8 W/kg
SAR(1 g) = 13.1 W/kg; SAR(10 g) = 6.1 W/kg
Maximum value of SAR (measured) = 22.0 W/kg





Impedance Measurement Plot for Head TSL




DASY5 Validation Report for Body TSL
Test Laboratory: CTTL, Beijing, China
DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN: 738
Communication System: UID 0, CW; Frequency: 2450 MHz; Duty Cycle: 1:1
Medium parameters used: $f = 2450$ MHz; $\sigma = 1.983$ S/m; $\epsilon_r = 52.51$; $\rho = 1000$ kg/m³
Phantom section: Center Section
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

- DASY5 Configuration:
- Probe: EX3DV4 - SN7433; ConvF(7.46, 7.46, 7.46); Calibrated: 9/26/2016;
 - Sensor-Surface: 1.4mm (Mechanical Surface Detection)
 - Electronics: DAE4 Sn1331; Calibrated: 1/19/2017
 - Phantom: Triple Flat Phantom 5.1C; Type: QD 000 P51 CA; Serial: 1161/1
 - Measurement SW: DASY52, Version 52.10 (0); SEMCAD X Version 14.6.10 (7417)

Dipole Calibration/Zoom Scan (7x7x7) (7x7x7)/Cube 0; Measurement grid: dx=5mm, dy=5mm, dz=5mm
Reference Value = 96.41 V/m; Power Drift = -0.03 dB
Peak SAR (extrapolated) = 27.8 W/kg
SAR(1 g) = 13.2 W/kg; SAR(10 g) = 6.1 W/kg
Maximum value of SAR (measured) = 22.3 W/kg



Impedance Measurement Plot for Body TSL



D5GHzV2 Sn:1079 (1/4)

In Collaboration with
TTL S P E A G
CALIBRATION LABORATORY

In Collaboration with
ILAC-MRA
CALIBRATION
CNAS L6570

Address: No. 31 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
E-mail: cni@china.ttl.com.cn http://www.china.ttl.com.cn

Client: **SRTC** Certificate No: **Z17-97133**

CALIBRATION CERTIFICATE

Object: D5GHzV2 - SN: 1079

Calibration Procedure(s): FF-Z11-003-01
Calibration Procedures for dipole validation kits

Calibration date: September 25, 2017

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements(SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature(23±3)°C and humidity<70%.

Calibration Equipment used (MATE critical for calibration)

Primary Standards	ID #	Cal Date/(Calibrated by, Certificate No.)	Scheduled Calibration
Power Meter NRP2	102196	02-Mar-17 (CTTL, No.J17X01254)	Mar-18
Power sensor NRP-291	100599	02-Mar-17 (CTTL, No.J17X01254)	Mar-18
ReferenceProbe EXD3V4	SN 3846	13-Jan-17(CTTL-SPEAG.No.Z16-97251)	Jan-18
DAE4	SN 1331	19-Jan-17(CTTL-SPEAG.No.Z17-97015)	Jan-18

Secondary Standards	ID #	Cal Date/(Calibrated by, Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	13-Jan-17 (CTTL, No.J17X00285)	Jan-18
NetworkAnalyzer E5071C	MY46110673	13-Jan-17 (CTTL, No.J17X00285)	Jan-18

Calibrated by:	Name	Function	Signature
	Zhao Jing	SAR Test Engineer	
Reviewed by:	Yu Zongying	SAR Test Engineer	
Approved by:	Qi Dianyuan	SAR Project Leader	

Issued: September 28, 2017
This calibration certificate shall not be reproduced except in full without written approval of the laboratory

Certificate No: Z17-97133 Page 1 of 16

In Collaboration with
TTL S P E A G
CALIBRATION LABORATORY

In Collaboration with
ILAC-MRA
CALIBRATION
CNAS L6570

Address: No. 31 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
E-mail: cni@china.ttl.com.cn http://www.china.ttl.com.cn

Glossary:

TSL: tissue simulating liquid
CorrF: sensitivity in TSL / NCRMax, y, z
N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Measurement procedure for assessment of specific absorption rate of human exposure to radio frequency fields from hand-held and body-mounted wireless communication devices- Part 1: Device used next to the ear (Frequency range of 300MHz to 6GHz)", July 2016
- IEC 62209-2, "Procedure to measure the Specific Absorption Rate (SAR) For wireless communication devices used in close proximity to the human body (frequency range of 30MHz to 6GHz)", March 2010
- KDB65664, SAR Measurement Requirements for 100 MHz to 6 GHz.

Additional Documentation:

- DASY4/S System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: Z17-97133 Page 2 of 16

In Collaboration with
TTL S P E A G
CALIBRATION LABORATORY

In Collaboration with
ILAC-MRA
CALIBRATION
CNAS L6570

Address: No. 31 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
E-mail: cni@china.ttl.com.cn http://www.china.ttl.com.cn

Measurement Conditions
DASY system configuration, as far as not given on page 1.

DASY Version	DASY32	52, 10 0 1446
Extrapolation	Advanced Extrapolation	
Phantom	Triple Flat Phantom 5 1C	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy = 4 mm, dz = 1.4 mm	Graded Ratio = 1.4 (Z direction)
Frequency	5200 MHz ± 1 MHz 5300 MHz ± 1 MHz 5500 MHz ± 1 MHz 5600 MHz ± 1 MHz 5800 MHz ± 1 MHz	

Head TSL parameters at 5200 MHz
The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.0	4.66 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.7 ± 6 %	4.62 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL at 5200 MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	7.77 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	77.6 mW / g ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	100 mW input power	2.24 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	22.3 mW / g ± 24.2 % (k=2)

Certificate No: Z17-97133 Page 3 of 16

In Collaboration with
TTL S P E A G
CALIBRATION LABORATORY

In Collaboration with
ILAC-MRA
CALIBRATION
CNAS L6570

Address: No. 31 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
E-mail: cni@china.ttl.com.cn http://www.china.ttl.com.cn

Head TSL parameters at 5300 MHz
The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.9	4.76 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	36.1 ± 6 %	4.67 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL at 5300 MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	8.13 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	81.3 mW / g ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	100 mW input power	2.32 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	23.2 mW / g ± 24.2 % (k=2)

Head TSL parameters at 5500 MHz
The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.6	4.96 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	36.9 ± 6 %	4.93 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL at 5500 MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	6.24 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	62.4 mW / g ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	100 mW input power	2.37 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	23.8 mW / g ± 24.2 % (k=2)

Certificate No: Z17-97133 Page 4 of 16

D5GHzV2 Sn:1079 (2/4)



Head TSL parameters at 5600 MHz

The following parameters and calculations were applied:

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.5	5.07 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.7 ± 6 %	4.98 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL at 5600 MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	8.16 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	81.6 mW / g ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	100 mW input power	2.34 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	23.4 mW / g ± 24.2 % (k=2)

Head TSL parameters at 5800 MHz

The following parameters and calculations were applied:

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.3	5.27 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.8 ± 6 %	5.18 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL at 5800 MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	7.85 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	78.7 mW / g ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	100 mW input power	2.25 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	22.6 mW / g ± 24.2 % (k=2)

Certificate No: Z17-97133

Page 5 of 16



Body TSL parameters at 5200 MHz

The following parameters and calculations were applied:

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	49.0	5.30 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	49.5 ± 6 %	5.38 mho/m ± 6 %
Body TSL temperature change during test	<1.0 °C	---	---

SAR result with Body TSL at 5200 MHz

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	100 mW input power	7.52 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	75.4 mW / g ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Body TSL	Condition	
SAR measured	100 mW input power	2.12 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	21.3 mW / g ± 24.2 % (k=2)

Body TSL parameters at 5300 MHz

The following parameters and calculations were applied:

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	49.9	5.42 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	49.2 ± 6 %	5.50 mho/m ± 6 %
Body TSL temperature change during test	<1.0 °C	---	---

SAR result with Body TSL at 5300 MHz

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	100 mW input power	7.68 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	76.9 mW / g ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Body TSL	Condition	
SAR measured	100 mW input power	2.18 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	21.9 mW / g ± 24.2 % (k=2)

Certificate No: Z17-97133

Page 6 of 16



Body TSL parameters at 5500 MHz

The following parameters and calculations were applied:

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	48.6	5.65 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	49.0 ± 6 %	5.72 mho/m ± 6 %
Body TSL temperature change during test	<1.0 °C	---	---

SAR result with Body TSL at 5500 MHz

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	100 mW input power	8.22 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	82.4 mW / g ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Body TSL	Condition	
SAR measured	100 mW input power	2.35 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	23.6 mW / g ± 24.2 % (k=2)

Body TSL parameters at 5600 MHz

The following parameters and calculations were applied:

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	48.5	5.77 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	48.4 ± 6 %	5.73 mho/m ± 6 %
Body TSL temperature change during test	<1.0 °C	---	---

SAR result with Body TSL at 5600 MHz

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	100 mW input power	8.08 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	80.7 mW / g ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Body TSL	Condition	
SAR measured	100 mW input power	2.30 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	23.0 mW / g ± 24.2 % (k=2)

Certificate No: Z17-97133

Page 7 of 16



Body TSL parameters at 5800 MHz

The following parameters and calculations were applied:

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	48.2	6.00 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	49.0 ± 6 %	5.94 mho/m ± 6 %
Body TSL temperature change during test	<1.0 °C	---	---

SAR result with Body TSL at 5800 MHz

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	100 mW input power	7.73 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	77.6 mW / g ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Body TSL	Condition	
SAR measured	100 mW input power	2.17 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	21.8 mW / g ± 24.2 % (k=2)

Certificate No: Z17-97133

Page 8 of 16

D5GHzV2 Sn:1079 (3/4)



In Collaboration with
TTL Calibration Laboratory
Add: No.31 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42310633-2079 Fax: +86-10-42310633-2504
E-mail: cti@china.ttl.com http://www.china.ttl.com

Appendix (Additional assessments outside the scope of CNAS L0570)

Antenna Parameters with Head TSL at 5200 MHz

Impedance, transformed to feed point	47.6Ω - 8.77jΩ
Return Loss	-20.7dB

Antenna Parameters with Head TSL at 5300 MHz

Impedance, transformed to feed point	45.6Ω - 6.82jΩ
Return Loss	-21.4dB

Antenna Parameters with Head TSL at 5500 MHz

Impedance, transformed to feed point	50.7Ω - 7.14jΩ
Return Loss	-23.0dB

Antenna Parameters with Head TSL at 5600 MHz

Impedance, transformed to feed point	55.2Ω - 4.09jΩ
Return Loss	-24.1dB

Antenna Parameters with Head TSL at 5800 MHz

Impedance, transformed to feed point	52.0Ω - 8.20jΩ
Return Loss	-21.6dB

Antenna Parameters with Body TSL at 5200 MHz

Impedance, transformed to feed point	50.8Ω - 10.1jΩ
Return Loss	-20.0dB

Antenna Parameters with Body TSL at 5300 MHz

Impedance, transformed to feed point	48.5Ω - 8.56jΩ
Return Loss	-21.1dB

Certificate No: Z17-97133 Page 9 of 16



In Collaboration with
TTL Calibration Laboratory
Add: No.31 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42310633-2079 Fax: +86-10-42310633-2504
E-mail: cti@china.ttl.com http://www.china.ttl.com

Antenna Parameters with Body TSL at 5500 MHz

Impedance, transformed to feed point	54.9Ω - 8.65jΩ
Return Loss	-21.9dB

Antenna Parameters with Body TSL at 5600 MHz

Impedance, transformed to feed point	56.6Ω - 2.28jΩ
Return Loss	-23.7dB

Antenna Parameters with Body TSL at 5800 MHz

Impedance, transformed to feed point	56.7Ω - 8.10jΩ
Return Loss	-20.2dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.313 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
-----------------	-------

Certificate No: Z17-97133 Page 10 of 16



In Collaboration with
TTL Calibration Laboratory
Add: No.31 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42310633-2079 Fax: +86-10-42310633-2504
E-mail: cti@china.ttl.com http://www.china.ttl.com

DASY5 Validation Report for Head TSL

Date: 09.21.2017

Test Laboratory: TTL, Beijing, China

DUT: Dipole 5GHz; Type: D5GHzV2; Serial: D5GHzV2 - SN: 1079

Communication System: CW; Frequency: 5200 MHz; Frequency: 5300 MHz;

Frequency: 5500 MHz; Frequency: 5600 MHz; Frequency: 5800 MHz;

Medium parameters used: f = 5200 MHz; α = 4.616 mho/m; εr = 35.72; ρ = 1000 kg/m³; Medium parameters used: f = 5300 MHz; α = 4.668 mho/m; εr = 36.09; ρ = 1000 kg/m³; Medium parameters used: f = 5500 MHz; α = 4.934 mho/m; εr = 35.92; ρ = 1000 kg/m³; Medium parameters used: f = 5600 MHz; α = 4.984 mho/m; εr = 35.73; ρ = 1000 kg/m³; Medium parameters used: f = 5800 MHz; α = 5.159 mho/m; εr = 35.83; ρ = 1000 kg/m³.

Phantom section: Left Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY5 Configuration:

- Probe: EX3DV4 - SN3846; ConvF(5.37.5.37.5.37); Calibrated: 1/13/2017, ConvF(4.72.4.72.4.72);

- Calibrated: 1/13/2017, ConvF(4.72.4.72.4.72); Calibrated: 1/13/2017,

- ConvF(4.95.4.95.4.95); Calibrated: 1/13/2017,

- Sensor-Surface: 1.4mm (Mechanical Surface Detection)

- Electronics: DAE4 Sn1331; Calibrated: 2017/1/19

- Phantom: Triple Flat Phantom 5.1C; Type: QD 000 P51 CA; Serial: 1161/3

- Measurement SW: DASY52, Version 52.10 (0); SEMCAD X Version 14.6.10 (7417)

Dipole Calibration (Pin=100mW, d=10mm, f=5200 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 58.81 V/m; Power Drift = -0.01 dB Peak SAR (extrapolated) = 30.8 W/kg SAR(1 g) = 7.77 W/kg; SAR(10 g) = 2.24 W/kg Maximum value of SAR (measured) = 18.2 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5300 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 65.19 V/m; Power Drift = 0.05 dB Peak SAR (extrapolated) = 33.7 W/kg SAR(1 g) = 8.13 W/kg; SAR(10 g) = 2.32 W/kg Maximum value of SAR (measured) = 19.3 W/kg

Certificate No: Z17-97133 Page 11 of 16



In Collaboration with
TTL Calibration Laboratory
Add: No.31 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42310633-2079 Fax: +86-10-42310633-2504
E-mail: cti@china.ttl.com http://www.china.ttl.com

Dipole Calibration (Pin=100mW, d=10mm, f=5500 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 57.89 V/m; Power Drift = 0.02 dB Peak SAR (extrapolated) = 34.3 W/kg SAR(1 g) = 8.24 W/kg; SAR(10 g) = 2.37 W/kg Maximum value of SAR (measured) = 19.6 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5600 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 57.89 V/m; Power Drift = 0.04 dB Peak SAR (extrapolated) = 35.7 W/kg SAR(1 g) = 8.16 W/kg; SAR(10 g) = 2.34 W/kg Maximum value of SAR (measured) = 20.0 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg


Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of

D5GHzV2 Sn:1079 (4/4)



Impedance Measurement Plot for Head TSL



Certificate No: Z17-97133

Page 13 of 16



DASY5 Validation Report for Body TSL

Date: 09.25.2017

Test Laboratory: CTTL, Beijing, China

DUT: Dipole 5GHz; Type: D5GHzV2; Serial: D5GHzV2 - SN: 1079

Communication System: CW; Frequency: 5200 MHz; Frequency: 5300 MHz; Frequency: 5500 MHz; Frequency: 5600 MHz; Frequency: 5800 MHz; Medium parameters used: $f = 5200$ MHz; $\sigma = 5.382$ mho/m; $\epsilon_r = 49.47$; $\rho = 1000$ kg/m³; Medium parameters used: $f = 5300$ MHz; $\sigma = 5.498$ mho/m; $\epsilon_r = 49.21$; $\rho = 1000$ kg/m³; Medium parameters used: $f = 5500$ MHz; $\sigma = 5.722$ mho/m; $\epsilon_r = 49.03$; $\rho = 1000$ kg/m³; Medium parameters used: $f = 5600$ MHz; $\sigma = 5.733$ mho/m; $\epsilon_r = 48.37$; $\rho = 1000$ kg/m³; Medium parameters used: $f = 5800$ MHz; $\sigma = 5.935$ mho/m; $\epsilon_r = 48.99$; $\rho = 1000$ kg/m³.

Phantom section: Center Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY5 Configuration:

- Probe: EX3DV4 - SN3846; ConvF(4.954.954.95); Calibrated: 1/13/2017, ConvF(4.954.954.95); Calibrated: 1/13/2017; ConvF(4.184.184.18); Calibrated: 1/13/2017, ConvF(4.184.184.18); Calibrated: 1/13/2017, ConvF(4.534.534.53); Calibrated: 1/13/2017,
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1331; Calibrated: 2017/11/19
- Phantom: Triple Flat Phantom 5.1C; Type: QD 000 P51 CA; Serial: 1161/3
- Measurement SW: DASY52, Version 52.10 (0); SEMCAD X Version 14.6.10 (7417)

Dipole Calibration /Pin=100mW, d=10mm, f=5200 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 55.18 V/m; Power Drift = 0.01 dB Peak SAR (extrapolated) = 30.0 W/kg SAR(1 g) = 7.52 W/kg; SAR(10 g) = 2.12 W/kg Maximum value of SAR (measured) = 18.2 W/kg

Dipole Calibration /Pin=100mW, d=10mm, f=5300 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.94 V/m; Power Drift = 0.01 dB Peak SAR (extrapolated) = 31.9 W/kg SAR(1 g) = 7.88 W/kg; SAR(10 g) = 2.19 W/kg Maximum value of SAR (measured) = 18.3 W/kg

Certificate No: Z17-97133


Page 14 of 16



Dipole Calibration /Pin=100mW, d=10mm, f=5500 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 66.70 V/m; Power Drift = -0.03 dB Peak SAR (extrapolated) = 33.7 W/kg SAR(1 g) = 8.22 W/kg; SAR(10 g) = 2.35 W/kg Maximum value of SAR (measured) = 18.6 W/kg

Dipole Calibration /Pin=100mW, d=10mm, f=5600 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 67.75 V/m; Power Drift = 0.01 dB Peak SAR (extrapolated) = 34.2 W/kg SAR(1 g) = 8.08 W/kg; SAR(10 g) = 2.3 W/kg Maximum value of SAR (measured) = 18.3 W/kg

Dipole Calibration /Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 68.20 V/m; Power Drift = 0.05 dB Peak SAR (extrapolated) = 33.3 W/kg SAR(1 g) = 7.73 W/kg; SAR(10 g) = 2.17 W/kg Maximum value of SAR (measured) = 18.3 W/kg




Certificate No: Z17-97133

Page 15 of 16



Impedance Measurement Plot for Body TSL



Certificate No: Z17-97133

Page 16 of 16

-----End of the test report-----