CONTACTS

Company Olive Union, Inc.

Homepage www.oliveunion.com

Email office@oliveunion.com

Tel. + 82 - 2 - 3454 - 0701

Address Seocho-daero 286 Unit 706

Seocho-gu, Seoul,

Republic of Korea

Olive cradle is not a bacic component of Olive.

© 2017 by Olive Union, Inc. All rights reserved.

SPECIFICATIONS

Product Category Air Conduction

Hearing Aids

Model OLV - 01

Bluetooth

Version 4.1
Frequency Band 2.4GHz
Operating Distance Up to 10m

Audio

Channel 16 band Max Output 110dB Full-On Gain 35dB

FRR 315Hz – 7,100Hz

CAUTION

- This device may cause signal interference so it is not recommended for use around sensitive tele-communications equipment.
- Avoid exerting physical pressure on the Olive or the charging dock such as sitting on it.
- Avoid exposing the Olive to water (heavy snow or rain or extreme humidity).
- If you are currently undergoing any hearing treatment, please consult your audiologist before using the Olive.
- Please consult a physician or audiologist before using the Olive if the user is a minor or suffers from any medical conditions including mental disorders.
- Keep it out of reach of children under 5. If any parts are swallowed, seek immediate medical attention.

WARRANTY

The Olive comes with a limited warranty of one (1) year starting from the date of purchase. Unless otherwise specified, this warranty covers the hardware components of the Product as originally supplied and does not cover, or partially covers, software, consumable items, or accessories, even if packaged or sold together with the Product.

The limited warranty only covers product defects caused by workmanship or build materials. To make a claim, contact Olive Union Customer Support at office@oliveunion.com.

Olive Union may use rebuilt, reconditioned or new parts and components when repairing any product. Alternatively, we may replace the defective product entirely with a rebuilt, reconditioned or new Olive product.



User Manual

Usage Time

Charging Time 1H
Total Usage Time 16H

Standby Up to 10 days

Power

Battery Rechargable Li-ion

Polymer Battery

STORAGE AND CARRY

	Storage	Carry
Temperature	10 – 40°C	-20 – 60°C
Relative Humidity	10 – 80%	5 – 90%

- 7. Do not put the hearing aid in the microwave oven.
- 8. Avoid using maximum volume.

WARNING

- If you find any skin irritation or discharge from your ear, discontinue use until discussing with your physician.
- The Olive contains magnetic parts. If you have a pacemaker or any other internal medical devices, please consult your physician before using the Olive
- Use caution when near NMR, MRI, CT scanner, or other medical machines. Other low-energy electromagnetic machines such as anti-theft EAS will not affect the Olive but may cause temporary interference.

The limited warranty does not cover:

- Defects or damage resulting from accidents, neglect, misuse or abnormal use; abnormal conditions or improper storage; unusual physical, electrical or electromechanical stress.
- Scratches, dents and cosmetic damage, unless caused by Olive Union.
 Devices that have the social number removed.
- 3. Devices that have the serial number removed, defaced, damaged, altered or made illegible.
- 4. Ordinary wear and tear.
- Defects or damage resulting from the use of the Product in conjunction or connection with accessories, products, or ancillary/peripheral equipment not furnished or approved by Olive Union.
- Defects or damage resulting from cellular signal reception or transmission, or viruses and other software problems introduced into the Product.

FCC Information to User

This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference the relativistic receiption, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one of the following measures:

Incre as the separation between the equipment and raceiver.

Incre as the separation between the equipment and raceiver.

Caution
Modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment.

FCC Compliance Information: This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation