

Label:PN-P4H-BQ-00007
Size: 67.1mm*42.3mm*0.3mm

SUMMUS Summus Medical Laser, LLC

Rechargeable Lithium Ion Battery
Model: 18650-7000mAh-10.8V 3S2P
Rating: 10.8V, 7000mAh, 75.6Wh
Charge limit voltage: 12.6V
Implementation of standards: IEC62133-2
Battery Designation: 6ICR19/66-2


Lot No:

WARNING!
Do not disassemble, impact, squeeze, throw or pierce the battery.
Do not short circuit the battery or put it into the water or fire.
Use only the charging method specified by the manufacturer.
Use only the same type of battery to replace.
Do not store it in a high temperature environment (above 60°C)

Chengsen Analytic Electronic Technology Co., Ltd


Label:PN-P4H-BQ-00003
Size: 29.4mm*17.4mm*0.3mm

LASER APERTURE



Label:PN-P4H-BQ-00004
Size: 18.4mm*18.3mm*0.3mm

Fibre-optic handpiece



Label:PN-P4H-BQ-00006
Size: 13.85mm*13.85mm*0.3mm

STOP