## למתי) THE STANDARDS INSTITUTION OF ISRAEL

## Test Report No. 7012320620

Applicant: Check-Cap Ltd.

**Equipment Under Test:** 

C-Scan Cap transceiver

<u>Model:</u> SUB-108010-00

From The Standards Institution Of Israel Industry Division Electronics & Telematics Laboratory EMC Branch



Certificate Number: AT-1359

# מתי) THE STANDARDS INSTITUTION OF ISRAEL

(Electrical & Electronics Laboratory)

#### Test Report No .: 7012320620

Page 2 of 18 pages

<u>Title:</u> C-Scan Cap transceiver <u>Model:</u> SUB-108010-00

FCC ID: 2AQ3L-CAP108010-00

Applicant:	Check-Cap Ltd.
Address:	Aba Hushi 29 Ave., POB 1271, Isfiya, Israel
Sample for test selected by:	The customer
The date of tests:	17 September 2020

Description of Equipment Under Test (EUT):	C-Scan Cap transceiver.
Model:	SUB-108010-00
Software version of radio unit:	4.1.2
Hardware version of radio unit:	PCBA-108011-01
Manufactured by:	Check-Cap Ltd.

#### **Reference Documents:**

*	CFR 47 FCC:	Rules and Regulations; Part 15. "Radio frequency devices";
		Subpart C: "Intentional radiators"
		Section 15.209. "Radiated emission limits, general requirements".
		"Radiated Emission Limits, Additional Provisions";

This Test Report contains 18 pages	This Test Report applies only to the specimen tested and may not
and may be used only in full.	be applied to other specimens of the same product.



Page 3 of 18 pages

Test Report No.: 7012320620

<u>Title:</u> C-Scan Cap transceiver <u>Model:</u> SUB-108010-00

FCC ID: 2AQ3L-CAP108010-00

## Table of Contents

1.	EUT Description and operation	4
1.1.	. General description:	4
2.	Test summary	5
2.1. 2.2.	. Potential emission sources: . EUT setup and operation:	6 6
.3	Measurements and derived results	6
3.1. 3.2. 3.3. 3.4.	. Location of the Test Site: . Test condition: . Radiated emission test. . Test of field strength emission from intentional radiator.	6 6 7 8
4.	Appendix 1. Test equipment used	12
5.	Appendix 2: Antenna Factor and Cable Loss	13
6.	Appendix 3: Test setups photo.	17



Test Report No.: 7012320620

Page 4 of 18 pages

<u>Title:</u> C-Scan Cap transceiver <u>Model:</u> SUB-108010-00

FCC ID: 2AQ3L-CAP108010-00

## 1. EUT Description and operation

#### 1.1. General description:

\* Note: the applicant supplied all information in clause below.

The EUT, C-Scan Cap, is a standalone application used for acquiring data from human colon to analyze and identify polyps. EUT comprises the 433.5 MHz transceiver module. Upon startup Cap transmit RF packets utilizing low output power of the integrated RF transceiver. Internal Cap algorithm detects entry to human body and switches the RF operation to higher output power. C-Scan Cap powered by 3.8VDC Li-Ion battery.

Transmitter frequency:	433.5 MHz.
Type of modulation:	GFSK
Antenna type:	Internal integrated.

The EUT external view presented in photo # 1.



Photo 1. Capsule external view.



Test Report No .: 7012320620

Page 5 of 18 pages

<u>Title:</u> C-Scan Cap transceiver <u>Model:</u> SUB-108010-00

FCC ID: 2AQ3L-CAP108010-00

### 2. Test summary

Parameter	FCC Part 15 Reference paragraph	Verdict
Test of field strength emission from intentional radiator	"Radiated Emission Limits, General requirements"; Section 15.209.	Comply

Electronics & Telematics Laboratory

September 2020

Name: Michael Feldman Position: Test Technician

Name: Eng. Yuri Rozenberg Position: Head of EMC Branch

Measurement uncertainty.

The test equipment has been calibrated according to its recommended procedures and is within the manufacturer's published limit of error.

The laboratory calibrates its standards by a third party (traceable to NIST, USA) on a regular basis according to equipment manufacturer requirements.

In the following table the uncertainty calculation is given.

Type of disturbance Test description	Calculated uncertainty U LAB		
Radiated disturbance electric field strength in a SAR at 3 m distance 30 MHz – 1.0 GHz	±4.32 dB		
electric field strength in a FAR at 3 m distance 1.0 – 18 GHz. 18 – 40 GHz.	±4.47 dB ±2.78 dB		



#### Test Report No.: 7012320620

<u>Title:</u> C-Scan Cap transceiver <u>Model:</u> SUB-108010-00 Page 6 of 18 pages

FCC ID: 2AQ3L-CAP108010-00

#### Normative References.

FCC 47 CFR Part 15, Subpart C	Radio Frequency Devices Subpart C – Intentional Radiators
ANSI C63.4: 2009	American National Standard for Method of Measurements of Radio-Noise Emissions from Low- Voltage Electrical and Electronic Equipment in the Range of 9 kHz to 40 GHz
ANSI C63.10: 2013	American National Standard for Testing of Unlicensed Wireless Devices.

#### 2.1. Potential emission sources:

The potential emission sources are detailed in Table 1.

#### Table 1. Potential emission sources

Frequency Location			
8.192 kHz	Positioning system magnetic field frequency		
32.768 kHz	Microcontroller oscillator		
50.0 MHz	RF transceiver Lo oscillator		
433.5 MHz	RF carrier frequency signal		

#### 2.2. EUT setup and operation:

Test was performed in continuous transmission mode.

#### 3. Measurements and derived results

#### 3.1. Location of the Test Site:

Radiated test measurements were conducted in the Anechoic chamber at the EMC laboratory of the Standards Institution of Israel in Tel-Aviv.

#### 3.2. Test condition:

Temperature: 23 °C. Humidity: 64 %. Atmospheric pressure: 1011 mbar.



Test Report No.: 7012320620

Page 7 of 18 pages

<u>Title:</u> C-Scan Cap transceiver Model: SUB-108010-00

FCC ID: 2AQ3L-CAP108010-00

#### 3.3. Radiated emission test.

#### 3.3.1. <u>General:</u>

Per FCC Part 15 Subpart C Sections 15.209

- Initial scans were made using a peak detector but still using the appropriate ANSI IF bandwidth.
- ★ A tolerance limit was set 10 dB below the specification limit. Levels above the tolerance limit were retested using the Peak, QP or Average detectors.

#### 3.3.2. <u>Radiated emission measurements:</u>

Preliminary investigation was performed from the lowest radio frequency signal generated in the equipment up to ten harmonic of a carrier frequency.

The final radiated emission measurements were performed in the semi Anechoic chamber. Test was started with new fresh battery. The EUT was operated in continue transmition mode. The test was performed in two operation options. Without human body simulator and transmitter programmed output power for outside body condition and then inside human body simulator with increased output power.

For measurement purposes and to determine compliance with emission limits, the radiating characteristics of C-Scan Cap was placed in a test fixture to imitate the electrical characteristics to those of an ingested transmitter placed in a human body.

The C-Scan Cap positioned in its worst-case orientation in the middle of the torso simulator in order to reflect the capsule operation in its use case condition.

Capsule with transmitter located in centre inside human torso simulator for testing medical implant transmitters. It consist of cylindrical Plexiglas container measuring 30 cm in diameter and 76 cm long with sidewall thickness of 0.635 cm.

The container filled with a tissue-equivalent fluidic material matches the dielectric parameters of human body at frequency of operation. The permittivity and conductivity of this material match the dielectric parameters of the body tissue-equivalent properties in KDB 865664 at the frequency of operation.

The dielectric parameters were measured - Epsilon 57.5 Sigma 0.92 - and satisfy the required target values. All emissions measurements made using the above specification at a nominal tissue-equivalent material temperature of 23° C.

A low-loss Plexiglas vertical pin permits the capsule to be positioned vertically or horizontally for measurements. The above fixture was placed on a turntable at a nominal height above ground and at 3 m distance from the measurement antenna.



Test Report No.: 7012320620

Page 8 of 18 pages

<u>Title:</u> C-Scan Cap transceiver <u>Model:</u> SUB-108010-00

FCC ID: 2AQ3L-CAP108010-00

Active Loop, Biconilog and Double Ridged Guide Horn antennas were used. The levels were maximized by rotating turntable through 360° and changing antenna-to-EUT polarization from vertical to horizontal. The worst-case results were noted in tables.

#### 3.3.3. <u>Radiated emission test results:</u>

Final measurements result are presented in tables and plots ## 1 - 6 in section 3.4.

#### 3.4. Test of field strength emission from intentional radiator.

#### 3.4.1. <u>General:</u>

Per FCC Part 15 Subpart C clause 15.209(a).

#### 3.4.2. <u>Requirements:</u>

The emission limit shown in the table based on measurements employing CISPR quasipeak detector except for the frequency band above 1000 MHz based on the employing an average detector.

Frequency MHz	Field Strength limit. μV/m	Calculated Field Strength limit dBµV/m		
30 – 88	100	40.0		
88 – 216	150	43.5		
216 – 960	200	46.0		
*Above 960	500	54.0		

\* Peak field strength shall not exceed the maximum permitted specified limit by more than 20 dB.

Field strength limits are specified at a distance of 3 meters.



Test Report No.: 7012320620

Page 9 of 18 pages

Title: C-Scan Cap transceiver

Model: SUB-108010-00

FCC ID: 2AQ3L-CAP108010-00

#### 3.4.3. <u>Test results:</u>

Freq. MHz	Antenna Polariz. V/H	Antenna Height (m)	Turn table Angle (°)	QP./Avg. Emission Level dBμV/m	Limit @ 3 m dBµV/m	Margin dB	Reference to plot #
*433.5	V	1.5	292	45.4	46.0	0.6	2
*433.5	Н	1.2	166	42.7	46.0	3.3	3
*1301.0	V	1.2	283	34.2	54.0	19.8	4
**433.5	V	1.2	187	44.7	46.0	1.3	5
**433.5	Н	1.4	17	40.7	46.0	5.3	6
**1301.5	V	1.2	169	34.0	54.0	20.0	7

\*Results of test outside human body simulator.

\*\* Results of test inside human body simulator.

Emission level = E Reading (dB $\mu$ V) + Cable loss (dB) + Antenna Factor (dB/m).	
For Cable Loss and Antenna Factor refer to Appendix 2.	
Margin (dB) = Limit (dBμV/m) – Emission level (dBμV/m)	

The provided test results are for vertically and horizontally positioned transmitter. The results for both outside and inside body simulator test options and for all polarization presented in a table above and in plots ##1 - 7.

All received spurious emissions were found below the specified limit.



Plot # 1. Spurious emissions scan 0.15 MHz – 30 MHz.

Management: Tel: 972-3-6467800 Fax: 972-3-6467779 Electronics: Tel 972-3-6465050 Fax: 972-3-7454026 Alarms Systems Section: Tel: 972-3-6467626 Fax: 972-3-6465370 Electrical: Tel: 972-3-6465160, Fax: 972-3-6465287 42 Chaim Levanon St, Tel Aviv 6997701 Israel. www.sii.org.il



Page 10 of 18 pages

Scan >1/1 Free Run

Mkr1 433.45662 MH 46.859 dBµV/r

¢1

#### Title: C-Scan Cap transceiver Model: SUB-108010-00

FCC ID: 2AQ3L-CAP108010-00



RE result with Capsule outside human body simulator.

well Time 10.0 ms CISPR

Ref 69.99 dBµV/r



G

Atten: 6 dB







RE result with Capsule inside human body simulator.

### Test Report No .: 7012320620

Page 11 of 18 pages

Title: C-Scan Cap transceiver Model: SUB-108010-00

FCC ID: 2AQ3L-CAP108010-00



Plot # 5. Vertical polarization





Plot # 6. Horizontal polarization



Test Report No .: 7012320620

Page 12 of 18 pages

<u>Title:</u> C-Scan Cap transceiver <u>Model:</u> SUB-108010-00

FCC ID: 2AQ3L-CAP108010-00

## 4. Appendix 1. Test equipment used

All measurements equipment is on SII calibration schedule with a recalibration interval not exceeding one year.

Na	Description	Manuf	Due Calibration		
NO	Description	Name	Model	Serial No	date
1	MXE EMI Receiver 20 Hz -26.5 GHz	Agilent	N9038A	MY52130048	June 2021
2	Double Ridged Guide Antenna 0.75 – 18 GHz	ETS-Lindgren	3115	00143138	March 2021
3	Broadband Horn antenna 15 – 40 GHz	Schwarzbeck Mess-Electronik	BBHA 9170	9170-341	March 2021
4	Double Ridged Waveguide Horn Antenna 1 – 18 GHz	ETS-Lindgren	3117	00139055	March 2021
5	Antenna Biconilog 30 – 6000 MHz	ETS-Lindgren	31142D	0146490	March 2021
6	Spectrum analyzer 20 Hz-40 GHz	Rohde&Schwarz	ESU 40	100168	November 2020
7	EMI Analyser 9 kHz - 26.5 GHz	HP	E7405A	SII 4944	May 2021
8	Attenuator 3 dB DC – 12.4 GHz	HP	8491A	50469	October 2020
9	LISN 9 kHz – 30 MHz	FCC	LISN 250-32-4- 16	SII5023	October 2020
10	Transient limiter 0.009-200 MHz	HP	11947A	3107105	August 2021
11	Cable RF 1m	Huber-Suhner	Sucoflex 104PE	21325/4PE	October 2020
12	Cable RF 5m	Harbour Industries	Neoflex LLEF142	1802	July 2021
13	Cable RF 0.5m	Huber-Suhner	Multiflex 141	520201	October 2020
14	Active Loop antenna 1.0 kHz – 30 MHz	ETS-Lindgren	6507	00144641	December 2020

#### Test equipment used



Page 13 of 18 pages

Test Report No.: 7012320620

<u>Title:</u> C-Scan Cap transceiver <u>Model:</u> SUB-108010-00

FCC ID: 2AQ3L-CAP108010-00

## 5. Appendix 2: Antenna Factor and Cable Loss

Cable Loss. Mast 6 m set cable.

Point	Frequency, MHz	Cable Loss, dB	Point	Frequency, MHz	Cable Loss, dB
1	30	0.3	21	1000	2.5
2	50	0.4	22	1100	2.6
3	100	0.6	23	1200	2.8
4	150	0.8	24	1300	2.9
5	200	1.0	25	1400	3.1
6	250	1.1	26	1500	3.2
7	300	1.2	27	1600	3.3
8	350	1.3	28	1700	3.5
9	400	1.5	29	1800	3.6
10	450	1.6	30	1900	3.7
11	500	1.7	31	2000	3.9
12	550	1.8	32	2100	4.0
13	600	1.9	33	2200	4.1
14	650	1.9	34	2300	4.2
15	700	2.0	35	2400	4.4
16	750	2.1	36	2500	4.6
17	800	2.1	37	2600	4.7
18	850	2.2	38	2700	4.8
19	900	2.3	39	2800	4.9
20	950	2.4	40	2900	5.0



#### Test Report No.: 7012320620

Page 14 of 18 pages

<u>Title:</u> C-Scan Cap transceiver <u>Model:</u> SUB-108010-00

FCC ID: 2AQ3L-CAP108010-00

Biconilog Antenna, ETS-Lindgren mod. 31142D, S/N: 0146490 3m calibration.						
No.	f / MHz	AF / dB/m	f / MHz	AF / dB/m	f / MHz	AF / dB/m
1	30	18.7	250	12.0	2750	31.0
2	35	15.7	300	13.8	3000	31.2
3	40	12.9	400	16.2	3250	32.7
4	45	10.6	500	18.6	3500	34.5
5	50	9.0	600	20.2	3750	34.3
6	60	7.3	700	21.8	4000	34.5
7	70	7.7	800	22.9	4250	35.3
8	80	8.2	900	24.1	4500	35.5
9	90	9.2	1000	24.8	4750	36.1
10	100	9.4	1250	26.9	5000	37.4
11	120	8.5	1500	30.2	5250	38.4
12	140	8.5	1750	28.5	5000	39.9
13	160	9.1	2000	28.9	5750	38.2
14	180	10.5	2250	29.8	6000	39.1
15	200	10.9	2500	32.5		

## Antenna factor



#### Test Report No.: 7012320620

Page 15 of 18 pages

<u>Title:</u> C-Scan Cap transceiver <u>Model:</u> SUB-108010-00

FCC ID: 2AQ3L-CAP108010-00

#### **Antenna Factor**

Double Ridged Guide Antenna mfr ETS-Lindgren model 3115 1m calibration

Point	Frequency (MHz)	Antenna Factor (dB/m)
1	1000	23.7
2	2000	28.5
3	3000	29.6
4	4000	32.5
5	4500	32.6
6	5000	33.5
7	6000	36.1
8	6500	36.5
9	7000	37.3
10	7500	38.0
11	8000	37.3
12	8500	37.9
13	9000	38.1
14	9500	38.5
15	10000	38.7
16	10500	38.8
17	11000	38.6
18	11500	38.8
19	12000	38.9
20	12500	39.3
21	13000	40.2
22	13500	40.8
23	14000	40.6
24	14500	40.4
25	15000	39.6
26	15500	39.5
27	16000	39.8
28	16500	40.4
29	17000	41.3
30	17500	42.8
31	18000	43.2

Management: Tel: 972-3-6467800 Fax: 972-3-6467779 Electronics: Tel 972-3-6465050 Fax: 972-3-7454026 Alarms Systems Section: Tel: 972-3-6467626 Fax: 972-3-6465370 Electrical: Tel: 972-3-6465160, Fax: 972-3-6465287 42 Chaim Levanon St, Tel Aviv 6997701 Israel. www.sii.org.il



## למתי) THE STANDARDS INSTITUTION OF ISRAEL

Page 16 of 18 pages

### Test Report No.: 7012320620

<u>Title:</u> C-Scan Cap transceiver <u>Model:</u> SUB-108010-00

FCC ID: 2AQ3L-CAP108010-00

#### Cable Loss Type: Neoflex LLEF142; Ser.No.1802; 5 m length

Point	Frequency (GHz)	Cable Loss (dB)
0	0.0-1.0	1.3
1	1.0 - 3.0	2.4
2	3.0 - 5.0	3.2
3	5.0-7.0	4.0
4	7.0-9.0	4.4
5	9.0-10.0	4.7
6	10.0-12.0	5.2
7	12.0-14.0	5.9
8	14.0-16.0	6.1
9	16.2-18.00	6.6

#### Active Loop antenna mfr.ETS-Lindgren mod. 6507 S/N 144641.

Frequency, MHz	Magnetic Antenna factor dBS/m	Electric Antenna factor dB/m
0.009	-20.0	31.5
0.010	-21.0	30.5
0.020	-26.7	24.9
0.075	-32.4	19.1
0.100	-32.7	18.8
0.150	-32.9	18.6
0.250	-33.0	18.5
0.500	-33.0	18.5
0.750	-33.0	18.5
1.000	-32.8	18.7
2.000	-32.7	18.8
3.000	-32.9	18.7
4.000	-33.2	18.3
5.000	-33.4	18.2
10.000	-34.0	17.6
15.000	34.2	17.3
20.000	-34.4	17.1
25.000	-34.8	16.7
30.000	-35.0	16.5



## 6. Appendix 3: Test setups photo.

RE test without human body simulator.





Photo 2.



Photo 4.

Photo 3.



Photo 5. Capsule in vertical position.



#### Test Report No.: 7012320620

Page 18 of 18 pages

<u>Title:</u> C-Scan Cap transceiver <u>Model:</u> SUB-108010-00

FCC ID: 2AQ3L-CAP108010-00

RE test inside human body simulator.







Photo 8. Capsule in vertical position.



Photo 7.



Photo 9. Capsule in horizontal position.

#### End of the document