

Date: September 19, 2023

Federal Communications Commission  
Authorization and Evaluation Division  
7435 Oakland Mills Road  
Columbia, MD 21046

Subject: Class II Permissive Changes to ENV2SOM  
FCC ID: 2AQ9D-ENV2SOM

Hereby we, Enovate Medical LLC., request Class II Permissive Change to the authorized device (FCC ID: 2AQ9D-ENV2SOM) to certify the device for 'Portable' use due to a change to the module's RF Exposure, specifically a change in minimum separation distance between the user and the antenna.

- Minimum separation distance: 20cm -> 14.95cm

All Hardware and Software of the ENV2SOM are identical. Separation distance is the only modification.

If you have any comments or questions, please contact the undersigned.

Sincerely,



Cameron Boone  
Director of Engineering  
[Cameron.boone@enovatemedical.com](mailto:Cameron.boone@enovatemedical.com)  
1152 Park Avenue., Murfreesboro, TN 37129