

校准证书

CALIBRATION CERTIFICATE



证书编号: JL2402024851

第 1 页, 共 11 页

Page 1 of 11 Pages

客户名称 Name of Customer	: 信恒检测技术(深圳)有限公司
客户地址 Address of Customer	: 广东省深圳市宝安区松岗街道潭头社区潭头工业 城二区 1 栋厂房 101、201、301
计量器具名称 Name of Instrument	: Head and Torso Simulator 头部躯干模拟器
型号/规格 Type/Specification	: Type 4128C
出厂编号 Serial No	: 3107080
资产编号 Asset No	: BTF-EM-148
制造单位 Manufacture	: Brule&Kjaer
校准依据 Calibrated in Accordance to	: JJF1520-2015 声学用头和躯干模拟器校准规范

(校准专用章)
Stamp

校准日期 : 2024 年 01 月 30 日
Operation Date Year Month Day
建议复校日期: --- 年 -- 月 -- 日
Suggested Recal.Date Year Month Day
签发日期 : 2024 年 01 月 30 日
Issue Date Year Month Day

批准人 : 张瑞纹
Approved by
签名 : 张瑞纹
Signature
核验员 : 胡岩
Checked by
校准员 : 龙晓峰
Calibrated by

证书首页背面“重要声明”是证书的组成部分,任何未包含“重要声明”内容的复制均为不完整复制。

重要声明

Important statement

1. 本院(站)是由深圳市人民政府依据《中华人民共和国计量法》设置并由国家市场监督管理总局、广东省市场监督管理局依法授权的法定计量检定机构。
SMQ is a legal metrological verification organization established by the Shenzhen Municipal People's Government and authorized by the State Administration for Market Regulation and Guangdong Administration for Market Regulation according to the Law on Metrology of the People's Republic of China.
2. 本院(站)进行的检定、校准和检测均可溯源至国际单位制(SI)单位和社会公用计量标准。
All verifications, calibrations and tests made by SMQ are traceable to the International System of Units (SI) and social public measurement standards.
3. 未经本院(站)许可,不得部分复印、摘用或篡改本证书/报告的内容;复印证书/报告未重新加盖本院证书/报告专用章无效。
Copying or excerpting portion of, or altering the content of the certificate/report is not permitted without the written authorization of SMQ. Any copy of certificates/reports without the Dedicated certificate/report Seal is deemed to be invalid.
4. 如果要满足被校准仪器的技术指标,或者技术法规要求,在规定范围内适用,请在建议复校日期前校准(适用于校准报告)。
To ensure that the calibrated object is properly used under given conditions in compliance with technical specifications or regulations, recalibrate before the suggested date (Applicable to calibration report only).
5. 本证书/报告提供的结果仅对本次被检的计量器具有效。
The results provided by the certificate/report are only valid for the measuring object under test this time.
6. 证书/报告无检定员/校准员、核验员、批准人签字,或涂改,或未盖本院证书/报告专用章及骑缝章无效。
Any certificate/report having not been signed by relevant responsible engineer, reviewer or authorized approver, or having been altered without authorization, or without both the Dedicated certificate/report Seal and its across-page seal is deemed to be invalid.
7. 证书/报告更改后,发出的电子版证书/报告、证书/报告的扫描件及传真件将不被追回,委托方有义务将更改后的证书/报告提供给使用原证书/报告的相关方。
SMQ is not responsible for recalling the electronic version of the original certificate/report when any revision is made to them. The applicant assumes the responsibility of providing the revised version to any interested party who uses them.
8. 申领电子证书时,相关内容和效力以电子证书为准;电子证书和纸质证书同时申领时,电子证书仅作为纸质证书的副本,相关内容和效力以同编号纸质证书为准。
The relevant content and effectiveness is subject to the electronic version of the original report which was only applied for. When an electronic report and a paper report are applied for at the same time, the electronic report is only a copy of the paper report, and the relevant content and effectiveness is subject to the paper report.
9. 证书中二维码浏览和下载完整报告功能是应委托方选择所设,该二维码及其复制图能使任何人扫描获取完整的证书电子版,本证书持有人如需限制他人经该二维码获取证书内容,应自行遮盖或消除证书及其复制件所附二维码,我院对委托方选择证书二维码功能所致的信息泄露概不负责(适用于附二维码证书)。
The QR code has the function of browsing and downloading complete report. Setting this function or not is chosen by the customer. The QR code and its copy enable anyone to scan and obtain the complete electronic version of the test report. Thus, if the owner of this report needs to restrict others from obtaining the content of the test report through the QR code, he shall cover or remove the QR code attached to the test report and its copies by himself. SMQ assumes no responsibility for the information leakage caused by the customer's selection of the QR code function of the test report (This clause applies to certificates with QR code attached) ..

获得的国家、省、市专业站 Establishment of Following Institutions

国家数字电子产品质量检验检测中心
National Digital Electronic Product Testing Center (NETC)
国家营养食品质量检验检测中心(广东)
National Nutrition Food Testing Center (Guangdong)
国家体育用品质量检验检测中心(广东)
National Sports Product Testing Center (Guangdong)
国家环保产品质量检验检测中心(广东)
National Environmental Product Testing Center (Guangdong)
国家分布式光伏发电系统质量检验检测中心(广东)
National Distributed Photovoltaic Power Generation System Testing Center (Guangdong)
国家电动汽车电池及充电系统产业计量测试中心
National Metrology Center for Electric Vehicle Battery and Charging System Industry
国家民用无人机产品质量检验检测中心(广东)
National Civil Unmanned Aerial Vehicle Product Testing Center (Guangdong)
国家高新技术计量站
National Hi-tech Metrology Station
国家医疗器械产业计量测试中心
National Metrology and Testing Center for Medical Device Industry
国家计量数据科学中心深圳分中心
National Metrology Data Science Center (Shenzhen)
国家城市能源计量中心(深圳)
National Urban Energy Measurement Center (Shenzhen)
中国轻工业联合会家具质量监督检测深圳站
Accredited Testing Station (Shenzhen) for Furniture by China National Light Industry Council (CNLIC)
广东省质量监督食品检验站(深圳)
Guangdong Quality Supervision and Inspection Institution for Food (Shenzhen)
广东省质量监督生态纺织服装产品检验站(深圳)
Guangdong Quality Supervision and Inspection Institution for Ecological Textile and Garment Products (Shenzhen)
广东省质量监督皮革制品检验站
Guangdong Quality Supervision and Inspection Institution for Leather Products
广东省质量监督家具检验站(深圳)
Guangdong Quality Supervision and Inspection Institution for Furniture Products (Shenzhen)
广东省质量监督学生用品检验站(深圳)
Guangdong Quality Supervision and Inspection Institution for Student's Articles (Shenzhen)
广东省质量监督自行车检验站
Guangdong Quality Supervision and Inspection Institution for Bicycle Quality
广东省质量监督钟表检验站(深圳)
Guangdong Quality Supervision and Inspection Institution for Horological Products (Shenzhen)
广东省质量监督节能环保产品(安全性能)检验站(深圳)
Guangdong Quality Supervision and Inspection Institution for Safety Performance of Environmental Protection and Energy Saving Products (Shenzhen)
广东省质量监督眼镜检验站(深圳)
Guangdong Quality Supervision and Inspection Institution for Eyewear Products (Shenzhen)
广东省质量监督电磁兼容检验站
Guangdong Quality Supervision and Inspection Institution for Electromagnetic Compatibility
广东省质量监督综合布线系统检验站
Guangdong Quality Supervision and Inspection Institution for Generic Cabling System
深圳市纤维纺织检验所
Shenzhen Quality Inspection Institution for Fiber and Textile
深圳市建材产品质量监督检验站
Shenzhen Quality Supervision and Inspection Institution for Building Materials
深圳市消防产品燃烧性能检测中心
Shenzhen Testing Center for Burning Behavior of Fire Protection Products

联系方式 Contact Information

联系电话 Contact Tel.: 0755-26941696 26941546 (龙珠) 27591789 (宝安) 82426246 (八卦岭) 28932280 (龙岗) 0769-21684525 (东莞)
传真电话 Fax No.: 0755-27591716 (宝安) 82408176 (八卦岭) 28932840 (龙岗) 0769-21684527 (东莞)
投诉及证书/报告真伪查询电话 Complaint Tel.: 400-900-8999 按5 Email: complaint@smq.com.cn

校准证书

CALIBRATION CERTIFICATE

证书编号: JL2402024851
Certificate No

第 2 页, 共 11 页
Page 2 of 11 Pages


校准用主要计量标准装置信息

Main Standard Devices Used

名称 Equipment Name	测量范围 Measuring Range	不确定度/准确度等级/ 最大允许误差 Uncertainty/Accuracy Class/ Maximum Permissible Error	计量标准考核证书号 Certificate No	有效期至 Due Date
_____	_____	_____	_____	_____

校准用主要标准器信息

Main Standards of Measurement Used

名称 Equipment Name	测量范围 Measuring Range	不确定度/准确度等级/ 最大允许误差 Uncertainty/Accuracy Class/ Maximum Permissible Error	设备编号 Equipment No	有效期至 Due Date	溯源单位/ 证书号 Traceability to/ Certificate No
标准传声器	(20~20k)Hz	声压级 $U=0.2\text{dB}, k=2$	SB3979/01	2024-11-26	
噪声、振动测试系统	DC~51.2kHz	电平 $\pm 0.01\text{dB}$	SB5196	2025-01-22	深圳检测院 /JL2401848311

附加说明

Appended Directions

委托日期: 2024 年 01 月 26 日 环境条件: 温度 21.4°C 相对湿度 70%
Application Date Operation Environment

校准地点: 本院声学龙珠实验室
Operation Location

校准项目/参数说明: 本次校准根据委托方要求实施, 校准项目/参数详见校准结果
Items Instructions

符合性及限制使用说明: 参见校准结果使用(Use referring to the results of Calibration)
Statement of Compliance and Limitation

校准结果

RESULTS OF CALIBRATION

证书编号: JL2402024851
Certificate No

第 3 页, 共 11 页
Page 3 of 11 Pages

1 外观检查: 符合要求。

Appearance Check: Pass.

2 灵敏度 (参考频率: 1000 Hz): 见表1

Sensitivity of Ear Simulator(Ref. Freq:1000 Hz): See Table 1

表 1 Table 1

通道 Channel	传声器型号 Type	传声器编号 No.	灵敏度 (mV/Pa) Sensitivity
L通道	/	/	11.6
R通道	/	/	11.5

3 耳自由场频率响应(声源位于正前方1.0 m处):

Frequency Response of Ear Simulator (The distance is 1m) : See Table 2

表 2 Table 2

频率 Frequency	左耳频响 F.R.of Left	右耳频响 F.R.of Right	允许范围 M.P.E.
(Hz)	(dB)	(dB)	(dB)
100	0.5	0.1	+1.0 ~ -1.0
125	0.2	-0.1	+1.0 ~ -1.0
160	0.3	0.1	+1.0 ~ -1.5
200	0.7	0.0	+1.0 ~ -1.5
250	-0.1	-0.3	+1.5 ~ -1.0
315	-0.2	-0.5	+2.0 ~ -0.5
400	1.0	0.8	+2.5 ~ 0.0
500	1.8	1.3	+3.5 ~ +1.0
630	3.9	3.6	+4.0 ~ +1.5
800	2.8	2.8	+6.0 ~ +2.5
1000	4.9	5.0	+5.5 ~ +2.0
1250	3.7	3.9	+6.0 ~ +0.5

校准结果

RESULTS OF CALIBRATION

证书编号: JL2402024851
Certificate No

第 4 页, 共 11 页
Page 4 of 11 Pages

1600	4.5	4.8	+7.0 ~ +2.0
2000	10.0	10.8	+13.5 ~ +9.0
2500	16.4	15.9	+19.5 ~ +14.5
3150	17.1	15.2	+20.5 ~ +13.5
4000	13.8	12.5	+18.0 ~ +12.0
5000	9.0	8.6	+15.5 ~ +8.5
6300	6.8	6.1	+9.0 ~ +2.5
8000	2.2	2.1	+11.0 ~ -1.0
10000	7.3	6.1	+10.0 ~ +0.5

4 耳指向性:

Ear directivity

4.1 耳指向性0°:

Ear directivity of 0° : See Table 3

表 3 Table 3

频率 Frequency	左耳频响 F.R.of Left	右耳频响 F.R.of Right
(Hz)	(dB)	(dB)
100	71.2	70.7
125	69.4	69.2
160	67.7	67.4
200	67.4	66.8
250	64.4	64.2
315	66.0	65.7
400	64.4	64.2
500	67.7	67.2
630	69.1	68.8
800	71.7	71.7
1000	72.6	72.7
1250	71.3	71.4
1600	72.9	73.2
2000	78.0	78.8

校准结果

RESULTS OF CALIBRATION

证书编号: JL2402024851
Certificate No

第 5 页, 共 11 页
Page 5 of 11 Pages

2500	85.3	84.8
3150	93.9	92.0
4000	85.2	83.9
5000	80.8	80.4
6300	83.9	83.2
8000	68.2	68.2
10000	73.3	72.1

4.2耳指向性90°:

Ear directivity of 90 ° : See Table 4

表 4 Table 4

频率 Frequency	左耳频响 F.R.of Left	右耳频响 F.R.of Right
(Hz)	(dB)	(dB)
100	71.3	72.3
125	68.8	71.0
160	67.7	70.1
200	66.7	68.9
250	64.4	67.1
315	64.8	68.6
400	62.1	68.1
500	66.9	72.0
630	66.8	72.5
800	71.4	76.5
1000	69.1	76.2
1250	69.5	75.9
1600	71.1	78.0
2000	73.2	79.5
2500	76.4	84.7
3150	81.4	93.0
4000	71.9	85.4
5000	70.8	88.8
6300	73.4	93.3

校准结果

RESULTS OF CALIBRATION

证书编号: JL2402024851
Certificate No

第 6 页, 共 11 页
Page 6 of 11 Pages

8000	56.2	82.1
10000	57.6	76.3

4.3耳指向性180°:

Ear directivity of 180 ° : See Table 5

表 5 Table 5

频率 Frequency (Hz)	左耳频响 F.R.of Left (dB)	右耳频响 F.R.of Right (dB)
100	71.5	71.2
125	69.6	69.1
160	67.8	67.3
200	67.1	67.0
250	64.9	64.4
315	65.7	65.3
400	63.9	63.2
500	67.4	67.0
630	68.6	68.0
800	71.8	71.4
1000	73.1	72.8
1250	72.4	72.1
1600	73.5	73.5
2000	76.8	76.8
2500	82.5	81.8
3150	90.6	88.9
4000	83.4	82.0
5000	75.6	75.3
6300	77.1	77.3
8000	67.5	68.4
10000	68.2	67.1

校准结果

RESULTS OF CALIBRATION

证书编号: JL2402024851
Certificate No

第 7 页, 共 11 页
Page 7 of 11 Pages

4.4耳指向性270°:

Ear directivity of 270 ° : See Table 6

表 6 Table 6

频率 Frequency	左耳频响 F.R.of Left	右耳频响 F.R.of Right
(Hz)	(dB)	(dB)
100	72.2	70.3
125	71.0	68.0
160	69.9	66.6
200	69.0	66.1
250	67.5	63.9
315	69.1	64.3
400	68.6	61.8
500	72.3	66.3
630	72.7	66.2
800	76.9	71.2
1000	76.2	68.8
1250	76.0	69.3
1600	77.8	71.2
2000	79.0	73.8
2500	85.0	75.5
3150	94.5	79.4
4000	88.0	72.2
5000	89.3	69.5
6300	94.0	72.1
8000	82.0	55.3
10000	77.3	54.6

校准结果

RESULTS OF CALIBRATION

证书编号: JL2402024851
Certificate No

第 8 页, 共 11 页
Page 8 of 11 Pages

5 嘴参考点处的幅频特性:
Frequency Response of MPR: See Table 7

表 7 Table 7

频率 Frequency	声压级 SPL
(Hz)	(dB)
100	59.6
125	64.8
160	70.0
200	75.1
250	77.1
315	75.9
400	74.0
500	72.7
630	71.8
800	71.8
1000	72.4
1250	73.1
1600	75.8
2000	74.8
2500	68.8
3150	66.3
4000	65.9
5000	64.1
6300	67.8
8000	63.8
10000	64.5

校准结果

RESULTS OF CALIBRATION

证书编号: JL2402024851
Certificate No

第 9 页, 共 11 页
Page 9 of 11 Pages

6 嘴参考点处的失真:
THD of MPR: See Table 8

表 8 Table 8

频率 Frequency	失真 THD
(Hz)	(%)
100	10.50
125	6.20
160	2.87
200	1.32
250	1.30
315	1.62
400	1.60
500	2.64
630	3.05
800	2.50
1000	2.11
1250	1.62
1600	1.05
2000	1.82
2500	2.13
3150	3.57
4000	1.51
5000	3.40
6300	1.20
8000	0.34
10000	0.42

校准结果

RESULTS OF CALIBRATION

证书编号: JL2402024851
Certificate No

第 10 页, 共 11 页
Page 10 of 11 Pages

7 嘴自由场幅频特性 (传声器位于嘴前0.5 m) :
Amplitude-frequency characteristics: See Table 9

表 9 Table 9

频率 Frequency	声压级 SPL
(Hz)	(dB)
100	50.6
125	56.0
160	59.9
200	65.5
250	67.7
315	67.0
400	65.8
500	67.5
630	65.5
800	61.5
1000	63.6
1250	65.3
1600	70.0
2000	69.6
2500	63.1
3150	61.6
4000	62.1
5000	60.4
6300	64.5
8000	61.5
10000	61.3

附注:

Notes:

1. 参考频率声压级测量结果扩展不确定度:

Related Expanded Uncertainty of Sound Pressure Level at Reference Frequency:

校准结果

RESULTS OF CALIBRATION

证书编号: JL2402024851
Certificate No

第 11 页, 共 11 页
Page 11 of 11 Pages

$U=1.0$ dB, $k=2$ (20 Hz ~ 200 Hz) ;
 $U=0.8$ dB, $k=2$ (250 Hz ~ 400 Hz) ;
 $U=0.6$ dB, $k=2$ (500 Hz ~ 1.25 kHz) ;
 $U=1.0$ dB, $k=2$ (1.6 kHz ~ 10 kHz) 。
 $U=1.2$ dB, $k=2$ (12.5 kHz ~ 20 kHz) 。

2. 失真测量结果扩展不确定度:

Relative Expanded Uncertainty of Distortion:

$U_{rel} = 0.5\%$, $k=2$ (0.05%~30%, 20 Hz ~ 20 kHz) 。

3. 本证书中给出的测量不确定度依据 JJF 1059.1-2012 《测量不确定度评与表示》评定

The uncertainty given in this certificate is evaluated according to JJF 1059.1-2012 *Evaluation and Expression of Uncertainty in Measurement*.
