



RF Exposure Evaluation Declaration

FCC ID: 2AQVU0005

APPLICANT: JOYTECH HEALTHCARE CO., LTD.

Application Type: Certification

Product: Arm-type Fully Automatic Digital Blood Pressure Monitor

Model No.: DBP-1307b, DBP-1305b, DBP-1319b, DBP-1318b,
DBP-1332b, DBP-1333b, DBP-1257b, DBP-1358b,
DBP-1359b

FCC Classification: Digital Transmission System (DTS)

Test Procedure(s): KDB 447498 D01v06

Test Date: July 28, 2019

Reviewed By

Kevin Guo

(Kevin Guo)

Approved By

Robin Wu

(Robin Wu)



The test results relate only to the samples tested.

The test results shown in the test report are traceable to the national/international standards through the calibration of the equipment and evaluated measurement uncertainty herein.

The test report shall not be reproduced except in full without the written approval of MRT Technology (Suzhou) Co., Ltd.

Revision History

Report No.	Version	Description	Issue Date	Note
1905WSU015-U2	Rev. 01	Initial Report	10-11-2019	Valid

§2.1033 General Information

Applicant:	JOYTECH HEALTHCARE CO., LTD.
Applicant Address:	No. 365, Wuzhou Road, Yuhang Economic Development Zone, Hangzhou City, 311100 Zhejiang, China
Manufacturer:	JOYTECH HEALTHCARE CO., LTD.
Manufacturer Address:	No. 365, Wuzhou Road, Yuhang Economic Development Zone, Hangzhou City, 311100 Zhejiang, China
Test Site:	MRT Technology (Suzhou) Co., Ltd
Test Site Address:	D8 Building, No.2 Tian'edang Rd., Wuzhong Economic Development Zone, Suzhou, China
Test Device Serial No.:	N/A <input type="checkbox"/> Production <input checked="" type="checkbox"/> Pre-Production <input type="checkbox"/> Engineering

Test Facility / Accreditations

Measurements were performed at MRT Laboratory located in Tian'edang Rd., Suzhou, China.

- MRT facility is a FCC registered (MRT Reg. No. 893164) test facility with the site description report on file and has met all the requirements specified in ANSI C63.4-2014.
- MRT facility is an IC registered (MRT Reg. No. 11384A-1) test laboratory with the site description on file at Industry Canada.
- MRT facility is a VCCI registered (R-20025, G-20034, C-20020, T-20020) test laboratory with the site description on file at VCCI Council.
- MRT Lab is accredited to ISO 17025 by the American Association for Laboratory Accreditation (A2LA) under the American Association for Laboratory Accreditation Program (A2LA Cert. No. 3628.01) in EMC, Telecommunications, Radio and SAR testing.



1. PRODUCT INFORMATION

1.1. Equipment Description

Product Name:	Arm-type Fully Automatic Digital Blood Pressure Monitor
Model No.:	DBP-1307b, DBP-1305b, DBP-1319b, DBP-1318b, DBP-1332b, DBP-1333b, DBP-1257b, DBP-1358b, DBP-1359b
Bluetooth Specification:	v5.0
Accessory	
Adapter:	Model: MPSUL0801000 INPUT: 100-240V~ 50/60Hz 0.25A OUTPUT: 5V/1000mA

Note: There are the same schematics and PCB design except the appearance, appearance color and size of LED display between the models. We select the maximum size LED display sample (Model: DBP-1307b) for all testing.

1.2. Product Specification Subjective to this Report

Bluetooth Frequency:	2402~2480MHz
Channel Number:	GFSK
Data Rate:	1Mbps, 2Mbps
Antenna Type:	PCB Antenna
Antenna Gain:	0.5dBi

Note: For other features of this EUT, test report will be issued separately.

2. RF Exposure Evaluation

2.1. Limits

SAR Test Exclusion Thresholds for 100 MHz – 6 GHz and ≤ 50 mm

Approximate SAR Test Exclusion Power Thresholds at Selected Frequencies and Test Separation Distances are illustrated in the following Table. The equation and threshold in Note 1 must be applied to determine SAR test exclusion.

MHz	5	10	15	20	25	mm
150	39	77	116	155	194	SAR Test Exclusion Threshold (mW)
300	27	55	82	110	137	
450	22	45	67	89	112	
835	16	33	49	66	82	
900	16	32	47	63	79	
1500	12	24	37	49	61	
1900	11	22	33	44	54	
2450	10	19	29	38	48	
3600	8	16	24	32	40	
5200	7	13	20	26	33	
5400	6	13	19	26	32	
5800	6	12	19	25	31	

MHz	30	35	40	45	50	mm
150	232	271	310	349	387	SAR Test Exclusion Threshold (mW)
300	164	192	219	246	274	
450	134	157	179	201	224	
835	98	115	131	148	164	
900	95	111	126	142	158	
1500	73	86	98	110	122	
1900	65	76	87	98	109	
2450	57	67	77	86	96	
3600	47	55	63	71	79	
5200	39	46	53	59	66	
5400	39	45	52	58	65	
5800	37	44	50	56	62	

Note: The 1-g and 10-g SAR test exclusion thresholds for 100 MHz to 6 GHz at test separation distances ≤ 50 mm are determined by:

$[(\text{max. power of channel, including tune-up tolerance, mW}) / (\text{min. test separation distance, mm})] * \sqrt{f(\text{GHz})} \leq 3.0$ for 1-g SAR and ≤ 7.5 for 10-g extremity SAR, where

- $f(\text{GHz})$ is the RF channel transmit frequency in GHz
- Power and distance are rounded to the nearest mW and mm before calculation
- The result is rounded to one decimal place for comparison
- 3.0 and 7.5 are referred to as the numeric thresholds in the step 2 below

The test exclusions are applicable only when the minimum test separation distance is ≤ 50 mm and for transmission frequencies between 100 MHz and 6 GHz. When the minimum test separation distance is < 5 mm, a distance of 5 mm according to 5) in section 4.1 is applied to determine SAR test exclusion.

2.2. Test Procedure

Software provided by client enabled the EUT to transmit and receive data at lowest, middle and highest channel individually.

The temperature and related humidity: 20°C and 75% RH.

2.3. Test Result of RF Exposure Evaluation

Product	Arm-type Fully Automatic Digital Blood Pressure Monitor
Test Item	RF Exposure Evaluation

Test Mode	Frequency Band (MHz)	Maximum output power to antenna (dBm)	Maximum output power to antenna (mW)	SAR Test Exclusion Threshold (mW)
Bluetooth	2402 ~ 2480	3.80	2.3988	10

Note: Maximum output power to antenna (dBm) = Max conducted Power (dBm) + Antenna Gain (dBi), Max conducted Power = 3.30dBm, Antenna Gain = 0.5dBi

Per FCC KDB 447498 D01v06, the SAR exclusion threshold for distances < 50mm is defined by the following equation:

$$\frac{\text{Max Power of Channel (mW)}}{\text{Test Separation Dist (mm)}} * \sqrt{\text{Frequency (GHz)}} \leq 3.0$$

Based on the maximum conducted power of Bluetooth and the antenna to use separation distance, Bluetooth SAR was not required;

$$[(2.3988\text{mW}/5) * \sqrt{2.402}] = 0.7436 < 3.0$$

The Max $P_d = 0.7436 < 3.0$

Note: When the minimum test separation distance is < 5 mm, a distance of 5 mm is applied to determine SAR test exclusion.

_____ The End _____

Appendix A – EUT Photograph

Refer to “ 1905WSU015-UE” file.