

RF Exposure Evaluation Report

Product : Wrist-type Fully Automatic Digital
Blood Pressure Monitor

Trade mark : N/A

Model/Type reference : DBP-8297B, DBP-8299B

Serial Number : N/A

Report Number : EED32O81898001

FCC ID : 2AQVU0040

Date of Issue : Dec. 20, 2022

Test Standards : 47 CFR Part 1.1307
47 CFR Part 1.1310
47 CFR Part 2.1093
447498 D04 Interim General RF
Exposure Guidance v01

Test result : PASS

Prepared for:

JOYTECH HEALTHCARE CO., LTD.

**No.365, Wuzhou Road, Yuhang Economic Development Zone,
Hangzhou City, 311100 Zhejiang, China**

Prepared by:

**Centre Testing International Group Co., Ltd.
Hongwei Industrial Zone, Bao'an 70 District,
Shenzhen, Guangdong, China**

TEL: +86-755-3368 3668

FAX: +86-755-3368 3385



Compiled by:

mark.chen.

Mark Chen

Approved by:

Aaron Ma

Aaron Ma

Reviewed by:

Tom Chen

Tom Chen

Date:

Dec. 20, 2022

Check No.: 6044281122

2 Version

Version No.	Date	Description
00	Dec. 20, 2022	Original

3 Contents

	Page
1 COVER PAGE	1
2 VERSION	2
3 CONTENTS	3
4 GENERAL INFORMATION	4
4.1 CLIENT INFORMATION	4
4.2 GENERAL DESCRIPTION OF EUT	4
4.3 PRODUCT SPECIFICATION SUBJECTIVE TO THIS STANDARD	4
4.4 TEST LOCATION	5
4.5 DEVIATION FROM STANDARDS	5
4.6 ABNORMALITIES FROM STANDARD CONDITIONS	5
4.7 OTHER INFORMATION REQUESTED BY THE CUSTOMER	5
5 SAR EVALUATION	6
5.1 RF EXPOSURE COMPLIANCE REQUIREMENT	6
5.1.1 <i>Limits</i>	6
5.1.2 <i>Test Procedure</i>	6
5.1.3 <i>EUT RF Exposure Evaluation</i>	7

4 General Information

4.1 Client Information

Applicant:	JOYTECH HEALTHCARE CO., LTD.
Address of Applicant:	No.365, Wuzhou Road, Yuhang Economic Development Zone, Hangzhou City, 311100 Zhejiang, China
Manufacturer:	JOYTECH HEALTHCARE CO., LTD.
Address of Manufacturer:	No.365, Wuzhou Road, Yuhang Economic Development Zone, Hangzhou City, 311100 Zhejiang, China
Factory:	JOYTECH HEALTHCARE CO., LTD.
Address of Factory:	No.365, Wuzhou Road, Yuhang Economic Development Zone, Hangzhou City, 311100 Zhejiang, China

4.2 General Description of EUT

Product Name:	Wrist-type Fully Automatic Digital Blood Pressure Monitor
Model No.(EUT):	DBP-8297B, DBP-8299B
Test Model No.:	DBP-8297B
Trade Mark:	N/A
Device type:	Portable

4.3 Product Specification subjective to this standard

Frequency Range:	2402MHz~2480MHz
Modulation Type:	GFSK
Test Power Grade:	Default
Test Software of EUT:	PhyPluskit
Antenna Type:	PCB antenna
Antenna Gain:	0dBi
Power Supply:	DC 3.0V
Sample Received Date:	Nov. 29, 2022
Sample tested Date:	Nov. 29, 2022 to Dec. 03, 2022
Remark:	<p>Company Name and Address shown on Report, the sample(s) and sample Information was/ were provided by the applicant who should be responsible for the authenticity which CTI hasn't verified.</p> <p>Model No.:DBP-8297B, DBP-8299B</p> <p>Only the model DBP-8297B was tested, since the electrical circuit design, layout, components used and internal wiring were identical for the above models.In addition to the shell size, the LCD screen size is different.</p>

4.4 Test Location

All tests were performed at:

Centre Testing International Group Co., Ltd

Building C, Hongwei Industrial Park Block 70, Bao'an District, Shenzhen, China

Telephone: +86 (0) 755 33683668 Fax:+86 (0) 755 33683385

No tests were sub-contracted.

FCC Designation No.: CN1164

4.5 Deviation from Standards

None.

4.6 Abnormalities from Standard Conditions

None.

4.7 Other Information Requested by the Customer

None.

5 SAR Evaluation

5.1 RF Exposure Compliance Requirement

5.1.1 Limits

The SAR-based exemption formula of § 1.1307(b)(3)(i)(B), repeated here as Formula (B.2), applies for single fixed, mobile, and portable RF sources with available maximum time-averaged power or effective radiated power (ERP), whichever is greater, of less than or equal to the threshold Pth (mW).

This method shall only be used at separation distances from 0.5 cm to 40 cm and at frequencies from 0.3 GHz to 6 GHz (inclusive). Pth is given by Formula

$$P_{th} \text{ (mW)} = \begin{cases} ERP_{20 \text{ cm}} (d/20 \text{ cm})^x & d \leq 20 \text{ cm} \\ ERP_{20 \text{ cm}} & 20 \text{ cm} < d \leq 40 \text{ cm} \end{cases}$$

where

$$x = -\log_{10} \left(\frac{60}{ERP_{20 \text{ cm}} \sqrt{f}} \right)$$

and f is in GHz, d is the separation distance (cm), and ERP20cm is per Formula (B.1).

$$P_{th} \text{ (mW)} = ERP_{20 \text{ cm}} \text{ (mW)} = \begin{cases} 2040f & 0.3 \text{ GHz} \leq f < 1.5 \text{ GHz} \\ 3060 & 1.5 \text{ GHz} \leq f \leq 6 \text{ GHz} \end{cases} \quad (\text{B.1})$$

The 1 mW Blanket Exemption of § 1.1307(b)(3)(i)(A) applies for single fixed, mobile, and portable RF sources with available maximum time-averaged power of no more than 1 mW, regardless of separation distance.

5.1.2 Test Procedure

Software provided by client enabled the EUT to transmit and receive data at lowest, middle and highest channel individually.

5.1.3 EUT RF Exposure Evaluation

For Stand alone:

For BLE

Frequency (MHz)	Separation distance (cm)	Max. Conducted Output power (dBm)	Antenna Gain (dBi)	EIRP (dBm)	ERP (dBm)	ERP (mW)	Limit (mW)	Result
2402	0.50	-1.69	0	-1.69	-3.84	0.413	2.788	PASS

Note:

- ① EIRP=conducted power+antenna gain;
- ② ERP=EIRP-2.15
- ③ Only the worst case data was recorded in the report.

The test report is effective only with both signature and specialized stamp, The result(s) shown in this report refer only to the sample(s) tested. Without written approval of CTI, this report can't be reproduced except in full.

*** End of Report ***